



The Dutch Accreditation Council RvA, by law appointed as the national accreditation body for The Netherlands, hereby declares that accreditation has been granted to:

Intermes B.V. Calibration Laboratory Hengelo

The organisation has demonstrated to be able to generate technical valid results in a competent way and work according to a management system.

This accreditation is based on an assessment against the requirements as laid down in ISO/IEC 17025:2005.

The accreditation covers the activities as specified in the authorized annex bearing the registration number.

The accreditation is valid provided that the organisation continues to meet the requirements.

The accreditation with registration number:

K 018

is granted on 22 December 2016

This declaration is valid until
1 December 2020

The accreditation has been granted for the first time on
15 September 1980

The Chief Executive

Ir. J.C. van der Poel

Annex to declaration of accreditation (scope of accreditation)

Normative document: EN ISO/IEC 17025:2005

Registration number: **K 018**

of **Intermes B.V.**
Calibration Laboratory

This annex is valid from: **08-11-2018** to **30-11-2020**

Replaces annex dated: **01-11-2017**

HCS code	Measured quantity, Instrument, Measure	Frequency	CMC ¹	Remarks	Location
LF 0 0	DC / LF				
LF 1 0	Direct voltage				HLO
	0 mV – 200mV		$1.0 \cdot 10^{-5} \cdot U$, minimum 0.1 μ V	Measuring (1)	
	0.2 V – 2 V		$7 \cdot 10^{-6} \cdot U$	Measuring (1)	

of **Intermes B.V.**
Calibration Laboratory

This annex is valid from: **08-11-2018** to **30-11-2020**

Replaces annex dated: **01-11-2017**

HCS code	Measured quantity, Instrument, Measure	Frequency	CMC ¹	Remarks	Location
	2 V – 20 V		$5 \cdot 10^{-6} \cdot U$	Measuring (1)	
	20 V – 200 V		$7 \cdot 10^{-6} \cdot U$	Measuring (1)	
	200 V – 1000 V		$5 \cdot 10^{-6} \cdot U$	Measuring (1)	
	0 mV – 220 mV		$2.0 \cdot 10^{-5} \cdot U$, minimum 1.5 μ V	Generate (1)	
	0.22 V – 2,2 V		$7 \cdot 10^{-6} \cdot U$	Generate (1)	
	2.2 V – 22 V		$1.0 \cdot 10^{-5} \cdot U$	Generate (1)	
	22 V – 220 V		$1.5 \cdot 10^{-5} \cdot U$	Generate (1)	
	220 V – 1100 V		$1.0 \cdot 10^{-5} \cdot U$	Generate (1)	
LF 2 0	Direct current				HLO
	1 μ A – 200 μ A		$1 \cdot 10^{-4} \cdot I$, minimum 0.5 nA	Measuring (1)	
	200 μ A – 20 mA		$3 \cdot 10^{-5} \cdot I$	Measuring (1)	
	20 mA – 200 mA		$5 \cdot 10^{-5} \cdot I$	Measuring (1)	
	0.2 A – 2 A		$2.0 \cdot 10^{-4} \cdot I$	Measuring (1)	
	2 A – 20 A		$2.0 \cdot 10^{-4} \cdot I$	Measuring (1)	
	0 μ A – 220 mA		$1.0 \cdot 10^{-4} \cdot I$, minimum 0.5 nA	Generate compliance < 0,5 V (1)	
	0.22 A – 2.2 A		$1 \cdot 10^{-4} \cdot I$	Generate compliance < 0.5 V (1)	
	2.2 A – 20 A		$2.0 \cdot 10^{-4} \cdot I$	Generate compliance < 2 V (1)	
	20 A – 1000 A		$5 \cdot 10^{-3} \cdot I$	Generate, with coils	
LF 3 0	Alternating voltage				HLO
	10 mV – 100 mV	20 Hz – 20 kHz	$1.2 \cdot 10^{-3} \cdot U$	Measuring (1)	
	10 mV – 100 mV	20 kHz – 100 kHz	$4 \cdot 10^{-3} \cdot U$	Measuring (1)	

of **Intermes B.V.**
Calibration Laboratory

This annex is valid from: **08-11-2018** to **30-11-2020**

Replaces annex dated: **01-11-2017**

HCS code	Measured quantity, Instrument, Measure	Frequency	CMC ¹	Remarks	Location
	100 mV – 200 mV	20 Hz – 20 kHz	$3 \cdot 10^{-4} \cdot U$	Measuring (1)	
	100 mV – 200 mV	20 kHz – 100 kHz	$4 \cdot 10^{-4} \cdot U$	Measuring (1)	
	0.2 V – 2 V	20 Hz – 10 kHz	$1.0 \cdot 10^{-4} \cdot U$	Measuring (1)	
	0.2 V – 2 V	10 kHz – 100 kHz	$5 \cdot 10^{-4} \cdot U$	Measuring (1)	
	2 V – 20 V	20 Hz – 10 kHz	$1.5 \cdot 10^{-4} \cdot U$	Measuring (1)	
	2 V – 20 V	10 kHz – 100 kHz	$3 \cdot 10^{-4} \cdot U$	Measuring (1)	
	20 V – 200 V	20 Hz – 10 kHz	$1.0 \cdot 10^{-4} \cdot U$	Measuring (1)	
	20 V – 200 V	10 kHz – 100 kHz	$2.0 \cdot 10^{-4} \cdot U$	Measuring (1)	
	200 V – 1000 V	55 Hz – 10 kHz	$1.5 \cdot 10^{-4} \cdot U$	Measuring (1)	
	200 V – 1000 V	10 kHz – 30 kHz	$1.5 \cdot 10^{-3} \cdot U$	Measuring (1)	
	1 kV – 100 kV	50 Hz	$1.0 \cdot 10^{-3} \cdot U$	Measuring (1)	
	2.2 mV – 22 mV	40 Hz – 20 kHz	$5 \cdot 10^{-4} \cdot U$	Generate (1)	
	22 mV – 220 V	40 Hz – 20 kHz	$1.0 \cdot 10^{-4} \cdot U$	Generate (1)	
	220 V – 1100 V	40 Hz – 1 kHz	$1.0 \cdot 10^{-4} \cdot U$	Generate (1)	
LF 4 0	Alternating current				HLO
	10 μ A – 100 μ A	55 Hz – 1 kHz	$4 \cdot 10^{-3} \cdot I$	Measuring (1)	
	100 μ A – 2 mA	55 Hz – 1 kHz	$1.5 \cdot 10^{-4} \cdot I$	Measuring (1)	
	2 mA – 20 mA	55 Hz – 1 kHz	$5 \cdot 10^{-4} \cdot I$	Measuring (1)	
	20 mA – 200 mA	55 Hz – 1 kHz	$1.0 \cdot 10^{-4} \cdot I$	Measuring (1)	
	0.2 A – 2 A	55 Hz – 300 Hz	$3 \cdot 10^{-4} \cdot I$	Measuring (1)	
	0.2 A – 2 A	300 Hz – 1 kHz	$2.0 \cdot 10^{-4} \cdot I$	Measuring (1)	
	2 A – 20 A	55 Hz – 1 kHz	$4 \cdot 10^{-4} \cdot I$	Measuring (1)	
	20 A – 600 A	50 Hz	$6 \cdot 10^{-4} \cdot I$	Measuring (1)	
	100 μ A – 220 mA	40 Hz – 1 kHz	$2.0 \cdot 10^{-4} \cdot I$	Generate (1)	
	0.22 A – 2.2 A	40 Hz – 1 kHz	$3 \cdot 10^{-4} \cdot I$	Generate (1)	
	2.2 A – 20 A	40 Hz – 440 Hz	$1.0 \cdot 10^{-3} \cdot I$	Generate (1)	
	20 A – 1000 A	45 – 60 Hz	$5 \cdot 10^{-3} \cdot I$	Generate, with coils	

of **Intermes B.V.**
Calibration Laboratory

This annex is valid from: **08-11-2018** to **30-11-2020**

Replaces annex dated: **01-11-2017**

HCS code	Measured quantity, Instrument, Measure	Frequency	CMC ¹	Remarks	Location
	20 A – 200 A	60 – 440 Hz	$7.5 \cdot 10^{-3} \cdot I$	Generate, with coils	
LF 6 1	Resistance				HLO
	100 $\mu\Omega$ - 1 m Ω		$3 \cdot 10^{-4} \cdot R$	Measuring (1)	
	1 m Ω - 100 m Ω		$1.5 \cdot 10^{-4} \cdot R$	Measuring (1)	
	100 m Ω - 1 Ω		$5 \cdot 10^{-5} \cdot R$	Measuring (1)	
	1 Ω – 2 Ω		$3.0 \cdot 10^{-5} \cdot R$	Measuring (1)	
	2 Ω – 2 k Ω		$1.0 \cdot 10^{-5} \cdot R$	Measuring (1)	
	2 k Ω – 20 k Ω		$5 \cdot 10^{-6} \cdot R$	Measuring (1)	
	20 k Ω – 2 M Ω		$1.0 \cdot 10^{-5} \cdot R$	Measuring (1)	
	2 M Ω - 20 M Ω		$3 \cdot 10^{-5} \cdot R$	Measuring (1)	
	20 M Ω - 200 M Ω		$1.0 \cdot 10^{-4} \cdot R$	Measuring (1)	
	200 M Ω – 2 G Ω		$1.5 \cdot 10^{-3} \cdot R$	Measuring (1)	
	0 Ω		70 $\mu\Omega$	Generate (1)	
	100 $\mu\Omega$, 1 m Ω , 10 m Ω		$1 \cdot 10^{-4} \cdot R$	Generate (1)	
	100 m Ω		$4 \cdot 10^{-5} \cdot R$	Generate (1)	
	1 Ω , 1.9 Ω		$8 \cdot 10^{-5} \cdot R$	Generate (1)	
	10 Ω		$2.5 \cdot 10^{-5} \cdot R$	Generate (1)	
	19 Ω , 100 Ω , 190 Ω , 1 k Ω , 1.9 k Ω , 10 k Ω , 19 k Ω , 100 k Ω , 190 k Ω		$2.0 \cdot 10^{-5} \cdot R$	Generate (1)	
	1 M Ω , 1.9 M Ω		$3 \cdot 10^{-5} \cdot R$	Generate (1)	
	10 M Ω		$4 \cdot 10^{-5} \cdot R$	Generate (1)	
	19 M Ω , 100 M Ω		$6 \cdot 10^{-5} \cdot R$	Generate (1)	

Annex to declaration of accreditation (scope of accreditation)
Normative document: EN ISO/IEC 17025:2005
Registration number: **K 018**

of **Intermes B.V.**
Calibration Laboratory

This annex is valid from: **08-11-2018** to **30-11-2020**

Replaces annex dated: **01-11-2017**

HCS code	Measured quantity, Instrument, Measure	Frequency	CMC¹	Remarks	Location
LF 6 5	LF Capacity				HLO
	2 nF, 10 nF, 20 nF, 200 nF	1 kHz	$1.0 \cdot 10^{-3} \cdot C$	Generate (1) only sine-shaped signals	

Annex to declaration of accreditation (scope of accreditation)
Normative document: EN ISO/IEC 17025:2005
Registration number: **K 018**

of **Intermes B.V.**
Calibration Laboratory

This annex is valid from: **08-11-2018** to **30-11-2020**

Replaces annex dated: **01-11-2017**

Remarks:

R = reading accuracy of the instrument

Temperature conditions for electrical calibrations is nominal 23 °C; temperature conditions for geometrical and torque calibrations is nominal 20 °C
, temperature conditions for pressure and temperature calibrations is nominal 21 °C

$p_e = p - p_{amb}$: p_e is overpressure, p_{amb} is ambient pressure

This list of calibrations is , unless otherwise stated, applicable for calibrations performed inside the IntermeS laboratory.

(1) Calibrations performed at customers' premises.