


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

|  |  |  |
|--|--|--|
|  <p><b>UKAS</b><br/>CALIBRATION</p> <p><b>0394</b></p> <p>Accredited to<br/><b>ISO/IEC 17025:2017</b></p> | <p><b>Trescal Limited (Trescal EMS – Rolls-Royce)</b></p> <p><b>Issue No: 059 Issue date: 11 August 2020</b></p> |  |
|  | <p><b>Trescal EMS</b><br/>Unit 2, Riverside Road<br/>Pride Park<br/>Derby<br/>DE24 8HY</p>                       | <p><b>Contact: Matt Gypps</b><br/>Tel: +44 (0) 1942 761226<br/>Fax: +44 (0) 2476 623626<br/>E-Mail: <a href="mailto:matt.gypps@trescal.com">matt.gypps@trescal.com</a><br/>Website: <a href="http://www.trescal.com">www.trescal.com</a></p> |
| <p><b>Calibration performed by the Organisations at the locations specified below</b></p>  |  |  |

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

| Location details   | Activity  | Location code |
|--|---|---------------|
| <p><b>Address (Pride Park)</b><br/>Trescal EMS<br/>Unit 2, Riverside Road<br/>Pride Park<br/>Derby<br/>DE24 8HY</p> <p><b>Local contact</b><br/>Trevor Smith<br/><br/>Tel: +44 (0) 1332 238102<br/>Email: <a href="mailto:Trevor.smith@trescal.com">Trevor.smith@trescal.com</a></p>   | <p><a href="#">Dimensional</a><br/><a href="#">Electrical</a><br/><a href="#">Humidity</a><br/><a href="#">Temperature</a><br/><a href="#">Torque</a></p> | Pride Park    |
| <p><b>Address (Ansty)</b><br/>Trescal EMS – Rolls-Royce<br/>Standards Room<br/>Building 6<br/>Ansty<br/>Coventry<br/>CV7 9JR</p> <p><b>Local contact</b><br/>David Williams<br/><br/>Tel: +44 (0) 2476 623625<br/>Fax: +44 (0) 2476 623626<br/>Email: <a href="mailto:David.williams2@rolls-royce.com">David.williams2@rolls-royce.com</a></p>     | <p><a href="#">Torque</a><br/><a href="#">Pressure</a></p>  | Ansty         |
| <p><b>Address (Inchinnan)</b><br/>Trescal EMS – Rolls-Royce<br/>Inchinnan Drive<br/>Inchinnan<br/>Renfrewshire<br/>PA4 9AF</p> <p><b>Local contact</b><br/>Robert Simpson<br/><br/>Tel: +44 (0) 141 626 8540<br/>Email: <a href="mailto:Robert.simpson@trescal.com">Robert.simpson@trescal.com</a></p>   | <p><a href="#">Dimensional</a><br/><a href="#">Torque</a></p>   | Inchinnan     |
| <p><b>Address (Washington)</b><br/>Trescal EMS – Rolls-Royce<br/>Calibration Laboratory<br/>Radial Park Road<br/>Washington<br/>Tyne and Wear<br/>NE38 9DA</p> <p><b>Local contact</b><br/>Robert Simpson Steve Jones<br/><br/>Tel: +44 (0) 191 297 3023<br/>Email: <a href="mailto:Robert.simpson@trescal.com">Robert.simpson@trescal.com</a></p> | <p><a href="#">Dimensional</a><br/><a href="#">Torque</a></p>   | Washington    |



0394

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Trescal Limited (Trescal EMS - Rolls-Royce)**

**Issue No: 059 Issue date: 11 August 2020**

Calibration performed by the Organisation at the locations specified

**Locations covered by the organisation and their relevant activities**

**Laboratory locations:**

| Location details   |  | Activity  | Location code |
|--|--|---|---------------|
| <b>Address (Bristol)</b><br>Trescal EMS – Rolls-Royce<br>Metrology Laboratory<br>(EW6/7) PO Box 3<br>Filton<br>Bristol<br>BS34 7QE         | <b>Local contact</b><br>Mr M Viney<br><br>Tel: +44 (0) 117 979 6099<br>Fax: +44 (0) 117 979 5038<br>Email: michael.viney@rolls-royce.com | <a href="#">Fuel Flow</a><br><a href="#">Torque</a> | Bristol       |
| <b>Address (Solihull)</b><br>Rolls-Royce<br>Derwent Building<br>5000 Solihull Parkway<br>Birmingham Business Park<br>Birmingham<br>B37 7YP | <b>Local contact</b><br>Jim Attwooll<br><br>Tel +44 (0) 121 2732781<br>Email: jim.attwooll@rolls-royce.com                               | Electrical DC&LF<br>Dimensional                     | Solihull      |

**Site activities performed away from the locations listed above:**

|  |   |  |      |
|--|---|--|------|
| All Rolls-Royce sites: The site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer. | <b>Local contact</b><br>Trevor Smith<br><br>Tel: +44 (0) 1332 238102<br><br>Email: Trevor.smith@trescal.com | <a href="#">Form</a><br><a href="#">Electrical</a> | Site |
|--|---|--|------|



0394

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by  
**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Trescal Limited (Trescal EMS - Rolls-Royce)**

**Issue No: 059 Issue date: 11 August 2020**

Calibration performed by the Organisation at the locations specified

### DETAIL OF ACCREDITATION

| Measured Quantity<br>Instrument or Gauge  | Range   | Calibration and<br>Measurement<br>Capability (CMC)<br>Expressed as an<br>Expanded<br>Uncertainty ( $k = 2$ ) | Remarks  | Location<br>Code |                     |
|---|---|--|--|------------------|---------------------|
| <b>LENGTH</b>   |   |  |  |                  |                     |
| RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES<br>UNLESS OTHERWISE STATED                                |   |  |  |                  |                     |
| Thread measuring cylinders  | BS3777:1964 and<br>BS 5590:1978 and specials<br>0.1 to 5.0 diameter   | 0.50 on diameter   | <b>NOTES</b><br><br>1 In addition to all items in the first column, other similar items, including parts of measuring instruments and machines, may be calibrated in accordance with the stated CMCs. Where the item or part calibrated is of lower quality due to wear, errors in geometry or form, or poor surface texture, or where any other factor adversely affects the measurement capability, greater uncertainties will be quoted.<br><br>2 The uncertainty quoted is for the departure from flatness, straightness, or squareness, i.e. the distance separating the two parallel planes which just enclose the surface under consideration.<br><br>3 All linear calibrations may be given in inch units.<br><br>4 Single start symmetrical thread forms only.<br><br>5 Single start symmetrical thread forms only.<br><br>6 By comparison with end standards using a length measuring machine. | Pride Park       |                     |
| Plain plug gauges (parallel), cylindrical setting standards, gear measuring cylinders and rollers. See Note 6 | 1 to 50 diameter<br>50 to 100<br>100 to 150<br>150 to 200<br>200 to 300                                     | 0.50<br>0.80<br>1.0<br>1.2<br>1.6  |  |                  | ] on diameter       |
| Plain ring gauges (parallel) and setting standards  | CCP 2.3.2, issue 11<br>1 to 50 diameter<br>50 to 100 diameter<br>100 to 150 diameter<br>150 to 200 diameter | 0.80<br>1.2<br>1.8<br>2.5  |  |                  | ] on diameter       |
| Length gauges, flat and spherical ended<br>See Note 6   | 0 m to 3 m  | 1.0 +<br>(5.0 x length in m)   |  |                  |                     |
| Length bars<br>Inspection and workshop grades 1 and 2   | BS 1790:1961<br>BS 5317:1976  | 0.45 +<br>(1.1 x length in m)  |  |                  |                     |
| Plain gap gauges (parallel)   | BS 969:2008<br>0.5 to 100<br>100 to 200<br>200 to 300   | 3.0<br>5.0<br>8.0  |  |                  |                     |
| Screw plug gauges (parallel) including check and setting plugs<br>See Notes 5 and 6                           | 1 to 100 diameter<br>100 to 300 diameter  | 2.5<br>5.0   |  |                  | ] On pitch diameter |
| Screw ring gauges (parallel)<br>See Notes 4 and 6   | 5 to 75 diameter<br>100 to 150 diameter<br>150 to 300 diameter  | 4.0<br>5.0<br>8.0  |  |                  | ] On pitch diameter |
| Screw pitch   | 0.2 to 8  | 1.5  |  |                  |                     |
| Screw flank angle   | 0° to 50°   | 5.0 minutes of arc   |  |                  |                     |
| Parallels   | BS 906:Parts 1 and 2:1992<br>5 to (50 x 100 x 400)  | 1.5 to 5.0   |  |                  |                     |



0394  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Trescal Limited (Trescal EMS - Rolls-Royce)**  
**Issue No: 059 Issue date: 11 August 2020**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge                                       | Range  | Calibration and<br>Measurement<br>Capability (CMC)<br>Expressed as an<br>Expanded<br>Uncertainty ( $k = 2$ )   | Remarks   | Location<br>Code |  |
|--|--|--|---|------------------|--|
| <b>LENGTH</b> (continued)  |  |  |   |                  |  |
| RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES<br>UNLESS OTHERWISE STATED |  |  |   |                  |  |
| Gauge blocks   |  | Class (see note)   | Note  | Pride Park       |  |
| Inch<br>(Steel)  | BS 4311-1:2007<br>0 in to 0.4 in<br>0.4 in to 1 in<br>Size 2 in<br>3 in<br>4 in            | C      D<br>3.0 $\mu$ in   4.0 $\mu$ in<br>4.0 $\mu$ in   5.0 $\mu$ in<br>5.0 $\mu$ in<br>6.0 $\mu$ in<br>7.0 $\mu$ in   | Class C uncertainties apply to the measurement of length by comparison with grade K standards of a similar material.<br>Class D uncertainties apply to the measurement of length by comparison with grade K standards of a dissimilar material. |                  |  |
| Millimetre<br>(Steel)  | BS EN ISO 3650:1999<br>0 to 10<br>10 to 25<br>Size 30, 40, 50<br>60, 70, 75<br>80, 90, 100 | C      D<br>0.080   0.10<br>0.10   0.13<br>0.12<br>0.15<br>0.18  | The uncertainties apply to new and used grade 0, 1 and 2 gauges to BS EN ISO 3650:1999 and BS 4311-1:2007.  |                  |  |
| Vee blocks   | BS 3731:1987<br>20 to 150 diameter, Vee capacity   | 2.5 to 5.0   |   |                  |  |
| Receiver, position and<br>profile gauges, jigs,<br>fixtures                    | 1500 x 750 x 750<br><br>1500 x 3200 x 1100   | From first principles:<br>Dependant on size and<br>features<br>Minimum per co-<br>ordinate:<br>3.0 + (10 x length in m)<br><br>Using CMM:<br>Dependant on size and<br>features<br>Minimum per co-<br>ordinate:<br>5.0 + (10 x length in m) |   |                  |  |
| <b>ANGLE</b>   |  |  |   |                  |  |
| Squares  |  |  |   |                  |  |
| Blade type   | BS 939:2007, CCP 2.4.17<br>issue 10<br>50 to 300<br>300 to 600                             | 3.0<br>5.0   |   |                  |  |
| Cylindrical  | BS 939:2007, CCP 2.4.17<br>issue 10<br>75 to 300<br>300 to 600                             | 2.0   On squareness<br>4.0   See Note 2  |   |                  |  |



0394  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Trescal Limited (Trescal EMS - Rolls-Royce)**  
**Issue No: 059 Issue date: 11 August 2020**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge                                       | Range   | Calibration and<br>Measurement<br>Capability (CMC)<br>Expressed as an<br>Expanded<br>Uncertainty ( $k = 2$ ) | Remarks   | Location<br>Code |
|--|---|--|---|------------------|
| <b>ANGLE</b> (continued)   |   |  |   | Pride Park       |
| RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES<br>UNLESS OTHERWISE STATED |   |  |   |                  |
| Block  | BS 939:2007<br>50 to 300<br>300 to 600                        | 3.0<br>5.0   |   |                  |
| Angle gauges   | NPL type<br>Other types                                       | 2.0 seconds of arc<br>3.0 seconds of arc   | In-house methods based<br>on MOY/SCMI/18                              |                  |
| Sine bars and tables   | BS 3064:1978<br>100 to 500 length                             | Linear dimensions:<br>1.0 + (10 x length in m)<br>Overall performance:<br>3.0 seconds of arc                 |   |                  |
| Sine centres   | 100 to 500 length or between<br>centres                       | Linear dimensions:<br>1.0+ (10 x length in m)<br>Overall performance<br>5.0 seconds of arc                   | In-house methods based<br>on BS 3064:1978                             |                  |
| Compound sine tables   | 100 to 500 length   | 5.0 seconds of arc   |   |                  |
| <b>FORM</b>  |   |  |   |                  |
| Straightedges<br>Cast iron<br>Steel<br>Granite                                 | BS 5204:Part 1:1975 and<br>BS 5204:Part 2:1977<br>0 m to 2m   | 1.0<br>+ (2.0 x length in m)<br>See Note 2   |   |                  |
| Roundness<br>External<br>Internal  | BS 3730:Part 2:1982<br>0 to 350 diameter<br>3 to 350 diameter | 0.050 on radius<br>0.050 on radius   |   |                  |
| Steel balls  | 1 to 25 diameter  | 0.50 on diameter   | By comparison with end standards<br>using a length measuring machine. |                  |
| <b>MEASURING INSTRUMENTS AND MACHINES</b>                                      |   |  |   |                  |
| Micrometers  |   |  |   |                  |
| External   | BS 870:2008,<br>CCP 2.4.1 issue 12<br>0 to 600                | Heads: 2.0<br>Setting and<br>Extension rods:<br>1.0<br>+ (5.0 x length in m)                                 |   |                  |
| Internal   | BS 959:2008<br>0 to 300                                       |  |   |                  |
| Depth  | BS 6468:2008<br>0 to 300                                      |  |   |                  |
| Micrometer heads   | BS 1734:1951; 0 to 100  | 1.0  |   |                  |
| Bench micrometer   | 0 to 100  | Overall performance<br>1.0   | In-house method based on<br>MOY/SCMI/22                               |                  |



0394  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Trescal Limited (Trescal EMS - Rolls-Royce)**  
**Issue No: 059 Issue date: 11 August 2020**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge                                       | Range   | Calibration and<br>Measurement<br>Capability (CMC)<br>Expressed as an<br>Expanded<br>Uncertainty ( $k = 2$ ) | Remarks   | Location<br>Code |
|--|---|--|---|------------------|
| <b>MEASURING INSTRUMENTS AND MACHINES</b> (continued)                          |   |  |   | Pride Park       |
| RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES<br>UNLESS OTHERWISE STATED |   |  |   |                  |
| Height setting micrometer  | 0 to 300  | Heads 1.0<br>Stepped column 1.6<br>Overall performance<br>2.0  | By comparison with end standards.   |                  |
| Riser Blocks   | 150<br>300  | 1.0<br>2.0   | By comparison with end standards.   |                  |
| Height gauges - (Simple)<br>including vernier, dial and<br>digital types       | BS EN ISO 13225:2012<br>0 to 300  | 4.0  |   |                  |
| Vernier gauges   |   |  |   |                  |
| Caliper<br>Height  | BS 887:2008<br>BS 1643:2008<br>0 to 1200  | Overall performance:<br>10 + (30 x length in m)  |   |                  |
| Depth  | BS 6365:2008<br>0 to 600  |  |   |                  |
| Dial gauges and dial test<br>indicators  | BS 907:2008 and BS<br>2795:1981<br>0 to 50  | 1.0  |   |                  |
| Dividing heads and<br>Rotary tables  | 100 to 750 capacity   | Linear dimensions<br>1.0 + (10 x length in m)<br>Overall angular<br>performance<br>3.0 seconds of arc        | By comparison, using angle standards<br>and autocollimator.   |                  |
| Spirit levels  | BS 958:1968 and BS<br>3509:1962<br>Nominal sensitivity 5 seconds<br>of arc to 60 minutes of arc | Mean sensitivity: 10 %<br>of nominal; minimum<br>0.50 seconds of arc   |   |                  |
| Clinometers  | 0° to 360°  | 10 seconds of arc  | In-house method based on<br>MOY/SCMI/36   |                  |
| Levels, electronic   | 0 seconds of arc to 10 minutes<br>of arc  | 1.0 % of range<br>minimum 0.50 seconds<br>of arc   | The quoted uncertainty will be<br>particularly dependent on the sensitivity<br>of the device. Using small angle<br>generator. |                  |
| Orifice plates   | BS EN ISO 5167-2:2003 (and<br>similar devices) Bore d<br>diameter 1.0 mm to 1 m                 | 4.0 +<br>(6.0 x length in m)   |   |                  |



0394  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Trescal Limited (Trescal EMS - Rolls-Royce)**  
**Issue No: 059 Issue date: 11 August 2020**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge                              | Range  | Calibration and<br>Measurement<br>Capability (CMC)<br>Expressed as an<br>Expanded<br>Uncertainty ( $k = 2$ ) | Remarks  | Location<br>Code |
|---|--|--|--|------------------|
| TORQUE<br><br>Hand torque tools<br>(excluding torque<br>screwdrivers) | BS EN ISO 6789:2017<br>And<br>BS EN ISO 6789:2003<br>(withdrawn and superseded)<br>and<br>CCP 3.6.6 Issue 9.0<br>1.0 N·m to 1000 N·m | 1.0 %  | The quoted uncertainty will be particularly dependent on the repeatability of the unit under test. |                  |



0394  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Trescal Limited (Trescal EMS - Rolls-Royce)**  
**Issue No: 059 Issue date: 11 August 2020**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge | Range  | Calibration and<br>Measurement<br>Capability (CMC)<br>Expressed as an<br>Expanded<br>Uncertainty ( $k = 2$ )                             | Remarks  | Location<br>Code |  |  |
|--|--|--|--|------------------|--|--|
| <b>ELECTRICAL MEASUREMENTS</b>           |  |  |  |                  |  |  |
| <b>DC VOLTAGE</b>                        |  |  |  |                  |  |  |
| Measurement                              | Up to 200 mV<br>200 mV to 2 V<br>2 V to 20 V<br>20 V to 200 V<br>200 V to 1 000 V  | 6.5 ppm + 1.3 $\mu$ V<br>5.1 ppm<br>6.1 ppm<br>9.4 ppm<br>9.6 ppm  | The stated CMCs are for a four-terminal configuration and may be increased if a two-terminal configuration is necessary. | Pride Park       |  |  |
| Generation                               | 0 mV to 2 mV<br>2 mV to 20 mV<br>20 mV to 200 mV<br>200 mV to 2 V<br>2 V to 20 V<br>20 V to 200 V<br>200 V to 1100 V   | 1.3 $\mu$ V<br>1.3 $\mu$ V<br>1.3 $\mu$ V<br>2.8 ppm + 0.90 $\mu$ V<br>2.2 ppm + 2.5 $\mu$ V<br>3.2 ppm + 39 $\mu$ V<br>5.6ppm + 0.39 mV |  |                  |  |  |
| <b>DC RESISTANCE</b>                     |  |  |  |                  |  |  |
| Measurement                              | 0 $\Omega$ to 20 $\Omega$<br>20 $\Omega$ to 200 $\Omega$<br>200 $\Omega$ to 200 k $\Omega$<br>200 k $\Omega$ to 2 M $\Omega$<br>2 M $\Omega$ to 20 M $\Omega$<br>20 M $\Omega$ to 200 M $\Omega$<br>200 M $\Omega$ to 2 G $\Omega$ | 30 ppm + 20 $\mu$ $\Omega$<br>13 ppm<br>14 ppm<br>24 ppm<br>55 ppm<br>450 ppm<br>0.50%   |  |                  |  |  |
| Generation                               |  |  |  |                  |  |  |
| Four terminal configuration              | 10 $\Omega$<br>100 $\Omega$<br>1 k $\Omega$<br>10 k $\Omega$<br>100 k $\Omega$<br>1 M $\Omega$<br>10 M $\Omega$<br>100 M $\Omega$  | 5.7 ppm<br>3.9 ppm<br>3.6 ppm<br>3.2 ppm<br>4.5 ppm<br>10 ppm<br>19 ppm<br>65 ppm  |  |                  |  |  |
| Two terminal configuration               | 0 $\Omega$ , 10 $\Omega$ and 100 $\Omega$<br>1 k $\Omega$<br>10 k $\Omega$<br>100 k $\Omega$<br>1 M $\Omega$<br>10 M $\Omega$<br>100 M $\Omega$  | 10 m $\Omega$<br>79 ppm<br>8.3 ppm<br>4.5 ppm<br>10 ppm<br>19 ppm<br>65 ppm  |  |                  |  |  |





0394  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Trescal Limited (Trescal EMS - Rolls-Royce)**  
**Issue No: 059 Issue date: 11 August 2020**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge | Range   | Calibration and<br>Measurement<br>Capability (CMC)<br>Expressed as an<br>Expanded<br>Uncertainty ( $k = 2$ ) | Remarks | Location<br>Code |
|--|---|--|---------|------------------|
| DC CURRENT                               |   |  |         | Pride Park       |
| Measurement                              | 10 $\mu$ A to 200 $\mu$ A<br>200 $\mu$ A to 200 mA<br>200 mA to 2 A                                   | 100 ppm<br>100 ppm<br>170 ppm  |         |                  |
| Generation                               | 10 $\mu$ A to 200 $\mu$ A<br>200 $\mu$ A to 2 mA<br>2 mA to 20 mA<br>20 mA to 200 mA<br>200 mA to 2 A | 22 ppm + 1.6 nA<br>15 ppm + 7.8 nA<br>15 ppm + 78 nA<br>15 ppm + 0.78 $\mu$ A<br>26 ppm + 16 $\mu$ A         |         |                  |
| AC VOLTAGE                               |   |  |         |                  |
| Measurement                              | 10 mV to 200 mV<br>10 Hz to 10 kHz<br>10 kHz to 30 kHz<br>30 kHz to 50 kHz                            | 390 ppm<br>640 ppm<br>0.17%  |         |                  |
|  | 200 mV to 2 V<br>10 Hz to 10 kHz<br>10 kHz to 30 kHz<br>30 kHz to 50 kHz                              | 190 ppm<br>270 ppm<br>870 ppm  |         |                  |
|  | 2 V to 20 V<br>10 Hz to 10 kHz<br>10 kHz to 30 kHz<br>30 kHz to 50 kHz                                | 180 ppm<br>270 ppm<br>870 ppm  |         |                  |
|  | 20 V to 200 V<br>10 Hz to 10 kHz<br>10 kHz to 30 kHz<br>30 kHz to 50 kHz                              | 190 ppm<br>270 ppm<br>870 ppm  |         |                  |
|  | 200 V to 300 V<br>40 Hz to 10 kHz<br>10 kHz to 30 kHz   | 250 ppm<br>390 ppm   |         |                  |
|  | 300 V to 1 kV<br>40 Hz to 10 kHz<br>10 kHz to 30 kHz  | 0.11 %<br>0.12 %   |         |                  |
|  | 200 V to 1 kV<br>30 kHz to 50 kHz   | 0.20 %   |         |                  |



0394  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Trescal Limited (Trescal EMS - Rolls-Royce)**  
**Issue No: 059 Issue date: 11 August 2020**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge | Range  | Calibration and<br>Measurement<br>Capability (CMC)<br>Expressed as an<br>Expanded<br>Uncertainty ( $k = 2$ ) | Remarks | Location<br>Code |
|--|--|--|---------|------------------|
| AC VOLTAGE (continued)                   |  |  |         | Pride Park       |
| Generation                               | 1 mV to 2 mV<br>20 Hz to 100 kHz   | 0.74% + 4.2 $\mu$ V  |         |                  |
|  | 2 mV to 20 mV<br>20 Hz to 100 kHz  | 0.032% + 4.2 $\mu$ V   |         |                  |
|  | 20 mV to 200 mV<br>20 Hz to 50 kHz<br>50 kHz to 100 kHz<br>100 kHz to 300 kHz<br>300 kHz to 1 MHz                        | 130 ppm + 7.0 $\mu$ V<br>0.044%<br>0.17%<br>0.83%  |         |                  |
|  | 200 mV to 2 V<br>10 Hz to 20 Hz<br>20 Hz to 50 kHz<br>50 kHz to 100 kHz<br>100 kHz to 300 kHz<br>300 kHz to 1 MHz        | 310 ppm<br>89 ppm<br>180 ppm<br>0.13 %<br>0.52%  |         |                  |
|  | 2 V to 20 V<br>10 Hz to 20 Hz<br>20 Hz to 50 kHz<br>50 kHz to 100 kHz<br>100 kHz to 300 kHz<br>300 kHz to 1 MHz          | 290 ppm<br>73 ppm<br>240 ppm<br>0.12%<br>0.52%   |         |                  |
|  | 20 V to 200 V<br>10 Hz to 20 Hz<br>20 Hz to 50 kHz<br>50 kHz to 100 kHz  | 220 ppm<br>100 ppm<br>200 ppm  |         |                  |
|  | 200 V to 1 kV<br>45 Hz to 33 kHz   | 130 ppm  |         |                  |
| AC CURRENT                               |  |  |         |                  |
| Measurement                              | 40 Hz to 1 kHz:<br>10 $\mu$ A to 200 $\mu$ A<br>200 $\mu$ A to 200 mA<br>200 mA to 2 A                                   | 370 ppm + 16 nA<br>840 ppm<br>660 ppm + 310 $\mu$ A  |         |                  |
| Generation                               | 40 Hz to 1 kHz:<br>10 $\mu$ A to 200 $\mu$ A<br>200 $\mu$ A to 2 mA<br>2 mA to 20 mA<br>20 mA to 200 mA<br>200 mA to 2 A | 90 ppm + 7.8 nA<br>85 ppm + 78 nA<br>85 ppm + 0.78 $\mu$ A<br>110 ppm + 7.8 $\mu$ A<br>370 ppm + 78 $\mu$ A  |         |                  |



0394  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Trescal Limited (Trescal EMS - Rolls-Royce)**  
**Issue No: 059 Issue date: 11 August 2020**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge   | Range   | Calibration and<br>Measurement<br>Capability (CMC)<br>Expressed as an<br>Expanded<br>Uncertainty ( $k = 2$ )  | Remarks  | Location<br>Code |
|--|---|---|--|------------------|
| <b>FREQUENCY</b>   |   |   |  | Pride Park       |
| Specific Values  | 1 MHz and 10 MHz  | 1.2 parts in $10^9$   | For calibrating oscillators  |                  |
| Other Values   | 0.1 Hz to 1 Hz<br>1 Hz to 10 Hz<br>10 Hz to 100 Hz<br>100 Hz to 1 kHz<br>1 kHz to 10 kHz<br>10 kHz to 100 kHz<br>100 kHz to 1 MHz<br>1 MHz to 60 MHz<br>60 MHz to 100 MHz<br>100 MHz to 150 MHz<br>150 MHz to 500 MHz   | 1.5 parts in $10^3$<br>1.5 parts in $10^4$<br>1.5 parts in $10^5$<br>1.5 parts in $10^6$<br>1.5 parts in $10^7$<br>1.7 parts in $10^8$<br>3.9 parts in $10^9$<br>2.5 parts in $10^9$<br>1.2 parts in $10^9$<br>2.4 parts in $10^9$<br>1.4 parts in $10^9$ | Measurement capability only above<br>60 MHz  |                  |
| <b>ELAPSED TIME</b>  |   |   |  |                  |
| Stop watches (mechanical<br>and electronic)  | $\pm 0.5$ s error / 24 hours<br>$\pm 2.0$ s error / 24 hours<br><br>10 s to 24 hours  | 0.062 s<br>0.090 s<br><br>0.41 s  | Time reference measurement<br>per 24 hour period<br>per 24 hour period<br><br>Real time measurement  |                  |
| <b>TEMPERATURE SIMULATION</b>  |   |   |  |                  |
| Temperature indicators and simulators (thermocouple type),<br>calibration by electrical simulation |   |   |  |                  |
| Base metal thermocouples   | Type J, -210 °C to 0 °C<br>Type J, 0 °C to 1200 °C<br><br>Type K, -270 °C to -200 °C<br>Type K, -200 °C to 0 °C<br>Type K, 0 °C to 1370 °C<br><br>Type N, -270 °C to -200 °C<br>Type N, -200 °C to 0 °C<br>Type N, 0 °C to 1300 °C<br><br>Type T, -270 °C to -200 °C<br>Type T, -200 °C to 0 °C<br>Type T, 0 °C to 400 °C | 0.064 °C<br>0.018 °C<br><br>0.23 °C<br>0.070 °C<br>0.022 °C<br><br>0.62 °C<br>0.084 °C<br>0.027 °C<br><br>0.19 °C<br>0.070 °C<br>0.020 °C   | excluding cold junction<br>compensation<br><br>excluding cold junction<br>compensation<br><br>excluding cold junction<br>compensation<br><br>excluding cold junction<br>compensation |                  |



0394  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Trescal Limited (Trescal EMS - Rolls-Royce)**  
**Issue No: 059 Issue date: 11 August 2020**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge   | Range                                       | Calibration and<br>Measurement<br>Capability (CMC)<br>Expressed as an<br>Expanded<br>Uncertainty ( $k = 2$ ) | Remarks                                 | Location<br>Code |
|--|---|--|---|------------------|
| Temperature indicators and simulators (thermocouple type),<br>calibration by electrical simulation (continued) |   |  |   | Pride Park       |
| Cold junction<br>compensation  | At ambient temperature of<br>20 °C ± 2.0 °C | 0.13 °C  |   |                  |
| Base metal thermocouples   | Type J, -210 °C to 0 °C                     | 0.14 °C  | including cold junction<br>compensation |                  |
|  | Type J, 0 °C to 1200 °C                     | 0.13 °C  |   |                  |
|  | Type K, -270 °C to -200 °C                  | 0.24 °C  | including cold junction<br>compensation |                  |
|  | Type K, -200 °C to 0 °C                     | 0.15 °C  |   |                  |
|  | Type K, 0 °C to 1370 °C                     | 0.13 °C  |   |                  |
| Type N, -270 °C to -200 °C   | Type N, -200 °C to 0 °C                     | 0.53 °C  | including cold junction<br>compensation |                  |
|  | Type N, -200 °C to 0 °C                     | 0.15 °C  |   |                  |
|  | Type N, 0 °C to 1300 °C                     | 0.13 °C  |   |                  |
| Type T, -270 °C to -200 °C   | Type T, -200 °C to 0 °C                     | 0.21 °C  | including cold junction<br>compensation |                  |
|  | Type T, -200 °C to 0 °C                     | 0.15 °C  |   |                  |
|  | Type T, 0 °C to 400 °C                      | 0.13 °C  |   |                  |
| Noble metal<br>thermocouples   | -50 °C to 0 °C                              | 0.19 °C  | excluding cold junction<br>compensation |                  |
|  | 0 °C to 250 °C<br>250 °C to 1760 °C         | 0.17 °C<br>0.089 °C  |   |                  |
| Cold junction<br>compensation  | At ambient temperature of<br>20 °C ± 2 °C   | 0.17 °C  |   |                  |
| Temperature indicators and simulators (thermocouple type),<br>calibration by electrical simulation             |   |  |   |                  |
| Noble metal<br>thermocouples   | -50 °C to 0 °C                              | 0.24 °C  | including cold junction<br>compensation |                  |
|  | 0 °C to 250 °C                              | 0.22 °C  |   |                  |
|  | 250 °C to 1760 °C                           | 0.18 °C  |   |                  |
| PRT simulation (Pt 100)  | -200 °C to 0 °C                             | 0.017 °C   |   |                  |
|  | 0 °C to 100 °C                              | 0.018 °C   |   |                  |
|  | 100 °C to 400 °C                            | 0.020 °C   |   |                  |
|  | 400 °C to 630 °C                            | 0.023 °C   |   |                  |
|  | 630 °C to 850 °C                            | 0.026 °C   |   |                  |



0394  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Trescal Limited (Trescal EMS - Rolls-Royce)**  
**Issue No: 059 Issue date: 11 August 2020**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge                  | Range                                    | Calibration and<br>Measurement<br>Capability (CMC)<br>Expressed as an<br>Expanded<br>Uncertainty ( $k = 2$ ) | Remarks  | Location<br>Code |
|---|--|--|--|------------------|
| <b>TEMPERATURE</b>  |  |  |  |                  |
| Thermocouples   |  |  |  |                  |
| Base metal  | -20 °C to +200 °C                        | 0.45 °C  | Calibration within both liquid and metal medium  | Pride Park       |
| Noble metal   | -20 °C to 200 °C                         | 0.92 °C  | Calibration within both liquid and metal medium  |                  |
| Resistance thermometers                                   | -20 °C to +200 °C                        | 0.070 °C   | Calibration within both liquid and metal medium  |                  |
| Electronic thermometers with sensors; analogue or digital | Ranges as per sensor                     | As per sensor type   | Calibration within both liquid and metal medium  |                  |
| <b>HUMIDITY</b>   |  |  |  |                  |
| Dew point   | -10 °C to 0 °C<br>0 °C to 20 °C          | 0.12 °C dp<br>0.10 °C dp   | By comparison with dew-point hygrometer and Platinum Resistance Thermometers           | Ansty            |
| Relative Humidity   | 5 %rh to 95 %rh                          | 2.0 %rh  | At air temperature 5 °C to 60 °C   |                  |
| Air Temperature   | 5 °C to 60 °C                            | 0.4 °C   |  |                  |
| <b>PRESSURE</b>   |  |  |  |                  |
| Methods consistent with EURAMET CG3                       |  |  |  |                  |
| Hydraulic pressure (Gauge)                                |  |  |  | Ansty            |
| Pressure indicating instruments and gauges                | 600 kPa to 120 MPa                       | 0.010 %  | Calibration of pressure measuring devices with an electrical output may be undertaken. |                  |
| Pneumatic pressure (Gauge)                                |  |  |  |                  |
| Pressure indicating instruments and gauges                | 3.70 kPa to 3.5 MPa                      | 0.010 %  |  |                  |
| Pneumatic pressure (Absolute)                             |  |  |  | Ansty            |
| Pressure indicating instruments and gauges                | 3.70 kPa to 3.5 MPa<br>75 kPa to 120 kPa | 0.010 % + 5.0 Pa<br>17 Pa  |  |                  |



0394  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Trescal Limited (Trescal EMS - Rolls-Royce)**  
**Issue No: 059 Issue date: 11 August 2020**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge  | Range  | Calibration and<br>Measurement<br>Capability (CMC)<br>Expressed as an<br>Expanded<br>Uncertainty ( $k = 2$ ) | Remarks  | Location<br>Code |  |
|---|--|--|--|------------------|--|
| TORQUE<br><br>Hand torque tools<br><br>LENGTH   | CCP 3.6.6 issue 9.0<br>0.113 N·m to 1356 N·m                               | 1.0 %  | The quoted uncertainty will be particularly dependent on the repeatability of the unit under test. | Ansty            |  |
| RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES<br>UNLESS OTHERWISE STATED                    |  |  |  | Inchinman        |  |
| Thread measuring cylinders  | BS3777:1964 and<br>BS 5590:1978 and specials<br>0.1 to 5.0 diameter        | 0.50 on diameter   | By comparison with end standards using a length measuring machine.                                 |                  |  |
| Plain plug gauges (parallel), cylindrical setting standards, gear measuring cylinders and rollers | 1 to 50 diameter<br>50 to 100 diameter<br>100 to 150 diameter              | 0.50<br>0.80<br>1.0 on diameter  |  |                  |  |
| Plain ring gauges (parallel) and setting standards  | CCP 2.3.2<br>1 to 50 diameter<br>50 to 100 diameter<br>100 to 150 diameter | 0.80<br>1.2<br>1.8 on diameter   |  |                  |  |
| Length gauges, flat and spherical ended   | 0 m to 1 m   | 1.0 +<br>(5.0 x length in m)   |  |                  | By comparison with end standards using a length measuring machine  |
| Plain gap gauges (parallel)   | BS 969:2008<br>0.5 to 100<br>100 to 200                                    | 3.0<br>5.0   |  |                  | Single start symmetrical thread forms only. By comparison with end standards using a length measuring machine. |
| Screw plug gauges (parallel) excluding check and setting plugs                                    | 1 to 100 diameter  | 2.5 on pitch diameter  |  |                  |  |
| Parallels   | BS 906:Parts 1 and 2:1992<br>5 to (50 x 100 x 400)                         | 1.5 to 5.0, dependant on size and grade  |  |                  |  |



0394  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Trescal Limited (Trescal EMS - Rolls-Royce)**  
**Issue No: 059 Issue date: 11 August 2020**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge                                       | Range  | Calibration and<br>Measurement<br>Capability (CMC)<br>Expressed as an<br>Expanded<br>Uncertainty ( $k = 2$ ) | Remarks                                 | Location<br>Code |
|--|--|--|---|------------------|
| RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES<br>UNLESS OTHERWISE STATED |  |  |   |                  |
| <b>FORM</b>  |  |  |   |                  |
| Surface plates<br>Granite<br>Cast iron   | BS 817:2008<br>160 x 100 to 4 m x 4 m                        | 1.5 +<br>(0.80 x diagonal in m)<br>See Note 2  |   |                  |
| Straightedges<br>Cast iron<br>Steel<br>Granite                                 | BS 5204:Part 1:1975 and<br>BS 5204:Part 2:1977<br>0 m to 2 m | 1.0<br>+ (2.0 x length in m)<br>See Note 2   |   |                  |
| <b>MEASURING INSTRUMENTS AND MACHINES</b>                                      |  |  |   |                  |
| <b>LENGTH</b>  |  |  |   |                  |
| Micrometers<br>External  | BS 870:2008,<br>CCP 2.4.1 issue12<br>0 to 300                | Heads: 2.0<br>Setting and<br>Extension rods:<br>1.0<br>+ (5.0 x length<br>in m)                              |   |                  |
| Internal   | BS 959:2008<br>0 to 300                                      |  |   |                  |
| Depth  | BS 6468:2008<br>0 to 300                                     |  |   |                  |
| Micrometer heads   | BS 1734:1951<br>0 to 100                                     | 1.0  |   |                  |
| Bench micrometer   | 0 to 100   | Overall performance<br>1.0   | In-house method based on<br>MOY/SCMI/22 |                  |
| Height setting micrometer  | 0 to 300   | Heads 1.0<br>Stepped column 1.6<br>Overall performance<br>2.0  | By comparison with end standards.       |                  |
| Riser Blocks for above   | 150<br>300   | 1.0<br>2.0   | By comparison with end standards.       |                  |
| Vernier gauges<br>Caliper  | BS 887:2008<br>0 to 300                                      | Overall<br>performance:<br>10 + (30 x<br>length in m)  |   |                  |
| Height   | BS 1643:2008<br>0 to 300                                     |  |   |                  |
| Depth  | BS 6365:2008<br>0 to 300                                     |  |   |                  |



0394  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Trescal Limited (Trescal EMS - Rolls-Royce)**  
**Issue No: 059 Issue date: 11 August 2020**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge                                       | Range   | Calibration and<br>Measurement<br>Capability (CMC)<br>Expressed as an<br>Expanded<br>Uncertainty ( $k = 2$ )   | Remarks   | Location<br>Code |
|--|---|--|---|------------------|
| MEASURING INSTRUMENTS AND MACHINES (continued)                                 |   |  |   | Inchman          |
| LENGTH (continued)   |   |  |   |                  |
| RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES<br>UNLESS OTHERWISE STATED |   |  |   |                  |
| Dial gauges and dial test indicators   | BS 907:2008 and BS 2795:1981<br>0 to 50         | 1.0  | By comparison, using angle standards and autocollimator.  |                  |
| Dividing heads / Rotary tables   | 0 to 750 capacity                               | <div style="border: 1px solid black; padding: 2px; display: inline-block;">           Linear dimensions<br/> <math>1.0 + (10 \times \text{length in m})</math><br/>           Overall angular performance<br/>           3 seconds of arc         </div> |   |                  |
| TORQUE   |   |  |   |                  |
| Hand torque tools  | CCP 3.6.6 issue 9.0<br>0.136 N·m to 677.91 N·m  | 1.0 %  | The quoted uncertainty will be particularly dependent on the repeatability of the unit under test |                  |
| LENGTH   |   |  |   |                  |
| RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES<br>UNLESS OTHERWISE STATED |   |  |   |                  |
| Micrometers<br>External  | BS 870:2008,<br>CCP 2.4.1 issue12<br>0 to 600   | Heads: 2.0<br>Setting and<br>Extension rods:<br>$1.0 + (5.0 \times \text{length in m})$  |   |                  |
| Internal<br>Depth  | BS 959:2008; 0 to 150<br>BS 6468:2008; 0 to 150 |  |   |                  |
| Vernier gauges<br>Caliper<br>Depth   | BS 887:2008; 0 to 600<br>BS 6365:2008; 0 to 150 | Overall performance:<br>$10 + (30 \times \text{length in m})$  |   |                  |
| Dial gauges and dial test indicators   | BS 907:2008 and BS 2795:1981<br>0 to 50         | 1.5  |   |                  |





0394  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Trescal Limited (Trescal EMS - Rolls-Royce)**  
**Issue No: 059 Issue date: 11 August 2020**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge                       | Range  | Calibration and<br>Measurement<br>Capability (CMC)<br>Expressed as an<br>Expanded<br>Uncertainty ( $k = 2$ )   | Remarks   | Location<br>Code |
|--|--|--|---|------------------|
| <b>TORQUE</b><br>Hand torque tools                             | CCP 3.6.6 issue 9.0<br>0.1 N·m to 1000 N·m   | 1.0 %  | The quoted uncertainty will be particularly dependent on the repeatability of the unit under test | Washington       |
| <b>FUEL FLOW</b><br>Flow rate - volume<br>Flow rate - mass     | 5 l/hr to 27000 l/hr<br>4 kg/hr to 21330 kg/hr   | 0.10 %<br>0.20 %   | Piston prover method<br><br>Calibration fluid AVTUR (Aviation fuel)                               | Bristol          |
| <b>TORQUE</b><br>Hand torque tools                             | CCP 3.6.6 issue 9.0<br>0.1 N·m to 1000 N·m   | 1.0 %  | The quoted uncertainty will be particularly dependent on the repeatability of the unit under test | Bristol          |
| <b>ELECTRICAL MEASUREMENTS</b><br>DC RESISTANCE<br>Measurement | 0 $\Omega$ to 20 $\Omega$<br>20 $\Omega$ to 200 $\Omega$<br>200 $\Omega$ to 2 k $\Omega$<br>2 k $\Omega$ to 20 k $\Omega$<br>20 k $\Omega$ to 200 k $\Omega$<br>200 k $\Omega$ to 2 M $\Omega$<br>2 M $\Omega$ to 20 M $\Omega$<br>20 M $\Omega$ to 200 M $\Omega$<br>200 M $\Omega$ to 1 G $\Omega$ | 28 ppm + 25 $\mu\Omega$<br>16 ppm + 100 $\mu\Omega$<br>13 ppm + 1.0 m $\Omega$<br>13 ppm + 10 m $\Omega$<br>16 ppm + 100 m $\Omega$<br>27 ppm + 2.0 $\Omega$<br>75 ppm + 100 $\Omega$<br>500 ppm + 12 k $\Omega$<br>1.0 % + 1.1 M $\Omega$ |   | Solihull         |
| DC VOLTAGE<br>Measurement                                      | 0 mV to 200 mV<br>200 mV to 2 V<br>2 V to 20 V<br>20 V to 200 V<br>200 V to 1 kV   | 11 ppm + 1.2 $\mu$ V<br>8.5 ppm + 0.9 $\mu$ V<br>8.5 ppm + 4.0 $\mu$ V<br>13 ppm + 60 $\mu$ V<br>13 ppm + 600 $\mu$ V  |   | Solihull         |
| DC CURRENT<br>Measurement                                      | 0 $\mu$ A to 200 $\mu$ A<br>200 $\mu$ A to 2 mA<br>2 mA to 20 mA<br>20 mA to 200 mA<br>200 mA to 2 A   | 140 ppm + 0.60 nA<br>130 ppm + 6.0 nA<br>130 ppm + 60 nA<br>130 ppm + 1.3 $\mu$ A<br>240 ppm + 25 $\mu$ A  |   | Solihull         |



0394  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Trescal Limited (Trescal EMS - Rolls-Royce)**  
**Issue No: 059 Issue date: 11 August 2020**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge | Range  | Calibration and<br>Measurement<br>Capability (CMC)<br>Expressed as an<br>Expanded<br>Uncertainty ( $k = 2$ ) | Remarks | Location<br>Code |
|--|--|--|---------|------------------|
| AC VOLTAGE<br>Measurement                | 10 mV to 200 mV<br>40 Hz to 10 kHz                 | 320 ppm + 5.0 $\mu$ V  |         | Solihull         |
|  | 200 mV to 2 V<br>40 Hz to 10 kHz                   | 210 ppm + 25 $\mu$ V   |         |                  |
|  | 2 V to 20 V<br>40 Hz to 10 kHz                     | 210 ppm + 250 $\mu$ V  |         |                  |
|  | 20 V to 200 V<br>40 Hz to 10 kHz                   | 210 ppm + 2.5 mV   |         |                  |
|  | 200 V to 1 kV<br>55 Hz to 1 kHz<br>1 kHz to 10 kHz | 360 ppm + 50 mV<br>450 ppm + 50 mV   |         |                  |
| AC CURRENT<br>Measurement                | 10 $\mu$ A to 200 $\mu$ A<br>55 Hz to 1 kHz        | 600 ppm + 25 nA  |         |                  |
|  | 200 $\mu$ A to 2 mA<br>55 Hz to 1 kHz              | 400 ppm + 250 nA   |         |                  |
|  | 2 mA to 20 mA<br>55 Hz to 1 kHz                    | 400 ppm + 2.5 $\mu$ A  |         |                  |
|  | 20 mA to 200 mA<br>55 Hz to 1 kHz                  | 400 ppm + 25 $\mu$ A   |         |                  |
|  | 200 mA to 2 A<br>55 Hz to 1 kHz                    | 900 ppm + 500 $\mu$ A  |         |                  |



0394  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Trescal Limited (Trescal EMS - Rolls-Royce)**  
**Issue No: 059 Issue date: 11 August 2020**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge | Range  | Calibration and<br>Measurement<br>Capability (CMC)<br>Expressed as an<br>Expanded<br>Uncertainty ( $k = 2$ )   | Remarks | Location<br>Code |
|--|--|--|---------|------------------|
| DC RESISTANCE<br>Generation              |  |  |         | Solihull         |
| Specific Values                          | 10 $\Omega$<br>100 $\Omega$<br>1 k $\Omega$<br>10 k $\Omega$<br>100 k $\Omega$<br>1 M $\Omega$<br>10 M $\Omega$<br>100 M $\Omega$  | 35 ppm<br>15 ppm<br>15 ppm<br>15 ppm<br>15 ppm<br>18 ppm<br>80 ppm<br>180 ppm  |         |                  |
| Other Values                             | 0 $\Omega$ to 11 $\Omega$<br>11 $\Omega$ to 33 $\Omega$<br>33 $\Omega$ to 110 $\Omega$<br>110 $\Omega$ to 330 $\Omega$<br>330 $\Omega$ to 1.1 k $\Omega$<br>1.1 k $\Omega$ to 3.3 k $\Omega$<br>3.3 k $\Omega$ to 11 k $\Omega$<br>11 k $\Omega$ to 33 k $\Omega$<br>33 k $\Omega$ to 110 k $\Omega$<br>110 k $\Omega$ to 330 k $\Omega$<br>330 k $\Omega$ to 1.1 M $\Omega$<br>1.1 M $\Omega$ to 3.3 M $\Omega$<br>3.3 M $\Omega$ to 11 M $\Omega$<br>11 M $\Omega$ to 33 M $\Omega$<br>33 M $\Omega$ to 110 M $\Omega$<br>110 M $\Omega$ to 330 M $\Omega$ | 180 ppm + 11 m $\Omega$<br>150 ppm + 19 m $\Omega$<br>110 ppm + 19 m $\Omega$<br>110 ppm + 19 m $\Omega$<br>110 ppm + 90 m $\Omega$<br>110 ppm + 90 m $\Omega$<br>110 ppm + 900 m $\Omega$<br>110 ppm + 900 m $\Omega$<br>140 ppm + 9.0 $\Omega$<br>150 ppm + 9.0 $\Omega$<br>180 ppm + 80 $\Omega$<br>200 ppm + 80 $\Omega$<br>710 ppm + 800 $\Omega$<br>0.14 % + 800 $\Omega$<br>0.60 % + 8.0 k $\Omega$<br>0.60 % + 21 k $\Omega$ |         |                  |
| DC VOLTAGE<br>Generation                 | 0 mV to 200 mV<br>200 mV to 2 V<br>2 V to 20 V<br>20 V to 200 V<br>200 V to 1 kV   | 12 ppm + 1.0 $\mu$ V<br>7.5 ppm + 1.5 $\mu$ V<br>6.0 ppm + 5.0 $\mu$ V<br>8.0 ppm + 70 $\mu$ V<br>10 ppm + 700 $\mu$ V   |         |                  |
| DC CURRENT<br>Generation                 | 0 $\mu$ A to 220 $\mu$ A<br>220 $\mu$ A to 2.2 mA<br>2.2 mA to 22 mA<br>22 mA to 220 mA<br>220 mA to 2.2 A<br>2.2 A to 11 A  | 70 ppm + 10 nA<br>60 ppm + 12 nA<br>60 ppm + 120 nA<br>70 ppm + 1.2 $\mu$ A<br>100 ppm + 35 $\mu$ A<br>710 ppm + 510 $\mu$ A   |         |                  |



0394

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Trescal Limited (Trescal EMS - Rolls-Royce)**

**Issue No: 059 Issue date: 11 August 2020**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge     | Range  | Calibration and<br>Measurement<br>Capability (CMC)<br>Expressed as an<br>Expanded<br>Uncertainty ( $k = 2$ )  | Remarks | Location<br>Code |
|--|--|---|---------|------------------|
| <b>AC VOLTAGE<br/>Generation</b>             | <i>40 Hz to 10 kHz</i><br>0.22 mV to 2.2 mV<br>2.2 mV to 22 mV<br>22 mV to 220 mV<br>220 mV to 2.2 V<br>2.2 V to 22 V<br>22 V to 220 V<br><br><i>55 Hz to 1 kHz</i><br>220 V to 1 kV | 700 ppm + 6.0 $\mu$ V<br>230 ppm + 7.0 $\mu$ V<br>140 ppm + 10 $\mu$ V<br>100 ppm + 14 $\mu$ V<br>100 ppm + 130 $\mu$ V<br>110 ppm + 1.5 mV<br><br>120 ppm + 8.0 mV |         | Solihull         |
| <b>AC CURRENT<br/>Generation</b>             | <i>55 Hz to 1 kHz</i><br>10 $\mu$ A to 220 $\mu$ A<br>220 $\mu$ A to 2.2 mA<br>2.2 mA to 22 mA<br>22 mA to 220 mA<br>220 mA to 2.2 A   | 260 ppm + 20 nA<br>250 ppm + 55 nA<br>200 ppm + 550 nA<br>200 ppm + 5.5 $\mu$ A<br>800 ppm + 55 $\mu$ A   |         |                  |
| <b>MEASURING INSTRUMENTS AND MACHINES</b>    |  |   |         |                  |
| Micrometers                                  |  |   |         |                  |
| External<br>Depth                            | As BS 870:2008 and above<br>As BS 6468:2008  | Heads: 2.0 between<br>any two points<br>Setting and extension<br>rods:<br>1.0 + 5.0 x length in m   |         |                  |
| Vernier gauges<br>Caliper<br>Height<br>Depth | As BS 887:2008<br>As BS 1643:2008<br>As BS 6365:2008   | Overall performance:<br>10 + (30 x length in m)   |         |                  |
| Dial gauges and dial test<br>indicators      | As BS 907:2008 and BS<br>2795:1981   | 1.0   |         |                  |



0394  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Trescal Limited (Trescal EMS - Rolls-Royce)**  
**Issue No: 059 Issue date: 11 August 2020**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge   | Range   | Calibration and<br>Measurement<br>Capability (CMC)<br>Expressed as an<br>Expanded<br>Uncertainty ( $k = 2$ )   | Remarks  | Location<br>Code |   |
|--|---|--|--|------------------|---|
| <b>FORM</b>  |   |  |  | Site             |   |
| RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES<br>UNLESS OTHERWISE STATED   |   |  |  |                  |   |
| Surface plates<br>Granite<br>Cast iron<br><br><b>ELECTRICAL</b><br><br>Temperature indicators and simulators (thermocouple type),<br>calibration by electrical simulation: | As BS 817:2008<br>160 x 100 to 4 m x 4 m  | 1.5 + (0.80 x diagonal<br>in m)<br>See Note 2  | Internal Reference junction enabled.<br>Ambient temperature range 18 °C<br>to 22°C (controlled customer<br>environment). | Site             |   |
| Base metal<br>thermocouple types   | Type J, -210 °C to 0 °C<br>Type J, 0 °C to 1200 °C<br><br>Type K, -270 °C to -200 °C<br>Type K, -200 °C to 0 °C<br>Type K, 0 °C to 1000 °C<br>Type K, 1000 °C to 1370 °C<br><br>Type N, -270 °C to -200 °C<br>Type N, -200 °C to -100 °C<br>Type N, -100 °C to 0 °C<br>Type N, 0 °C to 800 °C<br>Type N, 800 °C to 1300 °C<br><br>Type T, -270 °C to -200 °C<br>Type T, -200 °C to 0 °C<br>Type T, 0 °C to 400 °C | 0.36 °C<br>0.28 °C<br><br>4.6 °C<br>0.37 °C<br>0.29 °C<br>0.27 °C<br><br>1.9 °C<br>0.49 °C<br>0.34 °C<br>0.26 °C<br>0.24 °C<br><br>0.81 °C<br>0.36 °C<br>0.26 °C |  |                  |   |
| Noble metal thermocouple<br>types  | Type R, -50 °C to 0 °C<br>Type R, 0 °C to 150 °C<br>Type R, 150 °C to 400 °C<br>Type R, 400 °C to 1768 °C<br><br>Type S, -50 °C to 0 °C<br>Type S, 0 °C to 100 °C<br>Type S, 100 °C to 300 °C<br>Type S, 300 °C to 1768 °C  | 0.91 °C<br>0.71 °C<br>0.51 °C<br>0.62 °C<br><br>0.80 °C<br>0.66 °C<br>0.55 °C<br>0.48 °C   |  |                  |   |
| RTD Pt100  | Up to 0 °C<br>Up to 0 °C<br><br>0°C to 850 °C<br>0°C to 850 °C  | 0.072 °C<br>0.042 % + 0.072 °C<br><br>0.029 % + 0.075 °C<br>0.051 % + 0.075 °C   |  |                  | Ambient temperature range<br>18 °C to 28 °C<br>-10 °C to +50 °C<br><br>18 °C to 28 °C<br>-10 °C to +50 °C |



0394  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Trescal Limited (Trescal EMS - Rolls-Royce)**  
**Issue No: 059 Issue date: 11 August 2020**

Calibration performed by the Organisation at the locations specified

| Measured Quantity<br>Instrument or Gauge | Range  | Calibration and<br>Measurement<br>Capability (CMC)<br>Expressed as an<br>Expanded<br>Uncertainty ( $k = 2$ ) | Remarks   | Location<br>Code |  |
|--|--|--|---|------------------|--|
| DC Voltage                               | 0 V to 150 mV<br>0 V to 150 mV                                   | 0.023 % + 5.0 $\mu$ V<br>0.048 % + 5.0 $\mu$ V   | Ambient temperature range<br>18 °C to 28 °C<br>-10 °C to +50 °C | Site             |  |
|  | 0.15 V to 0.25 V<br>0.15 V to 0.25 V                             | 0.023 % + 8.4 $\mu$ V<br>0.048 % + 8.4 $\mu$ V   | 18 °C to 28 °C<br>-10 °C to +50 °C                              |                  |  |
|  | 0.25 V to 1 V<br>0.25 V to 1 V                                   | 0.023 % + 12 $\mu$ V<br>0.048 % + 12 $\mu$ V   | 18 °C to 28 °C<br>-10 °C to +50 °C                              |                  |  |
|  | 1 V to 25 V<br>1 V to 25 V                                       | 0.023 % + 0.65 mV<br>0.048 % + 0.65 mV   | 18 °C to 28 °C<br>-10 °C to +50 °C                              |                  |  |
|  | 25 V to 60 V<br>25 V to 60 V                                     | 0.023 % + 1.2 mV<br>0.048 % + 1.2 mV   | 18 °C to 28 °C<br>-10 °C to +50 °C                              |                  |  |
| DC Current                               | 0 to 25 mA<br>0 to 25 mA   | 0.025 % + 1.7 $\mu$ A<br>0.049 % + 1.7 $\mu$ A   | 18 °C to 28 °C<br>-10 °C to +50 °C                              |                  |  |
|  | 25 mA to 100 mA<br>25 mA to 100 mA                               | 0.025 % + 2.0 $\mu$ A<br>0.049 % + 2.0 $\mu$ A   | 18 °C to 28 °C<br>-10 °C to +50 °C                              |                  |  |
| DC Resistance                            | 0 $\Omega$ to 250 $\Omega$<br>0 $\Omega$ to 250 $\Omega$         | 0.023 % + 4.3 m $\Omega$<br>0.048 % + 4.3 m $\Omega$   | 18 °C to 28 °C<br>-10 °C to +50 °C                              |                  |  |
|  | 250 $\Omega$ to 2650 $\Omega$<br>250 $\Omega$ to 2650 $\Omega$   | 0.023 % + 11 m $\Omega$<br>0.048 % + 11 m $\Omega$   | 18 °C to 28 °C<br>-10 °C to +50 °C                              |                  |  |
|  | 2650 $\Omega$ to 4000 $\Omega$<br>2650 $\Omega$ to 4000 $\Omega$ | 0.023 % + 100 m $\Omega$<br>0.048 % + 100 m $\Omega$   | 18 °C to 28 °C<br>-10 °C to +50 °C                              |                  |  |
| END                                      |  |  |   |                  |  |



0394

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Trescal Limited (Trescal EMS - Rolls-Royce)**

**Issue No: 059 Issue date: 11 August 2020**

Calibration performed by the Organisation at the locations specified

## Appendix - Calibration and Measurement Capabilities

### Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

### Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

- As a single value that is valid throughout the range.
- As an explicit function of the measurand or of a parameter (see below).
- As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.
- As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.
- In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

### Expression of CMCs - symbols and units

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples are shown below. It should be noted that these expressions are *not* mathematical formulae but are instead written in a commonly used shorthand for expressing uncertainties - therefore, for purposes of clarity, an indication of how they are to be interpreted is also provided below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0  $\mu$ V

Over the range 100 mV to 1 V, the CMC is 0.0025 %·V + 5.0  $\mu$ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 %· $p$  + (0.12·10<sup>-6</sup>· $p$ ·10<sup>-6</sup>) + 4.0 Pa, where  $p$  is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means 1.5 · 0.01 ·  $i$ , where  $i$  is the instrument indication.