

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017 & KS Q ISO/IEC 17025:2017

Institute of Calibration Technology Co.,Ltd

72, Dongtancheomdansaneop 2-ro, Hwaseong-si, Gyeonggi-do, Republic of Korea

Phone : 82-31-379-5114, Fax : 82-31-379-5115 e-mail : sictadmin@sict.co.kr

Expanding Scope of Accreditation(Ne	7 Items
Expanding Scope of Accreditation(Ra	75 items
Change Scope of Accreditation(Uncer	181 Items

CALIBRATION

Valid To : Oct. 29. 2025

Accreditation No : KC01-018(1/208)

In recognition of the successful completion of the KOLAS evaluation process,
accreditation is granted to this laboratory to perform the following calibrations

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site
102. Linear dimension			10236	Coating thickness testers	Y	10413	Straight rules	N
10201	Balls	N	10237	Torque arms	N	10415	Test bars	N
10203	Electrical /Mechanical comparators	Y	103. Angle			105. Complex geometry		
10206	Dial/cylinder gauge testers	N	10302	Angle gauge blocks	N	10501	Base gauges for electric bulb	N
10207	Doctor blades	N	10303	Autocollimators	N	10502	Baench centers	Y
10208	Distance meters;electrooptic/ laser/ultrasonic	N	10304	Bevel protractors	Y	10503	Contact coordinatemeasuring machines	Y
10209	End bars	N	10306	Clinometers	N	10504	Non-contact coordinate measuringmachines	Y
10210	Extensometers, lineardisplacementtransducers	Y	10308	Fine angle generators, level comparators	N	10505	Gauge block accessories	N
10211	Filler gauges	Y	10310	Indexing tables	N	10511	Measuringmicroscopes, Profileprojectors	Y
10212	Film applicators	N	10311	Plate/Square/Electriclevels	N	10512	Microscopes, micro measuring	Y
10213	Gapgauges	N	10314	Penta-prisms	N	10514	Taper plug gauges	N
10214	Gauge blocks, by comparison	N	10315	Polygons	N	10515	Taper ring gauges	N
10216	Height gauges/measuring machines	Y	10316	Rotary tables	Y	10517	Stylus type roughness testers	Y
10219	Linear scales	N	10317	Sinebars,Plates,Tables, Centers	N	10518	Socket gauges for electric bulb	N
10220	Measuring machines, standard	Y	10318	Squareness testers, Right angle testers	Y	10519	Roughness standard /comparison specimens	N
10221	Micro scales/Standard scales	N	10319	Cylindrical squares	N	10525	Thread plug gauges	N
10223	Electronic micrometers	Y	10320	Precisionsquares	N	10526	Taper thread plug gauges	N
10224	Heightmicrometers, Riserblocks	N	10322	Angular dislplacement transducers	N	10527	Thread ring gauges	N
10225	Laser scan micrometers	Y	104. Form			10528	Taper threading gauges	N
10227	Standardtaperules, Peripheralgauges	N	10401	Form testers	Y	10529	V-blocks,Boxblocks	N
10228	Cylindricalplug/pingauges, Threadmeasuringwiregauges	Y	10404	Optical flats	N	106. Various dimensional		
10229	Radius gauges	N	10405	Optical parallels	N	10601	Inside/Outside/Gear tooth calipers,Calipergauges	Y
10230	Cylindrical ring gauges	N	10406	Parallel blocks	Y	10603	Cylinder/bore gauges	Y
10232	Step gauges	N	10407	Precision surface plates	Y	10604	Depthgauges,Depthmicrometers	Y
10233	Thickness gauges, taper	N	10408	Profile gauges	N	10605	Dial/digital gauges	Y
10234	Ultrasonic thickness gauges	Y	10409	Roundness measurement instruments	Y	10608	Grind gauges	N
10235	Ultrasonic/coating thickness specimens	N	10411	Roundness standard/ Roundness magnification standard specimens	N	10609	Microindicators, Testindicators	Y
			10412	Straight edges	Y			

Accreditation No : KC01-018(2/208)

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site
10610	Micrometer heads	Y	20503	Ionization gauges	N	20916	Gas flowmeters;turbine	N
10611	3-points, Micrometers	Y				20917	Liquid flowmeters;turbine	N
10612	Inside micrometers	Y	20504	Thermal conductivity gauge; Pirani, thermocouple, convectron etc.	N	20918	Gas flowmeters;ultrasonic	N
10613	Outside micrometers	Y				20919	Liquid flowmeters;ultrasonic	N
10615	Particle counters	N				20920	Gas flowmeters;variable area	N
10617	Standard sieves	N	20505	Standard leaks, Helium leak dectors	Y	20921	Liquid flowmeters;variable area	N
10620	Welding gauges	N				20922	Gas flowmeters;vortex	N
201. Mass			206. Volume			20923	Liquid flowmeters;vortex	N
20102	Auto-hopper scale balances	Y	20601	Volumetric glasswares	N	20925	Anemometers; vane, etc	N
20103	Auto-packer scale balances	Y	20602	Pycnometers	N	20999	Others ; Anemometers; ultrasonic waves	N
20104	Axle weigher balances	N	20604	Standard volume vessels	Y			
20106	Dial platform scale balances	Y	20605	Concrete air content meters	N	210. Hardness		
20107	Dial swing scale balances	Y	20606	Piston type volume meters	N	21001	Brinell hardness testers	Y
20109	Electric balances	Y	207. Density			21002	Rockwell hardness testers	Y
20111	Manual swing scale balances	Y	20702	Liquid density meters	N	21003	Shore hardness testers	Y
20112	Platform scale balances	Y	20704	Salinity meters	N	21004	Vickers hardness testers	Y
20113	Spring scale balances	Y	20705	Sucrose meters	N	21005	Durometer hardness testers	N
20116	Weights	Y	20706	Hydrometers; density, specific gravity, alcohol, API, baume, sugar, milk, soil salinity, LPG, etc.	N	21006	Leeb hardness testers	Y
202. Force						207. Density		
20202	Force measuring devices	N	20707	Chloride meters	N	30102	Frequency standards	N
20203	Tension/compression testing machines	Y	20799	Others ; Solid density	N	30103	General frequency sources	Y
20204	push-pull gauge	Y				30104	Frequency meters/counters	Y
203. Torque			208. Viscosity			30105	Time interval sources	Y
20302	Torque measuring devices	Y	20801	Kinematic viscometers; capillary, etc.	N	30106	Time interval meters /Stop watches/Timers	Y
20303	Torque wrenches/drivers	Y				20802	Dynamic viscometers; rotaional, etc	Y
204. Pressure			209. Fluid flow			30201	Standard RPM generators	Y
20401	Altimeters	Y	20901	Anemometers; hot-wire	N	30202	Contact type tachometers	Y
20402	Manometers	Y	20902	Anemometers; pitot tube, etc.	N	30203	Photo tachometers /stroboscopes	Y
20403	Pneumatic pressure ballances	N	20908	Gas flowmeters;differential pressure	N			
20404	Hydraulic pressure ballances	N				20909	Liquid flowmeters;differenia pressure	N
20405	Air data test systems	Y	20910	Liquid flowmeters; electromagnetic	N	30206	Wow-flutter meters	Y
20406	Absolutepressuregauges	Y				401. DC Voltage & current		
20407	Blood pressure gauges	Y	20911	Gas flowmeters;thermal mass, etc.	N	40102	Transconductance amplifiers	Y
20408	Compound pressure gauges	Y				20912	Liquid flowmeters; Coriolis, etc.	N
20409	Differential pressure gauges	Y	20914	Gas flowmeters;open channel, etc.	N	40104	Electricaltemperature calibrators	Y
20411	Gauge pressure gauges	Y				20915	Liquid flowmeters; positive displacement	N
205. Vacuum			20915	Liquid flowmeters; positive displacement	N	40106	Galvanometers /null detectors	Y
20501	Capacitance diaphragm gauges	N				40107	Potentiometers	Y
20502	Spinning rotor gauges	N						

Accreditation No : KC01-018(3/208)

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site
40108	DC power supplies	Y	40316	Current transformer	N	405. Low frequency electric & magnetic fields		
40110	DC voltage dividers	N	40318	AC voltmeters	Y			
40111	DC voltage standards	N	40319	Watt hour meters	N	40503	Flux meters	Y
40112	DC voltmeters	Y	40321	Ratio transformers	N	40504	Flux sources	N
40113	Static/Ionic voltmeters	N	404. Other DC & LF Measurements			40508	Magnetometers	Y
402. Resistance, Capacitance and Inductance			40401	LF amplifiers	Y	40510	Reference/standard magnets	N
			40402	DC/LF attenuators	Y	406. Radio frequency measurements		
40201	Capacitance bridges /indicators	Y	40403	Multimeter calibrators	N	40601	RF amplifiers	Y
40202	Decade capacitors	Y	40404	Oscilloscope calibrators	N	40602	Coaxial attenuators	Y
40204	Standard capacitors	Y	40405	CD/DVD meters/analyzers	Y	40605	Burst pulse generators	Y
40205	Earth testers	Y	40406	Video signal generators	Y	40606	Attenuator calibrators	N
40206	Inductance bridges /indicators	Y	40407	Audio distortion analyzers /meters	Y	40607	RF power meter calibrators	Y
40208	Inductors	Y	40408	LF filters	Y	40608	EMC transducers ; current probes, absorbing absorbing clamps, etc.	Y
40210	Insulation testers	Y	40409	LF/Audiosignal analyzers	Y	40610	Coaxial directional couplers /splitters	Y
40211	Q-meters	Y	40410	Line frequency meters	Y	40613	Electrostatic discharge generators	Y
40213	Resistance bridges & similar instruments	Y	40411	Function generators	Y	40614	EMC receivers	Y
40214	Resistance meters	Y	40412	Genescopes	Y	40615	RF filters	Y
40215	Resistors	Y	40413	AC/DC high voltages volt meters	Y	40616	RF impedance meters	N
40217	Impedance bridges/LCR meters	Y	40415	Jitter meters	Y	40617	RF impulse generators	Y
403. AC voltage, current & power			40416	Leakage current testers	Y	40618	Line impedance stabilization networks ; LISN, CDN, ISN, etc.	Y
40301	AC ammeters	Y	40417	Electronic AC/DC loads	Y	40619	Coaxial standard mismatches	Y
40302	Clamp ammeters/voltmeters	Y	40418	Modulation meters	Y	40621	Mobile communication test sets	Y
40303	AC voltage/current calibrators	Y	40419	Analogue/Digital multimeters	Y	40622	Modulation meters	Y
40304	Wattmeter calibrators	N	40420	Noise meters	Y	40623	Network analyzers	Y
40305	AC current shunts	Y	40421	Oscilloscopes	Y	40624	Noise figure meters	Y
40306	Phase angle generators, synchro resolve generators	N	40422	LF phase meters	Y	40625	Noise generators	N
40307	Voltage/current phase angle meters/synchro resolve meters	N	40424	Volt/Current recorders	Y	40626	Noise impulse simulators	Y
40308	Potential transformer test sets	Y	40425	Relay test sets	Y	40627	RF phase noise meters	N
40309	Potential transformer	N	40426	LF signal generators	Y	40628	Coaxial noise sources	N
40310	Power factor meters	Y	40427	LF spectrum analyzers	Y	40635	RF power meters	Y
40311	AC power meters	Y	40428	Spot generators	Y	40636	Diode power sensors	Y
40312	AC power supplies	Y	40429	Sweep generators	Y	40637	Thermocouple power sensors	Y
40313	Puncture/safety testers	Y	40430	Signal transducers	Y	40638	Pulse generators	Y
40314	Power recorders	Y	40432	Transistor curve tracers	Y	40639	Radar test sets	Y
40315	Current transformer test sets	Y	40434	AC/DC high voltage generators	Y	40640	RF signal generators	Y
			40435	AC/DC high voltage probes	Y	40641	RF spectrum analyzers	Y
			40436	Logic analyzers	Y	40643	Surge generators	Y
			40437	Telephone testers	Y			
			40438	Video signal analyzers	Y			
			40438	Video signal analyzers	Y			

Accreditation No : KC01-018(4/208)

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site
40644	SWR meters	N	50303	Psychrometers; assmann ventilated, PRT type, etc.	N	70213	Display color analyzers; luminance, chromaticity, white balance, etc.	Y
40645	RF terminations	Y						
40646	Coaxial thermistor mounts	Y						
40650	RF voltmeters	Y	50304	Temperature humidity recorders ; Hygrothermograph, etc	N	70214	Luminous intensity standard lamps	N
40651	Vector voltmeters	Y						
40652	Field strength meters	Y						
40653	AM/FM test sources	Y	50305	Transducers; dew-point /relative humidity	N	70215	Spectral irradiance standard lamps	N
40654	Dip simulators	Y						
407. Field strength & antennas			50306	Humidity generators; two-pressure, two-temperature, flow mixing humidity gererator, constant temperature and humidity chamber, etc.	Y	70216	Total spectral radiand flux standard lampa	N
40704	Loop antennas	N						
40705	Monopole antennas	N						
501. Contact thermometry			50401	Cereal moisture meters	Y	70217	Luminance standard sources	N
50101	Temperature generators: ovens, furnaces, isothermal liquid baths, ice-point baths, dry-block calibrators	Y						
50102	Temperature indicators /recorders/controllers, temperature calibrators	Y	50402	Wood moisture meters	N	70218	Spectral radiance standard lamps	N
50103	Glass thermometers; liquid-in-glass, Beckmann	N	50403	Paper moisture meters	N	70219	UV irradiance meters	N
50104	Resistance thermometers; SPRT, IPRT, thermistors, etc.	Y	504. Moisture			70220	Spectral irradiance meters	Y
50105	Thermal expansion thermometers ; bimetal, gas or liquid type	Y	50401	Cereal moisture meters	Y	70221	Total spectral radiant flux meters	Y
50106	Thermomecouples: noble metal, base metal, pure metal, special type, etc.	Y	50402	Wood moisture meters	N	70222	Spectral radiance meters	Y
50107	Temperature transducers	Y	50403	Paper moisture meters	N	703. Properties of materials		
50108	Primary fixed-point cells and apparatus	N	601. Sound in air			70301	Colorimeters; material color	Y
502. non contact thermometry			60102	Sound calibrator	N	70304	Color standard tiles	N
50203	Optical pyrometers	N	60104	Microphones	N	70306	Gloss meters	Y
50204	Standard radiation	N	60106	Sound level meters	Y	70307	Gloss standard plates	Y
50205	Thermal image apparatus	N	603. Vibration			70308	Haze meters	Y
50206	Blackbody furnaces	Y	60301	Vibration calibrators	N	70312	Lens testers	Y
50207	Others ; ear thermometers, etc.	N	60302	Vibration transducers	N	70315	Optical densitometers	Y
503. Humidity			60303	Vibration measuring instruments	N	70319	Reflectance meters	Y
50301	Dew-point hygrometers; chilled mirror, alumina thinfilm, etc.	N	701. Photometry			70321	Refractometers	Y
50302	Relative humidity hygrometers polimer thinfilm, hair, etc.	Y	70101	Illuminance meters	N	70323	Transmittance meters	Y
702. Properties of detectos & sources			70102	Luminance meters	N	70325	Spectrophotometers including FT-IR spectrophotometers	Y
70202	Color temperature meters	Y	70103	Total luminous flux meters	Y	70326	Wavelength reference material; absorption cell, bandpass filter, etc.	N
70203	Color temperature standard lamps	N	70104	Luminous intensity meters	Y			
70204	Colorimeters; source color	Y						
70207	Laser power meters	N						
70208	Standard LED light sources	N						
70209	Total luminous flux standard lamps	N						
70210	Optical detectors	N						
70211	Pyranometers and pyrhemometers	N						

Accreditation No : KC01-018(5/208)

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site
704. Fiber optics			901. Chemical analysis					
70402	Broadband light sources	Y	90101	Breath alcohol analyzer	N			
70410	Optical attenuators	Y	90102	Environmental air monitoring instruments	Y			
70412	Fiber-optic power meters	Y						
70413	Optical loss testers	Y	90103	Gas analyzers	Y			
70415	Optical multimeters	Y	90104	Exhaust gas test instruments	Y			
70416	Optical network analyzers	Y	90199	Others ; pH meter,	Y			
70417	Optical spectrum analyzers	Y				Electrical conductivity meter		
70418	Optical time domain reflectometers, OTDR	Y						
70423	Return loss meters	Y						
70429	Frequency stabilized lasers and LDs	Y						
70430	ASE light sources	Y						
70433	Optical power stabilized lasers and LDs	Y						

Note

1. This laboratory provides calibration services in permanent standard laboratory and at on-site.
2. Laboratory conducts on-site calibration should meet requirements of KOLAS-SR-007.
3. On-site calibration is allowed to items with marking 'Y', not allowed to items with marking 'N'.
4. Measurement uncertainty normally is quoted as an expanded uncertainty at a coverage probability of 95%, which usually requires the use of a coverage factor of $k=2$. It expresses the lowest uncertainty of measurement that can be provided by accredited calibration laboratories in normal conditions.
5. Due to the calibration environment such as reference standards or customers' facilities, it is note that uncertainty of measurement on a calibration certificate may be expressed larger than measurement uncertainty on scope of accreditation in general.

102. Linear dimension

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Balls	10201	(0.3 ~ 100) mm	$\sqrt{0.38^2 + (0.0046 \times l_0)^2}$ μm	Mesuring Machine, Standard/ SICT-CP-10201
Electrical/Mechanical comparators	10203	(0 ~ 5) mm	0.14 μm	Gauge Block/ SICT-CP-10203
Dial/cylinder gauge testers	10206	(0 ~ 25) mm	$\sqrt{0.12^2 + (0.0028 \times l_0)^2}$ μm	Laser Measurement Machine/ SICT-CP-10206
Doctor Blades	10207	(0 ~ 10) mm	1.6 μm	Electronic Micrometer/ SICT-CP-10207
Distance meters; electrooptic/laser/ultrasonic	10208	(0 ~ 40) m	$\sqrt{0.28^2 + (0.005 \times l_0)^2}$ mm	Laser interferometer/ SICT-CP-10208
End bars	10209	(25 ~ 500) mm (500 ~ 1 000) mm	$\sqrt{0.74^2 + (0.003 \times l_0)^2}$ μm $\sqrt{0.80^2 + (0.003 \times l_0)^2}$ μm	Gauge Block/ SICT-CP-10209
Extensometers, lineardisplacementtransducers	10210	(0 ~ 500) mm (500 ~ 1 000) mm	$\sqrt{1.9^2 + (0.0042 \times l_0)^2}$ μm $\sqrt{7.8^2 + (0.058 \times l_0)^2}$ μm	Gauge Block, Laser Measurement Machine/ SICT-CP-10210
Filler gauges	10211	(0 ~ 10) mm	1.2 μm	Mesuring Machine, Standard/ SICT-CP-10211
Film Applicators	10212	(0 ~ 10) mm	1.6 μm	Electronic Micrometer/ SICT-CP-10212
Gap gauges	10213	(1 ~ 500) mm	$\sqrt{0.72^2 + (0.0048 \times l_0)^2}$ μm	Gauge Block, contact coordinate measuring machines/ SICT-CP-10213
Gauge blocks, by comparison	10214	(0.5 ~ 100) mm	$\sqrt{80^2 + (1.3 \times l_0)^2}$ nm	Gauge Block Comparator/ SICT-CP-10214
Height gauges/measuring machines	10216	(0 ~ 1 500) mm	$\sqrt{0.68^2 + (0.0035 \times l_0)^2}$ μm	Gauge Block, Step gauge/ SICT-CP-10216
	10219	(0 ~ 40) m	$\sqrt{0.03^2 + (0.0027 \times l_0)^2}$ mm	Laser interferometer/ SICT-CP-10219
Measuring machines, standard	10220	(0 ~ 500) mm	$\sqrt{0.38^2 + (0.002 \times l_0)^2}$ μm	Gauge Block/ SICT-CP-10220
Micro scales/Standard scales	10221			Non-contact coordinate measuring machines, standard scale/ SICT-CP-10221
Micro scales		(0 ~ 1) mm	0.4 μm	
Standard scales		(0 ~ 600) mm	$\sqrt{0.51^2 + (0.0021 \times l_0)^2}$ μm	
Electronic micrometers	10223	(0 ~ 5) mm	0.14 μm	Gage Block/ SICT-CP-10223
Height micrometers,Riser blocks				Gauge Block Electronic Micrometer/ SICT-CP-10224
Block	10224	(0 ~ 600) mm	$\sqrt{1.1^2 + (0.0019 \times l_0)^2}$ μm	
Head		(0 ~ 25) mm	1.2 μm	
Laser scan micrometers	10225	(0.5 ~ 85) mm	$\sqrt{0.46^2 + (0.003 \times l_0)^2}$ μm	Cylindrical plug/pin gauge/ SICT-CP-10225
Standard tape rules,Peripheral gauges	10227	(0 ~ 40) m (40 ~ 80) m (80 ~ 100) m	$\sqrt{0.22^2 + (0.0046 \times l_0)^2}$ mm $\sqrt{0.25^2 + (0.0046 \times l_0)^2}$ mm $\sqrt{0.34^2 + (0.0046 \times l_0)^2}$ mm	Laser Measurement Machine/ SICT-CP-10227
Cylindrical plug/pin gauges, Thread measuring wire gauges	10228			Mesuring Machine, Standard/ SICT-CP-10228
Cylindrical plug/pin gauges		(0.01 ~ 200) mm	$\sqrt{0.42^2 + (0.003 \times l_0)^2}$ μm	
Thread measuring wire gauges		(0.1 ~ 10) mm	0.41 μm	
Radius gauges	10229	(0.1 ~ 100) mm	2.8 μm	non-contact coordinate measuring machines, standard scale/ SICT-CP-10229
Cylindrical ring gauges	10230	(1.0 ~ 200) mm	$\sqrt{0.69^2 + (0.0034 \times l_0)^2}$ μm	Mesuring Machine, Standard/ SICT-CP-10230

Note 1. l_0 unit : mm (10208,10227 unit : m)

102. Linear dimension

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Step gauges	10232	(0 ~ 1 500) mm	$\sqrt{0.76^2 + (0.0025 \times l_0)^2}$ μm	Gauge Block, Step gauge/ SICT-CP-10232
Taper thickness gauges	10233	(0 ~ 60) mm	4.3 μm	Profile Projector/ SICT-CP-10223
Ultrasonic thickness gauges	10234	(2.5 ~ 100) mm	3 μm	Ultrasonic Tester Blocks/ SICT-CP-10234
Ultrasonic/coating thickness specimens	10235	(0.01 ~ 2) mm	0.7 μm	Gauge Block, Mesuring Machine, Standard/ SICT-CP-10235
coating thickness specimens		(0.5 ~ 500) mm	$\sqrt{0.64^2 + (0.006 \times l_0)^2}$ μm	
Ultrasonic thickness specimens				
Coating thickness testers	10236	(0 ~ 2) mm	1.1 μm	Thickness specimens/ SICT-CP-10236
Torque arms	10237	(1 ~ 1 500) mm	$\sqrt{0.60^2 + (0.0061 \times l_0)^2}$ μm	Gauge Block, contact coordinate measuring machines/ SICT-CP-10237
Torque arms		(0 ~ 5) mm	1.2 μm	
Wires				

Note 1. l_0 unit : mm (10208,10227 unit : m)

103. Angle

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Angle gauge blocks Angle	10302	(0 ~ 360)°	0.6"	Indexing tables/ SICT-CP-10302
Autocollimators Angle	10303	± 500"	0.6"	Fine angle generators/ SICT-CP-10303
Bevel protractors Angle Accuracy Accessory Angle	10304	(0 ~ 360)° (0 ~ 90)°	0.9' 0.7'	Angle Gauge Block, Coordinate Measur Machine/ SICT-CP-10304
Clinometers Angle	10306	(0 ~ 360)°	3.3"	Rotary tables/ SICT-CP-10306
Fine angle generators, level comparators Angle	10308	± 1 000"	0.4"	Autocollimators/ SICT-CP-10308
Indexing tables Angle	10310	(0 ~ 360)°	0.5"	Polygons/ SICT-CP-10310
Plate/Square/Electric levels Angle Inclino meter Squareness	10311	(0 ~ 516)" (516 ~ 1 000)" (0 ~ 90)° (0 ~ 400) mm	0.4" 1.2" 0.05' 1.8 μm	Fine angle generators, Rotary tables/ SICT-CP-10311
Penta-prisms Angle	10314	90°	0.6"	Autocollimators/ SICT-CP-10314
Polygons Angle	10315	(0 ~ 360)°	0.4"	Indexing tables/ SICT-CP-10315
Rotary tables Angle	10316	(0 ~ 360)°	0.5"	Polygons/ SICT-CP-10316
Sinebars, Plates, Tables, Centers (Sinebars) distance, between two roller center parallelism, between two roller parallelism, between flat-two roller (Plates) Center length Flatness Parallelism	10317	(100 ~ 300) mm (100 ~ 300) mm (100 ~ 300) mm (100 ~ 300) mm (100 ~ 300) mm (100 ~ 300) mm (100 ~ 300) mm	$\sqrt{0.36^2 + (0.002 \times L_0)^2}$ μm 0.5 μm 0.6 μm $\sqrt{0.12^2 + (0.028 \times L_0)^2}$ μm 1.0 μm 1.2 μm	Mesuring Machine, Standard/ SICT-CP-10317
Squareness testers, Right angle testers	10318	(0 ~ 600) mm	2.0 μm	Cylindrical Square, Precision Square/ SICT-CP-10318

103. Angle

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Cylindrical squares	10319	(0 ~ 300) mm (300 ~ 600) mm	1.6 μm 1.9 μm	Cylindrical Square/ SICT-CP-10319
Precision squares Squareness Parallelism	10320	(0 ~ 600) mm (0 ~ 600) mm	2.9 μm 2.0 μm	contact coordinate measuring machines/ SICT-CP-10320
Angular displacement transducers Angle	10322	(0 ~ 360)°	2.9"	Rotary tables/ SICT-CP-10322

Note 1. l_0 unit : mm

104. Form

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Form testers Height length Width Angle	10401	(0 ~ 50) mm (0 ~ 50) mm 15° ~ 45°	1.0 μm 1.4 μm 2'	Form Standard Specimens Gage Block, Angle Gage Block/ SICT-CP-10401
Optical flats Flatness	10404	∅(0 ~ 60) mm ∅(60 ~ 100) mm	0.05 μm 0.08 μm	Optical Flat/ SICT-CP-10404
Optical parallels Flatness Parallelism	10405	∅(0 ~ 30) mm ∅(0 ~ 30) mm	0.05 μm 0.07 μm	Optical Flat,Gauge block comparator/ SICT-CP-10405
Parallel blocks Parallelism Flatness Length Difference	10406	(0 ~ 1 000) mm (0 ~ 1 000) mm (0 ~ 1 000) mm	1.5 μm 1.5 μm 2.2 μm	Electronic Micrometer/ SICT-CP-10406
Precision surface plates Flatness	10407	(2 000 × 2 000) mm (5 000 × 5 000) mm	2.0 μm 4.8 μm	Electronic Level/ SICT-CP-10407
Profile gauges	10408	(0 ~ 5) mm	0.7 μm	Gage Block/ SICT-CP-10408
Roundness measurement instruments Detector accuracy Rotational accuracy of spindle Rotational accuracy of axis	10409	(0 ~ 20) μm 360° 360°	0.51 μm 0.03 μm 0.04 μm	Roundness Standard Ball/ SICT-CP-10409
Roundness standard/Roundness magnification standard specimens Standard specimens Standard ball	10411	(0 ~ 300) μm 360°	0.52 μm 0.08 μm	Roundness Tester/ SICT-CP-10411
Straight edges Straightness Parallelism	10412	(0 ~ 2 000) mm (0 ~ 2 000) mm	1.8 μm 1.8 μm	Electronic levels/ SICT-CP-10412
Straight rules Length	10413	(0 ~ 2 000) mm	0.10 mm	LASER INTERFEROMETER/ SICT-CP-10413
Test bars Roundness Cylindricity Run-out	10415	(0 ~ 400) mm (0 ~ 400) mm (0 ~ 400) mm	0.6 μm 0.6 μm 1.1 μm	Roundness Tester,Electronic Micrometer/ SICT-CP-10415

Note 1. l_0 unit : mm

105. Complex geometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Base gauges for electric bulb Inside diameter of pass/stop and screw Pitch	10501	(1 ~ 50) mm (0.3 ~ 10) mm	$\sqrt{0.47^2 + (0.0028 \times l_0)^2}$ μm 1.6 μm	Mesuring Machine, Standard/ SICT-CP-10501
Bench centers Difference of both center Flatness of both bed Center height difference	10502	(0 ~ 200) mm (200 ~ 500) mm (0 ~ 500) mm (0 ~ 200) mm (200 ~ 500) mm	1.8 μm 3.4 μm 1.5 μm 1.8 μm 3.4 μm	Test Bar/ SICT-CP-10502
Contact coordinate measuring machines	10503	(0 ~ 1 500) mm (0 ~ 600) mm (0 ~ 600) mm	$\sqrt{0.56^2 + (0.0044 \times l_0)^2}$ μm 3.2 μm 1.2 μm	Step Gauge/ SICT-CP-10503
Non-contact coordinate measuring machines	10504	(0 ~ 1 000) mm	$\sqrt{0.43^2 + (0.0034 \times l_0)^2}$ μm	Standard Scale/ SICT-CP-10504
Gauge block accessories Round the ministry of Justice A he ministry of Justice,Base block Center point Plane figure Parallelism(triangle edge) Parallelism(equilibrium tide)	10505	(0 ~ 50) mm (0 ~ 50) mm (0 ~ 20) mm (0 ~ 50) mm (0 ~ 300) mm (0 ~ 150) mm	$\sqrt{0.32^2 + (0.0044 \times l_0)^2}$ μm $\sqrt{0.26^2 + (0.0044 \times l_0)^2}$ μm 1.1 μm 0.04 μm 0.4 μm 0.4 μm	Gauge Block/ SICT-CP-10505
Measuring microscopes,Profile projectors Length Magnification Angle	10511	(0 ~ 500) mm (5 ~ 100) 배 (0 ~ 360) °	$\sqrt{0.43^2 + (0.0034 \times l_0)^2}$ μm 0.04 % 0.9'	Standard Scale/ SICT-CP-10511
Microscopes, micro measuring	10512	(0 ~ 1) mm (1 ~ 50) mm	0.7 μm 3.0 μm	Standard Scale/ SICT-CP-10512
Taper plug gauges Small end diameter Big end diameter Plane angle Gage height	10514	(2 ~ 200) mm (2 ~ 200) mm (0 ~ 90) ° (2 ~ 200) mm	$\sqrt{1.3^2 + (0.0041 \times l_0)^2}$ μm $\sqrt{1.4^2 + (0.0041 \times l_0)^2}$ μm 5.9" $\sqrt{1.2^2 + (0.0044 \times l_0)^2}$ μm	Mesuring Machine, Standard/ SICT-CP-10514
Taper ring gauges Small end diameter Big end diameter Plane angle	10515	(5 ~ 200) mm (5 ~ 200) mm (0 ~ 90) °	2.5 μm 2.5 μm 0.006°	contact coordinate measuring machines/ SICT-CP-10515
Stylus type roughness testers Roughness parameter(Ra) Roughness parameter(Rz) Mean width(RSm) H,D	10517	(0 ~ 2) μm (2 ~ 10) μm (0 ~ 7) μm (7 ~ 30) μm (0 ~ 300) μm (0 ~ 6) μm (6 ~ 20) μm	9 nm 24 nm 77 nm 0.27 μm 1.3 μm 63 nm 97 nm	Roughness Specimen/ SICT-CP-10517
Socket gauges for electric bulb Outside diameter of pass/stop and screw Pitch	10518	(1 ~ 50) mm (0.3 ~ 10) mm	$\sqrt{0.44^2 + (0.0028 \times l_0)^2}$ μm 1.6 μm	Mesuring Machine, Standard/ SICT-CP-10518

Note 1. l_0 unit : mm

105. Complex geometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Roughness standard/comparison specimens Depth(H) Mean width(RSm) Roughness parameter(Ra) Roughness parameter(Rz)	10519	(0 ~ 6) μm (6 ~ 20) μm (0 ~ 300) μm (0 ~ 2) μm (2 ~ 10) μm (0 ~ 7) μm (7 ~ 30) μm	$\sqrt{(9.6 \times R)^2 + 12^2}$ nm $\sqrt{(8.94 \times R)^2 + 15^2}$ nm $\sqrt{(0.01 \times R)^2 + 0.51^2}$ μm $\sqrt{(10 \times R)^2 + 2.6^2}$ nm $\sqrt{(9.2 \times R)^2 + 7.4^2}$ nm $\sqrt{(29.2 \times R)^2 + 7.4^2}$ nm $\sqrt{(0.025 \times R)^2 + 0.096^2}$ μm	Roughness Tester/ SICT-CP-10519
Thread plug gauges Outside diameter Effective diameter Pitch Half angle	10525	(1 ~ 205) mm (1 ~ 210) mm (0.3 ~ 10) mm (0.5 ~ 45) °	1.7 μm 1.1 μm 1.2 μm 2'	Measuring Machine, Standard/ SICT-CP-10525
Taper thread plug gauges Half angle Pitch Gage length Notch length Small outside diameter Big outside diameter Small effective diameter Big effective diameter	10526	(0 ~ 45) ° (0.3 ~ 6) mm (2 ~ 50) mm (2 ~ 50) mm (2 ~ 200) mm (2 ~ 200) mm (2 ~ 200) mm (2 ~ 200) mm	2' 1.3 μm 2.6 μm 3.6 μm 2.3 μm 4.8 μm 2.9 μm 5.1 μm	Measuring Machine, Standard/ SICT-CP-10526
Thread ring gauges Minor diameter Effective diameter Pitch	10527	(3 ~ 200) mm (3 ~ 200) mm (0.3 ~ 10) mm	1.5 μm 2.3 μm 1.6 μm	Measuring Machine, Standard/ SICT-CP-10527
Taper thread ring gauges Alternation Thickness Notch length	10528	±3 mm (0 ~ 100) mm (0 ~ 100) mm	2.7 μm 2.3 μm 3.2 μm	Measuring Machine, Standard/ SICT-CP-10528
V-blocks, Boxblocks Plane figure Parallelism Difference of both part	10529	(5 ~ 300) mm (5 ~ 300) mm (5 ~ 300) mm	1.7 μm 2.0 μm 2.8 μm	contact coordinate measuring machines/ SICT-CP-10529

106. Various dimensional

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Inside/Outside/Gear tooth calipers, Caliper gauges Inside/Outside calipers Caliper gauges	10601	(0 ~ 2 000) mm (0 ~ 300) mm	$\sqrt{8.2^2 + (0.007 \times l_0)^2}$ μm $\sqrt{3.7^2 + (0.003 \times l_0)^2}$ μm	Gauge Block/ SICT-CP-10601
Cylinder/bore gauges Cylinder gauges Hole gauges	10603	(0 ~ 1 000) mm (0.1 ~ 25) mm	1.0 μm 3.3 μm	Dial Gauge Tester/ SICT-CP-10603
Depth gauges, Depth micrometers Depth micrometers Depth gauges	10604	(0 ~ 300) mm (0 ~ 1 000) mm	$\sqrt{0.86^2 + (0.003 \times l_0)^2}$ μm $\sqrt{5.9^2 + (0.004 \times l_0)^2}$ μm	Gauge Block/ SICT-CP-10604
Dial/digital gauges	10605	(0 ~ 100) mm (0 ~ 25) mm	$\sqrt{0.21^2 + (0.008 \times l_0)^2}$ μm $\sqrt{0.59^2 + (0.004 \times l_0)^2}$ μm	Dial Gauge Tester/
Grind gauges Depth Straightness	10608	(0 ~ 1) mm (0 ~ 150) mm	1.8 μm 2.5 μm	Electronic micrometer/ SICT-CP-10608
Micro indicators, Test indicators	10609	(0 ~ 5) mm	0.22 μm	Dial Gauge Tester/ SICT-CP-10609
Micrometer heads	10610	(0 ~ 50) mm	0.8 μm	Gauge Block/ SICT-CP-10610
3-points, Micrometers	10611	(2 ~ 200) mm (200 ~ 300) mm	$\sqrt{1.3^2 + (0.003 \times l_0)^2}$ μm 3 μm	Ring Gauge/ SICT-CP-10611
Inside micrometers Length Accuracy of scale Extension rod	10612	(5 ~ 300) mm (25 ~ 500) mm (13 ~ 500) mm	$\sqrt{1.1^2 + (0.004 \times l_0)^2}$ μm $\sqrt{1.1^2 + (0.004 \times l_0)^2}$ μm $\sqrt{1.2^2 + (0.004 \times l_0)^2}$ μm	Gauge Block/ SICT-CP-10612
Outside micrometers Outside micrometers V-anvil micrometers	10613	(0 ~ 25) mm (25 ~ 1 000) mm (1 ~ 85) mm	$\sqrt{0.2^2 + (0.003 \times l_0)^2}$ μm $\sqrt{0.83^2 + (0.003 \times l_0)^2}$ μm 0.8 μm	Gauge Block, cylindrical plug gauges/ SICT-CP-10613
Particle counters (Air) Flow Threshold voltage Counting efficiency (Liquid) Flow Threshold voltage	10615	(0.1 ~ 1) μm (0 ~ 100) L/min (0 ~ 10) V (0 ~ 110) % (0.05 ~ 25) μm (0 ~ 100) mL/min (0 ~ 10) V	0.09 L/min 0.42 mV 4.1 % 1.4 mL/min 0.42 mV	Particle calibration system/ SICT-CP-10615
Standard sieves Sieve opening Wire rod diameter	10617	(0.004 ~ 10) mm (0.004 ~ 130) mm	1.5 μm 2.4 μm	Non-contact coordinate measuring machines/ SICT-CP-10617
Welding gauges Height or depth Rule Angle	10620	(0 ~ 100) mm (0 ~ 100) mm (0 ~ 90)°	8.2 μm 6.0 μm 0.7'	Non-contact coordinate measuring machine, Gauge Block/ SICT-CP-10620

Note 1. l_0 unit : mm

201. Mass

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Auto-hopper scale balances	20102	(0 ~ 10) kg (10 ~ 50) kg (50 ~ 200) kg (200 ~ 500) kg (500 ~ 1 000) kg	1.2 g 14 g 21 g 49 g 0.10 kg	Hopper Scale Weight/ SICT-CP-20102
Auto-packer scale balances	20103	(0 ~ 10) kg (10 ~ 50) kg (50 ~ 200) kg	0.8 g 7.7 g 16 g	Weight/ SICT-CP-20103
Axle weigher balances Portable	20104	(100 ~ 500) kg (500 ~ 1 000) kg (1 000 ~ 2 000) kg (2 000 ~ 5 000) kg (5 000 ~ 10 000) kg (10 000 ~ 30 000) kg	0.2 kg 0.4 kg 1 kg 5 kg 10 kg 24 kg	Force Calibration Machine/ SICT-CP-20104
Dial platform scale balances	20106	(0 ~ 30) kg (30 ~ 60) kg (60 ~ 100) kg	42 g 0.08 kg 0.21 kg	Weight/ SICT-CP-20106
Dial swing scale balances	20107	(0 ~ 1) kg (1 ~ 10) kg (10 ~ 20) kg (20 ~ 50) kg (50 ~ 100) kg (100 ~ 200) kg (200 ~ 500) kg (500 ~ 1 000) kg (1 000 ~ 2 000) kg (2 000 ~ 5 000) kg	0.96 g 9.6 g 20 g 48 g 96 g 0.23 kg 0.48 kg 0.96 kg 1.9 kg 4.6 kg	Weight/ SICT-CP-20107
Electric balances	20109	(0 ~ 2) g (2 ~ 6) g (6 ~ 20) g (20 ~ 50) g (50 ~ 100) g (100 ~ 200) g (200 ~ 500) g (500 ~ 1 000) g (1 ~ 2) kg (2 ~ 5) kg (5 ~ 10) kg (10 ~ 25) kg (25 ~ 40) kg (40 ~ 60) kg (60 ~ 150) kg (150 ~ 600) kg (600 ~ 1 000) kg (1 000 ~ 2 000) kg (2 000 ~ 5 000) kg	7.0 µg 9.3 µg 14 µg 19 µg 29 µg 0.05 mg 0.10 mg 0.20 mg 0.5 mg 1.0 mg 3 mg 6 mg 16 mg 24 mg 0.30 g 1.2 g 2.0 g 38 g 0.10 kg	Weight/ SICT-CP-20109

201. Mass

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Manual swing scale balances	20111	(0 ~ 50) kg (50 ~ 200) kg (200 ~ 500) kg (500 ~ 1 000) kg	38 g 0.19 kg 0.38 kg 0.94 kg	Weight/ SICT-CP-20111
Platform scale balances	20112	(0 ~ 200) kg (200 ~ 500) kg (500 ~ 1 000) kg	46 g 0.09 kg 0.46 kg	Weight/ SICT-CP-20112
Spring scale balances	20113	(0 ~ 10) kg (10 ~ 50) kg (50 ~ 100) kg	21 g 0.08 kg 0.21 kg	Weight/ SICT-CP-20113
Weights less than class E2	20116	(1 mg ~ 20 kg) 1 mg 2 mg 5 mg 10 mg 20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg	1.0 µg 1.0 µg 1.0 µg 1.0 µg 1.2 µg 1.4 µg 1.7 µg 2.3 µg 2.9 µg 3.5 µg 4.6 µg 5.8 µg 7.0 µg 9.3 µg 12 µg 18 µg 36 µg 95 µg 0.12 mg 0.36 mg 0.94 mg 1.8 mg 3.7 mg	Weights, Mass Comparator/ SICT-CP-20116
less than class F2		(20 ~ 100) kg 50 kg 100 kg	0.12 g 0.21 g	
less than class M1		(100 ~ 200) kg 200 kg	1.0 g	
less than class F2		(200 ~ 1 000) kg 500 kg 1 000 kg	1.3 g 2.1 g	

202. Force

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Force measuring devices Force	20202	(0.4 ~ 20) N (20 ~ 50) N (50 ~ 100) N (100 ~ 200) N (200 ~ 500) N (0.5 ~ 1) kN (1 ~ 2) kN (2 ~ 5) kN (5 ~ 10) kN (10 ~ 20) kN (20 ~ 50) kN (50 ~ 100) kN (100 ~ 200) kN (200 ~ 500) kN (500 ~ 1 000) kN	6.0×10^{-4} 8.4×10^{-5} 8.5×10^{-5} 7.3×10^{-5} 6.9×10^{-5} 9.2×10^{-5} 9.0×10^{-5} 8.4×10^{-5} 8.7×10^{-5} 3.6×10^{-4} 4.1×10^{-4} 4.8×10^{-4} 4.5×10^{-4} 4.3×10^{-4} 4.5×10^{-4}	Force Calibration Machine/ SICT-CP-20202
Tension/compression testing machines tensile compression	20203	0.1 N ~ 2 kN (2 ~ 5) kN (5 ~ 10) kN (10 ~ 20) kN (20 ~ 50) kN (50 ~ 100) kN (100 ~ 300) kN (0.1 ~ 50) N (50 ~ 100) N (100 ~ 200) N (200 ~ 500) N (0.5 ~ 1) kN (1 ~ 2) kN (2 ~ 5) kN (5 ~ 10) kN (10 ~ 20) kN (20 ~ 50) kN (50 ~ 100) kN (100 ~ 300) kN (300 ~ 500) kN (500 ~ 1 000) kN (1 000 ~ 3 000) kN	1.2×10^{-3} 1.3×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.4×10^{-3} 1.2×10^{-3} 1.5×10^{-3} 1.3×10^{-3} 1.4×10^{-3} 1.3×10^{-3} 1.4×10^{-3} 1.3×10^{-3} 1.4×10^{-3} 1.3×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.4×10^{-3} 1.6×10^{-3}	Weights, Force Measuring Device/ SICT-CP-20203
push-pull gauge Force	20204	(0.02 ~ 0.2) N 0.2 N ~ 2 kN (2 ~ 5) kN	1.4×10^{-2} 1.3×10^{-3} 8.4×10^{-4}	Weights, Force Calibration Machine/ SICT-CP-20204

203. Torque

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Torque measuring devices Torque	20302	(0.001 ~ 1) N·m (1 ~ 10) N·m (10 ~ 20) N·m (20 ~ 50) N·m (50 ~ 100) N·m (100 ~ 200) N·m (200 ~ 500) N·m (500 ~ 1 000) N·m (1 000 ~ 2 000) N·m	3.3×10^{-3} 4.1×10^{-4} 7.5×10^{-4} 2.3×10^{-4} 3.4×10^{-4} 2.1×10^{-4} 2.2×10^{-4} 1.5×10^{-4} 1.6×10^{-4}	Torque Calibration Machine/ SICT-CP-20302
Torque wrenches/drivers Torque	20303	(0.02 ~ 0.1) N·m (0.1 ~ 0.5) N·m (0.5 ~ 1) N·m (1 ~ 2) N·m (2 ~ 5) N·m (5 ~ 10) N·m (10 ~ 20) N·m (20 ~ 50) N·m (50 ~ 100) N·m (100 ~ 200) N·m (200 ~ 500) N·m (500 ~ 1 000) N·m (1 000 ~ 2 000) N·m	1.4×10^{-2} 9.5×10^{-3} 7.8×10^{-3} 6.2×10^{-3} 4.6×10^{-3} 4.5×10^{-3} 4.7×10^{-3} 4.5×10^{-3} 4.9×10^{-3} 3.8×10^{-3} 3.7×10^{-3} 3.8×10^{-3} 2.8×10^{-3}	Torque Measuring Device/ SICT-CP-20303

204. Pressure

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Altimeters	20401	(-4 000 ~ 20 000) m (20 000 ~ 47 000) m	12 m 15 m	Digital Manometer, Air Dead Weight Tester/ SICT-CP-20401
Manometers	20402	(0 ~ 200) kPa	2.9×10^{-3}	Digital Manometer, Air Dead Weight Tester/ SICT-CP-20402
Pneumatic pressure ballances	20403	(5 ~ 7 000) kPa	6.1×10^{-5}	분동식 압력계/ SICT-CP-20403
Hydraulic pressure ballances	20404	(2 ~ 200) MPa	1.2×10^{-4}	분동식 압력계/ SICT-CP-20404
Air data test systems Static pressure Dynamic pressure	20405	(-2 500 ~ 20 000) m (20 000 ~ 30 500) m (0 ~ 342) km/hr (342 ~ 2 122) km/hr	0.8 m 7 m 0.1 km/hr 0.3 km/hr	Digital Manometer, Air Dead Weight Tester/ SICT-CP-20405
Absolute pressure gauges Dial, digital	20406	(5 ~ 7 000) kPa abs (7 ~ 200) MPa abs	8.4×10^{-5} 1.2×10^{-4}	Digital Manometer, Air Dead Weight Tester/ SICT-CP-20406
Blood pressure gauges	20407	(0 ~ 40) kPa	0.01 kPa	Digital Manometer/ SICT-CP-20407
Compound pressure gauges	20408	(-95 ~ 7 000) kPa	7.9×10^{-5}	Air Dead Weight Tester/ SICT-CP-20408
Differential pressure gauges	20409	(0 ~ 500) Pa (500 ~ 5 000) Pa (5 ~ 7 000) kPa	0.13 Pa 1.5 Pa 2.2×10^{-3}	Digital Manometer, Air Dead Weight Tester/ SICT-CP-20409
Gauge pressure gauges	20411	(0 ~ 500) Pa (500 ~ 5 000) Pa (5 ~ 7 000) kPa (7 ~ 200) MPa (200 ~ 500) MPa	0.13 Pa 1.3 Pa 7.9×10^{-5} 1.2×10^{-4} 3.5×10^{-4}	Digital Manometer, Air Dead Weight Tester, Oil Dead Weight Tester/ SICT-CP-20411
Pressure transducers/transmitters Absolute pressure Gauge pressure	20412	5 kPa abs ~ 200 MPa abs (0 ~ 500) Pa (500 ~ 5 000) Pa 5 kPa ~ 500 MPa	6.7×10^{-4} 0.13 Pa 1.2 Pa 6.7×10^{-4}	Digital Manometer, Air Dead Weight Tester, Oil Dead Weight Tester/ SICT-CP-20412
Dial type vacuum gauges	20413	(-95 ~ 0) kPa	0.059 kPa	Air Dead Weight Tester, SICT-CP-20413

205. Vacuum

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Capacitance diaphragm gauges Vacuum	20501	(0.133 ~ 13.3) Pa abs (13.3 ~ 133.3) Pa abs (133.3 ~ 1 333) Pa abs (1.333 ~ 13.332) kPa abs (13.332 ~ 133.322) kPa abs	0.051 Pa abs 0.49 Pa abs 1.5 Pa abs 12 Pa abs 17 Pa abs	Baratron gauge, SRG / SICT-CP-20501
Spinning rotor gauges Vacuum	20502	0.15 mPa abs ~ 0.01 Pa abs	3.4×10^{-2}	Baratron gauge, SRG / SICT-CP-20502
Ionization gauges Vacuum	20503	0.093 mPa abs ~ 0.15 mPa abs 0.15 mPa abs ~ 0.01 Pa abs	6.0×10^{-2} 3.5×10^{-2}	Baratron gauge, SRG, Ion / SICT-CP-20503
Thermal conductivity gauges; pirani, thermocouple, convector etc. Vacuum	20504	(0.133 ~ 13.3) Pa abs (13.3 ~ 133.3) Pa abs (133.3 ~ 1 333) Pa abs (1.333 ~ 13.332) kPa abs (13.332 ~ 133.322) kPa abs	0.051 Pa abs 0.49 Pa abs 1.5 Pa abs 13 Pa abs 17 Pa abs	Baratron gauge, SRG / SICT-CP-20504
Standard leaks, Helium leak detectors Leak	20505	22.0 $\mu\text{Pa m}^3/\text{s}$ 1.60 $\mu\text{Pa m}^3/\text{s}$ 0.51 $\mu\text{Pa m}^3/\text{s}$ 15.0 nPa m^3/s 6.4 nPa m^3/s 0.24 nPa m^3/s	4.8 $\mu\text{Pa m}^3/\text{s}$ 0.38 $\mu\text{Pa m}^3/\text{s}$ 0.098 $\mu\text{Pa m}^3/\text{s}$ 3.2 nPa m^3/s 1.3 nPa m^3/s 0.049 nPa m^3/s	Standard leaks, Helium leak detectors / SICT-CP-20505

206. Volume

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Volumetric glasswares	20601	(0 ~ 0.1) ml (0.1 ~ 0.5) ml (0.5 ~ 1) ml (1 ~ 2) ml (2 ~ 5) ml (2 ~ 10) ml (10 ~ 25) ml (25 ~ 50) ml (50 ~ 100) ml (100 ~ 200) ml (200 ~ 250) ml (250 ~ 500) ml (500 ~ 1 000) ml (1 000 ~ 2 000) ml (2 000 ~ 5 000) ml (5 000 ~ 10 000) ml	0.31 μl 0.40 μl 0.43 μl 1.1 μl 1.7 μl 2.5 μl 4.3 μl 5.1 μl 7.9 μl 13 μl 43 μl 69 μl 92 μl 0.17 ml 0.49 ml 0.87 ml	Weight, balances / SICT-CP-20601
Pycnometers	20602	(0 ~ 50) ml (50 ~ 100) ml (100 ~ 250) ml (250 ~ 500) ml	2.4 μl 4.4 μl 10 μl 21 μl	Weight, balances / SICT-CP-20602
Standard volume vessels	20604	(0 ~ 20) L (20 ~ 200) L (200 ~ 10 000) L	9.0×10^{-5} 1.3×10^{-4} 1.1×10^{-3}	Balances, Master Meter, Standard volume vessel/ SICT-CP-20604
Concrete air content meters	20605	(0 ~ 10) %	0.032 %	Weight, balances / SICT-CP-20605
Piston type volume meters	20606	(0 ~ 1) μl (1 ~ 2) μl (2 ~ 5) μl (5 ~ 10) μl (0.01 ~ 0.02) ml (0.02 ~ 0.05) ml (0.05 ~ 0.1) ml (0.1 ~ 0.2) ml (0.2 ~ 0.5) ml (0.5 ~ 1) ml (1 ~ 2) ml (2 ~ 5) ml (5 ~ 10) ml (10 ~ 20) ml (20 ~ 50) ml (50 ~ 100) ml	0.004 μl 0.005 μl 0.006 μl 0.008 μl 0.018 μl 0.035 μl 0.047 μl 0.14 μl 0.35 μl 0.65 μl 1.6 μl 1.9 μl 2.4 μl 5.0 μl 14 μl 64 μl	Weight, balances / SICT-CP-20606

207. Density

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Liquid density meters	20702	(0.7 ~ 1.4) g/cm ³	0.000 078 g/cm ³	Density standard materials/ SICT-CP-20702
Salinity meters	20704	(0 ~ 10) % (10 ~ 26) %	0.004 % 0.007 %	NaCl/ SICT-CP-20704
Sucrose meters	20705	(0 ~ 20) % (20 ~ 60) % (60 ~ 80) %	0.014 % 0.016 % 0.025 %	Sucrose/ SICT-CP-20705
Hydrometers: density, specific gravity, alcohol, API, baume, sugar, milk, soil, salinity, LPG, etc.	20706			
density		(0.600 ~ 0.700) g/cm ³ (0.700 ~ 0.800) g/cm ³ (0.800 ~ 0.900) g/cm ³ (0.900 ~ 1.000) g/cm ³ (1.000 ~ 1.100) g/cm ³ (1.100 ~ 1.200) g/cm ³ (1.200 ~ 1.300) g/cm ³ (1.300 ~ 1.400) g/cm ³ (1.400 ~ 1.500) g/cm ³ (1.500 ~ 1.600) g/cm ³ (1.600 ~ 1.700) g/cm ³ (1.700 ~ 1.800) g/cm ³ (1.800 ~ 1.900) g/cm ³ (1.900 ~ 2.000) g/cm ³ (2.000 ~ 2.200) g/cm ³ (2.200 ~ 3.000) g/cm ³ (3.000 ~ 3.600) g/cm ³ (3.600 ~ 4.000) g/cm ³	0.000 035 g/cm ³ 0.000 038 g/cm ³ 0.000 042 g/cm ³ 0.000 046 g/cm ³ 0.000 050 g/cm ³ 0.000 055 g/cm ³ 0.000 059 g/cm ³ 0.000 066 g/cm ³ 0.000 071 g/cm ³ 0.000 075 g/cm ³ 0.000 079 g/cm ³ 0.000 084 g/cm ³ 0.000 088 g/cm ³ 0.000 093 g/cm ³ 0.000 25 g/cm ³ 0.000 28 g/cm ³ 0.000 30 g/cm ³ 0.000 32 g/cm ³	Solid density standard material, Hydrostatic weighing Apparatus/ SICT-CP-20706-1
specific gravity		0.590 ~ 0.700 0.700 ~ 0.800 0.800 ~ 0.900 0.900 ~ 1.000 1.000 ~ 1.100 1.100 ~ 1.200 1.200 ~ 1.300 1.300 ~ 1.400 1.400 ~ 1.500 1.500 ~ 1.600 1.600 ~ 1.800 1.800 ~ 2.000 2.000 ~ 2.020 2.020 ~ 2.500 2.500 ~ 3.000	0.000 068 0.000 069 0.000 072 0.000 075 0.000 078 0.000 082 0.000 086 0.000 091 0.000 096 0.000 10 0.000 11 0.000 12 0.000 26 0.000 60 0.000 61	Solid density standard material, Hydrostatic weighing Apparatus/ SICT-CP-20706-2

* 20704, 20705, 20706, 20707 unit % is weight percent.

207. Density

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Hydrometers; density, specific gravity, alcohol, API, baume, sugar, milk, soil, salinity, LPG, etc.	20706	alcohol (Volumn)	(0 ~ 10) %	0.039 %	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-3
			(10 ~ 30) %	0.043 %	
			(30 ~ 40) %	0.038 %	
			(40 ~ 50) %	0.030 %	
			(50 ~ 60) %	0.025 %	
			(60 ~ 70) %	0.023 %	
			(70 ~ 80) %	0.020 %	
			(80 ~ 90) %	0.019 %	
			(90 ~ 100) %	0.017 %	
		API	-1 ~ 51	0,013	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-4
			51 ~ 101	0,014	
		Baumé-light	10 ~ 30	0,015	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-5
			30 ~ 40	0,016	
			40 ~ 60	0,018	
			60 ~ 70	0,019	
			70 ~ 100	0,12	
		Baumé - heavy	0 ~ 40	0,014	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-5
			40 ~ 75	0,013	
		sugar	(0 ~ 10) %	0.018 %	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-6
			(10 ~ 90) %	0.017 %	
		milk	(15 ~ 20)	0,081	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-7
			(20 ~ 40)	0,082	
		Bouyoucos	(-5.0 ~ 60.0) g/L	0.14 g/L	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-8
		salinity	(0 ~ 26.4) %	0.025 %	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-9
		LPG	(0.50 ~ 0.55) g/cm ³	0.000 065 g/cm ³	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-10
			(0.55 ~ 0.60) g/cm ³	0.000 066 g/cm ³	
			(0.60 ~ 0.65) g/cm ³	0.000 068 g/cm ³	

* 20704, 20705, 20706, 20707 unit % is weight percent.

207. Density

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Hydrometers; density, specific gravity, alcohol, API, baume, sugar, milk, soil, salinity, LPG, etc. Twaddell	20706	0 ~ 12 12 ~ 74 74 ~ 102 102 ~ 170 170 ~ 200	0,016 0,059 0,060 0,061 0,062	Solid density standard material, Hydrostatic weighing Apparatus/ SICT-CP-20706-11
Chloride meters	20707	(0 ~ 0.1) % (0.1 ~ 2.0) %	0.000 2 % 0.001 0 %	Chlorine standard liquid/ SICT-CP-20707
Solid density Stainless steel Glass	20799	(1 ~ 2) g (2 ~ 5) g (5 ~ 10) g (10 ~ 20) g (20 ~ 50) g (50 ~ 100) g (100 ~ 500) g (1 ~ 2) g (2 ~ 5) g (5 ~ 10) g (10 ~ 20) g (20 ~ 500) g	0.004 6 g/cm ³ 0.002 4 g/cm ³ 0.001 0 g/cm ³ 0.000 59 g/cm ³ 0.000 43 g/cm ³ 0.000 37 g/cm ³ 0.000 36 g/cm ³ 0.000 46 g/cm ³ 0.000 25 g/cm ³ 0.000 14 g/cm ³ 0.000 12 g/cm ³ 0.000 11 g/cm ³	Solid density standard material, Hydrostatic weighing Apparatus/ SICT-CP-20706-11

* 20704, 20705, 20706, 20707 unit % is weight percent.

208. Viscosity

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Kinematic viscometers; capillary, etc	20801	(2.5 ~ 100 000) mPa·s	1.3×10^{-2}	Viscosity CRM/ SICT-CP-20801
Dynamic viscometers; rotational, etc Viscosity	20802	(2.5 ~ 200 000) mPa·s	1.4×10^{-2}	Viscosity CRM/ SICT-CP-20802

209. Fluid flow

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Anemometers; hot-wire	20901	(0.1 ~ 1.0) m/s (1.0 ~ 2.0) m/s (2.0 ~ 70) m/s	8.7×10^{-2} 8.4×10^{-3} 4.8×10^{-3}	Wind Tunnel, Pitot tube, LDV/ SICT-CP-20901
Anemometers; pitot tube, etc.	20902	(0.1 ~ 1.0) m/s (1.0 ~ 2.0) m/s (2.0 ~ 70) m/s	8.7×10^{-2} 8.4×10^{-3} 4.8×10^{-3}	Wind Tunnel, Pitot tube, LDV/ SICT-CP-20902
Gas flowmeters;differential pressure	20908	$(1.2 \times 10^{-5} \sim 0.12) \text{ m}^3/\text{h}$ (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters;differential pressure	20909	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Liquid flowmeters; electromagnetic	20910	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters;thermal mass, etc.	20911	$(1.2 \times 10^{-5} \sim 0.12) \text{ m}^3/\text{h}$ (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters; Coriolis, etc.	20912	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters;open channel, etc.	20914	$(1.2 \times 10^{-5} \sim 0.12) \text{ m}^3/\text{h}$ (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters; positive displacement	20915	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters;turbine	20916	$(1.2 \times 10^{-5} \sim 0.12) \text{ m}^3/\text{h}$ (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters;turbine	20917	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters;ultrasonic	20918	$(1.2 \times 10^{-5} \sim 0.12) \text{ m}^3/\text{h}$ (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929

209. Fluid flow

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Liquid flowmeters;ultrasonic	20919	(0.01 ~ 50) m ³ /h	1.2×10^{-3}	Master Meter/SICT-CP-20926
		(0.000 12 ~ 0.01) m ³ /h	2.6×10^{-3}	Weight measuring method/ SICT-CP-20927
		(0.01 ~ 50) m ³ /h	7.0×10^{-4}	
Gas flowmeters;variable area	20920	(1.2×10^{-5} ~ 0.12) m ³ /h	1.9×10^{-3}	Sonic Nozzle/SICT-CP-20928
		(0.12 ~ 300) m ³ /h	2.0×10^{-3}	Master Meter/SICT-CP-20929
		(300 ~ 4 000) m ³ /h	3.6×10^{-3}	
Liquid flowmeters;variable area	20921	(0.01 ~ 50) m ³ /h	1.2×10^{-3}	Master Meter/SICT-CP-20926
		(0.000 12 ~ 0.01) m ³ /h	2.6×10^{-3}	Weight measuring method/ SICT-CP-20927
		(0.01 ~ 50) m ³ /h	7.0×10^{-4}	
Gas flowmeters;vortex	20922	(1.2×10^{-5} ~ 0.12) m ³ /h	1.9×10^{-3}	Sonic Nozzle/SICT-CP-20928
		(0.12 ~ 300) m ³ /h	2.0×10^{-3}	Master Meter/SICT-CP-20929
		(300 ~ 4 000) m ³ /h	3.6×10^{-3}	
Liquid flowmeters;vortex	20923	(0.01 ~ 50) m ³ /h	1.2×10^{-3}	Master Meter/SICT-CP-20926
		(0.000 12 ~ 0.01) m ³ /h	2.6×10^{-3}	Weight measuring method/ SICT-CP-20927
		(0.01 ~ 50) m ³ /h	7.0×10^{-4}	
Anemometers; vane, etc	20925	(0.1 ~ 1.0) m/s	8.7×10^{-2}	Wind Tunnel, Pitot tube, LDV/ SICT-CP-20925
		(1.0 ~ 2.0) m/s	8.4×10^{-3}	
		(2.0 ~ 70) m/s	4.8×10^{-3}	
Others; Ultrasonic current meter	20999	(0.1 ~ 1.0) m/s	8.7×10^{-2}	Wind Tunnel, Pitot tube, LDV/ SICT-CP-20999
		(1.0 ~ 2.0) m/s	8.4×10^{-3}	
		(2.0 ~ 70) m/s	4.8×10^{-3}	

210. Hardness

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Brinell hardness testers	21001	(100 ~ 250) HBW 10/3 000 (250 ~ 450) HBW 10/3 000 (450 ~ 650) HBW 10/3 000	3.1 HBW 10/3 000 4.9 HBW 10/3 000 8.2 HBW 10/3 000	Brinell Hardness CRM/ SICT-CP-21001
Rockwell hardness testers	21002	(20 ~ 70) HRC (20 ~ 100) HRBW (42 ~ 86) HR30N (29 ~ 82) HR30TW	0.45 HRC 0.80 HRBW 0.70 HR30N 1.1 HR30TW	Rockwell Hardness CRM/ SICT-CP-21002
Shore hardness testers	21003	(20 ~ 100) HS	0.9 HS	Shore Hardness CRM/ SICT-CP-21003
Vickers hardness testers	21004	(50 ~ 300) HV 0.2 (300 ~ 600) HV 0.2 (600 ~ 850) HV 0.2 (50 ~ 300) HV 0.3 (300 ~ 600) HV 0.3 (600 ~ 850) HV 0.5 (50 ~ 300) HV 0.5 (300 ~ 600) HV 0.5 (600 ~ 850) HV 1 (50 ~ 300) HV 10 (300 ~ 600) HV 10 (600 ~ 850) HV 10 (300 ~ 600) HV 30 (600 ~ 850) HV 30	6.0 HV 0.2 18 HV 0.2 27 HV 0.2 5.0 HV 0.3 14 HV 0.3 26 HV 0.5 6.0 HV 0.5 15 HV 0.5 20 HV 1 3.0 HV 10 8.0 HV 10 11 HV 10 8.0 HV 30 11 HV 30	Vickers Hardness CRM/ SICT-CP-21004
Durometer hardness testers	21005	(0 ~ 100) HDA (0 ~ 100) HDAM (0 ~ 100) HDAO (0 ~ 100) HDB (0 ~ 100) HDC (0 ~ 100) HDC2 (0 ~ 100) HDCS (0 ~ 100) HDD (0 ~ 100) HDD0 (0 ~ 100) HDE (0 ~ 100) HDE2 (0 ~ 100) HDF (0 ~ 100) HDF0 (0 ~ 100) HDM (0 ~ 100) HDO (0 ~ 100) HD00 (0 ~ 100) HD000 (0 ~ 100) HD000-S	0.4 HDA 0.8 HDAM 0.4 HDAO 0.4 HDB 0.3 HDC 0.6 HDC2 0.3 HDCS 0.3 HDD 0.3 HDD0 0.4 HDE 0.6 HDE2 0.6 HDF 0.6 HDF0 0.8 HDM 0.3 HDO 0.4 HD00 0.4 HD000 0.3 HD000-S	Durometer Calibration device/ SICT-CP-21005
Leeb hardness testers	21006	(400 ~ 1 000) HLD	4.6 HLD	Leeb Hardness CRM/ SICT-CP-21006

301. Time/frequency

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Frequency standards Time Base Frequency	30102	(1 ~ 100) MHz	4.3×10^{-12}	GPS Receiver, Universal Counter/ SICT-CP-30102
General frequency sources Time Base Frequency	30103	(10 ~ 100) kHz (0.1 ~ 100) MHz	1.3×10^{-11} 2.2×10^{-12}	GPS Receiver, Universal Counter/ SICT-CP-30103
Frequency meters/counters Time Base Frequency Input Frequency	30104	(1 ~ 10) MHz 0.1 Hz 0.1 Hz ~ 40 GHz	4.3×10^{-12} 6.4×10^{-10} 6.4×10^{-11}	GPS Receiver, Universal Counter/ SICT-CP-30104
Time interval sources Period Time interval	30105	1 ns ~ 10 s (1 ~ 100) ns 100 ns ~ 1 ms 1 ms ~ 10 s	6.1×10^{-9} 0.15 ns 1.3 ns 2.1 ns	GPS Receiver, Universal Counter/ SICT-CP-30105
Time interval meters /Stop watches/Timers Trigger Voltage Period Reference Frequency Relative Time Difference Time rate Timer Count	30106	(-5 ~ 5) V (5 ~ 100) ns (1 ~ 10) MHz day month (-9.95 ~ 9.95) s / day (-300 ~ 300) s / month (1 ~ 100) s (100 ~ 1 000) s (1 000 ~ 10 000) s ≥ 1	1.2×10^{-4} 6.2×10^{-5} ns 6.2×10^{-11} 1.1×10^{-7} 3.6×10^{-7} 6.1 ms 6.2 ms 5.8×10^{-6} 8.2×10^{-6} 5.8×10^{-5} 0.58	Stop Watch Calibrator/ SICT-CP-30106

302. Velocity & revolution

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Standard RPM generators Revolution Velocity Measurement	30201	(1 ~ 10 000) min ⁻¹	0.004 0 min ⁻¹	GPS Receiver, Synthesizer Function Generator/ SICT-CP-30201
Revolution Velocity Measurement (Centrifuge)		(30 ~ 5 000) min ⁻¹	0.059 min ⁻¹	
		(5 000 ~ 8 500) min ⁻¹	0.099 min ⁻¹	
		(8 500 ~ 50 000) min ⁻¹	0.59 min ⁻¹	
		(50 000 ~ 80 000) min ⁻¹	0.93 min ⁻¹	
		(80 000 ~ 99 000) min ⁻¹	1.1 min ⁻¹	
Contact type tachometers Revolution Velocity Measurement	30202	(1 ~ 10) min ⁻¹	0.10 min ⁻¹	GPS Receiver, Tachometer Cal System/ SICT-CP-30202
		(10 ~ 1 000) min ⁻¹	0.016 min ⁻¹	
		(1 000 ~ 5 000) min ⁻¹	0.063 min ⁻¹	
Photo tachometers/stroboscopes Revolution Velocity Measurement (Photo-tachometer)	30203	(1 ~ 999.99) min ⁻¹	0.006 1 min ⁻¹	GPS Receiver, Photo Signal Detector/ SICT-CP-30203
		(1 000.0 ~ 99 999.9) min ⁻¹	0.061 min ⁻¹	
		(100 000 ~ 600 000) min ⁻¹	0.61 min ⁻¹	
Revolution Velocity Measurement (Stroboscope)		(30 ~ 9 000) min ⁻¹	0.005 8 min ⁻¹	
		(9 000 ~ 90 000) min ⁻¹	0.058 min ⁻¹	
		(90 000 ~ 500 000) min ⁻¹	0.58 min ⁻¹	
Speed meters Speed Test	30204	(0 ~ 400) km/h	6.1×10^{-3} km/h	GPS Receiver, Synthesizer Function Generator/ SICT-CP-30204
Wow-flutter generators Carrier Frequency	30205	10 Hz ~ 99.99 kHz	6.2×10^{-6}	GPS Receiver, Universal Counter/ SICT-CP-30205
Function Frequency		1 Hz ~ 10 kHz	6.2×10^{-6}	
		(10 ~ 30) kHz	2.1×10^{-6}	
Wow/Flutter Deviation		(1 Hz ~ 100 Hz)	0.025 %	
Output Level		(1 ~ 10) mV	5.8×10^{-4}	
		(10 ~ 100) mV	1.7×10^{-4}	
	(0.1 ~ 6) V	1.3×10^{-4}		
CCIR Pulse		10 ms	1.0×10^{-2} ms	
		30 ms	3.0×10^{-2} ms	
		60 ms	6.0×10^{-2} ms	
		100 ms	1.0×10^{-1} ms	
Wow-flutter meters Wow/Flutter Deviation	30206	(0.1 ~ 0.3) %	0.019 %	GPS Receiver, Wow Flutter Calibrator/ SICT-CP-30206
		(0.3 ~ 3) %	0.020 %	
Carrier Frequency		3 kHz	6.2×10^{-5} kHz	
		3.15 kHz	6.2×10^{-5} kHz	
CCIR Pulse		(10 ~ 100) ms	0.59 %	
Output Voltage		(1 ~ 100) mV	6.8 μV	
		(0.1 ~ 1) V	9.8 μV	
		(1 ~ 10) V	76 μV	

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC ammeters DC Current	40101	(±) 0 nA (0 ~ 1) nA (1 ~ 10) nA (10 ~ 100) nA (0.1 ~ 1) μA (1 ~ 10) μA (10 ~ 100) μA (0.1 ~ 1) mA (1 ~ 10) mA (10 ~ 100) mA (100 ~ 200) mA (0.2 ~ 1) A (1 ~ 10) A (10 ~ 100) A	6 pA 6.6×10^{-3} 9.5×10^{-4} 5.1×10^{-4} 6.1×10^{-4} 2.9×10^{-4} 3.6×10^{-5} 2.9×10^{-5} 2.6×10^{-5} 1.9×10^{-5} 1.3×10^{-5} 2.1×10^{-5} 1.4×10^{-4} 1.5×10^{-4}	Calibrator/ SICT-CP-40101
Transconductance amplifiers DC Current AC Current	40102	(±) 10 μA ~ 10 A (10 ~ 50) A (50 ~ 100) A (10 μA) 10 Hz ~ 10 kHz (10 ~ 100) μA 10 Hz ~ 1 kHz (1 ~ 10) kHz (100 μA ~ 1 mA) 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (1 ~ 100) mA 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (100 mA ~ 1 A) 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (1 ~ 2) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (2 ~ 5) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	1.3×10^{-5} 4.3×10^{-5} 4.4×10^{-5} 2.6×10^{-3} 3.6×10^{-4} 6.4×10^{-4} 9.8×10^{-5} 7.5×10^{-5} 9.4×10^{-5} 7.8×10^{-5} 4.6×10^{-5} 4.2×10^{-5} 8.1×10^{-5} 4.9×10^{-5} 4.4×10^{-5} 7.9×10^{-5} 4.7×10^{-5} 4.5×10^{-5} 8.2×10^{-5} 5.2×10^{-5} 5.0×10^{-5}	AC-DC Active Current Shunt/ SICT-CP-40102

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Transconductance amplifiers AC Current	40102	(5 ~ 10) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 20) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (20 ~ 50) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (50 ~ 100) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (100 ~ 120) A 60 Hz	 8.6×10^{-5} 5.9×10^{-5} 7.8×10^{-5} 9.3×10^{-5} 6.8×10^{-5} 7.8×10^{-5} 1.0×10^{-4} 8.3×10^{-5} 1.1×10^{-4} 1.2×10^{-4} 9.7×10^{-5} 1.3×10^{-5} 4.6×10^{-4}	AC-DC Active Current Shunt/ SICT-CP-40102
DC voltage/current calibrators DC Voltage DC Current	40103	(±) 0 mV (0 ~ 1) mV (1 ~ 2) mV (2 ~ 10) mV (10 ~ 50) mV (50 ~ 100) mV (0.1 ~ 1) V (1 ~ 1 000) V (±) 0 nA (0 ~ 1) nA (1 ~ 100) nA 100 nA ~ 1 A (1 ~ 10) A (10 ~ 100) A	 0.24 μV 4.0×10^{-4} 2.0×10^{-4} 8.1×10^{-5} 4.1×10^{-5} 1.0×10^{-5} 6.3×10^{-6} 2.6×10^{-6} 7.0 pA 7.0×10^{-3} 4.7×10^{-3} 1.3×10^{-5} 1.3×10^{-4} 4.4×10^{-5}	Reference Multimeter/ SICT-CP-40103
Electrical temperature calibrators TEMPERATURE(SOURCE) T/C RTD	40104	(-9.835 ~ 0.000) mV 0.000 mV (0.000 ~ 13.421) mV (13.421 ~ 37.006) mV (37.006 ~ 61.017) mV (61.017 ~ 76.373) mV 0.999 Ω (0.999 ~ 2.499) Ω (2.499 ~ 4.322) Ω (4.322 ~ 100.000) Ω (100.000 ~ 177.155) Ω (177.155 ~ 313.708) Ω (313.708 ~ 627.422) Ω (627.422 ~ 3 233.3) Ω	0.42 μV 0.24 μV 0.42 μV 0.48 μV 0.53 μV 0.57 μV 0.063 mΩ 3.0×10^{-5} 1.9×10^{-5} 9.9×10^{-6} 8.8×10^{-6} 1.1×10^{-5} 9.2×10^{-6} 1.1×10^{-5}	디지털 멀티미터/ SICT-CP-40104

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrical temperature calibrators DC Voltage(SOURCE)	40104	(±) 0 mV (1 ~ 2) mV (2 ~ 3) mV (3 ~ 4) mV (4 ~ 5) mV (5 ~ 10) mV (10 ~ 50) mV (50 ~ 100) mV (0.1 ~ 0.2) V (0.2 ~ 0.3) V (0.3 ~ 0.6) V (0.6 ~ 1) V (1 ~ 6) V (6 ~ 10) V (10 ~ 70) V (70 ~ 100) V	0.24 μV 4.0×10^{-4} 2.0×10^{-4} 1.3×10^{-4} 1.0×10^{-4} 8.2×10^{-5} 4.2×10^{-5} 6.3×10^{-6} 6.2×10^{-5} 3.1×10^{-5} 3.1×10^{-5} 9.5×10^{-6} 3.1×10^{-5} 9.3×10^{-6} 6.2×10^{-5} 9.2×10^{-6}	디지털 멀티미터/ SICT-CP-40104
DC Current(SOURCE)		(±) 0 mA (0 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 7) mA (7 ~ 10) mA (10 ~ 20) mA (20 ~ 30) mA (30 ~ 40) mA (40 ~ 100) mA	0.064 μA 9.2×10^{-5} 6.2×10^{-5} 3.5×10^{-5} 2.3×10^{-5} 1.9×10^{-5} 3.3×10^{-5} 8.2×10^{-5} 7.0×10^{-5} 6.3×10^{-5}	
Resistance(SOURCE)		0 Ω (0 ~ 0.6) Ω (0.6 ~ 1) Ω (1 ~ 10) Ω (10 ~ 20) Ω (20 ~ 30) Ω (30 ~ 50) Ω (50 ~ 70) Ω (70 ~ 100) Ω (0.1 ~ 0.2) kΩ (0.2 ~ 0.3) kΩ (0.3 ~ 0.5) kΩ (0.5 ~ 0.8) kΩ (0.8 ~ 1) kΩ (1 ~ 2) kΩ (2 ~ 3) kΩ (3 ~ 5) kΩ (5 ~ 8) kΩ (8 ~ 10) kΩ (10 ~ 20) kΩ (20 ~ 30) kΩ (30 ~ 40) kΩ (40 ~ 50) kΩ (50 ~ 100) kΩ	0.061 mΩ 6.1×10^{-4} 8.9×10^{-5} 6.7×10^{-5} 3.3×10^{-5} 2.3×10^{-5} 1.8×10^{-5} 1.4×10^{-5} 1.2×10^{-5} 3.2×10^{-5} 2.3×10^{-5} 1.8×10^{-5} 1.4×10^{-5} 1.1×10^{-5} 3.2×10^{-5} 2.3×10^{-5} 1.8×10^{-5} 1.4×10^{-5} 1.1×10^{-5} 3.2×10^{-5} 2.4×10^{-5} 1.9×10^{-5} 1.6×10^{-5} 1.1×10^{-5}	

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Electrical temperature calibrators	40104			디지털 멀티미터/ SICT-CP-40104	
TEMPERATURE(MEASURE)					
T/C		(-9.835 ~ 0.000) mV	0.59 μ V		
		0.000 mV	0.50 μ V		
		(0.000 ~ 13.422) mV	0.62 μ V		
		(13.422 ~ 28.947) mV	0.75 μ V		
		(28.947 ~ 45.094) mV	0.88 μ V		
		(45.094 ~ 53.113) mV	0.95 μ V		
		(53.113 ~ 76.374) mV	1.1 μ V		
		RTD	0.998 Ω		0.24 m Ω
			(0.998 ~ 2.496) Ω		1.0×10^{-4}
			(2.496 ~ 4.315) Ω		7.1×10^{-5}
			(4.315 ~ 16.994) Ω		3.9×10^{-5}
			(16.994 ~ 249.580) Ω		3.5×10^{-5}
			(249.580 ~ 317.988) Ω		4.3×10^{-5}
			(317.988 ~ 390.474) Ω		4.0×10^{-5}
			(390.474 ~ 3 233.2) Ω		3.5×10^{-5}
DC Voltage(MEASURE)			(\pm)		
			0 mV		0.50 μ V
			(1 ~ 5) mV		5.2×10^{-4}
			(5 ~ 10) mV		9.3×10^{-5}
			(10 ~ 100) mV		5.9×10^{-5}
			(0.1 ~ 0.5) V		6.3×10^{-5}
			(0.5 ~ 0.8) V		1.3×10^{-5}
			(0.8 ~ 1) V		1.6×10^{-5}
			(1 ~ 10) V		6.6×10^{-6}
			(10 ~ 20) V		9.1×10^{-6}
			(20 ~ 40) V		7.9×10^{-6}
			(40 ~ 70) V		6.9×10^{-6}
			(70 ~ 100) V		6.4×10^{-6}
			(100 ~ 200) V		7.8×10^{-6}
			(200 ~ 300) V		2.2×10^{-5}
DC Current(MEASURE)			(\pm)		
			0 mA		0.062 μ A
			(0 ~ 0.1) mA		6.4×10^{-4}
			(0.1 ~ 0.2) mA		3.2×10^{-4}
			(0.2 ~ 0.3) mA		2.2×10^{-4}
		(0.3 ~ 0.7) mA	1.7×10^{-4}		
		(0.7 ~ 1) mA	9.3×10^{-5}		
		(1 ~ 2) mA	9.9×10^{-5}		
		(2 ~ 5) mA	7.6×10^{-5}		
		(5 ~ 10) mA	5.8×10^{-5}		
		(10 ~ 20) mA	9.9×10^{-5}		
		(20 ~ 30) mA	8.2×10^{-5}		
		(30 ~ 40) mA	7.4×10^{-5}		
		(40 ~ 50) mA	7.0×10^{-5}		
		(50 ~ 100) mA	6.7×10^{-5}		
		(100 ~ 130) mA	8.7×10^{-5}		

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrical temperature calibrators Resistance(MEASURE)	40104	0 Ω (0 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) kΩ (1 ~ 10) kΩ (10 ~ 20) kΩ (20 ~ 30) kΩ (30 ~ 40) kΩ (40 ~ 50) kΩ (50 ~ 100) kΩ	0.098 mΩ 6.4×10^{-5} 1.1×10^{-5} 9.6×10^{-6} 6.5×10^{-5} 6.1×10^{-5} 4.7×10^{-5} 4.0×10^{-5} 4.2×10^{-5} 3.9×10^{-5} 3.4×10^{-5}	디지털 멀티미터/ SICT-CP-40104
DC current shunts Resistance	40105	1 μΩ (0.001 ~ 0.01) mΩ (0.01 ~ 0.2) mΩ (0.2 ~ 1) mΩ (1 ~ 10) mΩ (10 ~ 100) mΩ (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) kΩ (1 ~ 10) kΩ (10 ~ 100) kΩ (0.1 ~ 1) MΩ (1 ~ 10) MΩ (10 ~ 100) MΩ	0.32 nΩ 2.8×10^{-4} 1.8×10^{-4} 1.5×10^{-4} 1.1×10^{-6} 2.8×10^{-6} 8.1×10^{-7} 1.3×10^{-6} 6.7×10^{-7} 6.2×10^{-7} 7.9×10^{-7} 2.0×10^{-6} 1.4×10^{-6} 5.3×10^{-6} 9.8×10^{-6}	Trans Conductance Amplifier/ SICT-CP-40105
Galvanometers/null detectors DC Voltage	40106	(±) (100 ~ 300) μV (0.3 ~ 1) mV 1 mV ~ 1 000 V	1.4×10^{-2} 1.2×10^{-2} 6.8×10^{-3}	Calibrator/ SICT-CP-40106
Potentiometers DC Voltage	40107	(100 ~ 300) μV (0.3 ~ 1) mV (1 ~ 3) mV 3 mV ~ 1 000 V	5.7×10^{-3} 2.2×10^{-3} 6.0×10^{-4} 3.0×10^{-4}	Calibrator/ SICT-CP-40107

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC power supplies DC Voltage	40108	(±) 0 mV (0 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 600) V (600 ~ 1 000) V	5.8 μV 5.8×10^{-4} 5.8×10^{-5} 7.5×10^{-6} 3.3×10^{-6} 7.7×10^{-6} 1.3×10^{-5} 6.6×10^{-5}	DC Electronics Load/ SICT-CP-40108
DC Current		(1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A (10 ~ 300) A (300 ~ 500) A (500 ~ 1 000) A (1 000 ~ 3 000) A	5.8×10^{-3} 5.9×10^{-4} 2.4×10^{-4} 3.1×10^{-4} 2.4×10^{-4} 2.6×10^{-4} 4.7×10^{-5} 5.1×10^{-4}	
Load regulation		(0 ~ 2) mV (2 ~ 20) mV (20 ~ 200) mV	0.16 mV 7.8×10^{-2} 8.2×10^{-3}	
Ripple		(0.1 ~ 0.4) mV (0.4 ~ 0.6) mV (0.6 ~ 1) mV (1 ~ 10) mV (10 ~ 50) mV	3.8×10^{-1} 1.1×10^{-1} 7.3×10^{-2} 4.4×10^{-2} 7.1×10^{-2}	
DC voltage dividers DC Voltage Ratio	40110	(±) (0.01 ~ 1 000) V (1 ~ 50) kV (50 ~ 100) kV	4.5×10^{-6} 8.8×10^{-5} 8.4×10^{-5}	Calibrator/ SICT-CP-40110
DC voltage standards DC Voltage	40111	1 V 1.018 V 10 V	1.6 μV 0.8 μV 3.1 μV	Null Detector/ SICT-CP-40111
DC voltmeters DC Voltage	40112	(±) 0 mV (0 ~ 1) mV (1 ~ 2) mV (2 ~ 5) mV (5 ~ 10) mV (10 ~ 20) mV (20 ~ 50) mV (50 ~ 80) mV (80 ~ 100) mV (0.1 ~ 1 000) V	0.51 μV 5.0×10^{-4} 2.5×10^{-4} 1.0×10^{-4} 6.2×10^{-5} 5.0×10^{-5} 2.5×10^{-5} 1.0×10^{-5} 6.2×10^{-6} 8.0×10^{-6}	Calibrator/ SICT-CP-40112
Static/Ionic voltmeters DC Voltage	40113	(±) 0 V 0 V ~ 50 kV	68 mV 1.3×10^{-2}	DC Power Supply/ SICT-CP-40113

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Capacitance bridges/indicators	40201	Frequency	50 Hz ~ 100 MHz	7.0×10^{-8}	Standard Capacitance Set/ SICT-CP-40201
		Capacitance	(1 pF) 50 Hz ~ 1 kHz 1 kHz ~ 1 MHz 2 MHz 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz (10 pF) 50 Hz ~ 5 MHz 10 MHz 13 MHz (100 pF) 50 Hz ~ 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz (1 000 pF) 50 Hz ~ 1 MHz 2 MHz 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz (10 nF) (50 ~ 100) Hz 100 Hz ~ 100 kHz (100 nF) (50 ~ 100) Hz 100 Hz ~ 100 kHz (1 μF) (50 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	3.5×10^{-4} 3.6×10^{-4} 3.6×10^{-4} 4.2×10^{-4} 5.4×10^{-4} 7.2×10^{-4} 9.4×10^{-4} 2.5×10^{-3} 3.6×10^{-3} 3.5×10^{-4} 3.7×10^{-4} 3.8×10^{-4} 3.5×10^{-4} 3.6×10^{-4} 3.7×10^{-4} 4.8×10^{-4} 6.0×10^{-4} 3.5×10^{-4} 3.8×10^{-4} 4.5×10^{-4} 5.6×10^{-4} 7.2×10^{-4} 2.0×10^{-3} 2.9×10^{-3} 3.0×10^{-4} 8.1×10^{-5} 3.0×10^{-4} 8.1×10^{-4} 5.1×10^{-4} 8.1×10^{-5} 1.0×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Capacitance bridges/indicators Capacitance	40201	(10 μF) 120 Hz ~ 1 kHz (100 μF) 120 Hz (1 mF) 120 Hz (3 mF) 120 Hz (10 mF) 120 Hz (30 mF) 120 Hz	1.2×10^{-3} 1.3×10^{-3} 1.4×10^{-3} 1.4×10^{-3} 1.4×10^{-3} 2.9×10^{-3}	Standard Capacitance Set/ SICT-CP-40201
Decade capacitors Capacitance	40202	(50 Hz ~ 20 kHz) 1 pF (1 ~ 10) pF (10 ~ 100) pF (100 ~ 1 000) pF 1 000 pF ~ 100 nF 100 nF ~ 1 μF (1 kHz) 1 pF (1 ~ 1 000) pF 1 000 pF ~ 100 nF 100 nF ~ 1 μF	5.7×10^{-5} 4.6×10^{-5} 3.8×10^{-5} 4.6×10^{-5} 2.9×10^{-4} 5.1×10^{-4} 2.5×10^{-5} 2.4×10^{-5} 5.5×10^{-5} 9.3×10^{-5}	Standard Capacitance Set/ SICT-CP-40202
Standard capacitors Capacitance	40204	(50 Hz ~ 20 kHz) 1 pF (1 ~ 10) pF (10 ~ 100) pF (100 ~ 1 000) pF 1 000 pF ~ 100 nF 100 nF ~ 1 μF (1 kHz) 1 pF (1 ~ 10) pF (10 ~ 100) pF (100 ~ 1 000) pF 1 000 pF ~ 100 nF 100 nF ~ 1 μF (1 pF) 1 kHz 1 kHz ~ 1 MHz (1 ~ 2) MHz (2 ~ 3) MHz (3 ~ 4) MHz (4 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz	5.2×10^{-5} 4.0×10^{-5} 3.0×10^{-5} 4.0×10^{-5} 2.9×10^{-4} 5.1×10^{-4} 9.1×10^{-6} 7.1×10^{-6} 6.1×10^{-6} 7.1×10^{-6} 5.0×10^{-5} 9.0×10^{-5} 2.4×10^{-4} 2.5×10^{-4} 3.3×10^{-4} 4.7×10^{-4} 6.7×10^{-4} 9.1×10^{-4} 2.5×10^{-3} 3.7×10^{-3}	Standard Capacitance Set/ Capacitance Bridge SICT-CP-40204

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Standard capacitors Capacitance	40204	(1 ~ 10) pF 1 kHz ~ 3 MHz (3 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz (10 ~ 100) pF 1 kHz ~ 1 MHz (1 ~ 3) MHz (3 ~ 4) MHz (4 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz (100 pF ~ 1 nF) 1 kHz 1 kHz ~ 1 MHz (1 ~ 2) MHz (2 ~ 3) MHz (3 ~ 4) MHz (4 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz (1 ~ 100) nF 120 Hz ~ 100 kHz (100 nF ~ 1 μF) 120 Hz 120 Hz ~ 10 kHz (10 ~ 100) kHz (1 ~ 10) μF 120 Hz ~ 1 kHz (30 μF) 120 Hz (100 μF) 120 Hz (300 μF) 120 Hz (1 mF) 120 Hz (3 mF) 120 Hz (10 mF) 120 Hz (30 mF) 120 Hz	2.3×10^{-4} 2.4×10^{-4} 2.6×10^{-4} 2.8×10^{-4} 2.3×10^{-4} 2.4×10^{-4} 2.5×10^{-4} 2.7×10^{-4} 4.0×10^{-4} 5.4×10^{-4} 2.3×10^{-4} 2.4×10^{-4} 2.8×10^{-4} 3.6×10^{-4} 5.0×10^{-4} 6.6×10^{-4} 1.9×10^{-3} 2.8×10^{-3} 2.3×10^{-4} 2.4×10^{-4} 2.3×10^{-4} 2.4×10^{-4} 1.2×10^{-3} 1.3×10^{-3} 1.3×10^{-3} 1.5×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.4×10^{-3} 2.9×10^{-3}	Standard Capacitance Set/ Capacitance Bridge SICT-CP-40204

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Earth testers	40205			Decade Resistor/ SICT-CP-40205
Test Voltage		1 V (1 ~ 10) V (10 ~ 50) V (50 ~ 100) V (100 ~ 500) V (500 ~ 1 000) V	6.4×10^{-3} 6.4×10^{-4} 1.3×10^{-3} 6.4×10^{-4} 1.6×10^{-4} 6.4×10^{-4}	
Resistance		1 mΩ (1 ~ 10) mΩ 10 mΩ ~ 100 kΩ	8.6×10^{-4} 7.2×10^{-4} 6.8×10^{-4}	
AC Current out		1 A (1 ~ 3) A (3 ~ 20) A (20 ~ 30) A (30 ~ 60) A (60 ~ 100) A (100 ~ 150) A (150 ~ 200) A	1.2×10^{-3} 1.5×10^{-3} 9.7×10^{-4} 1.0×10^{-3} 8.4×10^{-4} 1.0×10^{-3} 4.6×10^{-3} 3.7×10^{-3}	
Timer		1 s (1 ~ 100) s (100 ~ 1 000) s (1 000 ~ 10 000) s	5.8×10^{-6} 5.8×10^{-6} 8.2×10^{-6} 5.8×10^{-5}	
Inductance bridges/indicators	40206			Standard Inductor/ SICT-CP-40206
Frequency		50 Hz ~ 100 MHz	7.0×10^{-8}	
Inductance		(1 kHz) 100 μH 1 mH 10 mH 100 mH 1 H 10 H	1.9×10^{-4} 1.3×10^{-4} 1.3×10^{-4} 1.3×10^{-4} 1.3×10^{-4} 1.3×10^{-4} 1.3×10^{-4}	
Inductors	40208			Standard Inductor/ SICT-CP-40208
Standard Inductance		(1 kHz) 100 μH 1 mH 10 mH 100 mH 1 H 10 H	28 nH 0.24 μH 2.4 μH 24 μH 0.24 mH 2.5 mH	
Decade Inductance		(1 kHz) 100 μH ~ 10 H	3.5×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Insulation testers	40210			High Resistance Decade/ SICT-CP-40210
AC Voltage		1 V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V	8.8×10^{-5} 9.0×10^{-5} 1.0×10^{-4} 1.1×10^{-4}	
Insulation Voltage		1 V (1 ~ 10) V (10 ~ 25) V (25 ~ 50) V (50 ~ 100) V (100 ~ 250) V (250 ~ 500) V (500 ~ 1 000) V (1 000 ~ 5 000) V (5 000 ~ 10 000) V	6.4×10^{-4} 6.4×10^{-5} 2.5×10^{-4} 1.3×10^{-4} 6.4×10^{-5} 2.5×10^{-4} 1.3×10^{-4} 6.4×10^{-5} 6.5×10^{-3} 6.1×10^{-3}	
Insulation Resistance		1 k Ω (1 ~ 10) k Ω (10 ~ 100) k Ω (0.1 ~ 1) M Ω (1 ~ 10) M Ω (10 ~ 100) M Ω (0.1 ~ 1) G Ω (1 ~ 10) G Ω (10 ~ 100) G Ω (0.1 ~ 1) T Ω 10 T Ω	7.1×10^{-5} 3.7×10^{-5} 2.5×10^{-5} 3.1×10^{-5} 9.5×10^{-5} 2.4×10^{-5} 3.1×10^{-5} 6.1×10^{-5} 1.3×10^{-4} 2.6×10^{-4} 6.3×10^{-4}	
Q-meters	40211			Universal Counter/ SICT-CP-40211
Frequency Test Quality Factor		60 Hz ~ 100 MHz 0 ~ 1 000	7.0×10^{-8} 9.7×10^{-4}	
Resistance bridges & similar instruments Resistance(Rheostat Arm)	40213			Standard Resistance Set/ SICT-CP-40213
		1 m Ω (1 ~ 10) m Ω (10 ~ 100) m Ω (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) k Ω (1 ~ 10) k Ω (10 ~ 100) k Ω (0.1 ~ 1) M Ω (1 ~ 10) M Ω (10 ~ 100) M Ω	1.5×10^{-6} 7.3×10^{-7} 2.7×10^{-6} 5.6×10^{-7} 1.1×10^{-6} 3.4×10^{-7} 2.2×10^{-7} 5.4×10^{-7} 2.0×10^{-6} 1.3×10^{-6} 5.3×10^{-6} 9.8×10^{-6}	
Resistance(Ratio Arm)		1 m Ω (1 ~ 10) m Ω (10 ~ 100) m Ω (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) k Ω (1 ~ 10) k Ω (10 ~ 100) k Ω (0.1 ~ 1) M Ω (1 ~ 10) M Ω (10 ~ 100) M Ω	1.5×10^{-6} 7.3×10^{-7} 2.7×10^{-6} 5.6×10^{-7} 1.1×10^{-6} 3.4×10^{-7} 2.2×10^{-7} 5.4×10^{-7} 2.0×10^{-6} 1.3×10^{-6} 5.3×10^{-6} 9.8×10^{-6}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistance meters DC Resistance	40214	10 $\mu\Omega$	3.3×10^{-4}	Standard Resistance Set/ SICT-CP-40214
		100 $\mu\Omega$	4.7×10^{-4}	
		1 m Ω	1.3×10^{-6}	
		(1 ~ 10) m Ω	8.0×10^{-7}	
		(10 ~ 100) m Ω	2.7×10^{-6}	
		(0.1 ~ 1) Ω	5.6×10^{-7}	
		(1 ~ 10) Ω	1.1×10^{-6}	
		(10 ~ 100) Ω	3.3×10^{-7}	
		(0.1 ~ 1) k Ω	2.2×10^{-7}	
		(1 ~ 10) k Ω	5.3×10^{-7}	
		(10 ~ 100) k Ω	1.9×10^{-6}	
		(0.1 ~ 1) M Ω	1.2×10^{-6}	
		(1 ~ 10) M Ω	5.2×10^{-6}	
		(10 ~ 100) M Ω	9.8×10^{-6}	
		(0.1 ~ 1) G Ω	3.2×10^{-5}	
		(1 ~ 10) G Ω	6.2×10^{-5}	
		(10 ~ 100) G Ω	1.3×10^{-4}	
		(0.1 ~ 1) T Ω	2.7×10^{-4}	
		10 T Ω	6.4×10^{-4}	
		Frequency	1 kHz	
AC Voltage	10 mV	1.6×10^{-4}		
	(10 ~ 100) mV	7.9×10^{-5}		
	(0.1 ~ 10) V	8.2×10^{-5}		
AC Resistance	(1 kHz)			
	1 m Ω	5.0×10^{-3}		
	(1 ~ 10) m Ω	5.2×10^{-4}		
	(10 ~ 100) m Ω	3.3×10^{-4}		
	100 m Ω ~ 10 k Ω	1.3×10^{-4}		
	(10 ~ 100) k Ω	1.4×10^{-4}		
	(0.1 ~ 1) M Ω	3.0×10^{-4}		
(1 ~ 10) M Ω	2.9×10^{-3}			
Resistors DC Resistance	40215	1 m Ω	1.6×10^{-6}	Standard Resistance Set/ SICT-CP-40215
		(1 ~ 10) m Ω	1.1×10^{-6}	
		(10 ~ 100) m Ω	2.8×10^{-6}	
		(0.1 ~ 1) Ω	8.1×10^{-7}	
		(1 ~ 10) Ω	1.3×10^{-6}	
		(10 ~ 100) Ω	6.7×10^{-7}	
		(0.1 ~ 1) k Ω	6.2×10^{-7}	
		(1 ~ 10) k Ω	7.9×10^{-7}	
		(10 ~ 100) k Ω	2.0×10^{-6}	
		(0.1 ~ 1) M Ω	1.4×10^{-6}	
		(1 ~ 10) M Ω	5.2×10^{-6}	
		(10 ~ 100) M Ω	9.7×10^{-6}	
		(0.1 ~ 1) G Ω	2.3×10^{-4}	
		(1 ~ 10) G Ω	6.9×10^{-4}	
		(10 ~ 100) G Ω	9.3×10^{-4}	
		(0.1 ~ 1) T Ω	1.4×10^{-3}	
(1 ~ 10) T Ω	4.1×10^{-3}			
(10 ~ 100) T Ω	7.6×10^{-3}			

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistors AC Resistance	40215	(50 Hz ~ 1 kHz) 1 mΩ (1 ~ 10) mΩ (10 ~ 100) mΩ (0.1 ~ 1) Ω (1 ~ 100) Ω (10 Ω) 1 kHz 1 MHz 2 MHz 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz (10 ~ 100) Ω 1 kHz 1 MHz 2 MHz 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz (100 Ω ~ 1 kΩ) 1 kHz 100 kHz 1 MHz 2 MHz 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz (1 ~ 10) kΩ 1 kHz 100 kHz 1 MHz (10 ~ 100) kΩ 1 kHz 100 kHz 1 MHz (100 kΩ ~ 1 MΩ) 1 kHz (1 ~ 10) MΩ 1 kHz	1.0×10^{-3} 5.9×10^{-4} 3.9×10^{-4} 2.4×10^{-4} 2.5×10^{-4} 2.4×10^{-4} 4.0×10^{-4} 5.6×10^{-4} 6.5×10^{-4} 7.5×10^{-4} 1.0×10^{-3} 4.0×10^{-3} 6.0×10^{-3} 2.4×10^{-4} 4.0×10^{-4} 4.8×10^{-4} 5.6×10^{-4} 5.6×10^{-4} 5.6×10^{-4} 2.0×10^{-3} 3.0×10^{-3} 2.4×10^{-4} 4.0×10^{-4} 4.0×10^{-4} 4.0×10^{-4} 4.0×10^{-4} 4.8×10^{-4} 5.6×10^{-4} 2.0×10^{-3} 3.0×10^{-3} 2.4×10^{-4} 3.3×10^{-4} 4.0×10^{-4} 2.4×10^{-4} 4.0×10^{-4} 4.0×10^{-4} 3.8×10^{-4} 3.0×10^{-3}	Standard Resistance Set/ SICT-CP-40215

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistors Decade Resistance	40215	0 Ω (0 ~ 10) mΩ (10 ~ 100) mΩ (0.1 ~ 1) Ω (1 ~ 7) Ω (7 ~ 10) Ω (10 ~ 70) Ω (70 ~ 100) Ω (100 ~ 700) Ω (0.7 ~ 1) kΩ (1 ~ 7) kΩ (7 ~ 10) kΩ (10 ~ 70) kΩ (70 ~ 100) kΩ (100 ~ 600) kΩ (0.6 ~ 1) MΩ (1 ~ 7) MΩ (7 ~ 10) MΩ (10 ~ 100) MΩ (0.1 ~ 1) GΩ (1 ~ 10) GΩ (10 ~ 100) GΩ (0.1 ~ 1) TΩ (1 ~ 10) TΩ	64 μΩ 6.5 μΩ 64 μΩ 66 μΩ 3.9×10^{-5} 1.3×10^{-5} 2.0×10^{-5} 9.8×10^{-6} 1.9×10^{-5} 9.6×10^{-6} 3.5×10^{-5} 1.2×10^{-5} 2.0×10^{-5} 9.8×10^{-6} 2.9×10^{-5} 1.2×10^{-5} 8.0×10^{-5} 2.7×10^{-5} 2.0×10^{-4} 2.5×10^{-4} 7.0×10^{-4} 1.0×10^{-3} 1.5×10^{-3} 4.3×10^{-3}	Standard Resistance Set/ SICT-CP-40215
Impedance bridges/LCR meters Frequency AC Voltage Capacitance	40217	50 Hz ~ 100 MHz 1 mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 10) V (10 ~ 100) V (1 pF) 50 Hz ~ 1 kHz 1 kHz ~ 1 MHz 2 MHz 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz (10 pF) 50 Hz ~ 5 MHz 10 MHz 13 MHz	7.0×10^{-8} 2.1×10^{-3} 4.4×10^{-4} 8.8×10^{-5} 8.2×10^{-5} 8.9×10^{-5} 3.5×10^{-4} 3.6×10^{-4} 4.2×10^{-4} 5.4×10^{-4} 7.2×10^{-4} 9.4×10^{-4} 2.5×10^{-3} 3.6×10^{-3} 3.5×10^{-4} 3.7×10^{-4} 3.8×10^{-4}	Standard Capacitor Set, Standard Resistor Set, Standard Inductor/ SICT-CP-40217

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Impedance bridges/LCR meters Capacitance	40217	(100 pF)		Standard Capacitor Set, Standard Resistor Set, Standard Inductor/ SICT-CP-40217	
		50 Hz ~ 3 MHz	3.5×10^{-4}		
		4 MHz	3.6×10^{-4}		
		5 MHz	3.7×10^{-4}		
		10 MHz	4.8×10^{-4}		
		13 MHz	6.0×10^{-4}		
		(1 000 pF)			
		50 Hz ~ 1 MHz	3.5×10^{-4}		
		2 MHz	3.8×10^{-4}		
		3 MHz	4.5×10^{-4}		
		4 MHz	5.6×10^{-4}		
		5 MHz	7.2×10^{-4}		
		10 MHz	2.0×10^{-3}		
		13 MHz	2.9×10^{-3}		
		(10 nF)			
		(50 ~ 100) Hz	3.0×10^{-4}		
		100 Hz ~ 100 kHz	8.1×10^{-5}		
		(100 nF)			
		(50 ~ 100) Hz	3.0×10^{-4}		
		100 Hz ~ 100 kHz	8.1×10^{-5}		
		(1 μF)			
		(50 ~ 100) Hz	5.1×10^{-4}		
		100 Hz ~ 10 kHz	8.1×10^{-5}		
		(10 ~ 100) kHz	1.0×10^{-4}		
		(10 μF)			
		120 Hz ~ 1 kHz	1.2×10^{-3}		
(100 μF)					
120 Hz	1.3×10^{-3}				
(1 mF)					
120 Hz	1.4×10^{-3}				
(3 mF)					
120 Hz	1.4×10^{-3}				
(10 mF)					
120 Hz	1.4×10^{-3}				
(30 mF)					
120 Hz	2.9×10^{-3}				
(1 pF)					
1 kHz ~ 1 MHz	0.000 12				
1 MHz ~ 5 MHz	0.000 23				
5 MHz ~ 13 MHz	0.000 84				

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Impedance bridges/LCR meters	40217	Capacitance		Standard Capacitor Set, Standard Resistor Set, Standard Inductor/ SICT-CP-40217
		(10 pF) 1 kHz ~ 13 MHz	0.000 15	
		(100 pF) 1 kHz ~ 5 MHz	0.000 13	
		5 MHz ~ 13 MHz	0.000 27	
		Dissipation Factor		
		(1 pF) 1 kHz ~ 1 MHz	0.000 12	
		(1 ~ 5) MHz	0.000 23	
		(5 ~ 13) MHz	0.000 84	
		(10 pF) 1 kHz ~ 13 MHz	0.000 15	
		(100 pF) 1 kHz ~ 5 MHz	0.000 13	
		(5 ~ 13) MHz	0.000 27	
		(1 000 pF) 1 kHz ~ 1 MHz	0.000 12	
		(1 ~ 5) MHz	0.000 24	
		(5 ~ 13) MHz	0.000 86	
		(10 nF) 120 Hz ~ 100 kHz	0.000 46	
		(100 nF) 120 Hz ~ 100 kHz	0.000 58	
		(1 μF) 120 Hz ~ 100 kHz	0.000 81	
		Resistance		
		(1 mΩ) 50 Hz	6.0×10^{-3}	
		50 Hz ~ 1 kHz	5.0×10^{-3}	
		(10 mΩ) 50 Hz	1.0×10^{-3}	
		50 Hz ~ 1 kHz	5.2×10^{-4}	
		(100 mΩ) 50 Hz	7.1×10^{-4}	
		50 Hz ~ 1 kHz	3.3×10^{-4}	
		(1 Ω) 50 Hz	6.8×10^{-4}	
		50 Hz ~ 1 kHz	1.3×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Impedance bridges/LCR meters	40217	Resistance		Standard Capacitor Set, Standard Resistor Set, Standard Inductor/ SICT-CP-40217
		(10 Ω)		
		50 Hz	9.1×10^{-4}	
		50 Hz ~ 1 kHz	1.3×10^{-4}	
		1 kHz ~ 1 MHz	3.3×10^{-4}	
		(1 ~ 2) MHz	5.2×10^{-4}	
		(2 ~ 3) MHz	6.1×10^{-4}	
		(3 ~ 4) MHz	7.1×10^{-4}	
		(4 ~ 5) MHz	1.0×10^{-3}	
		(5 ~ 10) MHz	4.0×10^{-3}	
		(10 ~ 13) MHz	6.0×10^{-3}	
		(100 Ω)		
		50 Hz	6.2×10^{-4}	
		50 Hz ~ 1 kHz	1.3×10^{-4}	
		1 kHz ~ 1 MHz	3.3×10^{-4}	
		(1 ~ 2) MHz	4.2×10^{-4}	
		(2 ~ 5) MHz	5.2×10^{-4}	
		(5 ~ 10) MHz	2.0×10^{-3}	
		(10 ~ 13) MHz	3.0×10^{-3}	
		(1 kΩ)		
		1 kHz	1.3×10^{-4}	
		1 kHz ~ 3 MHz	3.3×10^{-4}	
		(3 ~ 4) MHz	4.2×10^{-4}	
		(4 ~ 5) MHz	5.2×10^{-4}	
		(5 ~ 10) MHz	2.0×10^{-3}	
		(10 ~ 13) MHz	3.0×10^{-3}	
		(10 kΩ)		
		1 kHz	1.3×10^{-4}	
		(1 ~ 100) kHz	2.4×10^{-4}	
		100 kHz ~ 1 MHz	3.3×10^{-4}	
		(100 kΩ)		
		1 kHz	1.4×10^{-4}	
		100 kHz ~ 1 MHz	3.3×10^{-4}	
		(1 MΩ)		
		1 kHz	3.0×10^{-4}	
		(10 MΩ)		
		1 kHz	2.9×10^{-3}	
		Inductance		
		(1 kHz)		
		100 μH	1.9×10^{-4}	
		1 mH	1.3×10^{-4}	
		10 mH	1.3×10^{-4}	
		100 mH	1.3×10^{-4}	
		1 H	1.3×10^{-4}	
		10 H	1.3×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Impedance bridges/LCR meters	40217	(±)		Standard Capacitor Set, Standard Resistor Set, Standard Inductor/ SICT-CP-40217
DC Bias		0 μV	0.68 μV	
		0 μV ~ 100 mV	1.1×10^{-5}	
		(0.1 ~ 1) V	7.5×10^{-6}	
		(1 ~ 10) V	7.2×10^{-6}	
		(10 ~ 100) V	8.2×10^{-6}	
DC Current		0 μA	5.8 μA	
		0 μA ~ 200 mA	4.1×10^{-5}	
		(0.2 ~ 2) A	3.6×10^{-5}	
		(2 ~ 20) A	1.9×10^{-4}	
	(20 ~ 100) A	1.4×10^{-4}		

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters AC Current	40302	(900 μ A ~ 1 mA)		Power Calibrator, Calibrator/ SICT-CP-40302
		40 Hz ~ 1 kHz	6.3×10^{-4}	
		(1 ~ 5) kHz	7.2×10^{-4}	
		(5 ~ 10) kHz	2.1×10^{-3}	
		(1 ~ 3) mA		
		40 Hz ~ 1 kHz	3.8×10^{-4}	
		(1 ~ 5) kHz	7.2×10^{-4}	
		(5 ~ 10) kHz	3.8×10^{-3}	
		(3 ~ 9) mA		
		40 Hz ~ 1 kHz	3.1×10^{-4}	
		(1 ~ 5) kHz	5.8×10^{-4}	
		(5 ~ 10) kHz	3.1×10^{-3}	
		(9 ~ 10) mA		
		40 Hz ~ 1 kHz	6.3×10^{-4}	
		(1 ~ 5) kHz	7.0×10^{-4}	
		(5 ~ 10) kHz	2.0×10^{-3}	
		(10 ~ 30) mA		
		40 Hz ~ 1 kHz	3.6×10^{-4}	
		(1 ~ 5) kHz	6.8×10^{-4}	
		(5 ~ 10) kHz	2.6×10^{-3}	
		(30 ~ 90) mA		
		40 Hz ~ 1 kHz	2.9×10^{-4}	
		(1 ~ 5) kHz	5.4×10^{-4}	
		(5 ~ 10) kHz	2.2×10^{-3}	
(90 ~ 100) mA				
40 Hz ~ 1 kHz	6.3×10^{-4}			
(1 ~ 5) kHz	6.9×10^{-4}			
(5 ~ 10) kHz	1.6×10^{-3}			
(100 ~ 300) mA				
40 Hz	3.3×10^{-4}			
40 Hz ~ 1 kHz	5.7×10^{-4}			
(1 ~ 5) kHz	1.2×10^{-3}			
(5 ~ 10) kHz	9.3×10^{-3}			
(300 ~ 900) mA				
40 Hz	5.1×10^{-4}			
40 Hz ~ 1 kHz	4.8×10^{-4}			
(1 ~ 5) kHz	1.0×10^{-3}			
(5 ~ 10) kHz	8.9×10^{-3}			
(900 mA ~ 1 A)				
40 Hz ~ 1 kHz	7.0×10^{-4}			
(1 ~ 5) kHz	9.0×10^{-4}			
(5 ~ 10) kHz	5.1×10^{-3}			

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters AC Current	40302	(1 ~ 2) A		Power Calibrator, Calibrator/ SICT-CP-40302
		40 Hz ~ 1 kHz	4.3×10^{-4}	
		(1 ~ 5) kHz	6.6×10^{-4}	
		(5 ~ 10) kHz	4.6×10^{-3}	
		(2 ~ 3) A		
		(40 ~ 100) Hz	5.7×10^{-4}	
		100 Hz ~ 5 kHz	9.7×10^{-4}	
		(5 ~ 10) kHz	4.5×10^{-3}	
		(3 ~ 9) A		
		(40 ~ 100) Hz	4.4×10^{-4}	
		100 Hz ~ 5 kHz	8.2×10^{-4}	
		(5 ~ 10) kHz	4.4×10^{-3}	
		(9 ~ 10) A		
		(40 ~ 100) Hz	6.4×10^{-4}	
		100 Hz ~ 5 kHz	8.5×10^{-4}	
		(5 ~ 10) kHz	4.3×10^{-3}	
		(10 ~ 30) A		
		(40 ~ 100) Hz	3.6×10^{-4}	
		100 Hz ~ 5 kHz	7.0×10^{-4}	
		(5 ~ 10) kHz	6.2×10^{-2}	
		(30 ~ 90) A		
		(40 ~ 100) Hz	2.5×10^{-4}	
		100 Hz ~ 5 kHz	6.4×10^{-4}	
		(5 ~ 10) kHz	5.5×10^{-2}	
(90 ~ 100) A				
(40 ~ 100) Hz	1.8×10^{-4}			
100 Hz ~ 5 kHz	5.4×10^{-4}			
(5 ~ 10) kHz	4.3×10^{-2}			
(100 ~ 1 000) A				
40 Hz	1.3×10^{-3}			
40 Hz ~ 1 kHz	2.5×10^{-3}			
(1 000 ~ 2 500) A				
(40 ~ 60) Hz	1.3×10^{-3}			
(2 500 ~ 3 000) A				
60 Hz	1.3×10^{-3}			
(3 000 ~ 10 000) A				
60 Hz	3.6×10^{-4}			

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters DC Current	40302	0 μA	9.2 nA	Power Calibrator, Calibrator/ SICT-CP-40302
		(0 ~ 0.1) μA	9.2×10^{-2}	
		(0.1 ~ 0.2) μA	4.6×10^{-2}	
		(0.2 ~ 0.3) μA	3.1×10^{-2}	
		(0.3 ~ 0.9) μA	2.3×10^{-2}	
		(0.9 ~ 1) μA	9.3×10^{-3}	
		(1 ~ 2) μA	4.7×10^{-3}	
		(2 ~ 3) μA	3.1×10^{-3}	
		(3 ~ 9) μA	2.3×10^{-3}	
		(9 ~ 10) μA	9.6×10^{-4}	
		(10 ~ 20) μA	5.1×10^{-4}	
		(20 ~ 30) μA	3.5×10^{-4}	
		(30 ~ 50) μA	2.7×10^{-4}	
		(50 ~ 90) μA	1.9×10^{-4}	
		(90 ~ 100) μA	6.2×10^{-4}	
		(100 ~ 200) μA	3.2×10^{-4}	
		(200 ~ 700) μA	2.2×10^{-4}	
		(700 ~ 900) μA	9.3×10^{-5}	
		(0.9 ~ 1) mA	6.1×10^{-4}	
		(1 ~ 2) mA	3.1×10^{-4}	
		(2 ~ 7) mA	2.2×10^{-4}	
		(7 ~ 9) mA	9.1×10^{-5}	
		(9 ~ 10) mA	6.1×10^{-4}	
		(10 ~ 20) mA	3.1×10^{-4}	
		(20 ~ 70) mA	2.2×10^{-4}	
		(70 ~ 90) mA	9.9×10^{-5}	
		(90 ~ 100) mA	6.1×10^{-4}	
		(100 ~ 200) mA	3.1×10^{-4}	
		(200 ~ 700) mA	2.5×10^{-4}	
		(700 ~ 900) mA	1.3×10^{-4}	
		(0.9 ~ 1) A	6.4×10^{-4}	
		(1 ~ 2) A	3.4×10^{-4}	
		(2 ~ 3) A	4.5×10^{-4}	
		(3 ~ 7) A	3.6×10^{-4}	
		(7 ~ 9) A	2.2×10^{-4}	
		(9 ~ 10) A	6.4×10^{-4}	
		(10 ~ 30) A	3.4×10^{-4}	
		(30 ~ 70) A	2.6×10^{-4}	
		(70 ~ 100) A	1.7×10^{-4}	
		(100 ~ 2 500) A	1.3×10^{-3}	
AC Voltage	40302	(1 mV)		
		40 Hz ~ 10 kHz	4.8×10^{-3}	
		(10 ~ 50) kHz	5.0×10^{-3}	
		(50 ~ 100) kHz	6.5×10^{-3}	
		(1 ~ 2) mV		
		40 Hz ~ 10 kHz	2.4×10^{-3}	
		(10 ~ 50) kHz	2.6×10^{-3}	
		(50 ~ 100) kHz	3.5×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters AC Voltage	40302	(2 ~ 5) mV		Power Calibrator, Calibrator/ SICT-CP-40302
		40 Hz ~ 10 kHz	1.7×10^{-3}	
		(10 ~ 50) kHz	1.9×10^{-3}	
		(50 ~ 100) kHz	2.7×10^{-3}	
		(5 ~ 7) mV		
		40 Hz ~ 10 kHz	8.9×10^{-4}	
		(10 ~ 50) kHz	1.0×10^{-3}	
		(50 ~ 100) kHz	1.6×10^{-3}	
		(7 ~ 9) mV		
		40 Hz ~ 10 kHz	6.9×10^{-4}	
		(10 ~ 50) kHz	8.4×10^{-4}	
		(50 ~ 100) kHz	1.4×10^{-3}	
		(9 ~ 10) mV		
		40 Hz ~ 10 kHz	8.3×10^{-4}	
		(10 ~ 50) kHz	9.4×10^{-4}	
		(50 ~ 100) kHz	1.3×10^{-3}	
		(10 ~ 30) mV		
		40 Hz ~ 10 kHz	4.5×10^{-4}	
		(10 ~ 50) kHz	5.6×10^{-4}	
		(50 ~ 100) kHz	1.1×10^{-3}	
		(30 ~ 50) mV		
		40 Hz ~ 10 kHz	3.2×10^{-4}	
		(10 ~ 50) kHz	4.0×10^{-4}	
		(50 ~ 100) kHz	9.0×10^{-4}	
(50 ~ 70) mV				
40 Hz ~ 10 kHz	2.3×10^{-4}			
(10 ~ 50) kHz	3.1×10^{-4}			
(50 ~ 100) kHz	7.1×10^{-4}			
(70 ~ 90) mV				
40 Hz ~ 10 kHz	1.9×10^{-4}			
(10 ~ 50) kHz	2.6×10^{-4}			
(50 ~ 100) kHz	6.2×10^{-4}			
(90 ~ 100) mV				
40 Hz ~ 10 kHz	1.6×10^{-4}			
(10 ~ 50) kHz	2.4×10^{-4}			
(50 ~ 100) kHz	5.7×10^{-4}			
(100 ~ 200) mV				
40 Hz ~ 10 kHz	1.1×10^{-4}			
(10 ~ 50) kHz	1.8×10^{-4}			
(50 ~ 100) kHz	4.6×10^{-4}			

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters AC Voltage	40302	(200 ~ 500) mV		Power Calibrator, Calibrator/ SICT-CP-40302
		40 Hz ~ 10 kHz	8.8×10^{-5}	
		(10 ~ 50) kHz	1.4×10^{-4}	
		(50 ~ 100) kHz	2.2×10^{-4}	
		(500 ~ 700) mV		
		40 Hz ~ 10 kHz	7.1×10^{-5}	
		(10 ~ 50) kHz	1.0×10^{-4}	
		(50 ~ 100) kHz	1.6×10^{-4}	
		(700 ~ 900) mV		
		40 Hz ~ 10 kHz	6.5×10^{-5}	
		(10 ~ 50) kHz	9.5×10^{-5}	
		(50 ~ 100) kHz	1.4×10^{-4}	
		(900 mV ~ 1 V)		
		40 Hz ~ 10 kHz	8.6×10^{-5}	
		(10 ~ 50) kHz	1.1×10^{-4}	
		(50 ~ 100) kHz	1.5×10^{-4}	
		(1 ~ 2) V		
		40 Hz ~ 10 kHz	6.4×10^{-5}	
		(10 ~ 50) kHz	9.0×10^{-5}	
		(50 ~ 100) kHz	1.2×10^{-4}	
		(2 ~ 5) V		
		40 Hz	1.3×10^{-4}	
		40 Hz ~ 10 kHz	9.7×10^{-5}	
		(10 ~ 50) kHz	1.5×10^{-4}	
(50 ~ 100) kHz	2.2×10^{-4}			
(5 ~ 7) V				
40 Hz	8.2×10^{-5}			
40 Hz ~ 10 kHz	6.8×10^{-5}			
(10 ~ 50) kHz	1.1×10^{-4}			
(50 ~ 100) kHz	1.5×10^{-4}			
(7 ~ 9) V				
40 Hz	7.0×10^{-5}			
40 Hz ~ 10 kHz	6.2×10^{-5}			
(10 ~ 50) kHz	9.9×10^{-5}			
(50 ~ 100) kHz	1.3×10^{-4}			
(9 ~ 10) V				
40 Hz	8.9×10^{-5}			
40 Hz ~ 10 kHz	8.4×10^{-5}			
(10 ~ 50) kHz	1.1×10^{-4}			
(50 ~ 100) kHz	1.4×10^{-4}			
(10 ~ 20) V				
40 Hz ~ 10 kHz	6.2×10^{-5}			
(10 ~ 50) kHz	9.0×10^{-5}			
(50 ~ 100) kHz	1.1×10^{-4}			

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters AC Voltage	40302	(20 ~ 50) V		Power Calibrator, Calibrator/ SICT-CP-40302
		40 Hz ~ 10 kHz	1.6×10^{-4}	
		(10 ~ 50) kHz	1.7×10^{-4}	
		(50 ~ 100) kHz	3.4×10^{-4}	
		(50 ~ 70) V		
		40 Hz	9.8×10^{-5}	
		40 Hz ~ 10 kHz	8.3×10^{-5}	
		(10 ~ 50) kHz	1.2×10^{-4}	
		(50 ~ 100) kHz	2.4×10^{-4}	
		(70 ~ 90) V		
		40 Hz	8.5×10^{-5}	
		40 Hz ~ 10 kHz	7.5×10^{-5}	
		(10 ~ 50) kHz	1.1×10^{-4}	
		(50 ~ 100) kHz	2.2×10^{-4}	
		(90 ~ 100) V		
		40 Hz ~ 10 kHz	9.9×10^{-5}	
		(10 ~ 50) kHz	1.2×10^{-4}	
		(50 ~ 100) kHz	2.2×10^{-4}	
		(100 ~ 200) V		
		40 Hz ~ 10 kHz	7.3×10^{-5}	
		(10 ~ 50) kHz	1.0×10^{-4}	
		(50 ~ 100) kHz	1.9×10^{-4}	
		(200 ~ 500) V		
		40 Hz ~ 1 kHz	1.3×10^{-4}	
		(1 ~ 10) kHz	2.2×10^{-4}	
		(10 ~ 20) kHz	6.6×10^{-4}	
		(500 ~ 1 000) V		
		40 Hz ~ 1 kHz	1.5×10^{-4}	
		(1 ~ 10) kHz	2.2×10^{-4}	
		(10 ~ 20) kHz	9.9×10^{-4}	
DC Voltage		0 mV	61 μ V	
		(0 ~ 10) mV	6.1×10^{-3}	
		(10 ~ 20) mV	3.1×10^{-3}	
		(20 ~ 30) mV	2.0×10^{-3}	
		(30 ~ 60) mV	1.5×10^{-3}	
		(60 ~ 70) mV	8.7×10^{-4}	
		(70 ~ 80) mV	7.6×10^{-4}	
		(80 ~ 100) mV	6.8×10^{-4}	
		(100 ~ 200) mV	3.3×10^{-5}	
		(200 ~ 300) mV	2.2×10^{-5}	
		(300 ~ 800) mV	1.7×10^{-5}	
		(800 ~ 900) mV	9.6×10^{-6}	
		(0.9 ~ 1) V	6.1×10^{-5}	
		(1 ~ 2) V	3.1×10^{-5}	
		(2 ~ 3) V	2.1×10^{-5}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters DC Voltage Resistance	40302	(3 ~ 6) V	1.6×10^{-5}	Power Calibrator, Calibrator/ SICT-CP-40302
		(6 ~ 7) V	9.8×10^{-6}	
		(7 ~ 9) V	8.8×10^{-6}	
		(9 ~ 10) V	6.1×10^{-5}	
		(10 ~ 20) V	3.1×10^{-5}	
		(20 ~ 30) V	2.2×10^{-5}	
		(30 ~ 80) V	1.7×10^{-5}	
		(80 ~ 90) V	9.3×10^{-6}	
		(90 ~ 100) V	6.1×10^{-5}	
		(100 ~ 200) V	3.2×10^{-5}	
		(200 ~ 300) V	2.3×10^{-5}	
		(300 ~ 500) V	1.8×10^{-5}	
		(500 ~ 900) V	1.3×10^{-5}	
		(900 ~ 1 000) V	6.2×10^{-5}	
		0 Ω	0.61 mΩ	
		(0 ~ 9) Ω	0.66 mΩ	
		(9 ~ 100) Ω	6.2 mΩ	
(100 ~ 900) Ω	9.2 mΩ			
(0.9 ~ 9) kΩ	92 mΩ			
(9 ~ 90) kΩ	1.1 Ω			
(0.090 ~ 1) MΩ	63 Ω			
(1 ~ 10) MΩ	0.77 kΩ			
(10 ~ 100) MΩ	13 kΩ			
AC voltage/current calibrators AC Voltage	40303	(1 mV)		Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
		10 Hz ~ 10 kHz	1.7×10^{-3}	
		(10 ~ 100) kHz	3.0×10^{-3}	
		100 kHz ~ 1 MHz	1.2×10^{-2}	
		(1 ~ 2) mV		
		10 Hz ~ 10 kHz	1.1×10^{-3}	
		(10 ~ 100) kHz	1.7×10^{-3}	
		100 kHz ~ 1 MHz	7.7×10^{-3}	
		(2 ~ 5) mV		
		10 Hz	6.4×10^{-4}	
		10 Hz ~ 10 kHz	5.8×10^{-4}	
		(10 ~ 100) kHz	1.0×10^{-3}	
		100 kHz ~ 1 MHz	5.4×10^{-3}	
		(5 ~ 10) mV		
		10 Hz	4.2×10^{-4}	
		10 Hz ~ 10 kHz	3.5×10^{-4}	
		(10 ~ 100) kHz	5.8×10^{-4}	
		100 kHz ~ 1 MHz	3.9×10^{-3}	
		(10 ~ 20) mV		
		10 Hz	1.8×10^{-4}	
		10 Hz ~ 10 kHz	1.4×10^{-4}	
		(10 ~ 100) kHz	2.2×10^{-4}	
		100 kHz ~ 1 MHz	2.2×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators AC Voltage	40303	(20 ~ 50) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	1.4×10^{-4} 9.2×10^{-5} 1.6×10^{-4} 1.4×10^{-3}	Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
		(50 ~ 100) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	1.1×10^{-4} 6.6×10^{-5} 1.2×10^{-4} 1.3×10^{-3}	
		(100 ~ 200) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	8.2×10^{-5} 3.9×10^{-5} 7.6×10^{-5} 1.1×10^{-3}	
		(200 ~ 500) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	7.8×10^{-5} 3.6×10^{-5} 7.1×10^{-5} 1.1×10^{-3}	
		(500 mV ~ 1 V) 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	7.6×10^{-5} 3.3×10^{-5} 6.6×10^{-5} 1.1×10^{-3}	
		(1 ~ 2) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	7.1×10^{-5} 2.7×10^{-5} 5.8×10^{-5} 1.0×10^{-3}	
		(2 ~ 5) V 10 Hz 10 Hz ~ 10 kHz 10 kHz ~ 100 kHz 100 kHz ~ 1 MHz	7.2×10^{-5} 2.6×10^{-5} 7.5×10^{-5} 1.4×10^{-3}	
		(5 ~ 20) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	7.2×10^{-5} 2.8×10^{-5} 7.5×10^{-5} 1.4×10^{-3}	
		(20 ~ 50) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	7.2×10^{-5} 3.0×10^{-5} 8.0×10^{-5}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators	40303	(50 ~ 200) V		Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
		10 Hz	7.4×10^{-5}	
AC Voltage	10 Hz ~ 10 kHz	3.3×10^{-5}		
	(10 ~ 100) kHz	8.5×10^{-5}		
	(200 ~ 1 000) V			
	10 Hz	7.7×10^{-5}		
	10 Hz ~ 10 kHz	3.3×10^{-5}		
	(10 ~ 100) kHz	5.8×10^{-4}		
AC Current	(10 μ A)			
	10 Hz ~ 10 kHz	2.6×10^{-3}		
	(10 ~ 100) μ A			
	10 Hz ~ 1 kHz	3.6×10^{-4}		
	(1 ~ 10) kHz	6.4×10^{-4}		
	(100 μ A ~ 1 mA)			
	10 Hz	9.8×10^{-5}		
	10 Hz ~ 1 kHz	7.5×10^{-5}		
	(1 ~ 10) kHz	9.4×10^{-5}		
	(1 ~ 100) mA			
	10 Hz	7.8×10^{-5}		
	10 Hz ~ 1 kHz	4.6×10^{-5}		
	(1 ~ 10) kHz	4.2×10^{-5}		
	(100 mA ~ 1 A)			
	10 Hz	8.1×10^{-5}		
	10 Hz ~ 1 kHz	4.9×10^{-5}		
	(1 ~ 10) kHz	4.4×10^{-5}		
	(1 ~ 2) A			
	10 Hz	7.9×10^{-5}		
	10 Hz ~ 1 kHz	4.7×10^{-5}		
	(1 ~ 10) kHz	4.5×10^{-5}		
	(2 ~ 5) A			
	10 Hz	8.2×10^{-5}		
	10 Hz ~ 1 kHz	5.2×10^{-5}		
	(1 ~ 10) kHz	5.0×10^{-5}		
	(5 ~ 10) A			
	10 Hz	8.6×10^{-5}		
	10 Hz ~ 1 kHz	5.9×10^{-5}		
	(1 ~ 10) kHz	7.8×10^{-5}		
	(10 ~ 20) A			
	10 Hz	9.3×10^{-5}		
	10 Hz ~ 1 kHz	6.8×10^{-5}		
	(1 ~ 10) kHz	7.8×10^{-5}		

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators AC Current	40303	(20 ~ 50) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (50 ~ 100) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (100 ~ 200) A 60 Hz	 1.0×10^{-4} 8.3×10^{-5} 1.1×10^{-4} 1.2×10^{-4} 9.7×10^{-5} 1.3×10^{-4} 4.5×10^{-4}	Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
Wattmeter calibrators AC Voltage	40304	(1 mV) 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (1 ~ 2) mV 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (2 ~ 5) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (5 ~ 10) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (10 ~ 20) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (20 ~ 50) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (50 ~ 100) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	 1.7×10^{-3} 3.0×10^{-3} 1.2×10^{-2} 1.1×10^{-3} 1.7×10^{-3} 7.7×10^{-3} 6.4×10^{-4} 5.8×10^{-4} 1.0×10^{-3} 5.4×10^{-3} 4.2×10^{-4} 3.5×10^{-4} 5.8×10^{-4} 3.9×10^{-3} 1.8×10^{-4} 1.4×10^{-4} 2.2×10^{-4} 2.2×10^{-3} 1.4×10^{-4} 9.2×10^{-5} 1.6×10^{-4} 1.4×10^{-3} 1.1×10^{-4} 6.6×10^{-5} 1.2×10^{-4} 1.3×10^{-3}	Power Standard, Counter/ SICT-CP-40304

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Wattmeter calibrators AC Voltage	40304	(100 ~ 200) mV		Power Standard, Counter/ SICT-CP-40304	
		10 Hz	8.2×10^{-5}		
		10 Hz ~ 10 kHz	3.9×10^{-5}		
		(10 ~ 100) kHz	7.6×10^{-5}		
		100 kHz ~ 1 MHz	1.1×10^{-3}		
		(200 ~ 500) mV			
		10 Hz	7.8×10^{-5}		
		10 Hz ~ 10 kHz	3.6×10^{-5}		
		(10 ~ 100) kHz	7.1×10^{-5}		
		100 kHz ~ 1 MHz	1.1×10^{-3}		
		(500 mV ~ 1 V)			
		10 Hz	7.6×10^{-5}		
		10 Hz ~ 10 kHz	3.3×10^{-5}		
		(10 ~ 100) kHz	6.6×10^{-5}		
		100 kHz ~ 1 MHz	1.1×10^{-3}		
		(1 ~ 2) V			
		10 Hz	7.1×10^{-5}		
		10 Hz ~ 10 kHz	2.7×10^{-5}		
		(10 ~ 100) kHz	5.8×10^{-5}		
		100 kHz ~ 1 MHz	1.0×10^{-3}		
		(2 ~ 5) V			
		10 Hz	7.2×10^{-5}		
		10 Hz ~ 10 kHz	2.6×10^{-5}		
		(10 ~ 100) kHz	7.5×10^{-5}		
		100 kHz ~ 1 MHz	1.4×10^{-3}		
		(5 ~ 20) V			
		10 Hz	7.2×10^{-5}		
		10 Hz ~ 10 kHz	2.8×10^{-5}		
		(10 ~ 100) kHz	7.5×10^{-5}		
		100 kHz ~ 1 MHz	1.4×10^{-3}		
		(20 ~ 50) V			
		10 Hz	7.2×10^{-5}		
		10 Hz ~ 10 kHz	3.0×10^{-5}		
		(10 ~ 100) kHz	8.0×10^{-5}		
		(50 ~ 200) V			
		10 Hz	7.4×10^{-5}		
		10 Hz ~ 10 kHz	3.3×10^{-5}		
		(10 ~ 100) kHz	8.5×10^{-5}		
		(200 ~ 1 000) V			
		10 Hz	7.7×10^{-5}		
10 Hz ~ 10 kHz	3.3×10^{-5}				
(10 ~ 100) kHz	5.8×10^{-4}				

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Wattmeter calibrators AC Current	40304	(10 μ A) 10 Hz ~ 10 kHz	2.6×10^{-3}	Power Standard, Counter/ SICT-CP-40304
		(10 ~ 100) μ A 10 Hz ~ 1 kHz (1 ~ 10) kHz	3.6×10^{-4} 6.4×10^{-4}	
		(100 μ A ~ 1 mA) 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	9.8×10^{-5} 7.5×10^{-5} 9.4×10^{-5}	
		(1 ~ 100) mA 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	7.8×10^{-5} 4.6×10^{-5} 4.2×10^{-5}	
		(100 mA ~ 1 A) 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	8.1×10^{-5} 4.9×10^{-5} 4.4×10^{-5}	
		(1 ~ 2) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	7.9×10^{-5} 4.7×10^{-5} 4.5×10^{-5}	
		(2 ~ 5) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	8.2×10^{-5} 5.2×10^{-5} 5.0×10^{-5}	
		(5 ~ 10) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	8.6×10^{-5} 5.9×10^{-5} 7.8×10^{-5}	
		(10 ~ 20) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	9.3×10^{-5} 6.8×10^{-5} 7.8×10^{-5}	
		(20 ~ 50) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	1.0×10^{-4} 8.3×10^{-5} 1.1×10^{-4}	
		(50 ~ 100) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	1.2×10^{-4} 9.7×10^{-5} 1.3×10^{-4}	
		(100 ~ 200) A 60 Hz	4.5×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Wattmeter calibrators	40304	AC Power	(50 ~ 60) Hz		Power Standard, Counter/ SICT-CP-40304
		0 mW	0.05 mW		
		(0 ~ 0.22) mW	3.2×10^{-1}		
		(0.22 ~ 1.1) mW	6.3×10^{-2}		
		(1.1 ~ 2.2) mW	3.1×10^{-2}		
		(2.2 ~ 11) mW	6.3×10^{-3}		
		(11 ~ 22) mW	3.1×10^{-3}		
		(22 ~ 44) mW	1.6×10^{-3}		
		(44 ~ 66) mW	1.1×10^{-3}		
		(66 ~ 88) mW	8.0×10^{-4}		
		(88 ~ 110) mW	6.4×10^{-4}		
		(110 ~ 480) mW	2.6×10^{-4}		
		(480 ~ 550) mW	2.3×10^{-4}		
		(0.55 ~ 1.1) W	1.4×10^{-4}		
		(1.1 ~ 5.5) W	1.6×10^{-4}		
		5.5 W ~ 1.1 kW	1.4×10^{-4}		
		(1.1 ~ 2.2) kW	1.5×10^{-4}		
		(2.2 ~ 24) kW	1.4×10^{-4}		
Power Factor	(50 ~ 60) Hz				
	(0 ~ 550) mW				
	-1 ~ 1	2.3×10^{-4}			
Harmonic Voltage	550 mW ~ 24 kW				
	-1 ~ 1	1.5×10^{-4}			
	(50 ~ 60) Hz				
Harmonic Current	(0.5 ~ 3) %	0.042 %			
	(3 ~ 10) %	0.052 %			
	(10 ~ 20) %	0.081 %			
Flicker	(50 ~ 60) Hz				
	(0.5 ~ 3) %	0.042 %			
	(3 ~ 10) %	0.052 %			
Frequency	(10 ~ 20) %	0.055 %			
	P_{st} (0.25 ~ 5), (50 Hz)				
	Modulation Frequency				
	8.333 mHz	2.7×10^{-3}			
	16.667 mHz	2.7×10^{-3}			
	58.333 mHz	2.7×10^{-3}			
	325.000 mHz	2.7×10^{-3}			
	916.667 mHz	2.7×10^{-3}			
	13.500 Hz	2.7×10^{-3}			
	33.333 Hz	2.7×10^{-3}			
(10 ~ 100) Hz	7.0×10^{-7}				
(100 ~ 400) Hz	3.8×10^{-7}				
400 Hz ~ 1 MHz	7.0×10^{-7}				

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC current shunts AC Resistance	40305	(100 ~ 200) A (60 Hz) 1 mΩ	8.4×10^{-4}	Reference Multimeter, Calibrator/ SICT-CP-40305
		(40 ~ 60) Hz (1 ~ 10) mΩ (10 ~ 100) mΩ	2.1×10^{-4} 3.5×10^{-4}	
		(60 Hz ~ 1 kHz) (1 ~ 10) mΩ (10 ~ 100) mΩ	5.7×10^{-4} 3.4×10^{-4}	
		(40 Hz ~ 1 kHz) 100 mΩ ~ 100 Ω 100 Ω ~ 10 kΩ	1.8×10^{-4} 2.3×10^{-4}	
AC Voltage dividers		(50 Hz) (1 ~ 100) kV	2.6×10^{-4}	
		(60 Hz) (1 ~ 100) kV	1.9×10^{-4}	
Phase angle generators, synchro resolve generators Phase	40306	(-180 ~ 180) ° 50 Hz (50 ~ 500) Hz (500 ~ 1 000) Hz	0.0016° 0.0031° 0.010°	전력 교정기/ SICT-CP-40307
Voltage/current phase angle meters/synchro resolve meters Phase	40307	(50 ~ 60) Hz (-180 ~ 180) °	0.0088°	Power Calibrator/ SICT-CP-40307
Potential transformer test set Ratio	40308	(110 ~ 1 100) V (-19.99 ~ 19.99) % (-680 ~ 680)'	0.020 % 0.70'	Standard Potential transforme, Ratio transformers/ SICT-CP-40308
		(1 100 ~ 22 900) V (-19.99 ~ 19.99) %	0.016 %	
		Phase (-680 ~ 680)'	0.50'	
Potential transforme Ratio	40309	(110 ~ 1 100) V (-19.99 ~ 19.99) % (-680 ~ 680)'	0.020 % 0.70'	Standard Potential transforme/ SICT-CP-40309
		(1 100 ~ 22 900) V (-19.99 ~ 19.99) %	0.016 %	
		Phase (-680 ~ 680)'	0.50'	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Power factor meters AC Power Factor	40310	(50 Hz, 60 Hz) -1 ~ 1	1.1×10^{-4}	Power Calibrator/ SICT-CP-40310
AC power meters AC Voltage	40311	(1 mV) 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (1 ~ 2) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (2 ~ 5) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (5 ~ 7) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (7 ~ 9) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (9 ~ 10) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (10 ~ 30) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (30 ~ 60) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (60 ~ 200) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	4.8×10^{-3} 5.0×10^{-3} 6.5×10^{-3} 2.4×10^{-3} 2.6×10^{-3} 3.5×10^{-3} 1.7×10^{-3} 1.9×10^{-3} 2.7×10^{-3} 8.9×10^{-4} 1.0×10^{-3} 1.6×10^{-3} 6.9×10^{-4} 8.4×10^{-4} 1.4×10^{-3} 5.7×10^{-4} 7.1×10^{-4} 1.2×10^{-3} 3.6×10^{-4} 4.7×10^{-4} 1.1×10^{-3} 2.9×10^{-4} 3.7×10^{-4} 8.8×10^{-4} 1.9×10^{-4} 2.7×10^{-4} 6.5×10^{-4}	Power Calibrator, Calibrator/ SICT-CP-40311

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters AC Voltage	40311	(200 ~ 300) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	8.6×10^{-5} 1.3×10^{-4} 2.2×10^{-4}	Power Calibrator, Calibrator/ SICT-CP-40311
		(300 ~ 600) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	7.6×10^{-5} 1.2×10^{-4} 1.9×10^{-4}	
		(600 mV ~ 1 V) 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	6.7×10^{-5} 9.8×10^{-5} 1.5×10^{-4}	
		(1 ~ 2) V 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	5.6×10^{-5} 8.5×10^{-5} 1.2×10^{-4}	
		(2 ~ 3) V 40 Hz 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	1.3×10^{-4} 9.5×10^{-5} 1.5×10^{-4} 2.2×10^{-4}	
		(3 ~ 5) V 40 Hz 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	1.1×10^{-4} 8.0×10^{-5} 1.3×10^{-4} 1.8×10^{-4}	
		(5 ~ 7) V 40 Hz 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	8.1×10^{-5} 6.7×10^{-5} 1.1×10^{-4} 1.5×10^{-4}	
		(7 ~ 20) V 40 Hz 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	7.0×10^{-5} 6.1×10^{-5} 9.9×10^{-5} 1.3×10^{-4}	
		(20 ~ 60) V 40 Hz 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	1.6×10^{-4} 1.2×10^{-4} 1.6×10^{-4} 3.4×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters AC Voltage	40311	(60 ~ 100) V 40 Hz 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	9.0×10^{-5} 7.8×10^{-5} 1.2×10^{-4} 2.3×10^{-4}	Power Calibrator, Calibrator/ SICT-CP-40311
		(100 ~ 200) V 40 Hz 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	6.7×10^{-5} 6.5×10^{-5} 9.9×10^{-5} 1.9×10^{-4}	
		(200 ~ 400) V 40 Hz 40 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 20) kHz	1.4×10^{-4} 1.1×10^{-4} 2.2×10^{-4} 6.6×10^{-4}	
		(400 ~ 500) V 40 Hz 40 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 20) kHz	1.3×10^{-4} 1.1×10^{-4} 2.1×10^{-4} 5.4×10^{-4}	
		(500 ~ 600) V 40 Hz 40 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 20) kHz	1.5×10^{-4} 1.3×10^{-4} 2.2×10^{-4} 9.9×10^{-4}	
		(600 ~ 700) V 40 Hz 40 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 20) kHz	1.4×10^{-4} 1.2×10^{-4} 2.2×10^{-4} 8.5×10^{-4}	
		(700 ~ 900) V 40 Hz 40 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 20) kHz	1.3×10^{-4} 1.1×10^{-4} 2.1×10^{-4} 7.5×10^{-4}	
		(900 ~ 1 000) V 40 Hz 40 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 20) kHz	1.2×10^{-4} 1.0×10^{-4} 2.1×10^{-4} 6.1×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters AC Current	40311	(100 μ A) 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	2.3×10^{-4} 4.9×10^{-4} 2.1×10^{-3}	Power Calibrator, Calibrator/ SICT-CP-40311
		(100 ~ 300) μ A 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	2.3×10^{-4} 4.6×10^{-4} 1.8×10^{-3}	
		(300 μ A ~ 2 mA) 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	1.9×10^{-4} 3.8×10^{-4} 2.1×10^{-3}	
		(2 ~ 4) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	3.2×10^{-4} 6.9×10^{-4} 3.8×10^{-3}	
		(4 ~ 7) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	2.3×10^{-4} 4.8×10^{-4} 2.7×10^{-3}	
		(7 ~ 20) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	1.9×10^{-4} 3.7×10^{-4} 2.1×10^{-3}	
		(20 ~ 30) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	2.9×10^{-4} 6.5×10^{-4} 2.6×10^{-3}	
		(30 ~ 60) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	2.4×10^{-4} 5.2×10^{-4} 2.2×10^{-3}	
		(60 ~ 200) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	1.8×10^{-4} 3.7×10^{-4} 1.7×10^{-3}	
		(200 ~ 300) mA 40 Hz (40 ~ 60) Hz 60 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	5.8×10^{-4} 1.2×10^{-4} 5.3×10^{-4} 1.2×10^{-3} 9.3×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters AC Current	40311	(300 ~ 500) mA 40 Hz (40 ~ 60) Hz 60 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (500 ~ 800) mA 40 Hz (40 ~ 60) Hz 60 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (800 mA ~ 2 A) 40 Hz (40 ~ 60) Hz 60 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (2 ~ 3) A 40 Hz (40 ~ 60) Hz (60 ~ 100) Hz 100 Hz ~ 5 kHz (5 ~ 10) kHz (3 ~ 6) A 40 Hz (40 ~ 60) Hz (60 ~ 100) Hz 100 Hz ~ 5 kHz (5 ~ 10) kHz (6 ~ 10) A 40 Hz (40 ~ 60) Hz (60 ~ 100) Hz 100 Hz ~ 5 kHz (5 ~ 10) kHz (10 ~ 50) A 40 Hz (40 ~ 60) Hz (60 ~ 100) Hz 100 Hz ~ 5 kHz (5 ~ 10) kHz (50 ~ 100) A (40 ~ 100) Hz 100 Hz ~ 5 kHz (5 ~ 10) kHz	 4.9×10^{-4} 1.1×10^{-4} 4.6×10^{-4} 1.0×10^{-3} 8.9×10^{-3} 4.0×10^{-4} 1.3×10^{-4} 3.9×10^{-4} 8.3×10^{-4} 8.6×10^{-3} 3.5×10^{-4} 9.5×10^{-5} 3.5×10^{-4} 7.2×10^{-4} 8.4×10^{-3} 5.3×10^{-4} 1.5×10^{-4} 5.3×10^{-4} 9.4×10^{-4} 4.5×10^{-3} 4.1×10^{-4} 1.3×10^{-4} 4.1×10^{-4} 8.1×10^{-4} 4.4×10^{-3} 2.6×10^{-4} 1.1×10^{-4} 2.6×10^{-4} 6.4×10^{-4} 4.3×10^{-3} 2.4×10^{-4} 1.1×10^{-4} 2.4×10^{-4} 6.6×10^{-4} 6.2×10^{-2} 1.9×10^{-4} 5.7×10^{-4} 4.8×10^{-2}	Power Calibrator, Calibrator/ SICT-CP-40311

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311	(100 ~ 1 000) A		Power Calibrator, Calibrator/ SICT-CP-40311
AC Current		(40 ~ 100) Hz	1.3×10^{-3}	
		100 Hz ~ 1 kHz	2.7×10^{-3}	
		(1 000 ~ 2 500) A		
		(40 ~ 60) Hz	1.2×10^{-3}	
		(2 500 ~ 3 000) A		
		60 Hz	1.3×10^{-3}	
AC Wattage		(50 ~ 60) Hz		
		0 mW	70 μ W	
		(0 ~ 0.22) mW	2.1×10^{-1}	
		(0.22 ~ 1.1) mW	4.1×10^{-2}	
		(1.1 ~ 2.2) mW	2.1×10^{-2}	
		(2.2 ~ 11) mW	4.1×10^{-3}	
		(11 ~ 22) mW	2.1×10^{-3}	
		(22 ~ 44) mW	1.0×10^{-3}	
		(44 ~ 66) mW	7.0×10^{-4}	
		(66 ~ 88) mW	5.3×10^{-4}	
		(88 ~ 110) mW	4.3×10^{-4}	
		(110 ~ 480) mW	2.1×10^{-4}	
		480 mW ~ 12 kW	1.2×10^{-4}	
		(12 ~ 24) kW	6.8×10^{-4}	
		(24 ~ 300) kW	1.2×10^{-3}	
		(300 ~ 600) kW	1.4×10^{-3}	
DC Voltage		0 mV	0.78 μ V	
		(0 ~ 1) mV	8.0×10^{-4}	
		(1 ~ 2) mV	4.0×10^{-4}	
		(2 ~ 3) mV	2.7×10^{-4}	
		(3 ~ 4) mV	2.0×10^{-4}	
		(4 ~ 8) mV	1.6×10^{-4}	
		(8 ~ 9) mV	9.4×10^{-5}	
		(9 ~ 10) mV	8.5×10^{-5}	
		(10 ~ 20) mV	4.5×10^{-5}	
		(20 ~ 30) mV	3.2×10^{-5}	
		(30 ~ 60) mV	2.6×10^{-5}	
		(60 ~ 200) mV	1.8×10^{-5}	
		(200 ~ 300) mV	9.3×10^{-6}	
		(300 ~ 400) mV	8.3×10^{-6}	
		(400 ~ 700) mV	7.8×10^{-6}	
		(700 ~ 900) mV	6.9×10^{-6}	
		(0.9 ~ 1) V	9.0×10^{-6}	
		(1 ~ 2) V	7.2×10^{-6}	
		(2 ~ 3) V	5.6×10^{-6}	
		(3 ~ 9) V	5.1×10^{-6}	
		(9 ~ 10) V	7.5×10^{-6}	
		(10 ~ 20) V	6.0×10^{-6}	
		(20 ~ 40) V	7.9×10^{-6}	
		(40 ~ 90) V	6.9×10^{-6}	
		(90 ~ 200) V	8.8×10^{-6}	
		(200 ~ 400) V	9.9×10^{-6}	
		(400 ~ 900) V	8.7×10^{-6}	
		(900 ~ 1 000) V	1.0×10^{-5}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
AC power meters	40311			Power Calibrator, Calibrator/ SICT-CP-40311	
DC Current		0 μ A	9.0 nA		
		(0 ~ 0.1) μ A	9.0×10^{-2}		
		(0.1 ~ 0.2) μ A	4.5×10^{-2}		
		(0.2 ~ 0.3) μ A	3.0×10^{-2}		
		(0.3 ~ 0.4) μ A	2.3×10^{-2}		
		(0.4 ~ 0.9) μ A	1.8×10^{-2}		
		(0.9 ~ 1) μ A	9.1×10^{-3}		
		(1 ~ 2) μ A	4.5×10^{-3}		
		(2 ~ 3) μ A	3.0×10^{-3}		
		(3 ~ 4) μ A	2.3×10^{-3}		
		(4 ~ 9) μ A	1.8×10^{-3}		
		(9 ~ 10) μ A	9.4×10^{-4}		
		(10 ~ 20) μ A	5.0×10^{-4}		
		(20 ~ 30) μ A	3.4×10^{-4}		
		(30 ~ 50) μ A	2.7×10^{-4}		
		(50 ~ 90) μ A	1.9×10^{-4}		
		(90 ~ 100) μ A	1.3×10^{-4}		
		(100 ~ 200) μ A	8.6×10^{-5}		
		(200 ~ 300) μ A	6.5×10^{-5}		
		(300 ~ 500) μ A	5.5×10^{-5}		
		(0.5 ~ 2) mA	5.1×10^{-5}		
		(2 ~ 3) mA	7.3×10^{-5}		
		(3 ~ 4) mA	6.3×10^{-5}		
		(4 ~ 7) mA	5.7×10^{-5}		
		(7 ~ 20) mA	4.9×10^{-5}		
		(20 ~ 30) mA	8.2×10^{-5}		
		(30 ~ 50) mA	7.4×10^{-5}		
		(50 ~ 100) mA	6.6×10^{-5}		
		(100 ~ 200) mA	5.7×10^{-5}		
		(200 ~ 600) mA	1.5×10^{-4}		
		(0.6 ~ 2) A	1.1×10^{-4}		
		(2 ~ 3) A	4.0×10^{-4}		
		(3 ~ 4) A	3.2×10^{-4}		
		(4 ~ 7) A	2.8×10^{-4}		
		(7 ~ 10) A	2.1×10^{-4}		
		(10 ~ 20) A	1.5×10^{-4}		
		(20 ~ 30) A	2.5×10^{-4}		
		(30 ~ 80) A	2.1×10^{-4}		
		(80 ~ 100) A	1.5×10^{-4}		
		(100 ~ 2 500) A	1.3×10^{-3}		
DC Wattage			0 mW		61 nW
			(0 ~ 1) mW		7.7×10^{-5}
			(1 ~ 10) mW		4.8×10^{-5}
			(10 ~ 100) mW		6.1×10^{-5}
			(0.1 ~ 100) W		1.1×10^{-4}
			(0.1 ~ 10) kW		1.9×10^{-4}
		(10 ~ 20) kW	1.5×10^{-4}		
		(20 ~ 1 000) kW	1.0×10^{-3}		
		(1 000 ~ 2 500) kW	1.3×10^{-3}		

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311			Power Calibrator, Calibrator/ SICT-CP-40311
Harmonic Voltage		(50 ~ 60) Hz (0.5 ~ 3) % (3 ~ 5) % (5 ~ 10) % (10 ~ 20) %	0.030 % 0.033 % 0.042 % 0.065 %	
Harmonic Current		(50 ~ 60) Hz (0.5 ~ 3) % (3 ~ 10) % (5 ~ 20) %	0.030 % 0.032 % 0.038 %	
Flicker		P_{st} (0.25 ~ 5), (50 Hz) Modulation Frequency 8.333 mHz 16.667 mHz 58.333 mHz 325 mHz 916.667 mHz 13.5 Hz 33.333 Hz	3.5×10^{-2} 3.5×10^{-2} 3.5×10^{-2} 3.5×10^{-2} 3.5×10^{-2} 3.5×10^{-2} 3.5×10^{-2}	
Frequency		10 Hz ~ 10 MHz	1.3×10^{-4}	
Power Factor		(50 ~ 60) Hz -1 ~ 1	1.1×10^{-4}	
Current burden		(50 ~ 60) Hz 1.25 VA (1.25 ~ 3.75) VA (3.75 ~ 5) VA (5 ~ 10) VA (10 ~ 100) VA	2.9×10^{-3} 1.5×10^{-3} 8.0×10^{-4} 5.1×10^{-4} 4.0×10^{-4}	
Current burden factor		0.5 ~ 1	2.2×10^{-4}	
Voltage burden		(50 ~ 60) Hz 1.25 VA (1.25 ~ 3.75) VA (3.75 ~ 5) VA (5 ~ 10) VA (10 ~ 100) VA	7.4×10^{-4} 4.6×10^{-4} 3.7×10^{-4} 3.0×10^{-4} 3.1×10^{-4}	
Voltage burden factor		0.5 ~ 1	1.6×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
AC power supplies	AC Voltage	(10 mV) 40 Hz ~ 5 kHz	2.2×10^{-4}	Voltage Standard, Multimeter, Current Shunt/ SICT-CP-40312	
		(10 ~ 100) mV 40 Hz ~ 5 kHz	7.9×10^{-5}		
		(100 mV ~ 1 V) 40 Hz ~ 5 kHz	6.7×10^{-5}		
		(1 ~ 10) V 40 Hz ~ 5 kHz	6.8×10^{-5}		
		(10 ~ 100) V 40 Hz ~ 5 kHz	7.0×10^{-5}		
		(100 ~ 600) V 40 Hz ~ 5 kHz	4.1×10^{-5}		
		(600 ~ 1 000) V 40 Hz ~ 5 kHz	7.1×10^{-5}		
		Frequency	10 Hz		9.5×10^{-5}
			(10 ~ 50) Hz		1.9×10^{-5}
			(50 ~ 100) Hz		7.7×10^{-6}
			(0.1 ~ 1) kHz		8.4×10^{-7}
			(1 ~ 5) kHz		3.8×10^{-7}
		AC Current	(1 mA) (50 ~ 60) Hz		6.4×10^{-4}
	(1 ~ 10) mA (50 ~ 60) Hz		3.6×10^{-4}		
	(10 ~ 100) mA (50 ~ 60) Hz		2.4×10^{-4}		
	(100 mA ~ 1 A) (50 ~ 60) Hz		2.1×10^{-4}		
	(1 ~ 10) A (50 ~ 60) Hz		2.3×10^{-4}		
	(10 ~ 20) A (50 ~ 60) Hz		4.0×10^{-4}		
	(20 ~ 30) A (50 ~ 60) Hz		6.4×10^{-4}		
	(30 ~ 50) A (50 ~ 60) Hz		4.2×10^{-4}		

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power supplies	40312			Voltage Standard, Multimeter, Current Shunt/ SICT-CP-40312
DC Voltage		(±) 0 mV (0 ~ 10) mV (10 ~ 100) mV (0.1 ~ 100) V (100 ~ 600) V (600 ~ 1 000) V	5.8 μV 5.8×10^{-4} 5.8×10^{-5} 7.7×10^{-6} 1.3×10^{-5} 6.6×10^{-5}	
DC Current		(1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A (10 ~ 300) A (300 ~ 500) A (500 ~ 1 000) A (1 000 ~ 3 000) A	5.8×10^{-3} 5.9×10^{-4} 2.4×10^{-4} 3.1×10^{-4} 2.4×10^{-4} 2.6×10^{-4} 4.7×10^{-5} 5.1×10^{-4}	
Load Regulation		(0 ~ 2) mV (2 ~ 20) mV (20 ~ 200) mV	0.16 mV 7.8×10^{-2} 8.2×10^{-3}	
Ripple		(0.1 ~ 0.4) mV (0.4 ~ 0.6) mV (0.6 ~ 1) mV (1 ~ 10) mV (10 ~ 50) mV	3.8×10^{-1} 1.1×10^{-1} 7.3×10^{-2} 4.4×10^{-2} 7.1×10^{-2}	
Harmonic Voltage		(50 ~ 60) Hz 0.5 % (0.5 ~ 10) % (10 ~ 20) %	0.050 % 0.051 % 0.082 %	
Harmonic Current		(50 ~ 60) Hz 0.5 % (0.5 ~ 20) %	0.050 % 0.051 %	
Puncture/safety testers	40313			AC/DC Kilovoltmeter, High Voltage Digital Meter, Reference Multimeter/ SICT-CP-40313
DC Voltage		(±) 0 kV (0 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 2) kV (2 ~ 100) kV (100 ~ 200) kV	0.58 V 1.2×10^{-3} 6.1×10^{-4} 3.0×10^{-4} 2.3×10^{-4} 1.2×10^{-2}	
AC Voltage		(50 ~ 60) Hz 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 100) kV (100 ~ 200) kV	0.58 V 1.2×10^{-3} 6.2×10^{-4} 5.7×10^{-4} 1.2×10^{-2}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Puncture/safety testers	40313			AC/DC Kilovoltmeter, High Voltage Digital Meter, Reference Multimeter/ SICT-CP-40313
AC Breaking Current		(50 ~ 60) Hz		
		0.1 mA	5.3×10^{-4}	
		(0.1 ~ 0.5) mA	4.4×10^{-4}	
		(0.5 ~ 1) mA	7.3×10^{-4}	
		(1 ~ 2) mA	7.1×10^{-4}	
		(2 ~ 5) mA	4.4×10^{-4}	
		(5 ~ 10) mA	3.6×10^{-4}	
		(10 ~ 20) mA	7.1×10^{-4}	
		(20 ~ 50) mA	4.4×10^{-4}	
		(50 ~ 100) mA	7.3×10^{-4}	
DC Breaking Current		0.1 mA	3.9×10^{-4}	
		(0.1 ~ 0.5) mA	1.3×10^{-4}	
		(0.5 ~ 1) mA	6.4×10^{-4}	
		(1 ~ 2) mA	3.2×10^{-4}	
		(2 ~ 5) mA	1.3×10^{-4}	
		(5 ~ 10) mA	6.5×10^{-5}	
		(10 ~ 20) mA	3.3×10^{-4}	
		(20 ~ 50) mA	1.4×10^{-4}	
		(50 ~ 100) mA	6.4×10^{-4}	
Resistance		1 mΩ	8.6×10^{-4}	
		(1 ~ 10) mΩ	7.2×10^{-4}	
		10 mΩ ~ 100 kΩ	6.8×10^{-4}	
Insulation Voltage		1 V	6.4×10^{-4}	
		(1 ~ 10) V	6.4×10^{-5}	
		(10 ~ 25) V	2.5×10^{-4}	
		(25 ~ 50) V	1.3×10^{-4}	
		(50 ~ 100) V	6.4×10^{-5}	
		(100 ~ 250) V	2.5×10^{-4}	
		(250 ~ 500) V	1.3×10^{-4}	
		(500 ~ 1 000) V	6.4×10^{-5}	
		(1 000 ~ 5 000) V	6.5×10^{-3}	
		(5 000 ~ 10 000) V	6.1×10^{-3}	
Insulation Resistance		1 kΩ	7.1×10^{-5}	
		(1 ~ 10) kΩ	3.7×10^{-5}	
		(10 ~ 100) kΩ	2.5×10^{-5}	
		(0.1 ~ 1) MΩ	3.1×10^{-5}	
		(1 ~ 10) MΩ	9.5×10^{-5}	
		(10 ~ 100) MΩ	2.4×10^{-5}	
		(0.1 ~ 1) GΩ	3.1×10^{-5}	
		(1 ~ 10) GΩ	6.1×10^{-5}	
		(10 ~ 100) GΩ	1.3×10^{-4}	
		(0.1 ~ 1) TΩ	2.6×10^{-4}	
		10 TΩ	6.3×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Puncture/safety testers	40313			AC/DC Kilovoltmeter, High Voltage Digital Meter, Reference Multimeter/ SICT-CP-40313
Leakage current(DC)		0 μA	7.0 nA	
		(0 ~ 1) μA	7.0×10^{-3}	
		(1 ~ 2) μA	3.6×10^{-3}	
		(2 ~ 5) μA	1.4×10^{-3}	
		(5 ~ 10) μA	7.4×10^{-4}	
		(10 ~ 20) μA	4.0×10^{-4}	
		(20 ~ 50) μA	1.8×10^{-4}	
		(50 ~ 100) μA	1.3×10^{-4}	
		(100 ~ 200) μA	8.5×10^{-5}	
		(0.2 ~ 100) mA	6.1×10^{-4}	
Leakage current(AC)		(20 μA)		
		10 Hz	1.3×10^{-3}	
		(10 ~ 20) Hz	8.5×10^{-4}	
		20 Hz ~ 1 kHz	7.0×10^{-4}	
		(1 ~ 5) kHz	1.3×10^{-3}	
		(5 ~ 10) kHz	5.5×10^{-3}	
		(20 ~ 50) μA		
		10 Hz	6.8×10^{-4}	
		(10 ~ 20) Hz	4.4×10^{-4}	
		20 Hz ~ 1 kHz	3.4×10^{-4}	
		(1 ~ 5) kHz	6.8×10^{-4}	
		(5 ~ 10) kHz	2.8×10^{-3}	
		(50 ~ 100) μA		
		10 Hz	4.9×10^{-4}	
		(10 ~ 20) Hz	3.2×10^{-4}	
		20 Hz ~ 1 kHz	2.3×10^{-4}	
		(1 ~ 5) kHz	4.9×10^{-4}	
		(5 ~ 10) kHz	4.0×10^{-4}	
		(100 ~ 200) μA		
		10 Hz	3.9×10^{-4}	
		(10 ~ 20) Hz	2.5×10^{-4}	
		20 Hz ~ 1 kHz	1.7×10^{-4}	
		(1 ~ 5) kHz	4.0×10^{-4}	
		(5 ~ 10) kHz	1.7×10^{-3}	
		(200 ~ 500) μA		
		10 Hz	4.4×10^{-4}	
		(10 ~ 20) Hz	3.2×10^{-4}	
		20 Hz ~ 1 kHz	2.4×10^{-4}	
		(1 ~ 5) kHz	5.4×10^{-4}	
		(5 ~ 10) kHz	2.8×10^{-3}	
		500 μA ~ 1 mA		
		10 Hz	7.0×10^{-4}	
		(10 ~ 20) Hz	6.6×10^{-4}	
		20 Hz ~ 1 kHz	6.3×10^{-4}	
		(1 ~ 5) kHz	7.2×10^{-4}	
		(5 ~ 10) kHz	2.1×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Current transformer test set	40315	(5 ~ 1 500) A (-19.99 ~ 19.99) %	0.020 %	Current transforme, Ratio transformers/ SICT-CP-40315
Ratio		(-680 ~ 680)'	0.70'	
Current transformer	40316	(5 ~ 10 000) A (-19.99 ~ 19.99) %	0.020 %	Current transforme/ SICT-CP-40316
Ratio		(-680 ~ 680)'	0.70'	
Phase		(AC) 2 ~ 50	0.10 %	
Current Coil		(DC) 2 ~ 50	0.10 %	
transducers		(±) (10 A) 50 : 1 ~ 5 000 : 1	1.2×10^{-4}	
		(10 ~ 1 000) A 50 : 1 ~ 5 000 : 1	1.9×10^{-4}	
	(1 000 ~ 2 000) A 50 : 1 ~ 5 000 : 1	2.5×10^{-4}		
AC voltmeters	40318	(600 μV) 1 kHz	7.8×10^{-3}	Reference Multimeter, Calibrator/ SICT-CP-40318
AC Voltage		(600 μV ~ 1 mV) 10 Hz	5.0×10^{-3}	
		10 Hz ~ 10 kHz	4.8×10^{-3}	
		(10 ~ 100) kHz	6.5×10^{-3}	
		(1 ~ 3) mV 10 Hz	2.0×10^{-3}	
		10 Hz ~ 10 kHz	1.7×10^{-3}	
		(10 ~ 100) kHz	2.8×10^{-3}	
		(3 ~ 10) mV 10 Hz	7.7×10^{-4}	
		10 Hz ~ 10 kHz	5.7×10^{-4}	
		(10 ~ 100) kHz	1.2×10^{-3}	
		(10 ~ 30) mV 10 Hz	8.0×10^{-4}	
		10 Hz ~ 10 kHz	3.7×10^{-4}	
		(10 ~ 100) kHz	1.1×10^{-3}	
		(30 ~ 100) mV 10 Hz	4.3×10^{-4}	
		10 Hz ~ 10 kHz	1.7×10^{-4}	
		(10 ~ 100) kHz	5.7×10^{-4}	
		(100 mV ~ 10 V) 10 Hz	4.9×10^{-4}	
		10 Hz ~ 10 kHz	1.1×10^{-4}	
	(10 ~ 100) kHz	2.6×10^{-4}		

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltmeters AC Voltage	40318	(10 ~ 100) V		Reference Multimeter, Calibrator/ SICT-CP-40318
		10 Hz	5.3×10^{-4}	
		10 Hz ~ 10 kHz	1.3×10^{-4}	
		(10 ~ 100) kHz	3.6×10^{-4}	
		(100 ~ 1 000) V		
		50 Hz	3.7×10^{-4}	
		50 Hz ~ 1 kHz	1.1×10^{-4}	
		(1 mV)		
		100 kHz	6.5×10^{-3}	
		100 kHz ~ 1 MHz	2.7×10^{-2}	
		(1 ~ 10) mV		
		100 kHz	1.2×10^{-3}	
		100 kHz ~ 1 MHz	5.6×10^{-3}	
		(10 ~ 100) mV		
		100 kHz	5.7×10^{-4}	
		100 kHz ~ 1 MHz	3.7×10^{-3}	
		(100 mV ~ 1 V)		
		100 kHz	1.5×10^{-4}	
		100 kHz ~ 1 MHz	2.3×10^{-3}	
		(1 ~ 10) V		
		100 kHz	4.7×10^{-5}	
		100 kHz ~ 1 MHz	7.0×10^{-4}	
		(10 ~ 20) V		
		100 kHz	7.6×10^{-5}	
		100 kHz ~ 1 MHz	1.3×10^{-3}	
		(25 mV)		
		1 MHz	1.9×10^{-2}	
		(1 ~ 30) MHz	2.3×10^{-2}	
		(25 ~ 100) mV		
		1 MHz	2.3×10^{-2}	
		(1 ~ 30) MHz	2.8×10^{-2}	
		(100 ~ 300) mV		
		1 MHz	3.7×10^{-2}	
		(1 ~ 30) MHz	4.0×10^{-2}	
		(300 mV ~ 1 V)		
		1 MHz	2.4×10^{-2}	
		(1 ~ 30) MHz	2.7×10^{-2}	
		(1 ~ 2) V		
		1 MHz	1.5×10^{-2}	
		(1 ~ 30) MHz	1.8×10^{-2}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
AC voltmeters AC Output Voltage	40318	(1 mV) 10 Hz ~ 10 kHz (10 ~ 100) kHz	1.9×10^{-3} 3.1×10^{-3}	Reference Multimeter, Calibrator/ SICT-CP-40318	
		(1 ~ 10) mV 10 Hz ~ 10 kHz (10 ~ 100) kHz	2.6×10^{-4} 3.7×10^{-4}		
		(10 ~ 100) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	9.2×10^{-5} 4.8×10^{-5} 9.1×10^{-5}		
		(100 mV ~ 1 V) 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	7.2×10^{-5} 1.8×10^{-5} 5.9×10^{-5}		
DC Output Voltage		1 mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V	7.5×10^{-4} 7.6×10^{-5} 9.4×10^{-6} 2.2×10^{-5}		
Watt hour meters Watt Hour	40319	(50 ~ 60) Hz 0 mWh (0 ~ 480) mWh 480 mWh ~ 12 kWh (12 ~ 24) kWh (24 ~ 300) kWh (300 ~ 600) kWh	44 μ Wh 4.0×10^{-4} 3.6×10^{-4} 7.6×10^{-4} 1.3×10^{-3} 1.4×10^{-3}		Power Calibrator/ SICT-CP-40319
		(DC) 0 mWh (0 ~ 1) mWh (1 ~ 100) mWh 100 mWh ~ 100 Wh 100 Wh ~ 10 kWh (10 ~ 20) kWh (20 ~ 1 000) kWh (1 000 ~ 2 500) kWh	61 nWh 3.5×10^{-4} 3.4×10^{-4} 3.5×10^{-4} 3.8×10^{-4} 3.7×10^{-4} 1.1×10^{-3} 1.3×10^{-3}		

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Ratio transformers	40321	Ratio		Calibrator/ SICT-CP-40321
		PT (±) (0.1 ~ 1.000) % (1.000 ~ 19.00) %	0.006 % 0.01 %	
		CT (±) (0.1 ~ 1.000) % (1.000 ~ 19.00) %	0.019 % 0.02 %	
		Phase		
		PT (±) (0.040 ~ 1.999) ' (1.999 ~ 19.99) ' (19.99 ~ 199.9) ' (199.9 ~ 600) '	0.060 ' 0.06 ' 0.2 ' 1 '	
		CT (±) (0.040 ~ 1.999) ' (1.999 ~ 19.99) ' (19.99 ~ 199.9) ' (199.9 ~ 600) '	0.060 ' 0.06 ' 0.2 ' 1 '	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF amplifiers Amplifier	40401	(DC) 1 mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 1 000) V (10 Hz ~ 10 kHz) 1 mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 1 000) V (10 ~ 100) kHz 1 mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 1 000) V	0.4 μ V 7.4×10^{-5} 6.1×10^{-5} 6.0×10^{-5} 1.0×10^{-4} 1.7μ V 2.6×10^{-4} 1.1×10^{-4} 9.0×10^{-5} 1.0×10^{-4} 3.1μ V 3.7×10^{-4} 1.1×10^{-4} 8.0×10^{-5} 1.0×10^{-4}	Reference Multimeter/ SICT-CP-40401
DC/LF attenuators Attenuation	40402	10 Hz ~ 100 kHz (0 ~ -20) dB (-20 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	0.001 9 dB 0.001 7 dB 0.005 5 dB 0.008 7 dB	Reference Multimeter/ SICT-CP-40402
Multimeter calibrators DC Voltage DC Current AC Voltage	40403	(\pm) 0 mV (0 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V (\pm) 1 nA (1 ~ 100) nA 100 nA ~ 10 A (10 ~ 50) A (50 ~ 100) A (1 mV) 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (1 ~ 2) mV 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	0.05 μ V 3.3×10^{-6} 1.4×10^{-6} 1.0×10^{-6} 1.6×10^{-6} 2.1×10^{-6} 7.0 pA 4.7×10^{-3} 1.2×10^{-5} 4.0×10^{-5} 4.4×10^{-5} 1.7×10^{-3} 3.0×10^{-3} 1.2×10^{-2} 1.1×10^{-3} 1.7×10^{-3} 7.7×10^{-3}	Reference Multimeter/ SICT-CP-40403

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators AC Voltage	40403	(2 ~ 5) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	6.4×10^{-4} 5.8×10^{-4} 1.0×10^{-3} 5.4×10^{-3}	Reference Multimeter/ SICT-CP-40403
		(5 ~ 10) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	4.2×10^{-4} 3.5×10^{-4} 5.8×10^{-4} 3.9×10^{-3}	
		(10 ~ 20) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	1.8×10^{-4} 1.4×10^{-4} 2.2×10^{-4} 2.2×10^{-3}	
		(20 ~ 50) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	1.4×10^{-4} 9.2×10^{-5} 1.6×10^{-4} 1.4×10^{-3}	
		(50 ~ 100) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	1.1×10^{-4} 6.6×10^{-5} 1.2×10^{-4} 1.3×10^{-3}	
		(100 ~ 200) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	8.2×10^{-5} 3.9×10^{-5} 7.6×10^{-5} 1.1×10^{-3}	
		(200 ~ 500) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	7.8×10^{-5} 3.6×10^{-5} 7.1×10^{-5} 1.1×10^{-3}	
		(0.5 ~ 1) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	7.6×10^{-5} 3.3×10^{-5} 6.6×10^{-5} 1.1×10^{-3}	
		(1 ~ 2) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	7.1×10^{-5} 2.7×10^{-5} 5.8×10^{-5} 1.0×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators	40403	AC Voltage	(2 ~ 5) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	Reference Multimeter/ SICT-CP-40403
			(5 ~ 20) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	
			(20 ~ 50) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	
			(50 ~ 200) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	
			(200 ~ 1 000) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	
		AC Current	(10 μA) 10 Hz ~ 10 kHz	
			(10 ~ 100) μA 10 Hz ~ 1 kHz (1 ~ 10) kHz	
			(0.1 ~ 1) mA 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	
			(1 ~ 100) mA 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	
			(0.1 ~ 1) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	
			(1 ~ 2) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators	40403	AC Current		Reference Multimeter/ SICT-CP-40403
		Resistance		
		(2 ~ 5) A		
		10 Hz	8.2×10^{-5}	
		10 Hz ~ 1 kHz	5.2×10^{-5}	
		(1 ~ 10) kHz	5.0×10^{-5}	
		(5 ~ 10) A		
		10 Hz	8.6×10^{-5}	
		10 Hz ~ 1 kHz	5.9×10^{-5}	
		(1 ~ 10) kHz	7.8×10^{-5}	
		(10 ~ 20) A		
		10 Hz	9.3×10^{-5}	
		10 Hz ~ 1 kHz	6.8×10^{-5}	
		(1 ~ 10) kHz	7.8×10^{-5}	
		(20 ~ 50) A		
		10 Hz	1.0×10^{-4}	
		10 Hz ~ 1 kHz	8.3×10^{-5}	
		(1 ~ 10) kHz	1.1×10^{-4}	
		(50 ~ 100) A		
		10 Hz	1.2×10^{-4}	
		10 Hz ~ 1 kHz	9.7×10^{-5}	
		(1 ~ 10) kHz	1.3×10^{-4}	
		(100 ~ 200) A		
		60 Hz	4.5×10^{-4}	
		0 Ω	0.14 μΩ	
		(0 ~ 1) Ω	6.6×10^{-6}	
		(1 ~ 1.9) Ω	8.4×10^{-6}	
		(1.9 ~ 10) Ω	3.6×10^{-6}	
		(10 ~ 19) Ω	2.6×10^{-6}	
		(19 ~ 100) Ω	2.8×10^{-6}	
		(0.1 ~ 1) kΩ	2.5×10^{-6}	
		(1 ~ 1.9) kΩ	3.8×10^{-6}	
		(1.9 ~ 10) kΩ	2.0×10^{-6}	
		(10 ~ 19) kΩ	1.3×10^{-6}	
		(19 ~ 100) kΩ	1.9×10^{-6}	
		(100 ~ 190) kΩ	2.0×10^{-6}	
		(0.19 ~ 1) MΩ	2.9×10^{-6}	
		(1 ~ 1.9) MΩ	3.1×10^{-6}	
		(1.9 ~ 10) MΩ	3.6×10^{-6}	
		(10 ~ 19) MΩ	2.9×10^{-6}	
		(19 ~ 100) MΩ	1.5×10^{-5}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators Multimeter calibrators(property) (Digital sampling) AC Voltage	40403	(1 mV) 0.1 Hz ~ 3 kHz	8.4×10^{-4}	Reference Multimeter/ SICT-CP-40403
		(1 ~ 2) mV 0.1 Hz ~ 3 kHz	4.2×10^{-4}	
		(2 ~ 3) mV 0.1 Hz ~ 3 kHz	2.8×10^{-4}	
		(3 ~ 5) mV 0.1 Hz ~ 3 kHz	1.7×10^{-4}	
		(5 ~ 10) mV 0.1 Hz ~ 3 kHz	8.8×10^{-5}	
		(10 ~ 20) mV 0.1 Hz ~ 3 kHz	4.8×10^{-5}	
		(20 ~ 30) mV 0.1 Hz ~ 3 kHz	3.6×10^{-5}	
		(30 ~ 50) mV 0.1 Hz ~ 3 kHz	3.0×10^{-5}	
		(50 ~ 100) mV 0.1 Hz ~ 3 kHz	2.6×10^{-5}	
		(100 ~ 200) mV 0.1 Hz ~ 3 kHz	4.8×10^{-5}	
		(200 ~ 300) mV 0.1 Hz ~ 3 kHz	3.6×10^{-5}	
		(300 ~ 500) mV 0.1 Hz ~ 3 kHz	2.8×10^{-5}	
		(500 mV ~ 1 V) 0.1 Hz ~ 3 kHz	2.4×10^{-5}	
		(1 ~ 2) V 0.1 Hz ~ 3 kHz	4.8×10^{-5}	
		(2 ~ 3) V 0.1 Hz ~ 3 kHz	3.6×10^{-5}	
(3 ~ 5) V 0.1 Hz ~ 3 kHz	3.0×10^{-5}			
(5 ~ 10) V 0.1 Hz ~ 3 kHz	2.6×10^{-5}			

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Oscilloscope calibrators Sine Wave Generator Time Marker Generator Impedance Mesurement	40404	(600 mV ~ 1 V) 50 kHz (50 ~ 500) kHz 0.5 MHz ~ 1 GHz (1 ~ 6) GHz (0.1 ~ 1) ns (1 ~ 10) ns (10 ~ 100) ns 0.1 μs ~ 10 ms (10 ~ 100) ms (0.1 ~ 1) s (1 ~ 5) s (50 ~ 75) Ω 75 Ω ~ 1 MΩ	0.58 mV 1.0×10^{-3} 1.7×10^{-2} 1.9×10^{-2} 5.8×10^{-8} 6.5×10^{-9} 3.1×10^{-9} 5.8×10^{-8} 6.1×10^{-9} 5.8×10^{-8} 1.2×10^{-8} 1.7×10^{-4} 2.1×10^{-4}	Calibrator/ SICT-CP-40404
CD/DVD meters/analyzers Jitter	40405	(1.0 ~ 60.0) ns 1 % 2 % 4 % 8 % 10 % 15 %	1.7×10^{-3} 0.05 % 0.09 % 0.19 % 0.36 % 0.44 % 0.67 %	Modulation Domain Analyzer/ SICT-CP-40405
Video signal generators NTSC, PAL Multiburst NTSC, PAL, SECAM Pulse and Bar NTSC, PAL , SECAM Frequency Video frequency Video level TTL Sync level D-TV Level	40406	(0.1 ~ 1) MHz (1 ~ 2) MHz (2 ~ 6) MHz (0 ~ 300) ns (0 ~ 1 000) mV 1 Hz ~ 10 MHz (10 ~ 100) Hz 100 Hz ~ 500 MHz (30 ~ 600) mV (600 ~ 1 200) mV (1 ~ 5) V (30 ~ 600) mV (600 ~ 1 200) mV	6.0×10^{-2} 6.2×10^{-3} 3.1×10^{-3} 4.2×10^{-4} 3.5×10^{-3} 1.6×10^{-9} 6.2×10^{-8} 6.2×10^{-9} 2.6×10^{-3} 2.3×10^{-3} 2.7×10^{-3} 2.6×10^{-3} 2.3×10^{-3}	Video Measurement/ SICT-CP-40406

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Video signal generators NTSC, PAL, H-Timing(Level)	40406	(0 ~ 100) mV	2.6×10^{-3}	Video Measurement/ SICT-CP-40406	
		(100 ~ 1 000) mV	3.4×10^{-3}		
(Time)		(0 ~ 254) ns	1.2×10^{-2}		
		(254 ~ 300) ns	3.8×10^{-3}		
		300 ns ~ 3 μ s	3.2×10^{-3}		
		(3 ~ 7) μ s	7.4×10^{-3}		
		(7 ~ 10) μ s	4.2×10^{-3}		
NTSC, PAL Color Bar(Luminance Level)		(0 ~ 100) mV	0.06 mV		
		(100 ~ 1 000) mV	3.4×10^{-3}		
NTSC, PAL Color Bar(Chrominance Level)		(0 ~ 100) mV	0.06 mV		
		(100 ~ 1 000) mV	3.4×10^{-3}		
NTSC, PAL Color Bar(Phase)		(0 ~ 360)°	0.13°		
SECAM Color Bar Frequency		(D'R & D'B) (3 ~ 5) MHz	1.2×10^{-3}		
RF Output frequency		10 kHz ~ 10 MHz (10 ~ 100) MHz (100 ~ 1 000) MHz	6.0×10^{-4} 6.0×10^{-5} 6.0×10^{-6}		
RF Output level	(0.1 ~ 10) mV (10 ~ 500) mV	1.4×10^{-2} 1.3×10^{-2}			
Sound Frequency	10 Hz ~ 100 kHz 100 kHz ~ 1 MHz	6.1×10^{-8} 6.1×10^{-7}			
Audio distortion analyzers/meters	40407	Input Frequency	1 Hz ~ 200 kHz	6.1×10^{-7}	Calibrator/ SICT-CP-40407
		Output Level Flatness Test	(10 ~ 100) kHz	0.008 3 dB	
		Input DC Voltage	0 mV (0 ~ 1) mV (1 ~ 10) mV 10 mV ~ 300 V	0.27 μ V 5.8×10^{-3} 5.8×10^{-4} 5.8×10^{-4}	
		Input Distortion	(100 Hz ~ 10 kHz) (-10 ~ -40) dB (-40 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	0.005 8 dB 0.006 0 dB 0.006 8 dB 0.012 dB 0.028 dB	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Audio distortion analyzers/meters	40407	(10 kHz ~ 50 kHz)		Calibrator/ SICT-CP-40407
Input Distortion		(-10 ~ -40) dB	0.005 9 dB	
		(-40 ~ -50) dB	0.006 3 dB	
		(-50 ~ -60) dB	0.008 2 dB	
		(-60 ~ -70) dB	0.019 dB	
		(-70 ~ -80) dB	0.052 dB	
Input AC Voltage		(10 ~ 100) Hz		
		(1 ~ 10) mV	9.0×10^{-4}	
		(10 ~ 100) mV	4.0×10^{-4}	
		100 mV ~ 100 V	4.2×10^{-4}	
		(100 ~ 300) V	5.3×10^{-4}	
		(100 Hz ~ 1 kHz)		
		(1 ~ 10) mV	8.4×10^{-4}	
		(10 ~ 100) mV	1.8×10^{-4}	
		100 mV ~ 10 V	1.1×10^{-4}	
		(10 ~ 100) V	1.0×10^{-4}	
		(100 ~ 300) V	2.3×10^{-4}	
		(1 ~ 10) kHz		
		(1 ~ 10) mV	8.4×10^{-4}	
		(10 ~ 100) mV	1.8×10^{-4}	
		100 mV ~ 10 V	1.5×10^{-4}	
		(10 ~ 100) V	2.7×10^{-4}	
		(10 ~ 100) kHz		
		(1 ~ 10) mV	1.4×10^{-3}	
	(10 ~ 100) mV	7.6×10^{-4}		
	100 mV ~ 1 V	4.8×10^{-4}		
	(1 ~ 10) V	4.1×10^{-4}		
	(10 ~ 100) V	3.4×10^{-4}		
Input Attenuation	(10 Hz)			
	(30 ~ -50) dB	0.006 8 dB		
	(-50 ~ -60) dB	0.016 dB		
	(-60 ~ -80) dB	0.052 dB		
	(10 Hz ~ 10 kHz)			
	(30 ~ -60) dB	0.008 3 dB		
	(-60 ~ -70) dB	0.014 dB		
	(-70 ~ -80) dB	0.042 dB		
	(10 ~ 100) kHz			
	(30 ~ -50) dB	0.009 1 dB		
	(-50 ~ -70) dB	0.023 dB		
	(-70 ~ -80) dB	0.057 dB		
Input Impedance	300 Ω ~ 200 kΩ	3.1×10^{-4}		
Input Filter	(10 Hz ~ 100 kHz)			
	1 V	8.3×10^{-4}		
(Distortion meter calibrator) Distortion	(400 Hz , 1 kHz)			
	(-10 ~ -20) dB	0.15 dB		
	(-20 ~ -40) dB	0.14 dB		
	(-40 ~ -60) dB	0.17 dB		
	(-60 ~ -80) dB	0.26 dB		

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF filters Filter	40408	10 Hz ~ 50 kHz (50 ~ 100) kHz (100 ~ 150) kHz	5.8×10^{-4} 1.2×10^{-3} 5.8×10^{-3}	Audio Analyzer/ SICT-CP-40408
LF/Audiosignalanalyzers Output Frequency AC Output Level AC Output Level Flatness Output Attenuation Output DC Offset Output Impedance Input Frequency AC Input Level Flatness DC Input Level Input Distortion	40409	1 Hz ~ 200 kHz (10 ~ 100) Hz (1 ~ 10) mV 10 mV ~ 30 V (-20 ~ 10) dBm (100 Hz ~ 10 kHz) (1 ~ 10) mV 10 mV ~ 30 V (-20 ~ 10) dBm (10 ~ 100) kHz (1 ~ 10) mV 10 mV ~ 30 V (-20 ~ 10) dBm 10 Hz ~ 100 kHz (0 ~ -60) dB (±) 0 mV (0 ~ 1) mV (1 ~ 10) mV 10 mV ~ 50 V 5 Ω (10 ~ 600) Ω 1 Hz ~ 200 kHz 10 Hz ~ 100 kHz (±) 0 mV (0 ~ 1) mV (1 ~ 10) mV 10 mV ~ 300 V (100 Hz ~ 10 kHz) (-10 ~ -40) dB (-40 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	5.8×10^{-6} 8.7×10^{-4} 9.4×10^{-5} 0.005 8 dB 8.7×10^{-4} 6.5×10^{-5} 0.005 8 dB 8.7×10^{-4} 9.4×10^{-5} 0.005 8 dB 0.007 1 dB 0.005 8 dB 0.7 μV 1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5} 1.2×10^{-3} 6.0×10^{-4} 6.1×10^{-7} 0.008 3 dB 0.27 μV 5.8×10^{-3} 5.8×10^{-4} 5.8×10^{-4} 0.005 8 dB 0.006 0 dB 0.006 8 dB 0.012 dB 0.028 dB	Calibrator, Reference Multimeter/ SICT-CP-40409

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF/Audiosignalanalyzers	40409			Calibrator, Reference Multimeter/ SICT-CP-40409
Input Distortion		(10 kHz ~ 50 kHz) (-10 ~ -40) dB (-40 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	0.005 9 dB 0.006 3 dB 0.008 2 dB 0.019 dB 0.052 dB	
AC Input Level		(10 ~ 100) Hz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 100 V (100 ~ 300) V (100 Hz ~ 1 kHz) (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 10 V (10 ~ 100) V (100 ~ 300) V (1 ~ 10) kHz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 10 V (10 ~ 100) V (10 ~ 100) kHz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 1 V (1 ~ 10) V (10 ~ 100) V	9.0×10^{-4} 4.0×10^{-4} 4.2×10^{-4} 5.3×10^{-4} 8.4×10^{-4} 1.8×10^{-4} 1.1×10^{-4} 1.0×10^{-4} 2.3×10^{-4} 8.4×10^{-4} 1.8×10^{-4} 1.5×10^{-4} 2.7×10^{-4} 1.4×10^{-3} 7.6×10^{-4} 4.1×10^{-4} 3.4×10^{-4} 2.6×10^{-4}	
Input Attenuation		(10 Hz) (30 ~ -50) dB (-50 ~ -60) dB (-60 ~ -80) dB (10 Hz ~ 10 kHz) (30 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB (10 ~ 100) kHz (30 ~ -50) dB (-50 ~ -70) dB (-70 ~ -80) dB	0.006 8 dB 0.016 dB 0.052 dB 0.008 3 dB 0.014 dB 0.042 dB 0.009 1 dB 0.023 dB 0.057 dB	
Input Impedance		300 Ω ~ 200 kΩ	3.1×10^{-4}	
Input Filter		(10 Hz ~ 100 kHz) 1 V	8.3×10^{-4}	
Line frequency meters	40410			Calibrator/ SICT-CP-40410
Frequency		16 Hz ~ 1 kHz	1.3×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Function generators	40411			Audio Analyzer, Digital Multimeter/ SICT-CP-40411
Frequency		(0.01 ~ 0.1) Hz	5.8×10^{-6}	
		(0.1 ~ 1) Hz	5.8×10^{-7}	
		1 Hz ~ 1 GHz	5.8×10^{-9}	
		(1 ~ 4) GHz	1.5×10^{-8}	
Output Level		(10 ~ 100) Hz		
		1 mV	1.0×10^{-3}	
		(1 ~ 10) mV	1.0×10^{-4}	
		10 mV ~ 100 V	7.0×10^{-5}	
		(100 Hz ~ 10 kHz)		
		1 mV	1.0×10^{-3}	
		(1 ~ 10) mV	1.0×10^{-4}	
		10 mV ~ 100 V	3.0×10^{-5}	
		(10 ~ 100) kHz		
		1 mV	1.0×10^{-3}	
		(1 ~ 10) mV	1.0×10^{-4}	
		10 mV ~ 100 V	8.0×10^{-5}	
DC Offset		(\pm)		
		0 mV	0.7 μ V	
		(0 ~ 1) mV	0.7 μ V	
		(1 ~ 10) mV	1.0×10^{-4}	
		10 mV ~ 20 V	6.0×10^{-5}	
Level Flatness		(100 mV)		
		(10 ~ 100) Hz	0.099 dB	
		100 Hz ~ 10 kHz	0.083 dB	
		(10 ~ 100) kHz	0.095 dB	
		(100 mV ~ 1 V)		
		(10 ~ 100) Hz	0.005 4 dB	
		100 Hz ~ 10 kHz	0.001 1 dB	
		(10 ~ 100) kHz	0.007 2 dB	
		(1 ~ 30) V		
		(10 ~ 100) Hz	0.021 dB	
		100 Hz ~ 10 kHz	0.015 dB	
		(10 ~ 100) kHz	0.027 dB	
Attenuation		(10 Hz ~ 100 kHz)		
		(0 ~ 80) dB	0.006 1 dB	
Distortion		(20 Hz ~ 1 kHz)		
		(3.16 ~ 0.010) %	1.5×10^{-1}	
		(1 ~ 100) kHz		
		(3.16 ~ 0.010) %	3.2×10^{-1}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Function generators Rise Fall Time	40411	100 μs ~ 100 ns	7.0×10^{-4}	Audio Analyzer, Digital Multimeter/ SICT-CP-40411
		(100 ~ 10) ns	7.8×10^{-4}	
		(10 ~ 1) ns	4.7×10^{-3}	
		1 ns ~ 100 ps	4.6×10^{-2}	
Duty cycle		(1 ~ 99) %	0.006 1 %	
High Frequency Flatness Test	40412	(100 kHz ~ 80 MHz)	0.11 dB	Signal Generator/ SICT-CP-40412
		(0 ~ 20) dBm		
FM Modulation		(0.1 ~ 400) kHz		
AM Modulation	(0.1 ~ 100) %	1.2×10^{-2}		
Genescopes Marker Frequency	40412	9 kHz ~ 10 MHz	2.8×10^{-6}	Signal Generator/ SICT-CP-40412
		(10 ~ 200) MHz	6.4×10^{-7}	
RF Level	40413	9 kHz ~ 200 MHz	0.31 dB	Calibrator/ SICT-CP-40413
	(100 ~ 50) dBμV			
AC/DC high voltages volt meters DC Voltage	40413	(±)	0.58 V	
		0 kV	1.2×10^{-3}	
		(0 ~ 0.5) kV	6.1×10^{-4}	
		(0.5 ~ 1) kV	4.4×10^{-4}	
		(1 ~ 2) kV	3.4×10^{-4}	
		(2 ~ 100) kV		
AC Voltage		(50 Hz)	0.58 V	
		0.01 kV	1.2×10^{-3}	
		(0.01 ~ 0.5) kV	6.2×10^{-4}	
		(0.5 ~ 1) kV	5.5×10^{-4}	
	(1 ~ 2) kV	5.3×10^{-4}		
	(2 ~ 3) kV	5.0×10^{-4}		
	(3 ~ 15 kV	5.7×10^{-4}		
	(15 ~ 100) kV			
	(60 Hz)	0.58 V		
	0.01 kV	1.2×10^{-3}		
	(0.01 ~ 0.5) kV	6.2×10^{-4}		
	(0.5 ~ 1) kV	5.5×10^{-4}		
	(1 ~ 2) kV	4.7×10^{-4}		
	(2 ~ 3) kV	4.5×10^{-4}		
	(3 ~ 15 kV	5.4×10^{-4}		
	(15 ~ 100) kV			

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Jitter meters CD/DVD Jitter VTR Jitter	40415	(1 ~ 20) ns (20 ~ 60) ns 0.05 μs (0.05 ~ 0.1) μs (0.1 ~ 0.2) μs (0.2 ~ 0.5) μs (0.5 ~ 0.7) μs 1 % 2 % 4 % 8 % 10 % 15 %	1.7×10^{-3} 1.6×10^{-3} 0.66 ns 0.77 ns 1.2 ns 2.8 ns 4.3 ns 0.05 % 0.09 % 0.19 % 0.36 % 0.44 % 0.67 %	Modulation Domain Analyzer/ SICT-CP-40415
Leakage current testers DC Current AC Current	40416	0 μA (0 ~ 1) μA (1 ~ 2) μA (2 ~ 5) μA (5 ~ 10) μA (10 ~ 20) μA (20 ~ 50) μA (50 ~ 100) μA (100 ~ 200) μA (0.2 ~ 100) mA (20 μA) 10 Hz (10 ~ 20) Hz (0.02 ~ 1) kHz (1 ~ 5) kHz (5 ~ 10) kHz (20 ~ 50) μA 10 Hz (10 ~ 20) Hz (0.02 ~ 1) kHz (1 ~ 5) kHz (5 ~ 10) kHz (50 ~ 100) μA 10 Hz (10 ~ 20) Hz (0.02 ~ 1) kHz (1 ~ 5) kHz (5 ~ 10) kHz	7.0 nA 2.4×10^{-3} 3.6×10^{-3} 1.4×10^{-3} 7.4×10^{-4} 4.0×10^{-4} 1.8×10^{-4} 1.3×10^{-4} 8.5×10^{-5} 6.1×10^{-4} 26 nA 8.5×10^{-4} 7.0×10^{-4} 1.3×10^{-3} 5.5×10^{-3} 6.8×10^{-4} 4.4×10^{-4} 3.4×10^{-4} 6.8×10^{-4} 2.8×10^{-3} 4.9×10^{-4} 3.2×10^{-4} 2.3×10^{-4} 4.9×10^{-4} 4.0×10^{-4}	Calibrator/ SICT-CP-40416

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers	40416	(100 ~ 200) μ A		Calibrator/ SICT-CP-40416
		AC Current		
		10 Hz	3.9×10^{-4}	
		(10 ~ 20) Hz	2.5×10^{-4}	
		(0.02 ~ 1) kHz	1.7×10^{-4}	
		(1 ~ 5) kHz	4.0×10^{-4}	
		(5 ~ 10) kHz	1.7×10^{-3}	
		(200 ~ 500) μ A		
		10 Hz	4.4×10^{-4}	
		(10 ~ 20) Hz	3.2×10^{-4}	
		(0.02 ~ 1) kHz	2.4×10^{-4}	
		(1 ~ 5) kHz	5.4×10^{-4}	
		(5 ~ 10) kHz	2.8×10^{-3}	
		(0.5 ~ 1) mA		
		10 Hz	7.0×10^{-4}	
		(10 ~ 20) Hz	6.6×10^{-4}	
		(0.02 ~ 1) kHz	6.3×10^{-4}	
		(1 ~ 5) kHz	7.2×10^{-4}	
		(5 ~ 10) kHz	2.1×10^{-3}	
		(1 ~ 100) mA		
		10 Hz	7.0×10^{-4}	
		(10 ~ 20) Hz	6.6×10^{-4}	
		(0.02 ~ 1) kHz	6.3×10^{-4}	
		(1 ~ 5) kHz	7.0×10^{-4}	
		(5 ~ 10) kHz	2.7×10^{-3}	
		DC Voltage		
		0 V	0.06 mV	
		(0 ~ 0.1) V	6.0×10^{-4}	
		(0.1 ~ 0.2) V	3.0×10^{-4}	
		(0.2 ~ 0.5) V	1.2×10^{-4}	
		(0.5 ~ 1) V	6.0×10^{-5}	
		(1 ~ 2) V	3.1×10^{-4}	
		(2 ~ 5) V	1.2×10^{-4}	
		(5 ~ 10) V	6.1×10^{-5}	
		(10 ~ 20) V	3.1×10^{-5}	
		(20 ~ 50) V	1.4×10^{-5}	
		(50 ~ 100) V	8.8×10^{-6}	
		(100 ~ 200) V	3.1×10^{-5}	
		(200 ~ 300) V	2.3×10^{-5}	
		(300 ~ 500) V	1.2×10^{-4}	
		(500 ~ 1 000) V	5.8×10^{-5}	
		AC Voltage		
		0.1 V		
		10 Hz	0.074 mV	
		(0.01 ~ 50) kHz	6.5×10^{-4}	
		(50 ~ 100) kHz	8.3×10^{-4}	
		(100 ~ 300) kHz	1.2×10^{-3}	
		(300 ~ 500) kHz	2.0×10^{-3}	
		(0.5 ~ 1) MHz	3.6×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers	40416	AC Voltage	(5 ~ 10) V	
			10 Hz	3.3×10^{-4}
		(10 ~ 20) Hz	1.4×10^{-4}	
		(0.02 ~ 20) kHz	8.9×10^{-5}	
		(20 ~ 100) kHz	1.4×10^{-4}	
		(100 ~ 300) kHz	3.9×10^{-4}	
		(300 ~ 500) kHz	1.4×10^{-3}	
		(0.5 ~ 1) MHz	2.2×10^{-3}	
		(10 ~ 20) V		
		10 Hz	3.1×10^{-4}	
		(10 ~ 20) Hz	1.2×10^{-4}	
		(0.02 ~ 20) kHz	6.0×10^{-5}	
		(20 ~ 50) kHz	9.0×10^{-5}	
		(50 ~ 100) kHz	1.1×10^{-4}	
		(20 ~ 50) V		
		10 Hz	4.2×10^{-4}	
		(10 ~ 20) Hz	2.2×10^{-4}	
		(0.02 ~ 50) kHz	1.8×10^{-4}	
		(50 ~ 100) kHz	2.8×10^{-4}	
		(50 ~ 100) V		
		10 Hz	3.4×10^{-4}	
		(0.01 ~ 50) kHz	1.4×10^{-4}	
		(50 ~ 100) kHz	2.2×10^{-4}	
		(100 ~ 1 000) V		
		(0.05 ~ 1) kHz	1.1×10^{-4}	
		Resistance		
		100 mΩ	7.7 μΩ	
		1 Ω ~ 10 kΩ	6.2×10^{-5}	
Input Voltage to Output Current Display(U1)	40416	20 Hz		
		(4.75 ~ 5.25) mA	0.006 3 mA	
		50 Hz		
		(4.77 ~ 5.28) mA	0.006 1 mA	
		60 Hz		
		(4.77 ~ 5.28) mA	0.006 1 mA	
		100 Hz		
		(4.85 ~ 5.36) mA	0.006 1 mA	
		200 Hz		
		(5.11 ~ 5.65) mA	0.006 1 mA	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers Input Voltage to Output Current Display(U1)	40416	500 Hz (6.64 ~ 7.34) mA	0.006 1 mA	Calibrator/ SICT-CP-40416
		1 kHz (9.70 ~ 10.73) mA	0.006 1 mA	
		2 kHz (14.07 ~ 15.56) mA	0.006 2 mA	
		5 kHz (17.82 ~ 19.70) mA	0.006 2 mA	
		10 kHz (18.66 ~ 20.63) mA	0.006 2 mA	
		20 kHz (18.92 ~ 20.92) mA	0.006 2 mA	
		50 kHz (19.00 ~ 21.00) mA	0.006 4 mA	
		100 kHz (19.00 ~ 21.00) mA	0.006 7 mA	
		200 kHz (19.00 ~ 21.00) mA	0.010 mA	
		500 kHz (19.00 ~ 21.00) mA	0.030 mA	
		1 MHz (19.00 ~ 21.00) mA	0.046 mA	
Input Voltage to Output Current Display(U2)		20 Hz (4.75 ~ 5.25) mA	0.006 3 mA	
		50 Hz (4.77 ~ 5.28) mA	0.006 1 mA	
		60 Hz (4.77 ~ 5.28) mA	0.006 1 mA	
		100 Hz (4.80 ~ 5.30) mA	0.006 1 mA	
		200 Hz (4.92 ~ 5.44) mA	0.006 1 mA	
		500 Hz (5.37 ~ 5.93) mA	0.006 1 mA	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Leakage current testers	40416	1 kHz (5.56 ~ 6.14) mA	0.006 1 mA	Calibrator/ SICT-CP-40416	
Input Voltage to Output Current Display(U2)		2 kHz (4.68 ~ 5.17) mA	0.006 1 mA		
		5 kHz (2.53 ~ 2.80) mA	0.000 63 mA		
		10 kHz (1.35 ~ 1.49) mA	0.000 62 mA		
		20 kHz (0.683 ~ 0.755) mA	0.000 61 mA		
		50 kHz (274.57 ~ 303.47) mA	0.029 μA		
		100 kHz (137.48 ~ 151.95) μA	0.020 μA		
		200 kHz (68.82 ~ 76.06) μA	0.030 μA		
		500 kHz (27.43 ~ 30.32) μA	0.042 μA		
		1 MHz (13.71 ~ 15.16) μA	0.033 μA		
		Input Voltage to Output Current Display(U3)	20 Hz (4.75 ~ 5.25) mA		0.006 3 mA
		50 Hz (4.77 ~ 5.28) mA	0.006 1 mA		
		60 Hz (4.77 ~ 5.28) mA	0.006 1 mA		
		100 Hz (4.80 ~ 5.30) mA	0.006 1 mA		
200 Hz (4.95 ~ 5.47) mA	0.006 1 mA				
500 Hz (5.65 ~ 6.25) mA	0.006 1 mA				
1 kHz (6.60 ~ 7.29) mA	0.006 1 mA				

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Leakage current testers	40416			Calibrator/ SICT-CP-40416	
Input Voltage to Output Current Display(U3)		2 kHz (7.14 ~ 7.89) mA	0.006 1 mA		
		5 kHz (5.31 ~ 5.87) mA	0.006 1 mA		
		10 kHz (3.12 ~ 3.45) mA	0.000 64 mA		
		20 kHz (1.63 ~ 1.81) mA	0.000 62 mA		
		50 kHz (0.664 ~ 0.734) mA	0.000 62 mA		
		100 kHz (322.16 ~ 367.12) μA	0.046 μA		
		200 kHz (166.03 ~ 183.81) μA	0.070 μA		
		500 kHz (66.37 ~ 73.35) μA	0.10 μA		
		1 MHz (33.14 ~ 36.63) μA	0.08 μA		
	Input Voltage to Output Voltage Ratio(U1)		4.00 (20 Hz)	1.3×10^{-4}	
			3.98 (50 Hz)	6.5×10^{-5}	
			3.97 (60 Hz)	6.5×10^{-5}	
			3.92 (100 Hz)	6.5×10^{-5}	
			3.72 (200 Hz)	6.5×10^{-5}	
			2.87 (500 Hz)	6.4×10^{-5}	
			1.96 (1 kHz)	6.4×10^{-5}	
		1.96 (2 kHz)	6.4×10^{-5}		
		1.96 (5 kHz)	6.4×10^{-5}		
		1.96 (10 kHz)	6.4×10^{-5}		
		1.00 (20 kHz)	6.7×10^{-5}		
		1.00 (50 kHz)	9.6×10^{-5}		
		1.00 (100 kHz)	1.2×10^{-4}		
	1.00 (200 kHz)	4.2×10^{-4}			
	1.00 (500 kHz)	1.5×10^{-3}			
	1.00 (1 MHz)	2.6×10^{-3}			

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers Input Voltage to Output Voltage Ratio(U2)	40416	4.00 (20 Hz) 3.99 (50 Hz) 3.99 (60 Hz) 3.96 (100 Hz) 3.87 (200 Hz) 3.54 (500 Hz) 3.43 (1 kHz) 4.06 (2 kHz) 7.50 (5 kHz) 14.1 (10 kHz) 27.8 (20 kHz) 69.2 (50 kHz) 138 (100 kHz) 272 (200 kHz) 691 (500 kHz) 1 382 (1 MHz)	1.3×10^{-4} 6.5×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.6×10^{-5} 7.0×10^{-5} 3.6×10^{-5} 4.8×10^{-5} 1.0×10^{-4} 2.7×10^{-4} 1.1×10^{-3} 3.0×10^{-3}	Calibrator/ SICT-CP-40416
Input Voltage to Output Voltage Ratio(U3)		4.00 (20 Hz) 3.99 (50 Hz) 3.98 (60 Hz) 3.95 (100 Hz) 3.83 (200 Hz) 2.36 (500 Hz) 2.87 (1 kHz) 2.65 (2 kHz) 3.57 (5 kHz) 6.09 (10 kHz) 11.6 (20 kHz) 28.7 (50 kHz) 57.2 (100 kHz) 114 (200 kHz) 286 (500 kHz) 572 (1 MHz)	1.3×10^{-4} 6.5×10^{-5} 6.5×10^{-5} 6.6×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.4×10^{-5} 6.4×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.9×10^{-5} 3.8×10^{-5} 7.9×10^{-5} 1.9×10^{-4} 6.1×10^{-4} 2.3×10^{-3}	
mAs Meter		1 mAs (1 ~ 2 000) mAs (2 000 ~ 9 999) mAs	1.2×10^{-3} 1.0×10^{-3} 1.1×10^{-3}	
Electronic AC/DC loads DC Voltage	40417	0 mV (0 ~ 5) mV (5 ~ 20) mV (20 ~ 100) mV (0.1 ~ 1) V (1 ~ 2) V (2 ~ 4) V (4 ~ 7) V (7 ~ 9) V (9 ~ 10) V (10 ~ 50) V (50 ~ 100) V (100 ~ 200) V (200 ~ 400) V (400 ~ 1 000) V	0.058 mV 5.8×10^{-2} 5.8×10^{-3} 1.2×10^{-3} 6.2×10^{-5} 3.2×10^{-5} 2.1×10^{-5} 1.3×10^{-5} 9.1×10^{-6} 7.9×10^{-6} 3.1×10^{-5} 1.0×10^{-5} 3.4×10^{-5} 2.5×10^{-5} 1.6×10^{-5}	Calibrator/ SICT-CP-40417

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Electronic AC/DC loads	40417	DC Current	1 mA	5.8 μA	Calibrator/ SICT-CP-40417
		(1 ~ 2) mA	2.9×10^{-3}		
		(2 ~ 5) mA	1.2×10^{-3}		
		(5 ~ 20) mA	5.8×10^{-4}		
		(20 ~ 50) mA	1.2×10^{-4}		
		(50 ~ 100) mA	5.9×10^{-5}		
		(0.1 ~ 0.2) A	2.9×10^{-4}		
		(0.2 ~ 0.4) A	1.9×10^{-4}		
		(0.4 ~ 0.6) A	1.2×10^{-4}		
		(0.6 ~ 0.8) A	8.4×10^{-5}		
		(0.8 ~ 1) A	6.6×10^{-5}		
		(1 ~ 3) A	5.1×10^{-5}		
		(3 ~ 6) A	2.6×10^{-5}		
		(6 ~ 10) A	1.6×10^{-5}		
		(10 ~ 40) A	4.0×10^{-5}		
		(40 ~ 100) A	2.9×10^{-5}		
		(100 ~ 1 000) A	1.4×10^{-4}		
		(1 000 ~ 2 000) A	4.9×10^{-4}		
		Charge voltage	0 mV	0.058 mV	
		(0 ~ 5) mV	5.8×10^{-2}		
		(5 ~ 20) mV	5.8×10^{-3}		
		(20 ~ 100) mV	1.2×10^{-3}		
		(0.1 ~ 1) V	6.2×10^{-5}		
		(1 ~ 2) V	3.2×10^{-5}		
		(2 ~ 4) V	2.1×10^{-5}		
		(4 ~ 7) V	1.3×10^{-5}		
		(7 ~ 9) V	9.1×10^{-6}		
		(9 ~ 10) V	7.9×10^{-6}		
		(10 ~ 50) V	3.1×10^{-5}		
		(50 ~ 100) V	1.0×10^{-5}		
		(100 ~ 200) V	3.4×10^{-5}		
		(200 ~ 400) V	2.5×10^{-5}		
		(400 ~ 1 000) V	1.6×10^{-5}		
		(1 000 ~ 1 200) V	1.0×10^{-3}		
		(1 200 ~ 1 400) V	9.2×10^{-4}		
		(1 400 ~ 1 500) V	8.7×10^{-4}		
		Charge and Discharge Current	(±)		
		1 mA	5.8 μA		
		(1 ~ 2) mA	2.9×10^{-3}		
		(2 ~ 5) mA	1.2×10^{-3}		
		(5 ~ 20) mA	5.8×10^{-4}		
		(20 ~ 50) mA	1.2×10^{-4}		
		(50 ~ 100) mA	5.9×10^{-5}		
		(0.1 ~ 0.2) A	2.9×10^{-4}		
		(0.2 ~ 0.4) A	1.9×10^{-4}		
		(0.4 ~ 0.6) A	1.2×10^{-4}		
		(0.6 ~ 0.8) A	8.4×10^{-5}		
		(0.8 ~ 1) A	6.6×10^{-5}		
		(1 ~ 3) A	5.1×10^{-5}		
		(3 ~ 6) A	2.6×10^{-5}		
		(6 ~ 10) A	1.6×10^{-5}		
		(10 ~ 40) A	4.0×10^{-5}		
		(40 ~ 100) A	2.9×10^{-5}		
		(100 ~ 1 000) A	1.4×10^{-4}		
		(1 000 ~ 3 000) A	4.9×10^{-4}		

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Electronic AC/DC loads	40417	Resistance	0.1 Ω	0.58 mΩ	Calibrator/ SICT-CP-40417
			(0.1 ~ 1) Ω	2.1×10^{-3}	
		(1 ~ 2) Ω	4.0×10^{-3}		
		(2 ~ 4) Ω	2.3×10^{-3}		
		(4 ~ 500) Ω	1.6×10^{-3}		
		(0.5 ~ 2) kΩ	9.0×10^{-4}		
		(2 ~ 10) kΩ	2.0×10^{-4}		
		AC Voltage			
		(0.001 V)			
		(40 ~ 400) Hz	0.61 mV		
		(0.001 ~ 0.1) V			
		(40 ~ 400) Hz	6.1×10^{-3}		
		(0.1 ~ 0.2) V			
		(40 ~ 400) Hz	3.1×10^{-3}		
		(0.2 ~ 0.5) V			
		(40 ~ 400) Hz	1.2×10^{-3}		
		(0.5 ~ 2) V			
		(40 ~ 400) Hz	6.2×10^{-4}		
		(2 ~ 3) V			
		(40 ~ 400) Hz	2.8×10^{-4}		
		(3 ~ 7) V			
		(40 ~ 400) Hz	2.3×10^{-4}		
		(7 ~ 20) V			
		(40 ~ 50) Hz	1.5×10^{-4}		
		(50 ~ 400) Hz	9.8×10^{-5}		
		(20 ~ 80) V			
		(40 ~ 50) Hz	2.1×10^{-4}		
		(50 ~ 400) Hz	1.2×10^{-4}		
		(80 ~ 200) V			
		(40 ~ 400) Hz	1.3×10^{-4}		
		(200 ~ 500) V			
		(50 ~ 400) Hz	1.8×10^{-4}		
		AC Current			
		(1 mA)			
		(40 ~ 400) Hz	0.58 mA		
		(1 ~ 100) mA			
		(40 ~ 400) Hz	5.8×10^{-2}		
		(100 mA ~ 0.2 A)			
		(40 ~ 400) Hz	5.8×10^{-3}		

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.			
Electronic AC/DC loads AC Current AC Resistance	40417	(0.2 ~ 0.6) A (40 ~ 400) Hz	2.0×10^{-3}	Calibrator/ SICT-CP-40417			
		(0.6 ~ 2) A (40 ~ 400) Hz	9.4×10^{-4}				
		(2 ~ 5) A (40 ~ 400) Hz	1.2×10^{-3}				
		(5 ~ 20) A (40 ~ 400) Hz	1.0×10^{-3}				
		(1 Ω) (40 ~ 400) Hz	1.0 mΩ				
		(1 ~ 50) Ω (40 ~ 400) Hz	1.5×10^{-3}				
		(50 ~ 100) Ω (40 ~ 400) Hz	1.1×10^{-3}				
		(100 Ω ~ 10 kΩ) (40 ~ 400) Hz	1.7×10^{-3}				
		<hr/>					
		Modulation meters Frequency Modulation Amplitude Modulation Phase Modulation	40418		0 kHz (0 ~ 400) kHz	1 Hz 1.2×10^{-2}	Measuring Receiver/ SICT-CP-40418
0 % (0 ~ 100) %	0.01 % 1.2×10^{-2}						
0 rad (0 ~ 400) rad	1.2 mrad 1.2×10^{-2}						
<hr/>							
Analogue/Digital multimeters DC Voltage	40419	(±) 0 mV	0.43 μV	Calibrator/ SICT-CP-40419			
		(0 ~ 1) mV	5.0×10^{-4}				
		(1 ~ 2) mV	2.5×10^{-4}				
		(2 ~ 5) mV	1.0×10^{-4}				
		(5 ~ 10) mV	5.0×10^{-5}				
		(10 ~ 15) mV	3.3×10^{-5}				
		(15 ~ 20) mV	2.5×10^{-6}				
		(20 ~ 50) mV	1.2×10^{-5}				
		(0.05 ~ 0.2) V	8.0×10^{-6}				
		(0.2 ~ 0.5) V	4.8×10^{-6}				
		(0.5 ~ 1) V	3.8×10^{-6}				
		(1 ~ 2) V	4.0×10^{-6}				
		(2 ~ 5) V	2.6×10^{-6}				
		(5 ~ 10) V	2.3×10^{-6}				
		(10 ~ 20) V	6.0×10^{-6}				
		(20 ~ 50) V	4.0×10^{-6}				
(50 ~ 100) V	3.5×10^{-6}						
(100 ~ 200) V	8.0×10^{-6}						
(200 ~ 500) V	5.2×10^{-6}						
(500 ~ 1 000) V	4.5×10^{-6}						

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters AC Voltage	40419	(0.6 mV)		Calibrator/ SICT-CP-40419
		1 kHz	4.1 μ V	
		(1 mV)		
		10 Hz	4.2 μ V	
		(10 ~ 40) Hz	4.2 μ V	
		(0.04 ~ 20) kHz	4.1 μ V	
		(20 ~ 50) kHz	4.2 μ V	
		(50 ~ 100) kHz	5.5 μ V	
		(100 ~ 300) kHz	11 μ V	
		(300 ~ 500) kHz	21 μ V	
		(0.5 ~ 1) MHz	23 μ V	
		(1 ~ 2) mV		
		10 Hz	2.2×10^{-3}	
		(10 ~ 40) Hz	2.2×10^{-3}	
		(0.04 ~ 20) kHz	2.1×10^{-3}	
		(20 ~ 50) kHz	2.2×10^{-3}	
		(50 ~ 100) kHz	3.0×10^{-3}	
		(100 ~ 300) kHz	6.0×10^{-3}	
		(300 ~ 500) kHz	1.1×10^{-2}	
		(0.5 ~ 1) MHz	1.3×10^{-2}	
		(2 ~ 5) mV		
		10 Hz	1.1×10^{-3}	
		(10 ~ 40) Hz	9.2×10^{-4}	
		(0.04 ~ 20) kHz	9.0×10^{-4}	
		(20 ~ 50) kHz	1.0×10^{-3}	
		(50 ~ 100) kHz	1.5×10^{-3}	
		(100 ~ 300) kHz	3.0×10^{-3}	
		(300 ~ 500) kHz	5.2×10^{-3}	
		(0.5 ~ 1) MHz	6.8×10^{-3}	
		(5 ~ 10) mV		
		10 Hz	6.3×10^{-4}	
		(10 ~ 40) Hz	5.0×10^{-4}	
		(0.04 ~ 20) kHz	4.9×10^{-4}	
		(20 ~ 50) kHz	5.9×10^{-4}	
		(50 ~ 100) kHz	9.5×10^{-4}	
		(100 ~ 300) kHz	1.9×10^{-3}	
		(300 ~ 500) kHz	3.2×10^{-3}	
		(0.5 ~ 1) MHz	4.3×10^{-3}	
		(10 ~ 15) mV		
		10 Hz	4.8×10^{-4}	
		(10 ~ 40) Hz	3.6×10^{-4}	
		(0.04 ~ 20) kHz	3.5×10^{-4}	
		(20 ~ 50) kHz	4.5×10^{-4}	
		(50 ~ 100) kHz	7.6×10^{-4}	
		(100 ~ 300) kHz	1.5×10^{-3}	
		(300 ~ 500) kHz	2.5×10^{-3}	
		(0.5 ~ 1) MHz	3.7×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters AC Voltage	40419	(15 ~ 20) mV		Calibrator/ SICT-CP-40419
		10 Hz	4.1×10^{-4}	
		(10 ~ 40) Hz	2.9×10^{-4}	
		(0.04 ~ 20) kHz	2.8×10^{-4}	
		(20 ~ 50) kHz	3.8×10^{-4}	
		(50 ~ 100) kHz	6.7×10^{-4}	
		(100 ~ 300) kHz	1.4×10^{-3}	
		(300 ~ 500) kHz	2.2×10^{-3}	
		(0.5 ~ 1) MHz	3.4×10^{-3}	
		(20 ~ 50) mV		
		10 Hz	4.6×10^{-4}	
		(10 ~ 40) Hz	2.6×10^{-4}	
		(0.04 ~ 20) kHz	2.1×10^{-4}	
		(20 ~ 50) kHz	2.7×10^{-4}	
		(50 ~ 100) kHz	6.6×10^{-4}	
		(100 ~ 300) kHz	1.0×10^{-3}	
		(300 ~ 500) kHz	1.6×10^{-3}	
		(0.5 ~ 1) MHz	3.3×10^{-3}	
		(50 ~ 100) mV		
		10 Hz	3.3×10^{-4}	
		(10 ~ 40) Hz	1.6×10^{-4}	
		(0.04 ~ 20) kHz	1.3×10^{-4}	
		(20 ~ 50) kHz	1.9×10^{-4}	
		(50 ~ 100) kHz	4.8×10^{-4}	
		(100 ~ 300) kHz	7.6×10^{-4}	
		(300 ~ 500) kHz	1.3×10^{-3}	
		(0.5 ~ 1) MHz	2.8×10^{-3}	
		(100 ~ 150) mV		
		10 Hz	2.9×10^{-4}	
		(10 ~ 40) Hz	1.3×10^{-4}	
		(0.04 ~ 20) kHz	1.0×10^{-4}	
		(20 ~ 50) kHz	1.6×10^{-4}	
		(50 ~ 100) kHz	4.1×10^{-4}	
		(100 ~ 300) kHz	6.8×10^{-4}	
		(300 ~ 500) kHz	1.3×10^{-3}	
		(0.5 ~ 1) MHz	2.6×10^{-3}	
		(150 ~ 200) mV		
		10 Hz	2.7×10^{-4}	
		(10 ~ 40) Hz	1.2×10^{-4}	
		(0.04 ~ 20) kHz	9.2×10^{-5}	
		(20 ~ 50) kHz	1.4×10^{-4}	
		(50 ~ 100) kHz	3.9×10^{-4}	
		(100 ~ 300) kHz	6.4×10^{-4}	
		(300 ~ 500) kHz	1.2×10^{-3}	
		(0.5 ~ 1) MHz	2.6×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters AC Voltage	40419	(0.2 ~ 0.5) V		Calibrator/ SICT-CP-40419
		10 Hz	3.0×10^{-4}	
		(10 ~ 20) Hz	1.3×10^{-4}	
		(20 ~ 40) Hz	9.0×10^{-5}	
		(0.04 ~ 20) kHz	6.6×10^{-5}	
		(20 ~ 50) kHz	1.0×10^{-4}	
		(50 ~ 100) kHz	1.6×10^{-4}	
		(100 ~ 300) kHz	4.6×10^{-4}	
		(300 ~ 500) kHz	1.2×10^{-3}	
		(0.5 ~ 1) MHz	2.7×10^{-3}	
		(0.5 ~ 1) V		
		10 Hz	2.5×10^{-4}	
		(10 ~ 20) Hz	1.0×10^{-4}	
		(20 ~ 40) Hz	5.8×10^{-5}	
		(0.04 ~ 20) kHz	4.9×10^{-5}	
		(20 ~ 50) kHz	7.7×10^{-5}	
		(50 ~ 100) kHz	1.2×10^{-4}	
		(100 ~ 300) kHz	3.6×10^{-4}	
		(300 ~ 500) kHz	1.0×10^{-3}	
		(0.5 ~ 1) MHz	2.5×10^{-3}	
		(1 ~ 2) V		
		10 Hz	2.2×10^{-4}	
		(10 ~ 20) Hz	8.5×10^{-5}	
		(20 ~ 40) Hz	4.5×10^{-5}	
		(0.04 ~ 20) kHz	4.2×10^{-5}	
		(20 ~ 50) kHz	6.8×10^{-5}	
		(50 ~ 100) kHz	9.6×10^{-5}	
		(100 ~ 300) kHz	3.2×10^{-4}	
		(300 ~ 500) kHz	9.0×10^{-4}	
		(0.5 ~ 1) MHz	2.4×10^{-3}	
		(2 ~ 5) V		
		10 Hz	3.0×10^{-4}	
		(10 ~ 20) Hz	1.3×10^{-4}	
		(20 ~ 40) Hz	8.2×10^{-5}	
		(0.04 ~ 20) kHz	6.2×10^{-5}	
		(20 ~ 50) kHz	1.0×10^{-4}	
		(50 ~ 100) kHz	1.4×10^{-4}	
		(100 ~ 300) kHz	4.4×10^{-4}	
		(300 ~ 500) kHz	1.3×10^{-3}	
		(0.5 ~ 1) MHz	2.2×10^{-3}	
		(5 ~ 10) V		
		10 Hz	2.5×10^{-4}	
		(10 ~ 20) Hz	9.8×10^{-5}	
		(20 ~ 40) Hz	5.4×10^{-5}	
		(0.04 ~ 20) kHz	4.7×10^{-5}	
		(20 ~ 50) kHz	7.7×10^{-5}	
		(50 ~ 100) kHz	1.1×10^{-4}	
		(100 ~ 300) kHz	3.2×10^{-4}	
(300 ~ 500) kHz	1.0×10^{-3}			
(0.5 ~ 1) MHz	1.7×10^{-3}			

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters AC Voltage	40419	(10 ~ 15) V		Calibrator/ SICT-CP-40419
		10 Hz	2.3×10^{-4}	
		(10 ~ 20) Hz	8.7×10^{-5}	
		(20 ~ 40) Hz	4.6×10^{-5}	
		(0.04 ~ 20) kHz	4.3×10^{-5}	
		(20 ~ 50) kHz	7.1×10^{-5}	
		(50 ~ 100) kHz	9.5×10^{-5}	
		(100 ~ 300) kHz	2.9×10^{-4}	
		(300 ~ 500) kHz	9.4×10^{-4}	
		(0.5 ~ 1) MHz	1.5×10^{-3}	
		(15 ~ 20) V		
		10 Hz	2.2×10^{-4}	
		(10 ~ 20) Hz	8.5×10^{-5}	
		(20 ~ 40) Hz	4.3×10^{-5}	
		(0.04 ~ 20) kHz	4.1×10^{-5}	
		(20 ~ 50) kHz	6.8×10^{-5}	
		(50 ~ 100) kHz	9.0×10^{-5}	
		(100 ~ 300) kHz	2.8×10^{-4}	
		(300 ~ 500) kHz	9.1×10^{-4}	
		(0.5 ~ 1) MHz	1.4×10^{-3}	
		(20 ~ 50) V		
		10 Hz	3.2×10^{-4}	
		(10 ~ 20) Hz	1.4×10^{-4}	
		(20 ~ 40) Hz	9.8×10^{-5}	
		(0.04 ~ 20) kHz	7.4×10^{-5}	
		(20 ~ 50) kHz	1.1×10^{-5}	
		(50 ~ 100) kHz	2.1×10^{-4}	
		(50 ~ 100) V		
		10 Hz	2.5×10^{-4}	
		(10 ~ 20) Hz	1.0×10^{-4}	
		(20 ~ 40) Hz	6.5×10^{-5}	
		(0.04 ~ 20) kHz	5.6×10^{-5}	
		(20 ~ 50) kHz	8.5×10^{-5}	
		(50 ~ 100) kHz	1.6×10^{-4}	
		(100 ~ 200) V		
		10 Hz	2.3×10^{-4}	
(10 ~ 20) Hz	9.3×10^{-5}			
(20 ~ 40) Hz	5.6×10^{-5}			
(0.04 ~ 20) kHz	5.1×10^{-5}			
(20 ~ 50) kHz	7.9×10^{-5}			
(50 ~ 100) kHz	1.4×10^{-4}			
(200 ~ 500) V				
50 Hz ~ 1 kHz	6.7×10^{-5}			
(500 ~ 1 000) V				
50 Hz ~ 1 kHz	6.3×10^{-5}			

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Analogue/Digital multimeters	40419	Resistance	0 Ω	0.001 0 mΩ	Calibrator/ SICT-CP-40419
			0 Ω ~ 10 kΩ	1.2×10^{-6}	
		(10 ~ 100) kΩ	1.4×10^{-6}		
		(0.1 ~ 1) MΩ	7.2×10^{-6}		
		(1 ~ 10) MΩ	7.7×10^{-6}		
		(10 ~ 100) MΩ	1.2×10^{-5}		
		(0.1 ~ 1) GΩ	3.2×10^{-4}		
		DC Current	(±)		
		0 nA	6.0 nA		
		(0 ~ 1) nA	6.9×10^{-3}		
		(1 ~ 100) nA	4.6×10^{-3}		
		(0.1 ~ 1) μA	6.0×10^{-3}		
		(1 ~ 2) μA	3.1×10^{-3}		
		(2 ~ 5) μA	1.2×10^{-3}		
		(5 ~ 10) μA	6.3×10^{-4}		
		(10 ~ 20) μA	3.5×10^{-4}		
		(20 ~ 50) μA	1.6×10^{-4}		
		(50 ~ 100) μA	9.4×10^{-5}		
		(100 ~ 200) μA	6.3×10^{-5}		
		(0.2 ~ 0.5) mA	4.8×10^{-5}		
		(0.5 ~ 1) mA	3.5×10^{-5}		
		(1 ~ 1.5) mA	3.1×10^{-5}		
		(1.5 ~ 2) mA	3.0×10^{-5}		
		(2 ~ 5) mA	4.4×10^{-5}		
		(5 ~ 10) mA	3.2×10^{-5}		
		(10 ~ 15) mA	2.9×10^{-5}		
		(15 ~ 20) mA	2.8×10^{-5}		
		(20 ~ 50) mA	5.6×10^{-5}		
		(50 ~ 100) mA	4.4×10^{-5}		
		(100 ~ 150) mA	4.1×10^{-5}		
		(150 ~ 200) mA	3.9×10^{-5}		
		(0.2 ~ 0.5) A	9.4×10^{-5}		
		(0.5 ~ 1) A	6.9×10^{-5}		
		(1 ~ 1.5) A	6.1×10^{-5}		
		(1.5 ~ 2) A	5.8×10^{-5}		
		(2 ~ 3) A	3.3×10^{-4}		
		(3 ~ 5) A	2.4×10^{-4}		
		(5 ~ 10) A	1.6×10^{-4}		
		(10 ~ 20) A	1.2×10^{-4}		
		(20 ~ 30) A	2.4×10^{-4}		

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters AC Current	40419	(20 μA)		Calibrator/ SICT-CP-40419
		1 kHz	11 nA	
		10 kHz	81 nA	
		(20 ~ 50) μA		
		10 Hz	1.4×10^{-4}	
		(10 ~ 20) Hz	9.0×10^{-5}	
		20 Hz ~ 1 kHz	7.3×10^{-5}	
		(1 ~ 5) kHz	1.5×10^{-4}	
		(5 ~ 10) kHz	5.5×10^{-4}	
		(50 ~ 100) μA		
		10 Hz	1.9×10^{-4}	
		(10 ~ 20) Hz	1.2×10^{-4}	
		20 Hz ~ 1 kHz	9.5×10^{-5}	
		(1 ~ 5) kHz	2.0×10^{-4}	
		(5 ~ 10) kHz	7.5×10^{-4}	
		(0.1 ~ 0.2) mA		
		10 Hz	3.0×10^{-4}	
		(10 ~ 20) Hz	1.9×10^{-4}	
		20 Hz ~ 1 kHz	1.4×10^{-5}	
		(1 ~ 5) kHz	3.1×10^{-4}	
		(5 ~ 10) kHz	1.2×10^{-3}	
		(0.2 ~ 0.5) mA		
		10 Hz	9.0×10^{-5}	
		(10 ~ 20) Hz	7.0×10^{-5}	
		20 Hz ~ 1 kHz	6.0×10^{-5}	
		(1 ~ 5) kHz	1.2×10^{-4}	
		(5 ~ 10) kHz	5.7×10^{-4}	
		(0.5 ~ 1) mA		
		10 Hz	1.4×10^{-4}	
		(10 ~ 20) Hz	1.0×10^{-4}	
		20 Hz ~ 1 kHz	8.0×10^{-5}	
		(1 ~ 5) kHz	1.6×10^{-4}	
		(5 ~ 10) kHz	7.6×10^{-4}	
		(1 ~ 2) mA		
		10 Hz	2.4×10^{-4}	
		(10 ~ 20) Hz	1.6×10^{-4}	
		20 Hz ~ 1 kHz	1.2×10^{-4}	
		(1 ~ 5) kHz	2.4×10^{-4}	
		(5 ~ 10) kHz	1.2×10^{-3}	
		(2 ~ 5) mA		
		10 Hz	9.0×10^{-5}	
		(10 ~ 20) Hz	7.0×10^{-5}	
		20 Hz ~ 1 kHz	5.2×10^{-5}	
		(1 ~ 5) kHz	1.1×10^{-4}	
		(5 ~ 10) kHz	5.4×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Analogue/Digital multimeters AC Current	40419	(5 ~ 10) mA		Calibrator/ SICT-CP-40419	
		10 Hz	1.4×10^{-4}		
		(10 ~ 20) Hz	1.0×10^{-4}		
		20 Hz ~ 1 kHz	7.3×10^{-5}		
		(1 ~ 5) kHz	1.4×10^{-4}		
		(5 ~ 10) kHz	7.2×10^{-4}		
		(10 ~ 20) mA			
		10 Hz	2.4×10^{-4}		
		(10 ~ 20) Hz	1.6×10^{-4}		
		20 Hz ~ 1 kHz	1.2×10^{-4}		
		(1 ~ 5) kHz	2.2×10^{-4}		
		(5 ~ 10) kHz	1.1×10^{-3}		
		(20 ~ 50) mA			
		10 Hz	1.0×10^{-4}		
		(10 ~ 20) Hz	8.0×10^{-5}		
		20 Hz ~ 1 kHz	4.8×10^{-5}		
		(1 ~ 5) kHz	1.1×10^{-4}		
		(5 ~ 10) kHz	4.0×10^{-4}		
		(50 ~ 100) mA			
		10 Hz	1.4×10^{-4}		
		(10 ~ 20) Hz	1.0×10^{-4}		
		20 Hz ~ 1 kHz	6.8×10^{-5}		
		(1 ~ 5) kHz	1.4×10^{-4}		
		(5 ~ 10) kHz	6.0×10^{-4}		
		(0.1 ~ 0.2) A			
		10 Hz	2.4×10^{-4}		
		(10 ~ 20) Hz	1.6×10^{-4}		
		20 Hz ~ 1 kHz	1.1×10^{-4}		
		(1 ~ 5) kHz	2.1×10^{-4}		
		(5 ~ 10) kHz	1.0×10^{-3}		
		(0.2 ~ 1) A			
		40 Hz	1.4×10^{-4}		
		40 Hz ~ 1 kHz	1.4×10^{-4}		
		(1 ~ 5) kHz	2.6×10^{-4}		
		(5 ~ 10) kHz	2.7×10^{-3}		
		(1 ~ 2) A			
		40 Hz ~ 1 kHz	2.4×10^{-4}		
		(1 ~ 5) kHz	4.2×10^{-4}		
		(5 ~ 10) kHz	5.2×10^{-3}		
		(2 ~ 3) A			
		(40 ~ 100) Hz	1.8×10^{-4}		
		100 Hz ~ 1 kHz	1.9×10^{-4}		
		(1 ~ 10) kHz	9.9×10^{-4}		

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Analogue/Digital multimeters	40419	AC Current	(3 ~ 5) A		Calibrator/ SICT-CP-40419
			(40 ~ 100) Hz	2.4×10^{-4}	
			100 Hz ~ 1 kHz	2.5×10^{-4}	
			(1 ~ 10) kHz	1.6×10^{-3}	
			(5 ~ 10) A		
			(40 ~ 100) Hz	4.2×10^{-4}	
			100 Hz ~ 1 kHz	4.2×10^{-4}	
			(1 ~ 10) kHz	3.1×10^{-3}	
			(10 ~ 20) A		
			(40 ~ 60) Hz	6.0×10^{-4}	
			(60 ~ 100) Hz	7.0×10^{-4}	
			100 Hz ~ 1 kHz	1.3×10^{-3}	
			(1 ~ 10) kHz	2.3×10^{-3}	
			(20 ~ 30) A		
			(40 ~ 60) Hz	8.0×10^{-4}	
			(60 ~ 100) Hz	1.2×10^{-3}	
			100 Hz ~ 1 kHz	3.9×10^{-3}	
			(1 ~ 10) kHz	4.5×10^{-3}	
		Frequency	10 Hz ~ 10 MHz	6.4×10^{-7}	
(Digital Sampling)		AC Voltage	(1 mV)		
			0.1 Hz ~ 3 kHz	8.4×10^{-4}	
			(1 mV ~ 2 mV)		
			0.1 Hz ~ 3 kHz	4.2×10^{-4}	
			(2 mV ~ 3 mV)		
			0.1 Hz ~ 3 kHz	2.8×10^{-4}	
			(3 mV ~ 5 mV)		
			0.1 Hz ~ 3 kHz	1.7×10^{-4}	
			(5 mV ~ 10 mV)		
			0.1 Hz ~ 3 kHz	8.8×10^{-5}	
			(10 mV ~ 20 mV)		
			0.1 Hz ~ 3 kHz	4.8×10^{-5}	
			(20 mV ~ 30 mV)		
			0.1 Hz ~ 3 kHz	3.6×10^{-5}	
			(30 mV ~ 50 mV)		
			0.1 Hz ~ 3 kHz	3.0×10^{-5}	
			(50 mV ~ 100 mV)		
			0.1 Hz ~ 3 kHz	2.6×10^{-5}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters (Digital Sampling) AC Voltage	40419	(100 mV ~ 200 mV) 0.1 Hz ~ 3 kHz	4.8×10^{-5}	Calibrator/ SICT-CP-40419
		(200 mV ~ 300 mV) 0.1 Hz ~ 3 kHz	3.6×10^{-5}	
		(300 mV ~ 500 mV) 0.1 Hz ~ 3 kHz	2.8×10^{-5}	
		(500 mV ~ 1 V) 0.1 Hz ~ 3 kHz	2.4×10^{-5}	
		(1 V ~ 2 V) 0.1 Hz ~ 3 kHz	4.8×10^{-5}	
		(2 V ~ 3 V) 0.1 Hz ~ 3 kHz	3.6×10^{-5}	
		(3 V ~ 5 V) 0.1 Hz ~ 3 kHz	3.0×10^{-5}	
		(5 V ~ 10 V) 0.1 Hz ~ 3 kHz	2.6×10^{-5}	
		(10 V ~ 30 V) 10 Hz ~ 3 kHz	3.6×10^{-5}	
		(30 V ~ 50 V) 10 Hz ~ 3 kHz	2.8×10^{-5}	
		(50 V ~ 100 V) 10 Hz ~ 3 kHz	2.4×10^{-5}	
		(100 V ~ 200 V) 10 Hz ~ 3 kHz	4.8×10^{-5}	
		(200 V ~ 1 000 V) 50 Hz ~ 1 kHz	2.4×10^{-5}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Noise meters AC Voltage Test	40420	(600 μV) 1 kHz	7.8×10^{-3}	Calibrator/ SICT-CP-40420
		(600 μV ~ 20 mV) (10 ~ 40) Hz 40 Hz ~ 20 kHz (20 ~ 100) kHz (100 ~ 300) kHz 300 kHz ~ 1 MHz	2.6×10^{-3} 2.5×10^{-3} 3.5×10^{-3} 7.0×10^{-3} 1.5×10^{-2}	
		(20 ~ 200) mV (10 ~ 40) Hz 40 Hz ~ 20 kHz (20 ~ 100) kHz 100 kHz ~ 1 MHz	5.2×10^{-4} 3.4×10^{-4} 8.8×10^{-4} 4.3×10^{-3}	
		(200 mV ~ 2 V) (10 ~ 40) Hz 40 Hz ~ 20 kHz (20 ~ 100) kHz (100 ~ 300) kHz 300 kHz ~ 1 MHz	3.5×10^{-4} 1.5×10^{-4} 4.6×10^{-4} 8.8×10^{-4} 3.4×10^{-3}	
		(2 ~ 20) V (10 ~ 40) Hz 40 Hz ~ 100 kHz (100 ~ 300) kHz 300 kHz ~ 1 MHz	3.0×10^{-4} 1.2×10^{-4} 4.4×10^{-4} 2.2×10^{-3}	
		(20 ~ 200) V (10 ~ 40) Hz 40 Hz ~ 20 kHz (20 ~ 100) kHz	3.0×10^{-4} 1.2×10^{-4} 1.8×10^{-4}	
		(200 ~ 500) V 50 Hz ~ 1 kHz	3.8×10^{-4}	
		(500 ~ 1 000) V 50 Hz ~ 1 kHz	3.7×10^{-4}	
		(25 ~ 500) mV (1 ~ 30) MHz	2.1×10^{-2}	
		(500 mV ~ 2 V) (0.1 ~ 30) MHz	2.1×10^{-2}	
Weighting Test		(DIN/NOISE) 31.5 Hz ~ 10 kHz (JIS A) 31.5 Hz ~ 16 kHz (CCIR) 31.5 Hz ~ 31.5 kHz (CCIR/ARM) 31.5 Hz ~ 31.5 kHz	0.12 dB 0.12 dB 0.12 dB 0.12 dB	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Noise meters	40420	(10 mV) 1 kHz	2.8×10^{-3}	Calibrator/ SICT-CP-40420
AC Voltage Output		(10 mV ~ 1 V) 20 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz	8.5×10^{-5} 1.3×10^{-4} 6.9×10^{-4}	
DC Voltage Output		0 mV 100 mV ~ 1 V	0.99 μ V 1.1×10^{-5}	
Oscilloscopes	40421	50 Ω 75 Ω 1 M Ω	3.5×10^{-5} 2.7×10^{-5} 2.5×10^{-5}	Calibration Generator/ SICT-CP-40421
Impedance Measure		(\pm) 0 mV (0 ~ 1) mV (1 ~ 5) mV (5 ~ 10) mV (10 ~ 100) mV (100 ~ 900) mV (0.9 ~ 9) V (9 ~ 200) V	0.79 μ V 8.0×10^{-4} 4.1×10^{-4} 1.7×10^{-4} 8.5×10^{-5} 1.5×10^{-5} 9.1×10^{-6} 9.5×10^{-6}	
DC Voltage		(1 kHz) 1 mV (1 ~ 25) mV (0.025 ~ 0.5) V (0.5 ~ 2.2) V (2.2 ~ 130) V	6.5×10^{-3} 8.8×10^{-4} 9.1×10^{-4} 6.8×10^{-4} 8.4×10^{-4}	
AC Voltage(Square wave)		100 ps (100 ~ 200) ps 200 ps ~ 20 ms 20 ms ~ 5 s	6.2×10^{-7} 3.1×10^{-7} 1.7×10^{-6} 1.6×10^{-5}	
Time Marker		(40 Hz ~ 20 kHz) 100 mV 100 mV ~ 12 V	3.2×10^{-5} 1.9×10^{-5}	
CAL Output Amplitude		100 Hz ~ 10 MHz	6.2×10^{-7}	
CAL Output Frequency		50 kHz 50 kHz ~ 1 MHz 1 MHz ~ 1 GHz (1 ~ 4) GHz (4 ~ 18) GHz (18 ~ 25) GHz (25 ~ 33) GHz (33 ~ 40) GHz	2.3×10^{-2} 4.7×10^{-2} 1.5×10^{-2} 1.8×10^{-2} 3.2×10^{-2} 5.5×10^{-2} 5.8×10^{-2} 6.0×10^{-2}	
Sinewave Signal Generator Level				

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Oscilloscopes AC Voltage(Sine wave)	40421	(10 ~ 40) Hz 2 mV (2 ~ 20) mV (20 ~ 200) mV (0.2 ~ 20) V (20 ~ 100) V (40 Hz ~ 1 kHz) 2 mV (2 ~ 20) mV (20 ~ 800) mV (0.8 ~ 20) V (20 ~ 200) V (1 ~ 50) kHz 2 mV (2 ~ 20) mV (20 ~ 200) mV (0.2 ~ 20) V (20 ~ 100) V (50 ~ 100) kHz 2 mV (2 ~ 20) mV (20 ~ 200) mV (0.2 ~ 20) V (20 ~ 100) V	5.3 μV 1.9×10^{-3} 6.8×10^{-4} 4.3×10^{-4} 4.5×10^{-4} 5.0 μV 1.7×10^{-3} 3.3×10^{-4} 1.7×10^{-4} 1.8×10^{-4} 5.2 μV 1.9×10^{-3} 4.8×10^{-4} 1.3×10^{-4} 1.7×10^{-4} 7.1 μV 2.7×10^{-3} 9.0×10^{-4} 1.9×10^{-4} 3.0×10^{-4}	Calibration Generator/ SICT-CP-40421
LF phase meters Phase Test	40422	(1 Hz ~ 200 kHz) (-180 ~ 180) °	0.074 °	Multi Function Generator/ SICT-CP-40422
Volt/Current recorders DC Voltage DC Current	40424	(±) (0 ~ 100) μV (0.1 ~ 1) mV (1 ~ 10) mV (0.01 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V (±) (0 ~ 1) nA (1 ~ 100) nA (0.1 ~ 1) μA (1 ~ 10) μA (10 ~ 100) μA (0.1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 100) A	0.51 μV 5.2×10^{-4} 5.9×10^{-5} 6.7×10^{-6} 4.3×10^{-6} 6.3×10^{-6} 8.7×10^{-6} 6.9 pA 4.6×10^{-3} 2.3×10^{-3} 7.2×10^{-4} 1.4×10^{-4} 7.6×10^{-5} 8.4×10^{-5} 1.2×10^{-4} 2.1×10^{-4}	Calibrator/ SICT-CP-40424

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Relay test sets	40425			Digital Multimeter/ SICT-CP-40425
DC Voltage		1 mV 1 mV ~ 1 V (1 ~ 100) V (100 ~ 1 000) V	6 μV 7.0×10^{-4} 7.0×10^{-5} 2.2×10^{-5}	
DC Current		1 mA 1 mA ~ 1 A (1 ~ 20) A (20 ~ 100) A	58 μA 6.0×10^{-4} 2.5×10^{-4} 4.0×10^{-4}	
AC Voltage		(1 mV) 20 Hz ~ 100 kHz (1 ~ 100) mV 20 Hz ~ 10 kHz (10 ~ 100) kHz (100 mV ~ 1 V) 20 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (1 ~ 10) V 20 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (10 ~ 100) V 20 Hz ~ 10 kHz (10 ~ 100) kHz (100 ~ 1 000) V 50 Hz ~ 10 kHz (10 ~ 30) kHz	58 μV 6.1×10^{-4} 1.2×10^{-3} 7.6×10^{-4} 1.3×10^{-3} 2.0×10^{-2} 2.2×10^{-4} 1.1×10^{-3} 2.1×10^{-2} 1.8×10^{-4} 1.1×10^{-3} 1.9×10^{-4} 4.5×10^{-4}	
AC Current		(1 mA) 40 Hz ~ 10 kHz (1 ~ 100) mA 40 Hz ~ 10 kHz (100 mA ~ 1 A) 40 Hz ~ 10 kHz (1 ~ 10) A 40 Hz ~ 10 kHz (10 ~ 100) A 40 Hz ~ 10 kHz	58 μA 8.6×10^{-4} 9.1×10^{-4} 9.9×10^{-4} 2.3×10^{-4}	
Timer		(1 ~ 100) s	5.8×10^{-6}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF signal generators	40426			Audio Analyzer, Digital Multimeter/ SICT-CP-40426
Frequency Test		(0.1 ~ 1) Hz 1 Hz ~ 100 MHz	5.8×10^{-7} 5.8×10^{-9}	
Output Level Test		(10 ~ 100) Hz 1 mV (1 ~ 10) mV 10 mV ~ 100 V	1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5}	
		(100 Hz ~ 10 kHz) 1 mV (1 ~ 10) mV 10 mV ~ 100 V	1.0×10^{-3} 1.0×10^{-4} 3.0×10^{-5}	
		(10 ~ 100) kHz 1 mV (1 ~ 10) mV 10 mV ~ 100 V	1.0×10^{-3} 1.0×10^{-4} 8.0×10^{-5}	
DC Offset		(±) 0 mV (0 ~ 1) mV (1 ~ 10) mV 10 mV ~ 20 V	0.7 μV 0.7 μV 1.0×10^{-4} 6.0×10^{-5}	
Output Level Flatness Test		(100 mV) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	0.099 dB 0.083 dB 0.095 dB	
		(100 mV ~ 1 V) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	0.005 4 dB 0.001 1 dB 0.007 2 dB	
		(1 ~ 30) V (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	0.021 dB 0.015 dB 0.027 dB	
Attenuator Test		(10 Hz ~ 100 kHz) (0 ~ 80) dB	0.006 1 dB	
Distortion		(20 Hz ~ 1 kHz) (3.16 ~ 0.010) % (1 ~ 100) kHz (3.16 ~ 0.010) %	1.5×10^{-1} 3.2×10^{-1}	
Rise/Fall Time		100 μs ~ 100 ns (100 ~ 10) ns (10 ~ 1) ns 1 ns ~ 100 ps	7.0×10^{-4} 7.8×10^{-4} 4.7×10^{-3} 4.6×10^{-2}	
Duty cycle		(1 ~ 99) %	0.006 1 %	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF spectrum analyzers	40427			Synthesizer Function Generator/ SICT-CP-40427
Reference Frequency		10 MHz	7.7×10^{-12}	
Center Frequency		10 Hz	6.1×10^{-5}	
		(10 ~ 100) Hz	6.1×10^{-6}	
		100 Hz ~ 1 kHz	6.1×10^{-7}	
		1 kHz ~ 1 MHz	6.1×10^{-8}	
		(1 ~ 100) MHz	6.1×10^{-9}	
		100 MHz ~ 1 GHz	6.1×10^{-8}	
Frequency Range		10 Hz	1.1×10^{-3}	
		(10 ~ 100) Hz	1.1×10^{-4}	
		100 Hz ~ 1 kHz	1.1×10^{-5}	
		(1 ~ 100) kHz	1.1×10^{-4}	
		100 kHz ~ 1 MHz	1.1×10^{-6}	
		(1 ~ 100) MHz	1.1×10^{-7}	
		100 MHz ~ 1 GHz	1.1×10^{-9}	
Resolution Bandwidth		100 Hz	3.3×10^{-2}	
		100 Hz ~ 3 kHz	3.1×10^{-2}	
		(3 ~ 300) kHz	3.3×10^{-2}	
		300 kHz ~ 1 MHz	3.4×10^{-2}	
Absolute Amplitude		(-60 dBV)		
		10 Hz	0.043 dB	
		10 Hz ~ 10 kHz	0.042 dB	
		(10 ~ 100) kHz	0.056 dB	
		(-60 ~ -50) dBV		
		10 Hz	0.017 dB	
		10 Hz ~ 10 kHz	0.015 dB	
		(10 ~ 100) kHz	0.022 dB	
		(-50 ~ -40) dBV		
		10 Hz	0.009 dB	
		10 Hz ~ 10 kHz	0.009 8 dB	
		(10 ~ 100) kHz	0.012 dB	
		(-40 ~ -30) dBV		
		10 Hz	0.009 dB	
		10 Hz ~ 10 kHz	0.006 8 dB	
		(10 ~ 100) kHz	0.011 dB	
	(-30 ~ 30) dBV			
	10 Hz	0.016 dB		
	10 Hz ~ 10 kHz	0.006 3 dB		
	(10 ~ 100) kHz	0.007 3 dB		
Referency Level		(-60 dBV)		
		10 Hz ~ 100 kHz	0.17 dB	
		(-60 ~ 30) dBV		
	10 Hz ~ 100 kHz	0.16 dB		

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF spectrum analyzers	40427			Synthesizer Function Generator/ SICT-CP-40427
Frequency Response		10 Hz 10 Hz ~ 100 kHz	0.009 1 dB 0.008 7 dB	
Logscale Fidelity		(0 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB (-80 ~ -90) dB	0.009 2 dB 0.012 dB 0.016 dB 0.042 dB	
Output frequency		10 Hz ~ 300 MHz	6.1×10^{-11}	
Input Impedance		1 M Ω (50 ~ 75) Ω	0.000 12 M Ω 0.000 7 Ω	
Output Voltage		10 mV 10 mV ~ 5 V	0.000 38 mV 8.8×10^{-5}	
Output Offset Voltage		(-30 ~ 30) V	6.7×10^{-6}	
Output Voltage Flatness		10 Hz ~ 100 kHz	0.000 67 dB	
Spot generators	40428			Audio Analyzer, Digital Multimeter/ SICT-CP-40428
Frequency		(0.1 ~ 1) Hz 1 Hz ~ 100 kHz	5.8×10^{-7} 5.8×10^{-9}	
Output Level		(10 ~ 100) Hz 1 mV (1 ~ 10) mV 10 mV ~ 10 V	1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5}	
		(100 Hz ~ 10 kHz) 1 mV (1 ~ 10) mV 10 mV ~ 10 V	1.0×10^{-3} 1.0×10^{-4} 3.0×10^{-5}	
		(10 ~ 100) kHz 1 mV (1 ~ 10) mV 10 mV ~ 10 V	1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5}	
Output Level Flatness		(100 mV) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	 0.099 dB 0.083 dB 0.095 dB	
		(100 mV ~ 1 V) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	 0.005 4 dB 0.001 1 dB 0.007 2 dB	
		(1 ~ 10) V (10 ~ 100) Hz 100 Hz ~ 100 kHz	 0.010 dB 0.011 dB	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Spot generators	40428	Attenuation (10 Hz ~ 100 kHz) (0 ~ 80) dB	0.006 1 dB	Audio Analyzer, Digital Multimeter/ SICT-CP-40428
		Distortion (20 Hz ~ 1 kHz) (3.16 ~ 0.010) %	1.5×10^{-1}	
		(1 ~ 100) kHz (3.16 ~ 0.010) %	3.2×10^{-1}	
Sweep generators	40429	Frequency (0.1 ~ 1) Hz 1 Hz ~ 100 kHz	5.8×10^{-7} 5.8×10^{-9}	Audio Analyzer, Digital Multimeter/ SICT-CP-40429
		Output Level Test (10 ~ 100) Hz 1 mV (1 ~ 10) mV 100 mV ~ 10 V	1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5}	
		(100 Hz ~ 10 kHz) 1 mV (1 ~ 100) mV 100 mV ~ 10 V	1.0×10^{-3} 1.0×10^{-4} 3.0×10^{-5}	
		(10 ~ 100) kHz 1 mV (1 ~ 100) mV 100 mV ~ 10 V	1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5}	
		Output Level Flatness (100 mV) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	0.099 dB 0.083 dB 0.095 dB	
		(100 mV ~ 1 V) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	0.005 4 dB 0.001 1 dB 0.007 2 dB	
		(1 ~ 10) V (10 ~ 100) Hz 100 Hz ~ 100 kHz	0.010 dB 0.011 dB	
		Attenuation (10 Hz ~ 10 kHz) (0 ~ 80) dB	0.006 1 dB	
		Distortion (20 Hz ~ 1 kHz) (3.16 ~ 0.010) %	1.5×10^{-1}	
		(1 ~ 100) kHz (3.16 ~ 0.010) %	3.2×10^{-1}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Signal transducers	40430			Digital Multimeter/ SICT-CP-40430
Voltage		1 mV (1 ~ 10) mV 10 mV ~ 100 V (100 ~ 300) V	5.2×10^{-4} 8.8×10^{-5} 3.1×10^{-5} 4.3×10^{-3}	
Current		10 μ A (10 ~ 100) μ A 100 μ A ~ 100 mA 100 mA ~ 20 A	9.5×10^{-4} 9.7×10^{-5} 7.0×10^{-5} 2.0×10^{-4}	
Frequency		(1 ~ 10) Hz 10 Hz ~ 100 kHz	3.1×10^{-4} 7.2×10^{-5}	
Transistor curve tracers	40432			Digital Multimeter/ SICT-CP-40432
DC Voltage(Source)		0 mV (0 ~ 100) mV 100 mV ~ 1 V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V	1.0 μ V 1.1×10^{-5} 1.0×10^{-5} 9.8×10^{-6} 1.1×10^{-5} 8.0×10^{-6}	
DC Current(Source)		0 nA (0 ~ 1) nA (1 ~ 100) nA (0.1 ~ 1) μ A (1 ~ 10) μ A 10 μ A ~ 10 mA (10 ~ 100) mA 100 mA ~ 10 A	0.12 nA 1.0×10^{-2} 8.0×10^{-3} 8.0×10^{-4} 9.0×10^{-5} 1.6×10^{-5} 5.0×10^{-5} 2.5×10^{-4}	
DC Voltage(Measure)		0 mV (0 ~ 100) mV 100 mV ~ 1 V (1 ~ 100) V (100 ~ 1 000) V	0.78 μ V 1.5×10^{-5} 6.0×10^{-4} 6.4×10^{-4} 6.2×10^{-5}	
AC/DC high voltage generators	40434			High Voltage Digital Meter/ SICT-CP-40434
DC Voltage		(\pm) 0 kV (0 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 2) kV (2 ~ 100) kV (100 ~ 200) kV	0.58 V 1.2×10^{-3} 6.1×10^{-4} 3.0×10^{-4} 2.3×10^{-4} 1.2×10^{-2}	
AC Voltage		(50 Hz) 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 100) kV (100 ~ 200) kV (60 Hz) 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 100) kV (100 ~ 200) kV	0.58 V 1.2×10^{-3} 6.1×10^{-4} 5.7×10^{-4} 1.2×10^{-2} 0.58 V 1.2×10^{-3} 6.2×10^{-4} 5.2×10^{-4} 1.2×10^{-2}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC/DC high voltage probes	40435			DC Power Supply/ SICT-CP-40435
DC Voltage		(±) 0 kV (0 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 100) kV	0.06 V 4.2×10^{-4} 2.1×10^{-4} 3.5×10^{-4}	
AC Voltage		(50 Hz) 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 3) kV (3 ~ 5) kV (5 ~ 10) kV (10 ~ 20) kV (20 ~ 100) kV (60 Hz) 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 3) kV (3 ~ 5) kV (5 ~ 10) kV (10 ~ 20) kV (20 ~ 100) kV	2.2 V 4.4×10^{-3} 2.3×10^{-3} 1.2×10^{-3} 7.5×10^{-4} 6.0×10^{-4} 4.0×10^{-4} 3.7×10^{-4} 2.2 V 4.4×10^{-3} 2.3×10^{-3} 1.2×10^{-3} 7.5×10^{-4} 6.0×10^{-4} 4.0×10^{-4} 3.7×10^{-4}	
Logic analyzers	40436			Calibrator/ SICT-CP-40436
DC Voltage		(0 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V	1.5 μV 8.8×10^{-6} 7.2×10^{-6}	
Clock frequency		10 MHz	7.7×10^{-12}	
Telephone testers	40437			Tone Pulse Simulator/ SICT-CP-40437
L1, L2 Output Voltage		(1 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 1 000) V	9.9×10^{-4} 1.1×10^{-5} 7.1×10^{-6} 8.5×10^{-6}	
Loop Current		(0.1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A	2.5×10^{-5} 5.2×10^{-5} 2.2×10^{-4} 4.7×10^{-4}	
Ring Output Voltage		(10 Hz ~ 20 kHz) 100 mV ~ 1 V (1 ~ 100) V (100 ~ 1 000) V	4.7×10^{-4} 2.9×10^{-4} 3.1×10^{-4}	
Ring Frequency		(1 ~ 1 000) Hz	7.0×10^{-5}	
D.T.M.F & Pulse		(+10 ~ -39.9) dBm	0.09 dB	
D.T.M.F & Frequency		(697 ~ 1 477) Hz	0.59 Hz	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Video signal analyzers	40438			Video Amplitude Calibration
Color Bar Decoding Accuracy(Gain)		(0 ~ 5) mV	2.5×10^{-1}	Fixture/ SICT-CP-40438
		(5 ~ 10) mV	5.0×10^{-2}	
		(10 ~ 100) mV	8.2×10^{-2}	
		(100 ~ 200) mV	4.1×10^{-2}	
		(200 ~ 500) mV	2.1×10^{-2}	
		(500 ~ 1 000) mV	8.5×10^{-3}	
Frequency		20 Hz ~ 5 MHz	5.8×10^{-6}	
Color Bar Decoding Accuracy(Phase)		(0 ~ 360)°	0.70°	
Measure Square Wave		(0 ~ 5) mV	9.4×10^{-2}	
		(5 ~ 10) mV	2.0×10^{-2}	
		(10 ~ 100) mV	9.9×10^{-3}	
		(100 ~ 300) mV	2.1×10^{-3}	
		(300 ~ 400) mV	1.5×10^{-3}	
		(400 ~ 600) mV	1.2×10^{-3}	
		(600 ~ 999.9) mV	9.4×10^{-4}	
Measure Sine Wave		No Filter, PAL NTS BW Lim, NTSC,PAL Chroma BP, NTSC,PAL (10 kHz ~ 10 MHz) 500 mV	7.0×10^{-3}	
Burst Frequency		(3 ~ 5) MHz	4.0×10^{-7}	
vertical Gain		(0 ~ 5) mV	9.4×10^{-2}	
		(5 ~ 10) mV	2.0×10^{-2}	
		(10 ~ 100) mV	9.9×10^{-3}	
		(100 ~ 300) mV	2.1×10^{-3}	
		(300 ~ 600) mV	1.5×10^{-3}	
		(600 ~ 999.9) mV	9.4×10^{-4}	
Horizontal Frequency		(20 ~ 100) Hz	3.1×10^{-3}	
		100 Hz ~ 10 kHz	6.1×10^{-4}	
		10 kHz ~ 10 MHz	6.1×10^{-5}	
Gain Frequency Response		Flat, Luminance, Chroma at (20 Hz ~ 20 MHz) 700 mV	7.0×10^{-3}	
Transient Response		(0 ~ 1 000) mV	1.3×10^{-2}	
(Video Noise)				
Luminance Volt Level		(0 ~ -30) dB	4.8×10^{-1}	
Chrominance AM/PM Level		(0 ~ -30) dB	6.7×10^{-1}	
Luminance Volt Level		(0 ~ 1 000) mV	1.7×10^{-5}	
Luminance Input Level		(0 ~ 1 000) mV	1.8×10^{-5}	
Chrominance Input Level		(0 ~ 1 000) mV	1.7×10^{-5}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Flux meters Flux	40503	0.1 mWb ~ 10 Wb	5.8×10^{-3}	Flux sources/ SICT-CP-40503
Flux sources Flux	40504	(0.1 ~ 50) mWb (0.05 ~ 0.1) Wb (0.1 ~ 10) Wb	6.6×10^{-4} 2.3×10^{-5} 1.4×10^{-5}	GPS receiver, Frequency counter/ SICT-CP-40504
Magnetometers Gauss	40508	(0 ~ 0.1) mT (0.1 ~ 0.5) mT (0.5 ~ 3) mT (3 ~ 5) mT (5 ~ 20) mT (20 ~ 30) mT (30 ~ 1 700) mT	7.1×10^{-2} 1.4×10^{-2} 7.0×10^{-3} 4.0×10^{-3} 3.0×10^{-3} 6.7×10^{-3} 6.4×10^{-3}	Helmholtz coil, Standard magnets/ SICT-CP-40508
Reference/standard magnets Gauss	40510	(1.5 ~ 30) mT (30 ~ 1 000) mT	7.3×10^{-3} 2.6×10^{-3}	Gaussmeters/ SICT-CP-40510

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF amplifiers	40601	Gain (0 ~ 80) dB 20 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 GHz (1 ~ 10) GHz (10 ~ 18) GHz (18 ~ 40) GHz	0.01 dB	Power Sensor, Attenuator/ SICT-CP-40601
			0.02 dB	
			0.10 dB	
			0.11 dB	
			0.14 dB	
			0.19 dB	
		Harmonics (9 kHz ~ 26.5 GHz) (0 ~ -100) dBc	0.98 dB	
		Reflection coefficient (0 ~ 1) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	4.2×10^{-3}	
			9.4×10^{-3}	
			1.5×10^{-2}	
		SWR (1 ~ ∞) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	9.0×10^{-3}	
	2.1×10^{-2}			
	3.3×10^{-2}			
Coaxial attenuators	40602	Attenuation (DC ~ 26.5 GHz) (0 ~ 10) dB (10 ~ 20) dB (20 ~ 30) dB (30 ~ 40) dB (40 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB (80 ~ 100) dB (100 ~ 110) dB (110 ~ 120) dB	0.040 dB	Power Sensor, Directional Coupler/ SICT-CP-40602
			0.042 dB	
			0.045 dB	
			0.050 dB	
			0.055 dB	
			0.074 dB	
			0.076 dB	
			0.089 dB	
			0.12 dB	
			0.14 dB	
			0.15 dB	
			(26.5 ~ 34) GHz (0 ~ 40) dB (40 ~ 50) dB (50 ~ 60) dB	
		0.32 dB		
		0.37 dB		
		(34 ~ 40) GHz (0 ~ 40) dB (40 ~ 50) dB (50 ~ 60) dB	0.35 dB	
			0.45 dB	
			0.53 dB	
	Reflection coefficient (0 ~ 1) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	4.2×10^{-3}		
		9.4×10^{-3}		
		1.5×10^{-2}		
	SWR (1 ~ ∞) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	9.7×10^{-3}		
		2.4×10^{-2}		
		3.8×10^{-2}		

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Burst pulse generators Burst Volatage	40605	50 Ω (±) 5 V (5 ~ 100) V (0.1 ~ 8) kV	2.0×10^{-2} 1.6×10^{-2} 2.0×10^{-2}	Digital Oscilloscope/ SICT-CP-40605
Rise/Fall Time		1 kΩ (±) 5 V ~ 8 kV	4.0×10^{-2}	
Pulse Width		1 ns (1 ~ 2) ns (2 ~ 4) ns 4 ns ~ 1 μs (1 ~ 2) μs (2 ~ 4) μs (4 ~ 10) μs	2.0×10^{-2} 6.8×10^{-3} 2.6×10^{-3} 1.5×10^{-3} 6.2×10^{-3} 2.6×10^{-3} 1.3×10^{-3}	
Time measurement by section		1 ns (1 ~ 2) ns 2 ns ~ 200 ms	6.0×10^{-3} 3.1×10^{-3} 1.5×10^{-3}	
Repeat Frequency		1 Hz ~ 25 MHz	1.6×10^{-3}	
Attenuator calibrators Attenuation	40606	(0 ~ 10) dB (10 ~ 20) dB (20 ~ 30) dB (30 ~ 40) dB (40 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB (80 ~ 90) dB (90 ~ 100) dB (100 ~ 110) dB (110 ~ 120) dB	0.024 dB 0.025 dB 0.027 dB 0.029 dB 0.031 dB 0.034 dB 0.036 dB 0.039 dB 0.042 dB 0.045 dB 0.048 dB 0.052 dB	Verification Kit/ SICT-CP-40606
RF power meter calibrators Output Power	40607	3 μW 10 μW 30 μW 100 μW 300 μW 1 mW 3 mW 10 mW 30 mW 100 mW	0.1 nW 0.2 nW 0.5 nW 1 nW 4 nW 0.18 μW 0.19 μW 0.2 μW 0.3 μW 1 μW	Digital Multimeter/ SICT-CP-40607

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
EMC transducers ; current probes, absorbing clamps, etc Transfer Impedance	40608	10 Hz ~ 50 MHz (50 ~ 200) MHz 200 MHz ~ 3 GHz	0.60 dB 1.1 dB 1.9 dB	Power Senso, Network analyzer/ SICT-CP-40608
Insertion Loss		30 MHz ~ 1 GHz	1.9 dB	
Electric Magnetic Near-Field		100 kHz ~ 1 GHz	1.9 dB	
Reflection coefficient		(0 ~ 1) 10 Hz ~ 1 GHz (1 ~ 3) GHz	4.2×10^{-3} 6.0×10^{-3}	
SWR		(1 ~ ∞) 10 Hz ~ 1 GHz (1 ~ 3) GHz	9.0×10^{-3} 1.3×10^{-2}	
Coaxial directional couplers/ splitters Coupling Factor	40610	(0 ~ 30) dB 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 3 GHz (3 ~ 8) GHz (8 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz	0.005 9 dB 0.007 3 dB 0.087 dB 0.095 dB 0.11 dB 0.14 dB 0.20 dB	Power Sensor, Synthesized Sweeper/ SICT-CP-40610
Reflection coefficient		(0 ~ 1) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	4.2×10^{-3} 9.4×10^{-3} 1.5×10^{-2}	
SWR		(1 ~ ∞) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	9.7×10^{-3} 2.4×10^{-2} 3.8×10^{-2}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.		
Electrostatic discharge generators Peak Current(Ip)	40613	(±)		Digital Oscilloscope/ SICT-CP-40613		
		(3.75 ~ 7.5) A	5.1×10^{-2}			
		(7.5 ~ 15) A	5.3×10^{-2}			
		(15 ~ 22.5) A	4.6×10^{-2}			
		(22.5 ~ 56.3) A	5.2×10^{-2}			
		(56.3 ~ 93.8) A	4.9×10^{-2}			
		(93.8 ~ 112.5) A	5.7×10^{-2}			
		(112.5 ~ 150) A	5.2×10^{-2}			
		Current I1 (30 ~ 60) ns	(±)			
			2 A		4.5×10^{-2}	
			(2 ~ 4) A		5.0×10^{-2}	
			(4 ~ 8) A		5.3×10^{-2}	
			(8 ~ 16) A		4.9×10^{-2}	
			(16 ~ 36) A		5.0×10^{-2}	
			(36 ~ 50) A		4.4×10^{-2}	
			(50 ~ 60) A		5.7×10^{-2}	
		Current I2 (60 ~ 130) ns	(±)			
			1 A		5.0×10^{-2}	
			(1 ~ 2) A		5.4×10^{-2}	
			(2 ~ 4) A		5.7×10^{-2}	
			(4 ~ 6) A		4.9×10^{-2}	
			(6 ~ 8) A		5.4×10^{-2}	
			(8 ~ 15) A		6.5×10^{-2}	
			(15 ~ 25) A		5.2×10^{-2}	
		Current I3 (360 ~ 800) ns	(±)			
			0.275 A		1.5×10^{-1}	
			(0.275 ~ 0.55) A		2.2×10^{-1}	
			(0.55 ~ 1.1) A		1.9×10^{-1}	
(1.1 ~ 1.65) A	1.6×10^{-1}					
(1.65 ~ 4.13) A	2.3×10^{-1}					
(4.13 ~ 6.88) A	1.5×10^{-1}					
(6.88 ~ 8.25) A	2.4×10^{-1}					
Current I4 (30 ~ 65) ns	(±)					
	0.15 A	1.3×10^{-1}				
	(0.15 ~ 0.3) A	1.9×10^{-1}				
	(0.3 ~ 0.6) A	3.0×10^{-1}				
	(0.6 ~ 1.2) A	2.1×10^{-1}				
	(1.2 ~ 2.25) A	2.5×10^{-1}				
	(2.25 ~ 2.7) A	2.0×10^{-1}				
	(2.7 ~ 3.75) A	1.5×10^{-1}				
(3.75 ~ 4.5) A	2.9×10^{-1}					
(4.5 ~ 6) A	2.0×10^{-1}					

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrostatic discharge generators Semiconductor Peak Current HBM	40613	(±) (0.083 ~ 0.17) A	4.8×10^{-2}	Digital Oscilloscope/ SICT-CP-40613
		(0.17 ~ 0.33) A	3.6×10^{-2}	
		(0.33 ~ 0.67) A	3.4×10^{-2}	
		(0.67 ~ 1.33) A	3.2×10^{-2}	
		(1.33 ~ 6.66) A	2.7×10^{-2}	
Semiconductor Peak Current MM		(±) (0.88 ~ 1.75) A	6.6×10^{-2}	
		(1.75 ~ 14) A	3.5×10^{-2}	
		(14 ~ 17.5) A	3.1×10^{-2}	
		(17.5 ~ 26.25) A	2.7×10^{-2}	
		(26.25 ~ 35) A	3.0×10^{-2}	
Time		0.1 ns	2.7×10^{-2}	
		0.1 ns ~ 1 ms	2.4×10^{-2}	
Peak Voltage		(±) 1 kV	3.0×10^{-2}	
	(1 ~ 35) kV	2.5×10^{-2}		
EMC receivers	40614	100 kHz ~ 1 GHz	6.1×10^{-10}	Network Analyzer , Pulse Generator/ SICT-CP-40614
Frequency accuracy		9 kHz ~ 50 MHz	0.011	
SWR		50 MHz ~ 8 GHz	0.028	
		(8 ~ 19) GHz	0.035	
		(19 ~ 26) GHz	0.045	
		(26 ~ 40) GHz	0.064	
IF Band Accuracy		1 Hz ~ 10 MHz	65 mHz	
IF Band Selectivity		1 Hz ~ 10 MHz	6.7×10^{-4}	
IF Band Linearity		1 Hz ~ 1 GHz	0.12 dB	
Frequency Response		10 Hz ~ 5 kHz	0.01 dB	
		5 kHz ~ 1 GHz	0.07 dB	
		(1 ~ 10) GHz	0.08 dB	
		(10 ~ 18) GHz	0.09 dB	
		(18 ~ 26.5) GHz	0.11 dB	
		(26.5 ~ 40) GHz	0.16 dB	
		(40 ~ 50) GHz	0.20 dB	
Frequency Response (CISPR)		9 kHz ~ 1 GHz	0.80 dB	
Display linearity accuracy	(80 ~ 50) dBμV	0.10 dB		
	(50 ~ 0) dBμV	0.14 dB		
Input Attenuation	(0 ~ 30) dB	0.15 dB		
	(30 ~ 70) dB	0.12 dB		
	(70 ~ 110) dB	0.10 dB		
Noise Indicator	DC ~ 26.5 GHz	0.16 dB		
Interference Immunity	9 kHz ~ 40 GHz	0.67 dB		

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF filters	40615	Reject Frequency	(9 ~ 90) kHz 0.024 kHz (90 ~ 900) kHz 0.24 kHz 900 kHz ~ 900 MHz 0.025 MHz 900 MHz ~ 18 GHz 0.068 MHz (18 ~ 50) GHz 0.12 MHz	Network Analyzer/ SICT-CP-40615
Insertion Loss		(9 kHz ~ 8 GHz)	(0 ~ 10) dB 0.13 dB (10 ~ 20) dB 0.14 dB (20 ~ 40) dB 0.15 dB (40 ~ 50) dB 0.16 dB (50 ~ 60) dB 0.18 dB (60 ~ 70) dB 0.23 dB (70 ~ 80) dB 0.66 dB (80 ~ 90) dB 1.7 dB (90 ~ 100) dB 4.3 dB (8 ~ 18) GHz (0 ~ 10) dB 0.23 dB (10 ~ 30) dB 0.24 dB (30 ~ 50) dB 0.25 dB (50 ~ 60) dB 0.26 dB (60 ~ 70) dB 0.31 dB (70 ~ 80) dB 0.73 dB (80 ~ 90) dB 1.7 dB (90 ~ 100) dB 4.3 dB (18 ~ 50) GHz (0 ~ 10) dB 0.48 dB (10 ~ 20) dB 0.51 dB (20 ~ 30) dB 0.52 dB (20 ~ 40) dB 0.53 dB (40 ~ 50) dB 0.54 dB (50 ~ 60) dB 0.59 dB (60 ~ 70) dB 0.78 dB (70 ~ 80) dB 1.6 dB (80 ~ 90) dB 2.8 dB (90 ~ 100) dB 6.0 dB	
RF impedance meters	40616	RF Level	(100 kHz ~ 18 GHz) (35 ~ 20) dBm 0.11 dB	Performance Kit/ SICT-CP-40616
			(20 Hz ~ 18 GHz) (20 ~ -70) dBm 0.12 dB	
Frequency		9 kHz ~ 0.1 MHz 0.1 MHz ~ 18 GHz 6.8 × 10 ⁻¹⁰ 6.2 × 10 ⁻¹¹		
Load Measurement		DC 10 Hz ~ 100 MHz 0.02 Ω 0.06 Ω (100 ~ 500) MHz 0.15 Ω 500 MHz ~ 1.8 GHz 0.21 Ω (1.8 ~ 3.0) GHz 0.41 Ω (3.0 ~ 18) GHz 1.1 Ω		

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF impulse generators Pulse Level	40617	9 kHz ~ 1 GHz	0.33 dB	Digital Oscilloscope/ SICT-CP-40617
Line impedance stabilization networks; LISN, CDN, ISN, etc. Impedance Phase Insertion Loss Decoupling attenuation(Isolation) Coupling/Decoupling network(Impedance) Coupling/Decoupling network (Insertion loss)	40618	9 kHz ~ 1 000 MHz 9 kHz ~ 1 000 MHz (0 ~ 100) dB 9 kHz ~ 100 MHz (100 ~ 1 000) MHz (0 ~ 100) dB (9 ~ 30) kHz (0.03 ~ 20) MHz (20 ~ 1 000) MHz 9 kHz ~ 1 000 MHz (0 ~ 100) dB 9 kHz ~ 1 000 MHz	2.0×10^{-2} 1.2 ° 0.07 dB 0.08 dB 0.21 dB 0.20 dB 0.21 dB 2.0×10^{-2} 0.10 dB	Impedance/Gain-Phase Analyzer, Calibration Kit/ SICT-CP-40618
Coaxial standard mismatches Reflection coefficient(<i>r</i>) SWR	40619	(0 ~ 1) (9 ~ 100) kHz 100 kHz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (1 ~ ∞) (9 ~ 100) kHz 100 kHz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz	0.004 3 0.006 0 0.009 5 0.016 0.008 6 0.012 0.019 0.032	Network Analyzer, Calibration Kit/ SICT-CP-40619
Mobile communication test sets RF Output Level	40621	(35 ~ 20) dBm 100 kHz ~ 18 GHz) (20 ~ -20) dBm 9 kHz ~ 1 GHz (1 ~ 18) GHz (18 ~ 40) GHz (40 ~ 50) GHz (-20 ~ -60) dBm 9 kHz ~ 1 GHz (1 ~ 18) GHz (18 ~ 40) GHz (40 ~ 50) GHz (-60 ~ -70) dBm 9 kHz ~ 1 GHz (1 ~ 18) GHz (18 ~ 40) GHz (40 ~ 50) GHz (-70 ~ -120) dBm 9 kHz ~ 26.5 GHz	0.08 dB 0.06 dB 0.11 dB 0.14 dB 0.18 dB 0.06 dB 0.11 dB 0.16 dB 0.21 dB 0.08 dB 0.10 dB 0.18 dB 0.21 dB 0.15 dB	Measuring Receiver, RF Signal Generator/ SICT-CP-40621

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Mobile communication test sets	40621			Measuring Receiver, RF Signal Generator/ SICT-CP-40621
Amplitude Modulation		(0.1 ~ 100) %	1.2×10^{-2}	
Frequency Modulation		(0.1 ~ 400) kHz	1.2×10^{-2}	
Phase Modulation		(0.1 ~ 400) rad	1.2×10^{-2}	
Distortion Harmonics of Modulation Rate Signal		≤ 20 %	2.3×10^{-2}	
Hamonics		(0 ~ -90) dB	0.36 dB	
Frequency Output Accuracy		9 kHz ~ 40 GHz	7.9×10^{-11}	
AC Output Level		(10 Hz ~ 100 kHz) (1 ~ 100) mV 100 mV ~ 100 V	5.2×10^{-4} 5.8×10^{-4}	
DC Output Level		1 mV ~ 1 V (1 ~ 10) V (10 ~ 100) V	1.5×10^{-6} 1.3×10^{-6} 1.8×10^{-6}	
AC Input Level		(10 Hz ~ 100 kHz) (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V	7.6×10^{-3} 1.3×10^{-3} 6.7×10^{-4} 1.7×10^{-4} 2.0×10^{-4}	
DC Input Level		(1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 100) V	1.3×10^{-3} 3.1×10^{-4} 6.2×10^{-5} 6.1×10^{-5}	
RF Input Level		(9 kHz ~ 18 GHz) (10 ~ -70) dBm (18 ~ 40) GHz (10 ~ -70) dBm	0.10 dB 0.16 dB	
Modulation meters	40622			Measuring Receiver/ SICT-CP-40622
Amplitude Modulation		0 kHz (0 ~ 400) kHz	0.01 % 1.2×10^{-2}	
Frequency Modulation		0 % (0 ~ 100) %	1 Hz 1.2×10^{-2}	
Phase Modulation		0 rad (0 ~ 400) rad	1.2 mrad 1.2×10^{-2}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Network analyzers	40623			Power Sensor, Verification Kit/ SICT-CP-40623
Frequency		10 Hz ~ 40 GHz	6.8×10^{-10}	
Source Power Level		(20 ~ -30) dBm		
		10 Hz ~ 100 Hz	0.01 dB	
		100 Hz ~ 500 MHz	0.08 dB	
		500 MHz ~ 10 GHz	0.09 dB	
		(10 ~ 18) GHz	0.11 dB	
		(18 ~ 27) GHz	0.13 dB	
		(27 ~ 40) GHz	0.16 dB	
		(40 ~ 50) GHz	0.21 dB	
		(50 ~ 80) GHz	0.45 dB	
		(80 ~ 110) GHz	0.53 dB	
		(-30 ~ -60) dBm		
		9 kHz ~ 500 MHz	0.08 dB	
		500 MHz ~ 10 GHz	0.09 dB	
		(10 ~ 18) GHz	0.11 dB	
		(18 ~ 27) GHz	0.13 dB	
		(27 ~ 40) GHz	0.16 dB	
Dynamic Range		(100 kHz ~ 18 GHz)		
		(0 ~ 10) dB	0.086 dB	
		(10 ~ 20) dB	0.087 dB	
		(20 ~ 30) dB	0.088 dB	
		(30 ~ 40) dB	0.091 dB	
		(40 ~ 50) dB	0.099 dB	
		(50 ~ 60) dB	0.11 dB	
		(60 ~ 70) dB	0.12 dB	
		(70 ~ 80) dB	0.14 dB	
		(80 ~ 90) dB	0.15 dB	
		(90 ~ 100) dB	0.18 dB	
Attenuation		(20 dB)		
		300 kHz ~ 1.5 GHz	0.050 dB	
		(1.5 ~ 8) GHz	0.051 dB	
		(8 ~ 18) GHz	0.055 dB	
		(18 ~ 26.5) GHz	0.067 dB	
		(40 dB)		
		300 kHz ~ 1.5 GHz	0.054 dB	
		(1.5 ~ 8) GHz	0.055 dB	
		(8 ~ 18) GHz	0.059 dB	
		(18 ~ 26.5) GHz	0.082 dB	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Network analyzers	40623	(± 180 °)		Power Sensor, Verification Kit/ SICT-CP-40623	
Phase		300 kHz ~ 45 MHz 45 MHz ~ 2.0 GHz (2.0 ~ 3.0) GHz (3.0 ~ 4.5) GHz (4.5 ~ 6.0) GHz (6.0 ~ 7.5) GHz (7.5 ~ 8.0) GHz (8.0 ~ 9.0) GHz (9.0 ~ 10.5) GHz (10.5 ~ 12.0) GHz (12.0 ~ 13.5) GHz (13.5 ~ 15.0) GHz (15.0 ~ 16.5) GHz (16.5 ~ 18.0) GHz (18.0 ~ 21.0) GHz (21.0 ~ 22.5) GHz (22.5 ~ 24.0) GHz (24.0 ~ 25.5) GHz (25.5 ~ 26.5) GHz	0.04° 0.09° 0.10° 0.11° 0.15° 0.19° 0.21° 0.22° 0.29° 0.28° 0.26° 0.27° 0.26° 0.29° 0.31° 0.28° 0.39° 0.33° 0.44°		
	SWR	(30 kHz ~ 2 GHz) 1.05 1.20 1.50 2.00 (2 ~ 18) GHz 1.05 1.20 1.50 2.00	0,021 0,021 0,021 0,021 0,018 0,018 0,018 0,024		
Noise figure meters	40624	Tuning Accuracy	10 MHz ~ 26.5 GHz	4.3×10^{-5}	Noise Source/ SICT-CP-40624
Referency frequency		10 MHz	6.1×10^{-10}		
Input VSWR		9 kHz ~ 1 GHz (1 ~ 20) GHz (20 ~ 26.5) GHz	0,008 0,019 0,03		
DC voltage		(0 ~ 28) V	0.000 18 V		
Range		(0 ~ 30) dB	0.052 dB		
Noise Figure		10 MHz ~ 8 GHz (8 ~ 18) GHz (18 ~ 26.5) GHz	0.12 dB 0.16 dB 0.37 dB		
Noise generators		40625	Noise Power	(-80 ~ -130) dBm/Hz	
Scale Fidelity	(0 ~ 50) dB		0.27 dB		

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Noise impulse simulators Peak Voltage Rise/Fall Time Pulse Width	40626	(±) 0.1 kV (0.1 ~ 5) kV 1 ns (1 ~ 2) ns (2 ~ 4) ns 10 ns (10 ~ 1 000) ns	4.0×10^{-2} 3.5×10^{-2} 6.0×10^{-3} 3.1×10^{-3} 1.5×10^{-3} 2.0×10^{-3} 1.5×10^{-3}	Digital Oscilloscope/ SICT-CP-40626
RF phase noise meters RF phase noise	40627	(Carrier Frequency) 100 MHz ~ 18 GHz (Offset Frequency) 10 Hz ~ 100 MHz	1.0 dB 1.0 dB	RF Signal analyzer/ SICT-CP-40627
Coaxial noise sources ENR SWR	40628	(4.5 ~ 16) dB (0.01 ~ 1) GHz (1 ~ 7) GHz (7 ~ 8) GHz (8 ~ 14) GHz (14 ~ 18) GHz (12 ~ 17) dB (0.01 ~ 1) GHz (1 ~ 2) GHz (2 ~ 6) GHz (6 ~ 7) GHz (7 ~ 12) GHz (12 ~ 18) GHz (18 ~ 26.5) GHz (0 ~ 1) (0.01 ~ 3) GHz (3 ~ 20) GHz (20 ~ 26.5) GHz	0.28 dB 0.27 dB 0.30 dB 0.31 dB 0.32 dB 0.31 dB 0.28 dB 0.30 dB 0.29 dB 0.40 dB 0.41 dB 0.47 dB 0.006 8 0,010 0,015	Coaxial noise sources, Noise figure analyzer/ SICT-CP-40628
RF power meters High power Zero Carryover Power Calibration Factor Power Ref. Output	40635	(0.1 ~ 500) W 10 kHz ~ 250 MHz (0.1 ~ 150) W (80 ~ 1 000) MHz (0.1 ~ 10) W (1 000 ~ 4 200) MHz 10 μW ~ 1 mW (1 ~ 100) mW 3 μW ~ 100 mW (88 ~ 100) % 50 MHz, 1 mW	2.6×10^{-2} 2.6×10^{-2} 2.7×10^{-2} 3 nW 0.01 mW 1.6×10^{-3} 0.5×10^{-3} 8 μW	Range Calibrator/ SICT-CP-40635

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Pulse generators	40638			Digital Oscilloscope/ SICT-CP-40638
Period		100 ps ~ 1 s	5.8×10^{-9}	
Frequency		1 Hz ~ 1 GHz (1 ~ 3.35) GHz	5.8×10^{-9} 1.7×10^{-6}	
Width		100 ps ~ 1 s	1.2×10^{-3}	
Delay Time		100 ps ~ 1 s	1.2×10^{-3}	
Double Pulse		100 ps ~ 1 s	1.2×10^{-3}	
Duty Cycle		(1 ~ 99) %	0.006 2 %	
DC Level		\pm (10 mV ~ 100 V)	5.8×10^{-4}	
Output Level		(100 Hz ~ 10 kHz) (10 ~ -20) dBm	0.018 dB	
Radar test sets	40639			Signal Analyzer, SART Generator, EPIRB Generator/ SICT-CP-40639
(Ship) RF Level		(20 ~ -20) dBm 20 Hz ~ 1 GHz (1 ~ 18) GHz	0.09 dB 0.13 dB	
		(-20 ~ -60) dBm 9 kHz ~ 1 GHz (1 ~ 18) GHz	0.06 dB 0.11 dB	
		(-60 ~ -120) dBm 10 MHz ~ 18 GHz	0.25 dB	
Amplitude Modulation		(0.1 ~ 100) %	1.2×10^{-2}	
Frequency Modulation		(0.1 ~ 400) kHz	1.2×10^{-2}	
Phase Modulation		(0.1 ~ 400) rad	1.2×10^{-2}	
Distortion of Modulation		(0 ~ 2) %	1.2×10^{-3}	
Hamonics		(9 kHz ~ 18 GHz) (0 ~ -110) dB	0.25 dB	
Frequency		9 kHz ~ 18 GHz	6.2×10^{-11}	
Pulse Period		1 ns ~ 10 ms	1.2×10^{-2}	
High power		(0.1 ~ 500) W 10 kHz ~ 250 MHz	2.6×10^{-2}	
		(0.1 ~ 150) W (80 ~ 1 000) MHz	2.6×10^{-2}	
		(0.1 ~ 10) W (1 000 ~ 4 200) MHz	2.7×10^{-2}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Radar test sets	40639			Signal Analyzer, SART Generator, EPIRB Generator/ SICT-CP-40639
(flight) Frequency(VOR/ILS/DME)		(74.6 ~ 1 150) MHz	8.2×10^{-8}	
Amplitude Modulation(VOR/ILS)		Localizer (108.1 ~ 111.95) MHz (0.1 ~ 20) %	0.62 %	
		Glideslope (330.95 ~ 334.70) MHz (20 ~ 40) %	0.84 %	
		Marker Beacon (74.6 ~ 75.4) MHz (40 ~ 95) %	1.4 %	
		VOR (108 ~ 117.95) MHz (0.1 ~ 30) %	0.62 %	
고주파 레벨(VOR/ILS)		Localizer (108.1 ~ 111.95) MHz (10 ~ -30) dBm (-30 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -110) dBm (-110 ~ -120) dBm	0.19 dB 0.23 dB 0.24 dB 0.34 dB 0.35 dB	
		Glideslope (330.95 ~ 334.70) MHz (10 ~ -30) dBm (-30 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -110) dBm (-110 ~ -120) dBm	0.19 dB 0.23 dB 0.24 dB 0.34 dB 0.35 dB	
DDM(VOR/ILS)		Localizer (108.1 ~ 111.95) MHz LEFT(-0.200 ~ -0.155) LEFT(-0.155 ~ -0.093) CENTER(0.000) RIGHT(0.093 ~ 0.155) RIGHT(0.155 ~ 0.200)	0.000 5 0.000 5 0.000 5 0.000 5 0.000 5	
		Glideslope (330.95 ~ 334.70) MHz DOWN(0.400 ~ 0.175) DOWN(0.175 ~ 0.091) CENTER(0.000) UP(-0.091 ~ -0.175) UP(-0.175 ~ -0.400)	0.000 5 0.000 5 0.000 5 0.000 5 0.000 5	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Radar test sets SDM(VOR/ILS) BEARING ANGLE(VOR)	40639	Localizer (108.1 ~ 111.95) MHz (0 ~ 40) % Glideslope (330.95 ~ 334.70) MHz (40 ~ 80) % VOR (108 ~ 117.95) MHz (0 ~ 360) °	0.87 % 1.2 % 0.04 °	Signal Analyzer, SART Generator, EPIRB Generator/ SICT-CP-40639
RF signal generators Frequency Modulation Amplitude Modulation Phase Modulation Pulse Modulation Distortion RF Level	40640	(0.1 ~ 400) kHz (0.1 ~ 100) % (0.1 ~ 400) rad (100 kHz ~ 12 000 MHz) Period (1 μs ~ 1 s) ton (100 ns ~ 100 μs) PRR ≤ 20 % (54 ~ 57) dBm 100 kHz ~ 500 MHz (51 ~ 54) dBm 100 kHz ~ 2 GHz (35 ~ 51) dBm 100 kHz ~ 18 GHz (20 ~ 35) dBm 100 kHz ~ 10 GHz (10 ~ 18) GHz (-30 ~ 20) dBm 20 Hz ~ 10 GHz (10 ~ 18) GHz (18 ~ 28) GHz (28 ~ 40) GHz (40 ~ 50) GHz (50 ~ 70) GHz (70 ~ 110) GHz (-30 ~ -60) dBm 20 Hz ~ 10 GHz (10 ~ 18) GHz (18 ~ 28) GHz (28 ~ 40) GHz (40 ~ 50) GHz	1.2 × 10 ⁻² 1.2 × 10 ⁻² 1.2 × 10 ⁻² 1.2 × 10 ⁻³ 1.2 × 10 ⁻³ 3.1 × 10 ⁻³ 2.3 × 10 ⁻² 0.35 dB 0.32 dB 0.32 dB 0.26 dB 0.27 dB 0.09 dB 0.11 dB 0.12 dB 0.16 dB 0.20 dB 0.29 dB 0.38 dB 0.10 dB 0.11 dB 0.12 dB 0.16 dB 0.21 dB	Measuring Receiver/ SICT-CP-40640

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF signal generators	40640	(-120 ~ -60) dBm	0.21 dB	Measuring Receiver/ SICT-CP-40640
RF Level		20 Hz ~ 4.2 GHz	0.23 dB	
		(4.2 ~ 8) GHz	0.24 dB	
		(8 ~ 12.4) GHz	0.27 dB	
		(12.4 ~ 18) GHz	0.31 dB	
		(18 ~ 26.5) GHz	0.37 dB	
Harmonic		(-10 ~ -110) dBc	0.37 dB	
Frequency		9 kHz ~ 40 GHz	2.1×10^{-11}	
RF spectrum analyzers	40641	(3 ~ 100) Hz	2.0×10^{-4}	Power Sensor, Synthesized Sweeper/ SICT-CP-40641
Center Frequency		(100 ~ 500) Hz	6.1×10^{-6}	
		(500 ~ 900) Hz	1.2×10^{-6}	
		900 Hz ~ 100 kHz	6.8×10^{-7}	
		0.1 MHz ~ 40 GHz	6.2×10^{-9}	
Frequency Counter		(3 ~ 100) Hz	2.0×10^{-4}	
		(100 ~ 500) Hz	6.1×10^{-6}	
		(500 ~ 900) Hz	1.2×10^{-6}	
		900 Hz ~ 100 kHz	6.8×10^{-7}	
		0.1 MHz ~ 40 GHz	6.2×10^{-9}	
Span		10 Hz ~ 100 kHz	7.6×10^{-3}	
		0.1 MHz ~ 40 GHz	7.7×10^{-6}	
RBW		1 Hz ~ 100 MHz	6.2×10^{-6}	
RBW Selectivity		1 Hz ~ 100 MHz	3.2×10^{-2}	
RBW Switching		1 Hz ~ 100 MHz	0.022 dB	
Scale Switching		1 dB ~ 10 dB scale/div	0.022 dB	
Scale Fidelity		(0 ~ -30) dB	0.073 dB	
		(-30 ~ -40) dB	0.077 dB	
		(-40 ~ -50) dB	0.082 dB	
		(-50 ~ -60) dB	0.095 dB	
		(-60 ~ -80) dB	0.13 dB	
		(-80 ~ -100) dB	0.17 dB	
Frequency Response		(10 ~ 100) Hz	0.01 dB	
		100 Hz ~ 1 GHz	0.15 dB	
	(1 ~ 6) GHz	0.16 dB		
	(6 ~ 10) GHz	0.17 dB		
	(10 ~ 18) GHz	0.19 dB		
	(18 ~ 26.5) GHz	0.21 dB		
	(26.5 ~ 30) GHz	0.23 dB		
	(30 ~ 35) GHz	0.28 dB		
	(35 ~ 40) GHz	0.32 dB		
Average Noise Level	DC ~ 40 GHz	0.17 dB		
Sideband Noise Level	(-30 ~ 30) kHz	0.33 dB		
CAL Output Freq. & Int. Frequency	DC ~ 1 GHz	6.2×10^{-9}		
CAL Output Level	(-20 ~ 20) dBm	0.09 dB		

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Surge generators	40643			Digital Oscilloscope/ SICT-CP-40643
Surge Voltage		(±) 2 V (2 ~ 5) V 5 V ~ 200 kV	7.0×10^{-2} 4.0×10^{-2} 3.5×10^{-2}	
Surge Current		(±) 5 A ~ 200 kA	3.6×10^{-2}	
Rise/Fall Time		1 ns (1 ~ 2) ns 2 ns ~ 10 s	6.0×10^{-3} 3.0×10^{-3} 2.0×10^{-3}	
Pulse Width		1 ns (1 ~ 2) ns 2 ns ~ 10 s	6.0×10^{-3} 3.0×10^{-3} 2.0×10^{-3}	
Time measurement by section		1 ns (1 ~ 2) ns 2 ns ~ 10 s	6.0×10^{-3} 3.0×10^{-3} 2.0×10^{-3}	
Frequency measurement by section		1 Hz ~ 25 MHz	1.6×10^{-3}	
Phase Shifting		at 50 Hz (0 ~ 360)° at 60 Hz (0 ~ 360)°	1.2° 1.4°	
SWR meters	40644			Coaxial Mismatch/ SICT-CP-40644
Frequency		9 kHz ~ 18 GHz	6.4×10^{-5}	
Output Level		30 kHz ~ 100 MHz 100 MHz ~ 10 GHz (10 ~ 18) GHz	0.06 dB 0.08 dB 0.09 dB	
SWR		(30 kHz ~ 30 MHz) 1.05 1.20 1.50 2.00 (30 MHz ~ 2 GHz) 1.05 1.20 1.50 2.00 (2 ~ 18) GHz 1.05 1.20 1.50 2.00	0,019 0,019 0,019 0,020 0,021 0,021 0,021 0,021 0,018 0,018 0,018 0,024	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF terminations (Open, Short, Phase)	40645	($\pm 180^\circ$) 10 Hz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.49° 0.61° 0.95° 1.2°	Network Analyzer, Coaxial Mismatch/ SICT-CP-40645
(Reflection coefficient)		(0 ~ 1) 10 Hz ~ 100 kHz 100 kHz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.004 3 0.006 0 0.009 5 0.016 0.019	
(SWR)		(1 ~ ∞) 10 Hz ~ 100 kHz 100 kHz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.008 6 0.012 0.019 0.032 0.038	
(Impedance)		(0.000 0 ~ 0.047 6) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.64 Ω 1.0 Ω 1.6 Ω 2.0 Ω	
		(0.047 6 ~ 0.090 9) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.71 Ω 1.1 Ω 1.8 Ω 2.2 Ω	
		(0.090 9 ~ 0.166 7) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.84 Ω 1.3 Ω 2.1 Ω 2.7 Ω	
		(0.166 7 ~ 0.230 8) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.99 Ω 1.6 Ω 2.5 Ω 3.1 Ω	
		(0.230 8 ~ 0.285 7) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	1.1 Ω 1.8 Ω 2.9 Ω 3.6 Ω	
		(0.285 7 ~ 0.333 4) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	1.3 Ω 2.1 Ω 3.3 Ω 4.2 Ω	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF terminations (Impedance Phase)	40645	(0.000 0 ~ 0.047 6, ±180 °) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz (0.047 6 ~ 0.090 9, ±180 °) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz (0.090 9 ~ 0.166 7, ±180 °) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz (0.166 7 ~ 0.230 8, ±180 °) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz (0.230 8 ~ 0.285 7, ±180 °) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz (0.285 7 ~ 0.333 4, ±180 °) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.68° 1.1° 1.7° 2.2° 0.67° 1.1° 1.7° 2.1° 0.66° 1.1° 1.7° 2.1° 0.65° 1.0° 1.6° 2.1° 0.64° 1.0° 1.6° 2.0° 0.62° 1.0° 1.6° 2.0°	Network Analyzer, Coaxial Mismatch/ SICT-CP-40645
Coaxial thermistor mounts Cal Factor	40646	(1 ~ 10) μW (9 ~ 100) kHz 100 kHz ~ 1 GHz (1 ~ 10) GHz (10 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz (10 μW ~ 10 mW) (9 ~ 100) kHz 100 kHz ~ 1 GHz (1 ~ 10) GHz (10 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz	0.4 × 10 ⁻² 0.9 × 10 ⁻² 1.1 × 10 ⁻² 1.9 × 10 ⁻² 2.8 × 10 ⁻² 4.0 × 10 ⁻² 0.4 × 10 ⁻² 0.9 × 10 ⁻² 1.1 × 10 ⁻² 1.9 × 10 ⁻² 2.6 × 10 ⁻² 3.5 × 10 ⁻²	Therimistor Mount, Synthesized Sweeper/ SICT-CP-40646

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Coaxial thermistor mounts Reflection coefficient SWR	40646	(0 ~ 1) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz (1 ~ ∞) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	 4.2×10^{-3} 9.4×10^{-3} 1.5×10^{-2} 9.7×10^{-3} 2.4×10^{-2} 3.8×10^{-2}	Synthesized Sweeper/ SICT-CP-40646
RF voltmeters RF Voltage	40650	3 V 1 V 300 mV 270 mV 240 mV 210 mV 180 mV 150 mV 120 mV 100 mV 90 mV 60 mV 30 mV 10 mV 3 mV 1 mV	4.2 mV 1.4 mV 0.36 mV 0.32 mV 0.28 mV 0.26 mV 0.24 mV 0.22 mV 0.17 mV 0.16 mV 0.13 mV 0.11 mV 0.048 mV 0.020 mV 0.018 mV 0.013 mV	RF Millivolt Meter Calibrator/ SICT-CP-40650
Vector voltmeters RF Voltage RF Phase	40651	3 V 1 V 300 mV 100 mV 30 mV 10 mV 3 mV 1 mV (0 ~ 270)°	4.2 mV 1.4 mV 0.36 mV 0.16 mV 0.048 mV 0.046 mV 0.022 mV 0.024 mV 0.006°	Signal Generator/ SICT-CP-40651
Field strength meters Center frequency Scale Fidelity Frequency response	40652	(9 ~ 100) kHz 0.1 MHz ~ 18 GHz (0 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -100) dB 9 kHz ~ 500 MHz 500 MHz ~ 18 GHz	 6.8×10^{-8} 6.2×10^{-9} 0.11 dB 0.12 dB 0.13 dB 0.18 dB 0.05 dB 0.08 dB	Signal Generator/ SICT-CP-40652
AM/FM test sources Output frequency	40653	(10 ~ 560) MHz	6.2×10^{-10}	Measuring Receiver/ SICT-CP-40653

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Dip simulators	40654			Digital Oscilloscope/ SICT-CP-40654
DC Voltage		1 V (1 ~ 5) V (5 ~ 10) V (10 ~ 50) V (50 ~ 100) V (100 ~ 500) V	1.1×10^{-5} 6.4×10^{-6} 5.4×10^{-6} 9.3×10^{-6} 8.0×10^{-6} 1.1×10^{-5}	
AC Voltage		(50 ~ 60) Hz 50 V (50 ~ 100) V (100 ~ 500) V	5.0×10^{-5} 2.5×10^{-5} 1.6×10^{-4}	
Frequency		50 Hz 60 Hz	8.4×10^{-6} 8.3×10^{-6}	
Dip DC Voltage		(0 ~ 50) V 0 % (0 ~ 120) %	0.2 V 3.4×10^{-2}	
Dip AC Voltage		(50 Hz ~ 60 Hz, 0 V ~ 400 V) 0 % (0 ~ 120) %	0.9 V 3.4×10^{-2}	
Time measurement by section		100 ns ~ 2 μs (2 ~ 4) μs (4 ~ 400) μs (0.4 ~ 2) ms 2 ms ~ 5 s	1.8×10^{-3} 2.0×10^{-3} 1.6×10^{-3} 2.0×10^{-3} 1.6×10^{-3}	
Inrush Current		(5 ~ 1 000) A	3.6×10^{-2}	
Phase Shifiting		at 50 Hz (0 ~ 360)° at 60 Hz (0 ~ 360)°	1.2° 1.4°	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Loop antennas Antenna Factor	40704	10 Hz ~ 30 MHz	1.2 dB	Signal generator1, Signal analyzer/ SICT-CP-40704
Monopole antennas Antenna Factor	40705	10 Hz ~ 30 MHz	1.4 dB	Signal generator1, Signal analyzer/ SICT-CP-40705

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistance thermometers: SPRT, IPRT, thermistors, etc.	50104	(-196 ~ 100) °C (100 ~ 300) °C (300 ~ 500) °C (500 ~ 660) °C (Fixed point) Ar TP Hg TP H ₂ O TP Ga MP In FP Sn FP Zn FP Al FP Ag FP	0.019 °C 0.020 °C 0.021 °C 0.044 °C 0.87 mK 1.1 mK 0.38 mK 1.1 mK 2.2 mK 2.2 mK 2.3 mK 2.7 mK 16 mK	SPRT, Fixed point/ SICT-CP-50104
Thermal expansion thermometers: bimetal, gas or liquid type bimetal	50105	(-196 ~ -70) °C (-70 ~ 100) °C (100 ~ 200) °C (200 ~ 500) °C (500 ~ 650) °C	0.6 °C 0.2 °C 0.3 °C 0.6 °C 1.2 °C	SPRT/ SICT-CP-50105
Thermocouples: noble metal, base metal, pure metal, special type, etc. Base metal Noble metal (Fixed point) H ₂ O ICE Point Sn FP Zn FP Al FP Ag FP Cu FP Co-C MP Fe MP	50106	(0 ~ 1 100) °C (1 100 ~ 1 300) °C (1 300 ~ 1 600) °C (-196 ~ -90) °C (-90 ~ 300) °C (300 ~ 500) °C (500 ~ 660) °C (660 ~ 900) °C (900 ~ 1 100) °C (1 100 ~ 1 300) °C 0.00 °C 231.928 °C 419.527 °C 660.323 °C 961.78 °C 1 084.62 °C 1 324 °C 1 534 °C	0.5 °C 1.7 °C 1.8 °C 0.4 °C 0.2 °C 0.3 °C 0.4 °C 1.1 °C 1.4 °C 1.8 °C 0.2 °C 0.2 °C 0.2 °C 0.2 °C 0.3 °C 0.3 °C 1.1 °C 1.6 °C	SPRT,Fixed point, STANDARD TC/ SICT-CP-50106

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Temperature transducers	50107	(-196 ~ 400) °C (400 ~ 500) °C (500 ~ 660) °C (660 ~ 800) °C (800 ~ 1 100) °C (1 100 ~ 1 300) °C (1 300 ~ 1 600) °C	0.031 °C 0.043 °C 0.072 °C 0.6 °C 0.7 °C 2.1 °C 2.2 °C	SPRT, THERMOCOUPLE, MULTIMETER SICT-CP-50107
Primary fixed-point cells and apparatus H ₂ O TP	50108	0.01 °C	0.24 mK	Triple-Point Cell SICT-CP-50108

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Optical pyrometers	50203	(900 ~ 1 800) °C	5 °C	Standard Lamp/ SICT-CP-50203
Standard radiation thermometers	50204	(-40 ~ -20) °C (-20 ~ 0) °C (0 ~ 10) °C (10 ~ 50) °C (50 ~ 100) °C (100 ~ 200) °C (200 ~ 300) °C (300 ~ 400) °C (400 ~ 500) °C (500 ~ 600) °C (600 ~ 700) °C (700 ~ 800) °C (800 ~ 900) °C (900 ~ 1 200) °C (1 200 ~ 1 400) °C (1 400 ~ 1 500) °C (1 500 ~ 1 600) °C (1 600 ~ 1 800) °C (1 800 ~ 2 000) °C (2 000 ~ 2 100) °C (2 100 ~ 2 200) °C (2 200 ~ 2 300) °C (2 300 ~ 2 400) °C	0.9 °C 0.7 °C 0.4 °C 0.3 °C 0.4 °C 0.5 °C 0.7 °C 0.8 °C 1.0 °C 1.1 °C 1.3 °C 1.6 °C 1.7 °C 1.8 °C 1.9 °C 2.0 °C 2.1 °C 2.5 °C 2.6 °C 4.2 °C 4.4 °C 4.6 °C 4.7 °C	Transfer Standard Pyrometer/ SICT-CP-50204
Thermal image apparatus	50205	(-40 ~ -20) °C (-20 ~ 0) °C (0 ~ 10) °C (10 ~ 50) °C (50 ~ 100) °C (100 ~ 200) °C (200 ~ 300) °C (300 ~ 400) °C (400 ~ 500) °C (500 ~ 600) °C (600 ~ 700) °C (700 ~ 800) °C (800 ~ 900) °C (900 ~ 1 200) °C (1 200 ~ 1 400) °C (1 400 ~ 1 500) °C (1 500 ~ 1 600) °C (1 600 ~ 1 800) °C (1 800 ~ 2 000) °C (2 000 ~ 2 100) °C (2 100 ~ 2 200) °C (2 200 ~ 2 300) °C (2 300 ~ 2 400) °C	0.9 °C 0.7 °C 0.4 °C 0.3 °C 0.4 °C 0.5 °C 0.7 °C 0.8 °C 1.0 °C 1.1 °C 1.3 °C 1.6 °C 1.7 °C 1.8 °C 1.9 °C 2.0 °C 2.1 °C 2.5 °C 2.6 °C 4.2 °C 4.4 °C 4.6 °C 4.7 °C	Transfer Standard Pyrometer/ SICT-CP-50205 SICT-CP-50205

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Blackbody furnaces	50206	(-40 ~ 0) °C (0 ~ 10) °C (10 ~ 50) °C (50 ~ 100) °C (100 ~ 200) °C (200 ~ 300) °C (300 ~ 400) °C (400 ~ 500) °C (500 ~ 600) °C (600 ~ 700) °C (700 ~ 1 100) °C (1 100 ~ 1 300) °C (1 300 ~ 1 500) °C (1 500 ~ 1 700) °C (1 700 ~ 1 800) °C (1 800 ~ 1 900) °C (1 900 ~ 2 000) °C (2 000 ~ 2 100) °C (2 100 ~ 2 200) °C (2 200 ~ 2 300) °C (2 300 ~ 2 400) °C	0.6 °C 0.4 °C 0.3 °C 0.4 °C 0.5 °C 0.7 °C 0.8 °C 0.9 °C 1.0 °C 1.1 °C 1.4 °C 1.5 °C 1.6 °C 1.8 °C 1.9 °C 2.0 °C 2.1 °C 3.7 °C 3.9 °C 4.1 °C 4.3 °C	Transfer Standard Pyrometer/ SICT-CP-50206
Others ; ear thermometers, etc.	50207	(30 ~ 45) °C	0.07 °C	Standard prt/ SICT-CP-50207

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Dew-point hygrometers; chilled mirror, alumina thinfilm, etc.	50301	(-90 ~ -80) °C D.P. (-80 ~ -70) °C D.P. (-70 ~ -50) °C D.P. (-50 ~ -20) °C D.P. (-20 ~ 90) °C D.P. (90 ~ 95) °C D.P.	0.60 °C D.P. 0.32 °C D.P. 0.20 °C D.P. 0.19 °C D.P. 0.13 °C D.P. 0.15 °C D.P.	Dewpoint Meter/ SICT-CP-50301
Relative humidity hygrometers; polimer thinfilm, hair, etc.	50302	humidity (3 ~ 60) % R.H. (60 ~ 90) % R.H. (90 ~ 98) % R.H. Temperature (-80 ~ 0) °C (0 ~ 80) °C (80 ~ 100) °C (100 ~ 180) °C	1.3 % R.H. 1.4 % R.H. 1.5 % R.H. 0.6 °C 0.3 °C 0.5 °C 1.5 °C	Dewpoint Meter/ SICT-CP-50302
Psychrometers: assmann ventilated, PRT type, etc.	50303	assmann ventilated (humidity) (10 ~ 50) % R.H. (50 ~ 70) % R.H. (70 ~ 90) % R.H. (90 ~ 95) % R.H. (Temperature) (0 ~ 50) °C PRT type (humidity) (10 ~ 50) % R.H. (50 ~ 80) % R.H. (80 ~ 98) % R.H. (Temperature) (0 ~ 80) °C (80 ~ 100) °C	1.3 % R.H. 1.4 % R.H. 1.5 % R.H. 1.6 % R.H. 0.3 °C 1.3 % R.H. 1.4 % R.H. 1.5 % R.H. 0.3 °C 0.5 °C	Dewpoint Meter/ SICT-CP-50303
Temperature humidity recorders; Hygrothermograph, etc	50304	Humidity (5 ~ 70) % R.H. (70 ~ 95) % R.H. Temperature (-20 ~ 80) °C	2.1 % R.H. 2.2 % R.H. 0.7 °C	Dewpoint Meter/ SICT-CP-50304

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Transducers; dew-point /relative humidity (Dew-point Transducers) Dew point	50305	(-90 ~ -80) °C D.P. (-80 ~ -70) °C D.P. (-70 ~ -60) °C D.P. (-60 ~ -40) °C D.P. (-40 ~ -20) °C D.P. (-20 ~ 0) °C D.P. (0 ~ 50) °C D.P. (50 ~ 90) °C D.P. (90 ~ 95) °C D.P.	0.60 °C D.P. 0.33 °C D.P. 0.22 °C D.P. 0.21 °C D.P. 0.20 °C D.P. 0.15 °C D.P. 0.14 °C D.P. 0.15 °C D.P. 0.17 °C D.P.	Dewpoint Meter/ SICT-CP-50305
(Relative humidity Transducers) Humidity		(3 ~ 50) % R.H. (50 ~ 80) % R.H. (80 ~ 98) % R.H.	1.3 % R.H. 1.4 % R.H. 1.5 % R.H.	
Temperature		(-80 ~ 0) °C (0 ~ 80) °C (80 ~ 100) °C (100 ~ 180) °C	0.7 °C 0.3 °C 0.5 °C 1.5 °C	
Humidity generators; two-pressure, two-temperature, flow mixing humidity generator, constant temperature and humidity chamber, etc. Dew point	50306	(-90 ~ -80) °C D.P. (-80 ~ -70) °C D.P. (-70 ~ -50) °C D.P. (-50 ~ -30) °C D.P. (-30 ~ -10) °C D.P. (-10 ~ 60) °C D.P. (60 ~ 80) °C D.P. (80 ~ 95) °C D.P.	0.60 °C D.P. 0.32 °C D.P. 0.19 °C D.P. 0.17 °C D.P. 0.16 °C D.P. 0.13 °C D.P. 0.14 °C D.P. 0.15 °C D.P.	Dewpoint Meter/ SICT-CP-50306
Humidity		(3 ~ 20) % R.H. (20 ~ 30) % R.H. (30 ~ 40) % R.H. (40 ~ 50) % R.H. (50 ~ 70) % R.H. (70 ~ 90) % R.H. (90 ~ 98) % R.H.	1.8 % R.H. 1.7 % R.H. 1.4 % R.H. 1.5 % R.H. 1.6 % R.H. 1.8 % R.H. 1.9 % R.H.	
Temperature		(-90 ~ 50) °C (50 ~ 100) °C (100 ~ 200) °C	0.4 °C 0.5 °C 0.6 °C	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Cereal moisture meters Moisture	50401	(9 ~ 20) % M.C.	0.7 % M.C.	Balance/ SICT-CP-50401
Wood moisture meters Moisture	50402	(8 ~ 25) % M.C.	2.5 % M.C.	Balance/ SICT-CP-50402
Paper moisture meters Moisture	50403	(8 ~ 20) % M.C.	3.4 % M.C.	Balance/ SICT-CP-50403

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Soundcalibrators	60102	(250 Hz) (94 ~ 114) dB (1 000 Hz) (94 ~ 114) dB	0.10 dB 0.10 dB	Referance microphone/ SICT-CP-60102
Microphones	60104	(250 Hz) (-60 ~ -20) dB	0.15 dB	Referance microphone/ SICT-CP-60104
Sound level meters	60106	(31.5 ~ 12 500) Hz 31.5 Hz 63 Hz 125 Hz 250 Hz 500 Hz 1 000 Hz 2 000 Hz 4 000 Hz 8 000 Hz 12 500 Hz	0.4 dB 0.3 dB 0.3 dB 0.2 dB 0.2 dB 0.2 dB 0.2 dB 0.2 dB 0.2 dB 0.4 dB 0.6 dB	Multifunction Acoustic Calibrator/ SICT-CP-60107

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Vibration calibrators	60301	10 Hz (10 ~ 2 500) Hz	1.6×10^{-2} 1.5×10^{-2}	Reference Accelerometer/ SICT-CP-60301
Vibration transducers	60302	0.5 Hz (0.5 ~ 20) Hz (20 ~ 1 250) Hz (1 250 ~ 2 500) Hz (2 500 ~ 5 000) Hz	1.5×10^{-2} 1.4×10^{-2} 1.1×10^{-2} 1.2×10^{-2} 2.4×10^{-2}	Reference Accelerometer/ SICT-CP-60302
Vibration measuring instruments Acceleration Velocity Displacement	60303	10 Hz 20 Hz (20 ~ 1 250) Hz (1 250 ~ 2 500) Hz 10 Hz 20 Hz (20 ~ 1 250) Hz (1 250 ~ 2 500) Hz (10 ~ 160) Hz (160 ~ 315) Hz (315 ~ 630) Hz	1.7×10^{-2} 1.6×10^{-2} 1.5×10^{-2} 1.6×10^{-2} 1.7×10^{-2} 1.6×10^{-2} 1.5×10^{-2} 1.6×10^{-2} 1.4×10^{-2} 2.1×10^{-2} 5.9×10^{-2}	Reference Accelerometer/ SICT-CP-60303

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Illuminance meters Illuminance	70101	(0.5 ~ 20 000) lx	1.7×10^{-2}	Illuminance Meters/ SICT-CP-70101
Luminance meters Luminance	70102	1 cd/m ² (1 ~ 10) cd/m ² (10 ~ 3 000) cd/m ² (3 000 ~ 15 000) cd/m ²	2.1×10^{-2} 1.7×10^{-2} 1.4×10^{-2} 1.6×10^{-2}	Luminance Standard Sources/ SICT-CP-70102
Total luminous flux meters Total luminous flux	70103	70 lm (70 ~ 4 650) lm	3.2×10^{-2} 1.5×10^{-2}	Total Luminous Flux Standard Lamps/ SICT-CP-70103
Luminance intensity meters Luminance	70104	(72 ~ 3 200) cd	3.7×10^{-2}	Luminous Intensity Standard Lamps, Illuminance Meters / SICT-CP-70104

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color temperature meters Color temperature Chromaticity	70202	(2 677 ~ 3 333) K x y	25 K 0,004 0,004	Color Temperature Standard Lamps/ SICT-CP-70202
Color temperature standard lamps Color temperature Chromaticity	70203	(2 677 ~ 3 333) K x y	27 K 0,005 0,005	Spectroradiometers/ SICT-CP-70203
Colorimeters; source color Luminance Chromaticity	70204	1 cd/m ² (1 ~ 10) cd/m ² (10 ~ 3 000) cd/m ² (3 000 ~ 15 000) cd/m ² (WHITE) x y (RED) x y (GREEN) x y (BLUE) x y (CIE Standard Illuminant A) x y	2.1 × 10 ⁻² 1.7 × 10 ⁻² 1.4 × 10 ⁻² 1.6 × 10 ⁻² 0.004 0.004 0.004 0.004 0.004 0.004 0.004 0.004 0.004 0.003 0.003	Luminance Standard Sources/ SICT-CP-70204
Laser power meters	70207	(405 nm) (0.75 ~ 9) mW (660 nm) (0.7 ~ 47) mW (785 nm) (0.7 ~ 46) mW (1 080 nm) (1 ~ 40) W	1.2 × 10 ⁻² 1.2 × 10 ⁻² 1.2 × 10 ⁻² 3.3 × 10 ⁻²	Optical Power Meters/ SICT-CP-70207
Standard LED light sources Total luminous flux	70208	(68.4 ~ 72.6) lm	3.8 × 10 ⁻²	Total Spectral Radiant Flux Meters/ SICT-CP-70208
Total luminous flux standard lamps Total luminous flux	70209	(320 ~ 10 000) lm	4.7 × 10 ⁻²	Total Luminous Flux Standard lamps/ SICT-CP-70209
Optical detectors Relative spectral responsivity	70210	(0 ~ 1) (300) nm (300 ~ 305) nm (305 ~ 310) nm (310 ~ 315) nm	8.8 × 10 ⁻² 8.0 × 10 ⁻² 7.2 × 10 ⁻² 6.5 × 10 ⁻²	Photodiodes/ SICT-CP-70210

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Optical detectors Relative spectral responsivity	70210	(315 ~ 320) nm (320 ~ 325) nm (325 ~ 330) nm (330 ~ 375) nm (375 ~ 435) nm (435 ~ 955) nm (955 ~ 965) nm (965 ~ 970) nm (970 ~ 975) nm (975 ~ 980) nm (980 ~ 985) nm (985 ~ 990) nm (990 ~ 995) nm (995 ~ 1 035) nm (1 035 ~ 1 055) nm (1 055 ~ 1 085) nm (1 090 ~ 1 100) nm	2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2} 2.9×10^{-2}	Photodiodes/ SICT-CP-70210
Pyranometers and pyrhemometers Irradiance responsivity	70211	(250 ~ 2 500) nm (1 000 ± 150)W/m ²	2.9×10^{-2}	Standard pyranometers/ SICT-CP-70211
Display color analyzers; luminance, chromaticity, white balance, etc. Luminance Chromaticity	70213	1 cd/m ² (1 ~ 5) cd/m ² (5 ~ 200) cd/m ² (WHITE) x y (RED) x y (GREEN) x y (BLUE) x y	3.8×10^{-2} 1.8×10^{-2} 1.7×10^{-2} 0.004 4 0.006 1 0.003 6 0.003 3 0.003 5 0.004 2 0.003 5 0.003 2	Luminance Meters/ SICT-CP-70213
Luminous intensity standard lamps Luminous intensity	70214	(10 ~ 20 000) cd	4.0×10^{-2}	Spectroradiometers/ SICT-CP-70214
Spectral irradiance standard lamps Spectral irradiance Illuminance	70215	250 nm (250 ~ 255) nm (255 ~ 300) nm (300 ~ 380) nm (380 ~ 450) nm (450 ~ 555) nm (555 ~ 660) nm (660 ~ 900) nm (900 ~ 1 600) nm (1 600 ~ 2 000) nm (2 000 ~ 2 300) nm (2 300 ~ 2 400) nm (800 ~ 7 200) lx	4.5×10^{-2} 4.4×10^{-2} 4.2×10^{-2} 3.5×10^{-2} 3.1×10^{-2} 2.9×10^{-2} 2.8×10^{-2} 2.7×10^{-2} 2.6×10^{-2} 2.9×10^{-2} 3.0×10^{-2} 3.4×10^{-2} 2.5×10^{-2}	Spectral Irradiance Standard Lamps/ SICT-CP-70215

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Total spectral radiant flux standard lamps Total spectral radiant flux	70216	350 nm (350 ~ 365) nm (365 ~ 380) nm (380 ~ 400) nm (400 ~ 455) nm (455 ~ 850) nm	6.7×10^{-2} 6.3×10^{-2} 5.8×10^{-2} 4.2×10^{-2} 3.9×10^{-2} 3.6×10^{-2}	Total Spectral Radiant Flux Standard Lamps/ SICT-CP-70216
Luminance standard lamps Luminance Chromaticity	70217	1 cd/m ² (1 ~ 10) cd/m ² (10 ~ 3 000) cd/m ² (3 000 ~ 15 000) cd/m ² (WHITE) x y (RED) x y (GREEN) x y (BLUE) x y (CIE Standard Illuminant A) x y	2.2×10^{-2} 1.8×10^{-2} 1.5×10^{-2} 1.8×10^{-2} 0.005 0.005 0.005 0.005 0.005 0.005 0.005 0.005 0.004 0.004	Luminance Standard Sources/ SICT-CP-70217
Spectral radiance standard lamps Spectral radiance	70218	300 nm (300 ~ 305) nm (305 ~ 310) nm (310 ~ 315) nm (315 ~ 320) nm (320 ~ 330) nm (330 ~ 340) nm (340 ~ 425) nm (425 ~ 470) nm (470 ~ 1 050) nm (1 050 ~ 1 600) nm	2.0×10^{-1} 1.7×10^{-1} 1.2×10^{-1} 8.9×10^{-2} 7.4×10^{-2} 4.8×10^{-2} 4.1×10^{-2} 3.5×10^{-2} 3.0×10^{-2} 2.8×10^{-2} 3.0×10^{-2}	Spectral Radiance Standard Sources/ SICT-CP-70218
UV irradiance meters Irradiance (UV Meter)	70219	(254 nm) 50 μW/cm ² ~ 3 mW/cm ² (365 nm) 10 μW/cm ² ~ 230 mW/cm ² (405 nm) 10 μW/cm ² ~ 230 mW/cm ²	3.7×10^{-2} 3.8×10^{-2} 3.8×10^{-2}	UV Meter Standard Detectors/ SICT-CP-70219

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Spectral irradiance meters Wavelength Spectral irradiance Illuminance	70220	(350 ~ 1 694) nm 250 nm (250 ~ 255) nm (255 ~ 300) nm (300 ~ 350) nm (350 ~ 380) nm (380 ~ 450) nm (450 ~ 550) nm (550 ~ 900) nm (900 ~ 1 615) nm (1 615 ~ 2 315) nm (2 315 ~ 2 365) nm (2 365 ~ 2 400) nm (800 ~ 7 200) lx	0.25 nm 4.5×10^{-2} 4.2×10^{-2} 4.1×10^{-2} 3.7×10^{-2} 3.5×10^{-2} 3.1×10^{-2} 2.9×10^{-2} 2.7×10^{-2} 2.6×10^{-2} 2.9×10^{-2} 3.4×10^{-2} 4.0×10^{-2} 1.9×10^{-2}	Spectral Irradiance Standard Lamps/ SICT-CP-70220
Total spectral radiant flux meters Wavelength Total spectral radiant flux	70221	(350 ~ 850) nm 350 nm (350 ~ 365) nm (365 ~ 375) nm (375 ~ 390) nm (390 ~ 445) nm (445 ~ 850) nm	0.25 nm 2.0×10^{-2} 1.8×10^{-2} 1.7×10^{-2} 1.6×10^{-2} 1.5×10^{-2} 1.4×10^{-2}	Total Spectral Radiant Flux Standard Lamps/ SICT-CP-70221
Spectral radiance meters Wavelength Spectral radiance	70222	(350 ~ 1 694) nm 300 nm (300 ~ 305) nm (305 ~ 310) nm (310 ~ 315) nm (315 ~ 320) nm (320 ~ 325) nm (325 ~ 335) nm (335 ~ 345) nm (345 ~ 405) nm (405 ~ 455) nm (455 ~ 755) nm (755 ~ 1 400) nm (1 400 ~ 1 525) nm (1 525 ~ 1 600) nm	0.25 nm 2.0×10^{-1} 1.7×10^{-1} 1.2×10^{-1} 8.8×10^{-2} 7.2×10^{-2} 5.5×10^{-2} 4.6×10^{-2} 3.7×10^{-2} 3.5×10^{-2} 3.0×10^{-2} 2.6×10^{-2} 2.7×10^{-2} 3.0×10^{-2} 2.8×10^{-2}	Spectral Radiance Standard Sources/ SICT-CP-70222

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source <small>T₁₀₀₀</small> A (2°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White	X		0.94	
	Y		0.85	
	Z		0.30	
2. L,Gray	X		0.64	
	Y		0.59	
	Z		0.22	
3. M,Gray	X		0.30	
	Y		0.27	
	Z		0.11	
4. D,Gray	X		0.11	
	Y		0.10	
	Z		0.04	
5. Red	X		0.36	
	Y		0.22	
	Z		0.06	
6. Yellow	X		0.78	
	Y		0.68	
	Z		0.08	
7. Green	X		0.19	
	Y		0.22	
	Z		0.08	
8. Cyan	X		0.17	
	Y		0.20	
	Z		0.16	
Included Reflectance Std. Light Source <small>T₁₀₀₀</small> A (2°)				
1. White	L*		0,36	
	a*		0,07	
	b*		0,07	
2. L,Gray	L*		0,32	
	a*		0,06	
	b*		0,06	
3. M,Gray	L*		0,25	
	a*		0,05	
	b*		0,05	
4. D,Gray	L*		0,17	
	a*		0,04	
	b*		0,04	
5. Red	L*		0,25	
	a*		0,23	
	b*		0,26	
6. Yellow	L*		0,34	
	a*		0,12	
	b*		0,37	
7. Green	L*		0,24	
	a*		0,11	
	b*		0,07	
8. Cyan	L*		0,23	
	a*		0,14	
	b*		0,14	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source <small>T_{vis}</small> A (2°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White	x		0,000 5	
	y		0,000 5	
2. L,Gray	x		0,000 5	
	y		0,000 5	
3. M,Gray	x		0,000 5	
	y		0,000 5	
4. D,Gray	x		0,000 5	
	y		0,000 5	
5. Red	x		0,000 8	
	y		0,000 5	
6. Yellow	x		0,000 5	
	y		0,000 5	
7. Green	x		0,000 5	
	y		0,000 5	
8. Cyan	x		0,000 5	
	y		0,000 5	
Included Reflectance Std. Light Source <small>T_{vis}</small> A (10°)				
1. White	X		0.95	
	Y		0.85	
	Z		0.30	
2. L,Gray	X		0.65	
	Y		0.59	
	Z		0.21	
3. M,Gray	X		0.30	
	Y		0.27	
	Z		0.11	
4. D,Gray	X		0.11	
	Y		0.10	
	Z		0.04	
5. Red	X		0.35	
	Y		0.22	
	Z		0.06	
6. Yellow	X		0.78	
	Y		0.66	
	Z		0.08	
7. Green	X		0.20	
	Y		0.22	
	Z		0.08	
8. Cyan	X		0.17	
	Y		0.20	
	Z		0.15	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source Type A (10°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White	L*		0,36	
	a*		0,06	
	b*		0,07	
2. L,Gray	L*		0,32	
	a*		0,06	
	b*		0,06	
3. M,Gray	L*		0,25	
	a*		0,05	
	b*		0,05	
4. D,Gray	L*		0,17	
	a*		0,04	
	b*		0,04	
5. Red	L*		0,25	
	a*		0,22	
	b*		0,26	
6. Yellow	L*		0,33	
	a*		0,12	
	b*		0,39	
7. Green	L*		0,24	
	a*		0,11	
	b*		0,07	
8. Cyan	L*		0,23	
	a*		0,13	
	b*		0.14	
Include Reflectance Std. Light Source Type A (10°)				
1. White	x		0,000 5	
	y		0,000 5	
2. I,Gray	x		0,000 5	
	y		0,000 5	
3. M,Gray	x		0,000 5	
	y		0,000 5	
4. D,Gray	x		0,000 5	
	y		0,000 5	
5. Red	x		0,000 7	
	y		0,000 5	
6. Yellow	x		0,000 5	
	y		0,000 5	
7. Green	x		0,000 5	
	y		0,000 5	
8. Cyan	x		0,000 5	
	y		0,000 5	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source T_{Vinc} C (2°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White	X		0.83	
	Y		0.85	
	Z		0.97	
2. L,Gray	X		0.58	
	Y		0.59	
	Z		0.69	
3. M,Gray	X		0.27	
	Y		0.28	
	Z		0.32	
4. D,Gray	X		0.10	
	Y		0.10	
	Z		0.11	
5. Red	X		0.25	
	Y		0.18	
	Z		0.15	
6. Yellow	X		0.56	
	Y		0.63	
	Z		0.22	
7. Green	X		0.18	
	Y		0.24	
	Z		0.21	
8. Cyan	X		0.20	
	Y		0.23	
	Z		0.49	
Included Reflectance Std. Light Source T_{Vinc} C (2°)				
1. White	L*		0,36	
	a*		0,09	
	b*		0,08	
2. L,Gray	L*		0,32	
	a*		0,08	
	b*		0,07	
3. M,Gray	L*		0,25	
	a*		0,06	
	b*		0,06	
4. D,Gray	L*		0,17	
	a*		0,04	
	b*		0,04	
5. Red	L*		0,24	
	a*		0,24	
	b*		0,21	
6. Yellow	L*		0,33	
	a*		0,16	
	b*		0,41	
7. Green	L*		0,24	
	a*		0,13	
	b*		0,08	
8. Cyan	L*		0,24	
	a*		0,13	
	b*		0,13	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source <small>T_{vis}</small> C (2°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White	x		0,000 5	
	y		0,000 5	
2. I, Gray	x		0,000 5	
	y		0,000 5	
3. M, Gray	x		0,000 5	
	y		0,000 5	
4. D, Gray	x		0,000 5	
	y		0,000 5	
5. Red	x		0,001 3	
	y		0,000 5	
6. Yellow	x		0,000 5	
	y		0,000 7	
7. Green	x		0,000 5	
	y		0,000 5	
8. Cyan	x		0,000 5	
	y		0,000 5	
Included Reflectance Std. Light Source <small>T_{vis}</small> C (10°)				
1. White	X		0.82	
	Y		0.85	
	Z		0.95	
2. L, Gray	X		0.57	
	Y		0.59	
	Z		0.68	
3. M, Gray	X		0.27	
	Y		0.28	
	Z		0.32	
4. D, Gray	X		0.10	
	Y		0.10	
	Z		0.11	
5. Red	X		0.24	
	Y		0.18	
	Z		0.14	
6. Yellow	X		0.58	
	Y		0.60	
	Z		0.20	
7. Green	X		0.18	
	Y		0.24	
	Z		0.19	
8. Cyan	X		0.20	
	Y		0.24	
	Z		0.48	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source <small>T_{VIS} 10°</small> C (10°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White	L*		0,36	
	a*		0,09	
	b*		0,07	
2. L,Gray	L*		0,32	
	a*		0,07	
	b*		0,07	
3. M,Gray	L*		0,25	
	a*		0,06	
	b*		0,05	
4. D,Gray	L*		0,17	
	a*		0,04	
	b*		0,04	
5. Red	L*		0,24	
	a*		0,23	
	b*		0,20	
6. Yellow	L*		0,32	
	a*		0,16	
	b*		0,42	
7. Green	L*		0,24	
	a*		0,12	
	b*		0,12	
8. Cyan	L*		0,24	
	a*		0,11	
	b*		0.15	
Included Reflectance Std. Light Source <small>T_{VIS} 10°</small> C (10°)				
1. White	x		0,000 5	
	y		0,000 5	
2. I,Gray	x		0,000 5	
	y		0,000 5	
3. M,Gray	x		0,000 5	
	y		0,000 5	
4. D,Gray	x		0,000 5	
	y		0,000 5	
5. Red	x		0,001 3	
	y		0,000 5	
6. Yellow	x		0,000 5	
	y		0,000 7	
7. Green	x		0,000 5	
	y		0,000 5	
8. Cyan	x		0,000 5	
	y		0,000 5	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source <small>T₁₀₀₀</small> D65 (2°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White	X		0.81	
	Y		0.85	
	Z		0.89	
2. L,Gray	X		0.56	
	Y		0.59	
	Z		0.64	
3. M,Gray	X		0.26	
	Y		0.28	
	Z		0.30	
4. D,Gray	X		0.09	
	Y		0.10	
	Z		0.10	
5. Red	X		0.25	
	Y		0.18	
	Z		0.14	
6. Yellow	X		0.58	
	Y		0.63	
	Z		0.20	
7. Green	X		0.17	
	Y		0.25	
	Z		0.20	
8. Cyan	X		0.19	
	Y		0.23	
	Z		0.46	
Included Reflectance Std. Light Source <small>T₁₀₀₀</small> D65 (2°)				
1. White	L*		0,36	
	a*		0,07	
	b*		0,07	
2. L,Gray	L*		0,32	
	a*		0,07	
	b*		0,06	
3. M,Gray	L*		0,25	
	a*		0,06	
	b*		0,05	
4. D,Gray	L*		0,17	
	a*		0,04	
	b*		0,04	
5. Red	L*		0,24	
	a*		0,25	
	b*		0,21	
6. Yellow	L*		0,33	
	a*		0,16	
	b*		0,40	
7. Green	L*		0,24	
	a*		0,13	
	b*		0,08	
8. Cyan	L*		0,24	
	a*		0,14	
	b*		0,13	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source <small>T₁₀₀₀</small> D65 (2°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White	x		0,000 5	
	y		0,000 5	
2. I, Gray	x		0,000 5	
	y		0,000 5	
3. M, Gray	x		0,000 5	
	y		0,000 5	
4. D, Gray	x		0,000 5	
	y		0,000 5	
5. Red	x		0,001 3	
	y		0,000 5	
6. Yellow	x		0,000 6	
	y		0,000 6	
7. Green	x		0,000 5	
	y		0,000 5	
8. Cyan	x		0,000 5	
	y		0,000 5	
Included Reflectance Std. Light Source <small>T₁₀₀₀</small> D65 (10°)				
1. White	X		0.80	
	Y		0.85	
	Z		0.88	
2. L, Gray	X		0.56	
	Y		0.59	
	Z		0.63	
3. M, Gray	X		0.26	
	Y		0.28	
	Z		0.29	
4. D, Gray	X		0.09	
	Y		0.10	
	Z		0.10	
5. Red	X		0.24	
	Y		0.18	
	Z		0.13	
6. Yellow	X		0.58	
	Y		0.60	
	Z		0.19	
7. Green	X		0.18	
	Y		0.25	
	Z		0.18	
8. Cyan	X		0.19	
	Y		0.24	
	Z		0.45	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source <small>T_{VIS}</small> D65 (10°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White	L*		0,36	
	a*		0,07	
	b*		0,07	
2. L,Gray	L*		0,32	
	a*		0,06	
	b*		0,06	
3. M,Gray	L*		0,25	
	a*		0,06	
	b*		0,05	
4. D,Gray	L*		0,17	
	a*		0,04	
	b*		0,04	
5. Red	L*		0,24	
	a*		0,24	
	b*		0,20	
6. Yellow	L*		0,32	
	a*		0,16	
	b*		0,41	
7. Green	L*		0,24	
	a*		0,12	
	b*		0,08	
8. Cyan	L*		0,24	
	a*		0,12	
	b*		0.12	
Included Reflectance Std. Light Source <small>T_{VIS}</small> D65 (10°)				
1. White	x		0,000 5	
	y		0,000 5	
2. I,Gray	x		0,000 5	
	y		0,000 5	
3. M,Gray	x		0,000 5	
	y		0,000 5	
4. D,Gray	x		0,000 5	
	y		0,000 5	
5. Red	x		0,001 3	
	y		0,000 5	
6. Yellow	x		0,000 6	
	y		0,000 6	
7. Green	x		0,000 5	
	y		0,000 5	
8. Cyan	x		0,000 5	
	y		0,000 5	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Exclude Reflectance Std. Light Source Type A (2°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White	X		0.90	
	Y		0.81	
	Z		0.29	
2. L,Gray	X		0.60	
	Y		0.55	
	Z		0.20	
3. M,Gray	X		0.26	
	Y		0.24	
	Z		0.10	
4. D,Gray	X		0.07	
	Y		0.06	
	Z		0.03	
5. Red	X		0.31	
	Y		0.20	
	Z		0.10	
6. Yellow	X		0.73	
	Y		0.64	
	Z		0.07	
7. Green	X		0.15	
	Y		0.19	
	Z		0.07	
8. Cyan	X		0.13	
	Y		0.16	
	Z		0.14	
Exclude Reflectance Std. Light Source Type A (2°)				
1. White	L*		0,35	
	a*		0,07	
	b*		0,07	
2. L,Gray	L*		0,31	
	a*		0,06	
	b*		0,06	
3. M,Gray	L*		0,24	
	a*		0,05	
	b*		0,05	
4. D,Gray	L*		0,14	
	a*		0,04	
	b*		0,04	
5. Red	L*		0,25	
	a*		0,30	
	b*		0,70	
6. Yellow	L*		0,33	
	a*		0,12	
	b*		0,45	
7. Green	L*		0,22	
	a*		0,12	
	b*		0,08	
8. Cyan	L*		0,22	
	a*		0,16	
	b*		0,15	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color	70301			Color Standard Tiles/ SICT-CP-70301
Exclude Reflectance Std. Light Source Type A (2°)				
1. White	x		0,000 5	
	y		0,000 5	
2. I,Gray	x		0,000 5	
	y		0,000 5	
3. M,Gray	x		0,000 5	
	y		0,000 5	
4. D,Gray	x		0,000 5	
	y		0,000 5	
5. Red	x		0,001 1	
	y		0,000 5	
6. Yellow	x		0,000 5	
	y		0,000 5	
7. Green	x		0,000 5	
	y		0,000 5	
8. Cyan	x		0,000 5	
	y		0,000 5	
Exclude Reflectance Std. Light Source Type A (10°)				
1. White	X		0.91	
	Y		0.81	
	Z		0.28	
2. L,Gray	X		0.61	
	Y		0.55	
	Z		0.20	
3. M,Gray	X		0.26	
	Y		0.24	
	Z		0.09	
4. D,Gray	X		0.07	
	Y		0.06	
	Z		0.03	
5. Red	X		0.30	
	Y		0.20	
	Z		0.10	
6. Yellow	X		0.74	
	Y		0.62	
	Z		0.07	
7. Green	X		0.15	
	Y		0.19	
	Z		0.06	
8. Cyan	X		0.13	
	Y		0.17	
	Z		0.14	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Excluded Reflectance Std. Light Source Type A (10°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White	L*		0,35	
	a*		0,06	
	b*		0,07	
2. L,Gray	L*		0,31	
	a*		0,05	
	b*		0,06	
3. M,Gray	L*		0,24	
	a*		0,05	
	b*		0,05	
4. D,Gray	L*		0,14	
	a*		0,04	
	b*		0,04	
5. Red	L*		0,25	
	a*		0,30	
	b*		0,70	
6. Yellow	L*		0,33	
	a*		0,12	
	b*		0,49	
7. Green	L*		0,22	
	a*		0,12	
	b*		0,08	
8. Cyan	L*		0,22	
	a*		0,14	
	b*		0.15	
Exclude Reflectance Std. Light Source Type A (10°)				
1. White	x		0,000 5	
	y		0,000 5	
2. I,Gray	x		0,000 5	
	y		0,000 5	
3. M,Gray	x		0,000 5	
	y		0,000 5	
4. D,Gray	x		0,000 5	
	y		0,000 5	
5. Red	x		0,001 1	
	y		0,000 5	
6. Yellow	x		0,000 5	
	y		0,000 5	
7. Green	x		0,000 5	
	y		0,000 5	
8. Cyan	x		0,000 5	
	y		0,000 5	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Excluded Reflectance Std. Light Source Type C (2°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White	X		0.79	
	Y		0.81	
	Z		0.92	
2. L,Gray	X		0.54	
	Y		0.55	
	Z		0.65	
3. M,Gray	X		0.23	
	Y		0.24	
	Z		0.28	
4. D,Gray	X		0.06	
	Y		0.06	
	Z		0.07	
5. Red	X		0.21	
	Y		0.20	
	Z		0.20	
6. Yellow	X		0.56	
	Y		0.60	
	Z		0.19	
7. Green	X		0.14	
	Y		0.20	
	Z		0.16	
8. Cyan	X		0.16	
	Y		0.19	
	Z		0.45	
Exclude Reflectance Std. Light Source Type C (2°)				
1. White	L*		0,35	
	a*		0,09	
	b*		0,07	
2. L,Gray	L*		0,31	
	a*		0,07	
	b*		0,06	
3. M,Gray	L*		0,24	
	a*		0,06	
	b*		0,05	
4. D,Gray	L*		0,14	
	a*		0,04	
	b*		0,04	
5. Red	L*		0,25	
	a*		0,40	
	b*		0,60	
6. Yellow	L*		0,32	
	a*		0,17	
	b*		0,52	
7. Green	L*		0,23	
	a*		0,14	
	b*		0,08	
8. Cyan	L*		0,23	
	a*		0,14	
	b*		0,14	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Excluded Reflectance Std. Light Source Type C (2°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White	x		0,000 5	
	y		0,000 5	
2. I,Gray	x		0,000 5	
	y		0,000 5	
3. M,Gray	x		0,000 5	
	y		0,000 5	
4. D,Gray	x		0,000 5	
	y		0,000 5	
5. Red	x		0,002 8	
	y		0,000 5	
6. Yellow	x		0,000 7	
	y		0,000 7	
7. Green	x		0,000 5	
	y		0,000 5	
8. Cyan	x		0,000 5	
	y		0,000 5	
Exclude Reflectance Std. Light Source Type C (10°)				
1. White	X		0.79	
	Y		0.81	
	Z		0.90	
2. L,Gray	X		0.54	
	Y		0.55	
	Z		0.63	
3. M,Gray	X		0.23	
	Y		0.24	
	Z		0.27	
4. D,Gray	X		0.06	
	Y		0.06	
	Z		0.07	
5. Red	X		0.20	
	Y		0.20	
	Z		0.20	
6. Yellow	X		0.55	
	Y		0.56	
	Z		0.17	
7. Green	X		0.14	
	Y		0.21	
	Z		0.14	
8. Cyan	X		0.17	
	Y		0.20	
	Z		0.44	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Excluded Reflectance Std. Light Source Type C (10°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White	L*		0,35	
	a*		0,08	
	b*		0,07	
2. L,Gray	L*		0,31	
	a*		0,07	
	b*		0,06	
3. M,Gray	L*		0,24	
	a*		0,06	
	b*		0,05	
4. D,Gray	L*		0,14	
	a*		0,04	
	b*		0,04	
5. Red	L*		0,30	
	a*		0,40	
	b*		0,60	
6. Yellow	L*		0,32	
	a*		0,17	
	b*		0,55	
7. Green	L*		0,23	
	a*		0,13	
	b*		0,14	
8. Cyan	L*		0,23	
	a*		0,12	
	b*		0.16	
Exclude Reflectance Std. Light Source Type C (10°)				
1. White	x		0,000 5	
	y		0,000 5	
2. I,Gray	x		0,000 5	
	y		0,000 5	
3. M,Gray	x		0,000 5	
	y		0,000 5	
4. D,Gray	x		0,000 5	
	y		0,000 5	
5. Red	x		0,002 9	
	y		0,000 5	
6. Yellow	x		0,000 6	
	y		0,000 8	
7. Green	x		0,000 5	
	y		0,000 5	
8. Cyan	x		0,000 5	
	y		0,000 5	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Excluded Reflectance Std. Light Source Type D65 (2°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White	X		0.77	
	Y		0.81	
	Z		0.85	
2. L,Gray	X		0.52	
	Y		0.55	
	Z		0.60	
3. M,Gray	X		0.23	
	Y		0.24	
	Z		0.26	
4. D,Gray	X		0.06	
	Y		0.06	
	Z		0.06	
5. Red	X		0.20	
	Y		0.20	
	Z		0.20	
6. Yellow	X		0.55	
	Y		0.60	
	Z		0.18	
7. Green	X		0.14	
	Y		0.21	
	Z		0.15	
8. Cyan	X		0.16	
	Y		0.19	
	Z		0.41	
Exclude Reflectance Std. Light Source Type D65 (2°)				
1. White	L*		0,35	
	a*		0,07	
	b*		0,07	
2. L,Gray	L*		0,31	
	a*		0,07	
	b*		0,06	
3. M,Gray	L*		0,24	
	a*		0,06	
	b*		0,05	
4. D,Gray	L*		0,14	
	a*		0,04	
	b*		0,04	
5. Red	L*		0,30	
	a*		0,40	
	b*		0,60	
6. Yellow	L*		0,32	
	a*		0,17	
	b*		0,51	
7. Green	L*		0,23	
	a*		0,14	
	b*		0,08	
8. Cyan	L*		0,23	
	a*		0,15	
	b*		0,14	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Excluded Reflectance Std. Light Source Type D65 (2°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White	x		0,000 5	
	y		0,000 5	
2. I,Gray	x		0,000 5	
	y		0,000 5	
3. M,Gray	x		0,000 5	
	y		0,000 5	
4. D,Gray	x		0,000 5	
	y		0,000 5	
5. Red	x		0,002 7	
	y		0,000 5	
6. Yellow	x		0,000 6	
	y		0,000 7	
7. Green	x		0,000 5	
	y		0,000 5	
8. Cyan	x		0,000 5	
	y		0,000 5	
Exclude Reflectance Std. Light Source Type D65 (10°)				
1. White	X		0.77	
	Y		0.81	
	Z		0.84	
2. L,Gray	X		0.52	
	Y		0.55	
	Z		0.59	
3. M,Gray	X		0.23	
	Y		0.24	
	Z		0.25	
4. D,Gray	X		0.06	
	Y		0.06	
	Z		0.06	
5. Red	X		0.20	
	Y		0.20	
	Z		0.20	
6. Yellow	X		0.54	
	Y		0.57	
	Z		0.17	
7. Green	X		0.14	
	Y		0.21	
	Z		0.14	
8. Cyan	X		0.16	
	Y		0.20	
	Z		0.41	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Excluded Reflectance Std. Light Source Type D65 (10°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		L*	0,35	
		a*	0,07	
		b*	0,07	
2. L,Gray		L*	0,31	
		a*	0,06	
		b*	0,06	
3. M,Gray		L*	0,24	
		a*	0,05	
		b*	0,05	
4. D,Gray		L*	0,14	
		a*	0,04	
		b*	0,04	
5. Red		L*	0,30	
		a*	0,40	
		b*	0,60	
6. Yellow		L*	0,32	
		a*	0,17	
		b*	0,55	
7. Green		L*	0,23	
		a*	0,13	
		b*	0,09	
8. Cyan		L*	0,23	
		a*	0,12	
		b*	0.13	
Exclude Reflectance Std. Light Source Type D65 (10°)				
1. White		x	0,000 5	
		y	0,000 5	
2. I,Gray		x	0,000 5	
		y	0,000 5	
3. M,Gray		x	0,000 5	
		y	0,000 5	
4. D,Gray		x	0,000 5	
		y	0,000 5	
5. Red		x	0,002 8	
		y	0,000 5	
6. Yellow		x	0,000 7	
		y	0,000 7	
7. Green		x	0,000 5	
		y	0,000 5	
8. Cyan		x	0,000 5	
		y	0,000 5	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Included Reflectance Std. Light Source <small>T₁₀₀₀</small> A (2°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White	X		0.94	
	Y		0.85	
	Z		0.30	
2. L,Gray	X		0.64	
	Y		0.59	
	Z		0.22	
3. M,Gray	X		0.30	
	Y		0.27	
	Z		0.11	
4. D,Gray	X		0.11	
	Y		0.10	
	Z		0.04	
5. Red	X		0.36	
	Y		0.22	
	Z		0.06	
6. Yellow	X		0.78	
	Y		0.68	
	Z		0.08	
7. Green	X		0.19	
	Y		0.22	
	Z		0.08	
8. Cyan	X		0.17	
	Y		0.20	
	Z		0.16	
Included Reflectance Std. Light Source <small>T₁₀₀₀</small> A (2°)				
1. White	L*		0,37	
	a*		0,07	
	b*		0,07	
2. L,Gray	L*		0,32	
	a*		0,06	
	b*		0,06	
3. M,Gray	L*		0,25	
	a*		0,05	
	b*		0,05	
4. D,Gray	L*		0,17	
	a*		0,04	
	b*		0,04	
5. Red	L*		0,25	
	a*		0,23	
	b*		0,26	
6. Yellow	L*		0,35	
	a*		0,12	
	b*		0,37	
7. Green	L*		0,24	
	a*		0,11	
	b*		0,07	
8. Cyan	L*		0,23	
	a*		0,14	
	b*		0,14	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Included Reflectance Std. Light Source <small>T_{vis}</small> A (2°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White	x		0.000 8	
	y		0.000 7	
2. L.Gray	x		0.000 8	
	y		0.000 7	
3. M.Gray	x		0.000 8	
	y		0.000 7	
4. D.Gray	x		0.000 8	
	y		0.000 7	
5. Red	x		0.001 0	
	y		0.000 7	
6. Yellow	x		0.000 8	
	y		0.000 8	
7. Green	x		0.000 7	
	y		0.000 8	
8. Cyan	x		0.000 7	
	y		0.000 7	
Included Reflectance Std. Light Source <small>T_{vis}</small> A (10°)				
1. White	X		0.95	
	Y		0.85	
	Z		0.30	
2. L.Gray	X		0.65	
	Y		0.59	
	Z		0.21	
3. M.Gray	X		0.30	
	Y		0.27	
	Z		0.11	
4. D.Gray	X		0.11	
	Y		0.10	
	Z		0.04	
5. Red	X		0.35	
	Y		0.22	
	Z		0.06	
6. Yellow	X		0.78	
	Y		0.66	
	Z		0.08	
7. Green	X		0.20	
	Y		0.22	
	Z		0.08	
8. Cyan	X		0.17	
	Y		0.20	
	Z		0.15	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Included Reflectance Std. Light Source Type A (10°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White	L*		0,37	
	a*		0,06	
	b*		0,07	
2. L,Gray	L*		0,32	
	a*		0,06	
	b*		0,06	
3. M,Gray	L*		0,25	
	a*		0,05	
	b*		0,05	
4. D,Gray	L*		0,17	
	a*		0,04	
	b*		0,04	
5. Red	L*		0,25	
	a*		0,22	
	b*		0,26	
6. Yellow	L*		0,34	
	a*		0,12	
	b*		0,39	
7. Green	L*		0,24	
	a*		0,11	
	b*		0,07	
8. Cyan	L*		0,23	
	a*		0,13	
	b*		0,14	
Include Reflectance Std. Light Source Type A (10°)				
1. White	x		0.000 8	
	y		0.000 7	
2. I,Gray	x		0.000 8	
	y		0.000 7	
3. M,Gray	x		0.000 8	
	y		0.000 7	
4. D,Gray	x		0.000 8	
	y		0.000 7	
5. Red	x		0.001 0	
	y		0.000 7	
6. Yellow	x		0.000 8	
	y		0.000 8	
7. Green	x		0.000 7	
	y		0.000 8	
8. Cyan	x		0.000 7	
	y		0.000 7	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Included Reflectance Std. Light Source T_{Vinc} C (2°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White	X		0.83	
	Y		0.85	
	Z		0.97	
2. L,Gray	X		0.58	
	Y		0.59	
	Z		0.69	
3. M,Gray	X		0.27	
	Y		0.28	
	Z		0.32	
4. D,Gray	X		0.10	
	Y		0.10	
	Z		0.11	
5. Red	X		0.25	
	Y		0.18	
	Z		0.15	
6. Yellow	X		0.57	
	Y		0.63	
	Z		0.22	
7. Green	X		0.18	
	Y		0.24	
	Z		0.21	
8. Cyan	X		0.20	
	Y		0.23	
	Z		0.49	
Included Reflectance Std. Light Source T_{Vinc} C (2°)				
1. White	L*		0,37	
	a*		0,09	
	b*		0,08	
2. L,Gray	L*		0,32	
	a*		0,08	
	b*		0,07	
3. M,Gray	L*		0,25	
	a*		0,06	
	b*		0,06	
4. D,Gray	L*		0,17	
	a*		0,04	
	b*		0,04	
5. Red	L*		0,24	
	a*		0,24	
	b*		0,21	
6. Yellow	L*		0,33	
	a*		0,16	
	b*		0,41	
7. Green	L*		0,24	
	a*		0,13	
	b*		0,08	
8. Cyan	L*		0,24	
	a*		0,13	
	b*		0,13	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Included Reflectance Std. Light Source <small>T_{vis}</small> C (2°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White	x		0.000 7	
	y		0.000 7	
2. I, Gray	x		0.000 7	
	y		0.000 7	
3. M, Gray	x		0.000 7	
	y		0.000 7	
4. D, Gray	x		0.000 7	
	y		0.000 7	
5. Red	x		0.001 4	
	y		0.000 7	
6. Yellow	x		0.000 8	
	y		0.000 9	
7. Green	x		0.000 7	
	y		0.000 7	
8. Cyan	x		0.000 7	
	y		0.000 7	
Included Reflectance Std. Light Source <small>T_{vis}</small> C (10°)				
1. White	X		0.82	
	Y		0.85	
	Z		0.95	
2. L, Gray	X		0.57	
	Y		0.59	
	Z		0.68	
3. M, Gray	X		0.27	
	Y		0.28	
	Z		0.32	
4. D, Gray	X		0.10	
	Y		0.10	
	Z		0.11	
5. Red	X		0.24	
	Y		0.18	
	Z		0.14	
6. Yellow	X		0.58	
	Y		0.60	
	Z		0.20	
7. Green	X		0.18	
	Y		0.24	
	Z		0.19	
8. Cyan	X		0.20	
	Y		0.24	
	Z		0.48	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Included Reflectance Std. Light Source <small>T_{VIS}</small> C (10°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White		L*	0,37	
		a*	0,09	
		b*	0,07	
2. L,Gray		L*	0,32	
		a*	0,07	
		b*	0,07	
3. M,Gray		L*	0,25	
		a*	0,06	
		b*	0,05	
4. D,Gray		L*	0,17	
		a*	0,04	
		b*	0,04	
5. Red		L*	0,24	
		a*	0,23	
		b*	0,20	
6. Yellow		L*	0,32	
		a*	0,16	
		b*	0,42	
7. Green		L*	0,24	
		a*	0,12	
		b*	0,12	
8. Cyan		L*	0,24	
		a*	0,11	
		b*	0.15	
Included Reflectance Std. Light Source <small>T_{VIS}</small> C (10°)				
1. White		x	0.000 7	
		y	0.000 7	
2. I,Gray		x	0.000 7	
		y	0.000 7	
3. M,Gray		x	0.000 7	
		y	0.000 7	
4. D,Gray		x	0.000 7	
		y	0.000 7	
5. Red		x	0.001 4	
		y	0.000 7	
6. Yellow		x	0.000 8	
		y	0.000 9	
7. Green		x	0.000 7	
		y	0.000 7	
8. Cyan		x	0.000 7	
		y	0.000 7	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Included Reflectance Std. Light Source <small>T₁₀₀₀</small> D65 (2°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White	X		0.81	
	Y		0.85	
	Z		0.89	
2. L,Gray	X		0.56	
	Y		0.59	
	Z		0.64	
3. M,Gray	X		0.26	
	Y		0.28	
	Z		0.30	
4. D,Gray	X		0.09	
	Y		0.10	
	Z		0.10	
5. Red	X		0.25	
	Y		0.18	
	Z		0.14	
6. Yellow	X		0.58	
	Y		0.63	
	Z		0.20	
7. Green	X		0.17	
	Y		0.25	
	Z		0.20	
8. Cyan	X		0.19	
	Y		0.23	
	Z		0.46	
Included Reflectance Std. Light Source <small>T₁₀₀₀</small> D65 (2°)				
1. White	L*		0,37	
	a*		0,07	
	b*		0,07	
2. L,Gray	L*		0,32	
	a*		0,07	
	b*		0,06	
3. M,Gray	L*		0,25	
	a*		0,06	
	b*		0,05	
4. D,Gray	L*		0,17	
	a*		0,04	
	b*		0,04	
5. Red	L*		0,24	
	a*		0,25	
	b*		0,21	
6. Yellow	L*		0,33	
	a*		0,16	
	b*		0,40	
7. Green	L*		0,24	
	a*		0,13	
	b*		0,08	
8. Cyan	L*		0,24	
	a*		0,14	
	b*		0,13	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Included Reflectance Std. Light Source <small>T_{View}</small> D65 (2°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White	x		0.000 7	
	y		0.000 7	
2. I, Gray	x		0.000 7	
	y		0.000 7	
3. M, Gray	x		0.000 7	
	y		0.000 7	
4. D, Gray	x		0.000 7	
	y		0.000 7	
5. Red	x		0.001 4	
	y		0.000 7	
6. Yellow	x		0.000 8	
	y		0.000 8	
7. Green	x		0.000 7	
	y		0.000 7	
8. Cyan	x		0.000 7	
	y		0.000 7	
Included Reflectance Std. Light Source <small>T_{View}</small> D65 (10°)				
1. White	X		0.80	
	Y		0.85	
	Z		0.88	
2. L, Gray	X		0.56	
	Y		0.59	
	Z		0.63	
3. M, Gray	X		0.26	
	Y		0.28	
	Z		0.29	
4. D, Gray	X		0.09	
	Y		0.10	
	Z		0.10	
5. Red	X		0.24	
	Y		0.18	
	Z		0.13	
6. Yellow	X		0.58	
	Y		0.60	
	Z		0.19	
7. Green	X		0.18	
	Y		0.25	
	Z		0.18	
8. Cyan	X		0.19	
	Y		0.24	
	Z		0.45	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Included Reflectance Std. Light Source <small>T_{VIS}</small> D65 (10°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White		L*	0,37	
		a*	0,07	
		b*	0,07	
2. L,Gray		L*	0,32	
		a*	0,06	
		b*	0,06	
3. M,Gray		L*	0,25	
		a*	0,06	
		b*	0,05	
4. D,Gray		L*	0,17	
		a*	0,04	
		b*	0,04	
5. Red		L*	0,24	
		a*	0,24	
		b*	0,20	
6. Yellow		L*	0,32	
		a*	0,16	
		b*	0,41	
7. Green		L*	0,24	
		a*	0,12	
		b*	0,08	
8. Cyan		L*	0,24	
		a*	0,12	
		b*	0.12	
Included Reflectance Std. Light Source <small>T_{VIS}</small> D65 (10°)				
1. White		x	0.000 7	
		y	0.000 7	
2. I,Gray		x	0.000 7	
		y	0.000 7	
3. M,Gray		x	0.000 7	
		y	0.000 7	
4. D,Gray		x	0.000 7	
		y	0.000 7	
5. Red		x	0.001 4	
		y	0.000 7	
6. Yellow		x	0.000 8	
		y	0.000 8	
7. Green		x	0.000 7	
		y	0.000 7	
8. Cyan		x	0.000 7	
		y	0.000 7	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles	70304			Color Standard Tiles/ SICT-CP-70304
Exclude Reflectance Std. Light Source Type A (2°)				
1. White	X		0.90	
	Y		0.81	
	Z		0.29	
2. L,Gray	X		0.60	
	Y		0.55	
	Z		0.20	
3. M,Gray	X		0.26	
	Y		0.24	
	Z		0.10	
4. D,Gray	X		0.07	
	Y		0.06	
	Z		0.03	
5. Red	X		0.31	
	Y		0.2	
	Z		0.1	
6. Yellow	X		0.73	
	Y		0.64	
	Z		0.07	
7. Green	X		0.15	
	Y		0.19	
	Z		0.07	
8. Cyan	X		0.13	
	Y		0.16	
	Z		0.14	
Exclude Reflectance Std. Light Source Type A (2°)				
1. White	L*		0,36	
	a*		0,07	
	b*		0,07	
2. L,Gray	L*		0,31	
	a*		0,06	
	b*		0,06	
3. M,Gray	L*		0,24	
	a*		0,05	
	b*		0,05	
4. D,Gray	L*		0,14	
	a*		0,04	
	b*		0,04	
5. Red	L*		0,25	
	a*		0,30	
	b*		0,70	
6. Yellow	L*		0,33	
	a*		0,12	
	b*		0,45	
7. Green	L*		0,22	
	a*		0,12	
	b*		0,08	
8. Cyan	L*		0,22	
	a*		0,16	
	b*		0,15	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles	70304			Color Standard Tiles/ SICT-CP-70304
Exclude Reflectance Std. Light Source Type A (2°)				
1. White	x		0.000 8	
	y		0.000 7	
2. I,Gray	x		0.000 8	
	y		0.000 7	
3. M,Gray	x		0.000 8	
	y		0.000 7	
4. D,Gray	x		0.000 8	
	y		0.000 7	
5. Red	x		0.001 3	
	y		0.000 7	
6. Yellow	x		0.000 8	
	y		0.000 8	
7. Green	x		0.000 7	
	y		0.000 8	
8. Cyan	x		0.000 7	
	y		0.000 7	
Exclude Reflectance Std. Light Source Type A (10°)				
1. White	X		0.91	
	Y		0.81	
	Z		0.28	
2. L,Gray	X		0.61	
	Y		0.55	
	Z		0.20	
3. M,Gray	X		0.26	
	Y		0.24	
	Z		0.09	
4. D,Gray	X		0.07	
	Y		0.06	
	Z		0.03	
5. Red	X		0.30	
	Y		0.20	
	Z		0.10	
6. Yellow	X		0.74	
	Y		0.62	
	Z		0.07	
7. Green	X		0.15	
	Y		0.19	
	Z		0.06	
8. Cyan	X		0.13	
	Y		0.17	
	Z		0.14	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Excluded Reflectance Std. Light Source Type A (10°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White	L*		0,36	
	a*		0,06	
	b*		0,07	
2. L,Gray	L*		0,31	
	a*		0,05	
	b*		0,06	
3. M,Gray	L*		0,24	
	a*		0,05	
	b*		0,05	
4. D,Gray	L*		0,14	
	a*		0,04	
	b*		0,04	
5. Red	L*		0,25	
	a*		0,30	
	b*		0,70	
6. Yellow	L*		0,33	
	a*		0,12	
	b*		0,49	
7. Green	L*		0,22	
	a*		0,12	
	b*		0,08	
8. Cyan	L*		0,22	
	a*		0,14	
	b*		0,15	
Exclude Reflectance Std. Light Source Type A (10°)				
1. White	x		0.000 6	
	y		0.000 6	
2. I,Gray	x		0.000 6	
	y		0.000 6	
3. M,Gray	x		0.000 6	
	y		0.000 6	
4. D,Gray	x		0.000 6	
	y		0.000 6	
5. Red	x		0.001 1	
	y		0.000 6	
6. Yellow	x		0.000 6	
	y		0.000 6	
7. Green	x		0.000 6	
	y		0.000 6	
8. Cyan	x		0.000 6	
	y		0.000 6	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Excluded Reflectance Std. Light Source Type C (2°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White	X		0.79	
	Y		0.81	
	Z		0.92	
2. L,Gray	X		0.54	
	Y		0.55	
	Z		0.65	
3. M,Gray	X		0.23	
	Y		0.24	
	Z		0.28	
4. D,Gray	X		0.06	
	Y		0.06	
	Z		0.07	
5. Red	X		0.21	
	Y		0.20	
	Z		0.20	
6. Yellow	X		0.56	
	Y		0.60	
	Z		0.19	
7. Green	X		0.14	
	Y		0.20	
	Z		0.16	
8. Cyan	X		0.16	
	Y		0.19	
	Z		0.45	
Exclude Reflectance Std. Light Source Type C (2°)				
1. White	L*		0,36	
	a*		0,09	
	b*		0,07	
2. L,Gray	L*		0,31	
	a*		0,07	
	b*		0,06	
3. M,Gray	L*		0,24	
	a*		0,06	
	b*		0,05	
4. D,Gray	L*		0,14	
	a*		0,04	
	b*		0,04	
5. Red	L*		0,25	
	a*		0,40	
	b*		0,60	
6. Yellow	L*		0,32	
	a*		0,17	
	b*		0,52	
7. Green	L*		0,23	
	a*		0,14	
	b*		0,08	
8. Cyan	L*		0,23	
	a*		0,14	
	b*		0,14	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Excluded Reflectance Std. Light Source Type C (2°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White	x		0,000 7	
	y		0,000 7	
2. I,Gray	x		0,000 7	
	y		0,000 7	
3. M,Gray	x		0,000 7	
	y		0,000 7	
4. D,Gray	x		0,000 7	
	y		0,000 7	
5. Red	x		0,002 8	
	y		0,000 7	
6. Yellow	x		0,000 9	
	y		0,000 9	
7. Green	x		0,000 7	
	y		0,000 7	
8. Cyan	x		0,000 7	
	y		0,000 7	
Excluded Reflectance Std. Light Source Type C (10°)				
1. White	X		0.79	
	Y		0.81	
	Z		0.90	
2. L,Gray	X		0.54	
	Y		0.55	
	Z		0.63	
3. M,Gray	X		0.23	
	Y		0.24	
	Z		0.27	
4. D,Gray	X		0.06	
	Y		0.06	
	Z		0.07	
5. Red	X		0.20	
	Y		0.20	
	Z		0.20	
6. Yellow	X		0.55	
	Y		0.56	
	Z		0.17	
7. Green	X		0.14	
	Y		0.21	
	Z		0.14	
8. Cyan	X		0.17	
	Y		0.20	
	Z		0.44	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Excluded Reflectance Std. Light Source Type C (10°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White	L*		0,36	
	a*		0,08	
	b*		0,07	
2. L,Gray	L*		0,31	
	a*		0,07	
	b*		0,06	
3. M,Gray	L*		0,24	
	a*		0,06	
	b*		0,05	
4. D,Gray	L*		0,14	
	a*		0,04	
	b*		0,04	
5. Red	L*		0,30	
	a*		0,40	
	b*		0,60	
6. Yellow	L*		0,32	
	a*		0,17	
	b*		0,55	
7. Green	L*		0,23	
	a*		0,13	
	b*		0,14	
8. Cyan	L*		0,23	
	a*		0,12	
	b*		0.16	
Exclude Reflectance Std. Light Source Type C (10°)				
1. White	x		0.000 7	
	y		0.000 7	
2. I,Gray	x		0.000 7	
	y		0.000 7	
3. M,Gray	x		0.000 7	
	y		0.000 7	
4. D,Gray	x		0.000 7	
	y		0.000 7	
5. Red	x		0.002 9	
	y		0.000 7	
6. Yellow	x		0.000 8	
	y		0.001 0	
7. Green	x		0.000 7	
	y		0.000 7	
8. Cyan	x		0.000 7	
	y		0.000 7	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Excluded Reflectance Std. Light Source D65 (2°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White	X		0.77	
	Y		0.81	
	Z		0.85	
2. L,Gray	X		0.52	
	Y		0.55	
	Z		0.60	
3. M,Gray	X		0.23	
	Y		0.24	
	Z		0.26	
4. D,Gray	X		0.06	
	Y		0.06	
	Z		0.06	
5. Red	X		0.20	
	Y		0.20	
	Z		0.20	
6. Yellow	X		0.55	
	Y		0.60	
	Z		0.18	
7. Green	X		0.14	
	Y		0.21	
	Z		0.15	
8. Cyan	X		0.16	
	Y		0.19	
	Z		0.41	
Exclude Reflectance Std. Light Source Type D65 (2°)				
1. White	L*		0,36	
	a*		0,07	
	b*		0,07	
2. L,Gray	L*		0,31	
	a*		0,07	
	b*		0,06	
3. M,Gray	L*		0,24	
	a*		0,06	
	b*		0,05	
4. D,Gray	L*		0,14	
	a*		0,04	
	b*		0,04	
5. Red	L*		0,30	
	a*		0,40	
	b*		0,60	
6. Yellow	L*		0,32	
	a*		0,17	
	b*		0,51	
7. Green	L*		0,23	
	a*		0,14	
	b*		0,08	
8. Cyan	L*		0,23	
	a*		0,15	
	b*		0,14	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Excluded Reflectance Std. Light Source Type D65 (2°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White	x		0.000 7	
	y		0.000 7	
2. I, Gray	x		0.000 7	
	y		0.000 7	
3. M, Gray	x		0.000 7	
	y		0.000 7	
4. D, Gray	x		0.000 7	
	y		0.000 7	
5. Red	x		0.002 7	
	y		0.000 7	
6. Yellow	x		0.000 8	
	y		0.000 9	
7. Green	x		0.000 7	
	y		0.000 8	
8. Cyan	x		0.000 7	
	y		0.000 7	
Exclude Reflectance Std. Light Source Type D65 (10°)				
1. White	X		0.77	
	Y		0.81	
	Z		0.84	
2. L, Gray	X		0.52	
	Y		0.55	
	Z		0.59	
3. M, Gray	X		0.23	
	Y		0.24	
	Z		0.25	
4. D, Gray	X		0.06	
	Y		0.06	
	Z		0.06	
5. Red	X		0.20	
	Y		0.20	
	Z		0.20	
6. Yellow	X		0.54	
	Y		0.57	
	Z		0.17	
7. Green	X		0.14	
	Y		0.21	
	Z		0.14	
8. Cyan	X		0.16	
	Y		0.20	
	Z		0.41	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Excluded Reflectance Std. Light Source Type D65 (10°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White	L*		0,36	
	a*		0,07	
	b*		0,07	
2. L,Gray	L*		0,31	
	a*		0,06	
	b*		0,06	
3. M,Gray	L*		0,24	
	a*		0,05	
	b*		0,05	
4. D,Gray	L*		0,14	
	a*		0,04	
	b*		0,04	
5. Red	L*		0,30	
	a*		0,40	
	b*		0,60	
6. Yellow	L*		0,32	
	a*		0,17	
	b*		0,55	
7. Green	L*		0,23	
	a*		0,13	
	b*		0,09	
8. Cyan	L*		0,23	
	a*		0,12	
	b*		0,13	
Exclude Reflectance Std. Light Source Type D65 (10°)				
1. White	x		0.000 7	
	y		0.000 7	
2. I,Gray	x		0.000 7	
	y		0.000 7	
3. M,Gray	x		0.000 7	
	y		0.000 7	
4. D,Gray	x		0.000 7	
	y		0.000 7	
5. Red	x		0.002 8	
	y		0.000 7	
6. Yellow	x		0.000 9	
	y		0.000 9	
7. Green	x		0.000 7	
	y		0.000 8	
8. Cyan	x		0.000 7	
	y		0.000 7	
Absolute Spectral Reflectance White Plate (Include, Exclude Reflectance)		(360 ~ 830) nm	0.007 6	Absolute Spectral Reflectance White Standard Plates/ SICT-CP-70304

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Gloss meters Gloss	70306	20 ° 60 ° 85 °	8.9×10^{-3} 9.5×10^{-3} 8.0×10^{-3}	Gloss Standard/ SICT-CP-70306
Gloss standard plates Gloss	70307	20° 60° 85°	9.5×10^{-3} 9.8×10^{-3} 8.3×10^{-3}	Gloss Meter/ SICT-CP-70307
Haze meters Haze Transmittance	70308	H-1 H-5 H-10 H-20 H-30 T-30 T-50 T-70 T-90	0.30 0.26 0.4 0.6 0.8 0.50 0.50 0.50 0.50	Haze Standard Plate, Transmittance Standard Plates/ SICT-CP-70308
Lens meters Vertex diopter	70312	-25 D ~ 25 D	0.03 D	Reference Lens/ SICT-CP-70312
Optical densitymeters Density	70315	1 Step ~ 10 Step 11 Step 11 Step ~ 15 Step	0.03 0.06 0.11	Density CRM/ SICT-CP-70315
Reflectance meters Reflectance	70319	380 nm ~ 780 nm	1.1×10^{-2}	Absolute Spectral Reflectance White Standard Plates/ SICT-CP-70319
Refractometers Refracto	70321	(1.332 99 ~ 1.505 80) nD 1.51 nD 1.62 nD	0.000 04 nD 0.000 2 nD 0.000 2 nD	Reference Refracto CRM/ SICT-CP-70321
Transmittance meters	70323	(0.1) (250 ~ 750) nm (0.5) (250 ~ 750) nm (0.9) (250 ~ 750) nm	6.1×10^{-3} 3.8×10^{-3} 2.2×10^{-3}	Transmittance Filter/ SICT-CP-70323

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Spectrophotometers including FT-IR spectrophotometers	70325			Wavelength Filters, Transmittance Filters, Absolute Spectral Reflectance White Standard Plates, Wavenumber Filters/ SICT-CP-70325	
		Wavelength	(240 ~ 750) nm		0.4 nm
		Transmittance	(0.1 ~ 0.3)		
			250 nm		8.1×10^{-3}
300 nm	7.8×10^{-3}				
350 nm	7.7×10^{-3}				
400 nm	5.4×10^{-3}				
450 nm	5.2×10^{-3}				
500 nm	5.2×10^{-3}				
550 nm	5.2×10^{-3}				
600 nm	5.2×10^{-3}				
650 nm	5.2×10^{-3}				
700 nm	5.2×10^{-3}				
750 nm	5.2×10^{-3}				
(0.3 ~ 0.5)					
250 nm	8.1×10^{-3}				
300 nm	7.8×10^{-3}				
350 nm	7.7×10^{-3}				
400 nm	5.2×10^{-3}				
450 nm	5.2×10^{-3}				
500 nm	5.2×10^{-3}				
550 nm	5.1×10^{-3}				
600 nm	5.2×10^{-3}				
650 nm	5.2×10^{-3}				
700 nm	5.2×10^{-3}				
750 nm	5.2×10^{-3}				
(0.5 ~ 0.9)					
250 nm	7.9×10^{-3}				
300 nm	8.2×10^{-3}				
350 nm	7.8×10^{-3}				
400 nm	5.2×10^{-3}				
450 nm	5.2×10^{-3}				
500 nm	5.2×10^{-3}				
550 nm	5.1×10^{-3}				
600 nm	5.1×10^{-3}				
650 nm	5.1×10^{-3}				
700 nm	5.2×10^{-3}				
750 nm	5.2×10^{-3}				
(0.01)					
440 nm	1.3×10^{-2}				
465 nm	8.4×10^{-3}				
546 nm	8.9×10^{-3}				
590 nm	1.0×10^{-2}				
635 nm	8.1×10^{-3}				

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.		
Spectrophotometers including FT-IR spectrophotometers	70325	Transmittance (0.03)	440 nm	7.6×10^{-3}	Wavelength Filters, Transmittance Filters, Absolute Spectral Reflectance White Standard Plates, Wavenumber Filters/ SICT-CP-70325	
			465 nm	5.7×10^{-3}		
			546 nm	5.8×10^{-3}		
			590 nm	6.3×10^{-3}		
			635 nm	5.9×10^{-3}		
			Absorbance (0.1 ~ 0.3)	250 nm		0.003 6
		300 nm		0.003 4		
		350 nm		0.003 5		
		400 nm		0.002 3		
		450 nm		0.002 3		
		500 nm		0.002 3		
		550 nm		0.002 3		
		600 nm		0.002 3		
		650 nm		0.002 3		
		700 nm		0.002 3		
		750 nm		0.002 3		
		(0.3 ~ 0.5)		250 nm		0.003 6
				300 nm		0.003 4
				350 nm		0.003 5
				400 nm		0.002 3
				450 nm		0.002 3
				500 nm		0.002 3
				550 nm		0.002 3
				600 nm		0.002 3
		(0.5 ~ 0.9)		250 nm		0.003 5
				300 nm		0.003 5
				350 nm		0.003 5
				400 nm		0.002 3
				450 nm		0.002 3
			500 nm	0.002 3		
			550 nm	0.002 3		
			600 nm	0.002 3		
			650 nm	0.002 3		
			700 nm	0.002 3		
			750 nm	0.002 3		

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Spectrophotometers including FT-IR spectrophotometers	70325			Wavelength Filters, Transmittance Filters, Absolute Spectral Reflectance White Standard Plates, Wavenumber Filters/ SICT-CP-70325
Absorbance		(0.01)		
		440 nm	0.003 8	
		465 nm	0.002 7	
		546 nm	0.002 9	
		590 nm	0.003 1	
		635 nm	0.002 7	
		(0.03)		
		440 nm	0.003 1	
		465 nm	0.002 4	
		546 nm	0.002 6	
		590 nm	0.002 6	
		635 nm	0.002 6	
Reflectance		(250 ~ 2 500) nm	1.4×10^{-2}	
Wavenumber		544.9 cm^{-1}	2.5 cm^{-1}	
		842.1 cm^{-1}	1.3 cm^{-1}	
		906.82 cm^{-1}	0.12 cm^{-1}	
		1 028.42 cm^{-1}	0.28 cm^{-1}	
		1 069.27 cm^{-1}	0.78 cm^{-1}	
		1 154.62 cm^{-1}	0.11 cm^{-1}	
		1 583.04 cm^{-1}	0.11 cm^{-1}	
		1 601.38 cm^{-1}	0.12 cm^{-1}	
		2 850.20 cm^{-1}	0.13 cm^{-1}	
		3 001.40 cm^{-1}	0.11 cm^{-1}	
		3 026.44 cm^{-1}	0.11 cm^{-1}	
		3 060.14 cm^{-1}	0.11 cm^{-1}	
		3 082.22 cm^{-1}	0.11 cm^{-1}	
Wavelength reference materials; absorption cell, bandpass filter, etc.	70326			Spectrophotometers, Absolute Spectral Reflectance White Standard Plates/ SICT-CP-70326
Wavelength		(240 ~750) nm	0.5 nm	
Transmittance		(0.1 ~ 0.3)		
		250 nm	8.5×10^{-3}	
		300 nm	8.1×10^{-3}	
		350 nm	8.1×10^{-3}	
		400 nm	5.9×10^{-3}	
		450 nm	5.7×10^{-3}	
		500 nm	5.7×10^{-3}	
		550 nm	5.7×10^{-3}	
		600 nm	5.7×10^{-3}	
		650 nm	5.7×10^{-3}	
		700 nm	5.7×10^{-3}	
		750 nm	5.7×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.		
Wavelength reference materials; absorption cell, bandpass filter, etc. Transmittance	70326	(0.3 ~ 0.5)		Spectrophotometers, Absolute Spectral Reflectance White Standard Plates/ SICT-CP-70326		
		250 nm	8.3×10^{-3}			
		300 nm	8.1×10^{-3}			
		350 nm	8.0×10^{-3}			
		400 nm	5.7×10^{-3}			
		450 nm	5.7×10^{-3}			
		500 nm	5.7×10^{-3}			
		550 nm	5.7×10^{-3}			
		600 nm	5.7×10^{-3}			
		650 nm	5.7×10^{-3}			
		700 nm	5.7×10^{-3}			
		750 nm	5.7×10^{-3}			
		(0.5 ~ 0.9)				
		250 nm	8.2×10^{-3}			
		300 nm	8.1×10^{-3}			
		350 nm	8.0×10^{-3}			
		400 nm	5.7×10^{-3}			
		450 nm	5.7×10^{-3}			
		500 nm	5.7×10^{-3}			
		550 nm	5.7×10^{-3}			
		600 nm	5.6×10^{-3}			
		650 nm	5.6×10^{-3}			
		700 nm	5.7×10^{-3}			
		750 nm	5.9×10^{-3}			
		Absorbance			(0.1 ~ 0.3)	
					250 nm	0.003 7
					300 nm	0.003 7
					350 nm	0.003 8
					400 nm	0.002 7
					450 nm	0.002 7
					500 nm	0.002 8
					550 nm	0.002 8
					600 nm	0.002 7
					650 nm	0.002 7
					700 nm	0.002 7
					750 nm	0.002 8
					(0.3 ~ 0.5)	
					250 nm	0.003 6
					300 nm	0.003 6
					350 nm	0.003 6
					400 nm	0.002 4
					450 nm	0.002 4
		500 nm	0.002 4			
		550 nm	0.002 4			
		600 nm	0.002 4			
		650 nm	0.002 4			
		700 nm	0.002 4			
		750 nm	0.002 5			

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Wavelength reference materials; absorption cell, bandpass filter, etc.	70326			Spectrophotometers, Absolute Spectral Reflectance White Standard Plates/ SICT-CP-70326
Absorbance		(0.5 ~ 0.9)		
		250 nm	0.003 6	
		300 nm	0.003 5	
		350 nm	0.003 5	
		400 nm	0.002 4	
		450 nm	0.002 4	
		500 nm	0.002 4	
		550 nm	0.002 4	
		600 nm	0.002 4	
		650 nm	0.002 4	
		700 nm	0.002 4	
		750 nm	0.002 4	
Reflectance		(360 ~ 830) nm	1.0×10^{-2}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Broadband light sources Wavelength output	70402	1 310 nm, 1 550 nm	0.058 nm	Optical spectrum analyzer, Optical powermeter/ SICT-CP-70402
Optical power output		1 310 nm, 1 550 nm (0 ~ -60) dBm	0.070 dB	
Optical attenuators Optical Attenuation	70410	1 310 nm, 1 550 nm (-60 ~ 0) dB	0.08 dB	Optical powermeter, Optical power stabilized lasers and LDs/ SICT-CP-70410
Fiber-optic power meters Absolute optical power	70412	1 310 nm, 1 550 nm (0 ~ -60) dBm	0.072 dB	Optical powermeter, Optical power stabilized lasers and LDs, Optical attenuator/ SICT-CP-70412
Optical Linearity		1 310 nm, 1 550 nm (0 ~ -60) dB	0.03 dB	
Optical loss testers Optical Attenuation	70413	1 310 nm, 1 550 nm (0 ~ -60) dB	0.03 dB	Optical attenuator/ SICT-CP-70413
Optical multimeters Absolute optical power measure	70415	1 310 nm, 1 550 nm (0 ~ -60) dBm	0.072 dB	Optical powermeter, Optical power stabilized lasers and LDs, Optical attenuator/ SICT-CP-70415
Linearity measure		1 310 nm, 1 550 nm (0 ~ -60) dB	0.03 dB	
Optical network analyzer (Optical multimeter) Absolute optical power	70416	1 310 nm, 1 550 nm (-60 ~ 0) dBm	0.072 dB	Optical powermeter, OTDR, Fiber reference, Wavelength meter Optical spectrum analyzer Optical attenuator Optical Returnloss generator/ SICT-CP-70416
(광)Optical spectrum analyzer Wavelength measure		1 310 nm 1 550 nm	0.058 nm 0.058 nm	
Resolution measure		Resolution: (0.1 ~ 1) nm 1 310 nm 1 550 nm	0.058 nm 0.058 nm	
Absolute optical power measure		1 310 nm, 1 550 nm (-60 ~ 0) dBm	0.072 dB	
(Optical attenuator) Optical Attenuation		1 310 nm, 1 550 nm (-60 ~ 0) dB	0.07 dB	
Return loss		1 310 nm, 1 550 nm (20 ~ 40) dB	0.8 dB	
(Optical time domain reflectometer) Wavelength output		1 310 nm, 1 550 nm	0.082 nm	
Optical Length measure		1 310 nm 3.3 km Fiber 13.4 km Fiber	0.081 m 0.34 m	
		1 550 nm 3.3 km Fiber 13.4 km Fiber	0.080 m 0.34 m	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Optical network analyzer Optical loss measure	70416	1 310 nm 7.20 dB Fiber 2.90 dB Fiber 1 550 nm 4.20 dB Fiber 1.60 dB Fiber	0.13 dB 0.05 dB 0.05 dB 0.05 dB	Optical powermeter, OTDR, Fiber reference, Wavelength meter Optical spectrum analyzer Optical attenuator Optical Returnloss generator/ SICT-CP-70416
Optical spectrum analyzers Wavelength measure Resolution measure Absolute optical power measure Linearity measure	70417	1 310 nm 1 550 nm 분해능: (0.1 ~ 1) nm 1 310 nm 1 550 nm 1 310 nm, 1 550 nm (-60 ~ 0) dBm 1 310 nm, 1 550 nm (-60 ~ 0) dB	0.058 nm 0.058 nm 0.058 nm 0.058 nm 0.072 dB 0.03 dB	Optical powermeter, Optical power stabilized lasers and LDs, Optical attenuator, Optical spectrum analyzer/ SICT-CP-70417
Optical time domain reflectometers, OTDR Wavelength output Optical Length measure Optical loss measure	70418	1 310 nm, 1 550 nm 1 310 nm 3.3 km Fiber 13.4 km Fiber 1 550 nm 3.3 km Fiber 13.4 km Fiber 1 310 nm 7.20 dB Fiber 2.90 dB Fiber 1 550 nm 4.20 dB Fiber 1.60 dB Fiber	0.08 nm 0.081 m 0.34 m 0.080 m 0.34 m 0.13 dB 0.05 dB 0.05 dB 0.05 dB	Optical length fiber reference, Optical fiberloss reference, Optical spectrum analyzer/ SICT-CP-70418
Return loss meters Return loss measure	70423	1 310 nm, 1 550 nm 20 dB ~ 40 dB	0.8 dB	Optical Returnloss generator SICT-CP-70423
Frequency stabilized lasers and LDs Wavelength optical power	70429	1 310 nm 1 550 nm 1 310 nm, 1 550 nm (-60 ~ 0) dBm	4 pm 4 pm 0.07 dB	Wavelength meter, Optical powermeter/ SICT-CP-70429

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
ASE light sources Wavelength output	70430	1 310 nm, 1 550 nm	0.058 nm	Optical spectrum analyzer, Optical powermeter/ SICT-CP-70430
Optical power output		1 310 nm, 1 550 nm (-60 ~ 0) dBm	0.07 dB	
Optical power stabilized lasers and LDs Wavelength output	70433	1 310 nm 1 550 nm	4 pm 4 pm	Wavelength meter, Optical powermeter/ SICT-CP-70433
Optical power output		1 310 nm, 1 550 nm (-60 ~ 0) dBm	0.07 dB	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.				
Breath alcohol analyzers	90101	(0.000 ~ 0.080) %BAC (0.080 ~ 0.190) %BAC	3.3×10^{-2} 2.1×10^{-2}	Standard gas/ SICT-CP-90101				
Dry process								
Wet process		(0.000 ~ 0.080) %BAC	2.9×10^{-2}					
		(0.080 ~ 0.150) %BAC	1.6×10^{-2}					
		(0.150 ~ 0.400) %BAC	1.3×10^{-2}					
Environmental air monitoring instruments	90102	Oxygen (0 ~ 700) $\mu\text{mol/mol}$ (0.07 ~ 1.5) cmol/mol (1.5 ~ 20) cmol/mol	1.0×10^{-2} 2.0×10^{-2} 1.2×10^{-2}	Standard gas/ SICT-CP-90102				
Carbon monoxide					(0 ~ 850) $\mu\text{mol/mol}$	2.2×10^{-2}		
					Carbon dioxide	(0 ~ 0.50) cmol/mol	2.0×10^{-2}	
(0.50 ~ 5.00) cmol/mol		1.5×10^{-2}						
(5.00 ~ 19.00) cmol/mol		2.1×10^{-2}						
Nitrogen monoxide		(0 ~ 850) $\mu\text{mol/mol}$	2.1×10^{-2}					
Isobutane		(0 ~ 0.8) cmol/mol	2.2×10^{-2}					
Methane		(0 ~ 2.0) cmol/mol	1.4×10^{-2}					
Hydrogen sulfide		(0 ~ 45) $\mu\text{mol/mol}$	3.6×10^{-2}					
Propane		(0 ~ 2 000) $\mu\text{mol/mol}$	3.0×10^{-2}					
Isobutylene		(0 ~ 25) $\mu\text{mol/mol}$	1.0×10^{-2}					
Ammonia		(0 ~ 50) $\mu\text{mol/mol}$	4.9×10^{-2}					
Sulfur dioxide		(0 ~ 850) $\mu\text{mol/mol}$	2.2×10^{-2}					
Nitrogen dioxide		(0 ~ 1 000) $\mu\text{mol/mol}$	1.0×10^{-2}					
Hydrogen		(0 ~ 500) $\mu\text{mol/mol}$	2.3×10^{-2}					
		(0.05 ~ 2.0) cmol/mol	2.1×10^{-2}					
Hydrogen chloride		(0 ~ 50) $\mu\text{mol/mol}$	4.8×10^{-2}					
Sulfur hexafluoride		(0 ~ 100) cmol/mol	0.1×10^{-2}					
Ozone		0.0 nmol/mol	2.2 nmol/mol					
		(0.0 ~ 1 000.0) nmol/mol	2.5×10^{-2}					
Gas analyzers		90103	Oxygen (0 ~ 700) $\mu\text{mol/mol}$ (0.07 ~ 1.5) cmol/mol (1.5 ~ 20) cmol/mol		1.0×10^{-2} 2.0×10^{-2} 1.2×10^{-2}	Standard gas/ SICT-CP-90103		
Carbon monoxide							(0 ~ 850) $\mu\text{mol/mol}$	2.2×10^{-2}
							Carbon dioxide	(0 ~ 0.50) cmol/mol
(0.50 ~ 5.00) cmol/mol	1.5×10^{-2}							
(5.00 ~ 19.00) cmol/mol	2.1×10^{-2}							
Nitrogen monoxide	(0 ~ 850) $\mu\text{mol/mol}$		2.1×10^{-2}					
Isobutane	(0 ~ 0.8) cmol/mol		2.2×10^{-2}					
Methane	(0 ~ 2.0) cmol/mol		1.4×10^{-2}					
Hydrogen sulfide	(0 ~ 45) $\mu\text{mol/mol}$		3.6×10^{-2}					
Propane	(0 ~ 2 000) $\mu\text{mol/mol}$		3.0×10^{-2}					
Isobutylene	(0 ~ 25) $\mu\text{mol/mol}$		1.0×10^{-2}					
Ammonia	(0 ~ 50) $\mu\text{mol/mol}$		4.9×10^{-2}					
Sulfur dioxide	(0 ~ 850) $\mu\text{mol/mol}$		2.2×10^{-2}					
Nitrogen dioxide	(0 ~ 1 000) $\mu\text{mol/mol}$		1.0×10^{-2}					
Hydrogen	(0 ~ 500) $\mu\text{mol/mol}$		2.3×10^{-2}					
	(0.05 ~ 2.0) cmol/mol		2.1×10^{-2}					
Hydrogen chloride	(0 ~ 50) $\mu\text{mol/mol}$		4.8×10^{-2}					
Sulfur hexafluoride	(0 ~ 100) cmol/mol		0.1×10^{-2}					
Ozone	0.0 nmol/mol		2.2 nmol/mol					
	(0.0 ~ 1 000.0) nmol/mol		2.5×10^{-2}					

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Exhaust gas test instruments	90104			Standard gas/ SICT-CP-90103
Oxygen		(0 ~ 1.5) $\mu\text{mol/mol}$	2.0×10^{-2}	
		(1.5 ~ 20) cmol/mol	1.1×10^{-2}	
Carbon monoxide		(0 ~ 5.0) cmol/mol	2.1×10^{-2}	
Carbon dioxide		(0 ~ 19) cmol/mol	2.0×10^{-2}	
Nitrogen monoxide		(0 ~ 2 000) $\mu\text{mol/mol}$	2.0×10^{-2}	
Isobutane		(0 ~ 0.8) cmol/mol	2.2×10^{-2}	
Methane		(0 ~ 2.0) cmol/mol	1.4×10^{-2}	
Propane		(0 ~ 2 000) $\mu\text{mol/mol}$	3.0×10^{-2}	
Ammonia		(0 ~ 50) $\mu\text{mol/mol}$	4.9×10^{-2}	
Sulfur dioxide		(0 ~ 850) $\mu\text{mol/mol}$	2.2×10^{-2}	
Nitrogen dioxide		(0 ~ 1 000) $\mu\text{mol/mol}$	1.0×10^{-2}	
Hydrogen		(0 ~ 500) $\mu\text{mol/mol}$	2.3×10^{-2}	
		(0.05 ~ 2.0) cmol/mol	2.1×10^{-2}	
Others: pH meter, Electrical conductivity meter	90104			CRM/ SICT-CP-90199
pH meter		(4 ~ 10) pH	0.013 pH	
Electrical conductivity meter		100 $\mu\text{S/cm}$	3.1 $\mu\text{S/cm}$	
		1 413 $\mu\text{S/cm}$	9.7 $\mu\text{S/cm}$	
		12.85 mS/cm	0.073 mS/cm	
		111.3 mS/cm	0.78 mS/cm	

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017 & KS Q ISO/IEC 17025:2017

Institute of Calibration Technology Co.,Ltd

47-22, Jingoksandan 2beon-ro, Gwangsan-gu, Gwangju

Phone : 82-62-953-5353, Fax : 82-62-953-5454 e-mail : sictadmin@sict.co.kr

Expanding Scope of Accreditation(Ra 6 items

Change Scope of Accreditation(Uncer 36 Items

CALIBRATION

Valid To : Oct. 29. 2025

Accreditation No : KC01-018(1/35)

In recognition of the successful completion of the KOLAS evaluation process,
accreditation is granted to this laboratory to perform the following calibrations

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site
102. Linear dimension			201. Mass			40424	Volt/Current recorders	Y
10211	Filler gauges	Y	20109	Electric balances	Y	501. Contact thermometry		
10216	Height gauges/measuring machines	Y	20112	Platform scale balances	Y	50101	Temperature generators: ovens, furnaces, isothermal liquid baths, ice-point baths, dry-block calibrators	Y
			20116	Weights	Y			
10228	Cylindrical plug/pin gauges, Thread measuring wire gauges	Y	401. DC Voltage & current					
			40101	DC ammeters	Y			
10234	Ultrasonic thickness gauges	Y	40103	DC voltage/current calibrators	Y	50102	Temperature indicators /recorders/controllers, temperature calibrators	Y
10236	Coating thickness testers	Y						
104. Form			40104	Electrical temperature calibrators	Y	50103	Glass thermometers; liquid-in-glass, Beckmann	N
10401	Form testers	Y						
10407	Precision surface plates	Y	40108	DC power supplies	Y	50104	Resistance thermometers; SPRT, IPRT, thermistors, etc.	Y
10409	Roundness measurement instruments	N	40112	DC voltmeters	Y			
			402. Resistance, Capacitance and Inductance					
10412	Straight edges	Y				50105	Thermal expansion thermometers ; bimetal, gas or liquid type	Y
105. Complex geometry			40205	Earth testers	Y			
10503	Contact coordinate measuring machines	Y	40210	Insulation testers	Y	50107	Temperature transducers	Y
			40214	Resistance meters	Y			
10504	Non-contact coordinate measuring machines	Y	40215	Resistors	Y	503. Humidity		
			403. AC voltage, current & power			50302	Relative humidity hygrometers polimer thin film, hair, etc.	Y
10511	Measuring microscopes, Profile projectors	Y	40301	AC ammeters	Y	50304	Temperature humidity recorders ; Hygrothermograph, etc	N
			40302	Clamp ammeters/voltmeters	Y			
10512	Microscopes, micro measuring	Y	40303	AC voltage/current calibrators	Y			
10517	Stylus type roughness testers	Y	40310	Power factor meters	Y	50305	Transducers; dew-point /relative humidity	N
106. Various dimensional			40311	AC power meters	Y			
10601	Inside/Outside/Gear tooth calipers, Caliper gauges	Y	40312	AC power supplies	Y	50306	Humidity generators; two-pressure, two-temperature, flow mixing humidity generator, constant temperature and humidity chamber, etc.	Y
			40313	Puncture/safety testers	Y			
10603	Cylinder/bore gauges	Y	40314	Power recorders	Y			
10604	Depth gauges, Depth micrometers	Y	40318	AC voltmeters	Y			
10605	Dial/digital gauges	Y	404. Other DC & LF Measurements					
10609	Microindicators, Test indicators	Y	40410	Line frequency meters	Y			
			40416	Leakage current testers	Y			
10611	3-points, Micrometers	Y	40417	Electronic AC/DC loads	Y			
10612	Inside micrometers	Y	40419	Analogue/Digital multimeters	Y			
10613	Outside micrometers	Y	40421	Oscilloscopes	Y			

Accreditation No : KC01-018(2/35)

Note

1. This laboratory provides calibration services in permanent standard laboratory and at on-site.
2. Laboratory conducts on-site calibration should meet requirements of KOLAS-SR-007.
3. On-site calibration is allowed to items with marking 'Y', not allowed to items with marking 'N'.
4. Measurement uncertainty normally is quoted as an expanded uncertainty at a coverage probability of 95%, which usually requires the use of a coverage factor of $k=2$. It expresses the lowest uncertainty of measurement that can be provided by accredited calibration laboratories in normal conditions.
5. Due to the calibration environment such as reference standards or customers' facilities, it is note that uncertainty of measurement on a calibration certificate may be expressed larger than measurement uncertainty on scope of accreditation in general.

102. Linear dimension

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Filler gauges	10211	(0.01 ~ 10) mm	1.2 μm	Micrometers/ SICT-CP-10211
Height gauges/measuring machines	10216	(0 ~ 1 010) mm	$\sqrt{1.1^2 + (0.0037 \times l_0)^2}$ μm	Gauge Block, Step gauge/ SICT-CP-10216
Cylindrical plug/pin gauges, Thread measuring wire gauges Cylindrical plug/pin gauges	10228	(1 ~ 20) mm	$\sqrt{0.48^2 + (0.0034 \times l_0)^2}$ μm	Laser Scan Micrometers/ SICT-CP-10228
Ultrasonic thickness gauges	10234	(2.5 ~ 100) mm	3 μm	Ultrasonic Tester Blocks/ SICT-CP-10234
Coating thickness testers	10236	(0 ~ 1.527) mm	1.1 μm	Thickness specimens/ SICT-CP-10236

Note 1. l_0 unit : mm

104. Form

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Form testers Height length Width Angle	10401	(0 ~ 50) mm (0 ~ 45) mm 15° ~ 45°	1.0 μm 1.7 μm 2'	Form Standard Specimens Gage Block, Angle Gage Block/ SICT-CP-10401
Precision surface plates Flatness	10407	(2 000 ~ 20 000) cm ²	2.0 μm	Electronic Level/ SICT-CP-10407
Roundness measurement instruments Detector accuracy Rotational accuracy of spindle Rotational accuracy of axis	10409	(0 ~ 15) μm 360° 360°	0.62 μm 0.09 μm 0.05 μm	Roundness Standard Ball/ SICT-CP-10409
Straight edges	10412	(0 ~ 1 500) mm (0 ~ 1 500) mm	1.9 μm 1.8 μm	Electronic Micrometers/ SICT-CP-10412

Note 1. l_0 unit : mm

105. Complex geometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Contact coordinate measuring machines	10503	(0 ~ 1 000) mm	$\sqrt{0.94^2 + (0.0048 \times l_0)^2}$ μm	Step Gauge/ SICT-CP-10503
Non-contact coordinate measuring machines	10504	(0 ~ 500) mm	$\sqrt{0.53^2 + (0.0036 \times l_0)^2}$ μm	Standard Scale/ SICT-CP-10504
Measuring microscopes, Profile projectors Length Magnification Angle	10511	(0 ~ 300) mm (5 ~ 100) 배 (0 ~ 360) °	$\sqrt{0.56^2 + (0.0036 \times l_0)^2}$ μm 0.05 % 0.9'	Standard Scale/ SICT-CP-10511
Microscopes, micro measuring	10512	(0 ~ 1) mm (1 ~ 50) mm	1.3 μm 3.0 μm	Standard Scale/ SICT-CP-10512
Stylus type roughness testers Ra Rz	10517	(0 ~ 4) μm (0 ~ 12) μm	추후 확정 추후 확정	Roughness Specimen/ SICT-CP-10517

Note 1. l_0 unit : mm

106. Various dimensional

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Inside/Outside/Geartooth calipers, Caliper gauges Inside/Outside calipers Caliper gauges	10601	(0 ~ 2 000) mm	$\sqrt{8.2^2 + (0.008 \ 1 \times l_0)^2}$ μm	Gauge Block/ SICT-CP-10601
		(0 ~ 100) mm	$\sqrt{2.5^2 + (0.004 \ 2 \times l_0)^2}$ μm	
		(100 ~ 300) mm	$\sqrt{8.2^2 + (0.004 \ 2 \times l_0)^2}$ μm	
Cylinder/bore gauges Cylinder gauges Hole gauges	10603	(0 ~ 400) mm	0.7 μm	Dial Gauge Tester/ SICT-CP-10603
		(0.1 ~ 25) mm	3.5 μm	
Depth gauges,Depth micrometers Depth micrometers Depth gauges	10604	(0 ~ 300) mm	$\sqrt{0.87^2 + (0.003 \ 3 \times l_0)^2}$ μm	Gauge Block/ SICT-CP-10604
		(0 ~ 600) mm	$\sqrt{6.0^2 + (0.007 \ 8 \times l_0)^2}$ μm	
Dial/digital gauges Dial/Digital gauges Digital thickness gauges	10605	(0 ~ 100) mm	$\sqrt{0.33^2 + (0.006 \ 8 \times l_0)^2}$ μm	Gauge Block/ SICT-CP-10605
		(0 ~ 25) mm	$\sqrt{0.82^2 + (0.006 \ 8 \times l_0)^2}$ μm	
Micro indicators,Test indicators	10609	(0 ~ 5) mm	0.29 μm	Dial Gauge Tester/ SICT-CP-10609
3-points, Micrometers	10611	(6 ~ 100) mm	$\sqrt{1.2^2 + (0.004 \ 1 \times l_0)^2}$ μm	Ring Gauge/ SICT-CP-10611
Inside micrometers Inside micrometer bar type micrometer(Accuracy of scale) bar type micrometer(Length of extension bars)	10612	(5 ~ 200) mm	$\sqrt{1.1^2 + (0.004 \ 2 \times l_0)^2}$ μm	Gauge Block/ SICT-CP-10612
		(25 ~ 200) mm	$\sqrt{1.1^2 + (0.004 \ 2 \times l_0)^2}$ μm	
		(13 ~ 150) mm	$\sqrt{1.5^2 + (0.004 \ 2 \times l_0)^2}$ μm	
Outside micrometers Outside micrometers V-anvil micrometers	10613	(0 ~ 25) mm	$\sqrt{0.22^2 + (0.003 \times l_0)^2}$ μm	Gauge Block, cylindrical plug gauges/ SICT-CP-10613
		(25 ~ 500) mm	$\sqrt{0.83^2 + (0.003 \ 1 \times l_0)^2}$ μm	
		(1 ~ 20) mm	0.97 μm	

Note 1. l_0 unit : mm

201. Mass

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electric balances	20109	(0 ~ 2) g (2 ~ 6) g (6 ~ 20) g (20 ~ 50) g (50 ~ 200) g (200 ~ 500) g (500 ~ 2 000) g (2 ~ 5) kg (5 ~ 10) kg (10 ~ 20) kg (20 ~ 30) kg (30 ~ 60) kg (60 ~ 100) kg (100 ~ 600) kg (600 ~ 1 000) kg	21 µg 40 µg 42 µg 61 µg 0.12 mg 0.33 mg 1.2 mg 3.2 mg 6.2 mg 12 mg 0.20 g 0.36 g 6.8 g 18 g 38 g	Weight/ SICT-CP-20109
Platform scale balances	20112	(0 ~ 200) kg (200 ~ 500) kg (500 ~ 1 000) kg	46 g 0.092 kg 0.46 kg	Weight/ SICT-CP-20112
Weights less than class F1	20116	(1 mg ~ 20 kg) 1 mg 2 mg 5 mg 10 mg 20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg	3.5 µg 3.5 µg 3.5 µg 3.5 µg 4.4 µg 5.3 µg 6.4 µg 7.5 µg 10 µg 12 µg 16 µg 19 µg 24 µg 31 µg 38 µg 64 µg 0.12 mg 0.39 mg 0.66 mg 1.2 mg 4.7 mg 7.1 mg 13 mg	Weight, Electronic Balance/ SICT-CP-20116

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC ammeters DC Current	40101	(±) 0 μA (0.1 ~ 1) μA (1 ~ 10) μA (10 ~ 100) μA (0.1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A (10 ~ 20) A	9.3 nA 9.4×10^{-3} 9.9×10^{-4} 1.5×10^{-4} 6.9×10^{-5} 8.0×10^{-5} 1.2×10^{-4} 6.5×10^{-4} 1.2×10^{-3}	Calibrator/ SICT-CP-40101
DC voltage/current calibrators DC Voltage DC Current	40103	(±) 0 mV (0 ~ 1) mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V (±) 0 μA (0 ~ 10) μA (10 ~ 100) μA (0.1 ~ 1) mA (1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A (10 ~ 100) A	0.51 μV 6.1×10^{-4} 6.2×10^{-5} 1.3×10^{-5} 6.6×10^{-6} 5.2×10^{-6} 7.2×10^{-6} 9.0×10^{-6} 6.9 nA 7.4×10^{-4} 1.1×10^{-4} 7.2×10^{-5} 4.7×10^{-5} 7.5×10^{-5} 2.4×10^{-4} 5.7×10^{-4} 2.8×10^{-4}	Reference Multimeter/ SICT-CP-40103
Electrical temperature calibrators TEMPERATURE(SOURCE) T/C RTD	40104	-9.835 mV (-9.835 ~ -5.237) mV (-5.237 ~ 0.000) mV (0.000 ~ 13.421) mV (13.421 ~ 68.788) mV (68.788 ~ 76.373) mV 1.000 Ω (1.000 ~ 2.499) Ω (2.499 ~ 16.996) Ω (16.996 ~ 100.000) Ω (100.000 ~ 249.584) Ω (249.584 ~ 3 233.3) Ω	0.62 μV 0.62 μV 0.52 μV 0.62 μV 1.2 μV 1.3 μV 0.063 mΩ 3.7×10^{-5} 3.0×10^{-5} 1.0×10^{-5} 2.8×10^{-5} 1.8×10^{-5}	Digital Multimeter/ SICT-CP-40104

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrical temperature calibrators	40104			Digital Multimeter/ SICT-CP-40104
DC VOLTAGE(SOURCE)		(±)		
		0 mV	8.1 μV	
		(0 ~ 10) mV	8.5×10^{-4}	
		(10 ~ 100) mV	1.2×10^{-4}	
		(0.1 ~ 1) V	4.6×10^{-5}	
		(1 ~ 10) V	4.0×10^{-5}	
		(10 ~ 100) V	5.9×10^{-5}	
DC CURRENT(SOURCE)		(±)		
		0 mA	0.14 μA	
		(0 ~ 1) mA	7.0×10^{-4}	
		(1 ~ 10) mA	1.0×10^{-3}	
		(10 ~ 100) mA	7.0×10^{-4}	
RESISTANCE(SOURCE)		0 Ω	0.072 mΩ	
		(0 ~ 1) Ω	6.7×10^{-5}	
	(1 ~ 10) Ω	4.6×10^{-4}		
	100 Ω ~ 50 kΩ	1.2×10^{-4}		
TEMPERATURE(MEASURE)				
T/C	-9.836 mV	0.84 μV		
	(-9.836 ~ -5.238) mV	0.83 μV		
	(-5.238 ~ 0.000) mV	0.80 μV		
	(0.000 ~ 0.002) mV	0.70 μV		
	(0.002 ~ 6.319) mV	0.81 μV		
	(6.319 ~ 13.421) mV	0.87 μV		
	(13.421 ~ 21.036) mV	0.94 μV		
	(21.036 ~ 28.946) mV	1.0 μV		
	(28.946 ~ 37.006) mV	1.1 μV		
	(37.006 ~ 53.112) mV	1.2 μV		
	(53.112 ~ 61.017) mV	1.3 μV		
	(61.017 ~ 76.373) mV	1.4 μV		
	RTD			
	0.999 Ω	0.24 mΩ		
	(0.999 ~ 2.497) Ω	1.0×10^{-4}		
	(2.497 ~ 4.316) Ω	7.1×10^{-5}		
	(4.316 ~ 16.995) Ω	3.9×10^{-5}		
	(16.995 ~ 177.156) Ω	3.4×10^{-5}		
	(177.156 ~ 249.584) Ω	3.5×10^{-5}		
	(249.584 ~ 3 233.3) Ω	4.3×10^{-5}		
DC VOLTAGE(MEASURE)	(±)			
	0 mV	1.2 μV		
	(0 ~ 10) mV	1.4×10^{-4}		
	(10 ~ 100) mV	3.5×10^{-5}		
	(0.1 ~ 1) V	6.0×10^{-5}		
	(1 ~ 10) V	1.7×10^{-5}		
	(10 ~ 200) V	2.3×10^{-5}		
	(200 ~ 300) V	2.9×10^{-5}		
DC CURRENT(MEASURE)	(±)			
	0 mA	0.09 μA		
	(0 ~ 1) mA	9.3×10^{-5}		
	(1 ~ 10) mA	7.0×10^{-5}		
	(10 ~ 50) mA	9.6×10^{-5}		
	(50 ~ 100) mA	8.1×10^{-5}		
	(100 ~ 130) mA	9.6×10^{-5}		

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrical temperature calibrators RESISTANCE (MEASURE)	40104	0 Ω (0 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) kΩ (1 ~ 10) kΩ (10 ~ 20) kΩ (20 ~ 30) kΩ (30 ~ 40) kΩ (40 ~ 50) kΩ	23 mΩ 1.3×10^{-4} 4.8×10^{-5} 3.4×10^{-5} 7.3×10^{-5} 7.0×10^{-5} 4.8×10^{-5} 4.0×10^{-5} 4.2×10^{-5} 3.9×10^{-5}	Digital Multimeter/ SICT-CP-40104
DC power supplies DC Voltage DC Current Load regulation Ripple	40108	(±) 0 mV (0 ~ 10) mV (10 ~ 100) mV 100 mV ~ 1 V (1 ~ 10) V (10 ~ 100) V (100 ~ 600) V (600 ~ 1 000) V 1 mA (1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 5) A (5 ~ 10) A (10 ~ 20) A (20 ~ 200) A 0 mV (0 ~ 2) mV (2 ~ 20) mV (20 ~ 200) mV 0.1 mV (0.1 ~ 0.2) mV (0.2 ~ 0.4) mV (0.4 ~ 0.6) mV (0.6 ~ 1) mV (1 ~ 10) mV (10 ~ 50) mV	5.8 μV 5.8×10^{-4} 5.9×10^{-5} 9.2×10^{-5} 5.3×10^{-5} 9.2×10^{-5} 1.7×10^{-5} 6.6×10^{-5} 5.8 μA 6.5×10^{-3} 6.5×10^{-4} 8.7×10^{-5} 1.3×10^{-4} 4.0×10^{-4} 3.1×10^{-4} 2.2×10^{-4} 0.97 mV 4.9×10^{-1} 3.3×10^{-1} 3.3×10^{-2} 0.097 mV 6.4×10^{-1} 2.5×10^{-1} 1.6×10^{-1} 9.6×10^{-2} 9.2×10^{-2} 1.9×10^{-1}	DC Electronics Load/ SICT-CP-40108
DC voltmeters DC Voltage	40112	(±) 0 mV (0 ~ 1) mV (1 ~ 2) mV (2 ~ 8) mV (8 ~ 10) mV (10 ~ 20) mV (20 ~ 50) mV (50 ~ 80) mV (80 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V	0.69 μV 8.0×10^{-4} 4.0×10^{-4} 1.6×10^{-4} 8.0×10^{-5} 4.5×10^{-5} 2.4×10^{-5} 1.9×10^{-5} 1.6×10^{-5} 9.4×10^{-6} 8.6×10^{-6} 1.0×10^{-5} 1.1×10^{-5}	Calibrator/ SICT-CP-40112

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Earth testers	40205			Decade Resistor/ SICT-CP-40205
Test Voltage		60 Hz 1 V (1 ~ 10) V (10 ~ 50) V (50 ~ 100) V (100 ~ 500) V (500 ~ 1 000) V	6.1×10^{-3} 6.2×10^{-4} 1.2×10^{-3} 6.2×10^{-4} 4.9×10^{-4} 7.7×10^{-4}	
Resistance		1 mΩ 1 mΩ ~ 1 Ω 1 Ω ~ 10 Ω 10 Ω ~ 100 kΩ	1.3×10^{-3} 1.3×10^{-3} 8.4×10^{-4} 6.5×10^{-4}	
AC Current out		60 Hz 1 A (1 ~ 3) A (3 ~ 10) A (10 ~ 20) A (20 ~ 30) A (30 ~ 60) A	3.1×10^{-3} 4.1×10^{-3} 3.1×10^{-3} 3.4×10^{-3} 6.4×10^{-3} 4.6×10^{-3}	
Timer		(1 ~ 100) s (100 ~ 1 000) s	5.8×10^{-6} 8.2×10^{-6}	
Insulation testers	40210			High Resistance Decade/ SICT-CP-40210
AC Voltage		60 Hz 1 V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V	1.2×10^{-4} 1.2×10^{-4} 1.3×10^{-4} 4.9×10^{-4}	
Insulation Voltage		1 V (1 ~ 10) V (10 ~ 25) V (25 ~ 50) V (50 ~ 100) V (100 ~ 250) V (250 ~ 500) V (500 ~ 1 000) V (1 000 ~ 5 000) V (5 000 ~ 10 000) V	6.1×10^{-4} 6.1×10^{-5} 2.5×10^{-4} 1.2×10^{-4} 6.2×10^{-5} 2.5×10^{-4} 1.2×10^{-4} 6.1×10^{-5} 6.5×10^{-3} 6.1×10^{-3}	
Insulation Resistance		1 kΩ 1 kΩ ~ 1 MΩ (1 ~ 10) MΩ (10 ~ 100) MΩ 100 MΩ ~ 1 GΩ (1 ~ 10) GΩ (10 ~ 100) GΩ 100 GΩ ~ 1 TΩ	1.4×10^{-4} 1.2×10^{-4} 3.6×10^{-4} 1.2×10^{-3} 2.3×10^{-3} 5.8×10^{-3} 5.9×10^{-3} 3.5×10^{-2}	
Resistance meters	40214			Standard Resistance Set/ SICT-CP-40214
Resistance		1 mΩ (1 ~ 10) mΩ (10 ~ 100) mΩ 100 mΩ ~ 1 MΩ	2.2×10^{-5} 2.1×10^{-5} 2.2×10^{-5} 2.1×10^{-5}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistors Decade Resistance	40215	0 Ω (0 ~ 1) mΩ (1 ~ 2) mΩ (2 ~ 3) mΩ (3 ~ 6) mΩ (6 ~ 8) mΩ (8 ~ 10) mΩ (10 ~ 30) mΩ (30 ~ 60) mΩ 60 mΩ ~ 0.1 Ω (0.1 ~ 0.3) Ω (0.3 ~ 0.6) Ω (0.6 ~ 1) Ω (1 ~ 3) Ω (3 ~ 10) Ω (10 ~ 300) Ω (300 ~ 1000) Ω (1 ~ 10) kΩ (10 ~ 100) kΩ 100 kΩ ~ 1 MΩ (1 ~ 2) MΩ (2 ~ 10) MΩ (10 ~ 100) MΩ 100 MΩ ~ 1 GΩ	62 μΩ 3.1×10^{-3} 6.2×10^{-3} 2.1×10^{-3} 1.5×10^{-3} 8.9×10^{-4} 6.9×10^{-4} 3.1×10^{-3} 1.6×10^{-3} 8.9×10^{-4} 3.1×10^{-4} 1.6×10^{-4} 9.0×10^{-5} 4.6×10^{-5} 2.5×10^{-5} 2.9×10^{-5} 1.6×10^{-5} 3.7×10^{-5} 3.0×10^{-5} 4.0×10^{-5} 8.5×10^{-5} 5.9×10^{-5} 6.5×10^{-4} 6.6×10^{-3}	Standard Resistance Set/ SICT-CP-40215

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC ammeters AC Current	40301	(100 μ A) 50 Hz ~ 1 kHz (0.1 ~ 1) mA 50 Hz ~ 1 kHz (1 ~ 10) mA (50 ~ 60) Hz 60 Hz ~ 1 kHz (10 ~ 100) mA (50 ~ 60) Hz 60 Hz ~ 1 kHz (0.1 ~ 1) A (50 ~ 60) Hz 60 Hz ~ 1 kHz (1 ~ 5) A (50 ~ 60) Hz 60 Hz ~ 1 kHz (5 ~ 10) A (50 ~ 60) Hz 60 Hz ~ 1 kHz (10 ~ 20) A (50 ~ 60) Hz (20 ~ 40) A (50 ~ 60) Hz	 2.7×10^{-3} 1.4×10^{-3} 6.1×10^{-4} 7.0×10^{-4} 4.7×10^{-4} 7.0×10^{-4} 2.1×10^{-4} 7.0×10^{-4} 2.6×10^{-4} 1.3×10^{-3} 3.0×10^{-4} 1.4×10^{-3} 4.8×10^{-4} 5.2×10^{-4}	Power Calibrator, Calibrator/ SICT-CP-40301
Clamp ammeters/voltmeters AC Current	40302	(50 ~ 60) Hz 100 μ A (100 ~ 200) μ A (200 ~ 300) μ A (300 ~ 400) μ A (400 ~ 500) μ A (500 ~ 600) μ A (600 ~ 900) μ A 900 μ A ~ 1 mA (1 ~ 2) mA (2 ~ 3) mA (3 ~ 4) mA (4 ~ 5) mA (5 ~ 6) mA (6 ~ 7) mA (7 ~ 8) mA (8 ~ 9) mA (9 ~ 10) mA (10 ~ 20) mA (20 ~ 30) mA (30 ~ 40) mA (40 ~ 50) mA (50 ~ 60) mA (60 ~ 70) mA	 0.27 μ A 2.1×10^{-3} 1.9×10^{-3} 1.7×10^{-3} 1.6×10^{-3} 1.5×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.3×10^{-3} 1.2×10^{-3} 1.1×10^{-3} 9.5×10^{-4} 8.7×10^{-4} 8.1×10^{-4} 7.6×10^{-4} 7.3×10^{-4} 9.3×10^{-4} 7.6×10^{-4} 6.5×10^{-4} 6.0×10^{-4} 5.7×10^{-4} 4.8×10^{-4} 4.6×10^{-4}	Power Calibrator, Calibrator/ SICT-CP-40302

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Clamp ammeters/voltmeters	40302	AC Current	(70 ~ 90) mA (90 ~ 100) mA (100 ~ 200) mA (200 ~ 300) mA (300 ~ 400) mA (400 ~ 500) mA (500 ~ 700) mA (700 ~ 900) mA 900 mA ~ 1 A (1 ~ 2) A (2 ~ 3) A (3 ~ 4) A (4 ~ 5) A (5 ~ 6) A (6 ~ 7) A (7 ~ 9) A (9 ~ 10) A (10 ~ 20) A (20 ~ 30) A (30 ~ 40) A (40 ~ 500) A (500 ~ 700) A (700 ~ 1 000) A	4.4×10^{-4} 7.5×10^{-4} 3.9×10^{-4} 3.1×10^{-4} 2.7×10^{-4} 2.5×10^{-4} 2.4×10^{-4} 2.2×10^{-4} 6.4×10^{-4} 3.7×10^{-4} 4.0×10^{-4} 3.3×10^{-4} 2.9×10^{-4} 3.4×10^{-4} 3.3×10^{-4} 3.1×10^{-4} 6.8×10^{-4} 6.4×10^{-4} 5.6×10^{-4} 5.2×10^{-4} 1.2×10^{-3} 1.4×10^{-3} 1.3×10^{-3}	Power Calibrator, Calibrator/ SICT-CP-40302
		DC Current	30 μ A (30 ~ 40) μ A (40 ~ 50) μ A (50 ~ 60) μ A (60 ~ 70) μ A (70 ~ 80) μ A (80 ~ 90) μ A (90 ~ 100) μ A (100 ~ 200) μ A (200 ~ 300) μ A (300 ~ 400) μ A (400 ~ 500) μ A (500 ~ 600) μ A (600 ~ 700) μ A (700 ~ 800) μ A (800 ~ 900) μ A 900 μ A ~ 1 mA (1 ~ 2) mA (2 ~ 4) mA (4 ~ 9) mA (9 ~ 10) mA (10 ~ 20) mA (20 ~ 70) mA (70 ~ 90) mA (90 ~ 100) mA (100 ~ 200) mA (200 ~ 400) mA	29 nA 7.7×10^{-4} 6.5×10^{-4} 5.7×10^{-4} 5.1×10^{-4} 4.7×10^{-4} 4.4×10^{-4} 7.3×10^{-4} 4.2×10^{-4} 3.2×10^{-4} 2.8×10^{-4} 2.5×10^{-4} 2.4×10^{-4} 2.3×10^{-4} 2.2×10^{-4} 2.1×10^{-4} 6.4×10^{-4} 3.6×10^{-4} 2.7×10^{-4} 1.8×10^{-4} 6.2×10^{-4} 3.3×10^{-4} 2.4×10^{-4} 1.5×10^{-4} 6.2×10^{-4} 3.3×10^{-4} 3.9×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Clamp ammeters/voltmeters	40302	DC Current		Power Calibrator, Calibrator/ SICT-CP-40302	
		(400 ~ 900) mA	3.5×10^{-4}		
		900 mA ~ 1 A	6.7×10^{-4}		
		(1 ~ 2) A	4.8×10^{-4}		
		(2 ~ 3) A	6.0×10^{-4}		
		(3 ~ 9) A	8.1×10^{-4}		
		(9 ~ 10) A	8.9×10^{-4}		
		(10 ~ 20) A	1.2×10^{-3}		
		(20 ~ 40) A	3.8×10^{-4}		
		(40 ~ 500) A	1.3×10^{-3}		
		(500 ~ 1 000) A	1.7×10^{-3}		
		AC Voltage	(50 ~ 60) Hz		
		10 mV	11 μ V		
		(10 ~ 20) mV	6.3×10^{-4}		
		(20 ~ 30) mV	4.7×10^{-4}		
		(30 ~ 40) mV	4.6×10^{-4}		
		(40 ~ 50) mV	4.0×10^{-4}		
		(50 ~ 60) mV	3.6×10^{-4}		
		(60 ~ 80) mV	3.3×10^{-4}		
		(80 ~ 90) mV	2.9×10^{-4}		
		(90 ~ 100) mV	2.8×10^{-4}		
		(100 ~ 200) mV	2.2×10^{-4}		
		(200 ~ 300) mV	2.0×10^{-4}		
		(300 ~ 400) mV	3.6×10^{-4}		
		(400 ~ 600) mV	3.2×10^{-4}		
		(600 ~ 700) mV	2.8×10^{-4}		
		(700 ~ 800) mV	2.6×10^{-4}		
		800 mV ~ 1 V	2.5×10^{-4}		
		(1 ~ 3) V	2.1×10^{-4}		
		(3 ~ 4) V	3.5×10^{-4}		
		(4 ~ 5) V	3.2×10^{-4}		
		(5 ~ 6) V	2.9×10^{-4}		
		(6 ~ 7) V	2.8×10^{-4}		
		(7 ~ 8) V	2.6×10^{-4}		
		(8 ~ 10) V	2.5×10^{-4}		
		(10 ~ 20) V	2.3×10^{-4}		
(20 ~ 40) V	1.9×10^{-4}				
(40 ~ 70) V	2.2×10^{-4}				
(70 ~ 90) V	1.9×10^{-4}				
(90 ~ 100) V	1.8×10^{-4}				
(100 ~ 200) V	1.5×10^{-4}				
(200 ~ 500) V	1.9×10^{-4}				
(500 ~ 700) V	2.6×10^{-4}				
(700 ~ 1 000) V	3.7×10^{-4}				
DC Voltage	10 mV	61 μ V			
(10 ~ 20) mV	3.1×10^{-3}				
(20 ~ 30) mV	2.0×10^{-3}				
(30 ~ 40) mV	1.5×10^{-3}				
(40 ~ 50) mV	1.2×10^{-3}				
(50 ~ 60) mV	1.0×10^{-3}				
(60 ~ 70) mV	8.7×10^{-4}				
(70 ~ 80) mV	7.6×10^{-4}				
(80 ~ 90) mV	6.8×10^{-4}				
(90 ~ 100) mV	7.0×10^{-5}				

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.			
Clamp ammeters/voltmeters DC Voltage	40302	(100 ~ 200) mV	4.2×10^{-5}	Power Calibrator, Calibrator/ SICT-CP-40302			
		(200 ~ 300) mV	3.4×10^{-5}				
		(300 ~ 400) mV	2.4×10^{-5}				
		(400 ~ 500) mV	2.2×10^{-5}				
		(500 ~ 600) mV	2.0×10^{-5}				
		(600 ~ 800) mV	1.8×10^{-5}				
		(800 ~ 900) mV	1.7×10^{-5}				
		900 mV ~ 1 V	6.3×10^{-5}				
		(1 ~ 2) V	3.4×10^{-5}				
		(2 ~ 4) V	2.5×10^{-5}				
		(4 ~ 5) V	2.2×10^{-5}				
		(5 ~ 6) V	2.0×10^{-5}				
		(6 ~ 7) V	1.9×10^{-5}				
		(7 ~ 9) V	1.8×10^{-5}				
		(9 ~ 10) V	6.3×10^{-5}				
		(10 ~ 20) V	3.4×10^{-5}				
		(20 ~ 30) V	2.5×10^{-5}				
		(30 ~ 40) V	3.0×10^{-6}				
		(40 ~ 50) V	2.7×10^{-5}				
		(50 ~ 60) V	2.6×10^{-5}				
		(60 ~ 70) V	2.5×10^{-5}				
		(70 ~ 90) V	2.4×10^{-5}				
		(90 ~ 100) V	6.5×10^{-5}				
		(100 ~ 200) V	3.8×10^{-5}				
		(200 ~ 400) V	3.0×10^{-5}				
		(400 ~ 500) V	2.8×10^{-5}				
		(500 ~ 600) V	2.6×10^{-5}				
		(600 ~ 700) V	2.5×10^{-5}				
		(700 ~ 900) V	2.4×10^{-5}				
		(900 ~ 1 000) V	6.5×10^{-5}				
		Resistance	40302		1 Ω	0.62 mΩ	
					(1 ~ 9) Ω	3.1×10^{-4}	
					(9 ~ 100) Ω	6.1×10^{-4}	
(100 ~ 900) Ω	4.7×10^{-5}						
900 Ω ~ 90 kΩ	7.0×10^{-5}						
90 kΩ ~ 1 MΩ	7.2×10^{-5}						
(1 ~ 10) MΩ	1.9×10^{-4}						
(10 ~ 100) MΩ	7.1×10^{-4}						
AC voltage/current calibrators AC Voltage	40303	(2 mV)		Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303			
		10 Hz	8.2×10^{-3}				
		(10 ~ 40) Hz	2.5×10^{-3}				
		40 Hz ~ 1 kHz	1.3×10^{-3}				
		(1 ~ 20) kHz	5.0×10^{-3}				
		(20 ~ 100) kHz	1.2×10^{-2}				
		(2 ~ 20) mV					
		10 Hz	9.5×10^{-4}				
		(10 ~ 40) Hz	3.5×10^{-4}				
		40 Hz ~ 1 kHz	2.3×10^{-4}				
		(1 ~ 20) kHz	8.1×10^{-4}				
		(20 ~ 100) kHz	2.0×10^{-3}				

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators AC Voltage	40303	(20 ~ 30) mV 10 Hz (10 ~ 40) Hz 40 Hz ~ 1 kHz (1 ~ 20) kHz (20 ~ 100) kHz (30 ~ 40) mV 10 Hz (10 ~ 40) Hz 40 Hz ~ 1 kHz (1 ~ 20) kHz (20 ~ 100) kHz (40 ~ 50) mV 10 Hz (10 ~ 40) Hz 40 Hz ~ 1 kHz (1 ~ 20) kHz (20 ~ 100) kHz (50 ~ 60) mV 10 Hz (10 ~ 40) Hz 40 Hz ~ 1 kHz (1 ~ 20) kHz (20 ~ 100) kHz (70 ~ 90) mV 10 Hz (10 ~ 40) Hz 40 Hz ~ 1 kHz (1 ~ 20) kHz (20 ~ 100) kHz (90 ~ 100) mV 10 Hz (10 ~ 40) Hz 40 Hz ~ 1 kHz (1 ~ 20) kHz (20 ~ 100) kHz (100 ~ 200) mV 10 Hz (10 ~ 40) Hz 40 Hz ~ 1 kHz (1 ~ 20) kHz (20 ~ 100) kHz (200 ~ 300) mV 10 Hz (10 ~ 40) Hz 40 Hz ~ 1 kHz (1 ~ 20) kHz (20 ~ 100) kHz	6.8×10^{-4} 2.7×10^{-4} 1.9×10^{-4} 6.5×10^{-4} 1.6×10^{-3} 5.4×10^{-4} 2.3×10^{-4} 1.7×10^{-4} 5.8×10^{-4} 1.4×10^{-3} 4.6×10^{-4} 2.1×10^{-4} 1.6×10^{-4} 5.3×10^{-4} 1.3×10^{-3} 4.1×10^{-4} 1.9×10^{-4} 1.5×10^{-4} 5.0×10^{-4} 1.2×10^{-3} 3.7×10^{-4} 1.8×10^{-4} 1.5×10^{-4} 4.8×10^{-4} 1.1×10^{-3} 3.0×10^{-4} 1.6×10^{-4} 1.4×10^{-4} 4.4×10^{-4} 1.0×10^{-3} 2.3×10^{-4} 1.4×10^{-4} 1.3×10^{-4} 3.9×10^{-4} 9.3×10^{-4} 8.6×10^{-4} 2.5×10^{-4} 2.0×10^{-4} 4.1×10^{-4} 1.4×10^{-3}	Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators AC Voltage	40303	(300 mV ~ 0.4 V) 10 Hz (10 ~ 40) Hz 40 Hz ~ 1 kHz (1 ~ 20) kHz (20 ~ 100) kHz (0.4 ~ 0.9) V 10 Hz (10 ~ 40) Hz 40 Hz ~ 1 kHz (1 ~ 20) kHz (20 ~ 100) kHz (0.9 ~ 2) V 10 Hz (10 ~ 40) Hz 40 Hz ~ 1 kHz (1 ~ 20) kHz (20 ~ 100) kHz (2 ~ 3) V 10 Hz (10 ~ 40) Hz 40 Hz ~ 1 kHz (1 ~ 20) kHz (20 ~ 100) kHz (3 ~ 9) V 10 Hz (10 ~ 40) Hz 40 Hz ~ 1 kHz (1 ~ 20) kHz (20 ~ 100) kHz (9 ~ 20) V 10 Hz (10 ~ 40) Hz 40 Hz ~ 1 kHz (1 ~ 20) kHz (20 ~ 100) kHz (20 ~ 30) V 10 Hz (10 ~ 40) Hz 40 Hz ~ 1 kHz (1 ~ 20) kHz (20 ~ 100) kHz (30 ~ 90) V 10 Hz (10 ~ 40) Hz 40 Hz ~ 1 kHz (1 ~ 20) kHz (20 ~ 100) kHz	6.6×10^{-4} 2.1×10^{-4} 1.6×10^{-4} 3.6×10^{-4} 1.2×10^{-3} 5.7×10^{-4} 1.9×10^{-4} 1.6×10^{-4} 3.3×10^{-4} 1.2×10^{-3} 4.1×10^{-4} 1.6×10^{-4} 1.1×10^{-4} 2.8×10^{-4} 8.2×10^{-4} 8.6×10^{-4} 2.3×10^{-4} 2.0×10^{-4} 4.1×10^{-4} 1.4×10^{-3} 6.6×10^{-4} 2.0×10^{-4} 1.6×10^{-4} 3.6×10^{-4} 1.2×10^{-3} 4.1×10^{-4} 1.5×10^{-4} 1.1×10^{-4} 2.9×10^{-4} 8.4×10^{-4} 8.8×10^{-4} 2.5×10^{-4} 2.0×10^{-4} 4.1×10^{-4} 1.4×10^{-3} 6.8×10^{-4} 2.1×10^{-4} 1.6×10^{-4} 3.6×10^{-4} 1.2×10^{-3}	Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators	40303			Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
AC Voltage		(90 ~ 200) V 10 Hz	8.6×10^{-4}	
		(10 ~ 40) Hz	2.3×10^{-4}	
		40 Hz ~ 1 kHz	2.0×10^{-4}	
		(1 ~ 20) kHz	4.1×10^{-4}	
		(20 ~ 100) kHz	1.4×10^{-3}	
		(200 ~ 300) V 50 Hz ~ 1 kHz	1.7×10^{-4}	
		(300 ~ 1 000) V 50 Hz	1.3×10^{-4}	
		50 Hz ~ 1 kHz	1.2×10^{-4}	
AC Current		(100 μ A)		
		10 Hz	7.1×10^{-4}	
		10 Hz ~ 1 kHz	5.6×10^{-4}	
		(1 ~ 5) kHz	7.1×10^{-4}	
		(5 ~ 10) kHz	2.3×10^{-3}	
		(100 μ A ~ 1 mA)		
		10 Hz	6.2×10^{-4}	
		10 Hz ~ 1 kHz	5.5×10^{-4}	
		(1 ~ 5) kHz	6.4×10^{-4}	
		(5 ~ 10) kHz	2.1×10^{-3}	
		(1 ~ 10) mA		
		10 Hz	6.2×10^{-4}	
		10 Hz ~ 1 kHz	5.5×10^{-4}	
		(1 ~ 5) kHz	6.2×10^{-4}	
		(5 ~ 10) kHz	3.2×10^{-3}	
		(10 ~ 100) mA		
		10 Hz	6.3×10^{-4}	
		10 Hz ~ 1 kHz	5.4×10^{-4}	
		(1 ~ 5) kHz	6.1×10^{-4}	
		(5 ~ 10) kHz	1.5×10^{-3}	
		(100 mA ~ 1 A)		
		40 Hz ~ 1 kHz	9.8×10^{-4}	
		(1 ~ 5) kHz	1.2×10^{-3}	
		(5 ~ 10) kHz	8.1×10^{-3}	
		(1 ~ 10) A		
		(40 ~ 100) Hz	1.3×10^{-3}	
		100 Hz ~ 1 kHz	1.5×10^{-3}	
		(10 ~ 20) A		
		(40 ~ 100) Hz	1.6×10^{-3}	
		100 Hz ~ 1 kHz	1.7×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311			Power Calibrator, Calibrator/ SICT-CP-40311
AC Current		(30 ~ 40) mA (40 ~ 60) mA (60 ~ 70) mA (70 ~ 100) mA (100 ~ 150) mA (150 ~ 300) mA (300 ~ 800) mA (800 mA ~ 2 A) (2 ~ 3) A (3 ~ 4) A (4 ~ 5) A (5 ~ 6) A (6 ~ 10) A (10 ~ 15) A (15 ~ 20) A (20 ~ 30) A (30 ~ 40) A (40 ~ 500) A (500 ~ 800) A (800 ~ 1 000) A	3.0×10^{-4} 2.7×10^{-4} 2.4×10^{-4} 2.3×10^{-4} 2.7×10^{-4} 2.5×10^{-4} 2.2×10^{-4} 2.1×10^{-4} 3.4×10^{-4} 2.9×10^{-4} 2.6×10^{-4} 3.3×10^{-4} 3.1×10^{-4} 6.2×10^{-4} 5.7×10^{-4} 5.2×10^{-4} 5.0×10^{-4} 1.3×10^{-3} 1.4×10^{-3} 1.3×10^{-3}	
AC Wattage		(50 ~ 60) Hz 24 mW (24 ~ 72) mW (72 ~ 120) mW (120 ~ 480) mW 480 mW ~ 2.4 W (2.4 ~ 24) W (24 ~ 120) W (120 ~ 240) W (240 ~ 480) W (480 ~ 600) W (600 ~ 960) W 960 W ~ 1.2 kW (1.2 ~ 2.4) kW (2.4 ~ 4.8) kW (4.8 ~ 9.6) kW (9.6 ~ 120) kW	0.42 mW 5.8×10^{-3} 3.5×10^{-3} 2.9×10^{-3} 1.1×10^{-3} 4.9×10^{-4} 2.1×10^{-4} 2.3×10^{-4} 2.6×10^{-4} 2.0×10^{-4} 2.4×10^{-4} 2.0×10^{-4} 3.0×10^{-4} 6.4×10^{-4} 5.8×10^{-4} 1.3×10^{-3}	
Frequency		25 Hz (25 ~ 60) Hz (60 ~ 100) Hz (100 ~ 200) Hz (200 ~ 400) Hz (400 ~ 1 000) Hz	9.1 mHz 1.7×10^{-4} 6.0×10^{-4} 3.9×10^{-4} 2.7×10^{-4} 1.3×10^{-3}	
Power Factor		(50 ~ 60) Hz 240 mW -1 ~ 1 -0.8, 0.8 -0.5, 0.5 -0.3, 0.3 -0.1, 0.1 240 mW ~ 120 kW -1 ~ 1 -0.8, 0.8 -0.5, 0.5 -0.3, 0.3 -0.1, 0.1	3.1×10^{-4} 3.9×10^{-4} 4.6×10^{-4} 5.6×10^{-4} 5.7×10^{-4} 2.1×10^{-4} 3.2×10^{-4} 4.6×10^{-4} 4.9×10^{-4} 5.0×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Puncture/safety testers	40313			AC/DC Kilovoltmeter, High Voltage Digital Meter, Reference Multimeter/ SICT-CP-40313
DC Breaking Current		0.1 mA (0.1 ~ 0.5) mA (0.5 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 10) mA (10 ~ 20) mA (20 ~ 50) mA (50 ~ 100) mA	0.61 μ A 1.2×10^{-3} 6.1×10^{-4} 3.1×10^{-4} 1.5×10^{-4} 7.7×10^{-5} 3.2×10^{-4} 1.8×10^{-4} 6.2×10^{-4}	
Resistance		1 m Ω 1 m Ω ~ 10 m Ω 10 m Ω ~ 100 k Ω	0.84 $\mu\Omega$ 7.0×10^{-4} 6.5×10^{-4}	
Insulation Voltage		1 V (1 ~ 10) V (10 ~ 25) V (25 ~ 50) V (50 ~ 100) V (100 ~ 250) V (250 ~ 500) V (500 ~ 1 000) V (1 000 ~ 2 000) V	0.62 mV 1.2×10^{-4} 2.8×10^{-4} 1.7×10^{-4} 1.2×10^{-4} 2.8×10^{-4} 1.7×10^{-4} 1.2×10^{-4} 6.9×10^{-3}	
Insulation Resistance		1 k Ω (1 ~ 100) k Ω 100 k Ω ~ 1 M Ω (1 ~ 10) M Ω (10 ~ 100) M Ω 100 M Ω ~ 1 G Ω (1 ~ 10) G Ω (10 ~ 100) G Ω 100 G Ω ~ 1 T Ω	0.14 Ω 1.3×10^{-4} 1.0×10^{-4} 9.1×10^{-4} 2.0×10^{-4} 2.6×10^{-4} 9.1×10^{-3} 1.5×10^{-3} 2.8×10^{-3}	
Leakage current(DC)		30 μ A (30 ~ 100) μ A 100 μ A ~ 10 mA (10 ~ 50) mA	29 nA 7.3×10^{-4} 6.4×10^{-4} 2.2×10^{-4}	
Leakage current(AC)		60 Hz 30 μ A (30 ~ 100) μ A 100 μ A ~ 1 mA (1 ~ 10) mA (10 ~ 50) mA	0.16 μ A 2.7×10^{-3} 1.5×10^{-3} 9.3×10^{-4} 9.5×10^{-4}	
Timer		1 s (1 ~ 100) s (100 ~ 1 000) s (1 000 ~ 10 000) s	5.8 μ s 5.8×10^{-6} 8.2×10^{-6} 5.8×10^{-5}	
Output AC Current		60 Hz 1 A (1 ~ 5) A (5 ~ 10) A (10 ~ 20) A (20 ~ 30) A (30 ~ 50) A (50 ~ 60) A	2.9 mA 2.1×10^{-3} 2.0×10^{-3} 3.0×10^{-3} 4.0×10^{-3} 3.1×10^{-3} 7.8×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
AC voltmeters AC Voltage	40318	(10 ~ 30) V 50 Hz ~ 1 kHz (1 ~ 10) kHz 50 Hz ~ 100 kHz	 2.0×10^{-4} 2.2×10^{-4} 1.2×10^{-3}	Reference Multimeter, Calibrator/ SICT-CP-40318	
		(30 ~ 100) V 50 Hz ~ 1 kHz (1 ~ 10) kHz 50 Hz ~ 100 kHz	 2.4×10^{-4} 2.5×10^{-4} 2.9×10^{-3}		
		(100 ~ 1 000) V 50 Hz ~ 1 kHz	 3.6×10^{-4}		
AC Output Voltage		(50 Hz ~ 1 kHz) 1 mV (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 1 V	 $5.4 \mu\text{V}$ 7.8×10^{-4} 2.0×10^{-4} 1.1×10^{-4}		
DC Output Voltage		1 mV (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 1 V	 $0.86 \mu\text{V}$ 8.7×10^{-5} 1.4×10^{-5} 9.0×10^{-5}		
Line frequency meters Frequency	40410	10 Hz (10 ~ 60) Hz (60 ~ 400) Hz 400 Hz ~ 1 kHz	 6.1 mHz 1.5×10^{-4} 1.2×10^{-4} 1.4×10^{-4}		Calibrator/ SICT-CP-40410
Leakage current testers DC Current	40416	(\pm) 0 μA (0 ~ 5) μA (5 ~ 10) μA (10 ~ 30) μA (30 ~ 50) μA (50 ~ 100) μA (100 ~ 200) μA (200 ~ 500) μA (0.5 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 10) mA (10 ~ 15) mA (15 ~ 20) mA	 9 nA 1.9×10^{-3} 9.9×10^{-4} 3.7×10^{-4} 2.4×10^{-4} 1.6×10^{-4} 1.3×10^{-4} 8.1×10^{-5} 6.1×10^{-4} 3.2×10^{-4} 1.5×10^{-4} 6.1×10^{-4} 4.1×10^{-4} 3.1×10^{-4}		Calibrator/ SICT-CP-40416
		(30 μA) 50 Hz ~ 1 kHz	 $0.16 \mu\text{A}$		
AC Current		(30 ~ 50) μA 50 Hz ~ 1 kHz	 3.9×10^{-3}		
		(50 ~ 100) μA 50 Hz ~ 1 kHz	 2.7×10^{-3}		

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.		
Leakage current testers	40416	AC Current	(100 ~ 200) μ A 50 Hz ~ 1 kHz	2.1×10^{-3}	Calibrator/ SICT-CP-40416	
			(200 ~ 500) μ A 50 Hz ~ 1 kHz	1.6×10^{-3}		
			(0.5 ~ 1) mA 50 Hz ~ 1 kHz	6.9×10^{-4}		
			(1 ~ 2) mA 50 Hz ~ 1 kHz	4.7×10^{-4}		
			(2 ~ 5) mA 50 Hz ~ 1 kHz	3.0×10^{-3}		
			(5 ~ 10) mA 50 Hz ~ 1 kHz	8.2×10^{-4}		
			(10 ~ 20) mA 50 Hz ~ 1 kHz	5.3×10^{-4}		
			(20 ~ 30) mA 50 Hz ~ 1 kHz	1.2×10^{-4}		
			DC Voltage	0 V		61 μ V
				(0 ~ 0.1) V		6.1×10^{-4}
				(0.1 ~ 0.2) V		3.1×10^{-4}
				(0.2 ~ 0.5) V		1.2×10^{-4}
		(0.5 ~ 1) V		6.1×10^{-4}		
		(1 ~ 2) V		3.1×10^{-4}		
		(2 ~ 5) V		1.2×10^{-4}		
		(5 ~ 10) V		6.2×10^{-5}		
		(10 ~ 20) V		3.3×10^{-5}		
		(20 ~ 50) V		1.7×10^{-5}		
		(50 ~ 100) V		1.2×10^{-5}		
		(100 ~ 200) V		3.4×10^{-5}		
		(200 ~ 300) V	2.5×10^{-5}			
		(300 ~ 500) V	1.2×10^{-4}			
		(500 ~ 1 000) V	6.2×10^{-5}			
		AC Voltage	(0.1 V)			
			50 Hz ~ 1 kHz	6.5×10^{-4}		
			(1 ~ 20) kHz	6.6×10^{-4}		
			(20 ~ 50) kHz	7.8×10^{-4}		
			(50 ~ 100) kHz	1.4×10^{-3}		
			(0.1 ~ 1) V			
			50 Hz ~ 1 kHz	1.2×10^{-4}		
			(1 ~ 20) kHz	1.3×10^{-4}		
			(20 ~ 50) kHz	1.8×10^{-4}		
			(50 ~ 100) kHz	3.8×10^{-4}		

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers AC Voltage	40416	(1 ~ 10) V 50 Hz ~ 1 kHz (1 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (10 ~ 100) V 50 Hz ~ 1 kHz (1 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) V 50 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 20) kHz (300 ~ 1 000) V 50 Hz ~ 10 kHz	 1.2×10^{-4} 1.3×10^{-4} 1.8×10^{-4} 3.4×10^{-4} 1.3×10^{-4} 1.2×10^{-4} 3.0×10^{-4} 6.8×10^{-4} 3.1×10^{-4} 3.4×10^{-4} 5.5×10^{-4} 3.7×10^{-4}	Calibrator/ SICT-CP-40416
Output Resistance		100 mΩ 100 mΩ ~ 1 Ω 1 Ω ~ 10 kΩ	8.6 μΩ 6.3×10^{-5} 6.2×10^{-5}	
Input Resistance		100 mΩ 100 mΩ ~ 100 Ω 100 Ω ~ 10 kΩ	24 μΩ 2.4×10^{-4} 6.3×10^{-5}	
Electronic AC/DC loads DC Voltage	40417	100 mV (0.1 ~ 1) V (1 ~ 2) V (2 ~ 4) V (4 ~ 6) V (6 ~ 8) V (8 ~ 10) V (10 ~ 50) V (50 ~ 100) V (100 ~ 200) V (200 ~ 400) V (400 ~ 1 000) V	6.3 μV 6.2×10^{-5} 3.2×10^{-5} 2.3×10^{-5} 1.5×10^{-5} 1.2×10^{-5} 1.1×10^{-5} 3.3×10^{-5} 1.2×10^{-5} 3.6×10^{-5} 2.6×10^{-5} 1.7×10^{-5}	Calibrator/ SICT-CP-40417
DC Current		100 mA 100 mA ~ 0.2 A (0.2 ~ 0.4) A (0.4 ~ 0.6) A (0.6 ~ 2) A (2 ~ 20) A (20 ~ 80) A (80 ~ 100) A	44 μA 4.8×10^{-4} 4.2×10^{-4} 3.8×10^{-4} 3.7×10^{-4} 8.1×10^{-4} 4.2×10^{-4} 4.1×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electronic AC/DC loads AC Voltage	40417	(50 ~ 60) Hz 0.1 V (0.1 ~ 0.2) V (0.2 ~ 0.5) V (0.5 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 200) V (200 ~ 500) V	0.61 mV 3.1×10^{-3} 1.2×10^{-3} 6.2×10^{-4} 3.5×10^{-4} 2.2×10^{-4} 1.9×10^{-4} 2.0×10^{-4}	Calibrator/ SICT-CP-40417
AC Current		(50 ~ 60) Hz 100 mA 100 mA ~ 0.3 A (0.3 ~ 0.4) A (0.4 ~ 0.5) A (0.5 ~ 0.6) A (0.6 ~ 0.8) A (0.8 ~ 2) A (2 ~ 4) A (4 ~ 5) A (5 ~ 7) A (7 ~ 9) A (9 ~ 10) A (10 ~ 13) A (13 ~ 16) A (16 ~ 20) A	2.5 mA 3.9×10^{-3} 3.2×10^{-3} 2.7×10^{-3} 3.5×10^{-3} 3.1×10^{-3} 2.7×10^{-3} 3.5×10^{-3} 2.6×10^{-3} 3.5×10^{-3} 2.9×10^{-3} 2.5×10^{-3} 4.5×10^{-3} 3.7×10^{-3} 3.2×10^{-3}	
Analogue/Digital multimeters DC Voltage	40419	(±) 0 mV (0 ~ 1) mV (1 ~ 2) mV (2 ~ 5) mV (5 ~ 10) mV (10 ~ 20) mV (20 ~ 50) mV (50 ~ 100) mV 100 mV ~ 0.2 V (0.2 ~ 0.5) V (0.5 ~ 1) V (1 ~ 2) V (2 ~ 5) V (5 ~ 10) V (10 ~ 20) V (20 ~ 50) V (50 ~ 100) V (100 ~ 200) V (200 ~ 500) V (500 ~ 1 000) V	0.7 μV 7.7×10^{-4} 3.9×10^{-4} 1.6×10^{-4} 8.4×10^{-5} 4.6×10^{-5} 2.4×10^{-5} 1.6×10^{-5} 1.7×10^{-5} 1.1×10^{-5} 9.4×10^{-6} 1.2×10^{-5} 9.3×10^{-6} 8.6×10^{-6} 1.6×10^{-5} 1.2×10^{-5} 1.0×10^{-5} 1.8×10^{-5} 1.2×10^{-5} 1.1×10^{-5}	Calibrator/ SICT-CP-40419

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Analogue/Digital multimeters	40419	AC Voltage	(3 mV) 50 Hz ~ 1 kHz	10 μV	Calibrator/ SICT-CP-40419
			(3 ~ 10) mV 50 Hz ~ 10 kHz (10 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz	9.0×10^{-4} 1.3×10^{-3} 1.3×10^{-3} 2.1×10^{-3}	
			(10 ~ 100) mV 50 Hz ~ 10 kHz (10 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz	2.5×10^{-4} 2.7×10^{-4} 5.4×10^{-4} 1.4×10^{-3}	
			(0.1 ~ 1) V 50 Hz ~ 10 kHz (10 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz	1.1×10^{-4} 1.3×10^{-4} 2.0×10^{-4} 4.2×10^{-4}	
			(1 ~ 10) V 50 Hz ~ 10 kHz (10 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz	1.1×10^{-4} 1.3×10^{-4} 2.0×10^{-4} 3.8×10^{-4}	
			(10 ~ 100) V 50 Hz ~ 20 kHz (20 ~ 50) kHz (50 ~ 100) kHz	1.2×10^{-4} 3.4×10^{-4} 8.1×10^{-4}	
			(100 ~ 300) V 50 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 20) kHz	2.3×10^{-4} 2.7×10^{-4} 5.1×10^{-4}	
			(300 ~ 1 000) V 50 Hz ~ 10 kHz	3.6×10^{-4}	
		Resistance	0 Ω	85 μΩ	
			(0 ~ 1) Ω	1.3×10^{-4}	
			(1 ~ 10) Ω	3.4×10^{-5}	
			(10 ~ 100) Ω	2.0×10^{-5}	
			100 Ω ~ 1 kΩ	1.6×10^{-5}	
			(1 ~ 10) kΩ	1.4×10^{-5}	
			(10 ~ 100) kΩ	1.7×10^{-5}	
			100 kΩ ~ 1 MΩ	2.4×10^{-5}	
			(1 ~ 10) MΩ	5.0×10^{-5}	
			(10 ~ 100) MΩ	1.3×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Analogue/Digital multimeters	40419	DC Current	(±) 10 µA (10 ~ 20) µA (20 ~ 50) µA (50 ~ 100) µA (0.1 ~ 0.2) mA (0.2 ~ 0.5) mA (0.5 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 10) mA (10 ~ 20) mA (20 ~ 50) mA (50 ~ 100) mA (0.1 ~ 0.2) A (0.2 ~ 0.5) A (0.5 ~ 1) A (1 ~ 2) A (2 ~ 5) A (5 ~ 10) A (10 ~ 20) A	10 nA 5.2×10^{-4} 2.5×10^{-4} 1.5×10^{-4} 1.0×10^{-4} 8.1×10^{-5} 6.8×10^{-5} 9.0×10^{-5} 8.1×10^{-5} 6.8×10^{-5} 9.0×10^{-5} 9.2×10^{-5} 8.0×10^{-5} 9.8×10^{-5} 1.5×10^{-4} 1.2×10^{-4} 4.6×10^{-4} 7.4×10^{-4} 6.5×10^{-4} 1.2×10^{-3}	Calibrator/ SICT-CP-40419
		AC Current	(30 µA) 50 Hz ~ 1 kHz (30 ~ 50) µA 50 Hz ~ 1 kHz (50 ~ 100) µA 50 Hz ~ 1 kHz (0.1 ~ 0.2) mA 50 Hz ~ 1 kHz (0.2 ~ 0.5) mA 50 Hz ~ 1 kHz (0.5 ~ 1) mA 50 Hz ~ 1 kHz (1 ~ 2) mA 50 Hz ~ 1 kHz (2 ~ 5) mA 50 Hz ~ 1 kHz (5 ~ 10) mA 50 Hz ~ 1 kHz (10 ~ 20) mA 50 Hz ~ 1 kHz	0.16 µA 3.9×10^{-3} 2.7×10^{-3} 3.6×10^{-5} 4.9×10^{-5} 2.4×10^{-4} 3.0×10^{-4} 1.4×10^{-4} 7.2×10^{-5} 3.0×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters AC Current	40419	(20 ~ 50) mA 50 Hz ~ 1 kHz	1.4×10^{-4}	Calibrator/ SICT-CP-40419
		(50 ~ 100) mA 50 Hz ~ 1 kHz	7.2×10^{-5}	
		100 mA ~ 0.2 A 50 Hz ~ 1 kHz	4.5×10^{-4}	
		(0.2 ~ 0.5) A 50 Hz ~ 1 kHz	2.6×10^{-4}	
		(0.5 ~ 1) A 50 Hz ~ 1 kHz	1.3×10^{-4}	
		(1 ~ 2) A 50 Hz ~ 1 kHz	7.6×10^{-4}	
		(2 ~ 5) A 50 Hz ~ 1 kHz	1.7×10^{-3}	
		(5 ~ 10) A 50 Hz ~ 1 kHz	1.4×10^{-3}	
		(10 ~ 15) A 50 Hz ~ 1 kHz	2.1×10^{-3}	
		(15 ~ 20) A 50 Hz ~ 1 kHz	2.0×10^{-3}	
Frequency		10 Hz ~ 1 MHz	6.1×10^{-7}	
Oscilloscopes Impedance Measure DC Voltage	40421	50 Ω	0.7 mΩ	Calibration Generator/ SICT-CP-40421
		75 Ω	0.8 mΩ	
		1 MΩ	13 Ω	
		1 mV	1.0 μV	
		(1 ~ 2) mV	4.9×10^{-4}	
		(2 ~ 5) mV	2.0×10^{-4}	
		(5 ~ 10) mV	1.0×10^{-4}	
		(10 ~ 20) mV	5.5×10^{-5}	
		(20 ~ 50) mV	2.7×10^{-5}	
		(50 ~ 100) mV	1.8×10^{-5}	
		100 mV ~ 0.5 V	1.6×10^{-5}	
		(0.5 ~ 1) V	9.4×10^{-6}	
		(1 ~ 2) V	1.1×10^{-5}	
		(2 ~ 5) V	9.2×10^{-6}	
		(5 ~ 10) V	8.6×10^{-6}	
(10 ~ 50) V	1.2×10^{-5}			
(50 ~ 200) V	1.5×10^{-5}			

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Oscilloscopes AC Voltage(Square wave)	40421	(1 kHz) 1 mV (1 ~ 2) mV (2 ~ 5) mV (5 ~ 10) mV (10 ~ 20) mV (20 ~ 50) mV (50 ~ 100) mV (100 ~ 200) mV (0.2 ~ 0.5) V (0.5 ~ 120) V	47 μV 2.4×10^{-2} 1.0×10^{-2} 5.8×10^{-3} 3.5×10^{-3} 2.1×10^{-3} 1.6×10^{-3} 1.4×10^{-3} 1.3×10^{-3} 1.2×10^{-3}	Calibration Generator/ SICT-CP-40421
Time Marker		5 ns 5 ns ~ 10 ms (10 ~ 20) ms 20 ms ~ 5 s	16 fs 3.0×10^{-6} 3.3×10^{-6} 3.3×10^{-5}	
CAL Output Amplitude		(50 Hz ~ 10 kHz) 100 mV (100 ~ 200) mV 200 mV ~ 1.2 V (1.2 ~ 2) V (2 ~ 4) V (4 ~ 12) V	36 μV 2.2×10^{-4} 1.7×10^{-4} 1.4×10^{-4} 1.2×10^{-4} 1.7×10^{-4}	
CAL Output Frequency		100 Hz ~ 10 MHz	6.5×10^{-7}	
Sine Wave Signal Generator Level		(600 mV) 50 kHz (0.05 ~ 100) MHz (100 ~ 500) MHz (0.5 ~ 1) GHz (1 ~ 3) GHz	25 mV 4.5×10^{-2} 7.2×10^{-2} 1.9×10^{-2} 2.3×10^{-2}	
Volt/Current recorders DC Voltage	40424	(±) 0 μV 0 μV ~ 10 mV (10 ~ 100) mV 100 mV ~ 1 V (1 ~ 10) V (10 ~ 100) V (100 ~ 500) V (500 ~ 1 000) V	0.70 μV 8.1×10^{-5} 1.6×10^{-5} 9.4×10^{-6} 8.6×10^{-6} 1.0×10^{-5} 1.2×10^{-5} 1.1×10^{-5}	Calibrator/ SICT-CP-40424
DC Current		(±) 0 μA (0 ~ 10) μA (10 ~ 100) μA 100 μA ~ 10 mA 10 mA ~ 1 A (1 ~ 10) A (10 ~ 20) A	9.3 nA 9.9×10^{-4} 1.6×10^{-4} 9.2×10^{-5} 2.0×10^{-4} 6.5×10^{-4} 1.2×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Temperature generators: ovens, furnaces, isothermal liquid baths, ice-point baths, dry-block calibrators	50101	0 ℃ (-90 ~ 250) ℃ (250 ~ 500) ℃ (500 ~ 650) ℃ (650 ~ 700) ℃ (700 ~ 1 300) ℃	0.01 ℃ 0.020 ℃ 0.024 ℃ 0.028 ℃ 1.3 ℃ 2.6 ℃	SPRT, STANDARD TC/ SICT-CP-50101
Temperature indicators/recorders /controllers, temperature calibrators Temperature indicators/recorders/controllers (With Sensor)	50102	(-45 ~ 0) ℃ (0 ~ 250) ℃ (250 ~ 650) ℃ (650 ~ 900) ℃ (900 ~ 1 100) ℃ (1 100 ~ 1 300) ℃	0.024 ℃ 0.031 ℃ 0.13 ℃ 1.4 ℃ 1.5 ℃ 2.6 ℃	SPRT, STANDARD TC/ SICT-CP-50102
(Without Sensor)		(-45 ~ 0) ℃ (0 ~ 100) ℃ (100 ~ 200) ℃ (200 ~ 300) ℃ (300 ~ 400) ℃ (400 ~ 500) ℃ (500 ~ 600) ℃ (600 ~ 700) ℃ (700 ~ 800) ℃ (800 ~ 1 100) ℃ (1 100 ~ 1 300) ℃	0.013 ℃ 0.018 ℃ 0.022 ℃ 0.027 ℃ 0.031 ℃ 0.035 ℃ 0.048 ℃ 0.052 ℃ 0.057 ℃ 0.08 ℃ 0.09 ℃	
Glass thermometers; liquid-in-glass, Beckmann liquid-in-glass	50103	(-45 ~ 0) ℃ (0 ~ 100) ℃ (100 ~ 200) ℃	0.048 ℃ 0.058 ℃ 0.062 ℃	SPRT/ SICT-CP-50103
Resistance thermometers; SPRT, IPRT, thermistors, etc. IPRT	50104	(-45 ~ 50) ℃ (50 ~ 250) ℃	0.024 ℃ 0.028 ℃	SPRT, Fixed point/ SICT-CP-50104
Thermal expansion thermometers; bimetal, gas or liquid type bimetal	50105	(-45 ~ 100) ℃ (100 ~ 300) ℃ (300 ~ 400) ℃ (400 ~ 650) ℃	0.4 ℃ 0.6 ℃ 1.5 ℃ 3.2 ℃	SPRT/ SICT-CP-50105

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Thermocouples: noble metal, base metal, pure metal, special type, etc. Base metal	50106	(-45 ~ 300) °C (300 ~ 400) °C (400 ~ 500) °C (500 ~ 650) °C (650 ~ 1 000) °C (1 000 ~ 1 100) °C (1 100 ~ 1 200) °C (1 200 ~ 1 300) °C	0.4 °C 0.6 °C 0.7 °C 0.8 °C 1.8 °C 2.0 °C 3.1 °C 3.6 °C	SPRT, Fixed point, STANDARD TC/ SICT-CP-50106
Temperature transducers	50107	(-45 ~ 100) °C (100 ~ 200) °C (200 ~ 250) °C (250 ~ 300) °C (300 ~ 500) °C (500 ~ 650) °C (650 ~ 700) °C (700 ~ 800) °C (800 ~ 1 000) °C (1 000 ~ 1 100) °C (1 100 ~ 1 300) °C	0.05 °C 0.06 °C 0.09 °C 0.14 °C 0.16 °C 0.18 °C 1.4 °C 1.6 °C 1.7 °C 2.8 °C 3.0 °C	SPRT, THERMOCOUPLE, MULTIMETER SICT-CP-50107

