



Organisme belge d'Accréditation
Belgische Accreditatieinstelling
Belgische Akkreditierungsstelle
Belgian Accreditation Body

EA MLA Signatory

Bijlage bij accreditatie-certificaat
Annexe au certificat d'accréditation
Annex to the accreditation certificate
Beilage zur Akkreditierungszertifikat

001-CAL

EN ISO/IEC 17025:2017

Versie / Version / Version / Fassung	20
Geldigheidsperiode / Validité / Validity / Gültigkeitsdauer	2022-01-13 - 2026-06-01

Maureen Logghe

Voorzitster van het Accreditatiebureau
La Présidente du Bureau d'Accréditation
Chair of the Accreditation Board
Vorsitzende des Akkreditierungsbüro

De accreditatie werd uitgereikt aan / L'accréditation est délivrée à /
The accreditation is granted to / Die akkreditierung wurde erteilt für:

TRESCAL nv
Vosstraat, 200
2600 Antwerpen

Activiteitencentra / Sites d'activités / Sites of activities / Standorte mit aktivitäten:

Locatie 1: BERCHEM	Vosstraat 200 2600 Antwerpen
Locatie 2: WELLIN	Rue Jean Meunier, 2 6922 Wellin
Locatie 3: LOUVAIN-LA-NEUVE	Rue du Bosquet, 7 1348 Ottignies-Louvain-la-Neuve

Volume Berchem

Calibration and Measurement Capabilities

Volume

Measure quantity, instrument or gauge	Range	expanded uncertainty (*)	Remarks	Calibration procedure
Volume measuring devices, pipettes	10 µl to 50 µl	0,40 µl	Gravimetric method	KI/02/KC/V.01
	50 µl to 100 µl	0,50 µl		
	100 µl to 500 µl	0,60 µl		
	500 µl to 1 ml	1,0 µl		
	1 ml to 10 ml	10 µl		
	10 ml to 25 ml	20 µl		
	25 ml to 50 ml	50 µl		
Volume measuring devices, glass cups, recipients, ...	50 ml to 100 ml	1,0 ml	Gravimetric method	KI/02/KC/V.51
	100 ml to 200 ml	1,5 ml		
	200 ml to 300 ml	2,0 ml		
	300 ml to 500 ml	2,5 ml		
	500 ml to 1000 ml	3,0 ml		
	1 l to 5 l	3,5 ml		

(*) the smallest uncertainty of measurement the laboratory can provide to its customers, expressed as the expanded uncertainty having a coverage probability of approximately 95%.