

Korea Laboratory Accreditation Scheme

CERTIFICATE OF ACCREDITATION

Institute of Calibration & Technology Co., Ltd.

Accreditation No. : KC01-018

Corporation Registration No. : 130111-0052843

Address of Laboratory : 1. 72, Dontancheomdansaneop 2-ro, Hwaseong-si, Gyeonggi-do,
Republic of Korea
1-① 47-22, Jingok Sandan 2beon-ro Gwangsan-gu
Gwangju, Republic of Korea

Date of Initial Accreditation : Feb. 23, 2001.

Validity of Accreditation : Oct. 30, 2021 ~ Oct. 29, 2025

Scope of Accreditation : Attached Annex

Date of issue : Sep. 15, 2021.

This Calibration laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025 : 2017. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to Joint ISO-ILAC-IAF Communiqué).



Head

Korea Laboratory Accreditation Scheme

Sanghoon Lee

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017 & KS Q ISO/IEC 17025:2017

Institute of Calibration Technology Co.,Ltd

72, Dongtancheomdansaneop 2-ro, Hwaseong-si, Gyeonggi-do, Republic of Korea
Phone : 82-31-379-5114, Fax : 82-31-379-5115 e-mail : sictadmin@sict.co.kr

CALIBRATION

Valid To : Oct. 29. 2025

Accreditation No : KC01-018(1/208)

In recognition of the successful completion of the KOLAS evaluation process,
accreditation is granted to this laboratory to perform the following calibrations

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site
102.	Linear dimension		10236	Coating thickness testers	Y	10413	Straight rules	N
10201	Balls	N	10237	Torque arms	N	10415	Test bars	N
10203	Electrical /Mechanical comparators	Y	103. Angle			105.	Complex geometry	
			10302	Angle gauge blocks	N	10501	Base gauges for electric bulb	N
10206	Dial/cylinder gauge testers	N	10303	Autocollimators	N	10502	Baench centers	Y
10207	Doctor blades	N	10304	Bevel protractors	Y	10503	Contact coordinate measuring machines	Y
10208	Distance meters;electrooptic/laser/ultrasonic	N	10306	Clinometers	N	10504	Non-contact coordinate measuringmachines	Y
10209	End bars	N	10308	Fine angle generators, level comparators	N			
10210	Extensometers, linear displacement transducers	Y	10310	Indexing tables	N	10505	Gauge block accessories	N
			10311	Plate/Square/Electric levels	N	10511	Measuringmicroscopes, Profileprojectors	Y
10211	Filler gauges	Y	10314	Penta-prisms	N			
10212	Film applicators	N	10315	Polygons	N	10512	Microscopes, micro measuring	Y
10213	Gapgauges	N	10316	Rotary tables	Y	10514	Taper plug gauges	N
10214	Gauge blocks, by comparison	N	10317	Sinebars,Plates,Tables, Centers	N	10515	Taper ring gauges	N
10216	Height gauges/measuring machines	Y	10318	Squareness testers, Right angle testers	Y	10517	Stylus type roughness testers	Y
10219	Linear scales	N				10518	Socket gauges for electric bulb	N
10220	Measuring machines, standard	Y	10319	Cylindrical squares	N			
10221	Micro scales/Standard scales	N	10320	Precisionsquares	N	10519	Roughness standard /comparison specimens	N
10223	Electronic micrometers	Y	10322	Angular displacement transducers	N	10525	Thread plug gauges	N
10224	Heightmicrometers, Riserblocks	N	104. Form			10526	Taper thread plug gauges	N
10225	Laser scan micrometers	Y	10401	Form testers	Y	10527	Thread ring gauges	N
10227	Standardtaperules, Peripheralgauges	N	10404	Optical flats	N	10528	Taper threadring gauges	N
			10405	Optical parallels	N	10529	V-blocks,Boxblocks	N
10228	Cylindricalplug/pingauges, Threadmeasuringwiregauges	Y	10406	Parallel blocks	Y	106.	Various dimensional	
			10407	Precision surface plates	Y	10601	Inside/Outside/Geartooth calipers,Calipergauges	Y
10229	Radius gauges	N	10408	Profile gauges	N			
10230	Cylindrical ring gauges	N	10409	Roundness measurement instruments	Y	10603	Cylinder/bore gauges	Y
10232	Step gauges	N				10604	Depthgauges,Depthmicrometers	Y
10233	Thickness gauges, taper	N	10411	Roundness standard/ Roundness magnification standard specimens	N	10605	Dial/digital gauges	Y
10234	Ultrasonic thickness gauges	Y				10608	Grind gauges	N
10235	Ultrasonic/coating thickness specimens	N	10412	Straight edges	Y	10609	Microindicators, Testindicators	Y

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site			
10610	Micrometer heads	Y	20503	Ionization gauges	N	20916	Gas flowmeters;turbine	N			
10611	3-points, Micrometers	Y				20917	Liquid flowmeters;turbine	N			
10612	Inside micrometers	Y	20504	Thermal conductivity gauge; Pirani, thermocouple, convectron etc.	N	20918	Gas flowmeters;ultrasonic	N			
10613	Outside micrometers	Y				20919	Liquid flowmeters;ultrasonic	N			
10615	Particle counters	N				20920	Gas flowmeters;variable area	N			
10617	Standard sieves	N	20505	Standard leaks, Helium leak dectors	Y	20921	Liquid flowmeters;variable area	N			
10620	Welding gauges	N				20922		N			
201. Mass			206. Volume			20922	Gas flowmeters;vortex	N			
20102	Auto-hopper scale balances	Y	20601	Volumetric glasswares	N	20923	Liquid flowmeters;vortex	N			
20103	Auto-packer scale balances	Y	20602	Pycnometers	N	20925	Anemometers; vane, etc	N			
20104	Axle weigher balances	N	20604	Standard volume vessels	Y	20999	Others ; Anemometers; ultrasonic waves	N			
20106	Dial platform scale balances	Y	20605	Concrete air content meters	N						
20107	Dial swing scale balances	Y	20606	Piston type volume meters	N	210. Hardness					
20109	Electric balances	Y	207. Density			21001	Brinell hardness testers	Y			
20111	Manual swing scale balances	Y	20702	Liquid density meters	N	21002	Rockwell hardness testers	Y			
20112	Platform scale balances	Y	20704	Salinity meters	N	21003	Shore hardness testers	Y			
20113	Spring scale balances	Y	20705	Sucrose meters	N	21004	Vickers hardness testers	Y			
20116	Weights	Y	20706	Hydrometers; density, specific gravity, alcohol, API, baume, sugar, milk, soil salinity, LPG, etc.	N	21005	Durometer hardness testers	N			
202. Force						21006	Leeb hardness testers	Y			
20202	Force measuring devices	N				301. Time/frequency					
20203	Tension/compression testing machines	Y	20707	Chloride meters	N	30102	Frequency standards	N			
20204	push-pull gauge	Y	20799	Others ; Solid density	N	30103	General frequency sources	Y			
203. Torque			208. Viscosity			30104	Frequency meters/counters	Y			
20302	Torque measuring devices	Y	20801	Kinematic viscometers; capillary, etc.	N	30105	Time interval sources	Y			
20303	Torque wrenches/drivers	Y				30106	Time interval meters /Stop watches/Timers	Y			
204. Pressure			20802	Dynamic viscometers; rotaional, etc	Y	302. Velocity & revolution					
20401	Altimeters	Y				30201	Standard RPM generators	Y			
20402	Manometers	Y	209. Fluid flow			30202	Contact type tachometers	Y			
20403	Pneumatic pressure ballances	N	20901	Anemometers; hot-wire	N	30203	Photo tachometers /stroboscopes	Y			
20404	Hydraulic pressure ballances	N	20902	Anemometers; pitot tube, etc.	N						
20405	Air data test systems	Y	20908	Gas flowmeters;differential pressure	N	30204	Speed meters	Y			
20406	Absolute pressure gauges	Y				30205	Wow-flutter generators	N			
20407	Blood pressure gauges	Y	20909	Liquid flowmeters;differentia pressure	N	30206	Wow-flutter meters	Y			
20408	Compound pressure gauges	Y				401. DC Voltage & current					
20409	Differential pressure gauges	Y	20910	Liquid flowmeters; electromagnetic	N	40101	DC ammeters	Y			
20411	Gauge pressure gauges	Y				40102	Transconductance amplifiers	Y			
20412	Pressure transducers/transmitters	Y	20911	Gas flowmeters;thermal mass, etc.	N	40103	DC voltage/current calibrators	Y			
20413	Dial type vacuum gauges	Y	20912	Liquid flowmeters; Coriolis, etc.	N	40104	Electrical temperature calibrators	Y			
205. Vacuum			20914	Gas flowmeters;open channel, etc.	N	40105	DC current shunts	Y			
20501	Capacitance diaphragm gauges	N				40106	Galvanometers /null detectors	Y			
20502	Spinning rotor gauges	N	20915	Liquid flowmeters; positive displacement	N	40107					

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site
40108	DC power supplies	Y	40316	Current transformer	N	405.	Low frequency electric & magnetic fields	
40110	DC voltage dividers	N	40318	AC voltmeters	Y			
40111	DC voltage standards	N	40319	Watt hour meters	N	40503	Flux meters	Y
40112	DC voltmeters	Y	40321	Ratio transformers	N	40504	Flux sources	N
40113	Static/Ionic voltmeters	N	404. Other DC & LF Measurements			40508	Magnetometers	Y
402. Resistance, Capacitance and Inductance			40401	LF amplifiers	Y	40510	Reference/standard magnets	N
			40402	DC/LF attenuators	Y	406. Radio frequency measurements		
40201	Capacitance bridges /indicators	Y	40403	Multimeter calibrators	N	40601	RF amplifiers	Y
			40404	Oscilloscope calibrators	N	40602	Coaxial attenuators	Y
40202	Decade capacitors	Y	40405	CD/DVD meters/analyzers	Y	40605	Burst pulse generators	Y
40204	Standard capacitors	Y	40406	Video signal generators	Y	40606	Attenuator calibrators	N
40205	Earth testers	Y	40407	Audio distortion analyzers /meters	Y	40607	RF power meter calibrators	Y
40206	Inductance bridges /indicators	Y	40408	LF filters		40608	EMC transducers ; current probes, absorbing absorbing clamps, etc.	Y
			40409	LF/Audiosignalanalyzers	Y			
40210	Insulation testers	Y	40410	Line frequency meters	Y	40610	Coaxial directional couplers /splitters	Y
40211	Q-meters	Y	40411	Function generators	Y			
40213	Resistance bridges & similar instruments	Y	40412	Genescopes	Y	40613	Electrostatic discharge generators	Y
			40413	AC/DC high voltages volt meters	Y	40614	EMC receivers	
40214	Resistance meters	Y	40415	Jitter meters	Y	40615	RF filters	Y
40215	Resistors	Y	40416	Leakage current testers	Y	40616	RF impedance meters	N
40217	Impedance bridges/LCR meters	Y	40417	Electronic AC/DC loads	Y	40617	RF impulse generators	Y
40301	AC ammeters	Y	40418	Modulation meters	Y	40618	Line impedance stabilization networks ; LISN, CDN, ISN, etc.	Y
40302	Clamp ammeters/voltmeters	Y	40419	Analogue/Digital multimeters	Y			
40303	AC voltage/current calibrators	Y	40420	Noise meters	Y			
			40421	Oscilloscopes	Y	40619	Coaxial standard mismatches	Y
40304	Wattmeter calibrators	N	40422	LF phase meters	Y	40621	Mobile communication test sets	Y
40305	AC current shunts	Y	40424	Volt/Current recorders	Y			
40306	Phase angle generators, synchro resolve generators	N	40425	Relay test sets	Y	40622	Modulation meters	Y
			40426	LF signal generators	Y	40623	Network analyzers	Y
40307	Voltage/current phase angle meters/synchro resolve meters	N	40427	LF spectrum analyzers	Y	40624	Noise figure meters	Y
			40428	Spot generators	Y	40625	Noise generators	N
40308	Potential transformer test sets	Y	40429	Sweep generators	Y	40626	Noise impulse simulators	Y
			40430	Signal transducers	Y	40627	RF phase noise meters	N
40309	Potential transformer	N	40432	Transistor curve tracers	Y	40628	Coaxial noise sources	N
40310	Power factor meters	Y	40434	AC/DC high voltage generators	Y	40635	RF power meters	Y
40311	AC power meters	Y	40435	AC/DC high voltage probes	Y	40636	Diode power sensors	Y
40312	AC power supplies	Y	40436	Logic analyzers	Y	40637	Thermocouple power sensors	Y
40313	Puncture/safety testers	Y	40437	Telephone testers	Y	40638	Pulse generators	Y
40314	Power recorders	Y	40438	Video signal analyzers	Y	40639	Radar test sets	Y
40315	Current transformer test sets	Y	40438	Video signal analyzers	Y	40640	RF signal generators	Y
						40641	RF spectrum analyzers	Y
						40643	Surge generators	Y

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site
40644	SWR meters	N	50303	Psychrometers; assmann ventilated, PRT type, etc.	N	70213	Display color analyzers; luminance, chromaticity, white balance, etc.	Y
40645	RF terminations	Y						
40646	Coaxial thermistor mounts	Y						
40650	RF voltmeters	Y	50304	Temperature humidity recorders ; Hygrothermograph, etc	N	70214	Luminous intensity standard lamps	N
40651	Vector voltmeters	Y						
40652	Field strength meters	Y						
40653	AM/FM test sources	Y	50305	Transducers; dew-point /relative humidity	N	70215	Spectral irradiance standard lamps	N
40654	Dip simulators	Y						
407. Field strength & antennas			50306	Humidity generators; two-pressure, two-temperature, flow mixing humidity gererator, constant temperature and humidity chamber, etc.	Y	70216	Total spectral radiand flux standard lampa	N
40704	Loop antennas	N				70217	Luminance standard sources	N
40705	Monopole antennas	N				70218	Spectral radiance standard lamps	N
501. Contact thermometry						70219	UV irradiance meters	N
50101	Temperature generators: ovens, furnaces, isothermal liquid baths,ice-point baths, dry-block calibrators	Y				70220	Spectral irradiance meters	Y
50102	Temperature indicators /recorders/controllers, temperature calibrators	Y		504. Moisture		70221	Total spectral radiant flux meters	Y
			50401	Cereal moisture meters	Y			
			50402	Wood moisture meters	N	70222	Spectral radiance meters	Y
			50403	Paper moisture meters	N		703. Properties of materials	
				601. Sound in air		70301	Colorimeters; material color	Y
50103	Glass thermometers; liquid-in-glass, Beckmann	N	60102	Sound calibrator	N	70304	Color standard tiles	N
			60104	Microphones	N	70306	Gloss meters	Y
50104	Resistance thermometers; SPRT, IPRT, thermistors, etc.	Y	60106	Sound level meters	Y	70307	Gloss standard plates	Y
				603. Vibration		70308	Haze meters	Y
50105	Thermal expansion thermometers ; bimetal, gas or liquid type	Y	60301	Vibration calibrators	N	70312	Lens testers	Y
			60302	Vibration transducers	N	70315	Optical densitometers	Y
			60303	Vibration measuring instruments	N	70319	Reflectance meters	Y
50106	Thermocouples: noble metal, base metal, pure metal, special type, etc.	Y				70321	Refractometers	Y
				701. Photometry		70323	Transmittance meters	Y
50107	Temperature transducers	Y	70101	Illuminance meters	N			
50108	Primary fixed-point cells and apparatus	N	70102	Luminance meters	N			
502. non contact thermometry			70103	Total luminous flux meters	Y			
			70104	Luminous intensity meters	Y			
				702. Properties of detectos & sources				
50203	Optical pyrometers	N						
50204	Standard radiation	N	70202	Color temperature meters	Y			
50205	Thermal image apparatus	N	70203	Color temperature standard lamps	N			
50206	Blackbody furnaces	Y						
50207	Others ; ear thermometers, etc.	N	70204	Colorimeters; source color	Y			
			70207	Laser power meters	N			
503. Humidity			70208	Standard LED light sources	N			
50301	Dew-point hygrometers; chilled mirror, alumina thinfilm, etc.	N	70209	Total luminous flux standard lamps	N			
			70210	Optical detectors	N			
50302	Relative humidity hygrometers polimer thinfilm, hair, etc.	Y	70211	Pyranometers and pyrheliometers	N			

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site
704. Fiber optics			901. Chemical analysis					
70402	Broadband light sources	Y	90101	Breath alcohol analyzer	N			
70410	Optical attenuators	Y	90102	Environmental air monitoring instruments	Y			
70412	Fiber-optic power meters	Y	90103	Gas analyzers	Y			
70413	Optical loss testers	Y	90104	Exhaust gas test instruments	Y			
70415	Optical multimeters	Y	90199	Others ; pH meter, Electrical conductivity meter	Y			
70416	Optical network analyzers	Y						
70417	Optical spectrum analyzers	Y						
70418	Optical time domain reflectometers, OTDR	Y						
70423	Return loss meters	Y						
70429	Frequency stabilized lasers and LDs	Y						
70430	ASE light sources	Y						
70433	Optical power stabilized lasers and LDs	Y						

Note

1. This laboratory provides calibration services in permanent standard laboratory and at on-site.
2. Laboratory conducts on-site calibration should meet requirements of KOLAS-SR-007.
3. On-site calibration is allowed to items with marking 'Y', not allowed to items with marking 'N'.
4. Measurement uncertainty normally is quoted as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k=2$. It expresses the lowest uncertainty of measurement that can be provided by accredited calibration laboratories in normal conditions.
5. Due to the calibration environment such as reference standards or customers' facilities, it is note that uncertainty of measurement on a calibration certificate may be expressed larger than measurement uncertainty on scope of accreditation in general.

102. Linear dimension

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Balls	10201	(0.3 ~ 100) mm	$\sqrt{0.38^2 + (0.0046 \times l_0)^2} \mu\text{m}$	Mesuring Machine, Standard/ SICT-CP-10201
Electrical/Mechanical comparators	10203	(0 ~ 5) mm	0.14 μm	Gauge Block/ SICT-CP-10203
Dial/cylinder gauge testers	10206	(0 ~ 25) mm	$\sqrt{0.12^2 + (0.0028 \times l_0)^2} \mu\text{m}$	Laser Measurement Machine/ SICT-CP-10206
Doctor Blades	10207	(0 ~ 10) mm	1.6 μm	Electronic Micrometer/ SICT-CP-10207
Distance meters; electrooptic/laser/ultrasonic	10208	(0 ~ 40) m	$\sqrt{0.28^2 + (0.005 \times l_0)^2} \text{ mm}$	Laser interferometer/ SICT-CP-10208
End bars	10209	(25 ~ 500) mm (500 ~ 1 000) mm	$\sqrt{0.74^2 + (0.003 \times l_0)^2} \mu\text{m}$ $\sqrt{0.80^2 + (0.003 \times l_0)^2} \mu\text{m}$	Gauge Block/ SICT-CP-10209
Extensometers, linear displacement transducers	10210	(0 ~ 500) mm (500 ~ 1 000) mm	$\sqrt{1.9^2 + (0.0042 \times l_0)^2} \mu\text{m}$ $\sqrt{7.8^2 + (0.058 \times l_0)^2} \mu\text{m}$	Gauge Block, Laser Measurement Machine/ SICT-CP-10210
Filler gauges	10211	(0 ~ 10) mm	1.2 μm	Mesuring Machine, Standard/ SICT-CP-10211
Film Applicators	10212	(0 ~ 10) mm	1.6 μm	Electronic Micrometer/ SICT-CP-10212
Gap gauges	10213	(1 ~ 500) mm	$\sqrt{0.72^2 + (0.0048 \times l_0)^2} \mu\text{m}$	Gauge Block, contact coordinate measuring machines/ SICT-CP-10213
Gauge blocks, by comparison	10214	(0.5 ~ 100) mm	$\sqrt{80^2 + (1.3 \times l_0)^2} \text{ nm}$	Gauge Block Comparator/ SICT-CP-10214
Height gauges/measuring machines	10216	(0 ~ 1 500) mm	$\sqrt{0.68^2 + (0.0035 \times l_0)^2} \mu\text{m}$	Gauge Block, Step gauge/ SICT-CP-10216
	10219	(0 ~ 40) m	$\sqrt{0.03^2 + (0.0027 \times l_0)^2} \text{ mm}$	Laser interferometer/ SICT-CP-10219
Measuring machines, standard	10220	(0 ~ 500) mm	$\sqrt{0.38^2 + (0.002 \times l_0)^2} \mu\text{m}$	Gauge Block/ SICT-CP-10220
Micro scales/Standard scales	10221			Non-contact coordinate measuring machines, standard scale/ SICT-CP-10221
Micro scales		(0 ~ 1) mm	0.4 μm	
Standard scales		(0 ~ 600) mm	$\sqrt{0.51^2 + (0.0021 \times l_0)^2} \mu\text{m}$	
Electronic micrometers	10223	(0 ~ 5) mm	0.14 μm	Gage Block/ SICT-CP-10223
Height micrometers,Riser blocks	10224			Gauge Block Electronic Micrometer/ SICT-CP-10224
Block	10224	(0 ~ 600) mm	$\sqrt{1.1^2 + (0.0019 \times l_0)^2} \mu\text{m}$	
Head		(0 ~ 25) mm	1.2 μm	
Laser scan micrometers	10225	(0.5 ~ 85) mm	$\sqrt{0.46^2 + (0.003 \times l_0)^2} \mu\text{m}$	Cylindrical plug/pin gauge/ SICT-CP-10225
Standard tape rules,Peripheral gauges	10227	(0 ~ 40) m (40 ~ 80) m (80 ~ 100) m	$\sqrt{0.22^2 + (0.0046 \times l_0)^2} \text{ mm}$ $\sqrt{0.25^2 + (0.0046 \times l_0)^2} \text{ mm}$ $\sqrt{0.34^2 + (0.0046 \times l_0)^2} \text{ mm}$	Laser Measurement Machine/ SICT-CP-10227
Cylindrical plug/pin gauges, Thread measuring wire gauges	10228			Mesuring Machine, Standard/ SICT-CP-10228
Cylindrical plug/pin gauges		(0.01 ~ 200) mm	$\sqrt{0.42^2 + (0.003 \times l_0)^2} \mu\text{m}$	
Thread measuring wire gauges		(0.1 ~ 10) mm	0.41 μm	
Radius gauges	10229	(0.1 ~ 100) mm	2.8 μm	Non-contact coordinate measuring machines,
Cylindrical ring gauges	10230	(1.0 ~ 200) mm	$\sqrt{0.69^2 + (0.0034 \times l_0)^2} \mu\text{m}$	Mesuring Machine, Standard/ SICT-CP-10230

Note 1. l_0 unit : mm (10208,10227 unit : m)

102. Linear dimension

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Step gauges	10232	(0 ~ 1 500) mm	$\sqrt{0.76^2 + (0.0025 \times l_0)^2} \mu\text{m}$	Gauge Block, Step gauge/ SICT-CP-10232
Taper thickness gauges	10233	(0 ~ 60) mm	4.3 μm	Profile Projector/ SICT-CP-10223
Ultrasonic thickness gauges	10234	(2.5 ~ 100) mm	3 μm	Ultrasonic Tester Blocks/ SICT-CP-10234
Ultrasonic/coating thickness specimens	10235			Gauge Block, Mesuring Machine, Standard/ SICT-CP-10235
coating thickness specimens		(0.01 ~ 2) mm	0.7 μm	
Ultrasonic thickness specimens		(0.5 ~ 500) mm	$\sqrt{0.64^2 + (0.006 \times l_0)^2} \mu\text{m}$	
Coating thickness testers	10236	(0 ~ 2) mm	1.1 μm	Thickness specimens/ SICT-CP-10236
Torque arms	10237			Gauge Block, contact coordinate measuring machines/ SICT-CP-10237
Torque arms		(1 ~ 1 500) mm	$\sqrt{0.60^2 + (0.0061 \times l_0)^2} \mu\text{m}$	
Wires		(0 ~ 5) mm	1.2 μm	

Note 1. l_0 unit : mm (10208, 10227 unit : m)

103. Angle

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Angle gauge blocks	10302			Indexing tables/ SICT-CP-10302
Angle		(0 ~ 360)°	0.6"	
Autocollimators	10303			Fine angle generators/ SICT-CP-10303
Angle		± 500"	0.6"	
Bevel protractors	10304			Angle Gauge Block, Coordinate Measur Machine/ SICT-CP-10304
Angle Accuracy		(0 ~ 360)°	0.9'	
Accessory Angle		(0 ~ 90)°	0.7'	
Clinometers	10306			Rotary tables/ SICT-CP-10306
Angle		(0 ~ 360)°	3.3"	
Fine angle generators, level comparators	10308			Autocollimators/ SICT-CP-10308
Angle		± 1 000"	0.4"	
Indexing tables	10310			Polygons/ SICT-CP-10310
Angle		(0 ~ 360)°	0.5"	
Plate/Square/Electric levels	10311			Fine angle generators, Rotary tables/ SICT-CP-10311
Angle		(0 ~ 516)'' (516 ~ 1 000)''	0.4" 1.2"	
Inclino meter		(0 ~ 90)°	0.05'	
Squareness		(0 ~ 400) mm	1.8 μm	
Penta-prisms	10314			Autocollimators/ SICT-CP-10314
Angle		90°	0.6"	
Polygons	10315			Indexing tables/ SICT-CP-10315
Angle		(0 ~ 360)°	0.4"	
Rotary tables	10316			Polygons/ SICT-CP-10316
Angle		(0 ~ 360)°	0.5"	
Sinebars, Plates, Tables, Centers	10317			Mesuring Machine, Standard/ SICT-CP-10317
(Sinebars)				
distance, between two roller center		(100 ~ 300) mm	$\sqrt{0.36^2 + (0.002 \times L_0)^2} \mu\text{m}$	
parallelism, between two roller		(100 ~ 300) mm	0.5 μm	
parallelism, between flat-two roller		(100 ~ 300) mm	0.6 μm	
(Plates)				
Center length		(100 ~ 300) mm	$\sqrt{0.12^2 + (0.028 \times L_0)^2} \mu\text{m}$	
Flatness		(100 ~ 300) mm	1.0 μm	
Parallelism		(100 ~ 300) mm	1.2 μm	
Squareness testers, Right angle testers	10318	(0 ~ 600) mm	2.0 μm	Cylindrical Square, Precision Square/ SICT-CP-10318

103. Angle

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Cylindrical squares	10319	(0 ~ 300) mm (300 ~ 600) mm	1.6 μm 1.9 μm	Cylindrical Square/ SICT-CP-10319
Precision squares	10320	(0 ~ 600) mm	2.9 μm	contact coordinate measuring machines/ SICT-CP-10320
Angular displacement transducers	10322	(0 ~ 360) $^{\circ}$	2.9"	Rotary tables/ SICT-CP-10322
Angle				

Note 1. l_0 unit : mm

104. Form

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Form testers	10401			
Height length		(0 ~ 50) mm	1.0 μm	Form Standard Specimens
Width		(0 ~ 50) mm	1.4 μm	Gage Block, Angle Gage Block/
Angle		15° ~ 45°	2'	SICT-CP-10401
Optical flats	10404			
Flatness		ϕ (0 ~ 60) mm	0.05 μm	Optical Flat/
		ϕ (60 ~ 100) mm	0.08 μm	SICT-CP-10404
Optical parallels	10405			
Flatness		ϕ (0 ~ 30) mm	0.05 μm	Optical Flat,Gauge block
Parallelism		ϕ (0 ~ 30) mm	0.07 μm	comparator/
				SICT-CP-10405
Parallel blocks	10406			
Parallelism		(0 ~ 1 000) mm	1.5 μm	Electronic Micrometer/
Flatness		(0 ~ 1 000) mm	1.5 μm	SICT-CP-10406
Length Difference		(0 ~ 1 000) mm	2.2 μm	
Precision surface plates	10407			
Flatness		(2 000 × 2 000) mm	2.0 μm	Electronic Level/
		(5 000 × 5 000) mm	4.8 μm	SICT-CP-10407
Profile gauges	10408			
		(0 ~ 5) mm	0.7 μm	Gage Block/
				SICT-CP-10408
Roundness measurement instruments	10409			
Dector accuracy		(0 ~ 20) μm	0.51 μm	Roundness Standard Ball/
Rotational accuracy of spindle		360°	0.03 μm	SICT-CP-10409
Rotational accuracy of axis		360°	0.04 μm	
Roundness standard/Roundness	10411			
magnification standard specimens				
Standard specimens		(0 ~ 300) μm	0.52 μm	Roundness Tester/
Standard ball		360°	0.08 μm	SICT-CP-10411
Straight edges	10412			
Straightness		(0 ~ 2 000) mm	1.8 μm	Electronic levels/
Parallelism		(0 ~ 2 000) mm	1.8 μm	SICT-CP-10412
Straight rules	10413			
Length		(0 ~ 2 000) mm	0.10 mm	LASER INTERFEROMETER/
				SICT-CP-10413
Test bars	10415			
Roundness		(0 ~ 400) mm	0.6 μm	Roundness Tester,Electronic
Cylindricity		(0 ~ 400) mm	0.6 μm	Micrometer/
Run-out		(0 ~ 400) mm	1.1 μm	SICT-CP-10415

Note 1. l₀ unit : mm

105. Complex geometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Base gauges for electric bulb Inside diameter of pass/stop and screw Pitch	10501	(1 ~ 50) mm (0.3 ~ 10) mm	$\sqrt{0.47^2 + (0.0028 \times l_0)^2}$ μm 1.6 μm	Mesuring Machine, Standard/ SICT-CP-10501
Bench centers Difference of both center Flatness of both bed Center height difference	10502	(0 ~ 200) mm (200 ~ 500) mm (0 ~ 500) mm (0 ~ 200) mm (200 ~ 500) mm	1.8 μm 3.4 μm 1.5 μm 1.8 μm 3.4 μm	Test Bar/ SICT-CP-10502
Contact coordinate measuring machines	10503	(0 ~ 1 500) mm (0 ~ 600) mm (0 ~ 600) mm	$\sqrt{0.56^2 + (0.0044 \times l_0)^2}$ μm 3.2 μm 1.2 μm	Step Gauge/ SICT-CP-10503
Non-contact coordinate measuring machines	10504	(0 ~ 1 000) mm	$\sqrt{0.43^2 + (0.0034 \times l_0)^2}$ μm	Standard Scale/ SICT-CP-10504
Gauge block accessories Round the ministry of Justice A he ministry of Justice,Base block Center point Plane figure Parallelism(triangle edge) Parallelism(equilibrium tide)	10505	(0 ~ 50) mm (0 ~ 50) mm (0 ~ 20) mm (0 ~ 50) mm (0 ~ 300) mm (0 ~ 150) mm	$\sqrt{0.32^2 + (0.0044 \times l_0)^2}$ μm $\sqrt{0.26^2 + (0.0044 \times l_0)^2}$ μm 1.1 μm 0.04 μm 0.4 μm 0.4 μm	Gauge Block/ SICT-CP-10505
Measuring microscopes,Profile projectors Length Magnification Angle	10511	(0 ~ 500) mm (5 ~ 100) 배 (0 ~ 360) °	$\sqrt{0.43^2 + (0.0034 \times l_0)^2}$ μm 0.04 % 0.9°	Standard Scale/ SICT-CP-10511
Microscopes, micro measuring	10512	(0 ~ 1) mm (1 ~ 50) mm	0.7 μm 3.0 μm	Standard Scale/ SICT-CP-10512
Taper plug gauges Small end diameter Big end diameter Plane angle Gage height	10514	(2 ~ 200) mm (2 ~ 200) mm (0 ~ 90) ° (2 ~ 200) mm	$\sqrt{1.3^2 + (0.0041 \times l_0)^2}$ μm $\sqrt{1.4^2 + (0.0041 \times l_0)^2}$ μm $\sqrt{1.2^2 + (0.0044 \times l_0)^2}$ μm	Measuring Machine, Standard/ SICT-CP-10514
Taper ring gauges Small end diameter Big end diameter Plane angle	10515	(5 ~ 200) mm (5 ~ 200) mm (0 ~ 90) °	2.5 μm 2.5 μm 0.006°	contact coordinate measuring machines/ SICT-CP-10515
Stylus type roughness testers Rounghness parameter(Ra) Rounghness parameter(Rz) Mean width(RSm) H,D	10517	(0 ~ 2) μm (2 ~ 10) μm (0 ~ 7) μm (7 ~ 30) μm (0 ~ 300) μm (0 ~ 6) μm (6 ~ 20) μm	9 nm 24 nm 77 nm 0.27 μm 1.3 μm 63 nm 97 nm	Roughness Specimen/ SICT-CP-10517
Socket gauges for electric bulb Outside diameter of pass/stop and screw Pitch	10518	(1 ~ 50) mm (0.3 ~ 10) mm	$\sqrt{0.44^2 + (0.0028 \times l_0)^2}$ μm 1.6 μm	Mesuring Machine, Standard/ SICT-CP-10518

Note 1. unit : mm

 l_0

105. Complex geometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Roughness standard/comparison specimens				
Depth(H)	10519	(0 ~ 6) μm (6 ~ 20) μm	$\sqrt{(9.6 \times R)^2 + 12^2} \text{ nm}$ $\sqrt{(8.94 \times R)^2 + 15^2} \text{ nm}$	Roughness Tester/ SICT-CP-10519
Mean width(RSm)		(0 ~ 300) μm	$\sqrt{(0.01 \times R)^2 + 0.51^2} \mu\text{m}$	
Rounghness parameter(Ra)		(0 ~ 2) μm (2 ~ 10) μm	$\sqrt{(10 \times R)^2 + 2.6^2} \text{ nm}$ $\sqrt{(9.2 \times R)^2 + 7.4^2} \text{ nm}$	
Rounghness parameter(Rz)		(0 ~ 7) μm (7 ~ 30) μm	$\sqrt{(29.2 \times R)^2 + 7.4^2} \text{ nm}$ $\sqrt{(0.025 \times R)^2 + 0.096^2} \mu\text{m}$	
Thread plug gauges	10525			Measuring Machine, Standard/ SICT-CP-10525
Outside diameter		(1 ~ 205) mm	1.7 μm	
Effective diameter		(1 ~ 210) mm	1.1 μm	
Pitch		(0.3 ~ 10) mm	1.2 μm	
Half angle		(0.5 ~ 45) $^\circ$	2'	
Taper thread plug gauges	10526			Measuring Machine, Standard/ SICT-CP-10526
Half angle		(0 ~ 45) $^\circ$	2'	
Pitch		(0.3 ~ 6) mm	1.3 μm	
Gage length		(2 ~ 50) mm	2.6 μm	
Notch length		(2 ~ 50) mm	3.6 μm	
Small outside diameter		(2 ~ 200) mm	2.3 μm	
Big outside diameter		(2 ~ 200) mm	4.8 μm	
Small effective diameter		(2 ~ 200) mm	2.9 μm	
Big effective diameter		(2 ~ 200) mm	5.1 μm	
Thread ring gauges	10527			Measuring Machine, Standard/ SICT-CP-10527
Minor diameter		(3 ~ 200) mm	1.5 μm	
Effective diameter		(3 ~ 200) mm	2.3 μm	
Pitch		(0.3 ~ 10) mm	1.6 μm	
Taper thread ring gauges	10528			Measuring Machine, Standard/ SICT-CP-10528
Alternation		$\pm 3 \text{ mm}$	2.7 μm	
Thinkness		(0 ~ 100) mm	2.3 μm	
Notch length		(0 ~ 100) mm	3.2 μm	
V-blocks,Boxblocks	10529			contact coordinate measuring machines/ SICT-CP-10529
Plane figure		(5 ~ 300) mm	1.7 μm	
Parallelism		(5 ~ 300) mm	2.0 μm	
Difference of both part		(5 ~ 300) mm	2.8 μm	

106. Various dimensional

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Inside/Outside/Geartooth calipers, Caliper gauges	10601	(0 ~ 2 000) mm (0 ~ 300) mm	$\sqrt{8.2^2 + (0.007 \times l_0)^2}$ μm	Gauge Block/ SICT-CP-10601
			$\sqrt{3.7^2 + (0.0032 \times l_0)^2}$ μm	
			μm	
Cylinder/bore gauges	10603	(0 ~ 1 000) mm (0.1 ~ 25) mm	1.0 μm	Dial Gauge Tester/ SICT-CP-10603
			3.3 μm	
			μm	
Depth gauges, Depth micrometers	10604	(0 ~ 300) mm (0 ~ 1 000) mm	$\sqrt{0.86^2 + (0.0034 \times l_0)^2}$ μm	Gauge Block/ SICT-CP-10604
			$\sqrt{5.9^2 + (0.0048 \times l_0)^2}$ μm	
			μm	
Dial/digital gauges	10605	(0 ~ 100) mm (0 ~ 25) mm	$\sqrt{0.21^2 + (0.0082 \times l_0)^2}$ μm	Dial Gauge Tester/
			$\sqrt{0.59^2 + (0.004 \times l_0)^2}$ μm	
			μm	
Grind gauges	10608	(0 ~ 1) mm (0 ~ 150) mm	1.8 μm	Electronic micrometer/ SICT-CP-10608
			2.5 μm	
Micro indicators, Test indicators	10609	(0 ~ 5) mm	0.22 μm	Dial Gauge Tester/ SICT-CP-10609
			μm	
Micrometer heads	10610	(0 ~ 50) mm	0.8 μm	Gauge Block/ SICT-CP-10610
			$\sqrt{1.3^2 + (0.003 \times 0.4 \times t)}$	
3-points, Micrometers	10611	(2 ~ 200) mm (200 ~ 300) mm	3 μm	Ring Gauge/ SICT-CP-10611
			μm	
Inside micrometers	10612	(5 ~ 300) mm (25 ~ 500) mm (13 ~ 500) mm	$\sqrt{1.1^2 + (0.0042 \times l_0)^2}$ μm	Gauge Block/ SICT-CP-10612
			$\sqrt{1.1^2 + (0.0042 \times l_0)^2}$ μm	
			$\sqrt{1.2^2 + (0.0048 \times l_0)^2}$ μm	
Outside micrometers	10613	(0 ~ 25) mm (25 ~ 1 000) mm (1 ~ 85) mm	$\sqrt{0.2^2 + (0.003 \times l_0)^2}$ μm	Gauge Block, cylindrical plug gauges/ SICT-CP-10613
			$\sqrt{0.83^2 + (0.003 \times l_0)^2}$ μm	
			0.8 μm	
Particle counters	10615	(Air) Flow Threshold voltage Counting efficiency	(0.1 ~ 1) μm	Particle calibration system/ SICT-CP-10615
			(0 ~ 100) L/min	
			(0 ~ 10) V	
			(0 ~ 110) %	
		(Liquid) Flow Threshold voltage	(0.05 ~ 25) μm	
			(0 ~ 100) mL/min	
			(0 ~ 10) V	1.4 mL/min
				0.42 mV
Standard sieves	10617	Sieve opening Wire rod diameter	1.5 μm	Non-contact coordinate measuring machines/ SICT-CP-10617
			2.4 μm	
Welding gauges	10620	Height or depth Rule Angle	8.2 μm	Non-contact coordinate measuring machine, Gauge Block/ SICT-CP-10620
			6.0 μm	
			0.7°	

Note 1. l_0 unit : mm

201. Mass

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Auto-hopper scale balances	20102	(0 ~ 10) kg (10 ~ 50) kg (50 ~ 200) kg (200 ~ 500) kg (500 ~ 1 000) kg	1.2 g 14 g 21 g 49 g 0.10 kg	Hopper Scale Weight/ SICT-CP-20102
Auto-packer scale balances	20103	(0 ~ 10) kg (10 ~ 50) kg (50 ~ 200) kg	0.8 g 7.7 g 16 g	Weight/ SICT-CP-20103
Axle weigher balances Portable	20104	(100 ~ 500) kg (500 ~ 1 000) kg (1 000 ~ 2 000) kg (2 000 ~ 5 000) kg (5 000 ~ 10 000) kg (10 000 ~ 30 000) kg	0.2 kg 0.4 kg 1 kg 5 kg 10 kg 24 kg	Force Calibration Machine/ SICT-CP-20104
Dial platform scale balances	20106	(0 ~ 30) kg (30 ~ 60) kg (60 ~ 100) kg	42 g 0.08 kg 0.21 kg	Weight/ SICT-CP-20106
Dial swing scale balances	20107	(0 ~ 1) kg (1 ~ 10) kg (10 ~ 20) kg (20 ~ 50) kg (50 ~ 100) kg (100 ~ 200) kg (200 ~ 500) kg (500 ~ 1 000) kg (1 000 ~ 2 000) kg (2 000 ~ 5 000) kg	0.96 g 9.6 g 20 g 48 g 96 g 0.23 kg 0.48 kg 0.96 kg 1.9 kg 4.6 kg	Weight/ SICT-CP-20107
Electric balances	20109	(0 ~ 2) g (2 ~ 6) g (6 ~ 20) g (20 ~ 50) g (50 ~ 100) g (100 ~ 200) g (200 ~ 500) g (500 ~ 1 000) g (1 ~ 2) kg (2 ~ 5) kg (5 ~ 10) kg (10 ~ 25) kg (25 ~ 40) kg (40 ~ 60) kg (60 ~ 150) kg (150 ~ 600) kg (600 ~ 1 000) kg (1 000 ~ 2 000) kg (2 000 ~ 5 000) kg	7.0 µg 9.3 µg 14 µg 19 µg 29 µg 0.05 mg 0.10 mg 0.20 mg 0.5 mg 1.0 mg 3 mg 6 mg 16 mg 24 mg 0.30 g 1.2 g 2.0 g 38 g 0.10 kg	Weight/ SICT-CP-20109

201. Mass

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Manual swing scale balances	20111	(0 ~ 50) kg (50 ~ 200) kg (200 ~ 500) kg (500 ~ 1 000) kg	38 g 0.19 kg 0.38 kg 0.94 kg	Weight/ SICT-CP-20111
Platform scale balances	20112	(0 ~ 200) kg (200 ~ 500) kg (500 ~ 1 000) kg	46 g 0.09 kg 0.46 kg	Weight/ SICT-CP-20112
Spring scale balances	20113	(0 ~ 10) kg (10 ~ 50) kg (50 ~ 100) kg	21 g 0.08 kg 0.21 kg	Weight/ SICT-CP-20113
Weights less than class E2	20116	(1 mg ~ 20 kg) 1 mg 2 mg 5 mg 10 mg 20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg less than class F2 less than class M1 less than class F2	1.0 µg 1.0 µg 1.0 µg 1.0 µg 1.2 µg 1.4 µg 1.7 µg 2.3 µg 2.9 µg 3.5 µg 4.6 µg 5.8 µg 7.0 µg 9.3 µg 12 µg 18 µg 36 µg 95 µg 0.12 mg 0.36 mg 0.94 mg 1.8 mg 3.7 mg (20 ~ 100) kg 50 kg 100 kg (100 ~ 200) kg 200 kg (200 ~ 1 000) kg 500 kg 1 000 kg	Weights, Mass Comparator/ SICT-CP-20116

202. Force

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Force measuring devices	20202	(0.4 ~ 20) N (20 ~ 50) N (50 ~ 100) N (100 ~ 200) N (200 ~ 500) N (0.5 ~ 1) kN (1 ~ 2) kN (2 ~ 5) kN (5 ~ 10) kN (10 ~ 20) kN (20 ~ 50) kN (50 ~ 100) kN (100 ~ 200) kN (200 ~ 500) kN (500 ~ 1 000) kN	6.0×10^{-4} 8.4×10^{-5} 8.5×10^{-5} 7.3×10^{-5} 6.9×10^{-5} 9.2×10^{-5} 9.0×10^{-5} 8.4×10^{-5} 8.7×10^{-5} 3.6×10^{-4} 4.1×10^{-4} 4.8×10^{-4} 4.5×10^{-4} 4.3×10^{-4} 4.5×10^{-4}	Force Calibration Machine/ SICT-CP-20202
Tension/compression testing machines	20203	0.1 N ~ 2 kN (2 ~ 5) kN (5 ~ 10) kN (10 ~ 20) kN (20 ~ 50) kN (50 ~ 100) kN (100 ~ 300) kN	1.2×10^{-3} 1.3×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.4×10^{-3}	Weights, Force Measuring Device/ SICT-CP-20203
		compression (0.1 ~ 50) N (50 ~ 100) N (100 ~ 200) N (200 ~ 500) N (0.5 ~ 1) kN (1 ~ 2) kN (2 ~ 5) kN (5 ~ 10) kN (10 ~ 20) kN (20 ~ 50) kN (50 ~ 100) kN (100 ~ 300) kN (300 ~ 500) kN (500 ~ 1 000) kN (1 000 ~ 3 000) kN	1.2×10^{-3} 1.5×10^{-3} 1.3×10^{-3} 1.4×10^{-3} 1.3×10^{-3} 1.4×10^{-3} 1.3×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.4×10^{-3} 1.6×10^{-3}	
push-pull gauge	20204	(0.02 ~ 0.2) N 0.2 N ~ 2 kN (2 ~ 5) kN	1.4×10^{-2} 1.3×10^{-3} 8.4×10^{-4}	Weights, Force Calibration Machine/ SICT-CP-20204

203. Torque

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Torque measuring devices	20302	(0.001 ~ 1) N·m (1 ~ 10) N·m (10 ~ 20) N·m (20 ~ 50) N·m (50 ~ 100) N·m (100 ~ 200) N·m (200 ~ 500) N·m (500 ~ 1 000) N·m (1 000 ~ 2 000) N·m	3.3×10^{-3} 4.1×10^{-4} 7.5×10^{-4} 2.3×10^{-4} 3.4×10^{-4} 2.1×10^{-4} 2.2×10^{-4} 1.5×10^{-4} 1.6×10^{-4}	Torque Calibration Machine/ SICT-CP-20302
Torque wrenches/drivers	20303	(0.02 ~ 0.1) N·m (0.1 ~ 0.5) N·m (0.5 ~ 1) N·m (1 ~ 2) N·m (2 ~ 5) N·m (5 ~ 10) N·m (10 ~ 20) N·m (20 ~ 50) N·m (50 ~ 100) N·m (100 ~ 200) N·m (200 ~ 500) N·m (500 ~ 1 000) N·m (1 000 ~ 2 000) N·m	1.4×10^{-2} 9.5×10^{-3} 7.8×10^{-3} 6.2×10^{-3} 4.6×10^{-3} 4.5×10^{-3} 4.7×10^{-3} 4.5×10^{-3} 4.9×10^{-3} 3.8×10^{-3} 3.7×10^{-3} 3.8×10^{-3} 2.8×10^{-3}	Torque Measuring Device/ SICT-CP-20303

204. Pressure

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Altimeters	20401	(-4 000 ~ 20 000) m (20 000 ~ 47 000) m	12 m 15 m	Digital Manometer, Air Dead Weight Tester/ SICT-CP-20401
Manometers	20402	(0 ~ 200) kPa	2.9×10^{-3}	Digital Manometer, Air Dead Weight Tester/ SICT-CP-20402
Pneumatic pressure ballances	20403	(5 ~ 7 000) kPa	6.1×10^{-5}	분동식 압력계/ SICT-CP-20403
Hydraulic pressure ballances	20404	(2 ~ 200) MPa	1.2×10^{-4}	분동식 압력계/ SICT-CP-20404
Air data test systems	20405	(-2 500 ~ 20 000) m (20 000 ~ 30 500) m	0.8 m 7 m	Digital Manometer, Air Dead Weight Tester/ SICT-CP-20405
		Dynamic pressure	(0 ~ 342) km/hr (342 ~ 2 122) km/hr	0.1 km/hr 0.3 km/hr
Absolute pressure gauges	20406	Dial,digital (5 ~ 7 000) kPa abs (7 ~ 200) MPa abs	8.4×10^{-5} 1.2×10^{-4}	Digital Manometer, Air Dead Weight Tester/ SICT-CP-20406
Blood pressure gauges	20407	(0 ~ 40) kPa	0.01 kPa	Digital Manometer/ SICT-CP-20407
Compound pressure gauges	20408	(-95 ~ 7 000) kPa	7.9×10^{-5}	Air Dead Weight Tester/ SICT-CP-20408
Differential pressure gauges	20409	(0 ~ 500) Pa (500 ~ 5 000) Pa (5 ~ 7 000) kPa	0.13 Pa 1.5 Pa 2.2×10^{-3}	Digital Manometer, Air Dead Weight Tester/ SICT-CP-20409
Gauge pressure gauges	20411	(0 ~ 500) Pa (500 ~ 5 000) Pa (5 ~ 7 000) kPa (7 ~ 200) MPa (200 ~ 500) MPa	0.13 Pa 1.3 Pa 7.9×10^{-5} 1.2×10^{-4} 3.5×10^{-4}	Digital Manometer, Air Dead Weight Tester, Oil Dead Weight Tester/ SICT-CP-20411
Pressure transducers/transmitters	20412	Absolute pressure Gauge pressure	5 kPa abs ~ 200 MPa abs (0 ~ 500) Pa (500 ~ 5 000) Pa 5 kPa ~ 500 MPa	6.7×10^{-4} 0.13 Pa 1.2 Pa 6.7×10^{-4}
Dial type vacuum gauges	20413	(-95 ~ 0) kPa	0.059 kPa	Air Dead Weight Tester, SICT-CP-20413

205. Vacuum

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Capacitance diaphragm gauges	20501	(0.133 ~ 13.3) Pa abs (13.3 ~ 133.3) Pa abs (133.3 ~ 1 333) Pa abs (1.333 ~ 13.332) kPa abs (13.332 ~ 133.322) kPa abs	0.051 Pa abs 0.49 Pa abs 1.5 Pa abs 12 Pa abs 17 Pa abs	Baratron gauge, SRG / SICT-CP-20501
Spinning rotor gauges	20502	0.15 mPa abs ~ 0.01 Pa abs	3.4×10^{-2}	Baratron gauge, SRG / SICT-CP-20502
Ionization gauges	20503	0.093 mPa abs ~ 0.15 mPa abs 0.15 mPa abs ~ 0.01 Pa abs	6.0×10^{-2} 3.5×10^{-2}	Baratron gauge, SRG, Ion / SICT-CP-20503
Thermal conductivity gauges: pirani, thermocouple, convectron etc.	20504	(0.133 ~ 13.3) Pa abs (13.3 ~ 133.3) Pa abs (133.3 ~ 1 333) Pa abs (1.333 ~ 13.332) kPa abs (13.332 ~ 133.322) kPa abs	0.051 Pa abs 0.49 Pa abs 1.5 Pa abs 13 Pa abs 17 Pa abs	Baratron gauge, SRG / SICT-CP-20504
Standard leaks, Helium leak detectors	20505	22.0 $\mu\text{Pa m}^3/\text{s}$ 1.60 $\mu\text{Pa m}^3/\text{s}$ 0.51 $\mu\text{Pa m}^3/\text{s}$ 15.0 $\text{nPa m}^3/\text{s}$ 6.4 $\text{nPa m}^3/\text{s}$ 0.24 $\text{nPa m}^3/\text{s}$	4.8 $\mu\text{Pa m}^3/\text{s}$ 0.38 $\mu\text{Pa m}^3/\text{s}$ 0.098 $\mu\text{Pa m}^3/\text{s}$ 3.2 $\text{nPa m}^3/\text{s}$ 1.3 $\text{nPa m}^3/\text{s}$ 0.049 $\text{nPa m}^3/\text{s}$	Standard leaks, Helium leak detectors / SICT-CP-20505

206. Volume

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Volumetric glasswares	20601	(0 ~ 0.1) ml (0.1 ~ 0.5) ml (0.5 ~ 1) ml (1 ~ 2) ml (2 ~ 5) ml (2 ~ 10) ml (10 ~ 25) ml (25 ~ 50) ml (50 ~ 100) ml (100 ~ 200) ml (200 ~ 250) ml (250 ~ 500) ml (500 ~ 1 000) ml (1 000 ~ 2 000) ml (2 000 ~ 5 000) ml (5 000 ~ 10 000) ml	0.31 µl 0.40 µl 0.43 µl 1.1 µl 1.7 µl 2.5 µl 4.3 µl 5.1 µl 7.9 µl 13 µl 43 µl 69 µl 92 µl 0.17 ml 0.49 ml 0.87 ml	Weight, balances / SICT-CP-20601
Pycnometers	20602	(0 ~ 50) ml (50 ~ 100) ml (100 ~ 250) ml (250 ~ 500) ml	2.4 µl 4.4 µl 10 µl 21 µl	Weight, balances / SICT-CP-20602
Standard volume vessels	20604	(0 ~ 20) L (20 ~ 200) L (200 ~ 10 000) L	9.0×10^{-5} 1.3×10^{-4} 1.1×10^{-3}	Balances, Master Meter, Standard volume vessel / SICT-CP-20604
Concrete air content meters	20605	(0 ~ 10) %	0.032 %	Weight, balances / SICT-CP-20605
Piston type volume meters	20606	(0 ~ 1) µl (1 ~ 2) µl (2 ~ 5) µl (5 ~ 10) µl (0.01 ~ 0.02) ml (0.02 ~ 0.05) ml (0.05 ~ 0.1) ml (0.1 ~ 0.2) ml (0.2 ~ 0.5) ml (0.5 ~ 1) ml (1 ~ 2) ml (2 ~ 5) ml (5 ~ 10) ml (10 ~ 20) ml (20 ~ 50) ml (50 ~ 100) ml	0.004 µl 0.005 µl 0.006 µl 0.008 µl 0.018 µl 0.035 µl 0.047 µl 0.14 µl 0.35 µl 0.65 µl 1.6 µl 1.9 µl 2.4 µl 5.0 µl 14 µl 64 µl	Weight, balances / SICT-CP-20606

207. Density

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Liquid density meters	20702	(0.7 ~ 1.4) g/cm ³	0.000 078 g/cm ³	Density standard materials/ SICT-CP-20702
Salinity meters	20704	(0 ~ 10) % (10 ~ 26) %	0.004 % 0.007 %	NaCl/ SICT-CP-20704
Sucrose meters	20705	(0 ~ 20) % (20 ~ 60) % (60 ~ 80) %	0.014 % 0.016 % 0.025 %	Sucrose/ SICT-CP-20705
Hydrometers; density, specific gravity, alcohol, API, baume, sugar, milk, soil, salinity, LPG, etc.	20706			
density		(0.600 ~ 0.700) g/cm ³	0.000 035 g/cm ³	Solid density standard meterial,
		(0.700 ~ 0.800) g/cm ³	0.000 038 g/cm ³	Hydrostatic weighing Apparatus/ SICT-CP-20706-1
		(0.800 ~ 0.900) g/cm ³	0.000 042 g/cm ³	
		(0.900 ~ 1.000) g/cm ³	0.000 046 g/cm ³	
		(1.000 ~ 1.100) g/cm ³	0.000 050 g/cm ³	
		(1.100 ~ 1.200) g/cm ³	0.000 055 g/cm ³	
		(1.200 ~ 1.300) g/cm ³	0.000 059 g/cm ³	
		(1.300 ~ 1.400) g/cm ³	0.000 066 g/cm ³	
		(1.400 ~ 1.500) g/cm ³	0.000 071 g/cm ³	
		(1.500 ~ 1.600) g/cm ³	0.000 075 g/cm ³	
		(1.600 ~ 1.700) g/cm ³	0.000 079 g/cm ³	
		(1.700 ~ 1.800) g/cm ³	0.000 084 g/cm ³	
		(1.800 ~ 1.900) g/cm ³	0.000 088 g/cm ³	
		(1.900 ~ 2.000) g/cm ³	0.000 093 g/cm ³	
		(2.000 ~ 2.200) g/cm ³	0.000 25 g/cm ³	
		(2.200 ~ 3.000) g/cm ³	0.000 28 g/cm ³	
		(3.000 ~ 3.600) g/cm ³	0.000 30 g/cm ³	
		(3.600 ~ 4.000) g/cm ³	0.000 32 g/cm ³	
specific gravity		0.590 ~ 0.700	0.000 068	Solid density standard meterial,
		0.700 ~ 0.800	0.000 069	Hydrostatic weighing Apparatus/ SICT-CP-20706-2
		0.800 ~ 0.900	0.000 072	
		0.900 ~ 1.000	0.000 075	
		1.000 ~ 1.100	0.000 078	
		1.100 ~ 1.200	0.000 082	
		1.200 ~ 1.300	0.000 086	
		1.300 ~ 1.400	0.000 091	
		1.400 ~ 1.500	0.000 096	
		1.500 ~ 1.600	0.000 10	
		1.600 ~ 1.800	0.000 11	
		1.800 ~ 2.000	0.000 12	
		2.000 ~ 2.020	0.000 26	
		2.020 ~ 2.500	0.000 60	
		2.500 ~ 3.000	0.000 61	

* 20704, 20705, 20706, 20707 unit % is weight percent.

207. Density

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Hydrometers; density, specific gravity, alcohol, API, baume, sugar, milk, soil, salinity, LPG, etc.	20706			
alcohol (Volumn)		(0 ~ 10) % (10 ~ 30) % (30 ~ 40) % (40 ~ 50) % (50 ~ 60) % (60 ~ 70) % (70 ~ 80) % (80 ~ 90) % (90 ~ 100) %	0.039 % 0.043 % 0.038 % 0.030 % 0.025 % 0.023 % 0.020 % 0.019 % 0.017 %	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-3
API		-1 ~ 51 51 ~ 101	0.013 0.014	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-4
Baumé-light		10 ~ 30 30 ~ 40 40 ~ 60 60 ~ 70 70 ~ 100	0.015 0.016 0.018 0.019 0.12	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-5
Baumé - heavy		0 ~ 40 40 ~ 75	0.014 0.013	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-5
sugar		(0 ~ 10) % (10 ~ 90) %	0.018 % 0.017 %	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-6
milk		(15 ~ 20) (20 ~ 40)	0.081 0.082	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-7
Bouyoucos		(-5.0 ~ 60.0) g/L	0.14 g/L	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-8
salinity		(0 ~ 26.4) %	0.025 %	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-9
LPG		(0.50 ~ 0.55) g/cm ³ (0.55 ~ 0.60) g/cm ³ (0.60 ~ 0.65) g/cm ³	0.000 065 g/cm ³ 0.000 066 g/cm ³ 0.000 068 g/cm ³	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-10

* 20704, 20705, 20706, 20707 unit % is weight percent.

207. Density

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Hydrometers; density, specific gravity, alcohol, API, baume, sugar, milk, soil, salinity, LPG, etc.	20706 Twadde11	0 ~ 12	0.016	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-11
		12 ~ 74	0.059	
		74 ~ 102	0.060	
		102 ~ 170	0.061	
		170 ~ 200	0.062	
Chloride meters	20707	(0 ~ 0.1) %	0.000 2 %	Chlorine standard liquid/ SICT-CP-20707
		(0.1 ~ 2.0) %	0.001 0 %	
Solid density	20799 Stainless steel	(1 ~ 2) g	0.004 6 g/cm ³	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-11
		(2 ~ 5) g	0.002 4 g/cm ³	
		(5 ~ 10) g	0.001 0 g/cm ³	
		(10 ~ 20) g	0.000 59 g/cm ³	
		(20 ~ 50) g	0.000 43 g/cm ³	
		(50 ~ 100) g	0.000 37 g/cm ³	
		(100 ~ 500) g	0.000 36 g/cm ³	
	Glass	(1 ~ 2) g	0.000 46 g/cm ³	
		(2 ~ 5) g	0.000 25 g/cm ³	
		(5 ~ 10) g	0.000 14 g/cm ³	
		(10 ~ 20) g	0.000 12 g/cm ³	
		(20 ~ 500) g	0.000 11 g/cm ³	

* 20704, 20705, 20706, 20707 unit % is weight percent.

208. Viscosity

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Kinematic viscometers; capillary,etc	20801	(2.5 ~ 100 000) mPa·s	1.3×10^{-2}	Viscosity CRM/ SICT-CP-20801
Dynamic viscometers; rotaional, etc Viscosity	20802	(2.5 ~ 200 000) mPa·s	1.4×10^{-2}	Viscosity CRM/ SICT-CP-20802

209. Fluid flow

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Anemometers; hot-wire	20901	(0.1 ~ 1.0) m/s (1.0 ~ 2.0) m/s (2.0 ~ 70) m/s	8.7×10^{-2} 8.4×10^{-3} 4.8×10^{-3}	Wind Tunnel, Pitot tube, LDV/ SICT-CP-20901
Anemometers; pitot tube, etc.	20902	(0.1 ~ 1.0) m/s (1.0 ~ 2.0) m/s (2.0 ~ 70) m/s	8.7×10^{-2} 8.4×10^{-3} 4.8×10^{-3}	Wind Tunnel, Pitot tube, LDV/ SICT-CP-20902
Gas flowmeters:differential pressure	20908	(1.2×10^{-5} ~ 0.12) m ³ /h (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters:differential pressure	20909	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Liquid flowmeters: electromagnetic	20910	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters:thermal mass, etc.	20911	(1.2×10^{-5} ~ 0.12) m ³ /h (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters: Coriolis, etc.	20912	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters:open channel, etc.	20914	(1.2×10^{-5} ~ 0.12) m ³ /h (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters: positive displacement	20915	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters:turbine	20916	(1.2×10^{-5} ~ 0.12) m ³ /h (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters:turbine	20917	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters:ultrasonic	20918	(1.2×10^{-5} ~ 0.12) m ³ /h (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929

209. Fluid flow

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Liquid flowmeters:ultrasonic	20919	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2 × 10 ⁻³ 2.6 × 10 ⁻³ 7.0 × 10 ⁻⁴	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters:variable area	20920	(1.2 × 10 ⁻⁵ ~ 0.12) m ³ /h (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9 × 10 ⁻³ 2.0 × 10 ⁻³ 3.6 × 10 ⁻³	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters:variable area	20921	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2 × 10 ⁻³ 2.6 × 10 ⁻³ 7.0 × 10 ⁻⁴	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters:vortex	20922	(1.2 × 10 ⁻⁵ ~ 0.12) m ³ /h (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9 × 10 ⁻³ 2.0 × 10 ⁻³ 3.6 × 10 ⁻³	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters:vortex	20923	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2 × 10 ⁻³ 2.6 × 10 ⁻³ 7.0 × 10 ⁻⁴	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Anemometers; vane, etc	20925	(0.1 ~ 1.0) m/s (1.0 ~ 2.0) m/s (2.0 ~ 70) m/s	8.7 × 10 ⁻² 8.4 × 10 ⁻³ 4.8 × 10 ⁻³	Wind Tunnel, Pitot tube, LDV/ SICT-CP-20925
Others; Ultrasonic current meter	20999	(0.1 ~ 1.0) m/s (1.0 ~ 2.0) m/s (2.0 ~ 70) m/s	8.7 × 10 ⁻² 8.4 × 10 ⁻³ 4.8 × 10 ⁻³	Wind Tunnel, Pitot tube, LDV/ SICT-CP-20999

210. Hardness

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Brinell hardness testers	21001	(100 ~ 250) HBW 10/3 000 (250 ~ 450) HBW 10/3 000 (450 ~ 650) HBW 10/3 000	3.1 HBW 10/3 000 4.9 HBW 10/3 000 8.2 HBW 10/3 000	Brinell Hardness CRM/ SICT-CP-21001
Rockwell hardness testers	21002	(20 ~ 70) HRC (20 ~ 100) HRBW (42 ~ 86) HR30N (29 ~ 82) HR30TW	0.45 HRC 0.80 HRBW 0.70 HR30N 1.1 HR30TW	Rockwell Hardness CRM/ SICT-CP-21002
Shore hardness testers	21003	(20 ~ 100) HS	0.9 HS	Shore Hardness CRM/ SICT-CP-21003
Vickers hardness testers	21004	(50 ~ 300) HV 0.2 (300 ~ 600) HV 0.2 (600 ~ 850) HV 0.2 (50 ~ 300) HV 0.3 (300 ~ 600) HV 0.3 (600 ~ 850) HV 0.5 (50 ~ 300) HV 0.5 (300 ~ 600) HV 0.5 (600 ~ 850) HV 1 (50 ~ 300) HV 10 (300 ~ 600) HV 10 (600 ~ 850) HV 10 (300 ~ 600) HV 30 (600 ~ 850) HV 30	6.0 HV 0.2 18 HV 0.2 27 HV 0.2 5.0 HV 0.3 14 HV 0.3 26 HV 0.5 6.0 HV 0.5 15 HV 0.5 20 HV 1 3.0 HV 10 8.0 HV 10 11 HV 10 8.0 HV 30 11 HV 30	Vickers Hardness CRM/ SICT-CP-21004
Durometer hardness testers	21005	(0 ~ 100) HDA (0 ~ 100) HDAM (0 ~ 100) HDAO (0 ~ 100) HDB (0 ~ 100) HDC (0 ~ 100) HDC2 (0 ~ 100) HDCS (0 ~ 100) HDD (0 ~ 100) HDDO (0 ~ 100) HDE (0 ~ 100) HDE2 (0 ~ 100) HDF (0 ~ 100) HDF0 (0 ~ 100) HDM (0 ~ 100) HDO (0 ~ 100) HD00 (0 ~ 100) HD000 (0 ~ 100) HD000-S	0.4 HDA 0.8 HDAM 0.4 HDAO 0.4 HDB 0.3 HDC 0.6 HDC2 0.3 HDCS 0.3 HDD 0.3 HDDO 0.4 HDE 0.6 HDE2 0.6 HDF 0.6 HDF0 0.8 HDM 0.3 HDO 0.4 HD00 0.4 HD000 0.3 HD000-S	Durometer Calibration device/ SICT-CP-21005
Leeb hardness testers	21006	(400 ~ 1 000) HLD	4.6 HLD	Leeb Hardness CRM/ SICT-CP-21006

301. Time/frequency

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Frequency standards Time Base Frequency	30102	(1 ~ 100) MHz	4.3×10^{-12}	GPS Receiver, Universal Counter/ SICT-CP-30102
General frequency sources Time Base Frequency	30103	(10 ~ 100) kHz (0.1 ~ 100) MHz	1.3×10^{-11} 2.2×10^{-12}	GPS Receiver, Universal Counter/ SICT-CP-30103
Frequency meters/counters Time Base Frequency	30104	(1 ~ 10) MHz	4.3×10^{-12}	GPS Receiver, Universal Counter/ SICT-CP-30104
		0.1 Hz 0.1 Hz ~ 40 GHz	6.4×10^{-10} 6.4×10^{-11}	
Time interval sources Period	30105	1 ns ~ 10 s	6.1×10^{-9}	GPS Receiver, Universal Counter/ SICT-CP-30105
		(1 ~ 100) ns	0.15 ns	
		100 ns ~ 1 ms	1.3 ns	
		1 ms ~ 10 s	2.1 ns	
Time interval meters /Stop watches/Timers Trigger Voltage	30106	(-5 ~ 5) V	1.2×10^{-4}	Stop Watch Calibrator/ SICT-CP-30106
		(5 ~ 100) ns	6.2×10^{-5} ns	
		(1 ~ 10) MHz	6.2×10^{-11}	
Reference Frequency		day	1.1×10^{-7}	
		month	3.6×10^{-7}	
Relative Time Difference				
		(-9.95 ~ 9.95) s / day	6.1 ms	
		(-300 ~ 300) s / month	6.2 ms	
Time rate				
		(1 ~ 100) s	5.8×10^{-6}	
		(100 ~ 1 000) s	8.2×10^{-6}	
		(1 000 ~ 10 000) s	5.8×10^{-5}	
Timer				
		≥ 1	0.58	
Count				

302. Velocity & revolution

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Standard RPM generators Revolution Velocity Measurement	30201	(1 ~ 10 000) min ⁻¹	0.004 0 min ⁻¹	GPS Receiver, Synthesizer Function Generator/ SICT-CP-30201
Revolution Velocity Measurement (Centrifuge)		(30 ~ 5 000) min ⁻¹ (5 000 ~ 8 500) min ⁻¹ (8 500 ~ 50 000) min ⁻¹ (50 000 ~ 80 000) min ⁻¹ (80 000 ~ 99 000) min ⁻¹	0.059 min ⁻¹ 0.099 min ⁻¹ 0.59 min ⁻¹ 0.93 min ⁻¹ 1.1 min ⁻¹	
Contact type tachometers Revolution Velocity Measurement	30202	(1 ~ 10) min ⁻¹ (10 ~ 1 000) min ⁻¹ (1 000 ~ 5 000) min ⁻¹	0.10 min ⁻¹ 0.016 min ⁻¹ 0.063 min ⁻¹	GPS Receiver, Tachometer Cal System/ SICT-CP-30202
Photo tachometers/stroboscopes Revolution Velocity Measurement (Photo-tachometer)	30203	(1 ~ 999.99) min ⁻¹ (1 000.0 ~ 99 999.9) min ⁻¹ (100 000 ~ 600 000) min ⁻¹	0.006 1 min ⁻¹ 0.061 min ⁻¹ 0.61 min ⁻¹	GPS Receiver, Photo Signal Detector/ SICT-CP-30203
Revolution Velocity Measurement (Stroboscope)		(30 ~ 9 000) min ⁻¹ (9 000 ~ 90 000) min ⁻¹ (90 000 ~ 500 000) min ⁻¹	0.005 8 min ⁻¹ 0.058 min ⁻¹ 0.58 min ⁻¹	
Speed meters Speed Test	30204	(0 ~ 400) km/h	6.1×10^{-3} km/h	GPS Receiver, Synthesizer SICT-CP-30204
Wow-flutter generators Carrier Frequency	30205	10 Hz ~ 99.99 kHz	6.2×10^{-6}	GPS Receiver, Universal Counter/ SICT-CP-30205
Function Frequency		1 Hz ~ 10 kHz (10 ~ 30) kHz	6.2×10^{-6} 2.1×10^{-6}	
Wow/Flutter Deviation		(1 Hz ~ 100 Hz) (0 ~ 3) %	0.025 %	
Output Level		(1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 6) V	5.8×10^{-4} 1.7×10^{-4} 1.3×10^{-4}	
CCIR Pulse		10 ms 30 ms 60 ms 100 ms	1.0×10^{-2} ms 3.0×10^{-2} ms 6.0×10^{-2} ms 1.0×10^{-1} ms	
Wow-flutter meters Wow/Flutter Deviation	30206	(0.1 ~ 0.3) % (0.3 ~ 3) %	0.019 % 0.020 %	GPS Receiver, Wow Flutter Calibrator/ SICT-CP-30206
Carrier Frequency		3 kHz 3.15 kHz	6.2×10^{-5} kHz 6.2×10^{-5} kHz	
CCIR Pulse		(10 ~ 100) ms	0.59 %	
Output Voltage		(1 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V	6.8 μ V 9.8 μ V 76 μ V	

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC ammeters	40101	(±) 0 nA (0 ~ 1) nA (1 ~ 10) nA (10 ~ 100) nA (0.1 ~ 1) μA (1 ~ 10) μA (10 ~ 100) μA (0.1 ~ 1) mA (1 ~ 10) mA (10 ~ 100) mA (100 ~ 200) mA (0.2 ~ 1) A (1 ~ 10) A (10 ~ 100) A	6 pA 6.6×10^{-3} 9.5×10^{-4} 5.1×10^{-4} 6.1×10^{-4} 2.9×10^{-4} 3.6×10^{-5} 2.9×10^{-5} 2.6×10^{-5} 1.9×10^{-5} 1.3×10^{-5} 2.1×10^{-5} 1.4×10^{-4} 1.5×10^{-4}	Calibrator/ SICT-CP-40101
Transconductance amplifiers	40102	(±) 10 μA ~ 10 A (10 ~ 50) A (50 ~ 100) A	1.3×10^{-5} 4.3×10^{-5} 4.4×10^{-5}	AC-DC Active Current Shunt/ SICT-CP-40102
		(10 μA) 10 Hz ~ 10 kHz	2.6×10^{-3}	
		(10 ~ 100) μA 10 Hz ~ 1 kHz (1 ~ 10) kHz	3.6×10^{-4} 6.4×10^{-4}	
		(100 μA ~ 1 mA) 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	9.8×10^{-5} 7.5×10^{-5} 9.4×10^{-5}	
		(1 ~ 100) mA 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	7.8×10^{-5} 4.6×10^{-5} 4.2×10^{-5}	
		(100 mA ~ 1 A) 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	8.1×10^{-5} 4.9×10^{-5} 4.4×10^{-5}	
		(1 ~ 2) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	7.9×10^{-5} 4.7×10^{-5} 4.5×10^{-5}	
		(2 ~ 5) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	8.2×10^{-5} 5.2×10^{-5} 5.0×10^{-5}	

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Transconductance amplifiers	40102	(5 ~ 10) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 20) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (20 ~ 50) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (50 ~ 100) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (100 ~ 120) A 60 Hz	8.6×10^{-5} 5.9×10^{-5} 7.8×10^{-5} 9.3×10^{-5} 6.8×10^{-5} 7.8×10^{-5} 1.0×10^{-4} 8.3×10^{-5} 1.1×10^{-4} 1.2×10^{-4} 9.7×10^{-5} 1.3×10^{-5} 4.6×10^{-4}	AC-DC Active Current Shunt/ SICT-CP-40102	
DC voltage/current calibrators	40103	(±) 0 mV (0 ~ 1) mV (1 ~ 2) mV (2 ~ 10) mV (10 ~ 50) mV (50 ~ 100) mV (0.1 ~ 1) V (1 ~ 1 000) V (±) 0 nA (0 ~ 1) nA (1 ~ 100) nA 100 nA ~ 1 A (1 ~ 10) A (10 ~ 100) A	0.24 μ V 4.0×10^{-4} 2.0×10^{-4} 8.1×10^{-5} 4.1×10^{-5} 1.0×10^{-5} 6.3×10^{-6} 2.6×10^{-6} 7.0 pA 7.0×10^{-3} 4.7×10^{-3} 1.3×10^{-5} 1.3×10^{-4} 4.4×10^{-5}	Reference Multimeter/ SICT-CP-40103	
Electrical temperature calibrators	40104	TEMPERATURE(SOURCE) T/C RTD	(-9.835 ~ 0.000) mV 0.000 mV (0.000 ~ 13.421) mV (13.421 ~ 37.006) mV (37.006 ~ 61.017) mV (61.017 ~ 76.373) mV 0.999 Ω (0.999 ~ 2.499) Ω (2.499 ~ 4.322) Ω (4.322 ~ 100.000) Ω (100.000 ~ 177.155) Ω (177.155 ~ 313.708) Ω (313.708 ~ 627.422) Ω (627.422 ~ 3 233.3) Ω	0.42 μ V 0.24 μ V 0.42 μ V 0.48 μ V 0.53 μ V 0.57 μ V 0.063 m Ω 3.0×10^{-5} 1.9×10^{-5} 9.9×10^{-6} 8.8×10^{-6} 1.1×10^{-5} 9.2×10^{-6} 1.1×10^{-5}	디지털 멀티미터/ SICT-CP-40104

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrical temperature calibrators				
DC Voltage(SOURCE)	40104	(±) 0 mV (1 ~ 2) mV (2 ~ 3) mV (3 ~ 4) mV (4 ~ 5) mV (5 ~ 10) mV (10 ~ 50) mV (50 ~ 100) mV (0.1 ~ 0.2) V (0.2 ~ 0.3) V (0.3 ~ 0.6) V (0.6 ~ 1) V (1 ~ 6) V (6 ~ 10) V (10 ~ 70) V (70 ~ 100) V	0.24 μ V 4.0×10^{-4} 2.0×10^{-4} 1.3×10^{-4} 1.0×10^{-4} 8.2×10^{-5} 4.2×10^{-5} 6.3×10^{-6} 6.2×10^{-5} 3.1×10^{-5} 3.1×10^{-5} 9.5×10^{-6} 3.1×10^{-5} 9.3×10^{-6} 6.2×10^{-5} 9.2×10^{-6}	디지털 멀티미터/ SICT-CP-40104
DC Current(SOURCE)		(±) 0 mA (0 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 7) mA (7 ~ 10) mA (10 ~ 20) mA (20 ~ 30) mA (30 ~ 40) mA (40 ~ 100) mA	0.064 μ A 9.2×10^{-5} 6.2×10^{-5} 3.5×10^{-5} 2.3×10^{-5} 1.9×10^{-5} 3.3×10^{-5} 8.2×10^{-5} 7.0×10^{-5} 6.3×10^{-5}	
Resistance(SOURCE)		0 Ω (0 ~ 0.6) Ω (0.6 ~ 1) Ω (1 ~ 10) Ω (10 ~ 20) Ω (20 ~ 30) Ω (30 ~ 50) Ω (50 ~ 70) Ω (70 ~ 100) Ω (0.1 ~ 0.2) k Ω (0.2 ~ 0.3) k Ω (0.3 ~ 0.5) k Ω (0.5 ~ 0.8) k Ω (0.8 ~ 1) k Ω (1 ~ 2) k Ω (2 ~ 3) k Ω (3 ~ 5) k Ω (5 ~ 8) k Ω (8 ~ 10) k Ω (10 ~ 20) k Ω (20 ~ 30) k Ω (30 ~ 40) k Ω (40 ~ 50) k Ω (50 ~ 100) k Ω	0.061 m Ω 6.1×10^{-4} 8.9×10^{-5} 6.7×10^{-5} 3.3×10^{-5} 2.3×10^{-5} 1.8×10^{-5} 1.4×10^{-5} 1.2×10^{-5} 3.2×10^{-5} 2.3×10^{-5} 1.8×10^{-5} 1.4×10^{-5} 1.1×10^{-5} 3.2×10^{-5} 2.3×10^{-5} 1.8×10^{-5} 1.4×10^{-5} 1.1×10^{-5} 3.2×10^{-5} 2.4×10^{-5} 1.9×10^{-5} 1.6×10^{-5} 1.1×10^{-5}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrical temperature calibrators	40104			디지털 멀티미터/ SICT-CP-40104
TEMPERATURE(MEASURE)				
T/C		(-9.835 ~ 0.000) mV 0.000 mV (0.000 ~ 13.422) mV (13.422 ~ 28.947) mV (28.947 ~ 45.094) mV (45.094 ~ 53.113) mV (53.113 ~ 76.374) mV	0.59 μ V 0.50 μ V 0.62 μ V 0.75 μ V 0.88 μ V 0.95 μ V 1.1 μ V	
RTD		0.998 Ω (0.998 ~ 2.496) Ω (2.496 ~ 4.315) Ω (4.315 ~ 16.994) Ω (16.994 ~ 249.580) Ω (249.580 ~ 317.988) Ω (317.988 ~ 390.474) Ω (390.474 ~ 3 233.2) Ω	0.24 $m\Omega$ 1.0×10^{-4} 7.1×10^{-5} 3.9×10^{-5} 3.5×10^{-5} 4.3×10^{-5} 4.0×10^{-5} 3.5×10^{-5}	
DC Voltage(MEASURE)		(±) 0 mV (1 ~ 5) mV (5 ~ 10) mV (10 ~ 100) mV (0.1 ~ 0.5) V (0.5 ~ 0.8) V (0.8 ~ 1) V (1 ~ 10) V (10 ~ 20) V (20 ~ 40) V (40 ~ 70) V (70 ~ 100) V (100 ~ 200) V (200 ~ 300) V	0.50 μ V 5.2×10^{-4} 9.3×10^{-5} 5.9×10^{-5} 6.3×10^{-5} 1.3×10^{-5} 1.6×10^{-5} 6.6×10^{-6} 9.1×10^{-6} 7.9×10^{-6} 6.9×10^{-6} 6.4×10^{-6} 7.8×10^{-6} 2.2×10^{-5}	
DC Current(MEASURE)		(±) 0 mA (0 ~ 0.1) mA (0.1 ~ 0.2) mA (0.2 ~ 0.3) mA (0.3 ~ 0.7) mA (0.7 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 10) mA (10 ~ 20) mA (20 ~ 30) mA (30 ~ 40) mA (40 ~ 50) mA (50 ~ 100) mA (100 ~ 130) mA	0.062 μ A 6.4×10^{-4} 3.2×10^{-4} 2.2×10^{-4} 1.7×10^{-4} 9.3×10^{-5} 9.9×10^{-5} 7.6×10^{-5} 5.8×10^{-5} 9.9×10^{-5} 8.2×10^{-5} 7.4×10^{-5} 7.0×10^{-5} 6.7×10^{-5} 8.7×10^{-5}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrical temperature calibrators				
Resistance(MEASURE)	40104	0 Ω (0 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) kΩ (1 ~ 10) kΩ (10 ~ 20) kΩ (20 ~ 30) kΩ (30 ~ 40) kΩ (40 ~ 50) kΩ (50 ~ 100) kΩ	0.098 mΩ 6.4×10^{-5} 1.1×10^{-5} 9.6×10^{-6} 6.5×10^{-5} 6.1×10^{-5} 4.7×10^{-5} 4.0×10^{-5} 4.2×10^{-5} 3.9×10^{-5} 3.4×10^{-5}	디지털 멀티미터/ SICT-CP-40104
DC current shunts				
Resistance	40105	1 μΩ (0.001 ~ 0.01) mΩ (0.01 ~ 0.2) mΩ (0.2 ~ 1) mΩ (1 ~ 10) mΩ (10 ~ 100) mΩ (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) kΩ (1 ~ 10) kΩ (10 ~ 100) kΩ (0.1 ~ 1) MΩ (1 ~ 10) MΩ (10 ~ 100) MΩ	0.32 nΩ 2.8×10^{-4} 1.8×10^{-4} 1.5×10^{-4} 1.1×10^{-6} 2.8×10^{-6} 8.1×10^{-7} 1.3×10^{-6} 6.7×10^{-7} 6.2×10^{-7} 7.9×10^{-7} 2.0×10^{-6} 1.4×10^{-6} 5.3×10^{-6} 9.8×10^{-6}	Trans Conductance Amplifier/ SICT-CP-40105
Galvanometers/null detectors				
DC Voltage	40106	(±) (100 ~ 300) μV (0.3 ~ 1) mV 1 mV ~ 1 000 V	1.4×10^{-2} 1.2×10^{-2} 6.8×10^{-3}	Calibrator/ SICT-CP-40106
Potentiometers				
DC Voltage	40107	(100 ~ 300) μV (0.3 ~ 1) mV (1 ~ 3) mV 3 mV ~ 1 000 V	5.7×10^{-3} 2.2×10^{-3} 6.0×10^{-4} 3.0×10^{-4}	Calibrator/ SICT-CP-40107

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC power supplies	40108	(±) 0 mV (0 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 600) V (600 ~ 1 000) V	5.8 μ V 5.8×10^{-4} 5.8×10^{-5} 7.5×10^{-6} 3.3×10^{-6} 7.7×10^{-6} 1.3×10^{-5} 6.6×10^{-5}	DC Electronics Load/ SICT-CP-40108
	DC Voltage	(1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A (10 ~ 300) A (300 ~ 500) A (500 ~ 1 000) A (1 000 ~ 3 000) A	5.8×10^{-3} 5.9×10^{-4} 2.4×10^{-4} 3.1×10^{-4} 2.4×10^{-4} 2.6×10^{-4} 4.7×10^{-5} 5.1×10^{-4}	
	DC Current	(0 ~ 2) mV (2 ~ 20) mV (20 ~ 200) mV	0.16 mV 7.8×10^{-2} 8.2×10^{-3}	
	Load regulation	(0.1 ~ 0.4) mV (0.4 ~ 0.6) mV (0.6 ~ 1) mV (1 ~ 10) mV (10 ~ 50) mV	3.8×10^{-1} 1.1×10^{-1} 7.3×10^{-2} 4.4×10^{-2} 7.1×10^{-2}	
	Ripple			
DC voltage dividers	40110	(±) (0.01 ~ 1 000) V (1 ~ 50) kV (50 ~ 100) kV	4.5×10^{-6} 8.8×10^{-5} 8.4×10^{-5}	Calibrator/ SICT-CP-40110
DC voltage standards	40111	1 V 1.018 V 10 V	1.6 μ V 0.8 μ V 3.1 μ V	Null Detector/ SICT-CP-40111
DC voltmeters	40112	(±) 0 mV (0 ~ 1) mV (1 ~ 2) mV (2 ~ 5) mV (5 ~ 10) mV (10 ~ 20) mV (20 ~ 50) mV (50 ~ 80) mV (80 ~ 100) mV (0.1 ~ 1 000) V	0.51 μ V 5.0×10^{-4} 2.5×10^{-4} 1.0×10^{-4} 6.2×10^{-5} 5.0×10^{-5} 2.5×10^{-5} 1.0×10^{-5} 6.2×10^{-6} 8.0×10^{-6}	Calibrator/ SICT-CP-40112
Static/Ionicvoltmeters	40113	(±) 0 V 0 V ~ 50 kV	68 mV 1.3×10^{-2}	DC Power Supply/ SICT-CP-40113

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Capacitance bridges/indicators	40201			Standard Capacitance Set/ SICT-CP-40201
Frequency		50 Hz ~ 100 MHz	7.0×10^{-8}	
Capacitance		(1 pF)		
		50 Hz ~ 1 kHz	3.5×10^{-4}	
		1 kHz ~ 1 MHz	3.6×10^{-4}	
		2 MHz	4.2×10^{-4}	
		3 MHz	5.4×10^{-4}	
		4 MHz	7.2×10^{-4}	
		5 MHz	9.4×10^{-4}	
		10 MHz	2.5×10^{-3}	
		13 MHz	3.6×10^{-3}	
		(10 pF)		
		50 Hz ~ 5 MHz	3.5×10^{-4}	
		10 MHz	3.7×10^{-4}	
		13 MHz	3.8×10^{-4}	
		(100 pF)		
		50 Hz ~ 3 MHz	3.5×10^{-4}	
		4 MHz	3.6×10^{-4}	
		5 MHz	3.7×10^{-4}	
		10 MHz	4.8×10^{-4}	
		13 MHz	6.0×10^{-4}	
		(1 000 pF)		
		50 Hz ~ 1 MHz	3.5×10^{-4}	
		2 MHz	3.8×10^{-4}	
		3 MHz	4.5×10^{-4}	
		4 MHz	5.6×10^{-4}	
		5 MHz	7.2×10^{-4}	
		10 MHz	2.0×10^{-3}	
		13 MHz	2.9×10^{-3}	
		(10 nF)		
		(50 ~ 100) Hz	3.0×10^{-4}	
		100 Hz ~ 100 kHz	8.1×10^{-5}	
		(100 nF)		
		(50 ~ 100) Hz	3.0×10^{-4}	
		100 Hz ~ 100 kHz	8.1×10^{-4}	
		(1 μF)		
		(50 ~ 100) Hz	5.1×10^{-4}	
		100 Hz ~ 10 kHz	8.1×10^{-5}	
		(10 ~ 100) kHz	1.0×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Capacitance bridges/indicators				
Capacitance	40201	(10 μ F) 120 Hz ~ 1 kHz	1.2×10^{-3}	Standard Capacitance Set/ SICT-CP-40201
		(100 μ F) 120 Hz	1.3×10^{-3}	
		(1 mF) 120 Hz	1.4×10^{-3}	
		(3 mF) 120 Hz	1.4×10^{-3}	
		(10 mF) 120 Hz	1.4×10^{-3}	
		(30 mF) 120 Hz	2.9×10^{-3}	
Decade capacitors				
Capacitance	40202	(50 Hz ~ 20 kHz) 1 pF (1 ~ 10) pF (10 ~ 100) pF (100 ~ 1 000) pF 1 000 pF ~ 100 nF 100 nF ~ 1 μ F	5.7×10^{-5} 4.6×10^{-5} 3.8×10^{-5} 4.6×10^{-5} 2.9×10^{-4} 5.1×10^{-4}	Standard Capacitance Set/ SICT-CP-40202
		(1 kHz) 1 pF (1 ~ 1 000) pF 1 000 pF ~ 100 nF 100 nF ~ 1 μ F	2.5×10^{-5} 2.4×10^{-5} 5.5×10^{-5} 9.3×10^{-5}	
Standard capacitors				
Capacitance	40204	(50 Hz ~ 20 kHz) 1 pF (1 ~ 10) pF (10 ~ 100) pF (100 ~ 1 000) pF 1 000 pF ~ 100 nF 100 nF ~ 1 μ F	5.2×10^{-5} 4.0×10^{-5} 3.0×10^{-5} 4.0×10^{-5} 2.9×10^{-4} 5.1×10^{-4}	Standard Capacitance Set/ Capacitance Bridge SICT-CP-40204
		(1 kHz) 1 pF (1 ~ 10) pF (10 ~ 100) pF (100 ~ 1 000) pF 1 000 pF ~ 100 nF 100 nF ~ 1 μ F	9.1×10^{-6} 7.1×10^{-6} 6.1×10^{-6} 7.1×10^{-6} 5.0×10^{-5} 9.0×10^{-5}	
		(1 pF) 1 kHz 1 kHz ~ 1 MHz (1 ~ 2) MHz (2 ~ 3) MHz (3 ~ 4) MHz (4 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz	2.4×10^{-4} 2.5×10^{-4} 3.3×10^{-4} 4.7×10^{-4} 6.7×10^{-4} 9.1×10^{-4} 2.5×10^{-3} 3.7×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Standard capacitors	40204	(1 ~ 10) pF 1 kHz ~ 3 MHz (3 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz (10 ~ 100) pF 1 kHz ~ 1 MHz (1 ~ 3) MHz (3 ~ 4) MHz (4 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz (100 pF ~ 1 nF) 1 kHz 1 kHz ~ 1 MHz (1 ~ 2) MHz (2 ~ 3) MHz (3 ~ 4) MHz (4 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz (1 ~ 100) nF 120 Hz ~ 100 kHz (100 nF ~ 1 μF) 120 Hz 120 Hz ~ 10 kHz (10 ~ 100) kHz (1 ~ 10) μF 120 Hz ~ 1 kHz (30 μF) 120 Hz (100 μF) 120 Hz (300 μF) 120 Hz (1 mF) 120 Hz (3 mF) 120 Hz (10 mF) 120 Hz (30 mF) 120 Hz	2.3 × 10 ⁻⁴ 2.4 × 10 ⁻⁴ 2.6 × 10 ⁻⁴ 2.8 × 10 ⁻⁴ 2.3 × 10 ⁻⁴ 2.4 × 10 ⁻⁴ 2.5 × 10 ⁻⁴ 2.7 × 10 ⁻⁴ 4.0 × 10 ⁻⁴ 5.4 × 10 ⁻⁴ 2.3 × 10 ⁻⁴ 2.4 × 10 ⁻⁴ 2.8 × 10 ⁻⁴ 3.6 × 10 ⁻⁴ 5.0 × 10 ⁻⁴ 6.6 × 10 ⁻⁴ 1.9 × 10 ⁻³ 2.8 × 10 ⁻³ 2.3 × 10 ⁻⁴ 2.4 × 10 ⁻⁴ 2.3 × 10 ⁻⁴ 2.4 × 10 ⁻⁴ 1.2 × 10 ⁻³ 1.3 × 10 ⁻³ 1.3 × 10 ⁻³ 1.5 × 10 ⁻³ 1.4 × 10 ⁻³ 1.5 × 10 ⁻³ 1.4 × 10 ⁻³ 2.9 × 10 ⁻³	Standard Capacitance Set/ Capacitance Bridge SICT-CP-40204

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Earth testers	40205			
Test Voltage		1 V (1 ~ 10) V (10 ~ 50) V (50 ~ 100) V (100 ~ 500) V (500 ~ 1 000) V	6.4×10^{-3} 6.4×10^{-4} 1.3×10^{-3} 6.4×10^{-4} 1.6×10^{-4} 6.4×10^{-4}	Decade Resistor/ SICT-CP-40205
Resistance		1 mΩ (1 ~ 10) mΩ 10 mΩ ~ 100 kΩ	8.6×10^{-4} 7.2×10^{-4} 6.8×10^{-4}	
AC Current out		1 A (1 ~ 3) A (3 ~ 20) A (20 ~ 30) A (30 ~ 60) A (60 ~ 100) A (100 ~ 150) A (150 ~ 200) A	1.2×10^{-3} 1.5×10^{-3} 9.7×10^{-4} 1.0×10^{-3} 8.4×10^{-4} 1.0×10^{-3} 4.6×10^{-3} 3.7×10^{-3}	
Timer		1 s (1 ~ 100) s (100 ~ 1 000) s (1 000 ~ 10 000) s	5.8×10^{-6} 5.8×10^{-6} 8.2×10^{-6} 5.8×10^{-5}	
Inductance bridges/indicators	40206			
Frequency		50 Hz ~ 100 MHz	7.0×10^{-8}	Standard Inductor/ SICT-CP-40206
Inductance		(1 kHz) 100 µH 1 mH 10 mH 100 mH 1 H 10 H	1.9×10^{-4} 1.3×10^{-4} 1.3×10^{-4} 1.3×10^{-4} 1.3×10^{-4} 1.3×10^{-4}	
Inductors	40208			
Standard Inductance		(1 kHz) 100 µH 1 mH 10 mH 100 mH 1 H 10 H	28 nH 0.24 µH 2.4 µH 24 µH 0.24 mH 2.5 mH	Standard Inductor/ SICT-CP-40208
Decade Inductance		(1 kHz) 100 µH ~ 10 H	3.5×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Insulation testers	40210			High Resistance Decade/ SICT-CP-40210
AC Voltage		1 V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V	8.8×10^{-5} 9.0×10^{-5} 1.0×10^{-4} 1.1×10^{-4}	
Insulation Voltage		1 V (1 ~ 10) V (10 ~ 25) V (25 ~ 50) V (50 ~ 100) V (100 ~ 250) V (250 ~ 500) V (500 ~ 1 000) V (1 000 ~ 5 000) V (5 000 ~ 10 000) V	6.4×10^{-4} 6.4×10^{-5} 2.5×10^{-4} 1.3×10^{-4} 6.4×10^{-5} 2.5×10^{-4} 1.3×10^{-4} 6.4×10^{-5} 6.5×10^{-3} 6.1×10^{-3}	
Insulation Resistance		1 kΩ (1 ~ 10) kΩ (10 ~ 100) kΩ (0.1 ~ 1) MΩ (1 ~ 10) MΩ (10 ~ 100) MΩ (0.1 ~ 1) GΩ (1 ~ 10) GΩ (10 ~ 100) GΩ (0.1 ~ 1) TΩ 10 TΩ	7.1×10^{-5} 3.7×10^{-5} 2.5×10^{-5} 3.1×10^{-5} 9.5×10^{-5} 2.4×10^{-5} 3.1×10^{-5} 6.1×10^{-5} 1.3×10^{-4} 2.6×10^{-4} 6.3×10^{-4}	
Q-meters	40211	60 Hz ~ 100 MHz 0 ~ 1 000	7.0×10^{-8} 9.7×10^{-4}	Universal Counter/ SICT-CP-40211
Resistance bridges & similar instruments	40213			Standard Resistance Set/ SICT-CP-40213
Resistance(Rheostat Arm)		1 mΩ (1 ~ 10) mΩ (10 ~ 100) mΩ (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) kΩ (1 ~ 10) kΩ (10 ~ 100) kΩ (0.1 ~ 1) MΩ (1 ~ 10) MΩ (10 ~ 100) MΩ	1.5×10^{-6} 7.3×10^{-7} 2.7×10^{-6} 5.6×10^{-7} 1.1×10^{-6} 3.4×10^{-7} 2.2×10^{-7} 5.4×10^{-7} 2.0×10^{-6} 1.3×10^{-6} 5.3×10^{-6} 9.8×10^{-6}	
Resistance(Ratio Arm)		1 mΩ (1 ~ 10) mΩ (10 ~ 100) mΩ (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) kΩ (1 ~ 10) kΩ (10 ~ 100) kΩ (0.1 ~ 1) MΩ (1 ~ 10) MΩ (10 ~ 100) MΩ	1.5×10^{-6} 7.3×10^{-7} 2.7×10^{-6} 5.6×10^{-7} 1.1×10^{-6} 3.4×10^{-7} 2.2×10^{-7} 5.4×10^{-7} 2.0×10^{-6} 1.3×10^{-6} 5.3×10^{-6} 9.8×10^{-6}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistance meters	40214			
DC Resistance		10 $\mu\Omega$ 100 $\mu\Omega$ 1 m Ω (1 ~ 10) m Ω (10 ~ 100) m Ω (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) k Ω (1 ~ 10) k Ω (10 ~ 100) k Ω (0.1 ~ 1) M Ω (1 ~ 10) M Ω (10 ~ 100) M Ω (0.1 ~ 1) G Ω (1 ~ 10) G Ω (10 ~ 100) G Ω (0.1 ~ 1) T Ω 10 T Ω	3.3×10^{-4} 4.7×10^{-4} 1.3×10^{-6} 8.0×10^{-7} 2.7×10^{-6} 5.6×10^{-7} 1.1×10^{-6} 3.3×10^{-7} 2.2×10^{-7} 5.3×10^{-7} 1.9×10^{-6} 1.2×10^{-6} 5.2×10^{-6} 9.8×10^{-6} 3.2×10^{-5} 6.2×10^{-5} 1.3×10^{-4} 2.7×10^{-4} 6.4×10^{-4}	Standard Resistance Set/ SICT-CP-40214
Frequency		1 kHz	7.0×10^{-8}	
AC Voltage		10 mV (10 ~ 100) mV (0.1 ~ 10) V	1.6×10^{-4} 7.9×10^{-5} 8.2×10^{-5}	
AC Resistance		(1 kHz) 1 m Ω (1 ~ 10) m Ω (10 ~ 100) m Ω 100 m Ω ~ 10 k Ω (10 ~ 100) k Ω (0.1 ~ 1) M Ω (1 ~ 10) M Ω	5.0×10^{-3} 5.2×10^{-4} 3.3×10^{-4} 1.3×10^{-4} 1.4×10^{-4} 3.0×10^{-4} 2.9×10^{-3}	
Resistors	40215			
DC Resistance		1 m Ω (1 ~ 10) m Ω (10 ~ 100) m Ω (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) k Ω (1 ~ 10) k Ω (10 ~ 100) k Ω (0.1 ~ 1) M Ω (1 ~ 10) M Ω (10 ~ 100) M Ω (0.1 ~ 1) G Ω (1 ~ 10) G Ω (10 ~ 100) G Ω (0.1 ~ 1) T Ω (1 ~ 10) T Ω (10 ~ 100) T Ω	1.6×10^{-6} 1.1×10^{-6} 2.8×10^{-6} 8.1×10^{-7} 1.3×10^{-6} 6.7×10^{-7} 6.2×10^{-7} 7.9×10^{-7} 2.0×10^{-6} 1.4×10^{-6} 5.2×10^{-6} 9.7×10^{-6} 2.3×10^{-4} 6.9×10^{-4} 9.3×10^{-4} 1.4×10^{-3} 4.1×10^{-3} 7.6×10^{-3}	Standard Resistance Set/ SICT-CP-40215

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistors	40215	(50 Hz ~ 1 kHz) 1 mΩ (1 ~ 10) mΩ (10 ~ 100) mΩ (0.1 ~ 1) Ω (1 ~ 100) Ω (10 Ω) 1 kHz 1 MHz 2 MHz 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz (10 ~ 100) Ω 1 kHz 1 MHz 2 MHz 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz (100 Ω ~ 1 kΩ) 1 kHz 100 kHz 1 MHz 2 MHz 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz (1 ~ 10) kΩ 1 kHz 100 kHz 1 MHz (10 ~ 100) kΩ 1 kHz 100 kHz 1 MHz (100 kΩ ~ 1 MΩ) 1 kHz (1 ~ 10) MΩ 1 kHz	1.0×10^{-3} 5.9×10^{-4} 3.9×10^{-4} 2.4×10^{-4} 2.5×10^{-4} 2.4×10^{-4} 4.0×10^{-4} 5.6×10^{-4} 6.5×10^{-4} 7.5×10^{-4} 1.0×10^{-3} 4.0×10^{-3} 6.0×10^{-3} 2.4×10^{-4} 4.0×10^{-4} 4.8×10^{-4} 5.6×10^{-4} 5.6×10^{-4} 5.6×10^{-4} 2.0×10^{-3} 3.0×10^{-3} 2.4×10^{-4} 4.0×10^{-4} 4.0×10^{-4} 4.0×10^{-4} 4.0×10^{-4} 4.8×10^{-4} 5.6×10^{-4} 2.0×10^{-3} 3.0×10^{-3} 2.4×10^{-4} 3.3×10^{-4} 4.0×10^{-4} 2.4×10^{-4} 4.0×10^{-4} 4.0×10^{-4} 3.8×10^{-4} 3.0×10^{-3}	Standard Resistance Set/ SICT-CP-40215

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistors	40215	0 Ω	64 μΩ	Standard Resistance Set/ SICT-CP-40215
		(0 ~ 10) mΩ	6.5 μΩ	
		(10 ~ 100) mΩ	64 μΩ	
		(0.1 ~ 1) Ω	66 μΩ	
		(1 ~ 7) Ω	3.9×10^{-5}	
		(7 ~ 10) Ω	1.3×10^{-5}	
		(10 ~ 70) Ω	2.0×10^{-5}	
		(70 ~ 100) Ω	9.8×10^{-6}	
		(100 ~ 700) Ω	1.9×10^{-5}	
		(0.7 ~ 1) kΩ	9.6×10^{-6}	
		(1 ~ 7) kΩ	3.5×10^{-5}	
		(7 ~ 10) kΩ	1.2×10^{-5}	
		(10 ~ 70) kΩ	2.0×10^{-5}	
		(70 ~ 100) kΩ	9.8×10^{-6}	
		(100 ~ 600) kΩ	2.9×10^{-5}	
		(0.6 ~ 1) MΩ	1.2×10^{-5}	
		(1 ~ 7) MΩ	8.0×10^{-5}	
		(7 ~ 10) MΩ	2.7×10^{-5}	
		(10 ~ 100) MΩ	2.0×10^{-4}	
		(0.1 ~ 1) GΩ	2.5×10^{-4}	
		(1 ~ 10) GΩ	7.0×10^{-4}	
		(10 ~ 100) GΩ	1.0×10^{-3}	
		(0.1 ~ 1) TΩ	1.5×10^{-3}	
		(1 ~ 10) TΩ	4.3×10^{-3}	
Impedance bridges/LCR meters	40217	50 Hz ~ 100 MHz	7.0×10^{-8}	Standard Capacitor Set, Standard Resistor Set, Standard Inductor/ SICT-CP-40217
		Frequency		
		1 mV	2.1×10^{-3}	
		(1 ~ 10) mV	4.4×10^{-4}	
		(10 ~ 100) mV	8.8×10^{-5}	
		(0.1 ~ 10) V	8.2×10^{-5}	
		(10 ~ 100) V	8.9×10^{-5}	
		AC Voltage		
		(1 pF)		
		50 Hz ~ 1 kHz	3.5×10^{-4}	
		1 kHz ~ 1 MHz	3.6×10^{-4}	
		2 MHz	4.2×10^{-4}	
		3 MHz	5.4×10^{-4}	
		4 MHz	7.2×10^{-4}	
		5 MHz	9.4×10^{-4}	
		10 MHz	2.5×10^{-3}	
		13 MHz	3.6×10^{-3}	
		Capacitance		
		(10 pF)		
		50 Hz ~ 5 MHz	3.5×10^{-4}	
		10 MHz	3.7×10^{-4}	
		13 MHz	3.8×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Impedance bridges/LCR meters	40217	(100 pF) 50 Hz ~ 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz	3.5×10^{-4} 3.6×10^{-4} 3.7×10^{-4} 4.8×10^{-4} 6.0×10^{-4}	Standard Capacitor Set, Standard Resistor Set, Standard Inductor/ SICT-CP-40217

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Impedance bridges/LCR meters				
Capacitance	40217	(10 pF) 1 kHz ~ 13 MHz	0.000 15	Standard Capacitor Set, Standard Resistor Set, Standard Inductor/ SICT-CP-40217
		(100 pF) 1 kHz ~ 5 MHz 5 MHz ~ 13 MHz	0.000 13 0.000 27	
Dissipation Factor		(1 pF) 1 kHz ~ 1 MHz (1 ~ 5) MHz (5 ~ 13) MHz	0.000 12 0.000 23 0.000 84	
		(10 pF) 1 kHz ~ 13 MHz	0.000 15	
		(100 pF) 1 kHz ~ 5 MHz (5 ~ 13) MHz	0.000 13 0.000 27	
		(1 000 pF) 1 kHz ~ 1 MHz (1 ~ 5) MHz (5 ~ 13) MHz	0.000 12 0.000 24 0.000 86	
		(10 nF) 120 Hz ~ 100 kHz	0.000 46	
		(100 nF) 120 Hz ~ 100 kHz	0.000 58	
		(1 μF) 120 Hz ~ 100 kHz	0.000 81	
Resistance		(1 mΩ) 50 Hz 50 Hz ~ 1 kHz	6.0 × 10 ⁻³ 5.0 × 10 ⁻³	
		(10 mΩ) 50 Hz 50 Hz ~ 1 kHz	1.0 × 10 ⁻³ 5.2 × 10 ⁻⁴	
		(100 mΩ) 50 Hz 50 Hz ~ 1 kHz	7.1 × 10 ⁻⁴ 3.3 × 10 ⁻⁴	
		(1 Ω) 50 Hz 50 Hz ~ 1 kHz	6.8 × 10 ⁻⁴ 1.3 × 10 ⁻⁴	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Impedance bridges/LCR meters	40217			
Resistance		(10 Ω)		Standard Capacitor Set, Standard Resistor Set,
		50 Hz	9.1×10^{-4}	Standard Inductor/ SICT-CP-40217
		50 Hz ~ 1 kHz	1.3×10^{-4}	
		1 kHz ~ 1 MHz	3.3×10^{-4}	
		(1 ~ 2) MHz	5.2×10^{-4}	
		(2 ~ 3) MHz	6.1×10^{-4}	
		(3 ~ 4) MHz	7.1×10^{-4}	
		(4 ~ 5) MHz	1.0×10^{-3}	
		(5 ~ 10) MHz	4.0×10^{-3}	
		(10 ~ 13) MHz	6.0×10^{-3}	
		(100 Ω)		
		50 Hz	6.2×10^{-4}	
		50 Hz ~ 1 kHz	1.3×10^{-4}	
		1 kHz ~ 1 MHz	3.3×10^{-4}	
		(1 ~ 2) MHz	4.2×10^{-4}	
		(2 ~ 5) MHz	5.2×10^{-4}	
		(5 ~ 10) MHz	2.0×10^{-3}	
		(10 ~ 13) MHz	3.0×10^{-3}	
		(1 kΩ)		
		1 kHz	1.3×10^{-4}	
		1 kHz ~ 3 MHz	3.3×10^{-4}	
		(3 ~ 4) MHz	4.2×10^{-4}	
		(4 ~ 5) MHz	5.2×10^{-4}	
		(5 ~ 10) MHz	2.0×10^{-3}	
		(10 ~ 13) MHz	3.0×10^{-3}	
		(10 kΩ)		
		1 kHz	1.3×10^{-4}	
		(1 ~ 100) kHz	2.4×10^{-4}	
		100 kHz ~ 1 MHz	3.3×10^{-4}	
		(100 kΩ)		
		1 kHz	1.4×10^{-4}	
		100 kHz ~ 1 MHz	3.3×10^{-4}	
		(1 MΩ)		
		1 kHz	3.0×10^{-4}	
		(10 MΩ)		
		1 kHz	2.9×10^{-3}	
Inductance		(1 kHz)		
		100 μH	1.9×10^{-4}	
		1 mH	1.3×10^{-4}	
		10 mH	1.3×10^{-4}	
		100 mH	1.3×10^{-4}	
		1 H	1.3×10^{-4}	
		10 H	1.3×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Impedance bridges/LCR meters	DC Bias	40217		
		(±)		Standard Capacitor Set,
		0 μV	0.68 μV	Standard Resistor Set,
		0 μV ~ 100 mV	1.1×10^{-5}	Standard Inductor/
		(0.1 ~ 1) V	7.5×10^{-6}	SICT-CP-40217
	DC Current	(1 ~ 10) V	7.2×10^{-6}	
		(10 ~ 100) V	8.2×10^{-6}	
		0 μA	5.8 μA	
		0 μA ~ 200 mA	4.1×10^{-5}	
		(0.2 ~ 2) A	3.6×10^{-5}	
		(2 ~ 20) A	1.9×10^{-4}	
		(20 ~ 100) A	1.4×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC ammeters				
AC Current	40301	(100 μ A) 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	2.2×10^{-4} 4.9×10^{-4} 2.1×10^{-3}	Power Calibrator, Calibrator/ SICT-CP-40301
		(100 μ A ~ 1 mA) 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	1.8×10^{-4} 3.8×10^{-4} 2.1×10^{-3}	
		(1 ~ 10) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	1.8×10^{-4} 3.4×10^{-4} 1.9×10^{-3}	
		(10 ~ 100) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	1.7×10^{-4} 3.2×10^{-4} 1.5×10^{-3}	
		(100 mA ~ 1 A) 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	3.5×10^{-4} 6.7×10^{-4} 8.3×10^{-3}	
		(1 ~ 10) A (40 ~ 100) Hz 100 Hz ~ 5 kHz	2.1×10^{-4} 5.8×10^{-4}	
		(10 ~ 20) A (40 ~ 100) Hz 100 Hz ~ 5 kHz	2.0×10^{-4} 5.3×10^{-4}	
		(20 ~ 100) A (40 ~ 100) Hz 100 Hz ~ 5 kHz	1.9×10^{-4} 5.9×10^{-4}	
		(100 ~ 200) A 60 Hz	8.5×10^{-4}	
Clamp ammeters/voltmeters				
AC Current	40302	(100 μ A) 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	6.5×10^{-4} 7.8×10^{-4} 2.1×10^{-3}	Power Calibrator, Calibrator/ SICT-CP-40302
		(100 ~ 300) μ A 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	3.8×10^{-4} 7.9×10^{-4} 4.0×10^{-3}	
		(300 ~ 900) μ A 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	3.1×10^{-4} 6.4×10^{-4} 3.3×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters	40302	(900 μ A ~ 1 mA) 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (1 ~ 3) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (3 ~ 9) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (9 ~ 10) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (10 ~ 30) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (30 ~ 90) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (90 ~ 100) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (100 ~ 300) mA 40 Hz 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (300 ~ 900) mA 40 Hz 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (900 mA ~ 1 A) 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	6.3 \times 10 ⁻⁴ 7.2 \times 10 ⁻⁴ 2.1 \times 10 ⁻³ 3.8 \times 10 ⁻⁴ 7.2 \times 10 ⁻⁴ 3.8 \times 10 ⁻³ 3.1 \times 10 ⁻⁴ 5.8 \times 10 ⁻⁴ 3.1 \times 10 ⁻³ 6.3 \times 10 ⁻⁴ 7.0 \times 10 ⁻⁴ 2.0 \times 10 ⁻³ 3.6 \times 10 ⁻⁴ 6.8 \times 10 ⁻⁴ 2.6 \times 10 ⁻³ 2.9 \times 10 ⁻⁴ 5.4 \times 10 ⁻⁴ 2.2 \times 10 ⁻³ 6.3 \times 10 ⁻⁴ 6.9 \times 10 ⁻⁴ 1.6 \times 10 ⁻³ 3.3 \times 10 ⁻⁴ 5.7 \times 10 ⁻⁴ 1.2 \times 10 ⁻³ 9.3 \times 10 ⁻³ 5.1 \times 10 ⁻⁴ 4.8 \times 10 ⁻⁴ 1.0 \times 10 ⁻³ 8.9 \times 10 ⁻³ 7.0 \times 10 ⁻⁴ 9.0 \times 10 ⁻⁴ 5.1 \times 10 ⁻³	Power Calibrator, Calibrator/ SICT-CP-40302

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters				
		40302		Power Calibrator, Calibrator/ SICT-CP-40302
AC Current				
		(1 ~ 2) A		
		40 Hz ~ 1 kHz	4.3×10^{-4}	
		(1 ~ 5) kHz	6.6×10^{-4}	
		(5 ~ 10) kHz	4.6×10^{-3}	
		(2 ~ 3) A		
		(40 ~ 100) Hz	5.7×10^{-4}	
		100 Hz ~ 5 kHz	9.7×10^{-4}	
		(5 ~ 10) kHz	4.5×10^{-3}	
		(3 ~ 9) A		
		(40 ~ 100) Hz	4.4×10^{-4}	
		100 Hz ~ 5 kHz	8.2×10^{-4}	
		(5 ~ 10) kHz	4.4×10^{-3}	
		(9 ~ 10) A		
		(40 ~ 100) Hz	6.4×10^{-4}	
		100 Hz ~ 5 kHz	8.5×10^{-4}	
		(5 ~ 10) kHz	4.3×10^{-3}	
		(10 ~ 30) A		
		(40 ~ 100) Hz	3.6×10^{-4}	
		100 Hz ~ 5 kHz	7.0×10^{-4}	
		(5 ~ 10) kHz	6.2×10^{-2}	
		(30 ~ 90) A		
		(40 ~ 100) Hz	2.5×10^{-4}	
		100 Hz ~ 5 kHz	6.4×10^{-4}	
		(5 ~ 10) kHz	5.5×10^{-2}	
		(90 ~ 100) A		
		(40 ~ 100) Hz	1.8×10^{-4}	
		100 Hz ~ 5 kHz	5.4×10^{-4}	
		(5 ~ 10) kHz	4.3×10^{-2}	
		(100 ~ 1 000) A		
		40 Hz	1.3×10^{-3}	
		40 Hz ~ 1 kHz	2.5×10^{-3}	
		(1 000 ~ 2 500) A		
		(40 ~ 60) Hz	1.3×10^{-3}	
		(2 500 ~ 3 000) A		
		60 Hz	1.3×10^{-3}	
		(3 000 ~ 10 000) A		
		60 Hz	3.6×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters	40302	0 μA	9.2 nA	Power Calibrator, Calibrator/ SICT-CP-40302
		(0 ~ 0.1) μA	9.2×10^{-2}	
		(0.1 ~ 0.2) μA	4.6×10^{-2}	
		(0.2 ~ 0.3) μA	3.1×10^{-2}	
		(0.3 ~ 0.9) μA	2.3×10^{-2}	
		(0.9 ~ 1) μA	9.3×10^{-3}	
		(1 ~ 2) μA	4.7×10^{-3}	
		(2 ~ 3) μA	3.1×10^{-3}	
		(3 ~ 9) μA	2.3×10^{-3}	
		(9 ~ 10) μA	9.6×10^{-4}	
		(10 ~ 20) μA	5.1×10^{-4}	
		(20 ~ 30) μA	3.5×10^{-4}	
		(30 ~ 50) μA	2.7×10^{-4}	
		(50 ~ 90) μA	1.9×10^{-4}	
		(90 ~ 100) μA	6.2×10^{-4}	
		(100 ~ 200) μA	3.2×10^{-4}	
		(200 ~ 700) μA	2.2×10^{-4}	
		(700 ~ 900) μA	9.3×10^{-5}	
		(0.9 ~ 1) mA	6.1×10^{-4}	
		(1 ~ 2) mA	3.1×10^{-4}	
		(2 ~ 7) mA	2.2×10^{-4}	
		(7 ~ 9) mA	9.1×10^{-5}	
		(9 ~ 10) mA	6.1×10^{-4}	
		(10 ~ 20) mA	3.1×10^{-4}	
		(20 ~ 70) mA	2.2×10^{-4}	
		(70 ~ 90) mA	9.9×10^{-5}	
		(90 ~ 100) mA	6.1×10^{-4}	
		(100 ~ 200) mA	3.1×10^{-4}	
		(200 ~ 700) mA	2.5×10^{-4}	
		(700 ~ 900) mA	1.3×10^{-4}	
		(0.9 ~ 1) A	6.4×10^{-4}	
		(1 ~ 2) A	3.4×10^{-4}	
		(2 ~ 3) A	4.5×10^{-4}	
		(3 ~ 7) A	3.6×10^{-4}	
		(7 ~ 9) A	2.2×10^{-4}	
		(9 ~ 10) A	6.4×10^{-4}	
		(10 ~ 30) A	3.4×10^{-4}	
		(30 ~ 70) A	2.6×10^{-4}	
		(70 ~ 100) A	1.7×10^{-4}	
		(100 ~ 2 500) A	1.3×10^{-3}	
AC Voltage		(1 mV)		
		40 Hz ~ 10 kHz	4.8×10^{-3}	
		(10 ~ 50) kHz	5.0×10^{-3}	
		(50 ~ 100) kHz	6.5×10^{-3}	
		(1 ~ 2) mV		
		40 Hz ~ 10 kHz	2.4×10^{-3}	
		(10 ~ 50) kHz	2.6×10^{-3}	
		(50 ~ 100) kHz	3.5×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters	40302			
	AC Voltage	(2 ~ 5) mV	1.7×10^{-3}	Power Calibrator, Calibrator/ SICT-CP-40302
		40 Hz ~ 10 kHz	1.9×10^{-3}	
		(10 ~ 50) kHz	2.7×10^{-3}	
		(50 ~ 100) kHz		
		(5 ~ 7) mV	8.9×10^{-4}	
		40 Hz ~ 10 kHz	1.0×10^{-3}	
		(10 ~ 50) kHz	1.6×10^{-3}	
		(50 ~ 100) kHz		
		(7 ~ 9) mV	6.9×10^{-4}	
		40 Hz ~ 10 kHz	8.4×10^{-4}	
		(10 ~ 50) kHz	1.4×10^{-3}	
		(50 ~ 100) kHz		
		(9 ~ 10) mV	8.3×10^{-4}	
		40 Hz ~ 10 kHz	9.4×10^{-4}	
		(10 ~ 50) kHz	1.3×10^{-3}	
		(50 ~ 100) kHz		
		(10 ~ 30) mV	4.5×10^{-4}	
		40 Hz ~ 10 kHz	5.6×10^{-4}	
		(10 ~ 50) kHz	1.1×10^{-3}	
		(50 ~ 100) kHz		
		(30 ~ 50) mV	3.2×10^{-4}	
		40 Hz ~ 10 kHz	4.0×10^{-4}	
		(10 ~ 50) kHz	9.0×10^{-4}	
		(50 ~ 100) kHz		
		(50 ~ 70) mV	2.3×10^{-4}	
		40 Hz ~ 10 kHz	3.1×10^{-4}	
		(10 ~ 50) kHz	7.1×10^{-4}	
		(50 ~ 100) kHz		
		(70 ~ 90) mV	1.9×10^{-4}	
		40 Hz ~ 10 kHz	2.6×10^{-4}	
		(10 ~ 50) kHz	6.2×10^{-4}	
		(50 ~ 100) kHz		
		(90 ~ 100) mV	1.6×10^{-4}	
		40 Hz ~ 10 kHz	2.4×10^{-4}	
		(10 ~ 50) kHz	5.7×10^{-4}	
		(50 ~ 100) kHz		
		(100 ~ 200) mV	1.1×10^{-4}	
		40 Hz ~ 10 kHz	1.8×10^{-4}	
		(10 ~ 50) kHz	4.6×10^{-4}	
		(50 ~ 100) kHz		

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters	40302	(200 ~ 500) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	8.8×10^{-5} 1.4×10^{-4} 2.2×10^{-4}	Power Calibrator, Calibrator/ SICT-CP-40302

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters	40302	(20 ~ 50) V		Power Calibrator, Calibrator/ SICT-CP-40302
		40 Hz ~ 10 kHz	1.6×10^{-4}	
		(10 ~ 50) kHz	1.7×10^{-4}	
		(50 ~ 100) kHz	3.4×10^{-4}	
		(50 ~ 70) V		
		40 Hz	9.8×10^{-5}	
		40 Hz ~ 10 kHz	8.3×10^{-5}	
		(10 ~ 50) kHz	1.2×10^{-4}	
		(50 ~ 100) kHz	2.4×10^{-4}	
		(70 ~ 90) V		
		40 Hz	8.5×10^{-5}	
		40 Hz ~ 10 kHz	7.5×10^{-5}	
		(10 ~ 50) kHz	1.1×10^{-4}	
		(50 ~ 100) kHz	2.2×10^{-4}	
		(90 ~ 100) V		
		40 Hz ~ 10 kHz	9.9×10^{-5}	
		(10 ~ 50) kHz	1.2×10^{-4}	
		(50 ~ 100) kHz	2.2×10^{-4}	
		(100 ~ 200) V		
		40 Hz ~ 10 kHz	7.3×10^{-5}	
		(10 ~ 50) kHz	1.0×10^{-4}	
		(50 ~ 100) kHz	1.9×10^{-4}	
DC Voltage		(200 ~ 500) V		
		40 Hz ~ 1 kHz	1.3×10^{-4}	
		(1 ~ 10) kHz	2.2×10^{-4}	
		(10 ~ 20) kHz	6.6×10^{-4}	
		(500 ~ 1 000) V		
		40 Hz ~ 1 kHz	1.5×10^{-4}	
		(1 ~ 10) kHz	2.2×10^{-4}	
		(10 ~ 20) kHz	9.9×10^{-4}	
		0 mV	$61 \mu\text{V}$	
		(0 ~ 10) mV	6.1×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters	40302	(3 ~ 6) V	1.6×10^{-5}	Power Calibrator, Calibrator/ SICT-CP-40302
		(6 ~ 7) V	9.8×10^{-6}	
		(7 ~ 9) V	8.8×10^{-6}	
		(9 ~ 10) V	6.1×10^{-5}	
		(10 ~ 20) V	3.1×10^{-5}	
		(20 ~ 30) V	2.2×10^{-5}	
		(30 ~ 80) V	1.7×10^{-5}	
		(80 ~ 90) V	9.3×10^{-6}	
		(90 ~ 100) V	6.1×10^{-5}	
		(100 ~ 200) V	3.2×10^{-5}	
		(200 ~ 300) V	2.3×10^{-5}	
		(300 ~ 500) V	1.8×10^{-5}	
		(500 ~ 900) V	1.3×10^{-5}	
		(900 ~ 1 000) V	6.2×10^{-5}	
	Resistance	0 Ω	0.61 mΩ	
		(0 ~ 9) Ω	0.66 mΩ	
		(9 ~ 100) Ω	6.2 mΩ	
		(100 ~ 900) Ω	9.2 mΩ	
		(0.9 ~ 9) kΩ	92 mΩ	
		(9 ~ 90) kΩ	1.1 Ω	
		(0.090 ~ 1) MΩ	63 Ω	
		(1 ~ 10) MΩ	0.77 kΩ	
		(10 ~ 100) MΩ	13 kΩ	
AC voltage/current calibrators	40303	(1 mV)		Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
		10 Hz ~ 10 kHz	1.7×10^{-3}	
		(10 ~ 100) kHz	3.0×10^{-3}	
		100 kHz ~ 1 MHz	1.2×10^{-2}	
		(1 ~ 2) mV		
		10 Hz ~ 10 kHz	1.1×10^{-3}	
		(10 ~ 100) kHz	1.7×10^{-3}	
		100 kHz ~ 1 MHz	7.7×10^{-3}	
		(2 ~ 5) mV		
		10 Hz	6.4×10^{-4}	
		10 Hz ~ 10 kHz	5.8×10^{-4}	
		(10 ~ 100) kHz	1.0×10^{-3}	
		100 kHz ~ 1 MHz	5.4×10^{-3}	
		(5 ~ 10) mV		
		10 Hz	4.2×10^{-4}	
		10 Hz ~ 10 kHz	3.5×10^{-4}	
		(10 ~ 100) kHz	5.8×10^{-4}	
		100 kHz ~ 1 MHz	3.9×10^{-3}	
		(10 ~ 20) mV		
		10 Hz	1.8×10^{-4}	
		10 Hz ~ 10 kHz	1.4×10^{-4}	
		(10 ~ 100) kHz	2.2×10^{-4}	
		100 kHz ~ 1 MHz	2.2×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators	40303			Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
AC Voltage		(20 ~ 50) mV	1.4×10^{-4}	
		10 Hz	9.2×10^{-5}	
		10 Hz ~ 10 kHz	1.6×10^{-4}	
		(10 ~ 100) kHz	1.4×10^{-3}	
		100 kHz ~ 1 MHz		
		(50 ~ 100) mV		
		10 Hz	1.1×10^{-4}	
		10 Hz ~ 10 kHz	6.6×10^{-5}	
		(10 ~ 100) kHz	1.2×10^{-4}	
		100 kHz ~ 1 MHz	1.3×10^{-3}	
		(100 ~ 200) mV		
		10 Hz	8.2×10^{-5}	
		10 Hz ~ 10 kHz	3.9×10^{-5}	
		(10 ~ 100) kHz	7.6×10^{-5}	
		100 kHz ~ 1 MHz	1.1×10^{-3}	
		(200 ~ 500) mV		
		10 Hz	7.8×10^{-5}	
		10 Hz ~ 10 kHz	3.6×10^{-5}	
		(10 ~ 100) kHz	7.1×10^{-5}	
		100 kHz ~ 1 MHz	1.1×10^{-3}	
		(500 mV ~ 1 V)		
		10 Hz	7.6×10^{-5}	
		10 Hz ~ 10 kHz	3.3×10^{-5}	
		(10 ~ 100) kHz	6.6×10^{-5}	
		100 kHz ~ 1 MHz	1.1×10^{-3}	
		(1 ~ 2) V		
		10 Hz	7.1×10^{-5}	
		10 Hz ~ 10 kHz	2.7×10^{-5}	
		(10 ~ 100) kHz	5.8×10^{-5}	
		100 kHz ~ 1 MHz	1.0×10^{-3}	
		(2 ~ 5) V		
		10 Hz	7.2×10^{-5}	
		10 Hz ~ 10 kHz	2.6×10^{-5}	
		10 kHz ~ 100 kHz	7.5×10^{-5}	
		100 kHz ~ 1 MHz	1.4×10^{-3}	
		(5 ~ 20) V		
		10 Hz	7.2×10^{-5}	
		10 Hz ~ 10 kHz	2.8×10^{-5}	
		(10 ~ 100) kHz	7.5×10^{-5}	
		100 kHz ~ 1 MHz	1.4×10^{-3}	
		(20 ~ 50) V		
		10 Hz	7.2×10^{-5}	
		10 Hz ~ 10 kHz	3.0×10^{-5}	
		(10 ~ 100) kHz	8.0×10^{-5}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators	40303			Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
AC Voltage		(50 ~ 200) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	7.4×10^{-5} 3.3×10^{-5} 8.5×10^{-5}	
		(200 ~ 1 000) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	7.7×10^{-5} 3.3×10^{-5} 5.8×10^{-4}	
AC Current		(10 μA) 10 Hz ~ 10 kHz	2.6×10^{-3}	
		(10 ~ 100) μA 10 Hz ~ 1 kHz (1 ~ 10) kHz	3.6×10^{-4} 6.4×10^{-4}	
		(100 μA ~ 1 mA) 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	9.8×10^{-5} 7.5×10^{-5} 9.4×10^{-5}	
		(1 ~ 100) mA 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	7.8×10^{-5} 4.6×10^{-5} 4.2×10^{-5}	
		(100 mA ~ 1 A) 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	8.1×10^{-5} 4.9×10^{-5} 4.4×10^{-5}	
		(1 ~ 2) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	7.9×10^{-5} 4.7×10^{-5} 4.5×10^{-5}	
		(2 ~ 5) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	8.2×10^{-5} 5.2×10^{-5} 5.0×10^{-5}	
		(5 ~ 10) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	8.6×10^{-5} 5.9×10^{-5} 7.8×10^{-5}	
		(10 ~ 20) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	9.3×10^{-5} 6.8×10^{-5} 7.8×10^{-5}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators				
AC Current	40303	(20 ~ 50) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (50 ~ 100) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (100 ~ 200) A 60 Hz	1.0×10^{-4} 8.3×10^{-5} 1.1×10^{-4} 1.2×10^{-4} 9.7×10^{-5} 1.3×10^{-4} 4.5×10^{-4}	Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
Wattmeter calibrators				
AC Voltage	40304	(1 mV) 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (1 ~ 2) mV 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (2 ~ 5) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (5 ~ 10) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (10 ~ 20) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (20 ~ 50) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (50 ~ 100) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	1.7×10^{-3} 3.0×10^{-3} 1.2×10^{-2} 1.1×10^{-3} 1.7×10^{-3} 7.7×10^{-3} 6.4×10^{-4} 5.8×10^{-4} 1.0×10^{-3} 5.4×10^{-3} 4.2×10^{-4} 3.5×10^{-4} 5.8×10^{-4} 3.9×10^{-3} 1.8×10^{-4} 1.4×10^{-4} 2.2×10^{-4} 2.2×10^{-3} 1.4×10^{-4} 9.2×10^{-5} 1.6×10^{-4} 1.4×10^{-3} 1.1×10^{-4} 6.6×10^{-5} 1.2×10^{-4} 1.3×10^{-3}	Power Standard, Counter/ SICT-CP-40304

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Wattmeter calibrators	40304			
AC Voltage		(100 ~ 200) mV		Power Standard, Counter/ SICT-CP-40304
		10 Hz	8.2×10^{-5}	
		10 Hz ~ 10 kHz	3.9×10^{-5}	
		(10 ~ 100) kHz	7.6×10^{-5}	
		100 kHz ~ 1 MHz	1.1×10^{-3}	
		(200 ~ 500) mV		
		10 Hz	7.8×10^{-5}	
		10 Hz ~ 10 kHz	3.6×10^{-5}	
		(10 ~ 100) kHz	7.1×10^{-5}	
		100 kHz ~ 1 MHz	1.1×10^{-3}	
		(500 mV ~ 1 V)		
		10 Hz	7.6×10^{-5}	
		10 Hz ~ 10 kHz	3.3×10^{-5}	
		(10 ~ 100) kHz	6.6×10^{-5}	
		100 kHz ~ 1 MHz	1.1×10^{-3}	
		(1 ~ 2) V		
		10 Hz	7.1×10^{-5}	
		10 Hz ~ 10 kHz	2.7×10^{-5}	
		(10 ~ 100) kHz	5.8×10^{-5}	
		100 kHz ~ 1 MHz	1.0×10^{-3}	
		(2 ~ 5) V		
		10 Hz	7.2×10^{-5}	
		10 Hz ~ 10 kHz	2.6×10^{-5}	
		(10 ~ 100) kHz	7.5×10^{-5}	
		100 kHz ~ 1 MHz	1.4×10^{-3}	
		(5 ~ 20) V		
		10 Hz	7.2×10^{-5}	
		10 Hz ~ 10 kHz	2.8×10^{-5}	
		(10 ~ 100) kHz	7.5×10^{-5}	
		100 kHz ~ 1 MHz	1.4×10^{-3}	
		(20 ~ 50) V		
		10 Hz	7.2×10^{-5}	
		10 Hz ~ 10 kHz	3.0×10^{-5}	
		(10 ~ 100) kHz	8.0×10^{-5}	
		(50 ~ 200) V		
		10 Hz	7.4×10^{-5}	
		10 Hz ~ 10 kHz	3.3×10^{-5}	
		(10 ~ 100) kHz	8.5×10^{-5}	
		(200 ~ 1 000) V		
		10 Hz	7.7×10^{-5}	
		10 Hz ~ 10 kHz	3.3×10^{-5}	
		(10 ~ 100) kHz	5.8×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Wattmeter calibrators	40304			
	AC Current	(10 μ A) 10 Hz ~ 10 kHz	2.6×10^{-3}	Power Standard, Counter/ SICT-CP-40304
		(10 ~ 100) μ A 10 Hz ~ 1 kHz (1 ~ 10) kHz	3.6×10^{-4} 6.4×10^{-4}	
		(100 μ A ~ 1 mA) 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	9.8×10^{-5} 7.5×10^{-5} 9.4×10^{-5}	
		(1 ~ 100) mA 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	7.8×10^{-5} 4.6×10^{-5} 4.2×10^{-5}	
		(100 mA ~ 1 A) 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	8.1×10^{-5} 4.9×10^{-5} 4.4×10^{-5}	
		(1 ~ 2) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	7.9×10^{-5} 4.7×10^{-5} 4.5×10^{-5}	
		(2 ~ 5) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	8.2×10^{-5} 5.2×10^{-5} 5.0×10^{-5}	
		(5 ~ 10) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	8.6×10^{-5} 5.9×10^{-5} 7.8×10^{-5}	
		(10 ~ 20) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	9.3×10^{-5} 6.8×10^{-5} 7.8×10^{-5}	
		(20 ~ 50) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	1.0×10^{-4} 8.3×10^{-5} 1.1×10^{-4}	
		(50 ~ 100) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	1.2×10^{-4} 9.7×10^{-5} 1.3×10^{-4}	
		(100 ~ 200) A 60 Hz	4.5×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Wattmeter calibrators	40304			Power Standard, Counter/ SICT-CP-40304
AC Power		(50 ~ 60) Hz 0 mW (0 ~ 0.22) mW (0.22 ~ 1.1) mW (1.1 ~ 2.2) mW (2.2 ~ 11) mW (11 ~ 22) mW (22 ~ 44) mW (44 ~ 66) mW (66 ~ 88) mW (88 ~ 110) mW (110 ~ 480) mW (480 ~ 550) mW (0.55 ~ 1.1) W (1.1 ~ 5.5) W 5.5 W ~ 1.1 kW (1.1 ~ 2.2) kW (2.2 ~ 24) kW	0.05 mW 3.2×10^{-1} 6.3×10^{-2} 3.1×10^{-2} 6.3×10^{-3} 3.1×10^{-3} 1.6×10^{-3} 1.1×10^{-3} 8.0×10^{-4} 6.4×10^{-4} 2.6×10^{-4} 2.3×10^{-4} 1.4×10^{-4} 1.6×10^{-4} 1.4×10^{-4} 1.5×10^{-4} 1.4×10^{-4}	
Power Factor		(50 ~ 60) Hz (0 ~ 550) mW -1 ~ 1 550 mW ~ 24 kW -1 ~ 1		2.3×10^{-4} 1.5×10^{-4}
Harmonic Voltage		(50 ~ 60) Hz (0.5 ~ 3) % (P_{st} 10) % (10 ~ 20) %		0.042 % 0.052 % 0.081 %
Harmonic Current		(50 ~ 60) Hz (0.5 ~ 3) % (3 ~ 10) % (10 ~ 20) %		0.042 % 0.052 % 0.055 %
Flicker		(0.25 ~ 5), (50 Hz) Modulation Frequency 8.333 mHz 16.667 mHz 58.333 mHz 325.000 mHz 916.667 mHz 13.500 Hz 33.333 Hz		2.7×10^{-3} 2.7×10^{-3} 2.7×10^{-3} 2.7×10^{-3} 2.7×10^{-3} 2.7×10^{-3} 2.7×10^{-3}
Frequency		(10 ~ 100) Hz (100 ~ 400) Hz 400 Hz ~ 1 MHz		7.0×10^{-7} 3.8×10^{-7} 7.0×10^{-7}

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC current shunts	40305	(100 ~ 200) A (60 Hz) 1 mΩ	8.4×10^{-4}	Reference Multimeter, Calibrator/ SICT-CP-40305
		(40 ~ 60) Hz (1 ~ 10) mΩ (10 ~ 100) mΩ	2.1×10^{-4} 3.5×10^{-4}	
		(60 Hz ~ 1 kHz) (1 ~ 10) mΩ (10 ~ 100) mΩ	5.7×10^{-4} 3.4×10^{-4}	
		(40 Hz ~ 1 kHz) 100 mΩ ~ 100 Ω 100 Ω ~ 10 kΩ	1.8×10^{-4} 2.3×10^{-4}	
		(50 Hz) (1 ~ 100) kV	2.6×10^{-4}	
		(60 Hz) (1 ~ 100) kV	1.9×10^{-4}	
Phase angle generators, synchro resolve generators	40306			전력 교정기/ SICT-CP-40307
		(-180 ~ 180) ° 50 Hz (50 ~ 500) Hz (500 ~ 1 000) Hz	0.001 6 ° 0.003 1 ° 0.010 °	
Voltage/current phase angle meters/synchro resolve meters	40307			Power Calibrator/ SICT-CP-40307
		(50 ~ 60) Hz (-180 ~ 180) °	0.008 8 °	
Potential transformer test set	40308			Standard Potential transforme, Ratio transformers/ SICT-CP-40308
		(110 ~ 1 100) V (-19.99 ~ 19.99) % (-680 ~ 680) ′	0.020 % 0.70'	
		(1 100 ~ 22 900) V (-19.99 ~ 19.99) %	0.016 %	
		(-680 ~ 680) ′	0.50'	
Potential transforme	40309			Standard Potential transforme/ SICT-CP-40309
		(110 ~ 1 100) V (-19.99 ~ 19.99) % (-680 ~ 680) ′	0.020 % 0.70'	
		(1 100 ~ 22 900) V (-19.99 ~ 19.99) %	0.016 %	
		(-680 ~ 680) ′	0.50'	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Power factor meters AC Power Factor	40310	(50 Hz, 60 Hz) -1 ~ 1	1.1×10^{-4}	Power Calibrator/ SICT-CP-40310
AC power meters AC Voltage	40311	(1 mV) 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	4.8×10^{-3} 5.0×10^{-3} 6.5×10^{-3}	Power Calibrator, Calibrator/ SICT-CP-40311
		(1 ~ 2) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	2.4×10^{-3} 2.6×10^{-3} 3.5×10^{-3}	
		(2 ~ 5) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	1.7×10^{-3} 1.9×10^{-3} 2.7×10^{-3}	
		(5 ~ 7) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	8.9×10^{-4} 1.0×10^{-3} 1.6×10^{-3}	
		(7 ~ 9) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	6.9×10^{-4} 8.4×10^{-4} 1.4×10^{-3}	
		(9 ~ 10) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	5.7×10^{-4} 7.1×10^{-4} 1.2×10^{-3}	
		(10 ~ 30) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	3.6×10^{-4} 4.7×10^{-4} 1.1×10^{-3}	
		(30 ~ 60) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	2.9×10^{-4} 3.7×10^{-4} 8.8×10^{-4}	
		(60 ~ 200) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	1.9×10^{-4} 2.7×10^{-4} 6.5×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311	(200 ~ 300) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (300 ~ 600) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (600 mV ~ 1 V) 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (1 ~ 2) V 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (2 ~ 3) V 40 Hz 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (3 ~ 5) V 40 Hz 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (5 ~ 7) V 40 Hz 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (7 ~ 20) V 40 Hz 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (20 ~ 60) V 40 Hz 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	8.6 × 10 ⁻⁵ 1.3 × 10 ⁻⁴ 2.2 × 10 ⁻⁴ 7.6 × 10 ⁻⁵ 1.2 × 10 ⁻⁴ 1.9 × 10 ⁻⁴ 6.7 × 10 ⁻⁵ 9.8 × 10 ⁻⁵ 1.5 × 10 ⁻⁴ 5.6 × 10 ⁻⁵ 8.5 × 10 ⁻⁵ 1.2 × 10 ⁻⁴ 1.3 × 10 ⁻⁴ 9.5 × 10 ⁻⁵ 1.5 × 10 ⁻⁴ 2.2 × 10 ⁻⁴ 1.1 × 10 ⁻⁴ 8.0 × 10 ⁻⁵ 1.3 × 10 ⁻⁴ 1.8 × 10 ⁻⁴ 8.1 × 10 ⁻⁵ 6.7 × 10 ⁻⁵ 1.1 × 10 ⁻⁴ 1.5 × 10 ⁻⁴ 7.0 × 10 ⁻⁵ 6.1 × 10 ⁻⁵ 9.9 × 10 ⁻⁵ 1.3 × 10 ⁻⁴ 1.6 × 10 ⁻⁴ 1.2 × 10 ⁻⁴ 1.6 × 10 ⁻⁴ 3.4 × 10 ⁻⁴	Power Calibrator, Calibrator/ SICT-CP-40311

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311			
AC Voltage		(60 ~ 100) V		Power Calibrator, Calibrator/ SICT-CP-40311
		40 Hz	9.0×10^{-5}	
		40 Hz ~ 10 kHz	7.8×10^{-5}	
		(10 ~ 50) kHz	1.2×10^{-4}	
		(50 ~ 100) kHz	2.3×10^{-4}	
		(100 ~ 200) V		
		40 Hz	6.7×10^{-5}	
		40 Hz ~ 10 kHz	6.5×10^{-5}	
		(10 ~ 50) kHz	9.9×10^{-5}	
		(50 ~ 100) kHz	1.9×10^{-4}	
		(200 ~ 400) V		
		40 Hz	1.4×10^{-4}	
		40 Hz ~ 1 kHz	1.1×10^{-4}	
		(1 ~ 10) kHz	2.2×10^{-4}	
		(10 ~ 20) kHz	6.6×10^{-4}	
		(400 ~ 500) V		
		40 Hz	1.3×10^{-4}	
		40 Hz ~ 1 kHz	1.1×10^{-4}	
		(1 ~ 10) kHz	2.1×10^{-4}	
		(10 ~ 20) kHz	5.4×10^{-4}	
		(500 ~ 600) V		
		40 Hz	1.5×10^{-4}	
		40 Hz ~ 1 kHz	1.3×10^{-4}	
		(1 ~ 10) kHz	2.2×10^{-4}	
		(10 ~ 20) kHz	9.9×10^{-4}	
		(600 ~ 700) V		
		40 Hz	1.4×10^{-4}	
		40 Hz ~ 1 kHz	1.2×10^{-4}	
		(1 ~ 10) kHz	2.2×10^{-4}	
		(10 ~ 20) kHz	8.5×10^{-4}	
		(700 ~ 900) V		
		40 Hz	1.3×10^{-4}	
		40 Hz ~ 1 kHz	1.1×10^{-4}	
		(1 ~ 10) kHz	2.1×10^{-4}	
		(10 ~ 20) kHz	7.5×10^{-4}	
		(900 ~ 1 000) V		
		40 Hz	1.2×10^{-4}	
		40 Hz ~ 1 kHz	1.0×10^{-4}	
		(1 ~ 10) kHz	2.1×10^{-4}	
		(10 ~ 20) kHz	6.1×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters				
				Power Calibrator, Calibrator/ SICT-CP-40311
AC Current	40311	(100 μ A) 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	2.3×10^{-4} 4.9×10^{-4} 2.1×10^{-3}	
		(100 ~ 300) μ A 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	2.3×10^{-4} 4.6×10^{-4} 1.8×10^{-3}	
		(300 μ A ~ 2 mA) 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	1.9×10^{-4} 3.8×10^{-4} 2.1×10^{-3}	
		(2 ~ 4) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	3.2×10^{-4} 6.9×10^{-4} 3.8×10^{-3}	
		(4 ~ 7) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	2.3×10^{-4} 4.8×10^{-4} 2.7×10^{-3}	
		(7 ~ 20) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	1.9×10^{-4} 3.7×10^{-4} 2.1×10^{-3}	
		(20 ~ 30) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	2.9×10^{-4} 6.5×10^{-4} 2.6×10^{-3}	
		(30 ~ 60) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	2.4×10^{-4} 5.2×10^{-4} 2.2×10^{-3}	
		(60 ~ 200) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	1.8×10^{-4} 3.7×10^{-4} 1.7×10^{-3}	
		(200 ~ 300) mA 40 Hz (40 ~ 60) Hz 60 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	5.8×10^{-4} 1.2×10^{-4} 5.3×10^{-4} 1.2×10^{-3} 9.3×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311	(300 ~ 500) mA 40 Hz (40 ~ 60) Hz 60 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (500 ~ 800) mA 40 Hz (40 ~ 60) Hz 60 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (800 mA ~ 2 A) 40 Hz (40 ~ 60) Hz 60 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (2 ~ 3) A 40 Hz (40 ~ 60) Hz (60 ~ 100) Hz 100 Hz ~ 5 kHz (5 ~ 10) kHz (3 ~ 6) A 40 Hz (40 ~ 60) Hz (60 ~ 100) Hz 100 Hz ~ 5 kHz (5 ~ 10) kHz (6 ~ 10) A 40 Hz (40 ~ 60) Hz (60 ~ 100) Hz 100 Hz ~ 5 kHz (5 ~ 10) kHz (10 ~ 50) A 40 Hz (40 ~ 60) Hz (60 ~ 100) Hz 100 Hz ~ 5 kHz (5 ~ 10) kHz (50 ~ 100) A (40 ~ 100) Hz 100 Hz ~ 5 kHz (5 ~ 10) kHz	4.9×10^{-4} 1.1×10^{-4} 4.6×10^{-4} 1.0×10^{-3} 8.9×10^{-3} 4.0×10^{-4} 1.3×10^{-4} 3.9×10^{-4} 8.3×10^{-4} 8.6×10^{-3} 3.5×10^{-4} 9.5×10^{-5} 3.5×10^{-4} 7.2×10^{-4} 8.4×10^{-3} 5.3×10^{-4} 1.5×10^{-4} 5.3×10^{-4} 9.4×10^{-4} 4.5×10^{-3} 4.1×10^{-4} 1.3×10^{-4} 4.1×10^{-4} 8.1×10^{-4} 4.4×10^{-3} 2.6×10^{-4} 1.1×10^{-4} 2.6×10^{-4} 6.4×10^{-4} 4.3×10^{-3} 2.4×10^{-4} 1.1×10^{-4} 2.4×10^{-4} 6.6×10^{-4} 6.2×10^{-2} 1.9×10^{-4} 5.7×10^{-4} 4.8×10^{-2}	Power Calibrator, Calibrator/ SICT-CP-40311

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311	(100 ~ 1 000) A (40 ~ 100) Hz 100 Hz ~ 1 kHz	1.3×10^{-3} 2.7×10^{-3}	Power Calibrator, Calibrator/ SICT-CP-40311
		(1 000 ~ 2 500) A (40 ~ 60) Hz	1.2×10^{-3}	
		(2 500 ~ 3 000) A 60 Hz	1.3×10^{-3}	
		(50 ~ 60) Hz		
		0 mW	$70 \mu\text{W}$	
		(0 ~ 0.22) mW	2.1×10^{-1}	
		(0.22 ~ 1.1) mW	4.1×10^{-2}	
		(1.1 ~ 2.2) mW	2.1×10^{-2}	
		(2.2 ~ 11) mW	4.1×10^{-3}	
		(11 ~ 22) mW	2.1×10^{-3}	
		(22 ~ 44) mW	1.0×10^{-3}	
		(44 ~ 66) mW	7.0×10^{-4}	
		(66 ~ 88) mW	5.3×10^{-4}	
		(88 ~ 110) mW	4.3×10^{-4}	
		(110 ~ 480) mW	2.1×10^{-4}	
		480 mW ~ 12 kW	1.2×10^{-4}	
		(12 ~ 24) kW	6.8×10^{-4}	
		(24 ~ 300) kW	1.2×10^{-3}	
		(300 ~ 600) kW	1.4×10^{-3}	
		0 mV	$0.78 \mu\text{V}$	
		(0 ~ 1) mV	8.0×10^{-4}	
		(1 ~ 2) mV	4.0×10^{-4}	
		(2 ~ 3) mV	2.7×10^{-4}	
		(3 ~ 4) mV	2.0×10^{-4}	
		(4 ~ 8) mV	1.6×10^{-4}	
		(8 ~ 9) mV	9.4×10^{-5}	
		(9 ~ 10) mV	8.5×10^{-5}	
		(10 ~ 20) mV	4.5×10^{-5}	
		(20 ~ 30) mV	3.2×10^{-5}	
		(30 ~ 60) mV	2.6×10^{-5}	
		(60 ~ 200) mV	1.8×10^{-5}	
		(200 ~ 300) mV	9.3×10^{-6}	
		(300 ~ 400) mV	8.3×10^{-6}	
		(400 ~ 700) mV	7.8×10^{-6}	
		(700 ~ 900) mV	6.9×10^{-6}	
		(0.9 ~ 1) V	9.0×10^{-6}	
		(1 ~ 2) V	7.2×10^{-6}	
		(2 ~ 3) V	5.6×10^{-6}	
		(3 ~ 9) V	5.1×10^{-6}	
		(9 ~ 10) V	7.5×10^{-6}	
		(10 ~ 20) V	6.0×10^{-6}	
		(20 ~ 40) V	7.9×10^{-6}	
		(40 ~ 90) V	6.9×10^{-6}	
		(90 ~ 200) V	8.8×10^{-6}	
		(200 ~ 400) V	9.9×10^{-6}	
		(400 ~ 900) V	8.7×10^{-6}	
		(900 ~ 1 000) V	1.0×10^{-5}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311	0 μA	9.0 nA	Power Calibrator, Calibrator/ SICT-CP-40311
		(0 ~ 0.1) μA	9.0×10^{-2}	
		(0.1 ~ 0.2) μA	4.5×10^{-2}	
		(0.2 ~ 0.3) μA	3.0×10^{-2}	
		(0.3 ~ 0.4) μA	2.3×10^{-2}	
		(0.4 ~ 0.9) μA	1.8×10^{-2}	
		(0.9 ~ 1) μA	9.1×10^{-3}	
		(1 ~ 2) μA	4.5×10^{-3}	
		(2 ~ 3) μA	3.0×10^{-3}	
		(3 ~ 4) μA	2.3×10^{-3}	
		(4 ~ 9) μA	1.8×10^{-3}	
		(9 ~ 10) μA	9.4×10^{-4}	
		(10 ~ 20) μA	5.0×10^{-4}	
		(20 ~ 30) μA	3.4×10^{-4}	
		(30 ~ 50) μA	2.7×10^{-4}	
		(50 ~ 90) μA	1.9×10^{-4}	
		(90 ~ 100) μA	1.3×10^{-4}	
		(100 ~ 200) μA	8.6×10^{-5}	
		(200 ~ 300) μA	6.5×10^{-5}	
		(300 ~ 500) μA	5.5×10^{-5}	
		(0.5 ~ 2) mA	5.1×10^{-5}	
		(2 ~ 3) mA	7.3×10^{-5}	
		(3 ~ 4) mA	6.3×10^{-5}	
		(4 ~ 7) mA	5.7×10^{-5}	
		(7 ~ 20) mA	4.9×10^{-5}	
		(20 ~ 30) mA	8.2×10^{-5}	
		(30 ~ 50) mA	7.4×10^{-5}	
		(50 ~ 100) mA	6.6×10^{-5}	
		(100 ~ 200) mA	5.7×10^{-5}	
		(200 ~ 600) mA	1.5×10^{-4}	
		(0.6 ~ 2) A	1.1×10^{-4}	
		(2 ~ 3) A	4.0×10^{-4}	
		(3 ~ 4) A	3.2×10^{-4}	
		(4 ~ 7) A	2.8×10^{-4}	
		(7 ~ 10) A	2.1×10^{-4}	
		(10 ~ 20) A	1.5×10^{-4}	
		(20 ~ 30) A	2.5×10^{-4}	
		(30 ~ 80) A	2.1×10^{-4}	
		(80 ~ 100) A	1.5×10^{-4}	
		(100 ~ 2 500) A	1.3×10^{-3}	
DC Wattage	40311	0 mW	61 nW	Power Calibrator, Calibrator/ SICT-CP-40311
		(0 ~ 1) mW	7.7×10^{-5}	
		(1 ~ 10) mW	4.8×10^{-5}	
		(10 ~ 100) mW	6.1×10^{-5}	
		(0.1 ~ 100) W	1.1×10^{-4}	
		(0.1 ~ 10) kW	1.9×10^{-4}	
		(10 ~ 20) kW	1.5×10^{-4}	
DC Current	40311	(20 ~ 1 000) kW	1.0×10^{-3}	Power Calibrator, Calibrator/ SICT-CP-40311
		(1 000 ~ 2 500) kW	1.3×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311			Power Calibrator, Calibrator/ SICT-CP-40311
Harmonic Voltage		(50 ~ 60) Hz ($\frac{P_{st}}{P_{st}} \sim 3$) % (3 ~ 5) % (5 ~ 10) % (10 ~ 20) %	0.030 % 0.033 % 0.042 % 0.065 %	
Harmonic Current		(50 ~ 60) Hz (0.5 ~ 3) % (3 ~ 10) % (5 ~ 20) %	0.030 % 0.032 % 0.038 %	
Flicker		(0.25 ~ 5), (50 Hz) Modulation Frequency 8.333 mHz 16.667 mHz 58.333 mHz 325 mHz 916.667 mHz 13.5 Hz 33.333 Hz	3.5×10^{-2} 3.5×10^{-2} 3.5×10^{-2} 3.5×10^{-2} 3.5×10^{-2} 3.5×10^{-2} 3.5×10^{-2}	
Frequency		10 Hz ~ 10 MHz	1.3×10^{-4}	
Power Factor		(50 ~ 60) Hz -1 ~ 1	1.1×10^{-4}	
Current burden		(50 ~ 60) Hz 1.25 VA (1.25 ~ 3.75) VA (3.75 ~ 5) VA (5 ~ 10) VA (10 ~ 100) VA	2.9×10^{-3} 1.5×10^{-3} 8.0×10^{-4} 5.1×10^{-4} 4.0×10^{-4}	
Current burden factor		0.5 ~ 1	2.2×10^{-4}	
Voltage burden		(50 ~ 60) Hz 1.25 VA (1.25 ~ 3.75) VA (3.75 ~ 5) VA (5 ~ 10) VA (10 ~ 100) VA	7.4×10^{-4} 4.6×10^{-4} 3.7×10^{-4} 3.0×10^{-4} 3.1×10^{-4}	
Voltage burden factor		0.5 ~ 1	1.6×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power supplies	40312	(10 mV) 40 Hz ~ 5 kHz	2.2×10^{-4}	Voltage Standard, Multimeter, Current Shunt/ SICT-CP-40312
		(10 ~ 100) mV 40 Hz ~ 5 kHz	7.9×10^{-5}	
		(100 mV ~ 1 V) 40 Hz ~ 5 kHz	6.7×10^{-5}	
		(1 ~ 10) V 40 Hz ~ 5 kHz	6.8×10^{-5}	
		(10 ~ 100) V 40 Hz ~ 5 kHz	7.0×10^{-5}	
		(100 ~ 600) V 40 Hz ~ 5 kHz	4.1×10^{-5}	
		(600 ~ 1 000) V 40 Hz ~ 5 kHz	7.1×10^{-5}	
		Frequency 10 Hz	9.5×10^{-5}	
		(10 ~ 50) Hz	1.9×10^{-5}	
		(50 ~ 100) Hz	7.7×10^{-6}	
AC Current	Frequency 10 Hz	(0.1 ~ 1) kHz	8.4×10^{-7}	
		(1 ~ 5) kHz	3.8×10^{-7}	
		AC Current (1 mA)	6.4×10^{-4}	
		(50 ~ 60) Hz	6.4×10^{-4}	
		(1 ~ 10) mA (50 ~ 60) Hz	3.6×10^{-4}	
		(10 ~ 100) mA (50 ~ 60) Hz	2.4×10^{-4}	
		(100 mA ~ 1 A) (50 ~ 60) Hz	2.1×10^{-4}	
		(1 ~ 10) A (50 ~ 60) Hz	2.3×10^{-4}	
		(10 ~ 20) A (50 ~ 60) Hz	4.0×10^{-4}	
		(20 ~ 30) A (50 ~ 60) Hz	6.4×10^{-4}	
		(30 ~ 50) A (50 ~ 60) Hz	4.2×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power supplies	DC Voltage	40312	(±) 0 mV (0 ~ 10) mV (10 ~ 100) mV (0.1 ~ 100) V (100 ~ 600) V (600 ~ 1 000) V	5.8 μ V 5.8×10^{-4} 5.8×10^{-5} 7.7×10^{-6} 1.3×10^{-5} 6.6×10^{-5}
			(1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A (10 ~ 300) A (300 ~ 500) A (500 ~ 1 000) A (1 000 ~ 3 000) A	5.8×10^{-3} 5.9×10^{-4} 2.4×10^{-4} 3.1×10^{-4} 2.4×10^{-4} 2.6×10^{-4} 4.7×10^{-5} 5.1×10^{-4}
			(0 ~ 2) mV (2 ~ 20) mV (20 ~ 200) mV	0.16 mV 7.8×10^{-2} 8.2×10^{-3}
			(0.1 ~ 0.4) mV (0.4 ~ 0.6) mV (0.6 ~ 1) mV (1 ~ 10) mV (10 ~ 50) mV	3.8×10^{-1} 1.1×10^{-1} 7.3×10^{-2} 4.4×10^{-2} 7.1×10^{-2}
			(50 ~ 60) Hz 0.5 % (0.5 ~ 10) % (10 ~ 20) %	0.050 % 0.051 % 0.082 %
	Ripple		(50 ~ 60) Hz 0.5 % (0.5 ~ 20) %	0.050 % 0.051 %
			(50 ~ 60) Hz 0.5 % (0.5 ~ 20) %	0.050 % 0.051 %
			(50 ~ 60) Hz 0.5 % (0.5 ~ 20) %	0.050 % 0.051 %
Puncture/safety testers	DC Voltage	40313	(±) 0 kV (0 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 2) kV (2 ~ 100) kV (100 ~ 200) kV	0.58 V 1.2×10^{-3} 6.1×10^{-4} 3.0×10^{-4} 2.3×10^{-4} 1.2×10^{-2}
			(50 ~ 60) Hz 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 100) kV (100 ~ 200) kV	0.58 V 1.2×10^{-3} 6.2×10^{-4} 5.7×10^{-4} 1.2×10^{-2}
			(50 ~ 60) Hz 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 100) kV (100 ~ 200) kV	0.58 V 1.2×10^{-3} 6.2×10^{-4} 5.7×10^{-4} 1.2×10^{-2}
			(50 ~ 60) Hz 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 100) kV (100 ~ 200) kV	0.58 V 1.2×10^{-3} 6.2×10^{-4} 5.7×10^{-4} 1.2×10^{-2}
			(50 ~ 60) Hz 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 100) kV (100 ~ 200) kV	0.58 V 1.2×10^{-3} 6.2×10^{-4} 5.7×10^{-4} 1.2×10^{-2}
			(50 ~ 60) Hz 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 100) kV (100 ~ 200) kV	0.58 V 1.2×10^{-3} 6.2×10^{-4} 5.7×10^{-4} 1.2×10^{-2}

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Puncture/safety testers				
AC Breaking Current	40313	(50 ~ 60) Hz 0.1 mA (0.1 ~ 0.5) mA (0.5 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 10) mA (10 ~ 20) mA (20 ~ 50) mA (50 ~ 100) mA	5.3 × 10 ⁻⁴ 4.4 × 10 ⁻⁴ 7.3 × 10 ⁻⁴ 7.1 × 10 ⁻⁴ 4.4 × 10 ⁻⁴ 3.6 × 10 ⁻⁴ 7.1 × 10 ⁻⁴ 4.4 × 10 ⁻⁴ 7.3 × 10 ⁻⁴	AC/DC Kilovoltmeter, High Voltage Digital Meter, Reference Multimeter/ SICT-CP-40313
DC Breaking Current		0.1 mA (0.1 ~ 0.5) mA (0.5 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 10) mA (10 ~ 20) mA (20 ~ 50) mA (50 ~ 100) mA	3.9 × 10 ⁻⁴ 1.3 × 10 ⁻⁴ 6.4 × 10 ⁻⁴ 3.2 × 10 ⁻⁴ 1.3 × 10 ⁻⁴ 6.5 × 10 ⁻⁵ 3.3 × 10 ⁻⁴ 1.4 × 10 ⁻⁴ 6.4 × 10 ⁻⁴	
Resistance		1 mΩ (1 ~ 10) mΩ 10 mΩ ~ 100 kΩ	8.6 × 10 ⁻⁴ 7.2 × 10 ⁻⁴ 6.8 × 10 ⁻⁴	
Insulation Voltage		1 V (1 ~ 10) V (10 ~ 25) V (25 ~ 50) V (50 ~ 100) V (100 ~ 250) V (250 ~ 500) V (500 ~ 1 000) V (1 000 ~ 5 000) V (5 000 ~ 10 000) V	6.4 × 10 ⁻⁴ 6.4 × 10 ⁻⁵ 2.5 × 10 ⁻⁴ 1.3 × 10 ⁻⁴ 6.4 × 10 ⁻⁵ 2.5 × 10 ⁻⁴ 1.3 × 10 ⁻⁴ 6.4 × 10 ⁻⁵ 6.5 × 10 ⁻³ 6.1 × 10 ⁻³	
Insulation Resistance		1 kΩ (1 ~ 10) kΩ (10 ~ 100) kΩ (0.1 ~ 1) MΩ (1 ~ 10) MΩ (10 ~ 100) MΩ (0.1 ~ 1) GΩ (1 ~ 10) GΩ (10 ~ 100) GΩ (0.1 ~ 1) TΩ 10 TΩ	7.1 × 10 ⁻⁵ 3.7 × 10 ⁻⁵ 2.5 × 10 ⁻⁵ 3.1 × 10 ⁻⁵ 9.5 × 10 ⁻⁵ 2.4 × 10 ⁻⁵ 3.1 × 10 ⁻⁵ 6.1 × 10 ⁻⁵ 1.3 × 10 ⁻⁴ 2.6 × 10 ⁻⁴ 6.3 × 10 ⁻⁴	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Puncture/safety testers	40313			AC/DC Kilovoltmeter, High Voltage Digital Meter, Reference Multimeter/ SICT-CP-40313
Leakage current(DC)		0 μA	7.0 nA	
		(0 ~ 1) μA	7.0×10^{-3}	
		(1 ~ 2) μA	3.6×10^{-3}	
		(2 ~ 5) μA	1.4×10^{-3}	
		(5 ~ 10) μA	7.4×10^{-4}	
		(10 ~ 20) μA	4.0×10^{-4}	
		(20 ~ 50) μA	1.8×10^{-4}	
		(50 ~ 100) μA	1.3×10^{-4}	
		(100 ~ 200) μA	8.5×10^{-5}	
		(0.2 ~ 100) mA	6.1×10^{-4}	
Leakage current(AC)		(20 μA)		
		10 Hz	1.3×10^{-3}	
		(10 ~ 20) Hz	8.5×10^{-4}	
		20 Hz ~ 1 kHz	7.0×10^{-4}	
		(1 ~ 5) kHz	1.3×10^{-3}	
		(5 ~ 10) kHz	5.5×10^{-3}	
		(20 ~ 50) μA		
		10 Hz	6.8×10^{-4}	
		(10 ~ 20) Hz	4.4×10^{-4}	
		20 Hz ~ 1 kHz	3.4×10^{-4}	
		(1 ~ 5) kHz	6.8×10^{-4}	
		(5 ~ 10) kHz	2.8×10^{-3}	
		(50 ~ 100) μA		
		10 Hz	4.9×10^{-4}	
		(10 ~ 20) Hz	3.2×10^{-4}	
		20 Hz ~ 1 kHz	2.3×10^{-4}	
		(1 ~ 5) kHz	4.9×10^{-4}	
		(5 ~ 10) kHz	4.0×10^{-4}	
		(100 ~ 200) μA		
		10 Hz	3.9×10^{-4}	
		(10 ~ 20) Hz	2.5×10^{-4}	
		20 Hz ~ 1 kHz	1.7×10^{-4}	
		(1 ~ 5) kHz	4.0×10^{-4}	
		(5 ~ 10) kHz	1.7×10^{-3}	
		(200 ~ 500) μA		
		10 Hz	4.4×10^{-4}	
		(10 ~ 20) Hz	3.2×10^{-4}	
		20 Hz ~ 1 kHz	2.4×10^{-4}	
		(1 ~ 5) kHz	5.4×10^{-4}	
		(5 ~ 10) kHz	2.8×10^{-3}	
		500 μA ~ 1 mA		
		10 Hz	7.0×10^{-4}	
		(10 ~ 20) Hz	6.6×10^{-4}	
		20 Hz ~ 1 kHz	6.3×10^{-4}	
		(1 ~ 5) kHz	7.2×10^{-4}	
		(5 ~ 10) kHz	2.1×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Puncture/safety testers				
Leakage current(AC)	40313	(1 ~ 100) mA 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	7.0×10^{-4} 6.6×10^{-4} 6.3×10^{-4} 7.0×10^{-4} 2.7×10^{-3}	AC/DC Kilovoltmeter, High Voltage Digital Meter, Reference Multimeter/ SICT-CP-40313
Output AC Current		(60 Hz) 1 A (1 ~ 3) A (3 ~ 20) A (20 ~ 30) A (30 ~ 60) A (60 ~ 100) A (100 ~ 150) A (150 ~ 200) A	1.2×10^{-3} 1.5×10^{-3} 9.7×10^{-4} 1.0×10^{-3} 8.4×10^{-4} 1.0×10^{-3} 4.6×10^{-3} 3.7×10^{-3}	
Timer		1 s (1 ~ 100) s (100 ~ 1 000) s (1 000 ~ 10 000) s	5.8×10^{-6} 5.8×10^{-6} 8.2×10^{-6} 5.8×10^{-5}	
Power recorders				
AC Wattage	40314	(50 ~ 60) Hz 0 mW (0 ~ 0.22) mW (0.22 ~ 1.1) mW (1.1 ~ 2.2) mW (2.2 ~ 11) mW (11 ~ 22) mW (22 ~ 44) mW (44 ~ 66) mW (66 ~ 88) mW (88 ~ 110) mW (110 ~ 480) mW 480 mW ~ 12 kW (12 ~ 24) kW (24 ~ 300) kW (300 ~ 600) kW	70 μ W 2.1×10^{-1} 4.1×10^{-2} 2.1×10^{-2} 4.1×10^{-3} 2.1×10^{-3} 1.0×10^{-3} 7.0×10^{-4} 5.3×10^{-4} 4.3×10^{-4} 2.1×10^{-4} 1.2×10^{-4} 6.8×10^{-4} 1.2×10^{-3} 1.4×10^{-3}	Power Energy Calibrator/ SICT-CP-40314
DC Wattage		0 mW (0 ~ 1) mW (1 ~ 10) mW (10 ~ 100) mW (0.1 ~ 100) W (0.1 ~ 10) kW (10 ~ 20) kW (20 ~ 1 000) kW (1 000 ~ 2 500) kW	61 nW 7.7×10^{-5} 4.8×10^{-5} 6.1×10^{-5} 1.1×10^{-4} 1.9×10^{-4} 1.5×10^{-4} 1.0×10^{-3} 1.3×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Current transformer test set	40315	(5 ~ 1 500) A (-19.99 ~ 19.99) %	0.020 %	Current transforme, Ratio transformers/ SICT-CP-40315
		(-680 ~ 680) '	0.70 '	
Current transformer	40316	(5 ~ 10 000) A (-19.99 ~ 19.99) %	0.020 %	Current transforme/ SICT-CP-40316
		(-680 ~ 680) '	0.70 '	
	Current Coil transducers	(AC) 2 ~ 50	0.10 %	
		(DC) 2 ~ 50	0.10 %	
		(±) (10 A) 50 : 1 ~ 5 000 : 1	1.2×10^{-4}	
		(10 ~ 1 000) A 50 : 1 ~ 5 000 : 1	1.9×10^{-4}	
		(1 000 ~ 2 000) A 50 : 1 ~ 5 000 : 1	2.5×10^{-4}	
AC voltmeters	40318	(600 µV) 1 kHz	7.8×10^{-3}	Reference Multimeter, Calibrator/ SICT-CP-40318
		(600 µV ~ 1 mV) 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	5.0×10^{-3} 4.8×10^{-3} 6.5×10^{-3}	
		(1 ~ 3) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	2.0×10^{-3} 1.7×10^{-3} 2.8×10^{-3}	
		(3 ~ 10) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	7.7×10^{-4} 5.7×10^{-4} 1.2×10^{-3}	
		(10 ~ 30) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	8.0×10^{-4} 3.7×10^{-4} 1.1×10^{-3}	
		(30 ~ 100) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	4.3×10^{-4} 1.7×10^{-4} 5.7×10^{-4}	
		(100 mV ~ 10 V) 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	4.9×10^{-4} 1.1×10^{-4} 2.6×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltmeters	40318	(10 ~ 100) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz (100 ~ 1 000) V 50 Hz 50 Hz ~ 1 kHz (1 mV) 100 kHz 100 kHz ~ 1 MHz (1 ~ 10) mV 100 kHz 100 kHz ~ 1 MHz (10 ~ 100) mV 100 kHz 100 kHz ~ 1 MHz (100 mV ~ 1 V) 100 kHz 100 kHz ~ 1 MHz (1 ~ 10) V 100 kHz 100 kHz ~ 1 MHz (10 ~ 20) V 100 kHz 100 kHz ~ 1 MHz (25 mV) 1 MHz (1 ~ 30) MHz (25 ~ 100) mV 1 MHz (1 ~ 30) MHz (100 ~ 300) mV 1 MHz (1 ~ 30) MHz (300 mV ~ 1 V) 1 MHz (1 ~ 30) MHz (1 ~ 2) V 1 MHz (1 ~ 30) MHz	5.3 × 10 ⁻⁴ 1.3 × 10 ⁻⁴ 3.6 × 10 ⁻⁴ 3.7 × 10 ⁻⁴ 1.1 × 10 ⁻⁴ 6.5 × 10 ⁻³ 2.7 × 10 ⁻² 1.2 × 10 ⁻³ 5.6 × 10 ⁻³ 5.7 × 10 ⁻⁴ 3.7 × 10 ⁻³ 1.5 × 10 ⁻⁴ 2.3 × 10 ⁻³ 4.7 × 10 ⁻⁵ 7.0 × 10 ⁻⁴ 7.6 × 10 ⁻⁵ 1.3 × 10 ⁻³ 1.9 × 10 ⁻² 2.3 × 10 ⁻² 2.3 × 10 ⁻² 2.8 × 10 ⁻² 3.7 × 10 ⁻² 4.0 × 10 ⁻² 2.4 × 10 ⁻² 2.7 × 10 ⁻² 1.5 × 10 ⁻² 1.8 × 10 ⁻²	Reference Multimeter, Calibrator/ SICT-CP-40318

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltmeters	40318			Reference Multimeter, Calibrator/ SICT-CP-40318
AC Output Voltage		(1 mV) 10 Hz ~ 10 kHz (10 ~ 100) kHz	1.9×10^{-3} 3.1×10^{-3}	
		(1 ~ 10) mV 10 Hz ~ 10 kHz (10 ~ 100) kHz	2.6×10^{-4} 3.7×10^{-4}	
		(10 ~ 100) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	9.2×10^{-5} 4.8×10^{-5} 9.1×10^{-5}	
		(100 mV ~ 1 V) 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	7.2×10^{-5} 1.8×10^{-5} 5.9×10^{-5}	
DC Output Voltage		1 mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V	7.5×10^{-4} 7.6×10^{-5} 9.4×10^{-6} 2.2×10^{-5}	
Watt hour meters	40319			Power Calibrator/ SICT-CP-40319
Watt Hour		(50 ~ 60) Hz 0 mWh (0 ~ 480) mWh 480 mWh ~ 12 kWh (12 ~ 24) kWh (24 ~ 300) kWh (300 ~ 600) kWh (DC) 0 mWh (0 ~ 1) mWh (1 ~ 100) mWh 100 mWh ~ 100 Wh 100 Wh ~ 10 kWh (10 ~ 20) kWh (20 ~ 1 000) kWh (1 000 ~ 2 500) kWh	44 μ Wh 4.0×10^{-4} 3.6×10^{-4} 7.6×10^{-4} 1.3×10^{-3} 1.4×10^{-3} 61 nWh 3.5×10^{-4} 3.4×10^{-4} 3.5×10^{-4} 3.8×10^{-4} 3.7×10^{-4} 1.1×10^{-3} 1.3×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Ratio transformers	Ratio	PT (±)		Calibrator/ SICT-CP-40321
		(0.1 ~ 1.000) %	0.006 %	
		(1.000 ~ 19.00) %	0.01 %	
		CT (±)		
		(0.1 ~ 1.000) %	0.019 %	
	Phase	(1.000 ~ 19.00) %	0.02 %	
		PT (±)		
		(0.040 ~ 1.999) '	0.060 '	
		(1.999 ~ 19.99) '	0.06 '	
		(19.99 ~ 199.9) '	0.2 '	
		(199.9 ~ 600) '	1 '	
		CT (±)		
		(0.040 ~ 1.999) '	0.060 '	
		(1.999 ~ 19.99) '	0.06 '	
		(19.99 ~ 199.9) '	0.2 '	
		(199.9 ~ 600) '	1 '	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF amplifiers	40401	(DC) 1 mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 1 000) V (10 Hz ~ 10 kHz) 1 mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 1 000) V (10 ~ 100) kHz 1 mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 1 000) V	0.4 μ V 7.4×10^{-5} 6.1×10^{-5} 6.0×10^{-5} 1.0×10^{-4} 1.7 μ V 2.6×10^{-4} 1.1×10^{-4} 9.0×10^{-5} 1.0×10^{-4} 3.1 μ V 3.7×10^{-4} 1.1×10^{-4} 8.0×10^{-5} 1.0×10^{-4}	Reference Multimeter/ SICT-CP-40401
DC/LF attenuators	40402	10 Hz ~ 100 kHz (0 ~ -20) dB (-20 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	0.001 9 dB 0.001 7 dB 0.005 5 dB 0.008 7 dB	Reference Multimeter/ SICT-CP-40402
Multimeter calibrators	40403	(DC Voltage) (±) 0 mV (0 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V (DC Current) (±) 1 nA (1 ~ 100) nA 100 nA ~ 10 A (10 ~ 50) A (50 ~ 100) A (AC Voltage) (1 mV) 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (1 ~ 2) mV 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	0.05 μ V 3.3×10^{-6} 1.4×10^{-6} 1.0×10^{-6} 1.6×10^{-6} 2.1×10^{-6} 7.0 pA 4.7×10^{-3} 1.2×10^{-5} 4.0×10^{-5} 4.4×10^{-5} 1.7 $\times 10^{-3}$ 3.0×10^{-3} 1.2×10^{-2} 1.1 $\times 10^{-3}$ 1.7×10^{-3} 7.7×10^{-3}	Reference Multimeter/ SICT-CP-40403

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators	40403	(2 ~ 5) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (5 ~ 10) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (10 ~ 20) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (20 ~ 50) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (50 ~ 100) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (100 ~ 200) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (200 ~ 500) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (0.5 ~ 1) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (1 ~ 2) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	6.4×10^{-4} 5.8×10^{-4} 1.0×10^{-3} 5.4×10^{-3} 4.2×10^{-4} 3.5×10^{-4} 5.8×10^{-4} 3.9×10^{-3} 1.8×10^{-4} 1.4×10^{-4} 2.2×10^{-4} 2.2×10^{-3} 1.4×10^{-4} 9.2×10^{-5} 1.6×10^{-4} 1.4×10^{-3} 1.1×10^{-4} 6.6×10^{-5} 1.2×10^{-4} 1.3×10^{-3} 8.2×10^{-5} 3.9×10^{-5} 7.6×10^{-5} 1.1×10^{-3} 7.8×10^{-5} 3.6×10^{-5} 7.1×10^{-5} 1.1×10^{-3} 7.6×10^{-5} 3.3×10^{-5} 6.6×10^{-5} 1.1×10^{-3} 7.1×10^{-5} 2.7×10^{-5} 5.8×10^{-5} 1.0×10^{-3}	Reference Multimeter/ SICT-CP-40403

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators	40403			
AC Voltage		(2 ~ 5) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (5 ~ 20) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (20 ~ 50) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz (50 ~ 200) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz (200 ~ 1 000) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	7.2×10^{-5} 2.6×10^{-5} 7.5×10^{-5} 1.4×10^{-3} 7.2×10^{-5} 2.8×10^{-5} 7.5×10^{-5} 1.4×10^{-3} 7.2×10^{-5} 3.0×10^{-5} 8.0×10^{-5} 7.4×10^{-5} 3.3×10^{-5} 8.5×10^{-5} 7.7×10^{-5} 3.3×10^{-5} 5.8×10^{-4}	Reference Multimeter/ SICT-CP-40403
AC Current		(10 μA) 10 Hz ~ 10 kHz (10 ~ 100) μA 10 Hz ~ 1 kHz (1 ~ 10) kHz (0.1 ~ 1) mA 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (1 ~ 100) mA 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (0.1 ~ 1) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (1 ~ 2) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	2.6×10^{-3} 3.6×10^{-4} 6.4×10^{-4} 9.8×10^{-5} 7.5×10^{-5} 9.4×10^{-5} 7.8×10^{-5} 4.6×10^{-5} 4.2×10^{-5} 8.1×10^{-5} 4.9×10^{-5} 4.4×10^{-5} 7.9×10^{-5} 4.7×10^{-5} 4.5×10^{-5}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators	40403	(2 ~ 5) A		Reference Multimeter/ SICT-CP-40403
		10 Hz	8.2×10^{-5}	
		10 Hz ~ 1 kHz	5.2×10^{-5}	
		(1 ~ 10) kHz	5.0×10^{-5}	
		(5 ~ 10) A		
		10 Hz	8.6×10^{-5}	
		10 Hz ~ 1 kHz	5.9×10^{-5}	
		(1 ~ 10) kHz	7.8×10^{-5}	
		(10 ~ 20) A		
		10 Hz	9.3×10^{-5}	
		10 Hz ~ 1 kHz	6.8×10^{-5}	
		(1 ~ 10) kHz	7.8×10^{-5}	
		(20 ~ 50) A		
		10 Hz	1.0×10^{-4}	
		10 Hz ~ 1 kHz	8.3×10^{-5}	
		(1 ~ 10) kHz	1.1×10^{-4}	
		(50 ~ 100) A		
		10 Hz	1.2×10^{-4}	
		10 Hz ~ 1 kHz	9.7×10^{-5}	
		(1 ~ 10) kHz	1.3×10^{-4}	
Resistance		(100 ~ 200) A		
		60 Hz	4.5×10^{-4}	
		0 Ω	$0.14 \mu\Omega$	
		(0 ~ 1) Ω	6.6×10^{-6}	
		(1 ~ 1.9) Ω	8.4×10^{-6}	
		(1.9 ~ 10) Ω	3.6×10^{-6}	
		(10 ~ 19) Ω	2.6×10^{-6}	
		(19 ~ 100) Ω	2.8×10^{-6}	
		(0.1 ~ 1) kΩ	2.5×10^{-6}	
		(1 ~ 1.9) kΩ	3.8×10^{-6}	
		(1.9 ~ 10) kΩ	2.0×10^{-6}	
		(10 ~ 19) kΩ	1.3×10^{-6}	
		(19 ~ 100) kΩ	1.9×10^{-6}	
		(100 ~ 190) kΩ	2.0×10^{-6}	
		(0.19 ~ 1) MΩ	2.9×10^{-6}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators	40403			Reference Multimeter/ SICT-CP-40403
Multimeter calibrators(property) (Digital sampling)				
AC Voltage		(1 mV) 0.1 Hz ~ 3 kHz	8.4×10^{-4}	
		(1 ~ 2) mV 0.1 Hz ~ 3 kHz	4.2×10^{-4}	
		(2 ~ 3) mV 0.1 Hz ~ 3 kHz	2.8×10^{-4}	
		(3 ~ 5) mV 0.1 Hz ~ 3 kHz	1.7×10^{-4}	
		(5 ~ 10) mV 0.1 Hz ~ 3 kHz	8.8×10^{-5}	
		(10 ~ 20) mV 0.1 Hz ~ 3 kHz	4.8×10^{-5}	
		(20 ~ 30) mV 0.1 Hz ~ 3 kHz	3.6×10^{-5}	
		(30 ~ 50) mV 0.1 Hz ~ 3 kHz	3.0×10^{-5}	
		(50 ~ 100) mV 0.1 Hz ~ 3 kHz	2.6×10^{-5}	
		(100 ~ 200) mV 0.1 Hz ~ 3 kHz	4.8×10^{-5}	
		(200 ~ 300) mV 0.1 Hz ~ 3 kHz	3.6×10^{-5}	
		(300 ~ 500) mV 0.1 Hz ~ 3 kHz	2.8×10^{-5}	
		(500 mV ~ 1 V) 0.1 Hz ~ 3 kHz	2.4×10^{-5}	
		(1 ~ 2) V 0.1 Hz ~ 3 kHz	4.8×10^{-5}	
		(2 ~ 3) V 0.1 Hz ~ 3 kHz	3.6×10^{-5}	
		(3 ~ 5) V 0.1 Hz ~ 3 kHz	3.0×10^{-5}	
		(5 ~ 10) V 0.1 Hz ~ 3 kHz	2.6×10^{-5}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators	40403			Reference Multimeter/ SICT-CP-40403
Multimeter calibrators(property) (Digital sampling)				
AC Voltage		(10 ~ 30) V 10 Hz ~ 3 kHz	3.6×10^{-5}	
		(30 ~ 50) V 10 Hz ~ 3 kHz	2.8×10^{-5}	
		(50 ~ 100) V 10 Hz ~ 3 kHz	2.4×10^{-5}	
		(100 ~ 200) V 10 Hz ~ 3 kHz	4.8×10^{-5}	
		(200 ~ 1 000) V 50 Hz ~ 1 kHz	2.4×10^{-5}	
Oscilloscope calibrators	40404			Calibrator/ SICT-CP-40404
DC Voltage Amplitude		(±) 0 mV (0 ~ 1) mV (1 ~ 2) mV (2 ~ 5) mV (5 ~ 10) mV (10 ~ 20) mV (20 ~ 50) mV (50 ~ 100) mV (100 ~ 200) mV (200 ~ 500) mV (0.5 ~ 1) V (1 ~ 2) V (2 ~ 5) V (5 ~ 10) V (10 ~ 20) V (20 ~ 50) V (50 ~ 100) V (100 ~ 200) V (200 ~ 500) V	0.50 µV 4.2×10^{-4} 2.1×10^{-5} 8.5×10^{-5} 4.3×10^{-5} 2.1×10^{-5} 8.5×10^{-6} 7.2×10^{-6} 3.8×10^{-6} 2.8×10^{-6} 6.3×10^{-6} 3.7×10^{-6} 2.9×10^{-6} 6.3×10^{-6} 3.8×10^{-6} 3.3×10^{-6} 6.4×10^{-6} 3.9×10^{-6} 3.8×10^{-6}	
AC Voltage Amplitude		(10 Hz ~ 10 kHz) 1 mV (1 ~ 2) mV (2 ~ 10) mV (10 ~ 500) mV (0.5 ~ 100) V (100 ~ 200) V	0.76 µV 8.3×10^{-5} 8.4×10^{-5} 6.0×10^{-5} 5.8×10^{-5} 4.0×10^{-5}	
Sine Wave Generator		(100 ~ 600) mV 50 kHz (50 ~ 500) kHz 0.5 MHz ~ 1 GHz (1 ~ 6) GHz	0.58 mV 1.0×10^{-3} 1.7×10^{-2} 1.9×10^{-2}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Oscilloscope calibrators				
Sine Wave Generator	40404	(600 mV ~ 1 V) 50 kHz (50 ~ 500) kHz 0.5 MHz ~ 1 GHz (1 ~ 6) GHz	0.58 mV 1.0×10^{-3} 1.7×10^{-2} 1.9×10^{-2}	Calibrator/ SICT-CP-40404
Time Marker Generator		(0.1 ~ 1) ns (1 ~ 10) ns (10 ~ 100) ns 0.1 μ s ~ 10 ms (10 ~ 100) ms (0.1 ~ 1) s (1 ~ 5) s	5.8×10^{-8} 6.5×10^{-9} 3.1×10^{-9} 5.8×10^{-8} 6.1×10^{-9} 5.8×10^{-8} 1.2×10^{-8}	
Impedance Mesurement		(50 ~ 75) Ω 75 Ω ~ 1 M Ω	1.7×10^{-4} 2.1×10^{-4}	
CD/DVD meters/analyzers				
Jitter	40405	(1.0 ~ 60.0) ns 1 % 2 % 4 % 8 % 10 % 15 %	1.7×10^{-3} 0.05 % 0.09 % 0.19 % 0.36 % 0.44 % 0.67 %	Modulation Domain Analyzer/ SICT-CP-40405
Video signal generators				
NTSC, PAL Multiburst	40406	(0.1 ~ 1) MHz (1 ~ 2) MHz (2 ~ 6) MHz	6.0×10^{-2} 6.2×10^{-3} 3.1×10^{-3}	Video Measurement / SICT-CP-40406
NTSC, PAL, SECAM Pulse and Bar		(0 ~ 300) ns (0 ~ 1 000) mV	4.2×10^{-4} 3.5×10^{-3}	
NTSC, PAL , SECAM Frequency		1 Hz ~ 10 MHz	1.6×10^{-9}	
Video frequency		(10 ~ 100) Hz 100 Hz ~ 500 MHz	6.2×10^{-8} 6.2×10^{-9}	
Video level		(30 ~ 600) mV (600 ~ 1 200) mV	2.6×10^{-3} 2.3×10^{-3}	
TTL Sync level		(1 ~ 5) V	2.7×10^{-3}	
D-TV Level		(30 ~ 600) mV (600 ~ 1 200) mV	2.6×10^{-3} 2.3×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Video signal generators NTSC, PAL, H-Timming(Level) (Time)	40406	(0 ~ 100) mV (100 ~ 1 000) mV (0 ~ 254) ns (254 ~ 300) ns 300 ns ~ 3 μ s (3 ~ 7) μ s (7 ~ 10) μ s	2.6×10^{-3} 3.4×10^{-3} 1.2×10^{-2} 3.8×10^{-3} 3.2×10^{-3} 7.4×10^{-3} 4.2×10^{-3}	Video Measurement/ SICT-CP-40406
NTSC, PAL Color Bar(Luminance Level)		(0 ~ 100) mV (100 ~ 1 000) mV	0.06 mV 3.4×10^{-3}	
NTSC, PAL Color Bar(Chrominance Level)		(0 ~ 100) mV (100 ~ 1 000) mV	0.06 mV 3.4×10^{-3}	
NTSC, PAL Color Bar(Phase)		(0 ~ 360)°	0.13°	
SECAM Color Bar Frequency		(D'R & D'B) (3 ~ 5) MHz	1.2×10^{-3}	
RF Output frequency		10 kHz ~ 10 MHz (10 ~ 100) MHz (100 ~ 1 000) MHz	6.0×10^{-4} 6.0×10^{-5} 6.0×10^{-6}	
RF Output level		(0.1 ~ 10) mV (10 ~ 500) mV	1.4×10^{-2} 1.3×10^{-2}	
Sound Frequency		10 Hz ~ 100 kHz 100 kHz ~ 1 MHz	6.1×10^{-8} 6.1×10^{-7}	
Audio distortion analyzers/meters Input Frequency	40407	1 Hz ~ 200 kHz	6.1×10^{-7}	Calibrator/ SICT-CP-40407
Input Level Flatness Test		(10 ~ 100) kHz	0.008 3 dB	
Input DC Voltage		0 mV (0 ~ 1) mV (1 ~ 10) mV 10 mV ~ 300 V	0.27 μ V 5.8×10^{-3} 5.8×10^{-4} 5.8×10^{-4}	
Input Distortion		(100 Hz ~ 10 kHz) (-10 ~ -40) dB (-40 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	0.005 8 dB 0.006 0 dB 0.006 8 dB 0.012 dB 0.028 dB	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Audio distortion analyzers/meters	40407			Calibrator/ SICT-CP-40407
Input Distortion		(10 kHz ~ 50 kHz) (-10 ~ -40) dB (-40 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	0.005 9 dB 0.006 3 dB 0.008 2 dB 0.019 dB 0.052 dB	
Input AC Voltage		(10 ~ 100) Hz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 100 V (100 ~ 300) V (100 Hz ~ 1 kHz) (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 10 V (10 ~ 100) V (100 ~ 300) V (1 ~ 10) kHz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 10 V (10 ~ 100) V (10 ~ 100) kHz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 1 V (1 ~ 10) V (10 ~ 100) V	9.0×10^{-4} 4.0×10^{-4} 4.2×10^{-4} 5.3×10^{-4} 8.4×10^{-4} 1.8×10^{-4} 1.1×10^{-4} 1.0×10^{-4} 2.3×10^{-4} 8.4×10^{-4} 1.8×10^{-4} 1.5×10^{-4} 2.7×10^{-4} 1.4×10^{-3} 7.6×10^{-4} 4.8×10^{-4} 4.1×10^{-4} 3.4×10^{-4}	
Input Attenuation		(10 Hz) (30 ~ -50) dB (-50 ~ -60) dB (-60 ~ -80) dB (10 Hz ~ 10 kHz) (30 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB (10 ~ 100) kHz (30 ~ -50) dB (-50 ~ -70) dB (-70 ~ -80) dB	0.006 8 dB 0.016 dB 0.052 dB 0.008 3 dB 0.014 dB 0.042 dB 0.009 1 dB 0.023 dB 0.057 dB	
Input Impedance		300 Ω ~ 200 kΩ	3.1×10^{-4}	
Input Filter		(10 Hz ~ 100 kHz) 1 V	8.3×10^{-4}	
(Distortion meter calibrator)				
Distortion		(400 Hz , 1 kHz) (-10 ~ -20) dB (-20 ~ -40) dB (-40 ~ -60) dB (-60 ~ -80) dB	0.15 dB 0.14 dB 0.17 dB 0.26 dB	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF filters	40408			
Filter		10 Hz ~ 50 kHz (50 ~ 100) kHz (100 ~ 150) kHz	5.8×10^{-4} 1.2×10^{-3} 5.8×10^{-3}	Audio Analyzer/ SICT-CP-40408
LF/Audiosignalanalyzers	40409			
Output Frequency		1 Hz ~ 200 kHz	5.8×10^{-6}	Calibrator, Reference Multimeter/ SICT-CP-40409
AC Output Level		(10 ~ 100) Hz (1 ~ 10) mV 10 mV ~ 30 V (-20 ~ 10) dBm (100 Hz ~ 10 kHz) (1 ~ 10) mV 10 mV ~ 30 V (-20 ~ 10) dBm (10 ~ 100) kHz (1 ~ 10) mV 10 mV ~ 30 V (-20 ~ 10) dBm	8.7×10^{-4} 9.4×10^{-5} 0.005 8 dB 8.7×10^{-4} 6.5×10^{-5} 0.005 8 dB 8.7×10^{-4} 9.4×10^{-5} 0.005 8 dB	
AC Output Level Flatness		10 Hz ~ 100 kHz	0.007 1 dB	
Output Attenuation		(0 ~ -60) dB	0.005 8 dB	
Output DC Offset		(±) 0 mV (0 ~ 1) mV (1 ~ 10) mV 10 mV ~ 50 V	0.7 μ V 1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5}	
Output Impedance		5 Ω (10 ~ 600) Ω	1.2×10^{-3} 6.0×10^{-4}	
Input Frequency		1 Hz ~ 200 kHz	6.1×10^{-7}	
AC Input Level Flatness		10 Hz ~ 100 kHz	0.008 3 dB	
DC Input Level		(±) 0 mV (0 ~ 1) mV (1 ~ 10) mV 10 mV ~ 300 V	0.27 μ V 5.8×10^{-3} 5.8×10^{-4} 5.8×10^{-4}	
Input Distortion		(100 Hz ~ 10 kHz) (-10 ~ -40) dB (-40 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	0.005 8 dB 0.006 0 dB 0.006 8 dB 0.012 dB 0.028 dB	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF/Audiosignalanalyzers	40409			Calibrator, Reference Multimeter/ SICT-CP-40409
Input Distortion		(10 kHz ~ 50 kHz) (-10 ~ -40) dB (-40 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	0.005 9 dB 0.006 3 dB 0.008 2 dB 0.019 dB 0.052 dB	
AC Input Level		(10 ~ 100) Hz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 100 V (100 ~ 300) V	9.0×10^{-4} 4.0×10^{-4} 4.2×10^{-4} 5.3×10^{-4}	
		(100 Hz ~ 1 kHz) (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 10 V (10 ~ 100) V	8.4×10^{-4} 1.8×10^{-4} 1.1×10^{-4} 1.0×10^{-4} 2.3×10^{-4}	
		(1 ~ 10) kHz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 10 V (10 ~ 100) V	8.4×10^{-4} 1.8×10^{-4} 1.5×10^{-4} 2.7×10^{-4}	
		(10 ~ 100) kHz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 1 V (1 ~ 10) V (10 ~ 100) V	1.4×10^{-3} 7.6×10^{-4} 4.1×10^{-4} 3.4×10^{-4} 2.6×10^{-4}	
Input Attenuation		(10 Hz) (30 ~ -50) dB (-50 ~ -60) dB (-60 ~ -80) dB	0.006 8 dB 0.016 dB 0.052 dB	
		(10 Hz ~ 10 kHz) (30 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	0.008 3 dB 0.014 dB 0.042 dB	
		(10 ~ 100) kHz (30 ~ -50) dB (-50 ~ -70) dB (-70 ~ -80) dB	0.009 1 dB 0.023 dB 0.057 dB	
Input Impedance		300 Ω ~ 200 kΩ	3.1×10^{-4}	
Input Filter		(10 Hz ~ 100 kHz) 1 V	8.3×10^{-4}	
Line frequency meters	40410			Calibrator/ SICT-CP-40410
Frequency		16 Hz ~ 1 kHz	1.3×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Function generators	40411			
Frequency		(0.01 ~ 0.1) Hz (0.1 ~ 1) Hz 1 Hz ~ 1 GHz (1 ~ 4) GHz	5.8×10^{-6} 5.8×10^{-7} 5.8×10^{-9} 1.5×10^{-8}	Audio Analyzer, Digital Multimeter/ SICT-CP-40411
Output Level		(10 ~ 100) Hz 1 mV (1 ~ 10) mV 10 mV ~ 100 V (100 Hz ~ 10 kHz) 1 mV (1 ~ 10) mV 10 mV ~ 100 V (10 ~ 100) kHz 1 mV (1 ~ 10) mV 10 mV ~ 100 V	1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5} 1.0×10^{-3} 1.0×10^{-4} 3.0×10^{-5} 1.0×10^{-3} 1.0×10^{-4} 8.0×10^{-5}	
DC Offset		(±) 0 mV (0 ~ 1) mV (1 ~ 10) mV 10 mV ~ 20 V	0.7 μ V 0.7 μ V 1.0×10^{-4} 6.0×10^{-5}	
Level Flatness		(100 mV) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz (100 mV ~ 1 V) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz (1 ~ 30) V (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	0.099 dB 0.083 dB 0.095 dB 0.005 4 dB 0.001 1 dB 0.007 2 dB 0.021 dB 0.015 dB 0.027 dB	
Attenuation		(10 Hz ~ 100 kHz) (0 ~ 80) dB	0.006 1 dB	
Distortion		(20 Hz ~ 1 kHz) (3.16 ~ 0.010) % (1 ~ 100) kHz (3.16 ~ 0.010) %	1.5×10^{-1} 3.2×10^{-1}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Function generators				
Rise Fall Time	40411	100 μ s ~ 100 ns (100 ~ 10) ns (10 ~ 1) ns 1 ns ~ 100 ps	7.0×10^{-4} 7.8×10^{-4} 4.7×10^{-3} 4.6×10^{-2}	Audio Analyzer, Digital Multimeter/ SICT-CP-40411
Duty cycle		(1 ~ 99) %	0.006 1 %	
High Frequency Flatness Test		(100 kHz ~ 80 MHz) (0 ~ 20) dBm	0.11 dB	
FM Modulation		(0.1 ~ 400) kHz	1.2×10^{-2}	
AM Modulation		(0.1 ~ 100) %	1.2×10^{-2}	
Genescopes				
Marker Frequency	40412	9 kHz ~ 10 MHz (10 ~ 200) MHz	2.8×10^{-6} 6.4×10^{-7}	Signal Generator/ SICT-CP-40412
RF Level		9 kHz ~ 200 MHz (100 ~ 50) dB μ V	0.31 dB	
AC/DC high voltages volt meters				
DC Voltage	40413	(\pm) 0 kV (0 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 2) kV (2 ~ 100) kV	0.58 V 1.2×10^{-3} 6.1×10^{-4} 4.4×10^{-4} 3.4×10^{-4}	Calibrator/ SICT-CP-40413
AC Voltage		(50 Hz) 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 2) kV (2 ~ 3) kV (3 ~ 15 kV (15 ~ 100) kV	0.58 V 1.2×10^{-3} 6.2×10^{-4} 5.5×10^{-4} 5.3×10^{-4} 5.0×10^{-4} 5.7×10^{-4}	
		(60 Hz) 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 2) kV (2 ~ 3) kV (3 ~ 15 kV (15 ~ 100) kV	0.58 V 1.2×10^{-3} 6.2×10^{-4} 5.5×10^{-4} 4.7×10^{-4} 4.5×10^{-4} 5.4×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Jitter meters	40415			Modulation Domain Analyzer/ SICT-CP-40415
CD/DVD Jitter		(1 ~ 20) ns (20 ~ 60) ns	1.7×10^{-3} 1.6×10^{-3}	
VTR Jitter		0.05 μs (0.05 ~ 0.1) μs (0.1 ~ 0.2) μs (0.2 ~ 0.5) μs (0.5 ~ 0.7) μs 1 % 2 % 4 % 8 % 10 % 15 %	0.66 ns 0.77 ns 1.2 ns 2.8 ns 4.3 ns 0.05 % 0.09 % 0.19 % 0.36 % 0.44 % 0.67 %	
Leakage current testers	40416			Calibrator/ SICT-CP-40416
DC Current		0 μA (0 ~ 1) μA (1 ~ 2) μA (2 ~ 5) μA (5 ~ 10) μA (10 ~ 20) μA (20 ~ 50) μA (50 ~ 100) μA (100 ~ 200) μA (0.2 ~ 100) mA	7.0 nA 2.4×10^{-3} 3.6×10^{-3} 1.4×10^{-3} 7.4×10^{-4} 4.0×10^{-4} 1.8×10^{-4} 1.3×10^{-4} 8.5×10^{-5} 6.1×10^{-4}	
AC Current		(20 μA) 10 Hz (10 ~ 20) Hz (0.02 ~ 1) kHz (1 ~ 5) kHz (5 ~ 10) kHz (20 ~ 50) μA 10 Hz (10 ~ 20) Hz (0.02 ~ 1) kHz (1 ~ 5) kHz (5 ~ 10) kHz (50 ~ 100) μA 10 Hz (10 ~ 20) Hz (0.02 ~ 1) kHz (1 ~ 5) kHz (5 ~ 10) kHz	26 nA 8.5×10^{-4} 7.0×10^{-4} 1.3×10^{-3} 5.5×10^{-3} 6.8×10^{-4} 4.4×10^{-4} 3.4×10^{-4} 6.8×10^{-4} 2.8×10^{-3} 4.9×10^{-4} 3.2×10^{-4} 2.3×10^{-4} 4.9×10^{-4} 4.0×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers				
AC Current	40416	(100 ~ 200) μ A 10 Hz (10 ~ 20) Hz (0.02 ~ 1) kHz (1 ~ 5) kHz (5 ~ 10) kHz (200 ~ 500) μ A 10 Hz (10 ~ 20) Hz (0.02 ~ 1) kHz (1 ~ 5) kHz (5 ~ 10) kHz (0.5 ~ 1) mA 10 Hz (10 ~ 20) Hz (0.02 ~ 1) kHz (1 ~ 5) kHz (5 ~ 10) kHz (1 ~ 100) mA 10 Hz (10 ~ 20) Hz (0.02 ~ 1) kHz (1 ~ 5) kHz (5 ~ 10) kHz	3.9×10^{-4} 2.5×10^{-4} 1.7×10^{-4} 4.0×10^{-4} 1.7×10^{-3} 4.4×10^{-4} 3.2×10^{-4} 2.4×10^{-4} 5.4×10^{-4} 2.8×10^{-3} 7.0×10^{-4} 6.6×10^{-4} 6.3×10^{-4} 7.2×10^{-4} 2.1×10^{-3} 7.0×10^{-4} 6.6×10^{-4} 6.3×10^{-4} 7.0×10^{-4} 2.7×10^{-3}	Calibrator/ SICT-CP-40416
DC Voltage		0 V (0 ~ 0.1) V (0.1 ~ 0.2) V (0.2 ~ 0.5) V (0.5 ~ 1) V (1 ~ 2) V (2 ~ 5) V (5 ~ 10) V (10 ~ 20) V (20 ~ 50) V (50 ~ 100) V (100 ~ 200) V (200 ~ 300) V (300 ~ 500) V (500 ~ 1 000) V	0.06 mV 6.0×10^{-4} 3.0×10^{-4} 1.2×10^{-4} 6.0×10^{-5} 3.1×10^{-4} 1.2×10^{-4} 6.1×10^{-5} 3.1×10^{-5} 1.4×10^{-5} 8.8×10^{-6} 3.1×10^{-5} 2.3×10^{-5} 1.2×10^{-4} 5.8×10^{-5}	
AC Voltage		0.1 V 10 Hz (0.01 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz	0.074 mV 6.5×10^{-4} 8.3×10^{-4} 1.2×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers	40416	(0.1 ~ 0.2) V 10 Hz (0.01 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (0.2 ~ 0.5) V 10 Hz (10 ~ 20) Hz (0.02 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (0.5 ~ 1) V 10 Hz (10 ~ 20) Hz (0.02 ~ 20) kHz (20 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (1 ~ 2) V 10 Hz (0.01 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (2 ~ 5) V 10 Hz (0.01 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz	4.5×10^{-4} 3.5×10^{-4} 5.5×10^{-4} 9.5×10^{-4} 1.8×10^{-3} 3.4×10^{-3} 4.0×10^{-4} 2.0×10^{-4} 1.7×10^{-4} 2.2×10^{-4} 6.0×10^{-4} 1.7×10^{-3} 2.8×10^{-3} 3.3×10^{-4} 1.4×10^{-4} 9.1×10^{-5} 1.5×10^{-4} 4.9×10^{-4} 1.4×10^{-3} 2.4×10^{-3} 4.3×10^{-4} 3.3×10^{-4} 5.5×10^{-4} 1.3×10^{-3} 2.1×10^{-3} 4.0×10^{-4} 2.0×10^{-4} 5.0×10^{-4} 1.7×10^{-3} 2.8×10^{-3}	Calibrator/ SICT-CP-40416

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers	40416			Calibrator/ SICT-CP-40416
AC Voltage		(5 ~ 10) V 10 Hz (10 ~ 20) Hz (0.02 ~ 20) kHz (20 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (10 ~ 20) V 10 Hz (10 ~ 20) Hz (0.02 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (20 ~ 50) V 10 Hz (10 ~ 20) Hz (0.02 ~ 50) kHz (50 ~ 100) kHz (50 ~ 100) V 10 Hz (0.01 ~ 50) kHz (50 ~ 100) kHz (100 ~ 1 000) V (0.05 ~ 1) kHz	3.3×10^{-4} 1.4×10^{-4} 8.9×10^{-5} 1.4×10^{-4} 3.9×10^{-4} 1.4×10^{-3} 2.2×10^{-3} 3.1×10^{-4} 1.2×10^{-4} 6.0×10^{-5} 9.0×10^{-5} 1.1×10^{-4} 4.2×10^{-4} 2.2×10^{-4} 1.8×10^{-4} 2.8×10^{-4} 3.4×10^{-4} 1.4×10^{-4} 2.2×10^{-4} 1.1×10^{-4}	
Resistance		100 mΩ 1 Ω ~ 10 kΩ	7.7 μΩ 6.2×10^{-5}	
Input Voltage to Output Current Display(U1)		20 Hz (4.75 ~ 5.25) mA 50 Hz (4.77 ~ 5.28) mA 60 Hz (4.77 ~ 5.28) mA 100 Hz (4.85 ~ 5.36) mA 200 Hz (5.11 ~ 5.65) mA	0.006 3 mA 0.006 1 mA 0.006 1 mA 0.006 1 mA 0.006 1 mA	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers Input Voltage to Output Current Display(U1)	40416	500 Hz (6.64 ~ 7.34) mA	0.006 1 mA	Calibrator/ SICT-CP-40416
		1 kHz (9.70 ~ 10.73) mA	0.006 1 mA	
		2 kHz (14.07 ~ 15.56) mA	0.006 2 mA	
		5 kHz (17.82 ~ 19.70) mA	0.006 2 mA	
		10 kHz (18.66 ~ 20.63) mA	0.006 2 mA	
		20 kHz (18.92 ~ 20.92) mA	0.006 2 mA	
		50 kHz (19.00 ~ 21.00) mA	0.006 4 mA	
		100 kHz (19.00 ~ 21.00) mA	0.006 7 mA	
		200 kHz (19.00 ~ 21.00) mA	0.010 mA	
		500 kHz (19.00 ~ 21.00) mA	0.030 mA	
		1 MHz (19.00 ~ 21.00) mA	0.046 mA	
Input Voltage to Output Current Display(U2)		20 Hz (4.75 ~ 5.25) mA	0.006 3 mA	
		50 Hz (4.77 ~ 5.28) mA	0.006 1 mA	
		60 Hz (4.77 ~ 5.28) mA	0.006 1 mA	
		100 Hz (4.80 ~ 5.30) mA	0.006 1 mA	
		200 Hz (4.92 ~ 5.44) mA	0.006 1 mA	
		500 Hz (5.37 ~ 5.93) mA	0.006 1 mA	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers Input Voltage to Output Current Display(U2)	40416	1 kHz (5.56 ~ 6.14) mA	0.006 1 mA	Calibrator/ SICT-CP-40416
		2 kHz (4.68 ~ 5.17) mA	0.006 1 mA	
		5 kHz (2.53 ~ 2.80) mA	0.000 63 mA	
		10 kHz (1.35 ~ 1.49) mA	0.000 62 mA	
		20 kHz (0.683 ~ 0.755) mA	0.000 61 mA	
		50 kHz (274.57 ~ 303.47) mA	0.029 μ A	
		100 kHz (137.48 ~ 151.95) μ A	0.020 μ A	
		200 kHz (68.82 ~ 76.06) μ A	0.030 μ A	
		500 kHz (27.43 ~ 30.32) μ A	0.042 μ A	
		1 MHz (13.71 ~ 15.16) μ A	0.033 μ A	
Input Voltage to Output Current Display(U3)		20 Hz (4.75 ~ 5.25) mA	0.006 3 mA	
		50 Hz (4.77 ~ 5.28) mA	0.006 1 mA	
		60 Hz (4.77 ~ 5.28) mA	0.006 1 mA	
		100 Hz (4.80 ~ 5.30) mA	0.006 1 mA	
		200 Hz (4.95 ~ 5.47) mA	0.006 1 mA	
		500 Hz (5.65 ~ 6.25) mA	0.006 1 mA	
		1 kHz (6.60 ~ 7.29) mA	0.006 1 mA	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers Input Voltage to Output Current Display(U3)	40416	2 kHz (7.14 ~ 7.89) mA	0.006 1 mA	Calibrator/ SICT-CP-40416
		5 kHz (5.31 ~ 5.87) mA	0.006 1 mA	
		10 kHz (3.12 ~ 3.45) mA	0.000 64 mA	
		20 kHz (1.63 ~ 1.81) mA	0.000 62 mA	
		50 kHz (0.664 ~ 0.734) mA	0.000 62 mA	
		100 kHz (322.16 ~ 367.12) µA	0.046 µA	
		200 kHz (166.03 ~ 183.81) µA	0.070 µA	
		500 kHz (66.37 ~ 73.35) µA	0.10 µA	
		1 MHz (33.14 ~ 36.63) µA	0.08 µA	
Input Voltage to Output Voltage Ratio(U1)		4.00 (20 Hz)	1.3×10^{-4}	
		3.98 (50 Hz)	6.5×10^{-5}	
		3.97 (60 Hz)	6.5×10^{-5}	
		3.92 (100 Hz)	6.5×10^{-5}	
		3.72 (200 Hz)	6.5×10^{-5}	
		2.87 (500 Hz)	6.4×10^{-5}	
		1.96 (1 kHz)	6.4×10^{-5}	
		1.96 (2 kHz)	6.4×10^{-5}	
		1.96 (5 kHz)	6.4×10^{-5}	
		1.96 (10 kHz)	6.4×10^{-5}	
		1.00 (20 kHz)	6.7×10^{-5}	
		1.00 (50 kHz)	9.6×10^{-5}	
		1.00 (100 kHz)	1.2×10^{-4}	
		1.00 (200 kHz)	4.2×10^{-4}	
		1.00 (500 kHz)	1.5×10^{-3}	
		1.00 (1 MHz)	2.6×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers				
Input Voltage to Output Voltage Ratio(U2)	40416	4.00 (20 Hz) 3.99 (50 Hz) 3.99 (60 Hz) 3.96 (100 Hz) 3.87 (200 Hz) 3.54 (500 Hz) 3.43 (1 kHz) 4.06 (2 kHz) 7.50 (5 kHz) 14.1 (10 kHz) 27.8 (20 kHz) 69.2 (50 kHz) 138 (100 kHz) 272 (200 kHz) 691 (500 kHz) 1 382 (1 MHz)	1.3×10^{-4} 6.5×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.6×10^{-5} 7.0×10^{-5} 3.6×10^{-5} 4.8×10^{-5} 1.0×10^{-4} 2.7×10^{-4} 1.1×10^{-3} 3.0×10^{-3}	Calibrator/ SICT-CP-40416
Input Voltage to Output Voltage Ratio(U3)		4.00 (20 Hz) 3.99 (50 Hz) 3.98 (60 Hz) 3.95 (100 Hz) 3.83 (200 Hz) 2.36 (500 Hz) 2.87 (1 kHz) 2.65 (2 kHz) 3.57 (5 kHz) 6.09 (10 kHz) 11.6 (20 kHz) 28.7 (50 kHz) 57.2 (100 kHz) 114 (200 kHz) 286 (500 kHz) 572 (1 MHz)	1.3×10^{-4} 6.5×10^{-5} 6.5×10^{-5} 6.6×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.4×10^{-5} 6.4×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.9×10^{-5} 3.8×10^{-5} 7.9×10^{-5} 1.9×10^{-4} 6.1×10^{-4} 2.3×10^{-3}	
mAs Meter		1 mAs (1 ~ 2 000) mAs (2 000 ~ 9 999) mAs	1.2×10^{-3} 1.0×10^{-3} 1.1×10^{-3}	
Electronic AC/DC loads				
DC Voltage	40417	0 mV (0 ~ 5) mV (5 ~ 20) mV (20 ~ 100) mV (0.1 ~ 1) V (1 ~ 2) V (2 ~ 4) V (4 ~ 7) V (7 ~ 9) V (9 ~ 10) V (10 ~ 50) V (50 ~ 100) V (100 ~ 200) V (200 ~ 400) V (400 ~ 1 000) V	0.058 mV 5.8×10^{-2} 5.8×10^{-3} 1.2×10^{-3} 6.2×10^{-5} 3.2×10^{-5} 2.1×10^{-5} 1.3×10^{-5} 9.1×10^{-6} 7.9×10^{-6} 3.1×10^{-5} 1.0×10^{-5} 3.4×10^{-5} 2.5×10^{-5} 1.6×10^{-5}	Calibrator/ SICT-CP-40417

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electronic AC/DC loads	40417			
DC Current		1 mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 20) mA (20 ~ 50) mA (50 ~ 100) mA (0.1 ~ 0.2) A (0.2 ~ 0.4) A (0.4 ~ 0.6) A (0.6 ~ 0.8) A (0.8 ~ 1) A (1 ~ 3) A (3 ~ 6) A (6 ~ 10) A (10 ~ 40) A (40 ~ 100) A (100 ~ 1 000) A (1 000 ~ 2 000) A	5.8 μA 2.9×10^{-3} 1.2×10^{-3} 5.8×10^{-4} 1.2×10^{-4} 5.9×10^{-5} 2.9×10^{-4} 1.9×10^{-4} 1.2×10^{-4} 8.4×10^{-5} 6.6×10^{-5} 5.1×10^{-5} 2.6×10^{-5} 1.6×10^{-5} 4.0×10^{-5} 2.9×10^{-5} 1.4×10^{-4} 4.9×10^{-4}	Calibrator/ SICT-CP-40417
Charge voltage		0 mV (0 ~ 5) mV (5 ~ 20) mV (20 ~ 100) mV (0.1 ~ 1) V (1 ~ 2) V (2 ~ 4) V (4 ~ 7) V (7 ~ 9) V (9 ~ 10) V (10 ~ 50) V (50 ~ 100) V (100 ~ 200) V (200 ~ 400) V (400 ~ 1 000) V (1 000 ~ 1 200) V (1 200 ~ 1 400) V (1 400 ~ 1 500) V	0.058 mV 5.8×10^{-2} 5.8×10^{-3} 1.2×10^{-3} 6.2×10^{-5} 3.2×10^{-5} 2.1×10^{-5} 1.3×10^{-5} 9.1×10^{-6} 7.9×10^{-6} 3.1×10^{-5} 1.0×10^{-5} 3.4×10^{-5} 2.5×10^{-5} 1.6×10^{-5} 1.0×10^{-3} 9.2×10^{-4} 8.7×10^{-4}	
Charge and Discharge Current		(±) 1 mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 20) mA (20 ~ 50) mA (50 ~ 100) mA (0.1 ~ 0.2) A (0.2 ~ 0.4) A (0.4 ~ 0.6) A (0.6 ~ 0.8) A (0.8 ~ 1) A (1 ~ 3) A (3 ~ 6) A (6 ~ 10) A (10 ~ 40) A (40 ~ 100) A (100 ~ 1 000) A (1 000 ~ 3 000) A	5.8 μA 2.9×10^{-3} 1.2×10^{-3} 5.8×10^{-4} 1.2×10^{-4} 5.9×10^{-5} 2.9×10^{-4} 1.9×10^{-4} 1.2×10^{-4} 8.4×10^{-5} 6.6×10^{-5} 5.1×10^{-5} 2.6×10^{-5} 1.6×10^{-5} 4.0×10^{-5} 2.9×10^{-5} 1.4×10^{-4} 4.9×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electronic AC/DC loads	40417			
Resistance		0.1 Ω (0.1 ~ 1) Ω (1 ~ 2) Ω (2 ~ 4) Ω (4 ~ 500) Ω (0.5 ~ 2) kΩ (2 ~ 10) kΩ	0.58 mΩ 2.1×10^{-3} 4.0×10^{-3} 2.3×10^{-3} 1.6×10^{-3} 9.0×10^{-4} 2.0×10^{-4}	Calibrator/ SICT-CP-40417
AC Voltage		(0.001 V) (40 ~ 400) Hz (0.001 ~ 0.1) V (40 ~ 400) Hz (0.1 ~ 0.2) V (40 ~ 400) Hz (0.2 ~ 0.5) V (40 ~ 400) Hz (0.5 ~ 2) V (40 ~ 400) Hz (2 ~ 3) V (40 ~ 400) Hz (3 ~ 7) V (40 ~ 400) Hz (7 ~ 20) V (40 ~ 50) Hz (50 ~ 400) Hz (20 ~ 80) V (40 ~ 50) Hz (50 ~ 400) Hz (80 ~ 200) V (40 ~ 400) Hz (200 ~ 500) V (50 ~ 400) Hz	0.61 mV 6.1×10^{-3} 3.1×10^{-3} 1.2×10^{-3} 6.2×10^{-4} 2.8×10^{-4} 2.3×10^{-4} 1.5×10^{-4} 9.8×10^{-5} 2.1×10^{-4} 1.2×10^{-4} 1.3×10^{-4} 1.8×10^{-4}	
AC Current		(1 mA) (40 ~ 400) Hz (1 ~ 100) mA (40 ~ 400) Hz (100 mA ~ 0.2 A) (40 ~ 400) Hz	0.58 mA 5.8×10^{-2} 5.8×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electronic AC/DC loads	AC Current	40417	(0.2 ~ 0.6) A (40 ~ 400) Hz	2.0×10^{-3}
			(0.6 ~ 2) A (40 ~ 400) Hz	9.4×10^{-4}
			(2 ~ 5) A (40 ~ 400) Hz	1.2×10^{-3}
			(5 ~ 20) A (40 ~ 400) Hz	1.0×10^{-3}
			(1 Ω) (40 ~ 400) Hz	$1.0 \text{ m}\Omega$
	AC Resistance		(1 ~ 50) Ω (40 ~ 400) Hz	1.5×10^{-3}
			(50 ~ 100) Ω (40 ~ 400) Hz	1.1×10^{-3}
			(100 Ω ~ 10 kΩ) (40 ~ 400) Hz	1.7×10^{-3}
Modulation meters	Frequency Modulation	40418	0 kHz (0 ~ 400) kHz	1 Hz 1.2×10^{-2}
			0 % (0 ~ 100) %	0.01 % 1.2×10^{-2}
			0 rad (0 ~ 400) rad	1.2 mrad 1.2×10^{-2}
	Amplitude Modulation			
Analogue/Digital multimeters	DC Voltage	40419	(±) 0 mV	$0.43 \mu\text{V}$
			(0 ~ 1) mV	5.0×10^{-4}
			(1 ~ 2) mV	2.5×10^{-4}
			(2 ~ 5) mV	1.0×10^{-4}
			(5 ~ 10) mV	5.0×10^{-5}
			(10 ~ 15) mV	3.3×10^{-5}
			(15 ~ 20) mV	2.5×10^{-6}
			(20 ~ 50) mV	1.2×10^{-5}
			(0.05 ~ 0.2) V	8.0×10^{-6}
			(0.2 ~ 0.5) V	4.8×10^{-6}
			(0.5 ~ 1) V	3.8×10^{-6}
			(1 ~ 2) V	4.0×10^{-6}
			(2 ~ 5) V	2.6×10^{-6}
			(5 ~ 10) V	2.3×10^{-6}
			(10 ~ 20) V	6.0×10^{-6}
			(20 ~ 50) V	4.0×10^{-6}
			(50 ~ 100) V	3.5×10^{-6}
			(100 ~ 200) V	8.0×10^{-6}
			(200 ~ 500) V	5.2×10^{-6}
			(500 ~ 1 000) V	4.5×10^{-6}

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters	40419	(0.6 mV) 1 kHz (1 mV) 10 Hz (10 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (1 ~ 2) mV 10 Hz (10 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (2 ~ 5) mV 10 Hz (10 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (5 ~ 10) mV 10 Hz (10 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (10 ~ 15) mV 10 Hz (10 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz	4.1 μ V 4.2 μ V 4.2 μ V 4.2 μ V 5.5 μ V 11 μ V 21 μ V 23 μ V 2.2 \times 10 ⁻³ 2.2 \times 10 ⁻³ 2.1 \times 10 ⁻³ 2.2 \times 10 ⁻³ 3.0 \times 10 ⁻³ 6.0 \times 10 ⁻³ 1.1 \times 10 ⁻² 1.3 \times 10 ⁻² 1.1 \times 10 ⁻³ 9.2 \times 10 ⁻⁴ 9.0 \times 10 ⁻⁴ 1.0 \times 10 ⁻³ 1.5 \times 10 ⁻³ 3.0 \times 10 ⁻³ 5.2 \times 10 ⁻³ 6.8 \times 10 ⁻³ 6.3 \times 10 ⁻⁴ 5.0 \times 10 ⁻⁴ 4.9 \times 10 ⁻⁴ 5.9 \times 10 ⁻⁴ 9.5 \times 10 ⁻⁴ 1.9 \times 10 ⁻³ 3.2 \times 10 ⁻³ 4.3 \times 10 ⁻³ 4.8 \times 10 ⁻⁴ 3.6 \times 10 ⁻⁴ 3.5 \times 10 ⁻⁴ 4.5 \times 10 ⁻⁴ 7.6 \times 10 ⁻⁴ 1.5 \times 10 ⁻³ 2.5 \times 10 ⁻³ 3.7 \times 10 ⁻³	Calibrator/ SICT-CP-40419

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters	40419			Calibrator/ SICT-CP-40419
AC Voltage		(15 ~ 20) mV		
		10 Hz	4.1×10^{-4}	
		(10 ~ 40) Hz	2.9×10^{-4}	
		(0.04 ~ 20) kHz	2.8×10^{-4}	
		(20 ~ 50) kHz	3.8×10^{-4}	
		(50 ~ 100) kHz	6.7×10^{-4}	
		(100 ~ 300) kHz	1.4×10^{-3}	
		(300 ~ 500) kHz	2.2×10^{-3}	
		(0.5 ~ 1) MHz	3.4×10^{-3}	
		(20 ~ 50) mV		
		10 Hz	4.6×10^{-4}	
		(10 ~ 40) Hz	2.6×10^{-4}	
		(0.04 ~ 20) kHz	2.1×10^{-4}	
		(20 ~ 50) kHz	2.7×10^{-4}	
		(50 ~ 100) kHz	6.6×10^{-4}	
		(100 ~ 300) kHz	1.0×10^{-3}	
		(300 ~ 500) kHz	1.6×10^{-3}	
		(0.5 ~ 1) MHz	3.3×10^{-3}	
		(50 ~ 100) mV		
		10 Hz	3.3×10^{-4}	
		(10 ~ 40) Hz	1.6×10^{-4}	
		(0.04 ~ 20) kHz	1.3×10^{-4}	
		(20 ~ 50) kHz	1.9×10^{-4}	
		(50 ~ 100) kHz	4.8×10^{-4}	
		(100 ~ 300) kHz	7.6×10^{-4}	
		(300 ~ 500) kHz	1.3×10^{-3}	
		(0.5 ~ 1) MHz	2.8×10^{-3}	
		(100 ~ 150) mV		
		10 Hz	2.9×10^{-4}	
		(10 ~ 40) Hz	1.3×10^{-4}	
		(0.04 ~ 20) kHz	1.0×10^{-4}	
		(20 ~ 50) kHz	1.6×10^{-4}	
		(50 ~ 100) kHz	4.1×10^{-4}	
		(100 ~ 300) kHz	6.8×10^{-4}	
		(300 ~ 500) kHz	1.3×10^{-3}	
		(0.5 ~ 1) MHz	2.6×10^{-3}	
		(150 ~ 200) mV		
		10 Hz	2.7×10^{-4}	
		(10 ~ 40) Hz	1.2×10^{-4}	
		(0.04 ~ 20) kHz	9.2×10^{-5}	
		(20 ~ 50) kHz	1.4×10^{-4}	
		(50 ~ 100) kHz	3.9×10^{-4}	
		(100 ~ 300) kHz	6.4×10^{-4}	
		(300 ~ 500) kHz	1.2×10^{-3}	
		(0.5 ~ 1) MHz	2.6×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters	40419			Calibrator/ SICT-CP-40419
AC Voltage		(0.2 ~ 0.5) V		
		10 Hz	3.0×10^{-4}	
		(10 ~ 20) Hz	1.3×10^{-4}	
		(20 ~ 40) Hz	9.0×10^{-5}	
		(0.04 ~ 20) kHz	6.6×10^{-5}	
		(20 ~ 50) kHz	1.0×10^{-4}	
		(50 ~ 100) kHz	1.6×10^{-4}	
		(100 ~ 300) kHz	4.6×10^{-4}	
		(300 ~ 500) kHz	1.2×10^{-3}	
		(0.5 ~ 1) MHz	2.7×10^{-3}	
		(0.5 ~ 1) V		
		10 Hz	2.5×10^{-4}	
		(10 ~ 20) Hz	1.0×10^{-4}	
		(20 ~ 40) Hz	5.8×10^{-5}	
		(0.04 ~ 20) kHz	4.9×10^{-5}	
		(20 ~ 50) kHz	7.7×10^{-5}	
		(50 ~ 100) kHz	1.2×10^{-4}	
		(100 ~ 300) kHz	3.6×10^{-4}	
		(300 ~ 500) kHz	1.0×10^{-3}	
		(0.5 ~ 1) MHz	2.5×10^{-3}	
		(1 ~ 2) V		
		10 Hz	2.2×10^{-4}	
		(10 ~ 20) Hz	8.5×10^{-5}	
		(20 ~ 40) Hz	4.5×10^{-5}	
		(0.04 ~ 20) kHz	4.2×10^{-5}	
		(20 ~ 50) kHz	6.8×10^{-5}	
		(50 ~ 100) kHz	9.6×10^{-5}	
		(100 ~ 300) kHz	3.2×10^{-4}	
		(300 ~ 500) kHz	9.0×10^{-4}	
		(0.5 ~ 1) MHz	2.4×10^{-3}	
		(2 ~ 5) V		
		10 Hz	3.0×10^{-4}	
		(10 ~ 20) Hz	1.3×10^{-4}	
		(20 ~ 40) Hz	8.2×10^{-5}	
		(0.04 ~ 20) kHz	6.2×10^{-5}	
		(20 ~ 50) kHz	1.0×10^{-4}	
		(50 ~ 100) kHz	1.4×10^{-4}	
		(100 ~ 300) kHz	4.4×10^{-4}	
		(300 ~ 500) kHz	1.3×10^{-3}	
		(0.5 ~ 1) MHz	2.2×10^{-3}	
		(5 ~ 10) V		
		10 Hz	2.5×10^{-4}	
		(10 ~ 20) Hz	9.8×10^{-5}	
		(20 ~ 40) Hz	5.4×10^{-5}	
		(0.04 ~ 20) kHz	4.7×10^{-5}	
		(20 ~ 50) kHz	7.7×10^{-5}	
		(50 ~ 100) kHz	1.1×10^{-4}	
		(100 ~ 300) kHz	3.2×10^{-4}	
		(300 ~ 500) kHz	1.0×10^{-3}	
		(0.5 ~ 1) MHz	1.7×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters	40419			
AC Voltage		(10 ~ 15) V	2.3×10^{-4}	Calibrator/ SICT-CP-40419
		10 Hz	8.7×10^{-5}	
		(10 ~ 20) Hz	4.6×10^{-5}	
		(20 ~ 40) Hz	4.3×10^{-5}	
		(0.04 ~ 20) kHz	7.1×10^{-5}	
		(20 ~ 50) kHz	9.5×10^{-5}	
		(50 ~ 100) kHz	2.9×10^{-4}	
		(100 ~ 300) kHz	9.4×10^{-4}	
		(300 ~ 500) kHz	1.5×10^{-3}	
		(0.5 ~ 1) MHz		
		(15 ~ 20) V	2.2×10^{-4}	
		10 Hz	8.5×10^{-5}	
		(10 ~ 20) Hz	4.3×10^{-5}	
		(20 ~ 40) Hz	4.1×10^{-5}	
		(0.04 ~ 20) kHz	6.8×10^{-5}	
		(50 ~ 100) kHz	9.0×10^{-5}	
		(100 ~ 300) kHz	2.8×10^{-4}	
		(300 ~ 500) kHz	9.1×10^{-4}	
		(0.5 ~ 1) MHz	1.4×10^{-3}	
		(20 ~ 50) V	3.2×10^{-4}	
		10 Hz	1.4×10^{-4}	
		(10 ~ 20) Hz	9.8×10^{-5}	
		(20 ~ 40) Hz	7.4×10^{-5}	
		(0.04 ~ 20) kHz	1.1×10^{-5}	
		(50 ~ 100) kHz	2.1×10^{-4}	
		(50 ~ 100) V	2.5×10^{-4}	
		10 Hz	1.0×10^{-4}	
		(10 ~ 20) Hz	6.5×10^{-5}	
		(20 ~ 40) Hz	5.6×10^{-5}	
		(0.04 ~ 20) kHz	8.5×10^{-5}	
		(50 ~ 100) kHz	1.6×10^{-4}	
		(100 ~ 200) V	2.3×10^{-4}	
		10 Hz	9.3×10^{-5}	
		(10 ~ 20) Hz	5.6×10^{-5}	
		(20 ~ 40) Hz	5.1×10^{-5}	
		(0.04 ~ 20) kHz	7.9×10^{-5}	
		(50 ~ 100) kHz	1.4×10^{-4}	
		(200 ~ 500) V	6.7×10^{-5}	
		50 Hz ~ 1 kHz		
		(500 ~ 1 000) V	6.3×10^{-5}	
		50 Hz ~ 1 kHz		

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters	Resistance	0 Ω	0.001 0 mΩ	Calibrator/ SICT-CP-40419
		0 Ω ~ 10 kΩ	1.2×10^{-6}	
		(10 ~ 100) kΩ	1.4×10^{-6}	
	DC Current	(0.1 ~ 1) MΩ	7.2×10^{-6}	
		(1 ~ 10) MΩ	7.7×10^{-6}	
		(10 ~ 100) MΩ	1.2×10^{-5}	
		(0.1 ~ 1) GΩ	3.2×10^{-4}	
		(±)		
		0 nA	6.0 nA	
		(0 ~ 1) nA	6.9×10^{-3}	
		(1 ~ 100) nA	4.6×10^{-3}	
		(0.1 ~ 1) μA	6.0×10^{-3}	
		(1 ~ 2) μA	3.1×10^{-3}	
		(2 ~ 5) μA	1.2×10^{-3}	
		(5 ~ 10) μA	6.3×10^{-4}	
		(10 ~ 20) μA	3.5×10^{-4}	
		(20 ~ 50) μA	1.6×10^{-4}	
		(50 ~ 100) μA	9.4×10^{-5}	
		(100 ~ 200) μA	6.3×10^{-5}	
		(0.2 ~ 0.5) mA	4.8×10^{-5}	
		(0.5 ~ 1) mA	3.5×10^{-5}	
		(1 ~ 1.5) mA	3.1×10^{-5}	
		(1.5 ~ 2) mA	3.0×10^{-5}	
		(2 ~ 5) mA	4.4×10^{-5}	
		(5 ~ 10) mA	3.2×10^{-5}	
		(10 ~ 15) mA	2.9×10^{-5}	
		(15 ~ 20) mA	2.8×10^{-5}	
		(20 ~ 50) mA	5.6×10^{-5}	
		(50 ~ 100) mA	4.4×10^{-5}	
		(100 ~ 150) mA	4.1×10^{-5}	
		(150 ~ 200) mA	3.9×10^{-5}	
		(0.2 ~ 0.5) A	9.4×10^{-5}	
		(0.5 ~ 1) A	6.9×10^{-5}	
		(1 ~ 1.5) A	6.1×10^{-5}	
		(1.5 ~ 2) A	5.8×10^{-5}	
		(2 ~ 3) A	3.3×10^{-4}	
		(3 ~ 5) A	2.4×10^{-4}	
		(5 ~ 10) A	1.6×10^{-4}	
		(10 ~ 20) A	1.2×10^{-4}	
		(20 ~ 30) A	2.4×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters	40419	(20 μ A) 1 kHz 10 kHz (20 ~ 50) μ A 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (50 ~ 100) μ A 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (0.1 ~ 0.2) mA 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (0.2 ~ 0.5) mA 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (0.5 ~ 1) mA 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (1 ~ 2) mA 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (2 ~ 5) mA 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	11 nA 81 nA 1.4×10^{-4} 9.0×10^{-5} 7.3×10^{-5} 1.5×10^{-4} 5.5×10^{-4} 1.9×10^{-4} 1.2×10^{-4} 9.5×10^{-5} 2.0×10^{-4} 7.5×10^{-4} 3.0×10^{-4} 1.9×10^{-4} 1.4×10^{-5} 3.1×10^{-4} 1.2×10^{-3} 9.0×10^{-5} 7.0×10^{-5} 6.0×10^{-5} 1.2×10^{-4} 5.7×10^{-4} 1.4×10^{-4} 1.0×10^{-4} 8.0×10^{-5} 1.6×10^{-4} 7.6×10^{-4} 2.4×10^{-4} 1.6×10^{-4} 1.2×10^{-4} 2.4×10^{-4} 1.2×10^{-3} 9.0×10^{-5} 7.0×10^{-5} 5.2×10^{-5} 1.1×10^{-4} 5.4×10^{-4}	Calibrator/ SICT-CP-40419

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters	40419	(5 ~ 10) mA 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (10 ~ 20) mA 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (20 ~ 50) mA 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (50 ~ 100) mA 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (0.1 ~ 0.2) A 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (0.2 ~ 1) A 40 Hz 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (1 ~ 2) A 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (2 ~ 3) A (40 ~ 100) Hz 100 Hz ~ 1 kHz (1 ~ 10) kHz	1.4 × 10 ⁻⁴ 1.0 × 10 ⁻⁴ 7.3 × 10 ⁻⁵ 1.4 × 10 ⁻⁴ 7.2 × 10 ⁻⁴ 2.4 × 10 ⁻⁴ 1.6 × 10 ⁻⁴ 1.2 × 10 ⁻⁴ 2.2 × 10 ⁻⁴ 1.1 × 10 ⁻³ 1.0 × 10 ⁻⁴ 8.0 × 10 ⁻⁵ 4.8 × 10 ⁻⁵ 1.1 × 10 ⁻⁴ 4.0 × 10 ⁻⁴ 1.4 × 10 ⁻⁴ 1.0 × 10 ⁻⁴ 6.8 × 10 ⁻⁵ 1.4 × 10 ⁻⁴ 6.0 × 10 ⁻⁴ 2.4 × 10 ⁻⁴ 1.6 × 10 ⁻⁴ 1.1 × 10 ⁻⁴ 2.1 × 10 ⁻⁴ 1.0 × 10 ⁻³ 1.4 × 10 ⁻⁴ 1.4 × 10 ⁻⁴ 2.6 × 10 ⁻⁴ 2.7 × 10 ⁻³ 2.4 × 10 ⁻⁴ 4.2 × 10 ⁻⁴ 5.2 × 10 ⁻³ 1.8 × 10 ⁻⁴ 1.9 × 10 ⁻⁴ 9.9 × 10 ⁻⁴	Calibrator/ SICT-CP-40419

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters				
AC Current	40419	(3 ~ 5) A (40 ~ 100) Hz 100 Hz ~ 1 kHz (1 ~ 10) kHz (5 ~ 10) A (40 ~ 100) Hz 100 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 20) A (40 ~ 60) Hz (60 ~ 100) Hz 100 Hz ~ 1 kHz (1 ~ 10) kHz (20 ~ 30) A (40 ~ 60) Hz (60 ~ 100) Hz 100 Hz ~ 1 kHz (1 ~ 10) kHz	2.4×10^{-4} 2.5×10^{-4} 1.6×10^{-3} 4.2×10^{-4} 4.2×10^{-4} 3.1×10^{-3} 6.0×10^{-4} 7.0×10^{-4} 1.3×10^{-3} 2.3×10^{-3} 8.0×10^{-4} 1.2×10^{-3} 3.9×10^{-3} 4.5×10^{-3}	Calibrator/ SICT-CP-40419
Frequency		10 Hz ~ 10 MHz	6.4×10^{-7}	
(Digital Smapling)				
AC Voltage		(1 mV) 0.1 Hz ~ 3 kHz (1 mV ~ 2 mV) 0.1 Hz ~ 3 kHz (2 mV ~ 3 mV) 0.1 Hz ~ 3 kHz (3 mV ~ 5 mV) 0.1 Hz ~ 3 kHz (5 mV ~ 10 mV) 0.1 Hz ~ 3 kHz (10 mV ~ 20 mV) 0.1 Hz ~ 3 kHz (20 mV ~ 30 mV) 0.1 Hz ~ 3 kHz (30 mV ~ 50 mV) 0.1 Hz ~ 3 kHz (50 mV ~ 100 mV) 0.1 Hz ~ 3 kHz	8.4×10^{-4} 4.2×10^{-4} 2.8×10^{-4} 1.7×10^{-4} 8.8×10^{-5} 4.8×10^{-5} 3.6×10^{-5} 3.0×10^{-5} 2.6×10^{-5}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters (Digital Smapling)	40419			Calibrator/ SICT-CP-40419
AC Voltage		(100 mV ~ 200 mV) 0.1 Hz ~ 3 kHz	4.8×10^{-5}	
		(200 mV ~ 300 mV) 0.1 Hz ~ 3 kHz	3.6×10^{-5}	
		(300 mV ~ 500 mV) 0.1 Hz ~ 3 kHz	2.8×10^{-5}	
		(500 mV ~ 1 V) 0.1 Hz ~ 3 kHz	2.4×10^{-5}	
		(1 V ~ 2 V) 0.1 Hz ~ 3 kHz	4.8×10^{-5}	
		(2 V ~ 3 V) 0.1 Hz ~ 3 kHz	3.6×10^{-5}	
		(3 V ~ 5 V) 0.1 Hz ~ 3 kHz	3.0×10^{-5}	
		(5 V ~ 10 V) 0.1 Hz ~ 3 kHz	2.6×10^{-5}	
		(10 V ~ 30 V) 10 Hz ~ 3 kHz	3.6×10^{-5}	
		(30 V ~ 50 V) 10 Hz ~ 3 kHz	2.8×10^{-5}	
		(50 V ~ 100 V) 10 Hz ~ 3 kHz	2.4×10^{-5}	
		(100 V ~ 200 V) 10 Hz ~ 3 kHz	4.8×10^{-5}	
		(200 V ~ 1 000 V) 50 Hz ~ 1 kHz	2.4×10^{-5}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Noise meters				
AC Voltage Test	40420	(600 μ V) 1 kHz	7.8×10^{-3}	Calibrator/ SICT-CP-40420
		(600 μ V ~ 20 mV) (10 ~ 40) Hz 40 Hz ~ 20 kHz (20 ~ 100) kHz (100 ~ 300) kHz 300 kHz ~ 1 MHz	2.6×10^{-3} 2.5×10^{-3} 3.5×10^{-3} 7.0×10^{-3} 1.5×10^{-2}	
		(20 ~ 200) mV (10 ~ 40) Hz 40 Hz ~ 20 kHz (20 ~ 100) kHz 100 kHz ~ 1 MHz	5.2×10^{-4} 3.4×10^{-4} 8.8×10^{-4} 4.3×10^{-3}	
		(200 mV ~ 2 V) (10 ~ 40) Hz 40 Hz ~ 20 kHz (20 ~ 100) kHz (100 ~ 300) kHz 300 kHz ~ 1 MHz	3.5×10^{-4} 1.5×10^{-4} 4.6×10^{-4} 8.8×10^{-4} 3.4×10^{-3}	
		(2 ~ 20) V (10 ~ 40) Hz 40 Hz ~ 100 kHz (100 ~ 300) kHz 300 kHz ~ 1 MHz	3.0×10^{-4} 1.2×10^{-4} 4.4×10^{-4} 2.2×10^{-3}	
		(20 ~ 200) V (10 ~ 40) Hz 40 Hz ~ 20 kHz (20 ~ 100) kHz	3.0×10^{-4} 1.2×10^{-4} 1.8×10^{-4}	
		(200 ~ 500) V 50 Hz ~ 1 kHz	3.8×10^{-4}	
		(500 ~ 1 000) V 50 Hz ~ 1 kHz	3.7×10^{-4}	
		(25 ~ 500) mV (1 ~ 30) MHz	2.1×10^{-2}	
		(500 mV ~ 2 V) (0.1 ~ 30) MHz	2.1×10^{-2}	
Weighting Test		(DIN/NOISE) 31.5 Hz ~ 10 kHz (JIS A) 31.5 Hz ~ 16 kHz (CCIR) 31.5 Hz ~ 31.5 kHz (CCIR/ARM) 31.5 Hz ~ 31.5 kHz	0.12 dB 0.12 dB 0.12 dB 0.12 dB	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Noise meters				
AC Voltage Output	40420	(10 mV) 1 kHz (10 mV ~ 1 V) 20 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz	2.8×10^{-3} 8.5×10^{-5} 1.3×10^{-4} 6.9×10^{-4}	Calibrator/ SICT-CP-40420
DC Voltage Output		0 mV 100 mV ~ 1 V	0.99 μ V 1.1×10^{-5}	
Oscilloscopes				
Impedance Measure	40421	50 Ω 75 Ω 1 M Ω	3.5×10^{-5} 2.7×10^{-5} 2.5×10^{-5}	Calibration Generator/ SICT-CP-40421
DC Voltage		(±) 0 mV (0 ~ 1) mV (1 ~ 5) mV (5 ~ 10) mV (10 ~ 100) mV (100 ~ 900) mV (0.9 ~ 9) V (9 ~ 200) V	0.79 μ V 8.0×10^{-4} 4.1×10^{-4} 1.7×10^{-4} 8.5×10^{-5} 1.5×10^{-5} 9.1×10^{-6} 9.5×10^{-6}	
AC Voltage(Square wave)		(1 kHz) 1 mV (1 ~ 25) mV (0.025 ~ 0.5) V (0.5 ~ 2.2) V (2.2 ~ 130) V	6.5×10^{-3} 8.8×10^{-4} 9.1×10^{-4} 6.8×10^{-4} 8.4×10^{-4}	
Time Marker		100 ps (100 ~ 200) ps 200 ps ~ 20 ms 20 ms ~ 5 s	6.2×10^{-7} 3.1×10^{-7} 1.7×10^{-6} 1.6×10^{-5}	
CAL Output Amplitude		(40 Hz ~ 20 kHz) 100 mV 100 mV ~ 12 V	3.2×10^{-5} 1.9×10^{-5}	
CAL Output Frequency		100 Hz ~ 10 MHz	6.2×10^{-7}	
Sinewave Signal Generator Level		50 kHz 50 kHz ~ 1 MHz 1 MHz ~ 1 GHz (1 ~ 4) GHz (4 ~ 18) GHz (18 ~ 25) GHz (25 ~ 33) GHz (33 ~ 40) GHz	2.3×10^{-2} 4.7×10^{-2} 1.5×10^{-2} 1.8×10^{-2} 3.2×10^{-2} 5.5×10^{-2} 5.8×10^{-2} 6.0×10^{-2}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Oscilloscopes				
AC Voltage(Sine wave)	40421	(10 ~ 40) Hz 2 mV (2 ~ 20) mV (20 ~ 200) mV (0.2 ~ 20) V (20 ~ 100) V (40 Hz ~ 1 kHz) 2 mV (2 ~ 20) mV (20 ~ 800) mV (0.8 ~ 20) V (20 ~ 200) V (1 ~ 50) kHz 2 mV (2 ~ 20) mV (20 ~ 200) mV (0.2 ~ 20) V (20 ~ 100) V (50 ~ 100) kHz 2 mV (2 ~ 20) mV (20 ~ 200) mV (0.2 ~ 20) V (20 ~ 100) V	5.3 μ V 1.9×10^{-3} 6.8×10^{-4} 4.3×10^{-4} 4.5×10^{-4} 5.0 μ V 1.7×10^{-3} 3.3×10^{-4} 1.7×10^{-4} 1.8×10^{-4} 5.2 μ V 1.9×10^{-3} 4.8×10^{-4} 1.3×10^{-4} 1.7×10^{-4} 7.1 μ V 2.7×10^{-3} 9.0×10^{-4} 1.9×10^{-4} 3.0×10^{-4}	Calibration Generator/ SICT-CP-40421
LF phase meters				
Phase Test	40422	(1 Hz ~ 200 kHz) (-180 ~ 180) °	0.074 °	Multi Function Generator/ SICT-CP-40422
Volt/Current recorders				
DC Voltage	40424	(±) (0 ~ 100) μ V (0.1 ~ 1) mV (1 ~ 10) mV (0.01 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V	0.51 μ V 5.2×10^{-4} 5.9×10^{-5} 6.7×10^{-6} 4.3×10^{-6} 6.3×10^{-6} 8.7×10^{-6}	Calibrator/ SICT-CP-40424
DC Current		(±) (0 ~ 1) nA (1 ~ 100) nA (0.1 ~ 1) μ A (1 ~ 10) μ A (10 ~ 100) μ A (0.1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 100) A	6.9 pA 4.6×10^{-3} 2.3×10^{-3} 7.2×10^{-4} 1.4×10^{-4} 7.6×10^{-5} 8.4×10^{-5} 1.2×10^{-4} 2.1×10^{-4}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Relay test sets	40425			Digital Multimeter/ SICT-CP-40425
DC Voltage		1 mV 1 mV ~ 1 V (1 ~ 100) V (100 ~ 1 000) V	6 µV 7.0×10^{-4} 7.0×10^{-5} 2.2×10^{-5}	
DC Current		1 mA 1 mA ~ 1 A (1 ~ 20) A (20 ~ 100) A	58 µA 6.0×10^{-4} 2.5×10^{-4} 4.0×10^{-4}	
AC Voltage		(1 mV) 20 Hz ~ 100 kHz (1 ~ 100) mV 20 Hz ~ 10 kHz (10 ~ 100) kHz (100 mV ~ 1 V) 20 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (1 ~ 10) V 20 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (10 ~ 100) V 20 Hz ~ 10 kHz (10 ~ 100) kHz (100 ~ 1 000) V 50 Hz ~ 10 kHz (10 ~ 30) kHz	58 µV 6.1×10^{-4} 1.2×10^{-3} 7.6×10^{-4} 1.3×10^{-3} 2.0×10^{-2} 2.2×10^{-4} 1.1×10^{-3} 2.1×10^{-2} 1.8×10^{-4} 1.1×10^{-3} 1.9×10^{-4} 4.5×10^{-4}	
AC Current		(1 mA) 40 Hz ~ 10 kHz (1 ~ 100) mA 40 Hz ~ 10 kHz (100 mA ~ 1 A) 40 Hz ~ 10 kHz (1 ~ 10) A 40 Hz ~ 10 kHz (10 ~ 100) A 40 Hz ~ 10 kHz	58 µA 8.6×10^{-4} 9.1×10^{-4} 9.9×10^{-4} 2.3×10^{-4}	
Timer		(1 ~ 100) s	5.8×10^{-6}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF signal generators	40426			
Frequency Test		(0.1 ~ 1) Hz 1 Hz ~ 100 MHz	5.8×10^{-7} 5.8×10^{-9}	Audio Analyzer, Digital Multimeter/ SICT-CP-40426
Output Level Test		(10 ~ 100) Hz 1 mV (1 ~ 10) mV 10 mV ~ 100 V	1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5}	
		(100 Hz ~ 10 kHz) 1 mV (1 ~ 10) mV 10 mV ~ 100 V	1.0×10^{-3} 1.0×10^{-4} 3.0×10^{-5}	
		(10 ~ 100) kHz 1 mV (1 ~ 10) mV 10 mV ~ 100 V	1.0×10^{-3} 1.0×10^{-4} 8.0×10^{-5}	
DC Offset		(±) 0 mV (0 ~ 1) mV (1 ~ 10) mV 10 mV ~ 20 V	0.7 μ V 0.7 μ V 1.0×10^{-4} 6.0×10^{-5}	
Output Level Flatness Test		(100 mV) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	0.099 dB 0.083 dB 0.095 dB	
		(100 mV ~ 1 V) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	0.005 4 dB 0.001 1 dB 0.007 2 dB	
		(1 ~ 30) V (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	0.021 dB 0.015 dB 0.027 dB	
Attenuator Test		(10 Hz ~ 100 kHz) (0 ~ 80) dB	0.006 1 dB	
Distortion		(20 Hz ~ 1 kHz) (3.16 ~ 0.010) %	1.5×10^{-1}	
		(1 ~ 100) kHz (3.16 ~ 0.010) %	3.2×10^{-1}	
Rise/Fall Time		100 μ s ~ 100 ns (100 ~ 10) ns (10 ~ 1) ns 1 ns ~ 100 ps	7.0×10^{-4} 7.8×10^{-4} 4.7×10^{-3} 4.6×10^{-2}	
Duty cycle		(1 ~ 99) %	0.006 1 %	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF spectrum analyzers	40427			Synthesizer Function Generator/ SICT-CP-40427
Reference Frequency		10 MHz	7.7×10^{-12}	
Center Frequency		10 Hz (10 ~ 100) Hz 100 Hz ~ 1 kHz 1 kHz ~ 1 MHz (1 ~ 100) MHz 100 MHz ~ 1 GHz	6.1×10^{-5} 6.1×10^{-6} 6.1×10^{-7} 6.1×10^{-8} 6.1×10^{-9} 6.1×10^{-8}	
Frequency Range		10 Hz (10 ~ 100) Hz 100 Hz ~ 1 kHz (1 ~ 100) kHz 100 kHz ~ 1 MHz (1 ~ 100) MHz 100 MHz ~ 1 GHz	1.1×10^{-3} 1.1×10^{-4} 1.1×10^{-5} 1.1×10^{-4} 1.1×10^{-6} 1.1×10^{-7} 1.1×10^{-9}	
Resolution Bandwidth		100 Hz 100 Hz ~ 3 kHz (3 ~ 300) kHz 300 kHz ~ 1 MHz	3.3×10^{-2} 3.1×10^{-2} 3.3×10^{-2} 3.4×10^{-2}	
Absolute Amplitude		(-60 dBV) 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz (-60 ~ -50) dBV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz (-50 ~ -40) dBV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz (-40 ~ -30) dBV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz (-30 ~ 30) dBV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz Reference Level	0.043 dB 0.042 dB 0.056 dB 0.017 dB 0.015 dB 0.022 dB 0.009 dB 0.009 8 dB 0.012 dB 0.009 dB 0.006 8 dB 0.011 dB 0.016 dB 0.006 3 dB 0.007 3 dB 0.17 dB 0.16 dB	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF spectrum analyzers	40427			Synthesizer Function Generator/ SICT-CP-40427
Frequency Response		10 Hz 10 Hz ~ 100 kHz	0.009 1 dB 0.008 7 dB	
Logscale Fidelity		(0 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB (-80 ~ -90) dB	0.009 2 dB 0.012 dB 0.016 dB 0.042 dB	
Output frequency		10 Hz ~ 300 MHz	6.1×10^{-11}	
Input Impedance		1 MΩ (50 ~ 75) Ω	0.000 12 MΩ 0.000 7 Ω	
Output Voltage		10 mV 10 mV ~ 5 V	0.000 38 mV 8.8×10^{-5}	
Output Offset Voltage		(-30 ~ 30) V	6.7×10^{-6}	
Output Voltage Flatness		10 Hz ~ 100 kHz	0.000 67 dB	
Spot generators	40428			Audio Analyzer, Digital Multimeter/ SICT-CP-40428
Frequency		(0.1 ~ 1) Hz 1 Hz ~ 100 kHz	5.8×10^{-7} 5.8×10^{-9}	
Output Level		(10 ~ 100) Hz 1 mV (1 ~ 10) mV 10 mV ~ 10 V	1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5}	
		(100 Hz ~ 10 kHz) 1 mV (1 ~ 10) mV 10 mV ~ 10 V	1.0×10^{-3} 1.0×10^{-4} 3.0×10^{-5}	
		(10 ~ 100) kHz 1 mV (1 ~ 10) mV 10 mV ~ 10 V	1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5}	
Output Level Flatness		(100 mV) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	0.099 dB 0.083 dB 0.095 dB	
		(100 mV ~ 1 V) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	0.005 4 dB 0.001 1 dB 0.007 2 dB	
		(1 ~ 10) V (10 ~ 100) Hz 100 Hz ~ 100 kHz	0.010 dB 0.011 dB	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Spot generators				
Attenuation	40428	(10 Hz ~ 100 kHz) (0 ~ 80) dB	0.006 1 dB	Audio Analyzer, Digital Multimeter/ SICT-CP-40428
Distortion		(20 Hz ~ 1 kHz) (3.16 ~ 0.010) %	1.5×10^{-1}	
		(1 ~ 100) kHz (3.16 ~ 0.010) %	3.2×10^{-1}	
Sweep generators				
Frequency	40429	(0.1 ~ 1) Hz 1 Hz ~ 100 kHz	5.8×10^{-7} 5.8×10^{-9}	Audio Analyzer, Digital Multimeter/ SICT-CP-40429
Output Level Test		(10 ~ 100) Hz 1 mV (1 ~ 10) mV 100 mV ~ 10 V	1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5}	
		(100 Hz ~ 10 kHz) 1 mV (1 ~ 100) mV 100 mV ~ 10 V	1.0×10^{-3} 1.0×10^{-4} 3.0×10^{-5}	
		(10 ~ 100) kHz 1 mV (1 ~ 100) mV 100 mV ~ 10 V	1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5}	
Output Level Flatness		(100 mV) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	0.099 dB 0.083 dB 0.095 dB	
		(100 mV ~ 1 V) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	0.005 4 dB 0.001 1 dB 0.007 2 dB	
		(1 ~ 10) V (10 ~ 100) Hz 100 Hz ~ 100 kHz	0.010 dB 0.011 dB	
Attenuation		(10 Hz ~ 10 kHz) (0 ~ 80) dB	0.006 1 dB	
Distortion		(20 Hz ~ 1 kHz) (3.16 ~ 0.010) %	1.5×10^{-1}	
		(1 ~ 100) kHz (3.16 ~ 0.010) %	3.2×10^{-1}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Signal transducers	40430			Digital Multimeter/ SICT-CP-40430
Voltage		1 mV (1 ~ 10) mV 10 mV ~ 100 V (100 ~ 300) V	5.2×10^{-4} 8.8×10^{-5} 3.1×10^{-5} 4.3×10^{-3}	
Current		10 μ A (10 ~ 100) μ A 100 μ A ~ 100 mA 100 mA ~ 20 A	9.5×10^{-4} 9.7×10^{-5} 7.0×10^{-5} 2.0×10^{-4}	
Frequency		(1 ~ 10) Hz 10 Hz ~ 100 kHz	3.1×10^{-4} 7.2×10^{-5}	
Transistor curve tracers	40432			Digital Multimeter/ SICT-CP-40432
DC Voltage(Source)		0 mV (0 ~ 100) mV 100 mV ~ 1 V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V	0.0 μ V 1.1×10^{-5} 1.0×10^{-5} 9.8×10^{-6} 1.1×10^{-5} 8.0×10^{-6}	
DC Current(Source)		0 nA (0 ~ 1) nA (1 ~ 100) nA (0.1 ~ 1) μ A (1 ~ 10) μ A 10 μ A ~ 10 mA (10 ~ 100) mA 100 mA ~ 10 A	0.12 nA 1.0×10^{-2} 8.0×10^{-3} 8.0×10^{-4} 9.0×10^{-5} 1.6×10^{-5} 5.0×10^{-5} 2.5×10^{-4}	
DC Voltage(Measure)		0 mV (0 ~ 100) mV 100 mV ~ 1 V (1 ~ 100) V (100 ~ 1 000) V	0.78 μ V 1.5×10^{-5} 6.0×10^{-4} 6.4×10^{-4} 6.2×10^{-5}	
AC/DC high voltage generators	40434			High Voltage Digital Meter/ SICT-CP-40434
DC Voltage		(\pm) 0 kV (0 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 2) kV (2 ~ 100) kV (100 ~ 200) kV	0.58 V 1.2×10^{-3} 6.1×10^{-4} 3.0×10^{-4} 2.3×10^{-4} 1.2×10^{-2}	
AC Voltage		(50 Hz) 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 100) kV (100 ~ 200) kV	0.58 V 1.2×10^{-3} 6.1×10^{-4} 5.7×10^{-4} 1.2×10^{-2}	
		(60 Hz) 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 100) kV (100 ~ 200) kV	0.58 V 1.2×10^{-3} 6.2×10^{-4} 5.2×10^{-4} 1.2×10^{-2}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC/DC high voltage probes	40435			DC Power Supply/ SICT-CP-40435
DC Voltage		(±) 0 kV (0 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 100) kV	0.06 V 4.2×10^{-4} 2.1×10^{-4} 3.5×10^{-4}	
AC Voltage		(50 Hz) 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 3) kV (3 ~ 5) kV (5 ~ 10) kV (10 ~ 20) kV (20 ~ 100) kV	2.2 V 4.4×10^{-3} 2.3×10^{-3} 1.2×10^{-3} 7.5×10^{-4} 6.0×10^{-4} 4.0×10^{-4} 3.7×10^{-4}	
		(60 Hz) 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 3) kV (3 ~ 5) kV (5 ~ 10) kV (10 ~ 20) kV (20 ~ 100) kV	2.2 V 4.4×10^{-3} 2.3×10^{-3} 1.2×10^{-3} 7.5×10^{-4} 6.0×10^{-4} 4.0×10^{-4} 3.7×10^{-4}	
Logic analyzers	40436			Calibrator/ SICT-CP-40436
DC Voltage		(0 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V	1.5 μ V 8.8×10^{-6} 7.2×10^{-6}	
Clock frequency		10 MHz	7.7×10^{-12}	
Telephone testers	40437			Tone Pulse Simulator/ SICT-CP-40437
L1, L2 Output Voltage		(1 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 1 000) V	9.9×10^{-4} 1.1×10^{-5} 7.1×10^{-6} 8.5×10^{-6}	
Loop Current		(0.1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A	2.5×10^{-5} 5.2×10^{-5} 2.2×10^{-4} 4.7×10^{-4}	
Ring Output Voltage		(10 Hz ~ 20 kHz) 100 mV ~ 1 V (1 ~ 100) V (100 ~ 1 000) V	4.7×10^{-4} 2.9×10^{-4} 3.1×10^{-4}	
Ring Frequency		(1 ~ 1 000) Hz	7.0×10^{-5}	
D.T.M.F & Pulse		(+10 ~ -39.9) dBm	0.09 dB	
D.T.M.F & Frequency		(697 ~ 1 477) Hz	0.59 Hz	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Video signal analyzers				
Color Bar Decoding Accuracy(Gain)	40438	(0 ~ 5) mV (5 ~ 10) mV (10 ~ 100) mV (100 ~ 200) mV (200 ~ 500) mV (500 ~ 1 000) mV	2.5×10^{-1} 5.0×10^{-2} 8.2×10^{-2} 4.1×10^{-2} 2.1×10^{-2} 8.5×10^{-3}	Video Amplitude Calibration Fixture/ SICT-CP-40438
Frequency		20 Hz ~ 5 MHz	5.8×10^{-6}	
Color Bar Decoding Accuracy(Phase)		(0 ~ 360)°	0.70°	
Measure Square Wave		(0 ~ 5) mV (5 ~ 10) mV (10 ~ 100) mV (100 ~ 300) mV (300 ~ 400) mV (400 ~ 600) mV (600 ~ 999.9) mV	9.4×10^{-2} 2.0×10^{-2} 9.9×10^{-3} 2.1×10^{-3} 1.5×10^{-3} 1.2×10^{-3} 9.4×10^{-4}	
Measure Sine Wave		No Filter, PAL NTS BW Lim, NTSC,PAL Chroma BP, NTSC,PAL (10 kHz ~ 10 MHz) 500 mV		
Burst Frequency		(3 ~ 5) MHz	7.0×10^{-3}	
vertical Gain		(0 ~ 5) mV (5 ~ 10) mV (10 ~ 100) mV (100 ~ 300) mV (300 ~ 600) mV (600 ~ 999.9) mV	4.0×10^{-7} 9.4×10^{-2} 2.0×10^{-2} 9.9×10^{-3} 2.1×10^{-3} 1.5×10^{-3} 9.4×10^{-4}	
Horizontal Frequency		(20 ~ 100) Hz 100 Hz ~ 10 kHz 10 kHz ~ 10 MHz	3.1×10^{-3} 6.1×10^{-4} 6.1×10^{-5}	
Gain Frequency Response		Flat, Luminance, Chroma at (20 Hz ~ 20 MHz) 700 mV		
Transient Response (Video Noise)		(0 ~ 1 000) mV	7.0×10^{-3}	
Luminance Volt Level		(0 ~ -30) dB	1.3×10^{-2}	
Chrominance AM/PM Level		(0 ~ -30) dB	4.8×10^{-1}	
Luminance Volt Level		(0 ~ 1 000) mV	6.7×10^{-1}	
Luminance Inputt Level		(0 ~ 1 000) mV	1.7×10^{-5}	
Chrominance Input Level		(0 ~ 1 000) mV	1.8×10^{-5}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Flux meters	40503			Flux sources/ SICT-CP-40503
Flux sources	40504	0.1 mWb ~ 10 Wb (0.1 ~ 50) mWb (0.05 ~ 0.1) Wb (0.1 ~ 10) Wb	5.8 × 10 ⁻³ 6.6 × 10 ⁻⁴ 2.3 × 10 ⁻⁵ 1.4 × 10 ⁻⁵	GPS receiver, Frequency counter/ SICT-CP-40504
Magnetometers	40508	(0 ~ 0.1) mT (0.1 ~ 0.5) mT (0.5 ~ 3) mT (3 ~ 5) mT (5 ~ 20) mT (20 ~ 30) mT (30 ~ 1 700) mT	7.1 × 10 ⁻² 1.4 × 10 ⁻² 7.0 × 10 ⁻³ 4.0 × 10 ⁻³ 3.0 × 10 ⁻³ 6.7 × 10 ⁻³ 6.4 × 10 ⁻³	Helmholtz coil, Standard magnets/ SICT-CP-40508
Reference/standard magnets	40510	(1.5 ~ 30) mT (30 ~ 1 000) mT	7.3 × 10 ⁻³ 2.6 × 10 ⁻³	Gaussmeters/ SICT-CP-40510

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF amplifiers	Gain	40601	(0 ~ 80) dB 20 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 GHz (1 ~ 10) GHz (10 ~ 18) GHz (18 ~ 40) GHz	0.01 dB 0.02 dB 0.10 dB 0.11 dB 0.14 dB 0.19 dB
			(9 kHz ~ 26.5 GHz) (0 ~ -100) dBc	0.98 dB
			(0 ~ 1) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	4.2×10^{-3} 9.4×10^{-3} 1.5×10^{-2}
			(1 ~ ∞) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	9.0×10^{-3} 2.1×10^{-2} 3.3×10^{-2}
	Reflection coefficient			
Coaxial attenuators	Attenuation	40602	(DC ~ 26.5 GHz) (0 ~ 10) dB (10 ~ 20) dB (20 ~ 30) dB (30 ~ 40) dB (40 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB (80 ~ 100) dB (100 ~ 110) dB (110 ~ 120) dB	0.040 dB 0.042 dB 0.045 dB 0.050 dB 0.055 dB 0.074 dB 0.076 dB 0.089 dB 0.12 dB 0.14 dB 0.15 dB
			(26.5 ~ 34) GHz (0 ~ 40) dB (40 ~ 50) dB (50 ~ 60) dB	0.26 dB 0.32 dB 0.37 dB
			(34 ~ 40) GHz (0 ~ 40) dB (40 ~ 50) dB (50 ~ 60) dB	0.35 dB 0.45 dB 0.53 dB
SWR	Reflection coefficient		(0 ~ 1) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	4.2×10^{-3} 9.4×10^{-3} 1.5×10^{-2}
SWR	SWR		(1 ~ ∞) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	9.7×10^{-3} 2.4×10^{-2} 3.8×10^{-2}

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Burst pulse generators	40605			Digital Oscilloscope/ SICT-CP-40605
Burst Volatage		50 Ω (±) 5 V (5 ~ 100) V (0.1 ~ 8) kV	2.0 × 10 ⁻² 1.6 × 10 ⁻² 2.0 × 10 ⁻²	
		1 kΩ (±) 5 V ~ 8 kV	4.0 × 10 ⁻²	
Rise/Fall Time		1 ns (1 ~ 2) ns (2 ~ 4) ns 4 ns ~ 1 μs (1 ~ 2) μs (2 ~ 4) μs (4 ~ 10) μs	2.0 × 10 ⁻² 6.8 × 10 ⁻³ 2.6 × 10 ⁻³ 1.5 × 10 ⁻³ 6.2 × 10 ⁻³ 2.6 × 10 ⁻³ 1.3 × 10 ⁻³	
Pulse Width		1 ns (1 ~ 2) ns 2 ns ~ 200 ms	6.0 × 10 ⁻³ 3.1 × 10 ⁻³ 1.5 × 10 ⁻³	
Time measurement by section		1 ns (1 ~ 2) ns 2 ns ~ 200 ms	6.0 × 10 ⁻³ 3.1 × 10 ⁻³ 1.5 × 10 ⁻³	
Repeat Frequency		1 Hz ~ 25 MHz	1.6 × 10 ⁻³	
Attenuator calibrators	40606			Verification Kit/ SICT-CP-40606
Attenuation		(0 ~ 10) dB (10 ~ 20) dB (20 ~ 30) dB (30 ~ 40) dB (40 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB (80 ~ 90) dB (90 ~ 100) dB (100 ~ 110) dB (110 ~ 120) dB	0.024 dB 0.025 dB 0.027 dB 0.029 dB 0.031 dB 0.034 dB 0.036 dB 0.039 dB 0.042 dB 0.045 dB 0.048 dB 0.052 dB	
RF power meter calibrators	40607			Digital Multimeter/ SICT-CP-40607
Output Power		3 μW 10 μW 30 μW 100 μW 300 μW 1 mW 3 mW 10 mW 30 mW 100 mW	0.1 nW 0.2 nW 0.5 nW 1 nW 4 nW 0.18 μW 0.19 μW 0.2 μW 0.3 μW 1 μW	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
EMC transducers ; current probes, absorbing clamps, etc	40608			Power Senso, Network analyzer/ SICT-CP-40608
Transfer Impedance		10 Hz ~ 50 MHz (50 ~ 200) MHz 200 MHz ~ 3 GHz	0.60 dB 1.1 dB 1.9 dB	
Insertion Loss		30 MHz ~ 1 GHz	1.9 dB	
Electric Magnetic Near-Field		100 kHz ~ 1 GHz	1.9 dB	
Reflection coefficient		(0 ~ 1) 10 Hz ~ 1 GHz (1 ~ 3) GHz	4.2×10^{-3} 6.0×10^{-3}	
SWR		(1 ~ ∞) 10 Hz ~ 1 GHz (1 ~ 3) GHz	9.0×10^{-3} 1.3×10^{-2}	
Coaxial directional couplers/ splitters	40610			Power Sensor, Synthesized Sweeper/ SICT-CP-40610
Coupling Factor		(0 ~ 30) dB 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 3 GHz (3 ~ 8) GHz (8 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz (30 ~ 80) dB 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 4 GHz (4 ~ 8) GHz (8 ~ 10) GHz (10 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz	0.005 9 dB 0.007 3 dB 0.087 dB 0.095 dB 0.11 dB 0.14 dB 0.20 dB 0.006 5 dB 0.021 dB 0.088 dB 0.092 dB 0.10 dB 0.11 dB 0.13 dB 0.20 dB	
Reflection coefficient		(0 ~ 1) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	4.2×10^{-3} 9.4×10^{-3} 1.5×10^{-2}	
SWR		(1 ~ ∞) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	9.7×10^{-3} 2.4×10^{-2} 3.8×10^{-2}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrostatic discharge generators				Digital Oscilloscope/ SICT-CP-40613
Peak Current(I_p)	40613	(±) (3.75 ~ 7.5) A (7.5 ~ 15) A (15 ~ 22.5) A (22.5 ~ 56.3) A (56.3 ~ 93.8) A (93.8 ~ 112.5) A (112.5 ~ 150) A	5.1×10^{-2} 5.3×10^{-2} 4.6×10^{-2} 5.2×10^{-2} 4.9×10^{-2} 5.7×10^{-2} 5.2×10^{-2}	
Current I_1 (30 ~ 60) ns		(±) 2 A (2 ~ 4) A (4 ~ 8) A (8 ~ 16) A (16 ~ 36) A (36 ~ 50) A (50 ~ 60) A (60 ~ 80) A	4.5×10^{-2} 5.0×10^{-2} 5.3×10^{-2} 4.9×10^{-2} 5.0×10^{-2} 4.4×10^{-2} 5.7×10^{-2} 5.2×10^{-2}	
Current I_2 (60 ~ 130) ns		(±) 1 A (1 ~ 2) A (2 ~ 4) A (4 ~ 6) A (6 ~ 8) A (8 ~ 15) A (15 ~ 25) A (25 ~ 30) A (30 ~ 40) A	5.0×10^{-2} 5.4×10^{-2} 5.7×10^{-2} 4.9×10^{-2} 5.4×10^{-2} 6.5×10^{-2} 5.2×10^{-2} 6.7×10^{-2} 6.1×10^{-2}	
Current I_3 (360 ~ 800) ns		(±) 0.275 A (0.275 ~ 0.55) A (0.55 ~ 1.1) A (1.1 ~ 1.65) A (1.65 ~ 4.13) A (4.13 ~ 6.88) A (6.88 ~ 8.25) A (8.25 ~ 11) A	1.5×10^{-1} 2.2×10^{-1} 1.9×10^{-1} 1.6×10^{-1} 2.3×10^{-1} 1.5×10^{-1} 2.4×10^{-1} 1.9×10^{-1}	
Current I_4 (30 ~ 65) ns		(±) 0.15 A (0.15 ~ 0.3) A (0.3 ~ 0.6) A (0.6 ~ 1.2) A (1.2 ~ 2.25) A (2.25 ~ 2.7) A (2.7 ~ 3.75) A (3.75 ~ 4.5) A (4.5 ~ 6) A	1.3×10^{-1} 1.9×10^{-1} 3.0×10^{-1} 2.1×10^{-1} 2.5×10^{-1} 2.0×10^{-1} 1.5×10^{-1} 2.9×10^{-1} 2.0×10^{-1}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrostatic discharge generators				
Semiconductor Peak Current HBM	40613	(±) (0.083 ~ 0.17) A (0.17 ~ 0.33) A (0.33 ~ 0.67) A (0.67 ~ 1.33) A (1.33 ~ 6.66) A	4.8×10^{-2} 3.6×10^{-2} 3.4×10^{-2} 3.2×10^{-2} 2.7×10^{-2}	Digital Oscilloscope/ SICT-CP-40613
Semiconductor Peak Current MM		(±) (0.88 ~ 1.75) A (1.75 ~ 14) A (14 ~ 17.5) A (17.5 ~ 26.25) A (26.25 ~ 35) A	6.6×10^{-2} 3.5×10^{-2} 3.1×10^{-2} 2.7×10^{-2} 3.0×10^{-2}	
Time		0.1 ns 0.1 ns ~ 1 ms	2.7×10^{-2} 2.4×10^{-2}	
Peak Voltage		(±) 1 kV (1 ~ 35) KV	3.0×10^{-2} 2.5×10^{-2}	
EMC receivers				
Frequency accuracy	40614	100 kHz ~ 1 GHz	6.1×10^{-10}	Network Analyzer, Pulse Generator/ SICT-CP-40614
SWR		9 kHz ~ 50 MHz 50 MHz ~ 8 GHz (8 ~ 19) GHz (19 ~ 26) GHz (26 ~ 40) GHz	0.011 0.028 0.035 0.045 0.064	
IF Band Accuracy		1 Hz ~ 10 MHz	65 mHz	
IF Band Selectivity		1 Hz ~ 10 MHz	6.7×10^{-4}	
IF Band Linearity		1 Hz ~ 1 GHz	0.12 dB	
Frequency Response		10 Hz ~ 5 kHz 5 kHz ~ 1 GHz (1 ~ 10) GHz (10 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz (40 ~ 50) GHz	0.01 dB 0.07 dB 0.08 dB 0.09 dB 0.11 dB 0.16 dB 0.20 dB	
Frequency Response (CISPR)		9 kHz ~ 1 GHz	0.80 dB	
Display linearity accuracy		(80 ~ 50) dB μ V (50 ~ 0) dB μ V	0.10 dB 0.14 dB	
Input Attenuation		(0 ~ 30) dB (30 ~ 70) dB (70 ~ 110) dB	0.15 dB 0.12 dB 0.10 dB	
Noise Indicator		DC ~ 26.5 GHz	0.16 dB	
Interference Immunity		9 kHz ~ 40 GHz	0.67 dB	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF filters	40615			
Reject Frequency		(9 ~ 90) kHz (90 ~ 900) kHz 900 kHz ~ 900 MHz 900 MHz ~ 18 GHz (18 ~ 50) GHz	0.024 kHz 0.24 kHz 0.025 MHz 0.068 MHz 0.12 MHz	Network Analyzer/ SICT-CP-40615
Insertion Loss		(9 kHz ~ 8 GHz) (0 ~ 10) dB (10 ~ 20) dB (20 ~ 40) dB (40 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB (80 ~ 90) dB (90 ~ 100) dB (8 ~ 18) GHz (0 ~ 10) dB (10 ~ 30) dB (30 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB (80 ~ 90) dB (90 ~ 100) dB (18 ~ 50) GHz (0 ~ 10) dB (10 ~ 20) dB (20 ~ 30) dB (20 ~ 40) dB (40 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB (80 ~ 90) dB (90 ~ 100) dB	0.13 dB 0.14 dB 0.15 dB 0.16 dB 0.18 dB 0.23 dB 0.66 dB 1.7 dB 4.3 dB 0.23 dB 0.24 dB 0.25 dB 0.26 dB 0.31 dB 0.73 dB 1.7 dB 4.3 dB 0.48 dB 0.51 dB 0.52 dB 0.53 dB 0.54 dB 0.59 dB 0.78 dB 1.6 dB 2.8 dB 6.0 dB	
RF impedance meters	40616			
RF Level		(100 kHz ~ 18 GHz) (35 ~ 20) dBm (20 Hz ~ 18 GHz) (20 ~ -70) dBm	0.11 dB	Performance Kit/ SICT-CP-40616
Frequency		9 kHz ~ 0.1 MHz 0.1 MHz ~ 18 GHz	6.8×10^{-10} 6.2×10^{-11}	
Load Measurement		DC 10 Hz ~ 100 MHz (100 ~ 500) MHz 500 MHz ~ 1.8 GHz (1.8 ~ 3.0) GHz (3.0 ~ 18) GHz	0.02 Ω 0.06 Ω 0.15 Ω 0.21 Ω 0.41 Ω 1.1 Ω	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF impulse generators Pulse Level	40617	9 kHz ~ 1 GHz	0.33 dB	Digital Oscilloscope/ SICT-CP-40617
Line impedance stabilization networks; LISN, CDN, ISN, etc.	40618			Impedance/Gain-Phase Analyzer, Calibration Kit/ SICT-CP-40618
Impedance		9 kHz ~ 1 000 MHz	2.0×10^{-2}	
Phase		9 kHz ~ 1 000 MHz	1.2 °	
Insertion Loss		(0 ~ 100) dB 9 kHz ~ 100 MHz (100 ~ 1 000) MHz	0.07 dB 0.08 dB	
Decoupling attenuation(Isolation)		(0 ~ 100) dB (9 ~ 30) kHz (0.03 ~ 20) MHz (20 ~ 1 000) MHz	0.21 dB 0.20 dB 0.21 dB	
Coupling/Decoupling network(Impedance)		9 kHz ~ 1 000 MHz	2.0×10^{-2}	
Coupling/Decoupling network (Insertion loss)		(0 ~ 100) dB 9 kHz ~ 1 000 MHz	0.10 dB	
Coaxial standard mismatches Reflection coefficient (Γ')	40619	(0 ~ 1) (9 ~ 100) kHz 100 kHz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz	0.004 3 0.006 0 0.009 5 0.016	Network Analyzer, Calibration Kit/ SICT-CP-40619
SWR		(1 ~ ∞) (9 ~ 100) kHz 100 kHz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz	0.008 6 0.012 0.019 0.032	
Mobile communication test sets RF Output Level	40621	(35 ~ 20) dBm 100 kHz ~ 18 GHz (20 ~ -20) dBm 9 kHz ~ 1 GHz (1 ~ 18) GHz (18 ~ 40) GHz (40 ~ 50) GHz (-20 ~ -60) dBm 9 kHz ~ 1 GHz (1 ~ 18) GHz (18 ~ 40) GHz (40 ~ 50) GHz (-60 ~ -70) dBm 9 kHz ~ 1 GHz (1 ~ 18) GHz (18 ~ 40) GHz (40 ~ 50) GHz (-70 ~ -120) dBm 9 kHz ~ 26.5 GHz	0.08 dB 0.06 dB 0.11 dB 0.14 dB 0.18 dB 0.06 dB 0.11 dB 0.16 dB 0.21 dB 0.08 dB 0.10 dB 0.18 dB 0.21 dB 0.15 dB	Measuring Receiver, RF Signal Generator/ SICT-CP-40621

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Mobile communication test sets	40621			Measuring Receiver, RF Signal Generator/ SICT-CP-40621
Amplitude Modulation		(0.1 ~ 100) %	1.2×10^{-2}	
Frequency Modulation		(0.1 ~ 400) kHz	1.2×10^{-2}	
Phase Modulation		(0.1 ~ 400) rad	1.2×10^{-2}	
Distortion Harmonics of Modulation Rate Signal		≤ 20 %	2.3×10^{-2}	
Harmonics		(0 ~ -90) dB	0.36 dB	
Frequency Output Accuracy		9 kHz ~ 40 GHz	7.9×10^{-11}	
AC Output Level		(10 Hz ~ 100 kHz) (1 ~ 100) mV 100 mV ~ 100 V	5.2×10^{-4} 5.8×10^{-4}	
DC Output Level		1 mV ~ 1 V (1 ~ 10) V (10 ~ 100) V	1.5×10^{-6} 1.3×10^{-6} 1.8×10^{-6}	
AC Input Level		(10 Hz ~ 100 kHz) (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V	7.6×10^{-3} 1.3×10^{-3} 6.7×10^{-4} 1.7×10^{-4} 2.0×10^{-4}	
DC Input Level		(1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 100) V	1.3×10^{-3} 3.1×10^{-4} 6.2×10^{-5} 6.1×10^{-5}	
RF Input Level		(9 kHz ~ 18 GHz) (10 ~ -70) dBm (18 ~ 40) GHz (10 ~ -70) dBm	0.10 dB 0.16 dB	
Modulation meters	40622			Measuring Receiver/ SICT-CP-40622
Amplitude Modulation		0 kHz (0 ~ 400) kHz	0.01 % 1.2×10^{-2}	
Frequency Modulation		0 % (0 ~ 100) %	1 Hz 1.2×10^{-2}	
Phase Modulation		0 rad (0 ~ 400) rad	1.2 mrad 1.2×10^{-2}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Network analyzers	40623	10 Hz ~ 40 GHz	6.8×10^{-10}	Power Sensor, Verification Kit/ SICT-CP-40623
Frequency		(20 ~ -30) dBm		
		10 Hz ~ 100 Hz	0.01 dB	
		100 Hz ~ 500 MHz	0.08 dB	
		500 MHz ~ 10 GHz	0.09 dB	
		(10 ~ 18) GHz	0.11 dB	
		(18 ~ 27) GHz	0.13 dB	
		(27 ~ 40) GHz	0.16 dB	
		(40 ~ 50) GHz	0.21 dB	
		(50 ~ 80) GHz	0.45 dB	
		(80 ~ 110) GHz	0.53 dB	
Source Power Level		(-30 ~ -60) dBm		
		9 kHz ~ 500 MHz	0.08 dB	
		500 MHz ~ 10 GHz	0.09 dB	
		(10 ~ 18) GHz	0.11 dB	
		(18 ~ 27) GHz	0.13 dB	
		(27 ~ 40) GHz	0.16 dB	
Dynamic Range		(100 kHz ~ 18 GHz)		
		(0 ~ 10) dB	0.086 dB	
		(10 ~ 20) dB	0.087 dB	
		(20 ~ 30) dB	0.088 dB	
		(30 ~ 40) dB	0.091 dB	
		(40 ~ 50) dB	0.099 dB	
		(50 ~ 60) dB	0.11 dB	
		(60 ~ 70) dB	0.12 dB	
		(70 ~ 80) dB	0.14 dB	
		(80 ~ 90) dB	0.15 dB	
		(90 ~ 100) dB	0.18 dB	
Attenuation		(20 dB)		
		300 kHz ~ 1.5 GHz	0.050 dB	
		(1.5 ~ 8) GHz	0.051 dB	
		(8 ~ 18) GHz	0.055 dB	
		(18 ~ 26.5) GHz	0.067 dB	
		(40 dB)		
		300 kHz ~ 1.5 GHz	0.054 dB	
		(1.5 ~ 8) GHz	0.055 dB	
		(8 ~ 18) GHz	0.059 dB	
		(18 ~ 26.5) GHz	0.082 dB	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Network analyzers	40623			Power Sensor, Verification Kit/ SICT-CP-40623
	Phase	(± 180 °) 300 kHz ~ 45 MHz 45 MHz ~ 2.0 GHz (2.0 ~ 3.0) GHz (3.0 ~ 4.5) GHz (4.5 ~ 6.0) GHz (6.0 ~ 7.5) GHz (7.5 ~ 8.0) GHz (8.0 ~ 9.0) GHz (9.0 ~ 10.5) GHz (10.5 ~ 12.0) GHz (12.0 ~ 13.5) GHz (13.5 ~ 15.0) GHz (15.0 ~ 16.5) GHz (16.5 ~ 18.0) GHz (18.0 ~ 21.0) GHz (21.0 ~ 22.5) GHz (22.5 ~ 24.0) GHz (24.0 ~ 25.5) GHz (25.5 ~ 26.5) GHz	0.04° 0.09° 0.10° 0.11° 0.15° 0.19° 0.21° 0.22° 0.29° 0.28° 0.26° 0.27° 0.26° 0.29° 0.31° 0.28° 0.39° 0.33° 0.44°	
	SWR	(30 kHz ~ 2 GHz) 1.05 1.20 1.50 2.00 (2 ~ 18) GHz 1.05 1.20 1.50 2.00	0.021 0.021 0.021 0.021 0.018 0.018 0.018 0.024	
Noise figure meters	40624			Noise Source/ SICT-CP-40624
	Tuning Accuracy	10 MHz ~ 26.5 GHz	4.3 × 10 ⁻⁵	
	Referency frequency	10 MHz	6.1 × 10 ⁻¹⁰	
	Input VSWR	9 kHz ~ 1 GHz (1 ~ 20) GHz (20 ~ 26.5) GHz	0.008 0.019 0.03	
	DC voltage	(0 ~ 28) V	0.000 18 V	
	Range	(0 ~ 30) dB	0.052 dB	
	Noise Figure	10 MHz ~ 8 GHz (8 ~ 18) GHz (18 ~ 26.5) GHz	0.12 dB 0.16 dB 0.37 dB	
Noise generators	40625			Spectrum Analyzer/ SICT-CP-40625
	Noise Power	(-80 ~ -130) dBm/Hz	0.10 dB	
	Scale Fidelity	(0 ~ 50) dB	0.27 dB	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Noise impulse simulators				
Peak Voltage	40626	(±) 0.1 kV (0.1 ~ 5) kV	4.0×10^{-2} 3.5×10^{-2}	Digital Oscilloscope/ SICT-CP-40626
Rise/Fall Time		1 ns (1 ~ 2) ns (2 ~ 4) ns	6.0×10^{-3} 3.1×10^{-3} 1.5×10^{-3}	
Pulse Width		10 ns (10 ~ 1 000) ns	2.0×10^{-3} 1.5×10^{-3}	
RF phase noise meters				
RF phase noise	40627	(Carrier Frequency) 100 MHz ~ 18 GHz (Offset Frequency) 10 Hz ~ 100 MHz	1.0 dB 1.0 dB	RF Signal analyzer/ SICT-CP-40627
Coaxial noise sources				
ENR	40628	(4.5 ~ 16) dB (0.01 ~ 1) GHz (1 ~ 7) GHz (7 ~ 8) GHz (8 ~ 14) GHz (14 ~ 18) GHz (12 ~ 17) dB (0.01 ~ 1) GHz (1 ~ 2) GHz (2 ~ 6) GHz (6 ~ 7) GHz (7 ~ 12) GHz (12 ~ 18) GHz (18 ~ 26.5) GHz	0.28 dB 0.27 dB 0.30 dB 0.31 dB 0.32 dB 0.31 dB 0.28 dB 0.30 dB 0.29 dB 0.40 dB 0.41 dB 0.47 dB	Coaxial noise sources, Noise figure analyzer/ SICT-CP-40628
SWR		(0 ~ 1) (0.01 ~ 3) GHz (3 ~ 20) GHz (20 ~ 26.5) GHz	0.006 8 0.010 0.015	
RF power meters				
High power	40635	(0.1 ~ 500) W 10 kHz ~ 250 MHz (0.1 ~ 150) W (80 ~ 1 000) MHz (0.1 ~ 10) W (1 000 ~ 4 200) MHz	2.6×10^{-2} 2.6×10^{-2} 2.7×10^{-2}	Range Calibrator/ SICT-CP-40635
Zero Carryover		10 μW ~ 1 mW (1 ~ 100) mW	3 nW 0.01 mW	
Power		3 μW ~ 100 mW	1.6×10^{-3}	
Calibration Factor		(88 ~ 100) %	0.5×10^{-3}	
Power Ref. Output		50 MHz, 1 mW	8 μW	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Diode power sensors	40636	(1 ~ 10) μW	0.4×10^{-2}	Thermistor Mount, Synthesized Sweeper/ SICT-CP-40636
		9 kHz ~ 100 kHz	0.9×10^{-2}	
		100 kHz ~ 1 GHz	1.1×10^{-2}	
		(1 ~ 10) GHz	1.9×10^{-2}	
		(10 ~ 18) GHz	2.8×10^{-2}	
		(18 ~ 26.5) GHz	4.0×10^{-2}	
		(26.5 ~ 40) GHz		
		(10 μW ~ 10 mW)		
		9 kHz ~ 100 kHz	0.4×10^{-2}	
		100 kHz ~ 1 GHz	0.9×10^{-2}	
Reflection coefficient		(1 ~ 10) GHz	1.1×10^{-2}	
		(10 ~ 18) GHz	1.9×10^{-2}	
		(18 ~ 26.5) GHz	2.6×10^{-2}	
		(26.5 ~ 40) GHz	3.5×10^{-2}	
		(0 ~ 1)		
		20 Hz ~ 1 GHz	4.2×10^{-3}	
		(1 ~ 20) GHz	9.4×10^{-3}	
		(20 ~ 40) GHz	1.5×10^{-2}	
SWR		(1 ~ ∞)		
		20 Hz ~ 1 GHz	9.7×10^{-3}	
		(1 ~ 20) GHz	2.4×10^{-2}	
		(20 ~ 40) GHz	3.8×10^{-2}	
Thermocouple power sensors	40637	(1 ~ 10) μW		Thermistor Mount, Synthesized Sweeper/ SICT-CP-40637
		9 kHz ~ 100 kHz	0.4×10^{-2}	
		100 kHz ~ 1 GHz	0.9×10^{-2}	
		(1 ~ 10) GHz	1.1×10^{-2}	
		(10 ~ 18) GHz	1.9×10^{-2}	
		(18 ~ 26.5) GHz	2.8×10^{-2}	
		(26.5 ~ 40) GHz	4.0×10^{-2}	
		(10 μW ~ 10 mW)		
		9 kHz ~ 100 kHz	0.4×10^{-2}	
		100 kHz ~ 1 GHz	0.9×10^{-2}	
Reflection coefficient		(1 ~ 10) GHz	1.1×10^{-2}	
		(10 ~ 18) GHz	1.9×10^{-2}	
		(18 ~ 26.5) GHz	2.6×10^{-2}	
		(26.5 ~ 40) GHz	3.5×10^{-2}	
		(0 ~ 1)		
		20 Hz ~ 1 GHz	4.2×10^{-3}	
		(1 ~ 20) GHz	9.4×10^{-3}	
		(20 ~ 40) GHz	1.5×10^{-2}	
SWR		(1 ~ ∞)		
		20 Hz ~ 1 GHz	9.7×10^{-3}	
		(1 ~ 20) GHz	2.4×10^{-2}	
		(20 ~ 40) GHz	3.8×10^{-2}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Pulse generators	40638			Digital Oscilloscope/ SICT-CP-40638
Period		100 ps ~ 1 s	5.8×10^{-9}	
Frequency		1 Hz ~ 1 GHz (1 ~ 3.35) GHz	5.8×10^{-9} 1.7×10^{-8}	
Width		100 ps ~ 1 s	1.2×10^{-3}	
Delay Time		100 ps ~ 1 s	1.2×10^{-3}	
Double Pulse		100 ps ~ 1 s	1.2×10^{-3}	
Duty Cycle		(1 ~ 99) %	0.006 2 %	
DC Level		$\pm(10 \text{ mV} \sim 100 \text{ V})$	5.8×10^{-4}	
Output Level		(100 Hz ~ 10 kHz) (10 ~ -20) dBm	0.018 dB	
Radar test sets	40639			Signal Analyzer, SART Generator, EPIRB Generator/ SICT-CP-40639
(Ship)				
RF Level		(20 ~ -20) dBm 20 Hz ~ 1 GHz (1 ~ 18) GHz	0.09 dB 0.13 dB	
		(-20 ~ -60) dBm 9 kHz ~ 1 GHz (1 ~ 18) GHz	0.06 dB 0.11 dB	
		(-60 ~ -120) dBm 10 MHz ~ 18 GHz	0.25 dB	
Amplitude Modulation		(0.1 ~ 100) %	1.2×10^{-2}	
Frequency Modulation		(0.1 ~ 400) kHz	1.2×10^{-2}	
Phase Modulation		(0.1 ~ 400) rad	1.2×10^{-2}	
Distortion of Modulation		(0 ~ 2) %	1.2×10^{-3}	
Harmonics		(9 kHz ~ 18 GHz) (0 ~ -110) dB	0.25 dB	
Frequency		9 kHz ~ 18 GHz	6.2×10^{-11}	
Pulse Period		1 ns ~ 10 ms	1.2×10^{-2}	
High power		(0.1 ~ 500) W 10 kHz ~ 250 MHz	2.6×10^{-2}	
		(0.1 ~ 150) W (80 ~ 1 000) MHz	2.6×10^{-2}	
		(0.1 ~ 10) W (1 000 ~ 4 200) MHz	2.7×10^{-2}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Radar test sets (flight) Frequency(VOR/ILS/DME)	40639	(74.6 ~ 1 150) MHz	8.2×10^{-8}	Signal Analyzer, SART Generator, EPIRB Generator/ SICT-CP-40639
Amplitude Modulation(VOR/ILS)		Localizer (108.1 ~ 111.95) MHz (0.1 ~ 20) %	0.62 %	
		Glideslope (330.95 ~ 334.70) MHz (20 ~ 40) %	0.84 %	
		Marker Beacon (74.6 ~ 75.4) MHz (40 ~ 95) %	1.4 %	
		VOR (108 ~ 117.95) MHz (0.1 ~ 30) %	0.62 %	
고주파 레벨(VOR/ILS)		Localizer (108.1 ~ 111.95) MHz (10 ~ -30) dBm (-30 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -110) dBm (-110 ~ -120) dBm	0.19 dB 0.23 dB 0.24 dB 0.34 dB 0.35 dB	
		Glideslope (330.95 ~ 334.70) MHz (10 ~ -30) dBm (-30 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -110) dBm (-110 ~ -120) dBm	0.19 dB 0.23 dB 0.24 dB 0.34 dB 0.35 dB	
DDM(VOR/ILS)		Localizer (108.1 ~ 111.95) MHz LEFT(-0.200 ~ -0.155) LEFT(-0.155 ~ -0.093) CENTER(0.000) RIGHT(0.093 ~ 0.155) RIGHT(0.155 ~ 0.200)	0.000 5 0.000 5 0.000 5 0.000 5 0.000 5	
		Glideslope (330.95 ~ 334.70) MHz DOWN(0.400 ~ 0.175) DOWN(0.175 ~ 0.091) CENTER(0.000) UP(-0.091 ~ -0.175) UP(-0.175 ~ -0.400)	0.000 5 0.000 5 0.000 5 0.000 5 0.000 5	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Radar test sets SDM(VOR/ILS)	40639	Localizer (108.1 ~ 111.95) MHz (0 ~ 40) % Glideslope (330.95 ~ 334.70) MHz (40 ~ 80) % BEARING ANGLE(VOR) VOR (108 ~ 117.95) MHz (0 ~ 360) °	0.87 % 1.2 % 0.04 °	Signal Analyzer, SART Generator, EPIRB Generator/ SICT-CP-40639
RF signal generators Frequency Modulation	40640	(0.1 ~ 400) kHz	1.2×10^{-2}	Measuring Receiver/ SICT-CP-40640
Amplitude Modulation		(0.1 ~ 100) %	1.2×10^{-2}	
Phase Modulation		(0.1 ~ 400) rad	1.2×10^{-2}	
Pulse Modulation		(100 kHz ~ 12 000 MHz) Period (1 μs ~ 1 s) ton (100 ns ~ 100 μs) PRR	1.2×10^{-3} 1.2×10^{-3} 3.1×10^{-3}	
Distortion		≤ 20 %	2.3×10^{-2}	
RF Level		(54 ~ 57) dBm 100 kHz ~ 500 MHz (51 ~ 54) dBm 100 kHz ~ 2 GHz (35 ~ 51) dBm 100 kHz ~ 18 GHz (20 ~ 35) dBm 100 kHz ~ 10 GHz (10 ~ 18) GHz (-30 ~ 20) dBm 20 Hz ~ 10 GHz (10 ~ 18) GHz (18 ~ 28) GHz (28 ~ 40) GHz (40 ~ 50) GHz (50 ~ 70) GHz (70 ~ 110) GHz (-30 ~ -60) dBm 20 Hz ~ 10 GHz (10 ~ 18) GHz (18 ~ 28) GHz (28 ~ 40) GHz (40 ~ 50) GHz	0.35 dB 0.32 dB 0.32 dB 0.26 dB 0.27 dB 0.09 dB 0.11 dB 0.12 dB 0.16 dB 0.20 dB 0.29 dB 0.38 dB 0.10 dB 0.11 dB 0.12 dB 0.16 dB 0.21 dB	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF signal generators				
RF Level	40640	(-120 ~ -60) dBm 20 Hz ~ 4.2 GHz (4.2 ~ 8) GHz (8 ~ 12.4) GHz (12.4 ~ 18) GHz (18 ~ 26.5) GHz	0.21 dB 0.23 dB 0.24 dB 0.27 dB 0.31 dB	Measuring Receiver/ SICT-CP-40640
Harmonic Frequency		(-10 ~ -110) dBc 9 kHz ~ 40 GHz	0.37 dB 2.1×10^{-11}	
RF spectrum analyzers				
Center Frequency	40641	(3 ~ 100) Hz (100 ~ 500) Hz (500 ~ 900) Hz 900 Hz ~ 100 kHz 0.1 MHz ~ 40 GHz	2.0×10^{-4} 6.1×10^{-6} 1.2×10^{-6} 6.8×10^{-7} 6.2×10^{-9}	Power Sensor, Synthesized Sweeper/ SICT-CP-40641
Frequency Counter		(3 ~ 100) Hz (100 ~ 500) Hz (500 ~ 900) Hz 900 Hz ~ 100 kHz 0.1 MHz ~ 40 GHz	2.0×10^{-4} 6.1×10^{-6} 1.2×10^{-6} 6.8×10^{-7} 6.2×10^{-9}	
Span		10 Hz ~ 100 kHz 0.1 MHz ~ 40 GHz	7.6×10^{-3} 7.7×10^{-6}	
RBW		1 Hz ~ 100 MHz	6.2×10^{-6}	
RBW Selectivity		1 Hz ~ 100 MHz	3.2×10^{-2}	
RBW Switching		1 Hz ~ 100 MHz	0.022 dB	
Scale Switching		1 dB ~ 10 dB scale/div	0.022 dB	
Scale Fidelity		(0 ~ -30) dB (-30 ~ -40) dB (-40 ~ -50) dB (-50 ~ -60) dB (-60 ~ -80) dB (-80 ~ -100) dB	0.073 dB 0.077 dB 0.082 dB 0.095 dB 0.13 dB 0.17 dB	
Frequency Response		(10 ~ 100) Hz 100 Hz ~ 1 GHz (1 ~ 6) GHz (6 ~ 10) GHz (10 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 30) GHz (30 ~ 35) GHz (35 ~ 40) GHz	0.01 dB 0.15 dB 0.16 dB 0.17 dB 0.19 dB 0.21 dB 0.23 dB 0.28 dB 0.32 dB	
Average Noise Level		DC ~ 40 GHz	0.17 dB	
Sideband Noise Level		(-30 ~ 30) kHz	0.33 dB	
CAL Output Freq. & Int. Frequency		DC ~ 1 GHz	6.2×10^{-9}	
CAL Output Level		(-20 ~ 20) dBm	0.09 dB	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Surge generators	40643	(±) 2 V (2 ~ 5) V 5 V ~ 200 kV	7.0×10^{-2} 4.0×10^{-2} 3.5×10^{-2}	Digital Oscilloscope/ SICT-CP-40643
		Surge Current	(±) 5 A ~ 200 kA	3.6×10^{-2}
		Rise/Fall Time	1 ns (1 ~ 2) ns 2 ns ~ 10 s	6.0×10^{-3} 3.0×10^{-3} 2.0×10^{-3}
		Pulse Width	1 ns (1 ~ 2) ns 2 ns ~ 10 s	6.0×10^{-3} 3.0×10^{-3} 2.0×10^{-3}
		Time measurement by section	1 ns (1 ~ 2) ns 2 ns ~ 10 s	6.0×10^{-3} 3.0×10^{-3} 2.0×10^{-3}
		Frequency measurement by section	1 Hz ~ 25 MHz	1.6×10^{-3}
		Phase Shifting	at 50 Hz (0 ~ 360)°	1.2°
			at 60 Hz (0 ~ 360)°	1.4°
SWR meters	40644	Frequency	9 kHz ~ 18 GHz	6.4×10^{-5}
		Output Level	30 kHz ~ 100 MHz	0.06 dB
			100 MHz ~ 10 GHz	0.08 dB
			(10 ~ 18) GHz	0.09 dB
		SWR	(30 kHz ~ 30 MHz)	
			1.05	0.019
			1.20	0.019
			1.50	0.019
			2.00	0.020
			(30 MHz ~ 2 GHz)	
			1.05	0.021
			1.20	0.021
		(2 ~ 18) GHz	1.50	0.021
			2.00	0.021
			1.05	0.018
			1.20	0.018
			1.50	0.018
			2.00	0.024

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF terminations	40645	(± 180 °)		Network Analyzer, Coaxial Mismatch/ SICT-CP-40645
(Open, Short, Phase)		10 Hz ~ 2 GHz	0.49°	
		(2 ~ 20) GHz	0.61°	
		(20 ~ 40) GHz	0.95°	
		(40 ~ 50) GHz	1.2°	
(Reflection coefficient)		(0 ~ 1)		
		10 Hz ~ 100 kHz	0.004 3	
		100 kHz ~ 500 MHz	0.006 0	
		500 MHz ~ 20 GHz	0.009 5	
		(20 ~ 40) GHz	0.016	
		(40 ~ 50) GHz	0.019	
(SWR)		(1 ~ ∞)		
		10 Hz ~ 100 kHz	0.008 6	
		100 kHz ~ 500 MHz	0.012	
		500 MHz ~ 20 GHz	0.019	
		(20 ~ 40) GHz	0.032	
		(40 ~ 50) GHz	0.038	
(Impedance)		(0.000 0 ~ 0.047 6)		
		10 Hz ~ 500 MHz	0.64 Ω	
		500 MHz ~ 20 GHz	1.0 Ω	
		(20 ~ 40) GHz	1.6 Ω	
		(40 ~ 50) GHz	2.0 Ω	
		(0.047 6 ~ 0.090 9)		
		10 Hz ~ 500 MHz	0.71 Ω	
		500 MHz ~ 20 GHz	1.1 Ω	
		(20 ~ 40) GHz	1.8 Ω	
		(40 ~ 50) GHz	2.2 Ω	
		(0.090 9 ~ 0.166 7)		
		10 Hz ~ 500 MHz	0.84 Ω	
		500 MHz ~ 20 GHz	1.3 Ω	
		(20 ~ 40) GHz	2.1 Ω	
		(40 ~ 50) GHz	2.7 Ω	
		(0.166 7 ~ 0.230 8)		
		10 Hz ~ 500 MHz	0.99 Ω	
		500 MHz ~ 20 GHz	1.6 Ω	
		(20 ~ 40) GHz	2.5 Ω	
		(40 ~ 50) GHz	3.1 Ω	
		(0.230 8 ~ 0.285 7)		
		10 Hz ~ 500 MHz	1.1 Ω	
		500 MHz ~ 20 GHz	1.8 Ω	
		(20 ~ 40) GHz	2.9 Ω	
		(40 ~ 50) GHz	3.6 Ω	
		(0.285 7 ~ 0.333 4)		
		10 Hz ~ 500 MHz	1.3 Ω	
		500 MHz ~ 20 GHz	2.1 Ω	
		(20 ~ 40) GHz	3.3 Ω	
		(40 ~ 50) GHz	4.2 Ω	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF terminations (Impedance Phase)	40645	(0.000 0 ~ 0.047 6, ±180°) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.68° 1.1° 1.7° 2.2°	Network Analyzer, Coaxial Mismatch/ SICT-CP-40645
		(0.047 6 ~ 0.090 9, ±180°) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.67° 1.1° 1.7° 2.1°	
		(0.090 9 ~ 0.166 7, ±180°) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.66° 1.1° 1.7° 2.1°	
		(0.166 7 ~ 0.230 8, ±180°) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.65° 1.0° 1.6° 2.1°	
		(0.230 8 ~ 0.285 7, ±180°) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.64° 1.0° 1.6° 2.0°	
		(0.285 7 ~ 0.333 4, ±180°) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.62° 1.0° 1.6° 2.0°	
Coaxial thermistor mounts	40646	(1 ~ 10) μW (9 ~ 100) kHz 100 kHz ~ 1 GHz (1 ~ 10) GHz (10 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz	0.4×10^{-2} 0.9×10^{-2} 1.1×10^{-2} 1.9×10^{-2} 2.8×10^{-2} 4.0×10^{-2}	Thermistor Mount, Synthesized Sweeper/ SICT-CP-40646
	Cal Factor	(10 μW ~ 10 mW) (9 ~ 100) kHz 100 kHz ~ 1 GHz (1 ~ 10) GHz (10 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz	0.4×10^{-2} 0.9×10^{-2} 1.1×10^{-2} 1.9×10^{-2} 2.6×10^{-2} 3.5×10^{-2}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Coaxial thermistor mounts	40646			Synthesized Sweeper/ SICT-CP-40646
Reflection coefficient		(0 ~ 1) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	4.2×10^{-3} 9.4×10^{-3} 1.5×10^{-2}	
SWR		(1 ~ ∞) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	9.7×10^{-3} 2.4×10^{-2} 3.8×10^{-2}	
RF voltmeters	40650			RF Millivolt Meter Calibrator/ SICT-CP-40650
RF Voltage		3 V 1 V 300 mV 270 mV 240 mV 210 mV 180 mV 150 mV 120 mV 100 mV 90 mV 60 mV 30 mV 10 mV 3 mV 1 mV	4.2 mV 1.4 mV 0.36 mV 0.32 mV 0.28 mV 0.26 mV 0.24 mV 0.22 mV 0.17 mV 0.16 mV 0.13 mV 0.11 mV 0.048 mV 0.020 mV 0.018 mV 0.013 mV	
Vector voltmeters	40651			Signal Generator/ SICT-CP-40651
RF Voltage		3 V 1 V 300 mV 100 mV 30 mV 10 mV 3 mV 1 mV	4.2 mV 1.4 mV 0.36 mV 0.16 mV 0.048 mV 0.046 mV 0.022 mV 0.024 mV	
RF Phase		(0 ~ 270)°	0.006°	
Field strength meters	40652			Signal Generator/
Center frequency		(9 ~ 100) kHz 0.1 MHz ~ 18 GHz	6.8×10^{-8} 6.2×10^{-9}	
Scale Fidelity		(0 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -100) dB	0.11 dB 0.12 dB 0.13 dB 0.18 dB	SICT-CP-40652
Frequency response		9 kHz ~ 500 MHz 500 MHz ~ 18 GHz	0.05 dB 0.08 dB	
AM/FM test sources	40653			Measuring Receiver/ SICT-CP-40653
Output frequency		(10 ~ 560) MHz	6.2×10^{-10}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Dip simulators	40654			Digital Oscilloscope/ SICT-CP-40654
DC Voltage		1 V (1 ~ 5) V (5 ~ 10) V (10 ~ 50) V (50 ~ 100) V (100 ~ 500) V	1.1×10^{-5} 6.4×10^{-6} 5.4×10^{-6} 9.3×10^{-6} 8.0×10^{-6} 1.1×10^{-5}	
AC Voltage		(50 ~ 60) Hz 50 V (50 ~ 100) V (100 ~ 500) V	5.0×10^{-5} 2.5×10^{-5} 1.6×10^{-4}	
Frequency		50 Hz 60 Hz	8.4×10^{-6} 8.3×10^{-6}	
Dip DC Voltage		(0 ~ 50) V 0 % (0 ~ 120) %	0.2 V 3.4×10^{-2}	
Dip AC Voltage		(50 Hz ~ 60 Hz, 0 V ~ 400 V) 0 % (0 ~ 120) %	0.9 V 3.4×10^{-2}	
Time measurement by section		100 ns ~ 2 μ s (2 ~ 4) μ s (4 ~ 400) μ s (0.4 ~ 2) ms 2 ms ~ 5 s	1.8×10^{-3} 2.0×10^{-3} 1.6×10^{-3} 2.0×10^{-3} 1.6×10^{-3}	
Inrush Current		(5 ~ 1 000) A	3.6×10^{-2}	
Phase Shifiting		at 50 Hz (0 ~ 360) $^{\circ}$ at 60 Hz (0 ~ 360) $^{\circ}$	1.2 $^{\circ}$ 1.4 $^{\circ}$	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Loop antennas Antenna Factor	40704	10 Hz ~ 30 MHz	1.2 dB	Signal generator1, Signal analyzer/ SICT-CP-40704
Monopole antennas Antenna Factor	40705	10 Hz ~ 30 MHz	1.4 dB	Signal generator1, Signal analyzer/ SICT-CP-40705

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Temperature generators: ovens, furnaces, isothermal liquid baths, ice-point baths, dry-block calibrators	50101	0 °C (-196 ~ -95) °C (-95 ~ -90) °C (-90 ~ 250) °C (250 ~ 550) °C (550 ~ 660) °C (660 ~ 1 100) °C (1 100 ~ 1 600) °C	0.010 °C 0.060 °C 0.030 °C 0.017 °C 0.020 °C 0.060 °C 0.7 °C 1.7 °C	SPRT, STANDARD TC/ SICT-CP-50101
Temperature indicators/recorders /controllers, temperature calibrators (Temperature indicators/recorders/controllers)	50102	With Sensor (-196 ~ 500) °C (500 ~ 660) °C (660 ~ 700) °C (700 ~ 900) °C (900 ~ 1 100) °C (1 100 ~ 1 400) °C (1 400 ~ 1 600) °C Without Sensor (-196 ~ 0) °C (0 ~ 100) °C (100 ~ 200) °C (200 ~ 300) °C (300 ~ 400) °C (400 ~ 500) °C (500 ~ 600) °C (600 ~ 700) °C (700 ~ 800) °C (800 ~ 1 300) °C (1 300 ~ 1 600) °C (temperature calibrators) Output (-196 ~ 500) °C (500 ~ 600) °C (600 ~ 800) °C (800 ~ 1 300) °C (1 300 ~ 1 600) °C Input (-196 ~ 500) °C (500 ~ 600) °C (600 ~ 800) °C (800 ~ 1 300) °C (1 300 ~ 1 600) °C	0.020 °C 0.045 °C 0.59 °C 0.60 °C 0.61 °C 2.0 °C 2.1 °C 0.010 °C 0.013 °C 0.018 °C 0.022 °C 0.025 °C 0.029 °C 0.033 °C 0.040 °C 0.044 °C 0.07 °C 0.09 °C 0.005 °C 0.006 °C 0.007 °C 0.08 °C 0.10 °C 0.03 °C 0.04 °C 0.05 °C 0.07 °C 0.09 °C	SPRT, STANDARD TC/ SICT-CP-50102
Glass thermometers; liquid-in-glass, Beckmann liquid-in-glass	50103	(-90 ~ -58) °C (-58 ~ 400) °C (400 ~ 500) °C	0.15 °C 0.04 °C 0.15 °C	SPRT/ SICT-CP-50103

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistance thermometers; SPRT, IPRT, thermistors, etc.	50104	(-196 ~ 100) °C (100 ~ 300) °C (300 ~ 500) °C (500 ~ 660) °C (Fixed point) Ar TP Hg TP H ₂ O TP Ga MP In FP Sn FP Zn FP Al FP Ag FP	0.019 °C 0.020 °C 0.021 °C 0.044 °C -189.344 2 °C -38.834 4 °C 0.01 °C 29.764 6 °C 156.598 5 °C 231.928 °C 419.527 °C 660.323 °C 961.78 °C	SPRT, Fixed point/ SICT-CP-50104
Thermal expansion thermometers; bimetal, gas or liquid type	50105	(-196 ~ -70) °C (-70 ~ 100) °C (100 ~ 200) °C (200 ~ 500) °C (500 ~ 650) °C	0.6 °C 0.2 °C 0.3 °C 0.6 °C 1.2 °C	SPRT/ SICT-CP-50105
Thermocouples: noble metal, base metal, pure metal, special type, etc.	50106	(0 ~ 1 100) °C (1 100 ~ 1 300) °C (1 300 ~ 1 600) °C (-196 ~ -90) °C (-90 ~ 300) °C (300 ~ 500) °C (500 ~ 660) °C (660 ~ 900) °C (900 ~ 1 100) °C (1 100 ~ 1 300) °C (Fixed point) H ₂ O ICE Point Sn FP Zn FP Al FP Ag FP Cu FP Co-C MP Fe MP	0.5 °C 1.7 °C 1.8 °C 0.4 °C 0.2 °C 0.3 °C 0.4 °C 1.1 °C 1.4 °C 1.8 °C 0.2 °C 0.2 °C 0.2 °C 0.2 °C 0.3 °C 0.3 °C 1.1 °C 1.6 °C	SPRT, Fixed point, STANDARD TC/ SICT-CP-50106

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Temperature transducers	50107	(-196 ~ 400) °C (400 ~ 500) °C (500 ~ 660) °C (660 ~ 800) °C (800 ~ 1 100) °C (1 100 ~ 1 300) °C (1 300 ~ 1 600) °C	0.031 °C 0.043 °C 0.072 °C 0.6 °C 0.7 °C 2.1 °C 2.2 °C	SPRT, THERMOCOUPLE, MULTIMETER SICT-CP-50107
Primary fixed-point cells and apparatus H ₂ O TP	50108	0.01 °C	0.24 mK	Triple-Point Cell SICT-CP-50108

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Optical pyrometers	50203	(900 ~ 1 800) °C	5 °C	Standard Lamp/ SICT-CP-50203
Standard radiation thermometers	50204	(-40 ~ -20) °C	0.9 °C	Transfer Standard Pyrometer/ SICT-CP-50204
		(-20 ~ 0) °C	0.7 °C	
		(0 ~ 10) °C	0.4 °C	
		(10 ~ 50) °C	0.3 °C	
		(50 ~ 100) °C	0.4 °C	
		(100 ~ 200) °C	0.5 °C	
		(200 ~ 300) °C	0.7 °C	
		(300 ~ 400) °C	0.8 °C	
		(400 ~ 500) °C	1.0 °C	
		(500 ~ 600) °C	1.1 °C	
		(600 ~ 700) °C	1.3 °C	
		(700 ~ 800) °C	1.6 °C	
		(800 ~ 900) °C	1.7 °C	
		(900 ~ 1 200) °C	1.8 °C	
		(1 200 ~ 1 400) °C	1.9 °C	
		(1 400 ~ 1 500) °C	2.0 °C	
		(1 500 ~ 1 600) °C	2.1 °C	
		(1 600 ~ 1 800) °C	2.5 °C	
		(1 800 ~ 2 000) °C	2.6 °C	
		(2 000 ~ 2 100) °C	4.2 °C	
		(2 100 ~ 2 200) °C	4.4 °C	
		(2 200 ~ 2 300) °C	4.6 °C	
		(2 300 ~ 2 400) °C	4.7 °C	
Thermal image apparatus	50205	(-40 ~ -20) °C	0.9 °C	Transfer Standard Pyrometer/ SICT-CP-50205
		(-20 ~ 0) °C	0.7 °C	
		(0 ~ 10) °C	0.4 °C	
		(10 ~ 50) °C	0.3 °C	
		(50 ~ 100) °C	0.4 °C	
		(100 ~ 200) °C	0.5 °C	
		(200 ~ 300) °C	0.7 °C	
		(300 ~ 400) °C	0.8 °C	
		(400 ~ 500) °C	1.0 °C	
		(500 ~ 600) °C	1.1 °C	
		(600 ~ 700) °C	1.3 °C	
		(700 ~ 800) °C	1.6 °C	
		(800 ~ 900) °C	1.7 °C	
		(900 ~ 1 200) °C	1.8 °C	SICT-CP-50205
		(1 200 ~ 1 400) °C	1.9 °C	
		(1 400 ~ 1 500) °C	2.0 °C	
		(1 500 ~ 1 600) °C	2.1 °C	
		(1 600 ~ 1 800) °C	2.5 °C	
		(1 800 ~ 2 000) °C	2.6 °C	
		(2 000 ~ 2 100) °C	4.2 °C	
		(2 100 ~ 2 200) °C	4.4 °C	
		(2 200 ~ 2 300) °C	4.6 °C	
		(2 300 ~ 2 400) °C	4.7 °C	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Blackbody furnaces	50206	(-40 ~ 0) °C (0 ~ 10) °C (10 ~ 50) °C (50 ~ 100) °C (100 ~ 200) °C (200 ~ 300) °C (300 ~ 400) °C (400 ~ 500) °C (500 ~ 600) °C (600 ~ 700) °C (700 ~ 1 100) °C (1 100 ~ 1 300) °C (1 300 ~ 1 500) °C (1 500 ~ 1 700) °C (1 700 ~ 1 800) °C (1 800 ~ 1 900) °C (1 900 ~ 2 000) °C (2 000 ~ 2 100) °C (2 100 ~ 2 200) °C (2 200 ~ 2 300) °C (2 300 ~ 2 400) °C	0.6 °C 0.4 °C 0.3 °C 0.4 °C 0.5 °C 0.7 °C 0.8 °C 0.9 °C 1.0 °C 1.1 °C 1.4 °C 1.5 °C 1.6 °C 1.8 °C 1.9 °C 2.0 °C 2.1 °C 3.7 °C 3.9 °C 4.1 °C 4.3 °C	Transfer Standard Pyrometer/ SICT-CP-50206
Others ; ear thermometers, etc.	50207	(30 ~ 45) °C	0.07 °C	Standard prt/ SICT-CP-50207

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Dew-point hygrometers; chilled mirror, alumina thinfilm, etc.	50301	(-90 ~ -80) °C D.P. (-80 ~ -70) °C D.P. (-70 ~ -50) °C D.P. (-50 ~ -20) °C D.P. (-20 ~ 90) °C D.P. (90 ~ 95) °C D.P.	0.60 °C D.P. 0.32 °C D.P. 0.20 °C D.P. 0.19 °C D.P. 0.13 °C D.P. 0.15 °C D.P.	Dewpoint Meter/ SICT-CP-50301
Relative humidity hygrometers; polimer thinfilm, hair, etc.	50302	(3 ~ 60) % R.H. (60 ~ 90) % R.H. (90 ~ 98) % R.H.	1.3 % R.H. 1.4 % R.H. 1.5 % R.H.	Dewpoint Meter/ SICT-CP-50302
		Temperature (-80 ~ 0) °C (0 ~ 80) °C (80 ~ 100) °C (100 ~ 180) °C	0.6 °C 0.3 °C 0.5 °C 1.5 °C	
Psychrometers; assmann ventilated, PRT type, etc.	50303			Dewpoint Meter/ SICT-CP-50303
assmann ventilated (humidity)		(10 ~ 50) % R.H. (50 ~ 70) % R.H. (70 ~ 90) % R.H. (90 ~ 95) % R.H.	1.3 % R.H. 1.4 % R.H. 1.5 % R.H. 1.6 % R.H.	
(Temperature)		(0 ~ 50) °C	0.3 °C	
PRT type (humidity)		(10 ~ 50) % R.H. (50 ~ 80) % R.H. (80 ~ 98) % R.H.	1.3 % R.H. 1.4 % R.H. 1.5 % R.H.	
(Temperature)		(0 ~ 80) °C (80 ~ 100) °C	0.3 °C 0.5 °C	
Temperature humidity recorders; Hygrothermograph, etc	50304			Dewpoint Meter/ SICT-CP-50304
Humidity		(5 ~ 70) % R.H. (70 ~ 95) % R.H.	2.1 % R.H. 2.2 % R.H.	
Temperature		(-20 ~ 80) °C	0.7 °C	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Transducers: dew-point /relative humidity (Dew-point Transducers)	50305			Dewpoint Meter/ SICT-CP-50305
Dew point		(-90 ~ -80) °C D.P. (-80 ~ -70) °C D.P. (-70 ~ -60) °C D.P. (-60 ~ -40) °C D.P. (-40 ~ -20) °C D.P. (-20 ~ 0) °C D.P. (0 ~ 50) °C D.P. (50 ~ 90) °C D.P. (90 ~ 95) °C D.P.	0.60 °C D.P. 0.33 °C D.P. 0.22 °C D.P. 0.21 °C D.P. 0.20 °C D.P. 0.15 °C D.P. 0.14 °C D.P. 0.15 °C D.P. 0.17 °C D.P.	
(Relative humidity Transducers)				
Humidity		(3 ~ 50) % R.H. (50 ~ 80) % R.H. (80 ~ 98) % R.H.	1.3 % R.H. 1.4 % R.H. 1.5 % R.H.	
Temperature		(-80 ~ 0) °C (0 ~ 80) °C (80 ~ 100) °C (100 ~ 180) °C	0.7 °C 0.3 °C 0.5 °C 1.5 °C	
Humidity generators; two-pressure, two-temperature, flow mixing humidity generator, constant temperature and humidity chamber, etc.	50306			Dewpoint Meter/ SICT-CP-50306
Dew point		(-90 ~ -80) °C D.P. (-80 ~ -70) °C D.P. (-70 ~ -50) °C D.P. (-50 ~ -30) °C D.P. (-30 ~ -10) °C D.P. (-10 ~ 60) °C D.P. (60 ~ 80) °C D.P. (80 ~ 95) °C D.P.	0.60 °C D.P. 0.32 °C D.P. 0.19 °C D.P. 0.17 °C D.P. 0.16 °C D.P. 0.13 °C D.P. 0.14 °C D.P. 0.15 °C D.P.	
Humidity		(3 ~ 20) % R.H. (20 ~ 30) % R.H. (30 ~ 40) % R.H. (40 ~ 50) % R.H. (50 ~ 70) % R.H. (70 ~ 90) % R.H. (90 ~ 98) % R.H.	1.8 % R.H. 1.7 % R.H. 1.4 % R.H. 1.5 % R.H. 1.6 % R.H. 1.8 % R.H. 1.9 % R.H.	
Temperature		(-90 ~ 50) °C (50 ~ 100) °C (100 ~ 200) °C	0.4 °C 0.5 °C 0.6 °C	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Cereal moisture meters Moisture	50401	(9 ~ 20) % M.C.	0.7 % M.C.	Balance/ SICT-CP-50401
Wood moisture meters Moisture	50402	(8 ~ 25) % M.C.	2.5 % M.C.	Balance/ SICT-CP-50402
Paper moisture meters Moisture	50403	(8 ~ 20) % M.C.	3.4 % M.C.	Balance/ SICT-CP-50403

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Soundcalibrators	60102	(250 Hz) (94 ~ 114) dB (1 000 Hz) (94 ~ 114) dB	0.10 dB 0.10 dB	Reference microphone/ SICT-CP-60102
Microphones	60104	(250 Hz) (-60 ~ -20) dB	0.15 dB	Reference microphone/ SICT-CP-60104
Sound level meters	60106	(31.5 ~ 12 500) Hz 31.5 Hz 63 Hz 125 Hz 250 Hz 500 Hz 1 000 Hz 2 000 Hz 4 000 Hz 8 000 Hz 12 500 Hz	0.4 dB 0.3 dB 0.3 dB 0.2 dB 0.2 dB 0.2 dB 0.2 dB 0.2 dB 0.2 dB 0.4 dB 0.6 dB	Multifunction Acoustic Calibrator/ SICT-CP-60107

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Vibration calibrators	60301	10 Hz (10 ~ 2 500) Hz	1.6×10^{-2} 1.5×10^{-2}	Reference Accelerometer/ SICT-CP-60301
Vibration transducers	60302	0.5 Hz (0.5 ~ 20) Hz (20 ~ 1 250) Hz (1 250 ~ 2 500) Hz (2 500 ~ 5 000) Hz	1.5×10^{-2} 1.4×10^{-2} 1.1×10^{-2} 1.2×10^{-2} 2.4×10^{-2}	Reference Accelerometer/ SICT-CP-60302
Vibration measuring instruments	60303			Reference Accelerometer/ SICT-CP-60303
Acceleration		10 Hz 20 Hz (20 ~ 1 250) Hz (1 250 ~ 2 500) Hz	1.7×10^{-2} 1.6×10^{-2} 1.5×10^{-2} 1.6×10^{-2}	
Velocity		10 Hz 20 Hz (20 ~ 1 250) Hz (1 250 ~ 2 500) Hz	1.7×10^{-2} 1.6×10^{-2} 1.5×10^{-2} 1.6×10^{-2}	
Displacement		(10 ~ 160) Hz (160 ~ 315) Hz (315 ~ 630) Hz	1.4×10^{-2} 2.1×10^{-2} 5.9×10^{-2}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Illuminance meters Illuminance	70101	(0.5 ~ 20 000) lx	1.7×10^{-2}	Illuminance Meters/ SICT-CP-70101
Luminance meters Luminance	70102	1 cd/m ² (1 ~ 10) cd/m ² (10 ~ 3 000) cd/m ² (3 000 ~ 15 000) cd/m ²	2.1 × 10 ⁻² 1.7 × 10 ⁻² 1.4 × 10 ⁻² 1.6 × 10 ⁻²	Luminance Standard Sources/ SICT-CP-70102
Total luminous flux meters Total luminous flux	70103	70 lm (70 ~ 4 650) lm	3.2 × 10 ⁻² 1.5 × 10 ⁻²	Total Luminous Flux Standard Lamps/ SICT-CP-70103
Luminance intensity meters Luminance	70104	(72 ~ 3 200) cd	3.7×10^{-2}	Luminous Intensity Standard Lamps, Illuminance Meters / SICT-CP-70104

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color temperature meters	70202			
Color temperature		(2 677 ~ 3 333) K	25 K	Color Temperature Standard Lamps/ SICT-CP-70202
Chromaticity		x y	0.004 0.004	
Color temperature standard lamps	70203			
Color temperature		(2 677 ~ 3 333) K	27 K	Spectroradiometers/ SICT-CP-70203
Chromaticity		x y	0.005 0.005	
Colorimeters; source color	70204			
Luminance		1 cd/m ² (1 ~ 10) cd/m ² (10 ~ 3 000) cd/m ² (3 000 ~ 15 000) cd/m ²	2.1 × 10 ⁻² 1.7 × 10 ⁻² 1.4 × 10 ⁻² 1.6 × 10 ⁻²	Luminance Standard Sources/ SICT-CP-70204
Chromaticity		(WHITE) x y (RED) x y (GREEN) x y (BLUE) x y (CIE Standard Illuminant A) x y	0.004 0.004 0.004 0.004 0.004 0.004 0.004 0.004 0.003 0.003	
Laser power meters	70207			
		(405 nm) (0.75 ~ 9) mW	1.2 × 10 ⁻²	Optical Power Meters/ SICT-CP-70207
		(660 nm) (0.7 ~ 47) mW	1.2 × 10 ⁻²	
		(785 nm) (0.7 ~ 46) mW	1.2 × 10 ⁻²	
		(1 080 nm) (1 ~ 40) W	3.3 × 10 ⁻²	
Standard LED light sources	70208			
Total luminous flux		(68.4 ~ 72.6) lm	3.8 × 10 ⁻²	Total Spectral Radiant Flux Meters/ SICT-CP-70208
Total luminous flux standard lamps	70209			
Total luminous flux		(320 ~ 10 000) lm	4.7 × 10 ⁻²	Total Luminous Flux Standard SICT-CP-70209
Optical detectors	70210			
Relative spectral responsivity		(0 ~ 1) (300) nm (300 ~ 305) nm (305 ~ 310) nm (310 ~ 315) nm	8.8 × 10 ⁻² 8.0 × 10 ⁻² 7.2 × 10 ⁻² 6.5 × 10 ⁻²	Photodiodes/ SICT-CP-70210

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Optical detectors Relative spectral responsivity	70210	(315 ~ 320) nm (320 ~ 325) nm (325 ~ 330) nm (330 ~ 375) nm (375 ~ 435) nm (435 ~ 955) nm (955 ~ 965) nm (965 ~ 970) nm (970 ~ 975) nm (975 ~ 980) nm (980 ~ 985) nm (985 ~ 990) nm (990 ~ 995) nm (995 ~ 1 035) nm (1 035 ~ 1 055) nm (1 055 ~ 1 085) nm (1 090 ~ 1 100) nm	2.9×10^{-2} 2.9×10^{-2}	Photodiodes/ SICT-CP-70210
Pyranometers and pyrheliometers Irradiance responsivity	70211	(250 ~ 2 500) nm (1 000 ± 150) W/m ²	2.9×10^{-2}	Standard pyranometers/ SICT-CP-70211
Display color analyzers; luminance, chromaticity, white balance, etc. Luminance	70213	1 cd/m ² (1 ~ 5) cd/m ² (5 ~ 200) cd/m ²	3.8×10^{-2} 1.8×10^{-2} 1.7×10^{-2}	Luminance Meters/ SICT-CP-70213
		(WHITE) x y (RED) x y (GREEN) x y (BLUE) x y	0.004 4 0.006 1 0.003 6 0.003 3 0.003 5 0.004 2 0.003 5 0.003 2	
Luminous intensity standard lamps Luminous intensity	70214	(10 ~ 20 000) cd	4.0×10^{-2}	Spectroradiometers/ SICT-CP-70214
Spectral irradiance standard lamps Spectral irradiance	70215	250 nm (250 ~ 255) nm (255 ~ 300) nm (300 ~ 380) nm (380 ~ 450) nm (450 ~ 555) nm (555 ~ 660) nm (660 ~ 900) nm (900 ~ 1 600) nm (1 600 ~ 2 000) nm (2 000 ~ 2 300) nm (2 300 ~ 2 400) nm	4.5×10^{-2} 4.4×10^{-2} 4.2×10^{-2} 3.5×10^{-2} 3.1×10^{-2} 2.9×10^{-2} 2.8×10^{-2} 2.7×10^{-2} 2.6×10^{-2} 2.9×10^{-2} 3.0×10^{-2} 3.4×10^{-2}	Spectral Irradiance Standard Lamps/ SICT-CP-70215
		Illuminance (800 ~ 7 200) lx	2.5×10^{-2}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Total spectral radiant flux standard lamps	70216			Total Spectral Radiant Flux Standard Lamps/ SICT-CP-70216
Total spectral radiant flux		350 nm (350 ~ 365) nm (365 ~ 380) nm (380 ~ 400) nm (400 ~ 455) nm (455 ~ 850) nm	6.7 × 10 ⁻² 6.3 × 10 ⁻² 5.8 × 10 ⁻² 4.2 × 10 ⁻² 3.9 × 10 ⁻² 3.6 × 10 ⁻²	
Luminance standard lamps	70217			Luminance Standard Sources/ SICT-CP-70217
Luminance		1 cd/m ² (1 ~ 10) cd/m ² (10 ~ 3 000) cd/m ² (3 000 ~ 15 000) cd/m ²	2.2 × 10 ⁻² 1.8 × 10 ⁻² 1.5 × 10 ⁻² 1.8 × 10 ⁻²	
Chromaticity		(WHITE) x y (RED) x y (GREEN) x y (BLUE) x y (CIE Standard Illuminant A) x y	0.005 0.005 0.005 0.005 0.005 0.005 0.005 0.005 0.004 0.004	
Spectral radiance standard lamps	70218			Spectral Radiance Standard Sources/ SICT-CP-70218
Spectral radiance		300 nm (300 ~ 305) nm (305 ~ 310) nm (310 ~ 315) nm (315 ~ 320) nm (320 ~ 330) nm (330 ~ 340) nm (340 ~ 425) nm (425 ~ 470) nm (470 ~ 1 050) nm (1 050 ~ 1 600) nm	2.0 × 10 ⁻¹ 1.7 × 10 ⁻¹ 1.2 × 10 ⁻¹ 8.9 × 10 ⁻² 7.4 × 10 ⁻² 4.8 × 10 ⁻² 4.1 × 10 ⁻² 3.5 × 10 ⁻² 3.0 × 10 ⁻² 2.8 × 10 ⁻² 3.0 × 10 ⁻²	
UV irradiance meters	70219			UV Meter Standard Detectors/ SICT-CP-70219
Irradiance (UV Meter)		(254 nm) 50 μW/cm ² ~ 3 mW/cm ² (365 nm) 10 μW/cm ² ~ 230 mW/cm ² (405 nm) 10 μW/cm ² ~ 230 mW/cm ²	3.7 × 10 ⁻² 3.8 × 10 ⁻² 3.8 × 10 ⁻²	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Spectral irradiance meters	70220			Spectral Irradiance Standard
Wavelength		(350 ~ 1 694) nm	0.25 nm	Lamps/ SICT-CP-70220
Spectral irradiance		250 nm	4.5×10^{-2}	
		(250 ~ 255) nm	4.2×10^{-2}	
		(255 ~ 300) nm	4.1×10^{-2}	
		(300 ~ 350) nm	3.7×10^{-2}	
		(350 ~ 380) nm	3.5×10^{-2}	
		(380 ~ 450) nm	3.1×10^{-2}	
		(450 ~ 550) nm	2.9×10^{-2}	
		(550 ~ 900) nm	2.7×10^{-2}	
		(900 ~ 1 615) nm	2.6×10^{-2}	
		(1 615 ~ 2 315) nm	2.9×10^{-2}	
		(2 315 ~ 2 365) nm	3.4×10^{-2}	
		(2 365 ~ 2 400) nm	4.0×10^{-2}	
Illuminance		(800 ~ 7 200) lx	1.9×10^{-2}	
Total spectral radiant flux meters	70221			Total Spectral Radiant Flux
Wavelength		(350 ~ 850) nm	0.25 nm	Standard Lamps/ SICT-CP-70221
Total spectral radiant flux		350 nm	2.0×10^{-2}	
		(350 ~ 365) nm	1.8×10^{-2}	
		(365 ~ 375) nm	1.7×10^{-2}	
		(375 ~ 390) nm	1.6×10^{-2}	
		(390 ~ 445) nm	1.5×10^{-2}	
		(445 ~ 850) nm	1.4×10^{-2}	
Spectral radiance meters	70222			Spectral Radiance Standard
Wavelength		(350 ~ 1 694) nm	0.25 nm	Sources/ SICT-CP-70222
Spectral radiance		300 nm	2.0×10^{-1}	
		(300 ~ 305) nm	1.7×10^{-1}	
		(305 ~ 310) nm	1.2×10^{-1}	
		(310 ~ 315) nm	8.8×10^{-2}	
		(315 ~ 320) nm	7.2×10^{-2}	
		(320 ~ 325) nm	5.5×10^{-2}	
		(325 ~ 335) nm	4.6×10^{-2}	
		(335 ~ 345) nm	3.7×10^{-2}	
		(345 ~ 405) nm	3.5×10^{-2}	
		(405 ~ 455) nm	3.0×10^{-2}	
		(455 ~ 755) nm	2.6×10^{-2}	
		(755 ~ 1 400) nm	2.7×10^{-2}	
		(1 400 ~ 1 525) nm	3.0×10^{-2}	
		(1 525 ~ 1 600) nm	2.8×10^{-2}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source A (2°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		X	0.94	
		Y	0.85	
		Z	0.30	
2. L,Gray		X	0.64	
		Y	0.59	
		Z	0.22	
3. M,Gray		X	0.30	
		Y	0.27	
		Z	0.11	
4. D,Gray		X	0.11	
		Y	0.10	
		Z	0.04	
5. Red		X	0.36	
		Y	0.22	
		Z	0.06	
6. Yellow		X	0.78	
		Y	0.68	
		Z	0.08	
7. Green		X	0.19	
		Y	0.22	
		Z	0.08	
8. Cyan		X	0.17	
		Y	0.20	
		Z	0.16	
Included Reflectance Std. Light Source A (2°)				
1. White		L*	0.36	
		a*	0.07	
		b*	0.07	
2. L,Gray		L*	0.32	
		a*	0.06	
		b*	0.06	
3. M,Gray		L*	0.25	
		a*	0.05	
		b*	0.05	
4. D,Gray		L*	0.17	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.25	
		a*	0.23	
		b*	0.26	
6. Yellow		L*	0.34	
		a*	0.12	
		b*	0.37	
7. Green		L*	0.24	
		a*	0.11	
		b*	0.07	
8. Cyan		L*	0.23	
		a*	0.14	
		b*	0.14	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source A (2°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		x	0.000 5	
		y	0.000 5	
2. L,Gray		x	0.000 5	
		y	0.000 5	
3. M,Gray		x	0.000 5	
		y	0.000 5	
4. D,Gray		x	0.000 5	
		y	0.000 5	
5. Red		x	0.000 8	
		y	0.000 5	
6. Yellow		x	0.000 5	
		y	0.000 5	
7. Green		x	0.000 5	
		y	0.000 5	
8. Cyan		x	0.000 5	
		y	0.000 5	
Included Reflectance Std. Light Source A (10°)				
1. White		X	0.95	
		Y	0.85	
		Z	0.30	
2. L,Gray		X	0.65	
		Y	0.59	
		Z	0.21	
3. M,Gray		X	0.30	
		Y	0.27	
		Z	0.11	
4. D,Gray		X	0.11	
		Y	0.10	
		Z	0.04	
5. Red		X	0.35	
		Y	0.22	
		Z	0.06	
6. Yellow		X	0.78	
		Y	0.66	
		Z	0.08	
7. Green		X	0.20	
		Y	0.22	
		Z	0.08	
8. Cyan		X	0.17	
		Y	0.20	
		Z	0.15	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source A (10°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		L*	0.36	
		a*	0.06	
		b*	0.07	
2. L,Gray		L*	0.32	
		a*	0.06	
		b*	0.06	
3. M,Gray		L*	0.25	
		a*	0.05	
		b*	0.05	
4. D,Gray		L*	0.17	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.25	
		a*	0.22	
		b*	0.26	
6. Yellow		L*	0.33	
		a*	0.12	
		b*	0.39	
7. Green		L*	0.24	
		a*	0.11	
		b*	0.07	
8. Cyan		L*	0.23	
		a*	0.13	
		b*	0.14	
Include Reflectance Std. Light Source Type A (10°)				
1. White		x	0.000 5	
		y	0.000 5	
2. I,Gray		x	0.000 5	
		y	0.000 5	
3. M,Gray		x	0.000 5	
		y	0.000 5	
4. D,Gray		x	0.000 5	
		y	0.000 5	
5. Red		x	0.000 7	
		y	0.000 5	
6. Yellow		x	0.000 5	
		y	0.000 5	
7. Green		x	0.000 5	
		y	0.000 5	
8. Cyan		x	0.000 5	
		y	0.000 5	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source C (2°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		X	0.83	
		Y	0.85	
		Z	0.97	
2. L,Gray		X	0.58	
		Y	0.59	
		Z	0.69	
3. M,Gray		X	0.27	
		Y	0.28	
		Z	0.32	
4. D,Gray		X	0.10	
		Y	0.10	
		Z	0.11	
5. Red		X	0.25	
		Y	0.18	
		Z	0.15	
6. Yellow		X	0.56	
		Y	0.63	
		Z	0.22	
7. Green		X	0.18	
		Y	0.24	
		Z	0.21	
8. Cyan		X	0.20	
		Y	0.23	
		Z	0.49	
Included Reflectance Std. Light Source C (2°)				
1. White		L*	0.36	
		a*	0.09	
		b*	0.08	
2. L,Gray		L*	0.32	
		a*	0.08	
		b*	0.07	
3. M,Gray		L*	0.25	
		a*	0.06	
		b*	0.06	
4. D,Gray		L*	0.17	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.24	
		a*	0.24	
		b*	0.21	
6. Yellow		L*	0.33	
		a*	0.16	
		b*	0.41	
7. Green		L*	0.24	
		a*	0.13	
		b*	0.08	
8. Cyan		L*	0.24	
		a*	0.13	
		b*	0.13	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source C (2 °)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		x	0.000 5	
		y	0.000 5	
2. I,Gray		x	0.000 5	
		y	0.000 5	
3. M,Gray		x	0.000 5	
		y	0.000 5	
4. D,Gray		x	0.000 5	
		y	0.000 5	
5. Red		x	0.001 3	
		y	0.000 5	
6. Yellow		x	0.000 5	
		y	0.000 7	
7. Green		x	0.000 5	
		y	0.000 5	
8. Cyan		x	0.000 5	
		y	0.000 5	
Included Reflectance Std. Light Source C (10 °)				
1. White		X	0.82	
		Y	0.85	
		Z	0.95	
2. L,Gray		X	0.57	
		Y	0.59	
		Z	0.68	
3. M,Gray		X	0.27	
		Y	0.28	
		Z	0.32	
4. D,Gray		X	0.10	
		Y	0.10	
		Z	0.11	
5. Red		X	0.24	
		Y	0.18	
		Z	0.14	
6. Yellow		X	0.58	
		Y	0.60	
		Z	0.20	
7. Green		X	0.18	
		Y	0.24	
		Z	0.19	
8. Cyan		X	0.20	
		Y	0.24	
		Z	0.48	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source C (10°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		L*	0.36	
		a*	0.09	
		b*	0.07	
2. L,Gray		L*	0.32	
		a*	0.07	
		b*	0.07	
3. M,Gray		L*	0.25	
		a*	0.06	
		b*	0.05	
4. D,Gray		L*	0.17	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.24	
		a*	0.23	
		b*	0.20	
6. Yellow		L*	0.32	
		a*	0.16	
		b*	0.42	
7. Green		L*	0.24	
		a*	0.12	
		b*	0.12	
8. Cyan		L*	0.24	
		a*	0.11	
		b*	0.15	
Included Reflectance Std. Light Source C (10°)				
1. White		x	0.000 5	
		y	0.000 5	
2. I,Gray		x	0.000 5	
		y	0.000 5	
3. M,Gray		x	0.000 5	
		y	0.000 5	
4. D,Gray		x	0.000 5	
		y	0.000 5	
5. Red		x	0.001 3	
		y	0.000 5	
6. Yellow		x	0.000 5	
		y	0.000 7	
7. Green		x	0.000 5	
		y	0.000 5	
8. Cyan		x	0.000 5	
		y	0.000 5	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source D65 (2°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		X	0.81	
		Y	0.85	
		Z	0.89	
2. L,Gray		X	0.56	
		Y	0.59	
		Z	0.64	
3. M,Gray		X	0.26	
		Y	0.28	
		Z	0.30	
4. D,Gray		X	0.09	
		Y	0.10	
		Z	0.10	
5. Red		X	0.25	
		Y	0.18	
		Z	0.14	
6. Yellow		X	0.58	
		Y	0.63	
		Z	0.20	
7. Green		X	0.17	
		Y	0.25	
		Z	0.20	
8. Cyan		X	0.19	
		Y	0.23	
		Z	0.46	
Included Reflectance Std. Light Source D65 (2°)				
1. White		L*	0.36	
		a*	0.07	
		b*	0.07	
2. L,Gray		L*	0.32	
		a*	0.07	
		b*	0.06	
3. M,Gray		L*	0.25	
		a*	0.06	
		b*	0.05	
4. D,Gray		L*	0.17	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.24	
		a*	0.25	
		b*	0.21	
6. Yellow		L*	0.33	
		a*	0.16	
		b*	0.40	
7. Green		L*	0.24	
		a*	0.13	
		b*	0.08	
8. Cyan		L*	0.24	
		a*	0.14	
		b*	0.13	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source D65 (2 °)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		x	0.000 5	
		y	0.000 5	
2. I,Gray		x	0.000 5	
		y	0.000 5	
3. M,Gray		x	0.000 5	
		y	0.000 5	
4. D,Gray		x	0.000 5	
		y	0.000 5	
5. Red		x	0.001 3	
		y	0.000 5	
6. Yellow		x	0.000 6	
		y	0.000 6	
7. Green		x	0.000 5	
		y	0.000 5	
8. Cyan		x	0.000 5	
		y	0.000 5	
Included Reflectance Std. Light Source D65 (10 °)				
1. White		X	0.80	
		Y	0.85	
		Z	0.88	
2. L,Gray		X	0.56	
		Y	0.59	
		Z	0.63	
3. M,Gray		X	0.26	
		Y	0.28	
		Z	0.29	
4. D,Gray		X	0.09	
		Y	0.10	
		Z	0.10	
5. Red		X	0.24	
		Y	0.18	
		Z	0.13	
6. Yellow		X	0.58	
		Y	0.60	
		Z	0.19	
7. Green		X	0.18	
		Y	0.25	
		Z	0.18	
8. Cyan		X	0.19	
		Y	0.24	
		Z	0.45	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Included Reflectance Std. Light Source D65 (10 °)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		L*	0.36	
		a*	0.07	
		b*	0.07	
2. L,Gray		L*	0.32	
		a*	0.06	
		b*	0.06	
3. M,Gray		L*	0.25	
		a*	0.06	
		b*	0.05	
4. D,Gray		L*	0.17	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.24	
		a*	0.24	
		b*	0.20	
6. Yellow		L*	0.32	
		a*	0.16	
		b*	0.41	
7. Green		L*	0.24	
		a*	0.12	
		b*	0.08	
8. Cyan		L*	0.24	
		a*	0.12	
		b*	0.12	
Included Reflectance Std. Light Source D65 (10 °)				
1. White		x	0.000 5	
		y	0.000 5	
2. I,Gray		x	0.000 5	
		y	0.000 5	
3. M,Gray		x	0.000 5	
		y	0.000 5	
4. D,Gray		x	0.000 5	
		y	0.000 5	
5. Red		x	0.001 3	
		y	0.000 5	
6. Yellow		x	0.000 6	
		y	0.000 6	
7. Green		x	0.000 5	
		y	0.000 5	
8. Cyan		x	0.000 5	
		y	0.000 5	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color	70301			Color Standard Tiles/ SICT-CP-70301
Exclude Reflectance Std. Light Source Type A (2°)				
1. White		X	0.90	
		Y	0.81	
		Z	0.29	
2. L,Gray		X	0.60	
		Y	0.55	
		Z	0.20	
3. M,Gray		X	0.26	
		Y	0.24	
		Z	0.10	
4. D,Gray		X	0.07	
		Y	0.06	
		Z	0.03	
5. Red		X	0.31	
		Y	0.20	
		Z	0.10	
6. Yellow		X	0.73	
		Y	0.64	
		Z	0.07	
7. Green		X	0.15	
		Y	0.19	
		Z	0.07	
8. Cyan		X	0.13	
		Y	0.16	
		Z	0.14	
Exclude Reflectance Std. Light Source Type A (2°)				
1. White		L*	0.35	
		a*	0.07	
		b*	0.07	
2. L,Gray		L*	0.31	
		a*	0.06	
		b*	0.06	
3. M,Gray		L*	0.24	
		a*	0.05	
		b*	0.05	
4. D,Gray		L*	0.14	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.25	
		a*	0.30	
		b*	0.70	
6. Yellow		L*	0.33	
		a*	0.12	
		b*	0.45	
7. Green		L*	0.22	
		a*	0.12	
		b*	0.08	
8. Cyan		L*	0.22	
		a*	0.16	
		b*	0.15	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Exclude Reflectance Std. Light Source Type A (2 °)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		x y	0.000 5 0.000 5	
2. I,Gray		x y	0.000 5 0.000 5	
3. M,Gray		x y	0.000 5 0.000 5	
4. D,Gray		x y	0.000 5 0.000 5	
5. Red		x y	0.001 1 0.000 5	
6. Yellow		x y	0.000 5 0.000 5	
7. Green		x y	0.000 5 0.000 5	
8. Cyan		x y	0.000 5 0.000 5	
Exclude Reflectance Std. Light Source Type A (10 °)				
1. White		X Y Z	0.91 0.81 0.28	
2. L,Gray		X Y Z	0.61 0.55 0.20	
3. M,Gray		X Y Z	0.26 0.24 0.09	
4. D,Gray		X Y Z	0.07 0.06 0.03	
5. Red		X Y Z	0.30 0.20 0.10	
6. Yellow		X Y Z	0.74 0.62 0.07	
7. Green		X Y Z	0.15 0.19 0.06	
8. Cyan		X Y Z	0.13 0.17 0.14	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Excluded Reflectance Std. Light Source A (10°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		L*	0.35	
		a*	0.06	
		b*	0.07	
2. L,Gray		L*	0.31	
		a*	0.05	
		b*	0.06	
3. M,Gray		L*	0.24	
		a*	0.05	
		b*	0.05	
4. D,Gray		L*	0.14	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.25	
		a*	0.30	
		b*	0.70	
6. Yellow		L*	0.33	
		a*	0.12	
		b*	0.49	
7. Green		L*	0.22	
		a*	0.12	
		b*	0.08	
8. Cyan		L*	0.22	
		a*	0.14	
		b*	0.15	
Exclude Reflectance Std. Light Source Type A (10°)				
1. White		x	0.000 5	
		y	0.000 5	
2. I,Gray		x	0.000 5	
		y	0.000 5	
3. M,Gray		x	0.000 5	
		y	0.000 5	
4. D,Gray		x	0.000 5	
		y	0.000 5	
5. Red		x	0.001 1	
		y	0.000 5	
6. Yellow		x	0.000 5	
		y	0.000 5	
7. Green		x	0.000 5	
		y	0.000 5	
8. Cyan		x	0.000 5	
		y	0.000 5	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Excluded Reflectance Std. Light Source C (2°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		X	0.79	
		Y	0.81	
		Z	0.92	
2. L,Gray		X	0.54	
		Y	0.55	
		Z	0.65	
3. M,Gray		X	0.23	
		Y	0.24	
		Z	0.28	
4. D,Gray		X	0.06	
		Y	0.06	
		Z	0.07	
5. Red		X	0.21	
		Y	0.20	
		Z	0.20	
6. Yellow		X	0.56	
		Y	0.60	
		Z	0.19	
7. Green		X	0.14	
		Y	0.20	
		Z	0.16	
8. Cyan		X	0.16	
		Y	0.19	
		Z	0.45	
Exclude Reflectance Std. Light Source Type C (2°)				
1. White		L*	0.35	
		a*	0.09	
		b*	0.07	
2. L,Gray		L*	0.31	
		a*	0.07	
		b*	0.06	
3. M,Gray		L*	0.24	
		a*	0.06	
		b*	0.05	
4. D,Gray		L*	0.14	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.25	
		a*	0.40	
		b*	0.60	
6. Yellow		L*	0.32	
		a*	0.17	
		b*	0.52	
7. Green		L*	0.23	
		a*	0.14	
		b*	0.08	
8. Cyan		L*	0.23	
		a*	0.14	
		b*	0.14	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Excluded Reflectance Std. Light Source C (2 °)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		x y	0.000 5 0.000 5	
2. I,Gray		x y	0.000 5 0.000 5	
3. M,Gray		x y	0.000 5 0.000 5	
4. D,Gray		x y	0.000 5 0.000 5	
5. Red		x y	0.002 8 0.000 5	
6. Yellow		x y	0.000 7 0.000 7	
7. Green		x y	0.000 5 0.000 5	
8. Cyan		x y	0.000 5 0.000 5	
Exclude Reflectance Std. Light Source Type C (10 °)				
1. White		X Y Z	0.79 0.81 0.90	
2. L,Gray		X Y Z	0.54 0.55 0.63	
3. M,Gray		X Y Z	0.23 0.24 0.27	
4. D,Gray		X Y Z	0.06 0.06 0.07	
5. Red		X Y Z	0.20 0.20 0.20	
6. Yellow		X Y Z	0.55 0.56 0.17	
7. Green		X Y Z	0.14 0.21 0.14	
8. Cyan		X Y Z	0.17 0.20 0.44	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Excluded Reflectance Std. Light Source C (10 °)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		L*	0.35	
		a*	0.08	
		b*	0.07	
2. L,Gray		L*	0.31	
		a*	0.07	
		b*	0.06	
3. M,Gray		L*	0.24	
		a*	0.06	
		b*	0.05	
4. D,Gray		L*	0.14	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.30	
		a*	0.40	
		b*	0.60	
6. Yellow		L*	0.32	
		a*	0.17	
		b*	0.55	
7. Green		L*	0.23	
		a*	0.13	
		b*	0.14	
8. Cyan		L*	0.23	
		a*	0.12	
		b*	0.16	
Exclude Reflectance Std. Light Source Type C (10 °)				
1. White		x	0.000 5	
		y	0.000 5	
2. I,Gray		x	0.000 5	
		y	0.000 5	
3. M,Gray		x	0.000 5	
		y	0.000 5	
4. D,Gray		x	0.000 5	
		y	0.000 5	
5. Red		x	0.002 9	
		y	0.000 5	
6. Yellow		x	0.000 6	
		y	0.000 8	
7. Green		x	0.000 5	
		y	0.000 5	
8. Cyan		x	0.000 5	
		y	0.000 5	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Excluded Reflectance Std. Light Source	70301			Color Standard Tiles/ SICT-CP-70301
D65 (2 °)				
1. White		X	0.77	
		Y	0.81	
		Z	0.85	
2. L,Gray		X	0.52	
		Y	0.55	
		Z	0.60	
3. M,Gray		X	0.23	
		Y	0.24	
		Z	0.26	
4. D,Gray		X	0.06	
		Y	0.06	
		Z	0.06	
5. Red		X	0.20	
		Y	0.20	
		Z	0.20	
6. Yellow		X	0.55	
		Y	0.60	
		Z	0.18	
7. Green		X	0.14	
		Y	0.21	
		Z	0.15	
8. Cyan		X	0.16	
		Y	0.19	
		Z	0.41	
Exclude Reflectance Std. Light Source Type				
D65 (2 °)				
1. White		L*	0.35	
		a*	0.07	
		b*	0.07	
2. L,Gray		L*	0.31	
		a*	0.07	
		b*	0.06	
3. M,Gray		L*	0.24	
		a*	0.06	
		b*	0.05	
4. D,Gray		L*	0.14	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.30	
		a*	0.40	
		b*	0.60	
6. Yellow		L*	0.32	
		a*	0.17	
		b*	0.51	
7. Green		L*	0.23	
		a*	0.14	
		b*	0.08	
8. Cyan		L*	0.23	
		a*	0.15	
		b*	0.14	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Excluded Reflectance Std. Light Source D65 (2 °)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		x	0.000 5	
		y	0.000 5	
2. I,Gray		x	0.000 5	
		y	0.000 5	
3. M,Gray		x	0.000 5	
		y	0.000 5	
4. D,Gray		x	0.000 5	
		y	0.000 5	
5. Red		x	0.002 7	
		y	0.000 5	
6. Yellow		x	0.000 6	
		y	0.000 7	
7. Green		x	0.000 5	
		y	0.000 5	
8. Cyan		x	0.000 5	
		y	0.000 5	
Exclude Reflectance Std. Light Source Type D65 (10 °)				
1. White		X	0.77	
		Y	0.81	
		Z	0.84	
2. L,Gray		X	0.52	
		Y	0.55	
		Z	0.59	
3. M,Gray		X	0.23	
		Y	0.24	
		Z	0.25	
4. D,Gray		X	0.06	
		Y	0.06	
		Z	0.06	
5. Red		X	0.20	
		Y	0.20	
		Z	0.20	
6. Yellow		X	0.54	
		Y	0.57	
		Z	0.17	
7. Green		X	0.14	
		Y	0.21	
		Z	0.14	
8. Cyan		X	0.16	
		Y	0.20	
		Z	0.41	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Excluded Reflectance Std. Light Source D65 (10 °)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		L*	0.35	
		a*	0.07	
		b*	0.07	
2. L,Gray		L*	0.31	
		a*	0.06	
		b*	0.06	
3. M,Gray		L*	0.24	
		a*	0.05	
		b*	0.05	
4. D,Gray		L*	0.14	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.30	
		a*	0.40	
		b*	0.60	
6. Yellow		L*	0.32	
		a*	0.17	
		b*	0.55	
7. Green		L*	0.23	
		a*	0.13	
		b*	0.09	
8. Cyan		L*	0.23	
		a*	0.12	
		b*	0.13	
Exclude Reflectance Std. Light Source Type D65 (10 °)				
1. White		x	0.000 5	
		y	0.000 5	
2. I,Gray		x	0.000 5	
		y	0.000 5	
3. M,Gray		x	0.000 5	
		y	0.000 5	
4. D,Gray		x	0.000 5	
		y	0.000 5	
5. Red		x	0.002 8	
		y	0.000 5	
6. Yellow		x	0.000 7	
		y	0.000 7	
7. Green		x	0.000 5	
		y	0.000 5	
8. Cyan		x	0.000 5	
		y	0.000 5	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Included Reflectance Std. Light Source A (2°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White		X	0.94	
		Y	0.85	
		Z	0.30	
2. L,Gray		X	0.64	
		Y	0.59	
		Z	0.22	
3. M,Gray		X	0.30	
		Y	0.27	
		Z	0.11	
4. D,Gray		X	0.11	
		Y	0.10	
		Z	0.04	
5. Red		X	0.36	
		Y	0.22	
		Z	0.06	
6. Yellow		X	0.78	
		Y	0.68	
		Z	0.08	
7. Green		X	0.19	
		Y	0.22	
		Z	0.08	
8. Cyan		X	0.17	
		Y	0.20	
		Z	0.16	
Included Reflectance Std. Light Source A (2°)				
1. White		L*	0.37	
		a*	0.07	
		b*	0.07	
2. L,Gray		L*	0.32	
		a*	0.06	
		b*	0.06	
3. M,Gray		L*	0.25	
		a*	0.05	
		b*	0.05	
4. D,Gray		L*	0.17	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.25	
		a*	0.23	
		b*	0.26	
6. Yellow		L*	0.35	
		a*	0.12	
		b*	0.37	
7. Green		L*	0.24	
		a*	0.11	
		b*	0.07	
8. Cyan		L*	0.23	
		a*	0.14	
		b*	0.14	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles	70304			Color Standard Tiles/ SICT-CP-70304
Included Reflectance Std. Light Source A (2°)				
1. White		x	0.000 8	
		y	0.000 7	
2. L,Gray		x	0.000 8	
		y	0.000 7	
3. M,Gray		x	0.000 8	
		y	0.000 7	
4. D,Gray		x	0.000 8	
		y	0.000 7	
5. Red		x	0.001 0	
		y	0.000 7	
6. Yellow		x	0.000 8	
		y	0.000 8	
7. Green		x	0.000 7	
		y	0.000 8	
8. Cyan		x	0.000 7	
		y	0.000 7	
Included Reflectance Std. Light Source A (10°)				
1. White		X	0.95	
		Y	0.85	
		Z	0.30	
2. L,Gray		X	0.65	
		Y	0.59	
		Z	0.21	
3. M,Gray		X	0.30	
		Y	0.27	
		Z	0.11	
4. D,Gray		X	0.11	
		Y	0.10	
		Z	0.04	
5. Red		X	0.35	
		Y	0.22	
		Z	0.06	
6. Yellow		X	0.78	
		Y	0.66	
		Z	0.08	
7. Green		X	0.20	
		Y	0.22	
		Z	0.08	
8. Cyan		X	0.17	
		Y	0.20	
		Z	0.15	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Included Reflectance Std. Light Source A (10°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White		L*	0.37	
		a*	0.06	
		b*	0.07	
2. L,Gray		L*	0.32	
		a*	0.06	
		b*	0.06	
3. M,Gray		L*	0.25	
		a*	0.05	
		b*	0.05	
4. D,Gray		L*	0.17	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.25	
		a*	0.22	
		b*	0.26	
6. Yellow		L*	0.34	
		a*	0.12	
		b*	0.39	
7. Green		L*	0.24	
		a*	0.11	
		b*	0.07	
8. Cyan		L*	0.23	
		a*	0.13	
		b*	0.14	
Include Reflectance Std. Light Source Type A (10°)				
1. White		x	0.000 8	
		y	0.000 7	
2. I,Gray		x	0.000 8	
		y	0.000 7	
3. M,Gray		x	0.000 8	
		y	0.000 7	
4. D,Gray		x	0.000 8	
		y	0.000 7	
5. Red		x	0.001 0	
		y	0.000 7	
6. Yellow		x	0.000 8	
		y	0.000 8	
7. Green		x	0.000 7	
		y	0.000 8	
8. Cyan		x	0.000 7	
		y	0.000 7	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles				
Included Reflectance Std. Light Source	70304			Color Standard Tiles/ SICT-CP-70304
C (2°)				
1. White		X	0.83	
		Y	0.85	
		Z	0.97	
2. L,Gray		X	0.58	
		Y	0.59	
		Z	0.69	
3. M,Gray		X	0.27	
		Y	0.28	
		Z	0.32	
4. D,Gray		X	0.10	
		Y	0.10	
		Z	0.11	
5. Red		X	0.25	
		Y	0.18	
		Z	0.15	
6. Yellow		X	0.57	
		Y	0.63	
		Z	0.22	
7. Green		X	0.18	
		Y	0.24	
		Z	0.21	
8. Cyan		X	0.20	
		Y	0.23	
		Z	0.49	
Included Reflectance Std. Light Source				
C (2°)				
1. White		L*	0.37	
		a*	0.09	
		b*	0.08	
2. L,Gray		L*	0.32	
		a*	0.08	
		b*	0.07	
3. M,Gray		L*	0.25	
		a*	0.06	
		b*	0.06	
4. D,Gray		L*	0.17	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.24	
		a*	0.24	
		b*	0.21	
6. Yellow		L*	0.33	
		a*	0.16	
		b*	0.41	
7. Green		L*	0.24	
		a*	0.13	
		b*	0.08	
8. Cyan		L*	0.24	
		a*	0.13	
		b*	0.13	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles				
Included Reflectance Std. Light Source C (2 °)	70304			Color Standard Tiles/ SICT-CP-70304
1. White		x	0.000 7	
		y	0.000 7	
2. I,Gray		x	0.000 7	
		y	0.000 7	
3. M,Gray		x	0.000 7	
		y	0.000 7	
4. D,Gray		x	0.000 7	
		y	0.000 7	
5. Red		x	0.001 4	
		y	0.000 7	
6. Yellow		x	0.000 8	
		y	0.000 9	
7. Green		x	0.000 7	
		y	0.000 7	
8. Cyan		x	0.000 7	
		y	0.000 7	
Included Reflectance Std. Light Source C (10 °)				
1. White		X	0.82	
		Y	0.85	
		Z	0.95	
2. L,Gray		X	0.57	
		Y	0.59	
		Z	0.68	
3. M,Gray		X	0.27	
		Y	0.28	
		Z	0.32	
4. D,Gray		X	0.10	
		Y	0.10	
		Z	0.11	
5. Red		X	0.24	
		Y	0.18	
		Z	0.14	
6. Yellow		X	0.58	
		Y	0.60	
		Z	0.20	
7. Green		X	0.18	
		Y	0.24	
		Z	0.19	
8. Cyan		X	0.20	
		Y	0.24	
		Z	0.48	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Included Reflectance Std. Light Source C (10°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White		L*	0.37	
		a*	0.09	
		b*	0.07	
2. L,Gray		L*	0.32	
		a*	0.07	
		b*	0.07	
3. M,Gray		L*	0.25	
		a*	0.06	
		b*	0.05	
4. D,Gray		L*	0.17	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.24	
		a*	0.23	
		b*	0.20	
6. Yellow		L*	0.32	
		a*	0.16	
		b*	0.42	
7. Green		L*	0.24	
		a*	0.12	
		b*	0.12	
8. Cyan		L*	0.24	
		a*	0.11	
		b*	0.15	
Included Reflectance Std. Light Source C (10°)				
1. White		x	0.000 7	
		y	0.000 7	
2. I,Gray		x	0.000 7	
		y	0.000 7	
3. M,Gray		x	0.000 7	
		y	0.000 7	
4. D,Gray		x	0.000 7	
		y	0.000 7	
5. Red		x	0.001 4	
		y	0.000 7	
6. Yellow		x	0.000 8	
		y	0.000 9	
7. Green		x	0.000 7	
		y	0.000 7	
8. Cyan		x	0.000 7	
		y	0.000 7	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Included Reflectance Std. Light Source D65 (2°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White		X	0.81	
		Y	0.85	
		Z	0.89	
2. L,Gray		X	0.56	
		Y	0.59	
		Z	0.64	
3. M,Gray		X	0.26	
		Y	0.28	
		Z	0.30	
4. D,Gray		X	0.09	
		Y	0.10	
		Z	0.10	
5. Red		X	0.25	
		Y	0.18	
		Z	0.14	
6. Yellow		X	0.58	
		Y	0.63	
		Z	0.20	
7. Green		X	0.17	
		Y	0.25	
		Z	0.20	
8. Cyan		X	0.19	
		Y	0.23	
		Z	0.46	
Included Reflectance Std. Light Source D65 (2°)				
1. White		L*	0.37	
		a*	0.07	
		b*	0.07	
2. L,Gray		L*	0.32	
		a*	0.07	
		b*	0.06	
3. M,Gray		L*	0.25	
		a*	0.06	
		b*	0.05	
4. D,Gray		L*	0.17	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.24	
		a*	0.25	
		b*	0.21	
6. Yellow		L*	0.33	
		a*	0.16	
		b*	0.40	
7. Green		L*	0.24	
		a*	0.13	
		b*	0.08	
8. Cyan		L*	0.24	
		a*	0.14	
		b*	0.13	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles				
Included Reflectance Std. Light Source	70304			Color Standard Tiles/ SICT-CP-70304
D65 (2 °)				
1. White		x	0.000 7	
		y	0.000 7	
2. I,Gray		x	0.000 7	
		y	0.000 7	
3. M,Gray		x	0.000 7	
		y	0.000 7	
4. D,Gray		x	0.000 7	
		y	0.000 7	
5. Red		x	0.001 4	
		y	0.000 7	
6. Yellow		x	0.000 8	
		y	0.000 8	
7. Green		x	0.000 7	
		y	0.000 7	
8. Cyan		x	0.000 7	
		y	0.000 7	
Included Reflectance Std. Light Source				
D65 (10 °)				
1. White		X	0.80	
		Y	0.85	
		Z	0.88	
2. L,Gray		X	0.56	
		Y	0.59	
		Z	0.63	
3. M,Gray		X	0.26	
		Y	0.28	
		Z	0.29	
4. D,Gray		X	0.09	
		Y	0.10	
		Z	0.10	
5. Red		X	0.24	
		Y	0.18	
		Z	0.13	
6. Yellow		X	0.58	
		Y	0.60	
		Z	0.19	
7. Green		X	0.18	
		Y	0.25	
		Z	0.18	
8. Cyan		X	0.19	
		Y	0.24	
		Z	0.45	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Included Reflectance Std. Light Source D65 (10 °)	70304			Color Standard Tiles/ SICT-CP-70304
1. White		L* a* b*	0.37 0.07 0.07	
2. L,Gray		L* a* b*	0.32 0.06 0.06	
3. M,Gray		L* a* b*	0.25 0.06 0.05	
4. D,Gray		L* a* b*	0.17 0.04 0.04	
5. Red		L* a* b*	0.24 0.24 0.20	
6. Yellow		L* a* b*	0.32 0.16 0.41	
7. Green		L* a* b*	0.24 0.12 0.08	
8. Cyan		L* a* b*	0.24 0.12 0.12	
Included Reflectance Std. Light Source D65 (10 °)				
1. White		x y	0.000 7 0.000 7	
2. I,Gray		x y	0.000 7 0.000 7	
3. M,Gray		x y	0.000 7 0.000 7	
4. D,Gray		x y	0.000 7 0.000 7	
5. Red		x y	0.001 4 0.000 7	
6. Yellow		x y	0.000 8 0.000 8	
7. Green		x y	0.000 7 0.000 7	
8. Cyan		x y	0.000 7 0.000 7	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles	70304			Color Standard Tiles/ SICT-CP-70304
Exclude Reflectance Std. Light Source Type A (2°)				
1. White		X	0.90	
		Y	0.81	
		Z	0.29	
2. L,Gray		X	0.60	
		Y	0.55	
		Z	0.20	
3. M,Gray		X	0.26	
		Y	0.24	
		Z	0.10	
4. D,Gray		X	0.07	
		Y	0.06	
		Z	0.03	
5. Red		X	0.31	
		Y	0.2	
		Z	0.1	
6. Yellow		X	0.73	
		Y	0.64	
		Z	0.07	
7. Green		X	0.15	
		Y	0.19	
		Z	0.07	
8. Cyan		X	0.13	
		Y	0.16	
		Z	0.14	
Exclude Reflectance Std. Light Source Type A (2°)				
1. White		L*	0.36	
		a*	0.07	
		b*	0.07	
2. L,Gray		L*	0.31	
		a*	0.06	
		b*	0.06	
3. M,Gray		L*	0.24	
		a*	0.05	
		b*	0.05	
4. D,Gray		L*	0.14	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.25	
		a*	0.30	
		b*	0.70	
6. Yellow		L*	0.33	
		a*	0.12	
		b*	0.45	
7. Green		L*	0.22	
		a*	0.12	
		b*	0.08	
8. Cyan		L*	0.22	
		a*	0.16	
		b*	0.15	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles				Color Standard Tiles/ SICT-CP-70304
Exclude Reflectance Std. Light Source Type A (2 °)	70304			
1. White		x	0.000 8	
		y	0.000 7	
2. I,Gray		x	0.000 8	
		y	0.000 7	
3. M,Gray		x	0.000 8	
		y	0.000 7	
4. D,Gray		x	0.000 8	
		y	0.000 7	
5. Red		x	0.001 3	
		y	0.000 7	
6. Yellow		x	0.000 8	
		y	0.000 8	
7. Green		x	0.000 7	
		y	0.000 8	
8. Cyan		x	0.000 7	
		y	0.000 7	
Exclude Reflectance Std. Light Source Type A (10 °)				
1. White		X	0.91	
		Y	0.81	
		Z	0.28	
2. L,Gray		X	0.61	
		Y	0.55	
		Z	0.20	
3. M,Gray		X	0.26	
		Y	0.24	
		Z	0.09	
4. D,Gray		X	0.07	
		Y	0.06	
		Z	0.03	
5. Red		X	0.30	
		Y	0.20	
		Z	0.10	
6. Yellow		X	0.74	
		Y	0.62	
		Z	0.07	
7. Green		X	0.15	
		Y	0.19	
		Z	0.06	
8. Cyan		X	0.13	
		Y	0.17	
		Z	0.14	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Excluded Reflectance Std. Light Source A (10°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White		L*	0.36	
		a*	0.06	
		b*	0.07	
2. L,Gray		L*	0.31	
		a*	0.05	
		b*	0.06	
3. M,Gray		L*	0.24	
		a*	0.05	
		b*	0.05	
4. D,Gray		L*	0.14	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.25	
		a*	0.30	
		b*	0.70	
6. Yellow		L*	0.33	
		a*	0.12	
		b*	0.49	
7. Green		L*	0.22	
		a*	0.12	
		b*	0.08	
8. Cyan		L*	0.22	
		a*	0.14	
		b*	0.15	
Exclude Reflectance Std. Light Source Type A (10°)				
1. White		x	0.000 6	
		y	0.000 6	
2. I,Gray		x	0.000 6	
		y	0.000 6	
3. M,Gray		x	0.000 6	
		y	0.000 6	
4. D,Gray		x	0.000 6	
		y	0.000 6	
5. Red		x	0.001 1	
		y	0.000 6	
6. Yellow		x	0.000 6	
		y	0.000 6	
7. Green		x	0.000 6	
		y	0.000 6	
8. Cyan		x	0.000 6	
		y	0.000 6	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Excluded Reflectance Std. Light Source C (2 °)	70304			Color Standard Tiles/ SICT-CP-70304
1. White		X	0.79	
		Y	0.81	
		Z	0.92	
2. L,Gray		X	0.54	
		Y	0.55	
		Z	0.65	
3. M,Gray		X	0.23	
		Y	0.24	
		Z	0.28	
4. D,Gray		X	0.06	
		Y	0.06	
		Z	0.07	
5. Red		X	0.21	
		Y	0.20	
		Z	0.20	
6. Yellow		X	0.56	
		Y	0.60	
		Z	0.19	
7. Green		X	0.14	
		Y	0.20	
		Z	0.16	
8. Cyan		X	0.16	
		Y	0.19	
		Z	0.45	
Exclude Reflectance Std. Light Source Type C (2 °)				
1. White		L*	0.36	
		a*	0.09	
		b*	0.07	
2. L,Gray		L*	0.31	
		a*	0.07	
		b*	0.06	
3. M,Gray		L*	0.24	
		a*	0.06	
		b*	0.05	
4. D,Gray		L*	0.14	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.25	
		a*	0.40	
		b*	0.60	
6. Yellow		L*	0.32	
		a*	0.17	
		b*	0.52	
7. Green		L*	0.23	
		a*	0.14	
		b*	0.08	
8. Cyan		L*	0.23	
		a*	0.14	
		b*	0.14	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Excluded Reflectance Std. Light Source C (2 °)	70304			Color Standard Tiles/ SICT-CP-70304
1. White		x	0.000 7	
		y	0.000 7	
2. I,Gray		x	0.000 7	
		y	0.000 7	
3. M,Gray		x	0.000 7	
		y	0.000 7	
4. D,Gray		x	0.000 7	
		y	0.000 7	
5. Red		x	0.002 8	
		y	0.000 7	
6. Yellow		x	0.000 9	
		y	0.000 9	
7. Green		x	0.000 7	
		y	0.000 7	
8. Cyan		x	0.000 7	
		y	0.000 7	
Exclude Reflectance Std. Light Source Type C (10 °)				
1. White		X	0.79	
		Y	0.81	
		Z	0.90	
2. L,Gray		X	0.54	
		Y	0.55	
		Z	0.63	
3. M,Gray		X	0.23	
		Y	0.24	
		Z	0.27	
4. D,Gray		X	0.06	
		Y	0.06	
		Z	0.07	
5. Red		X	0.20	
		Y	0.20	
		Z	0.20	
6. Yellow		X	0.55	
		Y	0.56	
		Z	0.17	
7. Green		X	0.14	
		Y	0.21	
		Z	0.14	
8. Cyan		X	0.17	
		Y	0.20	
		Z	0.44	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Excluded Reflectance Std. Light Source C (10°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White		L*	0.36	
		a*	0.08	
		b*	0.07	
2. L,Gray		L*	0.31	
		a*	0.07	
		b*	0.06	
3. M,Gray		L*	0.24	
		a*	0.06	
		b*	0.05	
4. D,Gray		L*	0.14	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.30	
		a*	0.40	
		b*	0.60	
6. Yellow		L*	0.32	
		a*	0.17	
		b*	0.55	
7. Green		L*	0.23	
		a*	0.13	
		b*	0.14	
8. Cyan		L*	0.23	
		a*	0.12	
		b*	0.16	
Exclude Reflectance Std. Light Source Type C (10°)				
1. White		x	0.000 7	
		y	0.000 7	
2. I,Gray		x	0.000 7	
		y	0.000 7	
3. M,Gray		x	0.000 7	
		y	0.000 7	
4. D,Gray		x	0.000 7	
		y	0.000 7	
5. Red		x	0.002 9	
		y	0.000 7	
6. Yellow		x	0.000 8	
		y	0.001 0	
7. Green		x	0.000 7	
		y	0.000 7	
8. Cyan		x	0.000 7	
		y	0.000 7	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Excluded Reflectance Std. Light Source	70304			Color Standard Tiles/ SICT-CP-70304
D65 (2 °)				
1. White		X	0.77	
		Y	0.81	
		Z	0.85	
2. L,Gray		X	0.52	
		Y	0.55	
		Z	0.60	
3. M,Gray		X	0.23	
		Y	0.24	
		Z	0.26	
4. D,Gray		X	0.06	
		Y	0.06	
		Z	0.06	
5. Red		X	0.20	
		Y	0.20	
		Z	0.20	
6. Yellow		X	0.55	
		Y	0.60	
		Z	0.18	
7. Green		X	0.14	
		Y	0.21	
		Z	0.15	
8. Cyan		X	0.16	
		Y	0.19	
		Z	0.41	
Exclude Reflectance Std. Light Source Type				
D65 (2 °)				
1. White		L*	0.36	
		a*	0.07	
		b*	0.07	
2. L,Gray		L*	0.31	
		a*	0.07	
		b*	0.06	
3. M,Gray		L*	0.24	
		a*	0.06	
		b*	0.05	
4. D,Gray		L*	0.14	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.30	
		a*	0.40	
		b*	0.60	
6. Yellow		L*	0.32	
		a*	0.17	
		b*	0.51	
7. Green		L*	0.23	
		a*	0.14	
		b*	0.08	
8. Cyan		L*	0.23	
		a*	0.15	
		b*	0.14	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles				
Excluded Reflectance Std. Light Source	70304			Color Standard Tiles/ SICT-CP-70304
D65 (2 °)				
1. White		x	0.000 7	
		y	0.000 7	
2. I,Gray		x	0.000 7	
		y	0.000 7	
3. M,Gray		x	0.000 7	
		y	0.000 7	
4. D,Gray		x	0.000 7	
		y	0.000 7	
5. Red		x	0.002 7	
		y	0.000 7	
6. Yellow		x	0.000 8	
		y	0.000 9	
7. Green		x	0.000 7	
		y	0.000 8	
8. Cyan		x	0.000 7	
		y	0.000 7	
Exclude Reflectance Std. Light Source Type				
D65 (10 °)				
1. White		X	0.77	
		Y	0.81	
		Z	0.84	
2. L,Gray		X	0.52	
		Y	0.55	
		Z	0.59	
3. M,Gray		X	0.23	
		Y	0.24	
		Z	0.25	
4. D,Gray		X	0.06	
		Y	0.06	
		Z	0.06	
5. Red		X	0.20	
		Y	0.20	
		Z	0.20	
6. Yellow		X	0.54	
		Y	0.57	
		Z	0.17	
7. Green		X	0.14	
		Y	0.21	
		Z	0.14	
8. Cyan		X	0.16	
		Y	0.20	
		Z	0.41	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Excluded Reflectance Std. Light Source D65 (10 °)	70304			Color Standard Tiles/ SICT-CP-70304
1. White		L* a* b*	0.36 0.07 0.07	
2. L,Gray		L* a* b*	0.31 0.06 0.06	
3. M,Gray		L* a* b*	0.24 0.05 0.05	
4. D,Gray		L* a* b*	0.14 0.04 0.04	
5. Red		L* a* b*	0.30 0.40 0.60	
6. Yellow		L* a* b*	0.32 0.17 0.55	
7. Green		L* a* b*	0.23 0.13 0.09	
8. Cyan		L* a* b*	0.23 0.12 0.13	
Exclude Reflectance Std. Light Source Type D65 (10 °)				
1. White	x y		0.000 7 0.000 7	
2. I,Gray	x y		0.000 7 0.000 7	
3. M,Gray	x y		0.000 7 0.000 7	
4. D,Gray	x y		0.000 7 0.000 7	
5. Red	x y		0.002 8 0.000 7	
6. Yellow	x y		0.000 9 0.000 9	
7. Green	x y		0.000 7 0.000 8	
8. Cyan	x y		0.000 7 0.000 7	
Absolute Spectral Reflectance White Plate (Include, Exclude Reflectance)		(360 ~ 830) nm	0.007 6	Absolute Spectral Reflectance White Standard Plates/ SICT-CP-70304

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Gloss meters	70306	20 ° 60 ° 85 °	8.9×10^{-3} 9.5×10^{-3} 8.0×10^{-3}	Gloss Standard/ SICT-CP-70306
Gloss standard plates	70307	20° 60° 85°	9.5×10^{-3} 9.8×10^{-3} 8.3×10^{-3}	Gloss Meter/ SICT-CP-70307
Haze meters	70308	Haze H-1 H-5 H-10 H-20 H-30 Transmittance T-30 T-50 T-70 T-90	0.30 0.26 0.4 0.6 0.8 0.50 0.50 0.50 0.50	Haze Standard Plate, Transmittance Standard Plates/ SICT-CP-70308
Lens meters	70312	Vertex diopter -25 D ~ 25 D	0.03 D	Reference Lens/ SICT-CP-70312
Optical densitymeters	70315	Density 1 Step ~ 10 Step 11 Step 11 Step ~ 15 Step	0.03 0.06 0.11	Density CRM/ SICT-CP-70315
Reflectance meters	70319	Reflectance 380 nm ~ 780 nm	1.1×10^{-2}	Absolute Spectral Reflectance White Standard Plates/ SICT-CP-70319
Refractometers	70321	Refracto (1.332 99 ~ 1.505 80) nD 1.51 nD 1.62 nD	0.000 04 nD 0.000 2 nD 0.000 2 nD	Reference Refracto CRM/ SICT-CP-70321
Transmittance meters	70323	(0.1) (250 ~ 750) nm (0.5) (250 ~ 750) nm (0.9) (250 ~ 750) nm	6.1×10^{-3} 3.8×10^{-3} 2.2×10^{-3}	Transmittance Filter/ SICT-CP-70323

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Spectrophotometers including FT-IR spectrophotometers	70325			Wavelength Filters, Transmittance Filters, Absolute Spectral Reflectance White Standard Plates, Wavenumber r:1+---/ SICT-CP-70325
Wavelength		(240 ~ 750) nm	0.4 nm	
Transmittance		(0.1 ~ 0.3)		
		250 nm	8.1×10^{-3}	
		300 nm	7.8×10^{-3}	
		350 nm	7.7×10^{-3}	
		400 nm	5.4×10^{-3}	
		450 nm	5.2×10^{-3}	
		500 nm	5.2×10^{-3}	
		550 nm	5.2×10^{-3}	
		600 nm	5.2×10^{-3}	
		650 nm	5.2×10^{-3}	
		700 nm	5.2×10^{-3}	
		750 nm	5.2×10^{-3}	
		(0.3 ~ 0.5)		
		250 nm	8.1×10^{-3}	
		300 nm	7.8×10^{-3}	
		350 nm	7.7×10^{-3}	
		400 nm	5.2×10^{-3}	
		450 nm	5.2×10^{-3}	
		500 nm	5.2×10^{-3}	
		550 nm	5.1×10^{-3}	
		600 nm	5.2×10^{-3}	
		650 nm	5.2×10^{-3}	
		700 nm	5.2×10^{-3}	
		750 nm	5.2×10^{-3}	
		(0.5 ~ 0.9)		
		250 nm	7.9×10^{-3}	
		300 nm	8.2×10^{-3}	
		350 nm	7.8×10^{-3}	
		400 nm	5.2×10^{-3}	
		450 nm	5.2×10^{-3}	
		500 nm	5.2×10^{-3}	
		550 nm	5.1×10^{-3}	
		600 nm	5.1×10^{-3}	
		650 nm	5.1×10^{-3}	
		700 nm	5.2×10^{-3}	
		750 nm	5.2×10^{-3}	
		(0.01)		
		440 nm	1.3×10^{-2}	
		465 nm	8.4×10^{-3}	
		546 nm	8.9×10^{-3}	
		590 nm	1.0×10^{-2}	
		635 nm	8.1×10^{-3}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Wavelength reference materials; absorption cell, bandpass filter, etc.	70326			Spectrophotometers, Absolute Spectral Reflectance White Standard Plates/ SICT-CP-70326
Transmittance		(0.3 ~ 0.5)		
		250 nm	8.3×10^{-3}	
		300 nm	8.1×10^{-3}	
		350 nm	8.0×10^{-3}	
		400 nm	5.7×10^{-3}	
		450 nm	5.7×10^{-3}	
		500 nm	5.7×10^{-3}	
		550 nm	5.7×10^{-3}	
		600 nm	5.7×10^{-3}	
		650 nm	5.7×10^{-3}	
		700 nm	5.7×10^{-3}	
		750 nm	5.7×10^{-3}	
		(0.5 ~ 0.9)		
		250 nm	8.2×10^{-3}	
		300 nm	8.1×10^{-3}	
		350 nm	8.0×10^{-3}	
		400 nm	5.7×10^{-3}	
		450 nm	5.7×10^{-3}	
		500 nm	5.7×10^{-3}	
		550 nm	5.7×10^{-3}	
		600 nm	5.6×10^{-3}	
		650 nm	5.6×10^{-3}	
		700 nm	5.7×10^{-3}	
		750 nm	5.9×10^{-3}	
Absorbance		(0.1 ~ 0.3)		
		250 nm	0.003 7	
		300 nm	0.003 7	
		350 nm	0.003 8	
		400 nm	0.002 7	
		450 nm	0.002 7	
		500 nm	0.002 8	
		550 nm	0.002 8	
		600 nm	0.002 7	
		650 nm	0.002 7	
		700 nm	0.002 7	
		750 nm	0.002 8	
		(0.3 ~ 0.5)		
		250 nm	0.003 6	
		300 nm	0.003 6	
		350 nm	0.003 6	
		400 nm	0.002 4	
		450 nm	0.002 4	
		500 nm	0.002 4	
		550 nm	0.002 4	
		600 nm	0.002 4	
		650 nm	0.002 4	
		700 nm	0.002 4	
		750 nm	0.002 5	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Wavelength reference materials; absorption cell, bandpass filter, etc.	70326			Spectrophotometers, Absolute Spectral Reflectance White Standard Plates/ SICT-CP-70326
Absorbance		(0.5 ~ 0.9)		
		250 nm	0.003 6	
		300 nm	0.003 5	
		350 nm	0.003 5	
		400 nm	0.002 4	
		450 nm	0.002 4	
		500 nm	0.002 4	
		550 nm	0.002 4	
		600 nm	0.002 4	
		650 nm	0.002 4	
		700 nm	0.002 4	
		750 nm	0.002 4	
Reflectance		(360 ~ 830) nm	1.0×10^{-2}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Broadband light sources Wavelength output	70402	1 310 nm, 1 550 nm	0.058 nm	Optical spectrum analyzer, Optical powermeter/ SICT-CP-70402
Optical power output		1 310 nm, 1 550 nm (0 ~ -60) dBm	0.070 dB	
Optical attenuators Optical Attenuation	70410	1 310 nm, 1 550 nm (-60 ~ 0) dB	0.08 dB	Optical powermeter, Optical power stabilized lasers and LDs/ SICT-CP-70410
Absolute optical power		1 310 nm, 1 550 nm (0 ~ -60) dBm	0.072 dB	
Fiber-optic power meters Optical Linearity	70412	1 310 nm, 1 550 nm (0 ~ -60) dB	0.03 dB	Optical powermeter, Optical power stabilized lasers and LDs, Optical attenuator/ SICT-CP-70412
Optical Attenuation		1 310 nm, 1 550 nm (0 ~ -60) dB	0.03 dB	
Optical loss testers Optical Attenuation	70413	1 310 nm, 1 550 nm (0 ~ -60) dB	0.03 dB	Optical attenuator/ SICT-CP-70413
Absolute optical power measure		1 310 nm, 1 550 nm (0 ~ -60) dBm	0.072 dB	
Optical multimeters Linearity measure	70415	1 310 nm, 1 550 nm (0 ~ -60) dB	0.03 dB	Optical powermeter, Optical power stabilized lasers and LDs, Optical attenuator/ SICT-CP-70415
Absolute optical power measure		1 310 nm, 1 550 nm (0 ~ -60) dBm	0.072 dB	
Optical network analyzer (Optical multimeter)	70416	1 310 nm, 1 550 nm (-60 ~ 0) dBm	0.072 dB	Optical powermeter, OTDR, Fiber reference, Wavelength meter Optical spectrum analyzer Optical attenuator Optical Returnloss generator/ SICT-CP-70416
Absolute optical power		1 310 nm 1 550 nm	0.058 nm 0.058 nm	
(Optical spectrum analyzer) Wavelength measure		Resolution: (0.1 ~ 1) nm		
Wavelength measure		1 310 nm 1 550 nm	0.058 nm 0.058 nm	
Absolute optical power measure		1 310 nm, 1 550 nm (-60 ~ 0) dBm	0.072 dB	
Absolute optical power measure		1 310 nm, 1 550 nm (-60 ~ 0) dBm	0.072 dB	
(Optical attenuator) Optical Attenuation		1 310 nm, 1 550 nm (-60 ~ 0) dB	0.07 dB	
Optical Attenuation		1 310 nm, 1 550 nm (-60 ~ 0) dB	0.07 dB	
Return loss (Optical time domain reflectometer)		1 310 nm, 1 550 nm (20 ~ 40) dB	0.8 dB	
Return loss		1 310 nm, 1 550 nm (20 ~ 40) dB	0.8 dB	
Wavelength output Optical Length measure		1 310 nm, 1 550 nm	0.082 nm	
Optical Length measure		1 310 nm 3.3 km Fiber 13.4 km Fiber	0.081 m 0.34 m	
Optical Length measure		1 550 nm 3.3 km Fiber 13.4 km Fiber	0.080 m 0.34 m	
Optical Length measure		1 550 nm 3.3 km Fiber 13.4 km Fiber	0.080 m 0.34 m	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Optical network analyzer Optical loss measure	70416	1 310 nm 7.20 dB Fiber 2.90 dB Fiber 1 550 nm 4.20 dB Fiber 1.60 dB Fiber	0.13 dB 0.05 dB 0.05 dB 0.05 dB	Optical powermeter, OTDR, Fiber reference, Wavelength meter Optical spectrum analyzer Optical attenuator Optical Returnloss generator/ SICT-CP-70416
Optical spectrum analyzers Wavelength measure	70417	1 310 nm 1 550 nm 분해능: (0.1 ~ 1) nm	0.058 nm 0.058 nm	Optical powermeter, Optical power stabilized lasers and LDs, Optical attenuator, Optical spectrum analyzer/ SICT-CP-70417
Resolution measure		1 310 nm 1 550 nm	0.058 nm 0.058 nm	
Absolute optical power measure		1 310 nm, 1 550 nm (-60 ~ 0) dBm	0.072 dB	
Linearity measure		1 310 nm, 1 550 nm (-60 ~ 0) dB	0.03 dB	
Optical time domain reflectometers, OTDR	70418			Optical length fiber reference, Optical fiberloss reference, Optical spectrum analyzer/ SICT-CP-70418
Wavelength output		1 310 nm, 1 550 nm	0.08 nm	
Optical Length measure		1 310 nm 3.3 km Fiber 13.4 km Fiber 1 550 nm 3.3 km Fiber 13.4 km Fiber	0.081 m 0.34 m 0.080 m 0.34 m	
Optical loss measure		1 310 nm 7.20 dB Fiber 2.90 dB Fiber 1 550 nm 4.20 dB Fiber 1.60 dB Fiber	0.13 dB 0.05 dB 0.05 dB 0.05 dB	
Return loss meters Return loss measure	70423	1 310 nm, 1 550 nm 20 dB ~ 40 dB	0.8 dB	Optical Returnloss generator SICT-CP-70423
Frequency stabilized lasers ands LDs Wavelength	70429	1 310 nm 1 550 nm	4 pm 4 pm	Wavelength meter, Optical powermeter/ SICT-CP-70429
optical power		1 310 nm, 1 550 nm (-60 ~ 0) dBm	0.07 dB	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
ASE light sources	70430	Wavelength output	1 310 nm, 1 550 nm	Optical spectrum analyzer, Optical powermeter/ SICT-CP-70430
		Optical power output	1 310 nm, 1 550 nm (-60 ~ 0) dBm	0.058 nm 0.07 dB
Optical power stabilized lasers and LDs	70433	Wavelength output	1 310 nm 1 550 nm	Wavelength meter, Optical powermeter/ SICT-CP-70433
		Optical power output	1 310 nm, 1 550 nm (-60 ~ 0) dBm	4 pm 4 pm 0.07 dB

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Breath alcohol analyzers				
Dry process	90101	(0.000 ~ 0.080) %BAC (0.080 ~ 0.190) %BAC	3.3×10^{-2} 2.1×10^{-2}	Standard gas/ SICT-CP-90101
Wet process		(0.000 ~ 0.080) %BAC (0.080 ~ 0.150) %BAC (0.150 ~ 0.400) %BAC	2.9×10^{-2} 1.6×10^{-2} 1.3×10^{-2}	
Environmental air monitoring instruments	90102			Standard gas/ SICT-CP-90102
Oxygen		(0 ~ 700) $\mu\text{mol/mol}$ (0.07 ~ 1.5) cmol/mol (1.5 ~ 20) cmol/mol	1.0×10^{-2} 2.0×10^{-2} 1.2×10^{-2}	
Carbon monoxide		(0 ~ 850) $\mu\text{mol/mol}$	2.2×10^{-2}	
Carbon dioxide		(0 ~ 0.50) cmol/mol (0.50 ~ 5.00) cmol/mol (5.00 ~ 19.00) cmol/mol	2.0×10^{-2} 1.5×10^{-2} 2.1×10^{-2}	
Nitrogen monoxide		(0 ~ 850) $\mu\text{mol/mol}$	2.1×10^{-2}	
Isobutane		(0 ~ 0.8) cmol/mol	2.2×10^{-2}	
Methane		(0 ~ 2.0) cmol/mol	1.4×10^{-2}	
Hydrogen sulfide		(0 ~ 45) $\mu\text{mol/mol}$	3.6×10^{-2}	
Propane		(0 ~ 2 000) $\mu\text{mol/mol}$	3.0×10^{-2}	
Isobutylene		(0 ~ 25) $\mu\text{mol/mol}$	1.0×10^{-2}	
Ammonia		(0 ~ 50) $\mu\text{mol/mol}$	4.9×10^{-2}	
Sulfur dioxide		(0 ~ 850) $\mu\text{mol/mol}$	2.2×10^{-2}	
Nitrogen dioxide		(0 ~ 1 000) $\mu\text{mol/mol}$	1.0×10^{-2}	
Hydrogen		(0 ~ 500) $\mu\text{mol/mol}$ (0.05 ~ 2.0) cmol/mol	2.3×10^{-2} 2.1×10^{-2}	
Hydrogen chloride		(0 ~ 50) $\mu\text{mol/mol}$	4.8×10^{-2}	
Sulfur hexafluoride		(0 ~ 100) cmol/mol	0.1×10^{-2}	
Ozone		0.0 nmol/mol (0.0 ~ 1 000.0) nmol/mol	2.2 nmol/mol 2.5×10^{-2}	
Gas analyzers	90103			Standard gas/ SICT-CP-90103
Oxygen		(0 ~ 700) $\mu\text{mol/mol}$ (0.07 ~ 1.5) cmol/mol (1.5 ~ 20) cmol/mol	1.0×10^{-2} 2.0×10^{-2} 1.2×10^{-2}	
Carbon monoxide		(0 ~ 850) $\mu\text{mol/mol}$	2.2×10^{-2}	
Carbon dioxide		(0 ~ 0.50) cmol/mol (0.50 ~ 5.00) cmol/mol (5.00 ~ 19.00) cmol/mol	2.0×10^{-2} 1.5×10^{-2} 2.1×10^{-2}	
Nitrogen monoxide		(0 ~ 850) $\mu\text{mol/mol}$	2.1×10^{-2}	
Isobutane		(0 ~ 0.8) cmol/mol	2.2×10^{-2}	
Methane		(0 ~ 2.0) cmol/mol	1.4×10^{-2}	
Hydrogen sulfide		(0 ~ 45) $\mu\text{mol/mol}$	3.6×10^{-2}	
Propane		(0 ~ 2 000) $\mu\text{mol/mol}$	3.0×10^{-2}	
Isobutylene		(0 ~ 25) $\mu\text{mol/mol}$	1.0×10^{-2}	
Ammonia		(0 ~ 50) $\mu\text{mol/mol}$	4.9×10^{-2}	
Sulfur dioxide		(0 ~ 850) $\mu\text{mol/mol}$	2.2×10^{-2}	
Nitrogen dioxide		(0 ~ 1 000) $\mu\text{mol/mol}$	1.0×10^{-2}	
Hydrogen		(0 ~ 500) $\mu\text{mol/mol}$ (0.05 ~ 2.0) cmol/mol	2.3×10^{-2} 2.1×10^{-2}	
Hydrogen chloride		(0 ~ 50) $\mu\text{mol/mol}$	4.8×10^{-2}	
Sulfur hexafluoride		(0 ~ 100) cmol/mol	0.1×10^{-2}	
Ozone		0.0 nmol/mol (0.0 ~ 1 000.0) nmol/mol	2.2 nmol/mol 2.5×10^{-2}	

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Exhaust gas test instruments	90104			
Oxygen		(0 ~ 1.5) $\mu\text{mol/mol}$ (1.5 ~ 20) cmol/mol	2.0×10^{-2} 1.1×10^{-2}	Standard gas/ SICT-CP-90103
Carbon monoxide		(0 ~ 5.0) cmol/mol	2.1×10^{-2}	
Carbon dioxide		(0 ~ 19) cmol/mol	2.0×10^{-2}	
Nitrogen monoxide		(0 ~ 2 000) $\mu\text{mol/mol}$	2.0×10^{-2}	
Isobutane		(0 ~ 0.8) cmol/mol	2.2×10^{-2}	
Methane		(0 ~ 2.0) cmol/mol	1.4×10^{-2}	
Propane		(0 ~ 2 000) $\mu\text{mol/mol}$	3.0×10^{-2}	
Ammonia		(0 ~ 50) $\mu\text{mol/mol}$	4.9×10^{-2}	
Sulfur dioxide		(0 ~ 850) $\mu\text{mol/mol}$	2.2×10^{-2}	
Nitrogen dioxide		(0 ~ 1 000) $\mu\text{mol/mol}$	1.0×10^{-2}	
Hydrogen		(0 ~ 500) $\mu\text{mol/mol}$ (0.05 ~ 2.0) cmol/mol	2.3×10^{-2} 2.1×10^{-2}	
Others; pH meter, Electrical conductivity meter	90104			
pH meter		(4 ~ 10) pH	0.013 pH	CRM/ SICT-CP-90199
Electrical conductivity meter		100 $\mu\text{S/cm}$ 1 413 $\mu\text{S/cm}$ 12.85 mS/cm 111.3 mS/cm	3.1 $\mu\text{S/cm}$ 9.7 $\mu\text{S/cm}$ 0.073 mS/cm 0.78 mS/cm	

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017 & KS Q ISO/IEC 17025:2017

Institute of Calibration Technology Co.,Ltd

47-22, Jingoksandan 2beon-ro, Gwangsan-gu, Gwangju, Republic of Korea

Phone : 82-62-953-5353, Fax : 82-62-953-5454 e-mail : sictadmin@sict.co.kr

CALIBRATION

Valid To : Oct. 29. 2025

Accreditation No : KC01-018(1/35)

In recognition of the successful completion of the KOLAS evaluation process,
accreditation is granted to this laboratory to perform the following calibrations

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site
102.	Linear dimension		201.	Mass		40424	Volt/Current recorders	Y
10211	Filler gauges	Y	20109	Electric balances	Y	50101	Contact thermometry	
10216	Height gauges/measuring machines	Y	20112	Platform scale balances	Y		Temperature generators: ovens, furnaces, isothermal liquid baths, ice-point baths, dry-block calibrators	Y
			20116	Weights	Y			
10228	Cylindrical plug/pingauges, Threadmeasuring wire gauges	Y	401.	DC Voltage & current		50102	Temperature indicators /recorders/controllers, temperature calibrators	Y
			40101	DC ammeters	Y			
10234	Ultrasonic thickness gauges	Y	40103	DC voltage/current calibrators	Y	50103	Glass thermometers; liquid-in-glass, Beckmann	N
10236	Coating thickness testers	Y	40104	Electrical temperature calibrators	Y	50104	Resistance thermometers; SPRT, IPRT, thermistors, etc.	Y
104.	Form					50105	Thermal expansion thermometers ; bimetal, gas or liquid type	Y
10401	Form testers	Y		Inductance				
10407	Precision surface plates	Y	40205	Earth testers	Y			
10409	Roundness measurement instruments	N	40108	DC power supplies	Y	50107	Temperature transducers	Y
			40112	DC voltmeters	Y			
10412	Straight edges	Y	40210	Insulation testers	Y	503.	Humidity	
105.	Complex geometry		40214	Resistance meters	Y	50302	Relative humidity hygrometers polymer thin film, hair, etc.	Y
10503	Contact coordinate measuring machines	Y	40215	Resistors	Y	50304	Temperature humidity recorders ; Hygrothermograph, etc	N
10504	Non-contact coordinate measuring machines	Y	403.	AC voltage, current & power		50305	Transducers; dew-point /relative humidity	N
10511	Measuring microscopes, Profile projectors	Y	40205	AC ammeters	Y			
10512	Microscopes, micro measuring	Y	40302	Clamp ammeters/voltmeters	Y			
10517	Stylus type roughness testers	Y	40303	AC voltage/current calibrators	Y			
106.	Various dimensional		40310	Power factor meters	Y			
10601	Inside/Outside/Gear tooth calipers, Caliper gauges	Y	40311	AC power meters	Y	50306	Humidity generators; two-pressure, two-temperature, flow mixing humidity generator, constant temperature and humidity chamber, etc.	Y
10603	Cylinder/bore gauges	Y	40312	AC power supplies	Y			
10604	Depth gauges, Depth micrometers	Y	40313	Puncture/safety testers	Y			
10605	Dial/digital gauges	Y	40314	Power recorders	Y			
10609	Microindicators, Test indicators	Y	40318	AC voltmeters	Y			
10611	3-points, Micrometers	Y	404.	Other DC & LF Measurements				
10612	Inside micrometers	Y	40410	Line frequency meters	Y			
10613	Outside micrometers	Y	40416	Leakage current testers	Y			
			40417	Electronic AC/DC loads	Y			
			40419	Analogue/Digital multimeters	Y			
			40421	Oscilloscopes	Y			

Note

1. This laboratory provides calibration services in permanent standard laboratory and at on-site.
2. Laboratory conducts on-site calibration should meet requirements of KOLAS-SR-007.
3. On-site calibration is allowed to items with marking 'Y', not allowed to items with marking 'N'.
4. Measurement uncertainty normally is quoted as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k=2$. It expresses the lowest uncertainty of measurement that can be provided by accredited calibration laboratories in normal conditions.
5. Due to the calibration environment such as reference standards or customers' facilities, it is noted that uncertainty of measurement on a calibration certificate may be expressed larger than measurement uncertainty on scope of accreditation in general.

102. Linear dimension

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Filler gauges	10211	(0.01 ~ 10) mm	1.2 μm	Micrometers/ SICT-CP-10211
Height gauges/measuring machines	10216	(0 ~ 1 010) mm	$\sqrt{1.1^2 + (0.0037 \times l_0)^2} \mu\text{m}$	Gauge Block, Step gauge/ SICT-CP-10216
Cylindrical plug/pin gauges, Thread measuring wire gauges	10228			Laser Scan Micrometers/ SICT-CP-10228
Cylindrical plug/pin gauges		(1 ~ 20) mm	$\sqrt{0.48^2 + (0.0034 \times l_0)^2} \mu\text{m}$	
Ultrasonic thickness gauges	10234	(2.5 ~ 100) mm	3 μm	Ultrasonic Tester Blocks/ SICT-CP-10234
Coating thickness testers	10236	(0 ~ 1.527) mm	1.1 μm	Thickness specimens/ SICT-CP-10236

Note 1. l_0 unit : mm

104. Form

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Form testers	10401			Form Standard Specimens Gage Block, Angle Gage Block/ SICT-CP-10401
Height length		(0 ~ 50) mm	1.0 μm	
Width		(0 ~ 45) mm	1.7 μm	
Angle		15° ~ 45°	2'	
Precision surface plates	10407			Electronic Level/ SICT-CP-10407
Flatness		(2 000 ~ 20 000) cm ²	2.0 μm	
Roundness measurement instruments	10409			Roundness Standard Ball/ SICT-CP-10409
Detector accuracy		(0 ~ 15) μm	0.62 μm	
Rotational accuracy of spindle		360°	0.09 μm	
Rotational accuracy of axis		360°	0.05 μm	
Straight edges	10412			Electronic Micrometers/ SICT-CP-10412
		(0 ~ 1 500) mm	1.9 μm	
		(0 ~ 1 500) mm	1.8 μm	

Note 1. l_0 unit : mm

105. Complex geometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Contact coordinate measuring machines	10503	(0 ~ 1 000) mm	$\sqrt{0.94^2 + (0.0048 \times l_0)^2} \mu\text{m}$	Step Gauge/ SICT-CP-10503
Non-contact coordinate measuring machines	10504	(0 ~ 500) mm	$\sqrt{0.53^2 + (0.0036 \times l_0)^2} \mu\text{m}$	Standard Scale/ SICT-CP-10504
Measuring microscopes, Profile projectors	10511			Standard Scale/ SICT-CP-10511
Length		(0 ~ 300) mm	$\sqrt{0.56^2 + (0.0036 \times l_0)^2} \mu\text{m}$	
Magnification		(5 ~ 100) 배	0.05 %	
Angle		(0 ~ 360) °	0.9°	
Microscopes, micro measuring	10512			Standard Scale/ SICT-CP-10512
		(0 ~ 1) mm	1.3 μm	
		(1 ~ 50) mm	3.0 μm	
Stylus type roughness testers	10517			Roughness Specimen/ SICT-CP-10517
Ra		(0 ~ 4) μm	추후확인	
Rz		(0 ~ 12) μm	추후확인	

Note 1. l_0 unit : mm

106. Various dimensional

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Inside/Outside/Geartooth calipers, Caliper gauges	10601	(0 ~ 2 000) mm (0 ~ 100) mm (100 ~ 300) mm	$\sqrt{8.2^2 + (0.008 1 \times l_0)^2} \mu\text{m}$ $\sqrt{2.5^2 + (0.004 2 \times l_0)^2} \mu\text{m}$ $\sqrt{8.2^2 + (0.004 2 \times l_0)^2} \mu\text{m}$	Gauge Block/ SICT-CP-10601
Cylinder/bore gauges	10603	(0 ~ 400) mm (0.1 ~ 25) mm	0.7 μm 3.5 μm	Dial Gauge Tester/ SICT-CP-10603
Depth gauges, Depth micrometers	10604	(0 ~ 300) mm (0 ~ 600) mm	$\sqrt{0.87^2 + (0.003 3 \times l_0)^2} \mu\text{m}$ $\sqrt{6.0^2 + (0.007 8 \times l_0)^2} \mu\text{m}$	Gauge Block/ SICT-CP-10604
Dial/digital gauges	10605	(0 ~ 100) mm (0 ~ 25) mm	$\sqrt{0.33^2 + (0.006 8 \times l_0)^2} \mu\text{m}$ $\sqrt{0.82^2 + (0.006 8 \times l_0)^2} \mu\text{m}$	Gauge Block/ SICT-CP-10605
Micro indicators, Test indicators	10609	(0 ~ 5) mm	0.29 μm	Dial Gauge Tester/ SICT-CP-10609
3-points, Micrometers	10611	(6 ~ 100) mm	$\sqrt{1.2^2 + (0.004 1 \times l_0)^2} \mu\text{m}$	Ring Gauge/ SICT-CP-10611
Inside micrometers	10612	(5 ~ 200) mm (25 ~ 200) mm (13 ~ 150) mm	$\sqrt{1.1^2 + (0.004 2 \times l_0)^2} \mu\text{m}$ $\sqrt{1.1^2 + (0.004 2 \times l_0)^2} \mu\text{m}$ $\sqrt{1.5^2 + (0.004 2 \times l_0)^2} \mu\text{m}$	Gauge Block/ SICT-CP-10612
Outside micrometers	10613	(0 ~ 25) mm (25 ~ 500) mm (1 ~ 20) mm	$\sqrt{0.22^2 + (0.003 \times l_0)^2} \mu\text{m}$ $\sqrt{0.83^2 + (0.003 1 \times l_0)^2} \mu\text{m}$ 0.97 μm	Gauge Block, cylindrical plug gauges/ SICT-CP-10613

Note 1. l_0 unit : mm

201. Mass

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electric balances	20109	(0 ~ 2) g (2 ~ 6) g (6 ~ 20) g (20 ~ 50) g (50 ~ 200) g (200 ~ 500) g (500 ~ 2 000) g (2 ~ 5) kg (5 ~ 10) kg (10 ~ 20) kg (20 ~ 30) kg (30 ~ 60) kg (60 ~ 100) kg (100 ~ 600) kg (600 ~ 1 000) kg	21 µg 40 µg 42 µg 61 µg 0.12 mg 0.33 mg 1.2 mg 3.2 mg 6.2 mg 12 mg 0.20 g 0.36 g 6.8 g 18 g 38 g	Weight/ SICT-CP-20109
Platform scale balances	20112	(0 ~ 200) kg (200 ~ 500) kg (500 ~ 1 000) kg	46 g 0.092 kg 0.46 kg	Weight/ SICT-CP-20112
Weights less than class F1	20116	(1 mg ~ 20 kg) 1 mg 2 mg 5 mg 10 mg 20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg	3.5 µg 3.5 µg 3.5 µg 3.5 µg 4.4 µg 5.3 µg 6.4 µg 7.5 µg 10 µg 12 µg 16 µg 19 µg 24 µg 31 µg 38 µg 64 µg 0.12 mg 0.39 mg 0.66 mg 1.2 mg 4.7 mg 7.1 mg 13 mg	Weight, Electronic Balance/ SICT-CP-20116

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC ammeters	40101	(±)		Calibrator/ SICT-CP-40101
		0 μA	9.3 nA	
		(0.1 ~ 1) μA	9.4 × 10 ⁻³	
		(1 ~ 10) μA	9.9 × 10 ⁻⁴	
		(10 ~ 100) μA	1.5 × 10 ⁻⁴	
		(0.1 ~ 10) mA	6.9 × 10 ⁻⁵	
		(10 ~ 100) mA	8.0 × 10 ⁻⁵	
		(0.1 ~ 1) A	1.2 × 10 ⁻⁴	
		(1 ~ 10) A	6.5 × 10 ⁻⁴	
		(10 ~ 20) A	1.2 × 10 ⁻³	
DC voltage/current calibrators	40103	(±)		Reference Multimeter/ SICT-CP-40103
		0 mV	0.51 μV	
		(0 ~ 1) mV	6.1 × 10 ⁻⁴	
		(1 ~ 10) mV	6.2 × 10 ⁻⁵	
		(10 ~ 100) mV	1.3 × 10 ⁻⁵	
		(0.1 ~ 1) V	6.6 × 10 ⁻⁶	
		(1 ~ 10) V	5.2 × 10 ⁻⁶	
		(10 ~ 100) V	7.2 × 10 ⁻⁶	
		(100 ~ 1 000) V	9.0 × 10 ⁻⁶	
		(±)		
DC Current	40103	0 μA	6.9 nA	
		(0 ~ 10) μA	7.4 × 10 ⁻⁴	
		(10 ~ 100) μA	1.1 × 10 ⁻⁴	
		(0.1 ~ 1) mA	7.2 × 10 ⁻⁵	
		(1 ~ 10) mA	4.7 × 10 ⁻⁵	
		(10 ~ 100) mA	7.5 × 10 ⁻⁵	
		(0.1 ~ 1) A	2.4 × 10 ⁻⁴	
		(1 ~ 10) A	5.7 × 10 ⁻⁴	
		(10 ~ 100) A	2.8 × 10 ⁻⁴	
		(±)		
Electrical temperature calibrators	40104			Digital Multimeter/ SICT-CP-40104
		TEMPERATURE(SOURCE)		
		T/C		
		-9.835 mV	0.62 μV	
		(-9.835 ~ -5.237) mV	0.62 μV	
		(-5.237 ~ 0.000) mV	0.52 μV	
		(0.000 ~ 13.421) mV	0.62 μV	
		(13.421 ~ 68.788) mV	1.2 μV	
		(68.788 ~ 76.373) mV	1.3 μV	
		RTD		
		1.000 Ω	0.063 mΩ	
		(1.000 ~ 2.499) Ω	3.7 × 10 ⁻⁵	

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrical temperature calibrators				Digital Multimeter/ SICT-CP-40104
DC VOLTAGE(SOURCE)	40104	(±) 0 mV (0 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V	8.1 µV 8.5×10^{-4} 1.2×10^{-4} 4.6×10^{-5} 4.0×10^{-5} 5.9×10^{-5}	
DC CURRENT(SOURCE)		(±) 0 mA (0 ~ 1) mA (1 ~ 10) mA (10 ~ 100) mA	0.14 µA 7.0×10^{-4} 1.0×10^{-3} 7.0×10^{-4}	
RESISTANCE(SOURCE)		0 Ω (0 ~ 1) Ω (1 ~ 10) Ω 100 Ω ~ 50 kΩ	0.072 mΩ 6.7×10^{-5} 4.6×10^{-4} 1.2×10^{-4}	
TEMPERATURE(MEASURE)	T/C	-9.836 mV (-9.836 ~ -5.238) mV (-5.238 ~ 0.000) mV (0.000 ~ 0.002) mV (0.002 ~ 6.319) mV (6.319 ~ 13.421) mV (13.421 ~ 21.036) mV (21.036 ~ 28.946) mV (28.946 ~ 37.006) mV (37.006 ~ 53.112) mV (53.112 ~ 61.017) mV (61.017 ~ 76.373) mV	0.84 µV 0.83 µV 0.80 µV 0.70 µV 0.81 µV 0.87 µV 0.94 µV 1.0 µV 1.1 µV 1.2 µV 1.3 µV 1.4 µV	
	RTD	0.999 Ω (0.999 ~ 2.497) Ω (2.497 ~ 4.316) Ω (4.316 ~ 16.995) Ω (16.995 ~ 177.156) Ω (177.156 ~ 249.584) Ω (249.584 ~ 3 233.3) Ω	0.24 mΩ 1.0×10^{-4} 7.1×10^{-5} 3.9×10^{-5} 3.4×10^{-5} 3.5×10^{-5} 4.3×10^{-5}	
DC VOLTAGE(MEASURE)		(±) 0 mV (0 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 200) V (200 ~ 300) V	1.2 µV 1.4×10^{-4} 3.5×10^{-5} 6.0×10^{-5} 1.7×10^{-5} 2.3×10^{-5} 2.9×10^{-5}	
DC CURRENT(MEASURE)		(±) 0 mA (0 ~ 1) mA (1 ~ 10) mA (10 ~ 50) mA (50 ~ 100) mA	0.09 µA 9.3×10^{-5} 7.0×10^{-5} 9.6×10^{-5} 8.1×10^{-5}	

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrical temperature calibrators RESISTANCE(MEASURE)	40104	0 Ω (0 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) kΩ (1 ~ 10) kΩ (10 ~ 20) kΩ (20 ~ 30) kΩ (30 ~ 40) kΩ (40 ~ 50) kΩ	23 mΩ 1.3×10^{-4} 4.8×10^{-5} 3.4×10^{-5} 7.3×10^{-5} 7.0×10^{-5} 4.8×10^{-5} 4.0×10^{-5} 4.2×10^{-5} 3.9×10^{-5}	Digital Multimeter/ SICT-CP-40104
DC power supplies	40108	(±) 0 mV (0 ~ 10) mV (10 ~ 100) mV 100 mV ~ 1 V (1 ~ 10) V (10 ~ 100) V (100 ~ 600) V (600 ~ 1 000) V	5.8 μV 5.8×10^{-4} 5.9×10^{-5} 9.2×10^{-5} 5.3×10^{-5} 9.2×10^{-5} 1.7×10^{-5} 6.6×10^{-5}	DC Electronics Load/ SICT-CP-40108
DC Current		1 mA (1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 5) A (5 ~ 10) A (10 ~ 20) A (20 ~ 200) A	5.8 μA 6.5×10^{-3} 6.5×10^{-4} 8.7×10^{-5} 1.3×10^{-4} 4.0×10^{-4} 3.1×10^{-4} 2.2×10^{-4}	
Load regulation		0 mV (0 ~ 2) mV (2 ~ 20) mV (20 ~ 200) mV	0.97 mV 4.9×10^{-1} 3.3×10^{-1} 3.3×10^{-2}	
Ripple		0.1 mV (0.1 ~ 0.2) mV (0.2 ~ 0.4) mV (0.4 ~ 0.6) mV (0.6 ~ 1) mV (1 ~ 10) mV (10 ~ 50) mV	0.097 mV 6.4×10^{-1} 2.5×10^{-1} 1.6×10^{-1} 9.6×10^{-2} 9.2×10^{-2} 1.9×10^{-1}	
DC voltmeters	40112	(±) 0 mV (0 ~ 1) mV (1 ~ 2) mV (2 ~ 8) mV (8 ~ 10) mV (10 ~ 20) mV (20 ~ 50) mV (50 ~ 80) mV (80 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V	0.69 μV 8.0×10^{-4} 4.0×10^{-4} 1.6×10^{-4} 8.0×10^{-5} 4.5×10^{-5} 2.4×10^{-5} 1.9×10^{-5} 1.6×10^{-5} 9.4×10^{-6} 8.6×10^{-6} 1.0×10^{-5}	Calibrator/ SICT-CP-40112

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Earth testers	40205	60 Hz		Decade Resistor/ SICT-CP-40205
		1 V	6.1×10^{-3}	
		(1 ~ 10) V	6.2×10^{-4}	
		(10 ~ 50) V	1.2×10^{-3}	
		(50 ~ 100) V	6.2×10^{-4}	
		(100 ~ 500) V	4.9×10^{-4}	
		(500 ~ 1 000) V	7.7×10^{-4}	
	Resistance	1 mΩ	1.3×10^{-3}	
		1 mΩ ~ 1 Ω	1.3×10^{-3}	
		1 Ω ~ 10 Ω	8.4×10^{-4}	
		10 Ω ~ 100 kΩ	6.5×10^{-4}	
AC Current out	40205	60 Hz		
		1 A	3.1×10^{-3}	
		(1 ~ 3) A	4.1×10^{-3}	
		(3 ~ 10) A	3.1×10^{-3}	
		(10 ~ 20) A	3.4×10^{-3}	
		(20 ~ 30) A	6.4×10^{-3}	
	AC Current out	(30 ~ 60) A	4.6×10^{-3}	
		(1 ~ 100) s	5.8×10^{-6}	
		(100 ~ 1 000) s	8.2×10^{-6}	
Insulation testers	40210	60 Hz		High Resistance Decade/ SICT-CP-40210
		1 V	1.2×10^{-4}	
		(1 ~ 10) V	1.2×10^{-4}	
		(10 ~ 100) V	1.3×10^{-4}	
		(100 ~ 1 000) V	4.9×10^{-4}	
	Insulation Voltage	1 V	6.1×10^{-4}	
		(1 ~ 10) V	6.1×10^{-5}	
		(10 ~ 25) V	2.5×10^{-4}	
		(25 ~ 50) V	1.2×10^{-4}	
		(50 ~ 100) V	6.2×10^{-5}	
		(100 ~ 250) V	2.5×10^{-4}	
		(250 ~ 500) V	1.2×10^{-4}	
		(500 ~ 1 000) V	6.1×10^{-5}	
		(1 000 ~ 5 000) V	6.5×10^{-3}	
		(5 000 ~ 10 000) V	6.1×10^{-3}	
	Insulation Resistance	1 kΩ	1.4×10^{-4}	
		1 kΩ ~ 1 MΩ	1.2×10^{-4}	
		(1 ~ 10) MΩ	3.6×10^{-4}	
		(10 ~ 100) MΩ	1.2×10^{-3}	
		100 MΩ ~ 1 GΩ	2.3×10^{-3}	
		(1 ~ 10) GΩ	5.8×10^{-3}	
Resistance meters	40214	(10 ~ 100) GΩ	5.9×10^{-3}	
		100 GΩ ~ 1 TΩ	3.5×10^{-2}	
		1 mΩ	2.2×10^{-5}	
		(1 ~ 10) mΩ	2.1×10^{-5}	
		(10 ~ 100) mΩ	2.2×10^{-5}	
		100 mΩ ~ 1 MΩ	2.1×10^{-5}	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistors Decade Resistance	40215	0 Ω (0 ~ 1) mΩ (1 ~ 2) mΩ (2 ~ 3) mΩ (3 ~ 6) mΩ (6 ~ 8) mΩ (8 ~ 10) mΩ (10 ~ 30) mΩ (30 ~ 60) mΩ 60 mΩ ~ 0.1 Ω (0.1 ~ 0.3) Ω (0.3 ~ 0.6) Ω (0.6 ~ 1) Ω (1 ~ 3) Ω (3 ~ 10) Ω (10 ~ 300) Ω (300 ~ 1000) Ω (1 ~ 10) kΩ (10 ~ 100) kΩ 100 kΩ ~ 1 MΩ (1 ~ 2) MΩ (2 ~ 10) MΩ (10 ~ 100) MΩ 100 MΩ ~ 1 GΩ	62 μΩ 3.1×10^{-3} 6.2×10^{-3} 2.1×10^{-3} 1.5×10^{-3} 8.9×10^{-4} 6.9×10^{-4} 3.1×10^{-3} 1.6×10^{-3} 8.9×10^{-4} 3.1×10^{-4} 1.6×10^{-4} 9.0×10^{-5} 4.6×10^{-5} 2.5×10^{-5} 2.9×10^{-5} 1.6×10^{-5} 3.7×10^{-5} 3.0×10^{-5} 4.0×10^{-5} 8.5×10^{-5} 5.9×10^{-5} 6.5×10^{-4} 6.6×10^{-3}	Standard Resistance Set/ SICT-CP-40215

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC ammeters	40301	(100 μ A) 50 Hz ~ 1 kHz	2.7×10^{-3}	Power Calibrator, Calibrator/ SICT-CP-40301
		(0.1 ~ 1) mA 50 Hz ~ 1 kHz	1.4×10^{-3}	
		(1 ~ 10) mA (50 ~ 60) Hz 60 Hz ~ 1 kHz	6.1×10^{-4} 7.0×10^{-4}	
		(10 ~ 100) mA (50 ~ 60) Hz 60 Hz ~ 1 kHz	4.7×10^{-4} 7.0×10^{-4}	
		(0.1 ~ 1) A (50 ~ 60) Hz 60 Hz ~ 1 kHz	2.1×10^{-4} 7.0×10^{-4}	
		(1 ~ 5) A (50 ~ 60) Hz 60 Hz ~ 1 kHz	2.6×10^{-4} 1.3×10^{-3}	
		(5 ~ 10) A (50 ~ 60) Hz 60 Hz ~ 1 kHz	3.0×10^{-4} 1.4×10^{-3}	
		(10 ~ 20) A (50 ~ 60) Hz	4.8×10^{-4}	
		(20 ~ 40) A (50 ~ 60) Hz	5.2×10^{-4}	
Clamp ammeters/voltmeters	40302	(50 ~ 60) Hz		Power Calibrator, Calibrator/ SICT-CP-40302
		100 μ A	0.27μ A	
		(100 ~ 200) μ A	2.1×10^{-3}	
		(200 ~ 300) μ A	1.9×10^{-3}	
		(300 ~ 400) μ A	1.7×10^{-3}	
		(400 ~ 500) μ A	1.6×10^{-3}	
		(500 ~ 600) μ A	1.5×10^{-3}	
		(600 ~ 900) μ A	1.4×10^{-3}	
		900 μ A ~ 1 mA	1.5×10^{-3}	
		(1 ~ 2) mA	1.3×10^{-3}	
		(2 ~ 3) mA	1.2×10^{-3}	
		(3 ~ 4) mA	1.1×10^{-3}	
		(4 ~ 5) mA	9.5×10^{-4}	
		(5 ~ 6) mA	8.7×10^{-4}	
		(6 ~ 7) mA	8.1×10^{-4}	
		(7 ~ 8) mA	7.6×10^{-4}	
		(8 ~ 9) mA	7.3×10^{-4}	
		(9 ~ 10) mA	9.3×10^{-4}	
		(10 ~ 20) mA	7.6×10^{-4}	
		(20 ~ 30) mA	6.5×10^{-4}	
		(30 ~ 40) mA	6.0×10^{-4}	
		(40 ~ 50) mA	5.7×10^{-4}	
		(50 ~ 60) mA	4.8×10^{-4}	
		(60 ~ 70) mA	4.6×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters	40302	(70 ~ 90) mA	4.4×10^{-4}	Power Calibrator, Calibrator/ SICT-CP-40302
		(90 ~ 100) mA	7.5×10^{-4}	
		(100 ~ 200) mA	3.9×10^{-4}	
		(200 ~ 300) mA	3.1×10^{-4}	
		(300 ~ 400) mA	2.7×10^{-4}	
		(400 ~ 500) mA	2.5×10^{-4}	
		(500 ~ 700) mA	2.4×10^{-4}	
		(700 ~ 900) mA	2.2×10^{-4}	
		900 mA ~ 1 A	6.4×10^{-4}	
		(1 ~ 2) A	3.7×10^{-4}	
		(2 ~ 3) A	4.0×10^{-4}	
		(3 ~ 4) A	3.3×10^{-4}	
		(4 ~ 5) A	2.9×10^{-4}	
		(5 ~ 6) A	3.4×10^{-4}	
		(6 ~ 7) A	3.3×10^{-4}	
		(7 ~ 9) A	3.1×10^{-4}	
		(9 ~ 10) A	6.8×10^{-4}	
		(10 ~ 20) A	6.4×10^{-4}	
		(20 ~ 30) A	5.6×10^{-4}	
		(30 ~ 40) A	5.2×10^{-4}	
		(40 ~ 500) A	1.2×10^{-3}	
		(500 ~ 700) A	1.4×10^{-3}	
		(700 ~ 1 000) A	1.3×10^{-3}	
DC Current	30 μA	30 μA	29 nA	
		(30 ~ 40) μA	7.7×10^{-4}	
		(40 ~ 50) μA	6.5×10^{-4}	
		(50 ~ 60) μA	5.7×10^{-4}	
		(60 ~ 70) μA	5.1×10^{-4}	
		(70 ~ 80) μA	4.7×10^{-4}	
		(80 ~ 90) μA	4.4×10^{-4}	
		(90 ~ 100) μA	7.3×10^{-4}	
		(100 ~ 200) μA	4.2×10^{-4}	
		(200 ~ 300) μA	3.2×10^{-4}	
		(300 ~ 400) μA	2.8×10^{-4}	
		(400 ~ 500) μA	2.5×10^{-4}	
		(500 ~ 600) μA	2.4×10^{-4}	
		(600 ~ 700) μA	2.3×10^{-4}	
		(700 ~ 800) μA	2.2×10^{-4}	
		(800 ~ 900) μA	2.1×10^{-4}	
		900 μA ~ 1 mA	6.4×10^{-4}	
		(1 ~ 2) mA	3.6×10^{-4}	
		(2 ~ 4) mA	2.7×10^{-4}	
		(4 ~ 9) mA	1.8×10^{-4}	
		(9 ~ 10) mA	6.2×10^{-4}	
		(10 ~ 20) mA	3.3×10^{-4}	
		(20 ~ 70) mA	2.4×10^{-4}	
		(70 ~ 90) mA	1.5×10^{-4}	
		(90 ~ 100) mA	6.2×10^{-4}	
		(100 ~ 200) mA	3.3×10^{-4}	
		(200 ~ 400) mA	3.9×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters	DC Current	(400 ~ 900) mA	3.5×10^{-4}	Power Calibrator, Calibrator/ SICT-CP-40302
		900 mA ~ 1 A	6.7×10^{-4}	
		(1 ~ 2) A	4.8×10^{-4}	
		(2 ~ 3) A	6.0×10^{-4}	
		(3 ~ 9) A	8.1×10^{-4}	
		(9 ~ 10) A	8.9×10^{-4}	
		(10 ~ 20) A	1.2×10^{-3}	
		(20 ~ 40) A	3.8×10^{-4}	
		(40 ~ 500) A	1.3×10^{-3}	
		(500 ~ 1 000) A	1.7×10^{-3}	
	AC Voltage	(50 ~ 60) Hz		
		10 mV	11 μ V	
		(10 ~ 20) mV	6.3×10^{-4}	
		(20 ~ 30) mV	4.7×10^{-4}	
		(30 ~ 40) mV	4.6×10^{-4}	
		(40 ~ 50) mV	4.0×10^{-4}	
		(50 ~ 60) mV	3.6×10^{-4}	
		(60 ~ 80) mV	3.3×10^{-4}	
		(80 ~ 90) mV	2.9×10^{-4}	
		(90 ~ 100) mV	2.8×10^{-4}	
		(100 ~ 200) mV	2.2×10^{-4}	
		(200 ~ 300) mV	2.0×10^{-4}	
		(300 ~ 400) mV	3.6×10^{-4}	
		(400 ~ 600) mV	3.2×10^{-4}	
		(600 ~ 700) mV	2.8×10^{-4}	
		(700 ~ 800) mV	2.6×10^{-4}	
		800 mV ~ 1 V	2.5×10^{-4}	
		(1 ~ 3) V	2.1×10^{-4}	
		(3 ~ 4) V	3.5×10^{-4}	
	DC Voltage	(4 ~ 5) V	3.2×10^{-4}	
		(5 ~ 6) V	2.9×10^{-4}	
		(6 ~ 7) V	2.8×10^{-4}	
		(7 ~ 8) V	2.6×10^{-4}	
		(8 ~ 10) V	2.5×10^{-4}	
		(10 ~ 20) V	2.3×10^{-4}	
		(20 ~ 40) V	1.9×10^{-4}	
		(40 ~ 70) V	2.2×10^{-4}	
		(70 ~ 90) V	1.9×10^{-4}	
		(90 ~ 100) V	1.8×10^{-4}	
		(100 ~ 200) V	1.5×10^{-4}	
		(200 ~ 500) V	1.9×10^{-4}	
		(500 ~ 700) V	2.6×10^{-4}	
		(700 ~ 1 000) V	3.7×10^{-4}	
		10 mV	61 μ V	
		(10 ~ 20) mV	3.1×10^{-3}	
		(20 ~ 30) mV	2.0×10^{-3}	
		(30 ~ 40) mV	1.5×10^{-3}	
		(40 ~ 50) mV	1.2×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters	40302	(100 ~ 200) mV	4.2×10^{-5}	Power Calibrator, Calibrator/ SICT-CP-40302
		(200 ~ 300) mV	3.4×10^{-5}	
		(300 ~ 400) mV	2.4×10^{-5}	
		(400 ~ 500) mV	2.2×10^{-5}	
		(500 ~ 600) mV	2.0×10^{-5}	
		(600 ~ 800) mV	1.8×10^{-5}	
		(800 ~ 900) mV	1.7×10^{-5}	
		900 mV ~ 1 V	6.3×10^{-5}	
		(1 ~ 2) V	3.4×10^{-5}	
		(2 ~ 4) V	2.5×10^{-5}	
		(4 ~ 5) V	2.2×10^{-5}	
		(5 ~ 6) V	2.0×10^{-5}	
		(6 ~ 7) V	1.9×10^{-5}	
		(7 ~ 9) V	1.8×10^{-5}	
		(9 ~ 10) V	6.3×10^{-5}	
		(10 ~ 20) V	3.4×10^{-5}	
		(20 ~ 30) V	2.5×10^{-5}	
		(30 ~ 40) V	3.0×10^{-6}	
		(40 ~ 50) V	2.7×10^{-5}	
		(50 ~ 60) V	2.6×10^{-5}	
		(60 ~ 70) V	2.5×10^{-5}	
		(70 ~ 90) V	2.4×10^{-5}	
		(90 ~ 100) V	6.5×10^{-5}	
		(100 ~ 200) V	3.8×10^{-5}	
		(200 ~ 400) V	3.0×10^{-5}	
		(400 ~ 500) V	2.8×10^{-5}	
		(500 ~ 600) V	2.6×10^{-5}	
		(600 ~ 700) V	2.5×10^{-5}	
		(700 ~ 900) V	2.4×10^{-5}	
		(900 ~ 1 000) V	6.5×10^{-5}	
Resistance	Resistance	1 Ω	0.62 mΩ	Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
		(1 ~ 9) Ω	3.1×10^{-4}	
		(9 ~ 100) Ω	6.1×10^{-4}	
		(100 ~ 900) Ω	4.7×10^{-5}	
		900 Ω ~ 90 kΩ	7.0×10^{-5}	
		90 kΩ ~ 1 MΩ	7.2×10^{-5}	
		(1 ~ 10) MΩ	1.9×10^{-4}	
		(10 ~ 100) MΩ	7.1×10^{-4}	
AC voltage/current calibrators	40303	(2 mV)		Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
		10 Hz	8.2×10^{-3}	
		(10 ~ 40) Hz	2.5×10^{-3}	
		40 Hz ~ 1 kHz	1.3×10^{-3}	
		(1 ~ 20) kHz	5.0×10^{-3}	
		(20 ~ 100) kHz	1.2×10^{-2}	
		(2 ~ 20) mV		
		10 Hz	9.5×10^{-4}	
		(10 ~ 40) Hz	3.5×10^{-4}	
		40 Hz ~ 1 kHz	2.3×10^{-4}	
		(1 ~ 20) kHz	8.1×10^{-4}	
		(20 ~ 100) kHz	2.0×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators	40303			Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
AC Voltage		(20 ~ 30) mV		
		10 Hz	6.8×10^{-4}	
		(10 ~ 40) Hz	2.7×10^{-4}	
		40 Hz ~ 1 kHz	1.9×10^{-4}	
		(1 ~ 20) kHz	6.5×10^{-4}	
		(20 ~ 100) kHz	1.6×10^{-3}	
		(30 ~ 40) mV		
		10 Hz	5.4×10^{-4}	
		(10 ~ 40) Hz	2.3×10^{-4}	
		40 Hz ~ 1 kHz	1.7×10^{-4}	
		(1 ~ 20) kHz	5.8×10^{-4}	
		(20 ~ 100) kHz	1.4×10^{-3}	
		(40 ~ 50) mV		
		10 Hz	4.6×10^{-4}	
		(10 ~ 40) Hz	2.1×10^{-4}	
		40 Hz ~ 1 kHz	1.6×10^{-4}	
		(1 ~ 20) kHz	5.3×10^{-4}	
		(20 ~ 100) kHz	1.3×10^{-3}	
		(50 ~ 60) mV		
		10 Hz	4.1×10^{-4}	
		(10 ~ 40) Hz	1.9×10^{-4}	
		40 Hz ~ 1 kHz	1.5×10^{-4}	
		(1 ~ 20) kHz	5.0×10^{-4}	
		(20 ~ 100) kHz	1.2×10^{-3}	
		(70 ~ 90) mV		
		10 Hz	3.7×10^{-4}	
		(10 ~ 40) Hz	1.8×10^{-4}	
		40 Hz ~ 1 kHz	1.5×10^{-4}	
		(1 ~ 20) kHz	4.8×10^{-4}	
		(20 ~ 100) kHz	1.1×10^{-3}	
		(90 ~ 100) mV		
		10 Hz	3.0×10^{-4}	
		(10 ~ 40) Hz	1.6×10^{-4}	
		40 Hz ~ 1 kHz	1.4×10^{-4}	
		(1 ~ 20) kHz	4.4×10^{-4}	
		(20 ~ 100) kHz	1.0×10^{-3}	
		(100 ~ 200) mV		
		10 Hz	2.3×10^{-4}	
		(10 ~ 40) Hz	1.4×10^{-4}	
		40 Hz ~ 1 kHz	1.3×10^{-4}	
		(1 ~ 20) kHz	3.9×10^{-4}	
		(20 ~ 100) kHz	9.3×10^{-4}	
		(200 ~ 300) mV		
		10 Hz	8.6×10^{-4}	
		(10 ~ 40) Hz	2.5×10^{-4}	
		40 Hz ~ 1 kHz	2.0×10^{-4}	
		(1 ~ 20) kHz	4.1×10^{-4}	
		(20 ~ 100) kHz	1.4×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators				
AC Voltage	40303	(300 mV ~ 0.4 V)		Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
		10 Hz	6.6×10^{-4}	
		(10 ~ 40) Hz	2.1×10^{-4}	
		40 Hz ~ 1 kHz	1.6×10^{-4}	
		(1 ~ 20) kHz	3.6×10^{-4}	
		(20 ~ 100) kHz	1.2×10^{-3}	
		(0.4 ~ 0.9) V		
		10 Hz	5.7×10^{-4}	
		(10 ~ 40) Hz	1.9×10^{-4}	
		40 Hz ~ 1 kHz	1.6×10^{-4}	
		(1 ~ 20) kHz	3.3×10^{-4}	
		(20 ~ 100) kHz	1.2×10^{-3}	
		(0.9 ~ 2) V		
		10 Hz	4.1×10^{-4}	
		(10 ~ 40) Hz	1.6×10^{-4}	
		40 Hz ~ 1 kHz	1.1×10^{-4}	
		(1 ~ 20) kHz	2.8×10^{-4}	
		(20 ~ 100) kHz	8.2×10^{-4}	
		(2 ~ 3) V		
		10 Hz	8.6×10^{-4}	
		(10 ~ 40) Hz	2.3×10^{-4}	
		40 Hz ~ 1 kHz	2.0×10^{-4}	
		(1 ~ 20) kHz	4.1×10^{-4}	
		(20 ~ 100) kHz	1.4×10^{-3}	
		(3 ~ 9) V		
		10 Hz	6.6×10^{-4}	
		(10 ~ 40) Hz	2.0×10^{-4}	
		40 Hz ~ 1 kHz	1.6×10^{-4}	
		(1 ~ 20) kHz	3.6×10^{-4}	
		(20 ~ 100) kHz	1.2×10^{-3}	
		(9 ~ 20) V		
		10 Hz	4.1×10^{-4}	
		(10 ~ 40) Hz	1.5×10^{-4}	
		40 Hz ~ 1 kHz	1.1×10^{-4}	
		(1 ~ 20) kHz	2.9×10^{-4}	
		(20 ~ 100) kHz	8.4×10^{-4}	
		(20 ~ 30) V		
		10 Hz	8.8×10^{-4}	
		(10 ~ 40) Hz	2.5×10^{-4}	
		40 Hz ~ 1 kHz	2.0×10^{-4}	
		(1 ~ 20) kHz	4.1×10^{-4}	
		(20 ~ 100) kHz	1.4×10^{-3}	
		(30 ~ 90) V		
		10 Hz	6.8×10^{-4}	
		(10 ~ 40) Hz	2.1×10^{-4}	
		40 Hz ~ 1 kHz	1.6×10^{-4}	
		(1 ~ 20) kHz	3.6×10^{-4}	
		(20 ~ 100) kHz	1.2×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators	AC Voltage	40303		Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
		(90 ~ 200) V		
		10 Hz	8.6×10^{-4}	
		(10 ~ 40) Hz	2.3×10^{-4}	
		40 Hz ~ 1 kHz	2.0×10^{-4}	
		(1 ~ 20) kHz	4.1×10^{-4}	
		(20 ~ 100) kHz	1.4×10^{-3}	
		(200 ~ 300) V		
		50 Hz ~ 1 kHz	1.7×10^{-4}	
		(300 ~ 1 000) V		
	AC Current	50 Hz	1.3×10^{-4}	
		50 Hz ~ 1 kHz	1.2×10^{-4}	
		(100 μA)		
		10 Hz	7.1×10^{-4}	
		10 Hz ~ 1 kHz	5.6×10^{-4}	
		(1 ~ 5) kHz	7.1×10^{-4}	
		(5 ~ 10) kHz	2.3×10^{-3}	
		(100 μA ~ 1 mA)		
		10 Hz	6.2×10^{-4}	
		10 Hz ~ 1 kHz	5.5×10^{-4}	
		(1 ~ 5) kHz	6.4×10^{-4}	
		(5 ~ 10) kHz	2.1×10^{-3}	
		(1 ~ 10) mA		
		10 Hz	6.2×10^{-4}	
		10 Hz ~ 1 kHz	5.5×10^{-4}	
		(1 ~ 5) kHz	6.2×10^{-4}	
		(5 ~ 10) kHz	3.2×10^{-3}	
		(10 ~ 100) mA		
		10 Hz	6.3×10^{-4}	
		10 Hz ~ 1 kHz	5.4×10^{-4}	
		(1 ~ 5) kHz	6.1×10^{-4}	
		(5 ~ 10) kHz	1.5×10^{-3}	
		(100 mA ~ 1 A)		
		40 Hz ~ 1 kHz	9.8×10^{-4}	
		(1 ~ 5) kHz	1.2×10^{-3}	
		(5 ~ 10) kHz	8.1×10^{-3}	
		(1 ~ 10) A		
		(40 ~ 100) Hz	1.3×10^{-3}	
		100 Hz ~ 1 kHz	1.5×10^{-3}	
		(10 ~ 20) A		
		(40 ~ 100) Hz	1.6×10^{-3}	
		100 Hz ~ 1 kHz	1.7×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Power factor meters	40310	(50 ~ 60) Hz		Power Calibrator/ SICT-CP-40310
AC Power Factor		-1 ~ 1	2.1×10^{-4}	
		-0.9 , 0.9	2.7×10^{-4}	
		-0.8 , 0.8	3.2×10^{-4}	
		-0.7 , 0.7	4.2×10^{-4}	
		-0.6 , 0.6	4.4×10^{-4}	
		-0.5 , 0.5	4.6×10^{-4}	
		-0.4 , 0.4	4.8×10^{-4}	
		-0.3 , 0.3	4.9×10^{-4}	
		-0.2 , 0.2	5.0×10^{-4}	
		-0.1 , 0.1	5.0×10^{-4}	
AC power meters	40311	(50 ~ 60) Hz		Power Calibrator, Calibrator/ SICT-CP-40311
AC Voltage		3 mV	$8 \mu\text{V}$	
		(3 ~ 10) mV	9.6×10^{-4}	
		(10 ~ 20) mV	5.6×10^{-4}	
		(20 ~ 40) mV	4.4×10^{-4}	
		(40 ~ 60) mV	3.8×10^{-4}	
		(60 ~ 90) mV	3.2×10^{-4}	
		(90 ~ 100) mV	2.4×10^{-4}	
		(100 ~ 200) mV	2.4×10^{-4}	
		(200 ~ 300) mV	1.6×10^{-4}	
		(300 ~ 400) mV	1.3×10^{-4}	
		(400 ~ 600) mV	1.2×10^{-4}	
		(600 ~ 700) mV	1.1×10^{-4}	
		(700 ~ 900) mV	1.0×10^{-4}	
		900 mV ~ 1 V	9.9×10^{-5}	
		(1 ~ 5) V	1.2×10^{-4}	
		(5 ~ 10) V	9.8×10^{-5}	
		(10 ~ 20) V	2.2×10^{-4}	
		(20 ~ 30) V	1.8×10^{-4}	
		(30 ~ 40) V	1.5×10^{-4}	
		(40 ~ 70) V	1.4×10^{-4}	
		(70 ~ 100) V	1.2×10^{-4}	
		(100 ~ 300) V	1.6×10^{-4}	
		(300 ~ 400) V	1.3×10^{-4}	
		(400 ~ 500) V	1.2×10^{-4}	
		(500 ~ 800) V	1.1×10^{-4}	
		(800 ~ 1 000) V	1.0×10^{-4}	
AC Current		(50 ~ 60) Hz		
		100 μA	$0.27 \mu\text{A}$	
		(100 ~ 200) μA	2.0×10^{-3}	
		(200 ~ 400) μA	1.8×10^{-3}	
		(400 ~ 500) μA	1.6×10^{-3}	
		(500 ~ 700) μA	1.5×10^{-3}	
		(700 ~ 900) μA	1.4×10^{-3}	
		900 μA ~ 1 mA	3.2×10^{-4}	
		(1 ~ 2) mA	3.5×10^{-4}	
		(2 ~ 5) mA	2.7×10^{-4}	
		(5 ~ 8) mA	2.5×10^{-4}	
		(8 ~ 10) mA	2.2×10^{-4}	
		(10 ~ 15) mA	4.4×10^{-4}	
		(15 ~ 20) mA	3.5×10^{-4}	
		(20 ~ 30) mA	3.6×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311	(30 ~ 40) mA	3.0×10^{-4}	Power Calibrator, Calibrator/ SICT-CP-40311
		(40 ~ 60) mA	2.7×10^{-4}	
		(60 ~ 70) mA	2.4×10^{-4}	
		(70 ~ 100) mA	2.3×10^{-4}	
		(100 ~ 150) mA	2.7×10^{-4}	
		(150 ~ 300) mA	2.5×10^{-4}	
		(300 ~ 800) mA	2.2×10^{-4}	
		(800 mA ~ 2 A)	2.1×10^{-4}	
		(2 ~ 3) A	3.4×10^{-4}	
		(3 ~ 4) A	2.9×10^{-4}	
		(4 ~ 5) A	2.6×10^{-4}	
		(5 ~ 6) A	3.3×10^{-4}	
		(6 ~ 10) A	3.1×10^{-4}	
		(10 ~ 15) A	6.2×10^{-4}	
		(15 ~ 20) A	5.7×10^{-4}	
		(20 ~ 30) A	5.2×10^{-4}	
		(30 ~ 40) A	5.0×10^{-4}	
		(40 ~ 500) A	1.3×10^{-3}	
		(500 ~ 800) A	1.4×10^{-3}	
		(800 ~ 1 000) A	1.3×10^{-3}	
AC Wattage		(50 ~ 60) Hz		
		24 mW	0.42 mW	
		(24 ~ 72) mW	5.8×10^{-3}	
		(72 ~ 120) mW	3.5×10^{-3}	
		(120 ~ 480) mW	2.9×10^{-3}	
		480 mW ~ 2.4 W	1.1×10^{-3}	
		(2.4 ~ 24) W	4.9×10^{-4}	
		(24 ~ 120) W	2.1×10^{-4}	
		(120 ~ 240) W	2.3×10^{-4}	
		(240 ~ 480) W	2.6×10^{-4}	
		(480 ~ 600) W	2.0×10^{-4}	
		(600 ~ 960) W	2.4×10^{-4}	
		960 W ~ 1.2 kW	2.0×10^{-4}	
		(1.2 ~ 2.4) kW	3.0×10^{-4}	
		(2.4 ~ 4.8) kW	6.4×10^{-4}	
		(4.8 ~ 9.6) kW	5.8×10^{-4}	
		(9.6 ~ 120) kW	1.3×10^{-3}	
Frequency		25 Hz	9.1 mHz	
		(25 ~ 60) Hz	1.7×10^{-4}	
		(60 ~ 100) Hz	6.0×10^{-4}	
		(100 ~ 200) Hz	3.9×10^{-4}	
		(200 ~ 400) Hz	2.7×10^{-4}	
		(400 ~ 1 000) Hz	1.3×10^{-3}	
Power Factor		(50 ~ 60) Hz		
		240 mW		
		-1 ~ 1	3.1×10^{-4}	
		-0.8, 0.8	3.9×10^{-4}	
		-0.5, 0.5	4.6×10^{-4}	
		-0.3, 0.3	5.6×10^{-4}	
		-0.1, 0.1	5.7×10^{-4}	
		240 mW ~ 120 kW		
		-1 ~ 1	2.1×10^{-4}	
		-0.8, 0.8	3.2×10^{-4}	
		-0.5, 0.5	4.6×10^{-4}	
		-0.3, 0.3	4.9×10^{-4}	
		-0.1, 0.1	5.0×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power supplies	40312	100 mV		Voltage Standard, Multimeter, Current Shunt / SICT-CP-40312
		40 Hz ~ 1 kHz	2.7×10^{-4}	
		(100 mV ~ 10 V)		
		40 Hz ~ 1 kHz	1.3×10^{-4}	
		(10 ~ 100) V		
		50 Hz ~ 1 kHz	1.4×10^{-4}	
		(100 ~ 500) V		
		40 Hz ~ 5 kHz	1.3×10^{-4}	
		(100 ~ 1 000) V		
		40 Hz ~ 5 kHz	1.4×10^{-4}	
Frequency		10 Hz	1.3×10^{-4}	
		10 Hz ~ 5 kHz	1.2×10^{-4}	
AC Current		(50 ~ 60) Hz		
		(1 mA)	3.7×10^{-3}	
		1 mA ~ 1 A	2.4×10^{-3}	
		(1 ~ 5) A	2.6×10^{-3}	
		(5 ~ 8) A	2.9×10^{-3}	
		(8 ~ 10) A	2.5×10^{-3}	
Puncture/safety testers	40313	0.1 kV	2.8 V	AC/DC Kilovoltmeter,
		(0.1 ~ 0.5) kV	1.0×10^{-2}	High Voltage Digital Meter,
		(0.5 ~ 1) kV	6.6×10^{-3}	Reference Multimeter /
		(1 ~ 2) kV	4.8×10^{-3}	SICT-CP-40313
		(2 ~ 3) kV	4.2×10^{-3}	
		(3 ~ 4) kV	3.9×10^{-3}	
		(4 ~ 5) kV	3.7×10^{-3}	
		(5 ~ 6) kV	3.6×10^{-3}	
		(6 ~ 10) kV	3.4×10^{-3}	
		60 Hz		
AC Voltage		0.1 kV	2.4 V	
		(0.1 ~ 0.5) kV	7.4×10^{-3}	
		(0.5 ~ 1) kV	4.8×10^{-3}	
		(1 ~ 2) kV	3.5×10^{-3}	
		(2 ~ 4) kV	3.0×10^{-3}	
		(4 ~ 5) kV	2.8×10^{-3}	
		(5 ~ 10) kV	2.6×10^{-3}	
		60 Hz		
AC Breaking Current		0.1 mA	0.67 μ A	
		(0.1 ~ 0.5) mA	1.5×10^{-3}	
		(0.5 ~ 1) mA	8.2×10^{-4}	
		(1 ~ 2) mA	5.3×10^{-4}	
		(2 ~ 5) mA	8.3×10^{-4}	
		(5 ~ 10) mA	5.5×10^{-4}	
		(10 ~ 20) mA	5.3×10^{-3}	
		(20 ~ 100) mA	8.3×10^{-4}	
		60 Hz		
		0.1 mA		

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Puncture/safety testers				
DC Breaking Current	40313	0.1 mA (0.1 ~ 0.5) mA (0.5 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 10) mA (10 ~ 20) mA (20 ~ 50) mA (50 ~ 100) mA	0.61 μ A 1.2×10^{-3} 6.1×10^{-4} 3.1×10^{-4} 1.5×10^{-4} 7.7×10^{-5} 3.2×10^{-4} 1.8×10^{-4} 6.2×10^{-4}	AC/DC Kilovoltmeter, High Voltage Digital Meter, Reference Multimeter/ SICT-CP-40313
Resistance		1 m Ω 1 m Ω ~ 10 m Ω 10 m Ω ~ 100 k Ω	0.84 μ Ω 7.0×10^{-4} 6.5×10^{-4}	
Insulation Voltage		1 V (1 ~ 10) V (10 ~ 25) V (25 ~ 50) V (50 ~ 100) V (100 ~ 250) V (250 ~ 500) V (500 ~ 1 000) V (1 000 ~ 2 000) V	0.62 mV 1.2×10^{-4} 2.8×10^{-4} 1.7×10^{-4} 1.2×10^{-4} 2.8×10^{-4} 1.7×10^{-4} 1.2×10^{-4} 6.9×10^{-3}	
Insulation Resistance		1 k Ω (1 ~ 100) k Ω 100 k Ω ~ 1 M Ω (1 ~ 10) M Ω (10 ~ 100) M Ω 100 M Ω ~ 1 G Ω (1 ~ 10) G Ω (10 ~ 100) G Ω 100 G Ω ~ 1 T Ω	0.14 Ω 1.3×10^{-4} 1.0×10^{-4} 9.1×10^{-4} 2.0×10^{-4} 2.6×10^{-4} 9.1×10^{-3} 1.5×10^{-3} 2.8×10^{-3}	
Leakage current(DC)		30 μ A (30 ~ 100) μ A 100 μ A ~ 10 mA (10 ~ 50) mA	29 nA 7.3×10^{-4} 6.4×10^{-4} 2.2×10^{-4}	
Leakage current(AC)		60 Hz 30 μ A (30 ~ 100) μ A 100 μ A ~ 1 mA (1 ~ 10) mA (10 ~ 50) mA	0.16 μ A 2.7×10^{-3} 1.5×10^{-3} 9.3×10^{-4} 9.5×10^{-4}	
Timer		1 s (1 ~ 100) s (100 ~ 1 000) s (1 000 ~ 10 000) s	5.8 μ s 5.8×10^{-6} 8.2×10^{-6} 5.8×10^{-5}	
Output AC Current		60 Hz 1 A (1 ~ 5) A (5 ~ 10) A (10 ~ 20) A (20 ~ 30) A (30 ~ 50) A (50 ~ 60) A	2.9 mA 2.1×10^{-3} 2.0×10^{-3} 3.0×10^{-3} 4.0×10^{-3} 3.1×10^{-3} 7.8×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Power recorders	40314	(50 ~ 60) Hz 240 mW 240 mW ~ 1.2 W (1.2 ~ 60) W 60 W ~ 2.4 kW (2.4 ~ 9.6) kW (9.6 ~ 120) kW	1.4 mW 3.2×10^{-3} 4.9×10^{-4} 3.0×10^{-4} 6.4×10^{-4} 3.5×10^{-3}	Power Energy Calibrator/ SICT-CP-40314
AC voltmeters	40318	1 mV 50 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz (1 ~ 3) mV 50 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz (3 ~ 10) mV 50 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz (10 ~ 30) mV 50 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz (30 ~ 100) mV 50 Hz ~ 10 kHz (10 ~ 100) kHz (100 ~ 300) mV 50 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz (300 mV ~ 1 V) 50 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz (1 ~ 3) V 50 Hz ~ 10 kHz (10 ~ 100) kHz (3 ~ 10) V 50 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz	7.4 $\times 10^{-3}$ 8.7×10^{-3} 3.2×10^{-2} 2.6×10^{-3} 3.0×10^{-3} 1.2×10^{-2} 9.6×10^{-4} 1.0×10^{-3} 6.1×10^{-3} 4.3×10^{-4} 4.4×10^{-4} 4.6×10^{-3} 2.8×10^{-4} 1.3×10^{-3} 2.0×10^{-4} 2.4×10^{-4} 1.4×10^{-3} 2.5×10^{-4} 2.6×10^{-4} 1.0×10^{-3} 2.1×10^{-4} 1.0×10^{-3} 2.4×10^{-4} 2.5×10^{-4} 1.2×10^{-3}	Reference Multimeter, Calibrator/ SICT-CP-40318 SICT-CP-40318 SICT-CP-40318

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
AC voltmeters	40318	(10 ~ 30) V		Reference Multimeter, Calibrator/ SICT-CP-40318	
		50 Hz ~ 1 kHz	2.0×10^{-4}		
		(1 ~ 10) kHz	2.2×10^{-4}		
		50 Hz ~ 100 kHz	1.2×10^{-3}		
		(30 ~ 100) V			
		50 Hz ~ 1 kHz	2.4×10^{-4}		
		(1 ~ 10) kHz	2.5×10^{-4}		
		50 Hz ~ 100 kHz	2.9×10^{-3}		
		(100 ~ 1 000) V			
		50 Hz ~ 1 kHz	3.6×10^{-4}		
AC Output Voltage		(50 Hz ~ 1 kHz)		Calibrator/ SICT-CP-40410	
		1 mV	$5.4 \mu\text{V}$		
		(1 ~ 10) mV	7.8×10^{-4}		
		(10 ~ 100) mV	2.0×10^{-4}		
		100 mV ~ 1 V	1.1×10^{-4}		
DC Output Voltage		1 mV	$0.86 \mu\text{V}$	Calibrator/ SICT-CP-40410	
		(1 ~ 10) mV	8.7×10^{-5}		
		(10 ~ 100) mV	1.4×10^{-5}		
		100 mV ~ 1 V	9.0×10^{-5}		
Line frequency meters	40410	10 Hz	6.1 mHz	Calibrator/ SICT-CP-40410	
		(10 ~ 60) Hz	1.5×10^{-4}		
		(60 ~ 400) Hz	1.2×10^{-4}		
		400 Hz ~ 1 kHz	1.4×10^{-4}		
Leakage current testers	40416	(±)		Calibrator/ SICT-CP-40416	
		0 μA	9 nA		
		(0 ~ 5) μA	1.9×10^{-3}		
		(5 ~ 10) μA	9.9×10^{-4}		
		(10 ~ 30) μA	3.7×10^{-4}		
		(30 ~ 50) μA	2.4×10^{-4}		
		(50 ~ 100) μA	1.6×10^{-4}		
		(100 ~ 200) μA	1.3×10^{-4}		
		(200 ~ 500) μA	8.1×10^{-5}		
		(0.5 ~ 1) mA	6.1×10^{-4}		
		(1 ~ 2) mA	3.2×10^{-4}		
		(2 ~ 5) mA	1.5×10^{-4}		
		(5 ~ 10) mA	6.1×10^{-4}		
		(10 ~ 15) mA	4.1×10^{-4}		
		(15 ~ 20) mA	3.1×10^{-4}		
	AC Current	(30 μA)		Calibrator/ SICT-CP-40416	
		50 Hz ~ 1 kHz	$0.16 \mu\text{A}$		
		(30 ~ 50) μA			
		50 Hz ~ 1 kHz	3.9×10^{-3}		
		(50 ~ 100) μA			
		50 Hz ~ 1 kHz	2.7×10^{-3}		

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers				Calibrator/ SICT-CP-40416
AC Current	40416	(100 ~ 200) μ A 50 Hz ~ 1 kHz	2.1×10^{-3}	
		(200 ~ 500) μ A 50 Hz ~ 1 kHz	1.6×10^{-3}	
		(0.5 ~ 1) mA 50 Hz ~ 1 kHz	6.9×10^{-4}	
		(1 ~ 2) mA 50 Hz ~ 1 kHz	4.7×10^{-4}	
		(2 ~ 5) mA 50 Hz ~ 1 kHz	3.0×10^{-3}	
		(5 ~ 10) mA 50 Hz ~ 1 kHz	8.2×10^{-4}	
		(10 ~ 20) mA 50 Hz ~ 1 kHz	5.3×10^{-4}	
		(20 ~ 30) mA 50 Hz ~ 1 kHz	1.2×10^{-4}	
DC Voltage		0 V	61 μ V	
		(0 ~ 0.1) V	6.1×10^{-4}	
		(0.1 ~ 0.2) V	3.1×10^{-4}	
		(0.2 ~ 0.5) V	1.2×10^{-4}	
		(0.5 ~ 1) V	6.1×10^{-4}	
		(1 ~ 2) V	3.1×10^{-4}	
		(2 ~ 5) V	1.2×10^{-4}	
		(5 ~ 10) V	6.2×10^{-5}	
		(10 ~ 20) V	3.3×10^{-5}	
		(20 ~ 50) V	1.7×10^{-5}	
		(50 ~ 100) V	1.2×10^{-5}	
		(100 ~ 200) V	3.4×10^{-5}	
		(200 ~ 300) V	2.5×10^{-5}	
		(300 ~ 500) V	1.2×10^{-4}	
		(500 ~ 1 000) V	6.2×10^{-5}	
AC Voltage		(0.1 V) 50 Hz ~ 1 kHz	6.5×10^{-4}	
		(1 ~ 20) kHz	6.6×10^{-4}	
		(20 ~ 50) kHz	7.8×10^{-4}	
		(50 ~ 100) kHz	1.4×10^{-3}	
		(0.1 ~ 1) V 50 Hz ~ 1 kHz	1.2×10^{-4}	
		(1 ~ 20) kHz	1.3×10^{-4}	
		(20 ~ 50) kHz	1.8×10^{-4}	
		(50 ~ 100) kHz	3.8×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Leakage current testers	40416	(1 ~ 10) V		Calibrator/ SICT-CP-40416	
		50 Hz ~ 1 kHz	1.2×10^{-4}		
		(1 ~ 20) kHz	1.3×10^{-4}		
		(20 ~ 50) kHz	1.8×10^{-4}		
		(50 ~ 100) kHz	3.4×10^{-4}		
		(10 ~ 100) V			
		50 Hz ~ 1 kHz	1.3×10^{-4}		
		(1 ~ 20) kHz	1.2×10^{-4}		
		(20 ~ 50) kHz	3.0×10^{-4}		
		(50 ~ 100) kHz	6.8×10^{-4}		
Output Resistance		(100 ~ 300) V			
		50 Hz ~ 1 kHz	3.1×10^{-4}		
		(1 ~ 10) kHz	3.4×10^{-4}		
Input Resistance		(10 ~ 20) kHz	5.5×10^{-4}		
		(300 ~ 1 000) V			
		50 Hz ~ 10 kHz	3.7×10^{-4}		
Electronic AC/DC loads	40417	100 mΩ	$8.6 \mu\Omega$	Calibrator/ SICT-CP-40417	
		100 mΩ ~ 1 Ω	6.3×10^{-5}		
		1 Ω ~ 10 kΩ	6.2×10^{-5}		
		100 mΩ	$24 \mu\Omega$		
		100 mΩ ~ 100 Ω	2.4×10^{-4}		
		100 Ω ~ 10 kΩ	6.3×10^{-5}		
		100 mV	$6.3 \mu\text{V}$		
		(0.1 ~ 1) V	6.2×10^{-5}		
		(1 ~ 2) V	3.2×10^{-5}		
		(2 ~ 4) V	2.3×10^{-5}		
DC Current		(4 ~ 6) V	1.5×10^{-5}		
		(6 ~ 8) V	1.2×10^{-5}		
		(8 ~ 10) V	1.1×10^{-5}		
		(10 ~ 50) V	3.3×10^{-5}		
		(50 ~ 100) V	1.2×10^{-5}		
		(100 ~ 200) V	3.6×10^{-5}		
		(200 ~ 400) V	2.6×10^{-5}		
		(400 ~ 1 000) V	1.7×10^{-5}		
`		100 mA	$44 \mu\text{A}$		
		100 mA ~ 0.2 A	4.8×10^{-4}		
		(0.2 ~ 0.4) A	4.2×10^{-4}		
		(0.4 ~ 0.6) A	3.8×10^{-4}		
		(0.6 ~ 2) A	3.7×10^{-4}		
		(2 ~ 20) A	8.1×10^{-4}		
		(20 ~ 80) A	4.2×10^{-4}		
		(80 ~ 100) A	4.1×10^{-4}		

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electronic AC/DC loads	AC Voltage	40417	(50 ~ 60) Hz	Calibrator/ SICT-CP-40417
			0.1 V	
			(0.1 ~ 0.2) V	
			(0.2 ~ 0.5) V	
			(0.5 ~ 1) V	
			(1 ~ 10) V	
			(10 ~ 100) V	
			(100 ~ 200) V	
			(200 ~ 500) V	
			2.0 × 10 ⁻⁴	
	AC Current		(50 ~ 60) Hz	
			100 mA	
			100 mA ~ 0.3 A	
			(0.3 ~ 0.4) A	
			(0.4 ~ 0.5) A	
			(0.5 ~ 0.6) A	
			(0.6 ~ 0.8) A	
			(0.8 ~ 2) A	
			(2 ~ 4) A	
			(4 ~ 5) A	
Analogue/Digital multimeters	DC Voltage	40419	(5 ~ 7) A	Calibrator/ SICT-CP-40419
			(7 ~ 9) A	
			(9 ~ 10) A	
			(10 ~ 13) A	
			(13 ~ 16) A	
			(16 ~ 20) A	
			3.2 × 10 ⁻³	
			(±)	
			0 mV	
			(0 ~ 1) mV	
			(1 ~ 2) mV	
			(2 ~ 5) mV	
			(5 ~ 10) mV	
			(10 ~ 20) mV	
			(20 ~ 50) mV	
			(50 ~ 100) mV	
			100 mV ~ 0.2 V	
			(0.2 ~ 0.5) V	
			(0.5 ~ 1) V	
			(1 ~ 2) V	
			(2 ~ 5) V	
			(5 ~ 10) V	
			(10 ~ 20) V	
			(20 ~ 50) V	
			(50 ~ 100) V	
			(100 ~ 200) V	
			(200 ~ 500) V	
			(500 ~ 1 000) V	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters				
AC Voltage	40419	(3 mV) 50 Hz ~ 1 kHz (3 ~ 10) mV 50 Hz ~ 10 kHz (10 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (10 ~ 100) mV 50 Hz ~ 10 kHz (10 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (0.1 ~ 1) V 50 Hz ~ 10 kHz (10 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (1 ~ 10) V 50 Hz ~ 10 kHz (10 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (10 ~ 100) V 50 Hz ~ 20 kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) V 50 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 20) kHz (300 ~ 1 000) V 50 Hz ~ 10 kHz	10 μ V 9.0×10^{-4} 1.3×10^{-3} 1.3×10^{-3} 2.1×10^{-3} 2.5×10^{-4} 2.7×10^{-4} 5.4×10^{-4} 1.4×10^{-3} 1.1×10^{-4} 1.3×10^{-4} 2.0×10^{-4} 4.2×10^{-4} 1.1×10^{-4} 1.3×10^{-4} 2.0×10^{-4} 3.8×10^{-4} 1.2×10^{-4} 3.4×10^{-4} 8.1×10^{-4} 2.3×10^{-4} 2.7×10^{-4} 5.1×10^{-4} 3.6×10^{-4}	Calibrator/ SICT-CP-40419
Resistance		0 Ω (0 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω 100 Ω ~ 1 k Ω (1 ~ 10) k Ω (10 ~ 100) k Ω 100 k Ω ~ 1 M Ω (1 ~ 10) M Ω (10 ~ 100) M Ω	85 μ Ω 1.3×10^{-4} 3.4×10^{-5} 2.0×10^{-5} 1.6×10^{-5} 1.4×10^{-5} 1.7×10^{-5} 2.4×10^{-5} 5.0×10^{-5} 1.3×10^{-4}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters	40419	(±)		Calibrator/ SICT-CP-40419
DC Current		10 μA	10 nA	
		(10 ~ 20) μA	5.2×10^{-4}	
		(20 ~ 50) μA	2.5×10^{-4}	
		(50 ~ 100) μA	1.5×10^{-4}	
		(0.1 ~ 0.2) mA	1.0×10^{-4}	
		(0.2 ~ 0.5) mA	8.1×10^{-5}	
		(0.5 ~ 1) mA	6.8×10^{-5}	
		(1 ~ 2) mA	9.0×10^{-5}	
		(2 ~ 5) mA	8.1×10^{-5}	
		(5 ~ 10) mA	6.8×10^{-5}	
		(10 ~ 20) mA	9.0×10^{-5}	
		(20 ~ 50) mA	9.2×10^{-5}	
		(50 ~ 100) mA	8.0×10^{-5}	
		(0.1 ~ 0.2) A	9.8×10^{-5}	
		(0.2 ~ 0.5) A	1.5×10^{-4}	
		(0.5 ~ 1) A	1.2×10^{-4}	
		(1 ~ 2) A	4.6×10^{-4}	
		(2 ~ 5) A	7.4×10^{-4}	
		(5 ~ 10) A	6.5×10^{-4}	
		(10 ~ 20) A	1.2×10^{-3}	
AC Current		(30 μA)		
		50 Hz ~ 1 kHz	0.16 μA	
		(30 ~ 50) μA		
		50 Hz ~ 1 kHz	3.9×10^{-3}	
		(50 ~ 100) μA		
		50 Hz ~ 1 kHz	2.7×10^{-3}	
		(0.1 ~ 0.2) mA		
		50 Hz ~ 1 kHz	3.6×10^{-5}	
		(0.2 ~ 0.5) mA		
		50 Hz ~ 1 kHz	4.9×10^{-5}	
		(0.5 ~ 1) mA		
		50 Hz ~ 1 kHz	2.4×10^{-4}	
		(1 ~ 2) mA		
		50 Hz ~ 1 kHz	3.0×10^{-4}	
		(2 ~ 5) mA		
		50 Hz ~ 1 kHz	1.4×10^{-4}	
		(5 ~ 10) mA		
		50 Hz ~ 1 kHz	7.2×10^{-5}	
		(10 ~ 20) mA		
		50 Hz ~ 1 kHz	3.0×10^{-4}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters				
AC Current	40419	(20 ~ 50) mA 50 Hz ~ 1 kHz	1.4×10^{-4}	Calibrator/ SICT-CP-40419
		(50 ~ 100) mA 50 Hz ~ 1 kHz	7.2×10^{-5}	
		100 mA ~ 0.2 A 50 Hz ~ 1 kHz	4.5×10^{-4}	
		(0.2 ~ 0.5) A 50 Hz ~ 1 kHz	2.6×10^{-4}	
		(0.5 ~ 1) A 50 Hz ~ 1 kHz	1.3×10^{-4}	
		(1 ~ 2) A 50 Hz ~ 1 kHz	7.6×10^{-4}	
		(2 ~ 5) A 50 Hz ~ 1 kHz	1.7×10^{-3}	
		(5 ~ 10) A 50 Hz ~ 1 kHz	1.4×10^{-3}	
		(10 ~ 15) A 50 Hz ~ 1 kHz	2.1×10^{-3}	
		(15 ~ 20) A 50 Hz ~ 1 kHz	2.0×10^{-3}	
Frequency		10 Hz ~ 1 MHz	6.1×10^{-7}	
Oscilloscopes				
Impedance Measure	40421	50 Ω 75 Ω 1 MΩ	0.7 mΩ 0.8 mΩ 13 Ω	Calibration Generator/ SICT-CP-40421
DC Voltage		1 mV (1 ~ 2) mV (2 ~ 5) mV (5 ~ 10) mV (10 ~ 20) mV (20 ~ 50) mV (50 ~ 100) mV 100 mV ~ 0.5 V (0.5 ~ 1) V (1 ~ 2) V (2 ~ 5) V (5 ~ 10) V (10 ~ 50) V (50 ~ 200) V	1.0 μV 4.9×10^{-4} 2.0×10^{-4} 1.0×10^{-4} 5.5×10^{-5} 2.7×10^{-5} 1.8×10^{-5} 1.6×10^{-5} 9.4×10^{-6} 1.1×10^{-5} 9.2×10^{-6} 8.6×10^{-6} 1.2×10^{-5} 1.5×10^{-5}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Oscilloscopes				Calibration Generator/ SICT-CP-40421
AC Voltage(Square wave)	40421	(1 kHz) 1 mV (1 ~ 2) mV (2 ~ 5) mV (5 ~ 10) mV (10 ~ 20) mV (20 ~ 50) mV (50 ~ 100) mV (100 ~ 200) mV (0.2 ~ 0.5) V (0.5 ~ 120) V	47 μ V 2.4×10^{-2} 1.0×10^{-2} 5.8×10^{-3} 3.5×10^{-3} 2.1×10^{-3} 1.6×10^{-3} 1.4×10^{-3} 1.3×10^{-3} 1.2×10^{-3}	
Time Marker		5 ns 5 ns ~ 10 ms (10 ~ 20) ms 20 ms ~ 5 s	16 fs 3.0×10^{-6} 3.3×10^{-6} 3.3×10^{-5}	
CAL Output Amplitude		(50 Hz ~ 10 kHz) 100 mV (100 ~ 200) mV 200 mV ~ 1.2 V (1.2 ~ 2) V (2 ~ 4) V (4 ~ 12) V	36 μ V 2.2×10^{-4} 1.7×10^{-4} 1.4×10^{-4} 1.2×10^{-4} 1.7×10^{-4}	
CAL Output Frequency		100 Hz ~ 10 MHz	6.5×10^{-7}	
Sine Wave Signal Generator Level		(600 mV) 50 kHz (0.05 ~ 100) MHz (100 ~ 500) MHz (0.5 ~ 1) GHz (1 ~ 3) GHz	25 mV 4.5×10^{-2} 7.2×10^{-2} 1.9×10^{-2} 2.3×10^{-2}	
Volt/Current recorders				Calibrator/ SICT-CP-40424
DC Voltage	40424	(±) 0 μ V 0 μ V ~ 10 mV (10 ~ 100) mV 100 mV ~ 1 V (1 ~ 10) V (10 ~ 100) V (100 ~ 500) V (500 ~ 1 000) V	0.70 μ V 8.1×10^{-5} 1.6×10^{-5} 9.4×10^{-6} 8.6×10^{-6} 1.0×10^{-5} 1.2×10^{-5} 1.1×10^{-5}	
DC Current		(±) 0 μ A (0 ~ 10) μ A (10 ~ 100) μ A 100 μ A ~ 10 mA 10 mA ~ 1 A (1 ~ 10) A (10 ~ 20) A	9.3 nA 9.9×10^{-4} 1.6×10^{-4} 9.2×10^{-5} 2.0×10^{-4} 6.5×10^{-4} 1.2×10^{-3}	

501. Contact thermometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Temperature generators: ovens, furnaces, isothermal liquid baths, ice-point baths, dry-block calibrators	50101	0 °C (-90 ~ 250) °C (250 ~ 500) °C (500 ~ 650) °C (650 ~ 700) °C (700 ~ 1 300) °C	0.01 °C 0.020 °C 0.024 °C 0.028 °C 1.3 °C 2.6 °C	SPRT, STANDARD TC/ SICT-CP-50101
Temperature indicators/recorders /controllers, temperature calibrators temperature indicators/recorders/controllers (With Sensor)	50102	(-45 ~ 0) °C (0 ~ 250) °C (250 ~ 650) °C (650 ~ 900) °C (900 ~ 1 100) °C (1 100 ~ 1 300) °C	0.024 °C 0.031 °C 0.13 °C 1.4 °C 1.5 °C 2.6 °C	SPRT, STANDARD TC/ SICT-CP-50102
(Without Sensor)		(-45 ~ 0) °C (0 ~ 100) °C (100 ~ 200) °C (200 ~ 300) °C (300 ~ 400) °C (400 ~ 500) °C (500 ~ 600) °C (600 ~ 700) °C (700 ~ 800) °C (800 ~ 1 100) °C (1 100 ~ 1 300) °C	0.013 °C 0.018 °C 0.022 °C 0.027 °C 0.031 °C 0.035 °C 0.048 °C 0.052 °C 0.057 °C 0.08 °C 0.09 °C	
Glass thermometers; liquid-in-glass, Beckmann liquid-in-glass	50103	(-45 ~ 0) °C (0 ~ 100) °C (100 ~ 200) °C	0.048 °C 0.058 °C 0.062 °C	SPRT/ SICT-CP-50103
Resistance thermometers: SPRT, IPRT, thermistors, etc. IPRT	50104	(-45 ~ 50) °C (50 ~ 250) °C	0.024 °C 0.028 °C	SPRT, Fixed point/ SICT-CP-50104
Thermal expansion thermometers; bimetal, gas or liquid type bimetal	50105	(-45 ~ 100) °C (100 ~ 300) °C (300 ~ 400) °C (400 ~ 650) °C	0.4 °C 0.6 °C 1.5 °C 3.2 °C	SPRT/ SICT-CP-50105

501. Contact thermometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Thermocouples: noble metal, base metal, pure metal, special type, etc.	50106	(-45 ~ 300) °C (300 ~ 400) °C (400 ~ 500) °C (500 ~ 650) °C (650 ~ 1 000) °C (1 000 ~ 1 100) °C (1 100 ~ 1 200) °C (1 200 ~ 1 300) °C	0.4 °C 0.6 °C 0.7 °C 0.8 °C 1.8 °C 2.0 °C 3.1 °C 3.6 °C	SPRT, Fixed point, STANDARD TC/ SICT-CP-50106
Temperature transducers	50107	(-45 ~ 100) °C (100 ~ 200) °C (200 ~ 250) °C (250 ~ 300) °C (300 ~ 500) °C (500 ~ 650) °C (650 ~ 700) °C (700 ~ 800) °C (800 ~ 1 000) °C (1 000 ~ 1 100) °C (1 100 ~ 1 300) °C	0.05 °C 0.06 °C 0.09 °C 0.14 °C 0.16 °C 0.18 °C 1.4 °C 1.6 °C 1.7 °C 2.8 °C 3.0 °C	SPRT, THERMOCOUPLE, MULTIMETER SICT-CP-50107

503. Humidity

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Relative humidity hygrometers; polymer thinfilm, hair, etc.	50302	(20 ~ 30) % R.H.	1.9 % R.H. 1.6 % R.H. 1.7 % R.H. 1.9 % R.H. 2.1 % R.H.	Dewpoint Meter/ SICT-CP-50302
	Temperature	(-20 ~ 50) °C	0.7 °C	
		(50 ~ 80) °C	0.9 °C	
		(80 ~ 100) °C	1.2 °C	
Temperature humidity recorders; Hygrothermograph, etc	50304	(20 ~ 30) % R.H.	2.4 % R.H. 2.6 % R.H.	Dewpoint Meter/ SICT-CP-50304
		(30 ~ 95) % R.H.		
	Temperature	(-20 ~ 50) °C	1.1 °C	
Transducers; dew-point /relative humidity	50305	(20 ~ 30) % R.H.	1.9 % R.H. 1.6 % R.H. 1.7 % R.H. 1.8 % R.H. 1.9 % R.H. 2.0 % R.H. 2.1 % R.H. 2.2 % R.H.	Dewpoint Meter/ SICT-CP-50305
	Temperature	(-20 ~ 50) °C	0.7 °C	
		(50 ~ 80) °C	0.9 °C	
		(80 ~ 100) °C	1.2 °C	
Humidity generators; two-pressure, two-temperature, flow mixing humidity generator, constant temperature and humidity chamber, etc. ensure/two-temperature/flow mixing humidity (humidity)	50306	(10 ~ 40) % R.H.	1.8 % R.H. 1.6 % R.H. 1.8 % R.H. 1.9 % R.H. 2.1 % R.H. 2.3 % R.H.	Dewpoint Meter/ SICT-CP-50306
	(Temperature)	(-75 ~ 200) °C	0.4 °C	