

Korea Laboratory Accreditation Scheme

CERTIFICATE OF ACCREDITATION

Institute of Calibration & Technology Co., Ltd.

Accreditation No. : KC01-018

Corporation Registration No. : 130111-0052843

Address of Laboratory : 1. 72, Dontancheomdansaneop 2-ro, Hwaseong-si, Gyeonggi-do,
Republic of Korea
1-① 47-22, Jingok Sandan 2beon-ro, Gwangsan-gu,
Gwangju, Republic of Korea

Date of Initial Accreditation : February 23, 2001.

Validity of Accreditation : October 30, 2021. ~ October 29, 2025.

Scope of Accreditation : Attached Annex

Date of issue : February 07, 2024.

This calibration laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to Joint ISO-ILAC-IAF Communiqué).



CHIN CHONGWOOK

Head

Korea Laboratory Accreditation Scheme

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017 & KS Q ISO/IEC 17025:2017

Institute of Calibration Technology Co.,Ltd

72, Dongtancheomdansaneop 2-ro, Hwaseong-si, Gyeonggi-do, Republic of Korea

Phone : 82-31-379-5114, Fax : 82-31-379-5115 e-mail : sictadmin@sict.co.kr

CALIBRATION

Valid To : Oct. 29. 2025

Accreditation No : KC01-018

In recognition of the successful completion of the KOLAS evaluation process,
accreditation is granted to this laboratory to perform the following calibrations

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site
102. Linear dimension			10235	Ultrasonic/coating thickness specimens	N	104. Form		
10201	Balls	N				10401	Form testers	Y
10203	Electrical/Mechanical comparators	Y	10236	Coating thickness testers	Y	10404	Optical flats	N
			10237	Torque arms	N	10405	Optical parallels	N
10204	Gauge block comparators	Y	10238	Width measuring specimens	N	10406	Parallel blocks	Y
10206	Dial/cylinder gauge testers	N	103. Angle			10407	Precision surface plates	Y
10207	Doctor blades	N	10302	Angle gauge blocks	N	10408	Profile gauges	N
10208	Distance meters; electrooptic /laser/ultrasonic	N	10303	Autocollimators	N	10409	Roundness measurement instruments	Y
			10304	Bevel protractors	Y			
10209	End bars	N	10306	Clinometers	N	10410	Form standard specimens	N
10210	Extensometers, linear displacement transducers	Y	10307	Collimators	Y	10411	Roundness standard/roundness magnification standard specimens	N
			10308	Fine angle generators, level comparators	N			
10211	Filler gauges	Y						
10212	Film applicators	N	10310	Indexing tables	N	10412	Straight edges	Y
10213	Gap gauges	N	10311	Plate/square/electric levels	N	10413	Straight rules	N
10214	Gauge blocks, by comparison	N	10312	Auto levels	N	10415	Test bars	N
10216	Height gauges/measuring machines	Y	10314	Penta-prisms	N	10416	Spherometers	N
			10315	Polygons	N	105. Complex geometry		
10219	Linear scales	N	10316	Rotary tables	Y	10501	Base gauges for electric bulb	N
10220	Standard measuring machines	Y	10317	Sine bars, plates, tables, centers	N	10502	Bench centers	Y
10221	Micro scales/Standard scales	N						
10223	Electronic micrometers	Y	10318	Squareness testers, right angle testers	Y	10503	Contact coordinate measuring machines	Y
10224	Height micrometers, riser blocks	N	10319	Cylindrical squares	N	10504	Non-contact coordinate measuring machines	Y
10225	Laser scan micrometers	Y	10320	Precision squares	N			
10227	Standard tape rules, peripheral gauges	N	10321	Theodolites, transits	N	10505	Gauge block accessories	N
			10322	Angular displacement transducers	Y	10508	Hardness indenters	N
10228	Cylindrical plug/pin gauges, Thread measuring wire gauges	Y	10323	Alignment telescopes, line of sight collimators	N	10511	Measuring microscopes, Profile projectors	Y
10229	Radius gauges	N				10512	Micro measuring microscopes	Y
10230	Cylindrical ring gauges	N	10324	Calibration system for survey instruments	Y	10513	Orifice plates	N
10231	Step blocks	N				10514	Taper plug gauges	N
10232	Step gauges	N	10325	Jig transits	N	10515	Taper ring gauges	N
10233	Taper thickness gauges	N	10326	Laser levels	N	10517	Stylus type roughness testers	Y
10234	Ultrasonic thickness gauges	Y	10327	Optical wedges	N			

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site		
10518	Socket gauges for electric bulb	N	202. Force			20706	Hydrometers; density, specific gravity, alcohol, API, baume, sugar, milk, soil, salinity, LPG, etc.	N		
10519	Roughness standard /comparison specimens		20202	Force measuring devices	N					
10525	Thread plug gauges	N	20204	push-pull gauge	Y	20707	Chloride meters	N		
10526	Taper thread plug gauges	N	203. Torque			20799	Others; Solid density	N		
10527	Thread ring gauges	N	20302	Torque measuring devices	Y	208. Viscosity				
10528	Taper thread ring gauges	N	20303	Torque wrenches/drivers	Y	20801	Kinematic viscometers; capillary, etc.	N		
10529	V-blocks, box blocks	N	204. Pressure							
10530	Position gauges	N	20401	Altimeters	Y	20802	Dynamic viscometers; rotaional, etc.	Y		
10531	SEM/TEM/SPM/AFM microscopes	Y	20402	Manometers	Y					
106. Various dimensional			20403	Pneumatic pressure ballances	N	209. Fluid flow				
10601	Inside/outside/gear tooth calipers, caliper gauges	Y	20404	Hydraulic pressure ballances	N	20901	Anemometers; hot-wire	N		
			20405	Air data test systems	Y	20902	Anemometers; pitot tube, etc.	N		
10603	Cylinder/bore gauges	Y	20406	Absolute pressure gauges	Y	20908	Gas flowmeters; differential pressure	N		
10604	Depth gauges, depth micrometers	Y	20407	Blood pressure gauges	Y					
			20408	Compound pressure gauges	Y	20909	Liquid flowmeters; differential pressure	N		
10605	Dial/digital gauges	Y	20409	Differential pressure gauges	Y					
10606	Geodesic baselines	Y	20411	Gauge pressure gauges	Y	20910	Liquid flowmeters; electromagnetic	N		
10608	Grind gauges	N	20412	Pressure transducers/transmitters	Y					
10609	Micro indicators, test indicators	Y			20911	Gas flowmeters; thermal mass, etc.	N			
10610	Micrometer heads	Y	20414	Water depth meters	Y	20912	Liquid flowmeters; Coriolis, etc.	N		
10611	3-points, micrometers	Y	205. Vacuum							
10612	Inside micrometers	Y	20501	Capacitance diaphragm gauges	N	20914	Gas flowmeters; positive displacement	N		
10613	Outside micrometers	Y	20502	Spinning rotor gauges	N					
10615	Particle counters	N	20503	Ionization gauges	N	20915	Liquid flowmeters; positive displacement	N		
10617	Standard sieves	N	20504	Thermal conductivity gauge; pirani, thermocouple, convectron, etc.	N					
10619	Water level meters	N			20916	Gas flowmeters; turbine	N			
10620	Welding gauges	N	20505	Standard leaks, Helium leak detectors	Y	20917	Liquid flowmeters; turbine	N		
201. Mass						20918	Gas flowmeters; ultrasonic	N		
20102	Auto-hopper scale balances	Y	206. Volume			20919	Liquid flowmeters; ultrasonic	N		
20103	Auto-packer scale balances	Y	20601	Volumetric glasswares	N					
20104	Axle weigher balances	N	20602	Pycnometers	N	20920	Gas flowmeters; variable area	N		
20106	Dial platform scale balances	Y	20603	Rain gauges	Y					
20107	Dial swing scale balances	Y	20604	Standard volume vessels	Y	20921	Liquid flowmeters; variable area	N		
20109	Electric balances	Y	20605	Concrete air content meters	N					
20111	Manual swing scale balances	Y	20606	Piston type volume meters	N	20922	Gas flowmeters; vortex	N		
20112	Platform scale balances	Y	207. Density			20923	Liquid flowmeters; vortex	N		
20113	Spring scale balances	Y	20702	Liquid density meters	N	20925	Anemometers; vane, etc	N		
20116	Weights	Y	20704	Salinity meters	N	20999	Others; Anemometers; ultrasonic waves	N		
			20705	Sucrose meters	N					

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site
210. Hardness			402. Resistance, Capacitance and Inductance			404. Other DC & LF measurements		
21001	Brinell hardness testers	Y	40201	Capacitance bridges /indicators	Y	40401	LF amplifiers	Y
21002	Rockwell hardness testers	Y	40202	Decade capacitors	Y	40402	DC/LF attenuators	Y
21003	Shore hardness testers	Y	40204	Standard capacitors	Y	40403	Multimeter calibrators	N
21004	Vickers hardness testers	Y	40205	Earth testers	Y	40404	Oscilloscope calibrators	N
21005	Durometer hardness testers	N	40206	Inductance bridges /indicators	Y	40405	CD/DVD meters/analyzers	Y
21006	Leeb hardness testers	Y	40208	Inductors	Y	40406	Video signal generators	Y
211. Impact			40210	Insulation testers	Y	40407	Audio distortion analyzers /meters	Y
21102	Charpy impact testers	Y	40211	Q-meters	Y	40410	Line frequency meters	Y
21103	Izod impact testers	Y	40213	Resistance bridges & similar instruments	Y	40411	Function generators	Y
30102	Frequency standards	N	40214	Resistance meters	Y	40412	Genescopes	Y
30103	General frequency sources	Y	40215	Resistors	Y	40413	AC/DC high voltage voltmeters	Y
30104	Frequency meters/counters	Y	40217	Impedance bridges/LCR meters	Y	40415	Jitter meters	Y
301. Time/frequency			403. AC voltage, current & power			40416	Leakage current testers	Y
30201	Standard RPM generators	Y	40301	AC ammeters	Y	40417	Electronic AC/DC loads	Y
30202	Contact type tachometers	Y	40302	Clamp ammeters/voltmeters	Y	40418	Modulation meters	Y
30203	Photo tachometers /stroboscopes	Y	40303	AC voltage/current calibrators	Y	40419	Analogue/digital multimeters	Y
30204	Speed meters	Y	40304	Wattmeter calibrators	N	40420	Noise meters	Y
30205	Wow-flutter generators	N	40305	AC current shunts	Y	40421	Oscilloscopes	Y
30206	Wow-flutter meters	Y	40306	Phase angle generators, synchro resolve generators	N	40422	LF phase meters	Y
401. DC Voltage & current			40307	Voltage/current phase angle meters/synchro resolve meters	N	40424	Voltage/current recorders	Y
40101	DC ammeters	Y	40308	Potential transformer test sets	Y	40425	Relay test sets	Y
40102	Transconductance amplifiers	Y	40309	Potential transformer	N	40426	LF signal generators	Y
40103	DC voltage/current calibrators	Y	40310	Power factor meters	Y	40427	LF spectrum analyzers	Y
40104	Electrical temperature calibrators	Y	40311	AC power meters	Y	40428	Spot generators	Y
40105	DC current shunts	Y	40312	AC power supplies	Y	40429	Sweep generators	Y
40106	Galvanometers /null detectors	Y	40313	Puncture/safety testers	Y	40430	Signal transducers	Y
40107	Potentiometers	Y	40314	Power recorders	Y	40432	Transistor curve tracers	Y
40108	DC power supplies	Y	40315	Current transformer test sets	Y	40434	AC/DC high voltage generators	Y
40110	DC voltage dividers	N	40316	Current/turn current coil transformers	N	40435	AC/DC high voltage probes	Y
40111	DC voltage standards	N	40318	AC voltmeters	Y	40436	Logic analyzers	Y
40112	DC voltmeters	Y	40319	Watt hour meters	N	40437	Telephone testers	Y
40113	Static/ionic voltmeters	N	40321	Ratio transformers	N	40438	Video signal analyzers	Y
			40503	Flux meters		405. Low frequency electric & magnetic field		
			40504	Flux sources		40505	Flux meters	N
			40506	Magnetometers		40507	Reference/standard magnets	Y
			40508	Ratio transformers		40509	Reference/standard magnets	N

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site
406. Radio frequency measurement			40652	Field strength meters	Y	50302	Relative humidity hygrometers; polymer thin film, hair, etc.	Y
40601 RF amplifiers	Y	40653	AM/FM test sources	Y				
40602 Coaxial attenuators	Y	40654	Dip simulators	Y				
40605 Burst pulse generators	Y	407. Field strength & antenna				50303	Psychrometers; Assmann ventilated, PRT type, etc.	N
40606 Attenuator calibrators	N	40702	Probes	N				
40607 RF power meter calibrators	Y	40703	Dipole antennas	N				
40608 EMC transducers; current probes, absorbing clamps, etc.	Y	40704	Loop antennas	N		50304	Temperature humidity recorders; hygrothermograph, etc	Y
		40705	Monopole antennas	N				
		40707	Horn antennas	N				
40610 Coaxial directional couplers /splitters	Y	501. Contact thermometry				50305	Transducers; dew-point /relative humidity	N
40613 Electrostatic discharge generators	Y	50101	Temperature generators: ovens, furnaces, isothermal liquid baths, ice-point baths, dry-block calibrators	Y		50306	Humidity generators; two-pressure, two-temperature, flow mixing humidity generator, constant temperature and humidity chamber, etc.	Y
40614 EMC receivers	Y	50102	Temperature indicators /recorders/controllers, temperature calibrators	Y				
40615 RF filters	Y	50103	Glass thermometers; liquid-in-glass, Beckmann	N	504. Moisture			
40616 RF impedance meters	N	50104	Resistance thermometers; SPRT, IPRT, thermistors, etc.	Y	50401	Cereal moisture meters	Y	
40617 RF impulse generators	Y	50105	Thermal expansion thermometers; bimetal, gas or liquid type	Y	50402	Wood moisture meters	N	
40618 Line impedance stabilization networks; LISN, CDN, ISN, etc.	Y	50106	Thermocouples: noble metal, base metal, pure metal, special type, etc.	Y	50403	Paper moisture meters	N	
40619 Coaxial standard mismatches	Y	50107	Temperature transducers	Y	601. Sound in air			
40621 Mobile communication test sets	Y	50108	Primary fixed-point cells and apparatus	N	60102	Sound calibrators	N	
40622 Modulation meters	Y				60104	Microphones	N	
40623 Network analyzers	Y				60106	Sound level meters	Y	
40624 Noise figure meters	Y				603. Vibration			
40625 Noise generators	N				60301	Vibration calibrators	N	
40626 Noise impulse simulators	Y				60302	Vibration transducers	N	
40627 RF phase noise meters	N				60303	Vibration measuring instruments	N	
40628 Coaxial noise sources	N				701. Photometry			
40635 RF power meters	Y				70101	Iluminance meters	N	
40636 Diode power sensors	Y				70102	Luminance meters	N	
40637 Thermocouple power sensors	Y	502. Non contact thermometry			70103	Total luminous flux meters	Y	
40638 Pulse generators	Y	50203	Optical pyrometers	N	70104	Luminous intensity meters	Y	
40639 Radar test sets	Y	50204	Standard radiation thermometers	N	702. Property of detectos & sources			
40640 RF signal generators	Y	50205	Thermal image apparatus	N				
40641 RF spectrum analyzers	Y	50206	Blackbody furnaces	Y	70202	Color temperature meters	Y	
40642 RF speed guns	Y	50207	Others; ear thermometers, etc.	N	70203	Color temperature standard lamps	N	
40643 Surge generators	Y				70204	Colorimeters; source color	Y	
40644 SWR meters	N				70207	Laser power meters	N	
40645 RF terminations	Y	503. Humidity			70208	Standard LED light sources	N	
40646 Coaxial thermistor mounts	Y	50301	Dew-point hygrometers; chilled mirror, alumina thin film, etc.	N				
40650 RF voltmeters	Y							
40651 Vector voltmeters	Y							

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site
70209	Total luminous flux standard lamps	N	703. Property of materials			70415	Optical multimeters	Y
			70301	Colorimeters; material color	Y	70416	Optical network analyzers	Y
70210	Optical detectors	N	70304	Color standard tiles	N	70417	Optical spectrum analyzers	Y
70211	Pyranometers and pyrheliometers	N	70306	Gloss meters	Y	70418	Optical time domain reflectometers; OTDR	Y
			70307	Gloss standard plates	Y			
70213	Display color analyzers; luminance, chromaticity, white balance, etc.	Y	70308	Haze meters	Y	70423	Return loss meters	Y
			70312	Lens meters	Y	70429	Frequency stabilized lasers and LDs	Y
			70315	Optical densitometers	Y			
70214	Luminous intensity standard lamps	N	70319	Reflectance meters	Y	70430	ASE light sources	Y
			70321	Refractometers	Y	70433	Optical power stabilized lasers and LDs	Y
70215	Spectral irradiance standard lamps	N	70323	Transmittance meters	Y	901. Chemical analysis		
			70325	Spectrophotometers including FT-IR spectrophotometers	Y	90101	Breath alcohol analyzers	N
70216	Total spectral radiant flux standard lamps	N	70326	Wavelength reference materials; absorption cell, bandpass filter, etc.	N	90102	Environmental air quality monitoring instruments	Y
						90103	Gas analyzers	Y
70217	Luminance standard sources	N	70326	Wavelength reference materials; absorption cell, bandpass filter, etc.	N	90104	Exhaust gas test instruments	Y
						90199	Others; pH meter, Electrical conductivity meter	Y
70220	Spectral irradiance meters	Y	70402	Broadband light sources	Y			
			70410	Optical attenuators	Y			
70221	Total spectral radiant flux meters	Y	70412	Fiber-optic power meters	Y			
			70413	Optical loss testers	Y			

Note

1. This laboratory provides calibration services in permanent standard laboratory and at on-site.
2. Laboratory conducts on-site calibration should meet requirements of KOLAS-SR-007.
3. On-site calibration is allowed to items with marking 'Y', not allowed to items with marking 'N'.
4. Measurement uncertainty normally is quoted as an expanded uncertainty at a coverage probability of 95%, which usually requires the use of a coverage factor of $k = 2$. It expresses the lowest uncertainty of measurement that can be provided by accredited calibration laboratories in normal conditions.
5. Due to the calibration environment such as reference standards or customers' facilities, it is note that uncertainty of measurement on a calibration certificate may be expressed larger than measurement uncertainty on scope of accreditation in general.

102. Linear dimension

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Balls	10201	(0.3 ~ 100) mm	$\sqrt{0.38^2 + (0.0046 \times l_0)^2}$ μm	Mesuring Machine, Standard/ SICT-CP-10201
Electrical/Mechanical comparators	10203	(0 ~ 5) mm	0.14 μm	Gauge Block/ SICT-CP-10203
Gauge block comparators (Differences of central length)	10204			Gauge Block/ SICT-CP-10204
Comparison		(0 ~ 10) μm	$\sqrt{24^2 + (0.33 \times l_0)^2}$ nm	
Direct		(0 ~ 10) mm	0.043 μm	
Dial/cylinder gauge testers	10206	(0 ~ 100) mm	$\sqrt{0.12^2 + (0.0030 \times l_0)^2}$ μm	Laser Measurement Machine/ SICT-CP-10206
Doctor blades	10207	(0 ~ 10) mm	1.6 μm	Electronic Micrometer/ SICT-CP-10207
Distance meters; electrooptic/laser/ultrasonic	10208	(0 ~ 40) m	$\sqrt{0.28^2 + (0.005 \times l_0)^2}$ mm	Laser interferometer/ SICT-CP-10208
End bars	10209	(25 ~ 1 500) mm	$\sqrt{0.6^2 + (0.0016 \times l_0)^2}$ μm	Linear measuring system/ SICT-CP-10209
Extensometers, linear displacement transducers	10210	(0 ~ 500) mm (500 ~ 1 000) mm	$\sqrt{1.9^2 + (0.0042 \times l_0)^2}$ μm $\sqrt{7.8^2 + (0.058 \times l_0)^2}$ μm	Gauge Block, Laser Measurement Machine/ SICT-CP-10210
Filler gauges	10211	(0 ~ 10) mm	1.2 μm	Mesuring Machine, Standard/ SICT-CP-10211
Film applicators	10212	(0 ~ 10) mm	1.6 μm	Electronic Micrometer/ SICT-CP-10212
Gap gauges	10213	(1 ~ 500) mm	$\sqrt{0.72^2 + (0.0048 \times l_0)^2}$ μm	Gauge Block, contact coordinate measuring machines/ SICT-CP-10213
Gauge blocks, by comparison	10214	(0.1 ~ 100) mm (100 ~ 250) mm (250 ~ 500) mm	$\sqrt{70^2 + (1.3 \times l_0)^2}$ nm $\sqrt{80^2 + (0.71 \times l_0)^2}$ nm $\sqrt{152^2 + (0.71 \times l_0)^2}$ nm	Gauge Block Comparator/ SICT-CP-10214
Height gauges/measuring machines	10216	(0 ~ 1 500) mm	$\sqrt{0.68^2 + (0.0035 \times l_0)^2}$ μm	Gauge Block, Step gauge/ SICT-CP-10216
Linear scales	10219	(0 ~ 40) m	$\sqrt{0.03^2 + (0.0027 \times l_0)^2}$ mm	Laser interferometer/ SICT-CP-10219
Standard measuring machines	10220	(0 ~ 500) mm	$\sqrt{0.38^2 + (0.002 \times l_0)^2}$ μm	Gauge Block/ SICT-CP-10220
Micro scales/Standard scales	10221	(0.01 ~ 1 500) mm	$\sqrt{0.17^2 + (0.0003 \times l_0)^2}$ μm	Linear measuring system/ SICT-CP-10221
Electronic micrometers	10223	(0 ~ 5) mm	0.14 μm	Gage Block/ SICT-CP-10223
Height micrometers, riser blocks Block	10224	(0 ~ 600) mm	$\sqrt{1.1^2 + (0.0019 \times l_0)^2}$ μm	Gauge Block Electronic Micrometer/ SICT-CP-10224
Head		(0 ~ 25) mm	1.2 μm	
Laser scan micrometers	10225	(0.5 ~ 85) mm	$\sqrt{0.46^2 + (0.003 \times l_0)^2}$ μm	Cylindrical plug/pin gauge/ SICT-CP-10225
Standard tape rules, peripheral gauges	10227	(0 ~ 40) m (40 ~ 80) m (80 ~ 100) m	$\sqrt{0.22^2 + (0.0046 \times l_0)^2}$ mm $\sqrt{0.25^2 + (0.0046 \times l_0)^2}$ mm $\sqrt{0.34^2 + (0.0046 \times l_0)^2}$ mm	Laser Measurement Machine/ SICT-CP-10227

Note 1. l_0 unit : mm (10208, 10227 unit : m)

102. Linear dimension

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Cylindrical plug/pin gauges, Thread measuring wire gauges	10228	(0.01 ~ 200) mm (ø 0.1 ~ ø 10) mm	$\sqrt{0.42^2 + (0.003 \times l_0)^2} \mu\text{m}$ 0.41 μm	Mesuring Machine, Standard/ SICT-CP-10228
Cylindrical plug/pin gauges Thread measuring wire gauges				Non-contact coordinate measuring machines,
Radius gauges	10229	(0.1 ~ 100) mm	2.8 μm	
Cylindrical ring gauges	10230	(1.0 ~ 100) mm (100 ~ 300) mm	$\sqrt{0.55^2 + (0.003 0 \times l)^2} \mu\text{m}$ $\sqrt{1.0^2 + (0.003 0 \times l)^2} \mu\text{m}$	Mesuring Machine, Standard/ SICT-CP-10230
Step blocks	10231	(0 ~ 300) μm	0.23 μm	Gauge Block/ SICT-CP-10231
Step gauges	10232	(0 ~ 1 510) mm	$\sqrt{0.28^2 + (0.000 95 \times l_0)^2} \mu\text{m}$	Linear measuring system/ SICT-CP-10232
Taper thickness gauges	10233	(0 ~ 60) mm	4.3 μm	Profile Projector/ SICT-CP-10223
Ultrasonic thickness gauges	10234	(2.5 ~ 100) mm	3 μm	Ultrasonic Tester Blocks/ SICT-CP-10234
Ultrasonic/coating thickness specimens coating thickness specimens Plateness Ultrasonic thickness specimens	10235	(0.01 ~ 25) mm	1.9 μm 0.8 μm $\sqrt{0.64^2 + (0.006 \times l_0)^2} \mu\text{m}$	Gauge Block,Mesuring Machine, Standard/ SICT-CP-10235
Coating thickness testers	10236	(0 ~ 25) mm	1.2 μm	Thickness specimens/ SICT-CP-10236
Torque arms Torque arms Wires	10237	(1 ~ 1 500) mm (0 ~ 5) mm	$\sqrt{0.60^2 + (0.006 1 \times l_0)^2} \mu\text{m}$ 1.2 μm	Gauge Block, contact coordinate measuring machines/ SICT-CP-10237
Width measuring specimens	10238	(0 ~ 200) mm (200 ~ 1 000) mm	$\sqrt{1.3^2 + (0.003 4 \times l)^2} \mu\text{m}$ $\sqrt{1.2^2 + (0.005 4 \times l)^2} \mu\text{m}$	Mesuring Machine, contact coordinate measuring machines/ SICT-CP-10237

Note 1. l_0 unit : mm (10208,10227 unit : m)

103. Angle

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Angle gauge blocks	10302	(0 ~ 360)°	0.6"	Indexing tables/ SICT-CP-10302
Autocollimators	10303	(-1 000 ~ 1 000) "	0.5"	Fine angle generators/ SICT-CP-10303
Bevel protractors	10304	(0 ~ 360)° (0 ~ 90)°	0.9' 0.7'	Angle Gauge Block, Coordinate Measur Machine/ SICT-CP-10304
Clinometers	10306	(0 ~ 360)°	3.3"	Rotary tables/ SICT-CP-10306
Collimators	10307	(-30 ~ 30)'	5"	Total station/ SICT-CP-10307
Fine angle generators, level comparators	10308	±1 000"	0.4"	Autocollimators/ SICT-CP-10308
Indexing tables	10310	(0 ~ 360)°	0.5"	Polygons/ SICT-CP-10310
Plate/Square/Electric levels	10311	(0 ~ 516) " (516 ~ 1 000) " (0 ~ 90)° (0 ~ 400) mm	0.4" 1.2" 0.05' 1.8 μm	Fine angle generators, Rotary tables/ SICT-CP-10311
Auto levels	10312	Straightness of Line of Sight Horizontal Angle Automatic level compensation range	0.03 mm 1.3' 5"	Calibration system for survey instrument, Indexing table, Rotary table/ SICT-CP-10312
Penta-prisms	10314	90°	0.6"	Autocollimators/ SICT-CP-10314
Polygons	10315	(0 ~ 360)°	0.4"	Indexing tables/ SICT-CP-10315
Rotary tables	10316	(0 ~ 360)°	0.5"	Polygons/ SICT-CP-10316
Sine bars, plates, tables, centers (Sinebars)	10317	distance,between two roller center parallelism,between two roller parallelism,between flat-two roller (Plates) Center length Flatness Parallelism	$\sqrt{0.36^2 + (0.002 \times L_0)^2} \mu\text{m}$ 0.5 μm 0.6 μm $\sqrt{0.12^2 + (0.028 \times L_0)^2} \mu\text{m}$ 1.0 μm 1.2 μm	Measuring Machine, Standard/ SICT-CP-10317
Squareness testers, right angle testers	10318	(0 ~ 600) mm	2.0 μm	Cylindrical Square, Precision Square/ SICT-CP-10318
Cylindrical squares	10319	(0 ~ 300) mm (300 ~ 600) mm	1.6 μm 1.9 μm	Cylindrical Square/ SICT-CP-10319
Precision squares	10320	Squareness Parallelism	2.9 μm 2.0 μm	contact coordinate measuring machines/ SICT-CP-10320
Theodolites, transits	10321	Straightness of Line of Sight Horizontal Angle Vertical Angle	0.09 mm 1.3" 1.3"	Calibration system for survey instrument, Indexing table/ SICT-CP-10321

103. Angle

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Angular displacement transducers Angle	10322	(0 ~ 360)°	2.9"	Rotary tables/ SICT-CP-10322
Alignment telescopes, line of sight collimators Straightness of Line of Sight Optical micrometer	10323	0 ~ ∞ (0 ~ 1.2) mm	0.022 mm 2 μm	Calibration system for survey instrument, Alignment telescope/ SICT-CP-10323
Calibration system for survey instruments Straightness of Line of Sight Horizontal Angle Vertical Angle	10324	(0.6 ~ ∞) m (0 ~ 360)° (-45 ~ 45)°	0.022 mm 2.1" 2.0"	Total station, Alignment telescope/ SICT-CP-10324
Jig transits Straightness of Line of Sight Horizontal Angle Vertical Angle	10325	(0.6 ~ ∞) m (0 ~ 360)° (-30 ~ 30)°	0.09 mm 1.3" 4.2"	Calibration system for survey instrument, Indexing table/ SICT-CP-10325
Laser levels Difference from absolute horizontal Difference to vertical about absolute horizontal Automatic level compensation range	10326	(0 ~ 4)' (0 ~ 4)' (-10 ~ 10)°	5" 8" 5"	Autocollimator, Rotary table/ SICT-CP-10326
Optical wedges Angular value on the wedge scale	10327	(-30 ~ 30)''	0.7"	Autocollimator, Rotary table/ SICT-CP-10327

104. Form

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Form testers	10401			Form Standard Specimens
Height length		(0 ~ 50) mm	1.0 μm	Gage Block, Angle Gage Block/
Width		(0 ~ 50) mm	1.4 μm	SICT-CP-10401
Angle		15° ~ 120°	2 °	
Optical flats	10404	\varnothing (0 ~ 60) mm \varnothing (60 ~ 100) mm	0.05 μm 0.08 μm	Optical Flat/ SICT-CP-10404
Optical parallels	10405	\varnothing (0 ~ 30) mm \varnothing (0 ~ 30) mm	0.05 μm 0.07 μm	Optical Flat,Gauge block comparator/ SICT-CP-10405
Parallel blocks	10406			Electronic Micrometer/ SICT-CP-10406
Parallelism		(0 ~ 1 000) mm	1.5 μm	
Flatness		(0 ~ 1 000) mm	1.5 μm	
Length Difference		(0 ~ 1 000) mm	2.2 μm	
Precision surface plates	10407			Electronic Level/ SICT-CP-10407
Flatness		(2 000 × 2 000) mm (5 000 × 5 000) mm	2.0 μm 4.8 μm	
Profile gauges	10408			Gage Block/ SICT-CP-10408
Roundness measurement instruments	10409			Roundness Standard Ball/ SICT-CP-10409
Detector accuracy		(0 ~ 20) μm	0.51 μm	
Rotational accuracy of spindle		360°	0.03 μm	
Rotational accuracy of axis		360°	0.04 μm	
Form standard specimens	10410			Standard measuring machine, Non-contact coordinate measuring machine/ SICT-CP-10410
Height		(0 ~ 50) mm	0.9 μm	
Width		(0 ~ 100) mm	$\sqrt{0.59^2 + (0.0079 \times l_0)^2} \mu\text{m}$	
Radius		(2.5 ~ 10) mm	1.0 μm	
Angle		(10 ~ 50) mm (0 ~ 90) °	1.4 μm 0.007 °	
Roundness standard/roundness magnification standard specimens	10411			Roundness Tester/ SICT-CP-10411
Standard specimens		(0 ~ 300) μm	0.52 μm	
Standard ball		360°	0.08 μm	
Straight edges	10412			Electronic levels/ SICT-CP-10412
Straightness		(0 ~ 2 000) mm	1.8 μm	
Parallelism		(0 ~ 2 000) mm	1.8 μm	
Straight rules	10413			LASER INTERFEROMETER/ SICT-CP-10413
Test bars	10415			Roundness Tester, Electronic Micrometer/
Roundness		(0 ~ 400) mm	0.6 μm	
Cylindricity		(0 ~ 400) mm	0.6 μm	SICT-CP-10415
Run-out		(0 ~ 400) mm	1.1 μm	
Spherometers	10416			Gauge Block/ SICT-CP-10416
Note 1. l_0 unit : mm				

105. Complex geometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Base gauges for electric bulb Inside diameter of pass/stop and screw Pitch	10501	(1 ~ 50) mm (0.3 ~ 10) mm	$\sqrt{0.47^2 + (0.0028 \times l_0)^2} \mu\text{m}$ 1.6 μm	Measuring Machine, Standard/ SICT-CP-10501
Bench centers Difference of both center Flatness of both bed Center height difference	10502	(0 ~ 200) mm (200 ~ 500) mm (0 ~ 500) mm (0 ~ 200) mm (200 ~ 500) mm	1.8 μm 3.4 μm 1.5 μm 1.8 μm 3.4 μm	Test Bar/ SICT-CP-10502
Contact coordinate measuring machines	10503	(0 ~ 1 500) mm (0 ~ 600) mm (0 ~ 600) mm	$\sqrt{0.56^2 + (0.0044 \times l_0)^2} \mu\text{m}$ 3.2 μm 1.2 μm	Step Gauge/ SICT-CP-10503
Non-contact coordinate measuring machines	10504	(0 ~ 1 000) mm	$\sqrt{0.43^2 + (0.0034 \times l_0)^2} \mu\text{m}$	Standard Scale/ SICT-CP-10504
Gauge block accessories Round the ministry of Justice A he ministry of Justice, Base block Center point Plane figure Parallelism(triangle edge) Parallelism(equilibrium tide)	10505	(0 ~ 50) mm (0 ~ 50) mm (0 ~ 20) mm (0 ~ 50) mm (0 ~ 300) mm (0 ~ 150) mm	$\sqrt{0.32^2 + (0.0044 \times l_0)^2} \mu\text{m}$ $\sqrt{0.26^2 + (0.0044 \times l_0)^2} \mu\text{m}$ 1.1 μm 0.04 μm 0.4 μm 0.4 μm	Gauge Block/ SICT-CP-10505
Hardness indenters Diameter Radius Angle	10508	(1 ~ 13) mm (0.2 ~ 6) mm (0 ~ 173)°	0.5 μm 1.0 μm 0.012 °	Standard measuring machine, Non-contact coordinate measuring machine/ SICT-CP-10410
Measuring microscopes, Profile projectors Length Magnification Angle	10511	(0 ~ 500) mm (5 ~ 100) 배 (0 ~ 360) °	$\sqrt{0.43^2 + (0.0034 \times l_0)^2} \mu\text{m}$ 0.04 % 0.9°	Standard Scale/ SICT-CP-10511
Micro measuring microscopes	10512	(0 ~ 1) mm (1 ~ 50) mm	0.7 μm 3.0 μm	Standard Scale/ SICT-CP-10512
Orifice plates Inside diameter Thickness	10513	(12.7 ~ 100) mm (100 ~ 300) mm (0 ~ 15) mm	$\sqrt{0.55^2 + (0.0030 \times l_0)^2} \mu\text{m}$ $\sqrt{1.0^2 + (0.0030 \times l_0)^2} \mu\text{m}$ 0.6 μm	Standard measuring machine/ SICT-CP-10513
Taper plug gauges Small end diameter Big end diameter Plane angle Gage height	10514	(2 ~ 200) mm (2 ~ 200) mm (0 ~ 90) ° (2 ~ 200) mm	$\sqrt{1.3^2 + (0.0041 \times l_0)^2} \mu\text{m}$ $\sqrt{1.4^2 + (0.0041 \times l_0)^2} \mu\text{m}$ 5.9° $\sqrt{1.2^2 + (0.0044 \times l_0)^2} \mu\text{m}$	Measuring Machine, Standard/ SICT-CP-10514
Taper ring gauges Small end diameter Big end diameter Plane angle	10515	(5 ~ 200) mm (5 ~ 200) mm (0 ~ 90) °	2.5 μm 2.5 μm 0.006°	contact coordinate measuring machines/ SICT-CP-10515

Note 1. l_0 unit : mm

105. Complex geometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Stylus type roughness testers Roughness parameter(Ra) Roughness parameter(Rz) Mean width(RSm) H,D	10517	(0 ~ 2) μm (2 ~ 10) μm (0 ~ 7) μm (7 ~ 30) μm (0 ~ 300) μm (0 ~ 6) μm (6 ~ 20) μm	9 nm 24 nm 77 nm 0.27 μm 1.3 μm 63 nm 97 nm	Roughness Specimen/ SICT-CP-10517
Socket gauges for electric bulb Outside diameter of pass/stop and screw Pitch	10518	(1 ~ 50) mm (0.3 ~ 10) mm	$\sqrt{0.44^2 + (0.0028 \times l_0)^2}$ μm 1.6 μm	Measuring Machine, Standard/ SICT-CP-10518
Roughness standard/comparison specimens Depth(H) Mean width(RSm) Roughness parameter(Ra) Roughness parameter(Rz)	10519	(0 ~ 6) μm (6 ~ 20) μm (0 ~ 300) μm (0 ~ 2) μm (2 ~ 10) μm (0 ~ 7) μm (7 ~ 30) μm	$\sqrt{(9.6 \times R)^2 + 12^2}$ nm $\sqrt{(8.94 \times R)^2 + 15^2}$ nm $\sqrt{(0.01 \times R)^2 + 0.51^2}$ μm $\sqrt{(10 \times R)^2 + 2.6^2}$ nm $\sqrt{(9.2 \times R)^2 + 7.4^2}$ nm $\sqrt{(29.2 \times R)^2 + 7.4^2}$ nm $\sqrt{(0.025 \times R)^2 + 0.096^2}$ μm	Roughness Tester/ SICT-CP-10519
Thread plug gauges Outside diameter Effective diameter Pitch Half angle	10525	(1 ~ 205) mm (1 ~ 210) mm (0.3 ~ 10) mm (0.5 ~ 45) °	1.7 μm 1.1 μm 1.2 μm 2'	Measuring Machine, Standard/ SICT-CP-10525
Taper thread plug gauges Half angle Pitch Gage length Notch length Small outside diameter Big outside diameter Small effective diameter Big effective diameter	10526	(0 ~ 45) ° (0.3 ~ 6) mm (2 ~ 50) mm (2 ~ 50) mm (2 ~ 200) mm (2 ~ 200) mm (2 ~ 200) mm (2 ~ 200) mm	2' 1.3 μm 2.6 μm 3.6 μm 2.3 μm 4.8 μm 2.9 μm 5.1 μm	Measuring Machine, Standard/ SICT-CP-10526
Thread ring gauges Minor diameter Effective diameter Pitch	10527	(3 ~ 200) mm (3 ~ 200) mm (0.3 ~ 10) mm	1.5 μm 2.3 μm 1.6 μm	Measuring Machine, Standard/ SICT-CP-10527
Taper thread ring gauges Alternation Thickness Notch length	10528	±3 mm (0 ~ 100) mm (0 ~ 100) mm	2.7 μm 2.3 μm 3.2 μm	Measuring Machine, Standard/ SICT-CP-10528
V-blocks, box blocks Plane figure Parallelism Difference of both part	10529	(5 ~ 300) mm (5 ~ 300) mm (5 ~ 300) mm	1.7 μm 2.0 μm 2.8 μm	contact coordinate measuring machines/ SICT-CP-10529
Position gauges Distance Diameter Angle	10530	(0 ~ 1 000) mm (2 ~ 12) mm (12 ~ 200) mm (0 ~ 360) °	$\sqrt{4.9^2 + (0.0054 \times l)^2}$ μm $\sqrt{2.4^2 + (0.0028 \times l)^2}$ μm $\sqrt{3.3^2 + (0.0028 \times l)^2}$ μm 5.2"	Contact coordinate measuring machine/ SICT-CP-10530
SEM/TEM/SPM/AFM microscopes Magnification	10531	(5 ~ 100) × (100 ~ 500 000) ×	0.003 5 0.003 0	Magnification reference specimen/ SICT-CP-10531

Note 1. l_0 unit : mm, R unit : μm

106. Various dimensional

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Inside/outside/gear tooth calipers, caliper gauges	10601			Gauge Block/ SICT-CP-10601
Inside/Outside calipers		(0 ~ 2 000) mm	$\sqrt{8.2^2 + (0.007 \times l_0)^2}$ μm	
Caliper gauges		(0 ~ 300) mm	$\sqrt{3.7^2 + (0.0032 \times l_0)^2}$ μm	
Cylinder/bore gauges	10603			Dial Gauge Tester/ SICT-CP-10603
Cylinder gauges		(0 ~ 1 000) mm	1.0 μm	
Hole gauges		(0.1 ~ 25) mm	3.3 μm	
Depth gauges, depth micrometers	10604			Gauge Block/ SICT-CP-10604
Depth micrometers		(0 ~ 300) mm	$\sqrt{0.86^2 + (0.0034 \times l_0)^2}$ μm	
Depth gauges		(0 ~ 1 000) mm	$\sqrt{5.9^2 + (0.0048 \times l_0)^2}$ μm	
Dial/digital gauges	10605			Dial Gauge Tester/
		(0 ~ 100) mm	$\sqrt{0.21^2 + (0.0082 \times l_0)^2}$ μm	
		(0 ~ 25) mm	$\sqrt{0.59^2 + (0.004 \times l_0)^2}$ μm	
Geodesic baselines	10606		$\sqrt{1.7^2 + 0.0033^2 \times l^2}$ mm	Total station/ SICT-CP-10606
Grind gauges	10608			Electronic micrometer/ SICT-CP-10608
Depth		(0 ~ 1) mm	1.8 μm	
Straightness		(0 ~ 150) mm	2.5 μm	
Micro indicators, test indicators	10609		0.22 μm	Dial Gauge Tester/ SICT-CP-10609
Micrometer heads	10610	(0 ~ 50) mm	0.8 μm	Gauge Block/ SICT-CP-10610
3-points, Micrometers	10611	(2 ~ 200) mm	$\sqrt{1.3^2 + (0.0034 \times l_0)^2}$ μm	Ring Gauge/ SICT-CP-10611
		(200 ~ 300) mm	3 μm	
Inside micrometers	10612			Gauge Block/
Length		(5 ~ 300) mm	$\sqrt{1.1^2 + (0.0042 \times l_0)^2}$ μm	SICT-CP-10612
Accuracy of scale		(25 ~ 500) mm	$\sqrt{1.1^2 + (0.0042 \times l_0)^2}$ μm	
Extension rod		(13 ~ 500) mm	$\sqrt{1.2^2 + (0.0048 \times l_0)^2}$ μm	
Outside micrometers	10613			Gauge Block, cylindrical plug gauges/
Outside micrometers		(0 ~ 25) mm	$\sqrt{0.2^2 + (0.003 \times l_0)^2}$ μm	SICT-CP-10613
V-anvil micrometers		(25 ~ 1 000) mm	$\sqrt{0.83^2 + (0.003 \times l_0)^2}$ μm	
		(1 ~ 85) mm	0.8 μm	
Particle counters	10615			Particle calibration system/ SICT-CP-10615
(Air)		(0.1 ~ 1) μm		
Flow		(0 ~ 100) L/min	0.09 L/min	
Threshold voltage		(0 ~ 10) V	0.42 mV	
Counting efficiency		(0 ~ 110) %	4.1 %	
(Liquid)		(0.05 ~ 25) μm		
Flow		(0 ~ 100) mL/min	1.4 mL/min	
Threshold voltage		(0 ~ 10) V	0.42 mV	

Note 1. l_0 unit : mm (10606 unit : m)

106. Various dimensional

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Standard sieves Sieve opening Wire rod diameter	10617	(0.004 ~ 10) mm (0.004 ~ 130) mm	1.5 μ m 2.4 μ m	Non-contact coordinate measuring machines/ SICT-CP-10617
Water level meters	10619	(0.05 ~ 6.5) m	1.6 mm	Laser distance meter/ SICT-CP-10619
Welding gauges Height or depth Rule Angle	10620	(0 ~ 100) mm (0 ~ 100) mm (0 ~ 90) $^{\circ}$	8.2 μ m 6.0 μ m 0.7'	Non-contact coordinate measuring machine,Gauge Block/ SICT-CP-10620
Optical micrometers Optical axis shift	10621	(0 ~ 1.2) mm (1.2 ~ 5) mm (5 ~ 10) mm	2 μ m 3 μ m 0.03 mm	Standard Scale/ SICT-CP-10621

201. Mass

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Auto-hopper scale balances	20102	(0 ~ 10) kg (10 ~ 50) kg (50 ~ 200) kg (200 ~ 500) kg (500 ~ 1 000) kg	1.2 g 14 g 21 g 49 g 0.10 kg	Hopper Scale Weight/ SICT-CP-20102
Auto-packer scale balances	20103	(0 ~ 10) kg (10 ~ 50) kg (50 ~ 200) kg	0.8 g 7.7 g 16 g	Weight/ SICT-CP-20103
Axle weigher balances Portable	20104	(100 ~ 500) kg (500 ~ 1 000) kg (1 000 ~ 2 000) kg (2 000 ~ 5 000) kg (5 000 ~ 10 000) kg (10 000 ~ 30 000) kg	0.2 kg 0.4 kg 1 kg 5 kg 10 kg 24 kg	Force Calibration Machine/ SICT-CP-20104
Dial platform scale balances	20106	(0 ~ 30) kg (30 ~ 60) kg (60 ~ 100) kg	42 g 0.08 kg 0.21 kg	Weight/ SICT-CP-20106
Dial swing scale balances	20107	(0 ~ 1) kg (1 ~ 10) kg (10 ~ 20) kg (20 ~ 50) kg (50 ~ 100) kg (100 ~ 200) kg (200 ~ 500) kg (500 ~ 1 000) kg (1 000 ~ 2 000) kg (2 000 ~ 5 000) kg	0.96 g 9.6 g 20 g 48 g 96 g 0.23 kg 0.48 kg 0.96 kg 1.9 kg 4.6 kg	Weight/ SICT-CP-20107
Electric balances	20109	(0 ~ 2) g (2 ~ 6) g (6 ~ 20) g (20 ~ 50) g (50 ~ 100) g (100 ~ 200) g (200 ~ 500) g (500 ~ 1 000) g (1 ~ 2) kg (2 ~ 5) kg (5 ~ 10) kg (10 ~ 25) kg (25 ~ 40) kg (40 ~ 60) kg (60 ~ 150) kg (150 ~ 600) kg (600 ~ 1 000) kg (1 000 ~ 2 000) kg (2 000 ~ 5 000) kg	7.0 µg 9.3 µg 14 µg 19 µg 29 µg 0.05 mg 0.10 mg 0.20 mg 0.5 mg 1.0 mg 3 mg 6 mg 16 mg 24 mg 0.30 g 1.2 g 2.0 g 38 g 0.10 kg	Weight/ SICT-CP-20109

201. Mass

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Manual swing scale balances	20111	(0 ~ 50) kg (50 ~ 200) kg (200 ~ 500) kg (500 ~ 1 000) kg	38 g 0.19 kg 0.38 kg 0.94 kg	Weight / SICT-CP-20111
Platform scale balances	20112	(0 ~ 200) kg (200 ~ 500) kg (500 ~ 1 000) kg	46 g 0.09 kg 0.46 kg	Weight / SICT-CP-20112
Spring scale balances	20113	(0 ~ 10) kg (10 ~ 50) kg (50 ~ 100) kg	21 g 0.08 kg 0.21 kg	Weight / SICT-CP-20113
Weights	20116	(1 mg ~ 20 kg) 1 mg 2 mg 5 mg 10 mg 20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg (20 ~ 100) kg 50 kg 100 kg (100 ~ 200) kg 200 kg (200 ~ 1 000) kg 500 kg 1 000 kg	(Class E2) 1.0 µg 1.0 µg 1.0 µg 1.0 µg 1.2 µg 1.4 µg 1.7 µg 2.3 µg 2.9 µg 3.5 µg 4.6 µg 5.8 µg 7.0 µg 9.3 µg 12 µg 18 µg 36 µg 95 µg 0.12 mg 0.36 mg 0.94 mg 1.8 mg 3.7 mg (Class F2) 0.12 g 0.21 g (Class M1) 1.0 g (Class F2) 1.3 g 2.1 g	Weights, Mass Comparator / SICT-CP-20116

202. Force

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Force measuring devices	20202	(0.4 ~ 20) N (20 ~ 50) N (50 ~ 100) N (100 ~ 200) N (200 ~ 500) N (0.5 ~ 1) kN (1 ~ 2) kN (2 ~ 5) kN (5 ~ 10) kN (10 ~ 20) kN (20 ~ 50) kN (50 ~ 100) kN (100 ~ 200) kN (200 ~ 500) kN (500 ~ 1 000) kN	6.0×10^{-4} 8.4×10^{-5} 8.5×10^{-5} 7.3×10^{-5} 6.9×10^{-5} 9.2×10^{-5} 9.0×10^{-5} 8.4×10^{-5} 8.7×10^{-5} 3.6×10^{-4} 4.1×10^{-4} 4.8×10^{-4} 4.5×10^{-4} 4.3×10^{-4} 4.5×10^{-4}	Force Calibration Machine/ SICT-CP-20202
Tension/compression testing machines	20203	0.1 N ~ 2 kN (2 ~ 5) kN (5 ~ 10) kN (10 ~ 20) kN (20 ~ 50) kN (50 ~ 100) kN (100 ~ 300) kN	1.2×10^{-3} 1.3×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.4×10^{-3}	Weights, Force Measuring Device/ SICT-CP-20203
push-pull gauge	20204	(0.02 ~ 0.2) N 0.2 N ~ 2 kN (2 ~ 5) kN	1.4×10^{-2} 1.3×10^{-3} 8.4×10^{-4}	Weights, Force Calibration Machine/ SICT-CP-20204

203. Torque

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Torque measuring devices	20302	(0.001 ~ 1) N·m	3.3×10^{-3}	Torque Calibration Machine/ SICT-CP-20302
		(1 ~ 10) N·m	4.1×10^{-4}	
		(10 ~ 20) N·m	7.5×10^{-4}	
		(20 ~ 50) N·m	2.3×10^{-4}	
		(50 ~ 100) N·m	3.4×10^{-4}	
		(100 ~ 200) N·m	2.1×10^{-4}	
		(200 ~ 500) N·m	2.2×10^{-4}	
		(500 ~ 1 000) N·m	1.5×10^{-4}	
		(1 000 ~ 2 000) N·m	1.6×10^{-4}	
Torque wrenches/drivers	20303	(0.02 ~ 0.1) N·m	1.4×10^{-2}	Torque Measuring Device/ SICT-CP-20303
		(0.1 ~ 0.5) N·m	9.5×10^{-3}	
		(0.5 ~ 1) N·m	7.8×10^{-3}	
		(1 ~ 2) N·m	6.2×10^{-3}	
		(2 ~ 5) N·m	4.6×10^{-3}	
		(5 ~ 10) N·m	4.5×10^{-3}	
		(10 ~ 20) N·m	4.7×10^{-3}	
		(20 ~ 50) N·m	4.5×10^{-3}	
		(50 ~ 100) N·m	4.9×10^{-3}	
		(100 ~ 200) N·m	3.8×10^{-3}	
		(200 ~ 500) N·m	3.7×10^{-3}	
		(500 ~ 1 000) N·m	3.8×10^{-3}	
		(1 000 ~ 2 000) N·m	2.8×10^{-3}	

204. Pressure

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Altimeters	20401	(-4 000 ~ 20 000) m (20 000 ~ 47 000) m	12 m 15 m	Digital Manometer, Air Dead Weight Tester/ SICT-CP-20401
Manometers	20402	(0 ~ 200) kPa	2.9×10^{-3}	Digital Manometer, Air Dead Weight Tester/ SICT-CP-20402
Pneumatic pressure balances	20403	(5 ~ 7 000) kPa	4.2×10^{-5}	Dead Weight Tester/ SICT-CP-20403
Hydraulic pressure balances	20404	(0.2 ~ 20) MPa (20 ~ 200) MPa	6.4×10^{-5} 7.0×10^{-5}	Dead Weight Tester/ SICT-CP-20404
Air data test systems	20405			Digital Manometer, Air Dead Weight Tester/ SICT-CP-20405
Static pressure		(-2 500 ~ 20 000) m (20 000 ~ 30 500) m	0.8 m 7 m	
Dynamic pressure		(0 ~ 342) km/hr (342 ~ 2 122) km/hr	0.1 km/hr 0.3 km/hr	
Absolute pressure gauges	20406			Digital Manometer, Air Dead Weight Tester/ SICT-CP-20406
Dial, digital		(5 ~ 350) kPa abs (350 ~ 7 000) kPa abs (7 ~ 200) MPa abs	4.0×10^{-5} 4.3×10^{-5} 7.4×10^{-5}	
Blood pressure gauges	20407	(0 ~ 40) kPa	0.01 kPa	Digital Manometer/ SICT-CP-20407
Compound pressure gauges	20408	(-95 ~ 7 000) kPa	4.4×10^{-5}	Air Dead Weight Tester/ SICT-CP-20408
Differential pressure gauges	20409			Digital Manometer, Air Dead Weight Tester/ SICT-CP-20409
		(0 ~ 500) Pa (500 ~ 5 000) Pa (5 ~ 350) kPa (350 ~ 7 000) kPa	0.10 Pa 1.0 Pa 4.2×10^{-5} 4.6×10^{-5}	
Gauge pressure gauges	20411			Digital Manometer, Air Dead Weight Tester, Oil Dead Weight Tester/ SICT-CP-20411
		(0 ~ 500) Pa (500 ~ 5 000) Pa (5 ~ 350) kPa (350 ~ 7 000) kPa (7 ~ 200) MPa (200 ~ 500) MPa	0.10 Pa 1.0 Pa 4.2×10^{-5} 4.6×10^{-5} 8.6×10^{-5} 2.6×10^{-4}	
Pressure transducers/transmitters	20412			Digital Manometer, Air Dead Weight Tester, Oil Dead Weight Tester/ SICT-CP-20412
Absolute pressure		(5 ~ 5 000) kPa abs (5 ~ 200) MPa abs	2.2×10^{-4} 2.4×10^{-4}	
Gauge pressure		(0 ~ 500) Pa (500 ~ 5 000) Pa (5 ~ 5 000) kPa (5 ~ 500) MPa	0.10 Pa 1.0 Pa 2.2×10^{-4} 2.4×10^{-4}	
Dial type vacuum gauges	20413	(-95 ~ 0) kPa	0.059 kPa	Air Dead Weight Tester, SICT-CP-20413
Water depth meters	20414	(0 ~ 198.12) m (198.12 ~ 350.52) m	0.062 m 0.46 m	Digital Manometer/ SICT-CP-20414

205. Vacuum

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Capacitance diaphragm gauges	20501	(0.133 ~ 13.3) Pa abs (13.3 ~ 133.3) Pa abs (133.3 ~ 1 333) Pa abs (1.333 ~ 13.332) kPa abs (13.332 ~ 133.322) kPa abs	0.051 Pa abs 0.49 Pa abs 1.5 Pa abs 12 Pa abs 17 Pa abs	Baratron gauge, SRG / SICT-CP-20501
Spinning rotor gauges	20502	0.15 mPa abs ~ 0.01 Pa abs	3.4×10^{-2}	Baratron gauge, SRG / SICT-CP-20502
Ionization gauges	20503	0.093 mPa abs ~ 0.15 mPa abs 0.15 mPa abs ~ 0.01 Pa abs	6.0×10^{-2} 3.5×10^{-2}	Baratron gauge, SRG, Ion / SICT-CP-20503
Thermal conductivity gauges; pirani, thermocouple, convectron, etc.	20504	(0.133 ~ 13.3) Pa abs (13.3 ~ 133.3) Pa abs (133.3 ~ 1 333) Pa abs (1.333 ~ 13.332) kPa abs (13.332 ~ 133.322) kPa abs	0.051 Pa abs 0.49 Pa abs 1.5 Pa abs 13 Pa abs 17 Pa abs	Baratron gauge, SRG / SICT-CP-20504
Standard leaks, Helium leak detectors	20505	22.0 $\mu\text{Pa m}^3/\text{s}$ 1.60 $\mu\text{Pa m}^3/\text{s}$ 0.51 $\mu\text{Pa m}^3/\text{s}$ 15.0 nPa m^3/s 6.4 nPa m^3/s 0.24 nPa m^3/s	4.8 $\mu\text{Pa m}^3/\text{s}$ 0.38 $\mu\text{Pa m}^3/\text{s}$ 0.098 $\mu\text{Pa m}^3/\text{s}$ 3.2 nPa m^3/s 1.3 nPa m^3/s 0.049 nPa m^3/s	Standard leaks, Helium leak detectors / SICT-CP-20505

206. Volume

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Volumetric glasswares	20601	(0 ~ 0.1) ml (0.1 ~ 0.5) ml (0.5 ~ 1) ml (1 ~ 2) ml (2 ~ 5) ml (2 ~ 10) ml (10 ~ 25) ml (25 ~ 50) ml (50 ~ 100) ml (100 ~ 200) ml (200 ~ 250) ml (250 ~ 500) ml (500 ~ 1 000) ml (1 000 ~ 2 000) ml (2 000 ~ 5 000) ml (5 000 ~ 10 000) ml	0.31 μ l 0.40 μ l 0.43 μ l 1.1 μ l 1.7 μ l 2.5 μ l 4.3 μ l 5.1 μ l 7.9 μ l 13 μ l 43 μ l 69 μ l 92 μ l 0.17 ml 0.49 ml 0.87 ml	Weight, balances / SICT-CP-20601
Pycnometers	20602	(0 ~ 50) ml (50 ~ 100) ml (100 ~ 250) ml (250 ~ 500) ml	2.4 μ l 4.4 μ l 10 μ l 21 μ l	Weight, balances / SICT-CP-20602
Rain gauges	20603	tipping bucket type : Rainfall intensity : (5 ~ 300) mm/h (0.1 ~ 1) mm standard type : (0.1 ~ 10) mm (10 ~ 50.8) mm (50.8 ~ 178) mm	1.2×10^{-3} 0.021 mm 0.054 mm 0.30 mm	Weight, balances / SICT-CP-20603
Standard volume vessels	20604	(0 ~ 20) L (20 ~ 200) L (200 ~ 10 000) L	9.0×10^{-5} 1.3×10^{-4} 1.1×10^{-3}	Balances, Master Meter, Standard volume vessel/ SICT-CP-20604
Concrete air content meters	20605	(0 ~ 10) %	0.032 %	Weight, balances / SICT-CP-20605
Piston type volume meters	20606	(0 ~ 1) μ l (1 ~ 2) μ l (2 ~ 5) μ l (5 ~ 10) μ l (0.01 ~ 0.02) ml (0.02 ~ 0.05) ml (0.05 ~ 0.1) ml (0.1 ~ 0.2) ml (0.2 ~ 0.5) ml (0.5 ~ 1) ml (1 ~ 2) ml (2 ~ 5) ml (5 ~ 10) ml (10 ~ 20) ml (20 ~ 50) ml (50 ~ 100) ml	0.004 μ l 0.005 μ l 0.006 μ l 0.008 μ l 0.018 μ l 0.035 μ l 0.047 μ l 0.14 μ l 0.35 μ l 0.65 μ l 1.6 μ l 1.9 μ l 2.4 μ l 5.0 μ l 14 μ l 64 μ l	Weight, balances / SICT-CP-20606

207. Density

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Liquid density meters	20702	(0.7 ~ 1.4) g/cm ³	0.000 078 g/cm ³	Density standard materials/ SICT-CP-20702
Salinity meters	20704	(0 ~ 10) % (10 ~ 26) %	0.004 % 0.007 %	NaCl/ SICT-CP-20704
Sucrose meters	20705	(0 ~ 20) % (20 ~ 60) % (60 ~ 80) %	0.014 % 0.016 % 0.025 %	Sucrose/ SICT-CP-20705
Hydrometers; density, specific gravity, alcohol, API, baume, sugar, milk, soil, salinity, LPG, etc.	20706			
density		(0.600 ~ 0.700) g/cm ³	0.000 035 g/cm ³	Solid density standard material,
		(0.700 ~ 0.800) g/cm ³	0.000 038 g/cm ³	Hydrostatic weighing Apparatus/ SICT-CP-20706-1
		(0.800 ~ 0.900) g/cm ³	0.000 042 g/cm ³	
		(0.900 ~ 1.000) g/cm ³	0.000 046 g/cm ³	
		(1.000 ~ 1.100) g/cm ³	0.000 050 g/cm ³	
		(1.100 ~ 1.200) g/cm ³	0.000 055 g/cm ³	
		(1.200 ~ 1.300) g/cm ³	0.000 059 g/cm ³	
		(1.300 ~ 1.400) g/cm ³	0.000 066 g/cm ³	
		(1.400 ~ 1.500) g/cm ³	0.000 071 g/cm ³	
		(1.500 ~ 1.600) g/cm ³	0.000 075 g/cm ³	
		(1.600 ~ 1.700) g/cm ³	0.000 079 g/cm ³	
		(1.700 ~ 1.800) g/cm ³	0.000 084 g/cm ³	
		(1.800 ~ 1.900) g/cm ³	0.000 088 g/cm ³	
		(1.900 ~ 2.000) g/cm ³	0.000 093 g/cm ³	
specific gravity		(2.000 ~ 2.200) g/cm ³	0.000 25 g/cm ³	
		(2.200 ~ 3.000) g/cm ³	0.000 28 g/cm ³	
		(3.000 ~ 3.600) g/cm ³	0.000 30 g/cm ³	
		(3.600 ~ 4.000) g/cm ³	0.000 32 g/cm ³	
		0.590 ~ 0.700	0.000 068	Solid density standard material,
		0.700 ~ 0.800	0.000 069	Hydrostatic weighing Apparatus/ SICT-CP-20706-2
		0.800 ~ 0.900	0.000 072	
		0.900 ~ 1.000	0.000 075	
		1.000 ~ 1.100	0.000 078	
		1.100 ~ 1.200	0.000 082	
		1.200 ~ 1.300	0.000 086	
		1.300 ~ 1.400	0.000 091	

* 20704, 20705, 20706, 20707 unit % is weight percent.

207. Density

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Hydrometers; density, specific gravity, alcohol, API, baume, sugar, milk, soil, salinity, LPG, etc.	20706			
alcohol (Volume)		(0 ~ 10) % (10 ~ 30) % (30 ~ 40) % (40 ~ 50) % (50 ~ 60) % (60 ~ 70) % (70 ~ 80) % (80 ~ 90) % (90 ~ 100) %	0.039 % 0.043 % 0.038 % 0.030 % 0.025 % 0.023 % 0.020 % 0.019 % 0.017 %	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-3
API		-1 ~ 51 51 ~ 101	0.013 0.014	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-4
Baumé-light		10 ~ 30 30 ~ 40 40 ~ 60 60 ~ 70 70 ~ 100	0.015 0.016 0.018 0.019 0.12	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-5
Baumé - heavy		0 ~ 40 40 ~ 75	0.014 0.013	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-5
sugar		(0 ~ 10) % (10 ~ 90) %	0.018 % 0.017 %	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-6
milk		(15 ~ 20) (20 ~ 40)	0.081 0.082	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-7
Bouyoucos		(-5.0 ~ 60.0) g/L	0.14 g/L	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-8
salinity		(0 ~ 26.4) %	0.025 %	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-9
LPG		(0.50 ~ 0.55) g/cm ³ (0.55 ~ 0.60) g/cm ³ (0.60 ~ 0.65) g/cm ³	0.000 065 g/cm ³ 0.000 066 g/cm ³ 0.000 068 g/cm ³	Solid density standard meterial, Hydrostatic weighing Apparatus/ SICT-CP-20706-10

* 20704, 20705, 20706, 20707 unit % is weight percent.

207. Density

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Hydrometers; density, specific gravity, alcohol, API, baume, sugar, milk, soil, salinity, LPG, etc.	20706			
Twadde11		0 ~ 12 12 ~ 74 74 ~ 102 102 ~ 170 170 ~ 200	0.016 0.059 0.060 0.061 0.062	Solid density standard materal, Hydrostatic weighing Apparatus/ SICT-CP-20706-11
Chloride meters	20707	(0 ~ 0.1) % (0.1 ~ 2.0) %	0.000 2 % 0.001 0 %	Chlorine standard liquid/ SICT-CP-20707
Others; Solid density	20799			
Stainless steel		(1 ~ 2) g (2 ~ 5) g (5 ~ 10) g (10 ~ 20) g (20 ~ 50) g (50 ~ 100) g (100 ~ 500) g	0.004 6 g/cm ³ 0.002 4 g/cm ³ 0.001 0 g/cm ³ 0.000 59 g/cm ³ 0.000 43 g/cm ³ 0.000 37 g/cm ³ 0.000 36 g/cm ³	Solid density standard materal, Hydrostatic weighing Apparatus/ SICT-CP-20706-11
Glass		(1 ~ 2) g (2 ~ 5) g (5 ~ 10) g (10 ~ 20) g (20 ~ 500) g	0.000 46 g/cm ³ 0.000 25 g/cm ³ 0.000 14 g/cm ³ 0.000 12 g/cm ³ 0.000 11 g/cm ³	

* 20704, 20705, 20706, 20707 unit % is weight percent.

208. Viscosity

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Kinematic viscometers; capillary, etc.	20801	(2.5 ~ 100 000) mPa·s	1.3×10^{-2}	Viscosity CRM/ SICT-CP-20801
Dynamic viscometers; rotaional, etc. Viscosity	20802	(2.5 ~ 200 000) mPa·s	1.7×10^{-2}	Viscosity CRM/ SICT-CP-20802

209. Fluid flow

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Anemometers; hot-wire	20901	(0.1 ~ 1.0) m/s (1.0 ~ 2.0) m/s (2.0 ~ 70) m/s	8.7×10^{-2} 8.4×10^{-3} 4.8×10^{-3}	Wind Tunnel, Pitot tube, LDV/ SICT-CP-20901
Anemometers; pitot tube, etc.	20902	(0.1 ~ 1.0) m/s (1.0 ~ 2.0) m/s (2.0 ~ 70) m/s	8.7×10^{-2} 8.4×10^{-3} 4.8×10^{-3}	Wind Tunnel, Pitot tube, LDV/ SICT-CP-20902
Gas flowmeters; differential pressure	20908	(1.2×10^{-5} ~ 0.12) m ³ /h (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters; differential pressure	20909	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Liquid flowmeters; electromagnetic	20910	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters; thermal mass, etc.	20911	(1.2×10^{-5} ~ 0.12) m ³ /h (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters; Coriolis, etc.	20912	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters; positive displacement	20914	(1.2×10^{-5} ~ 0.12) m ³ /h (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters; positive displacement	20915	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters; turbine	20916	(1.2×10^{-5} ~ 0.12) m ³ /h (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters; turbine	20917	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2×10^{-3} 2.6×10^{-3} 7.0×10^{-4}	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters; ultrasonic	20918	(1.2×10^{-5} ~ 0.12) m ³ /h (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929

209. Fluid flow

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Liquid flowmeters; ultrasonic	20919	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2 × 10 ⁻³ 2.6 × 10 ⁻³ 7.0 × 10 ⁻⁴	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters; variable area	20920	(1.2 × 10 ⁻⁵ ~ 0.12) m ³ /h (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9 × 10 ⁻³ 2.0 × 10 ⁻³ 3.6 × 10 ⁻³	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters; variable area	20921	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2 × 10 ⁻³ 2.6 × 10 ⁻³ 7.0 × 10 ⁻⁴	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Gas flowmeters; vortex	20922	(1.2 × 10 ⁻⁵ ~ 0.12) m ³ /h (0.12 ~ 300) m ³ /h (300 ~ 4 000) m ³ /h	1.9 × 10 ⁻³ 2.0 × 10 ⁻³ 3.6 × 10 ⁻³	Sonic Nozzle/SICT-CP-20928 Master Meter/SICT-CP-20929
Liquid flowmeters; vortex	20923	(0.01 ~ 50) m ³ /h (0.000 12 ~ 0.01) m ³ /h (0.01 ~ 50) m ³ /h	1.2 × 10 ⁻³ 2.6 × 10 ⁻³ 7.0 × 10 ⁻⁴	Master Meter/SICT-CP-20926 Weight measuring method/ SICT-CP-20927
Anemometers; vane, etc	20925	(0.1 ~ 1.0) m/s (1.0 ~ 2.0) m/s (2.0 ~ 70) m/s	8.7 × 10 ⁻² 8.4 × 10 ⁻³ 4.8 × 10 ⁻³	Wind Tunnel, Pitot tube, LDV/ SICT-CP-20925
Others: Anemometers; Ultrasonic current meter	20999	(0.1 ~ 1.0) m/s (1.0 ~ 2.0) m/s (2.0 ~ 70) m/s	8.7 × 10 ⁻² 8.4 × 10 ⁻³ 4.8 × 10 ⁻³	Wind Tunnel, Pitot tube, LDV/ SICT-CP-20999

210. Hardness

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Brinell hardness testers	21001	(100 ~ 250) HBW 10/3 000 (250 ~ 450) HBW 10/3 000 (450 ~ 650) HBW 10/3 000	3.1 HBW 10/3 000 4.9 HBW 10/3 000 8.2 HBW 10/3 000	Brinell Hardness CRM/ SICT-CP-21001
Rockwell hardness testers	21002	(20 ~ 70) HRC (20 ~ 100) HRBW (42 ~ 86) HR30N (29 ~ 82) HR30TW	0.45 HRC 0.80 HRBW 0.70 HR30N 1.1 HR30TW	Rockwell Hardness CRM/ SICT-CP-21002
Shore hardness testers	21003	(20 ~ 100) HS	0.9 HS	Shore Hardness CRM/ SICT-CP-21003
Vickers hardness testers	21004	(50 ~ 300) HV 0.2 (300 ~ 600) HV 0.2 (600 ~ 850) HV 0.2 (50 ~ 300) HV 0.3 (300 ~ 600) HV 0.3 (600 ~ 850) HV 0.5 (50 ~ 300) HV 0.5 (300 ~ 600) HV 0.5 (600 ~ 850) HV 1 (50 ~ 300) HV 10 (300 ~ 600) HV 10 (600 ~ 850) HV 10 (300 ~ 600) HV 30 (600 ~ 850) HV 30	6.0 HV 0.2 18 HV 0.2 27 HV 0.2 5.0 HV 0.3 14 HV 0.3 26 HV 0.5 6.0 HV 0.5 15 HV 0.5 20 HV 1 3.0 HV 10 8.0 HV 10 11 HV 10 8.0 HV 30 11 HV 30	Vickers Hardness CRM/ SICT-CP-21004
Durometer hardness testers	21005	(0 ~ 100) HDA (0 ~ 100) HDAM (0 ~ 100) HDAO (0 ~ 100) HDB (0 ~ 100) HDC (0 ~ 100) HDC2 (0 ~ 100) HDCS (0 ~ 100) HDD (0 ~ 100) HDDO (0 ~ 100) HDE (0 ~ 100) HDE2 (0 ~ 100) HDF (0 ~ 100) HDO (0 ~ 100) HDM (0 ~ 100) HDO (0 ~ 100) HD00 (0 ~ 100) HD000 (0 ~ 100) HD000-S	0.4 HDA 0.8 HDAM 0.4 HDAO 0.4 HDB 0.3 HDC 0.6 HDC2 0.3 HDCS 0.3 HDD 0.3 HDDO 0.4 HDE 0.6 HDE2 0.6 HDF 0.6 HDO 0.8 HDM 0.3 HDO 0.4 HD00 0.4 HD000 0.3 HD000-S	Durometer Calibration device/ SICT-CP-21005
Leeb hardness testers	21006	(400 ~ 1 000) HLD	4.6 HLD	Leeb Hardness CRM/ SICT-CP-21006

211. Impact

측정량/장비	분류번호	교정범위	측정불확도 (신뢰수준 약 95 %)	사용표준/측정방법 등
Charpy impact testers	21102	(0.5 ~ 900) J	-	Laser Distance Meter, Electronic Force Measuring Device/ SICT-CP-21102
Izod impact testers	21103	(0.5 ~ 900) J	-	Laser Distance Meter, Electronic Force Measuring Device/ SICT-CP-21103

301. Time/frequency

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Frequency standards Time Base Frequency	30102	(0.1 ~ 10) MHz	1.0×10^{-12}	GPS Receiver, Universal Counter/ SICT-CP-30102
General frequency sources Time Base Frequency	30103	10 kHz ~ 100 MHz	1.0×10^{-12}	GPS Receiver, Universal Counter/ SICT-CP-30103
Frequency meters/counters Time Base Frequency	30104	(1 ~ 10) MHz	1.0×10^{-12}	GPS Receiver, Universal Counter/ SICT-CP-30104
		0.01 Hz ~ 1 Hz	64 pHz	
		1 Hz ~ 60 GHz	6.4×10^{-11}	
Time interval sources Period	30105	1 ns ~ 10 s	6.1×10^{-9}	GPS Receiver, Universal Counter/ SICT-CP-30105
		(1 ~ 100) ns	0.15 ns	
		100 ns ~ 1 ms	1.3 ns	
		1 ms ~ 10 s	2.1 ns	
Time interval meters/stop watches/timers Trigger Voltage	30106	(-5 ~ 5) V	1.2×10^{-4}	Stop Watch Calibrator/ SICT-CP-30106
		(5 ~ 100) ns	6.2×10^{-5} ns	
		(1 ~ 10) MHz	6.2×10^{-11}	
		day	1.1×10^{-7}	
		month	3.6×10^{-7}	
		(-9.95 ~ 9.95) s / day	6.1 ms	
		(-300 ~ 300) s / month	6.2 ms	
		(1 ~ 100) s	5.8×10^{-6}	
		(100 ~ 1 000) s	8.2×10^{-6}	
		(1 000 ~ 10 000) s	5.8×10^{-5}	
		≥ 1	0.58	

302. Velocity & revolution

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Standard RPM generators Revolution Velocity Measurement	30201	(1 ~ 10 000) min ⁻¹	0.004 0 min ⁻¹	GPS Receiver, Synthesizer Function Generator/ SICT-CP-30201
Revolution Velocity Measurement (Centrifuge)		(30 ~ 5 000) min ⁻¹ (5 000 ~ 8 500) min ⁻¹ (8 500 ~ 50 000) min ⁻¹ (50 000 ~ 80 000) min ⁻¹ (80 000 ~ 99 000) min ⁻¹	0.059 min ⁻¹ 0.099 min ⁻¹ 0.59 min ⁻¹ 0.93 min ⁻¹ 1.1 min ⁻¹	
Contact type tachometers Revolution Velocity Measurement	30202	(1 ~ 10) min ⁻¹ (10 ~ 1 000) min ⁻¹ (1 000 ~ 5 000) min ⁻¹	0.10 min ⁻¹ 0.016 min ⁻¹ 0.063 min ⁻¹	GPS Receiver, Tachometer Cal System/ SICT-CP-30202
Photo tachometers/stroboscopes Revolution Velocity Measurement (Photo-tachometer)	30203	(1 ~ 999.99) min ⁻¹ (1 000.0 ~ 99 999.9) min ⁻¹ (100 000 ~ 600 000) min ⁻¹	0.006 1 min ⁻¹ 0.061 min ⁻¹ 0.61 min ⁻¹	GPS Receiver, Photo Signal Detector/ SICT-CP-30203
Revolution Velocity Measurement (Stroboscope)		(30 ~ 9 000) min ⁻¹ (9 000 ~ 90 000) min ⁻¹ (90 000 ~ 500 000) min ⁻¹	0.005 8 min ⁻¹ 0.058 min ⁻¹ 0.58 min ⁻¹	
Speed meters Speed Test	30204	(0 ~ 400) km/h	6.1 × 10 ⁻³ km/h	GPS Receiver, Synthesizer Function Generator/ SICT-CP-30204
Wow-flutter generators Carrier Frequency	30205	10 Hz ~ 99.99 kHz	6.2 × 10 ⁻⁶	GPS Receiver, Universal Counter/ SICT-CP-30205
Function Frequency		1 Hz ~ 10 kHz (10 ~ 30) kHz	6.2 × 10 ⁻⁶ 2.1 × 10 ⁻⁶	
Wow/Flutter Deviation		(1 Hz ~ 100 Hz) (0 ~ 3) %	0.025 %	
Output Level		(1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 6) V	5.8 × 10 ⁻⁴ 1.7 × 10 ⁻⁴ 1.3 × 10 ⁻⁴	
CCIR Pulse		10 ms 30 ms 60 ms 100 ms	1.0 × 10 ⁻² ms 3.0 × 10 ⁻² ms 6.0 × 10 ⁻² ms 1.0 × 10 ⁻¹ ms	
Wow-flutter meters Wow/Flutter Deviation	30206	(0.1 ~ 0.3) % (0.3 ~ 3) %	0.019 % 0.020 %	GPS Receiver, Wow Flutter Calibrator/ SICT-CP-30206
Carrier Frequency		3 kHz 3.15 kHz	6.2 × 10 ⁻⁵ kHz 6.2 × 10 ⁻⁵ kHz	
CCIR Pulse		(10 ~ 100) ms	0.59 %	
Output Voltage		(1 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V	6.8 µV 9.8 µV 76 µV	

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC ammeters	40101	(±) 0 pA (0 ~ 1) pA (1 ~ 10) pA (10 ~ 100) pA (0.1 ~ 100) nA (0.1 ~ 1) μA (1 ~ 10) μA (10 ~ 100) μA (0.1 ~ 1) mA (1 ~ 10) mA (10 ~ 100) mA (100 ~ 200) mA (0.2 ~ 1) A (1 ~ 10) A (10 ~ 100) A	13 fA 2.4×10^{-2} 9.4×10^{-3} 2.3×10^{-3} 8.5×10^{-4} 9.3×10^{-4} 4.5×10^{-5} 3.0×10^{-5} 2.5×10^{-5} 2.7×10^{-5} 2.0×10^{-5} 1.7×10^{-5} 3.0×10^{-5} 1.4×10^{-4} 1.5×10^{-4}	Calibrator/ SICT-CP-40101
Transconductance amplifiers	40102	(±) 10 μA ~ 10 A (10 ~ 50) A (50 ~ 100) A AC Current (10 μA) 10 Hz ~ 10 kHz (10 ~ 100) μA 10 Hz ~ 1 kHz (1 ~ 10) kHz (100 μA ~ 1 mA) 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (1 ~ 100) mA 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (100 mA ~ 1 A) 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (1 ~ 2) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (2 ~ 5) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	1.3×10^{-5} 4.3×10^{-5} 4.4×10^{-5} 2.6×10^{-3} 3.6×10^{-4} 6.4×10^{-4} 9.8×10^{-5} 7.5×10^{-5} 9.4×10^{-5} 7.8×10^{-5} 4.6×10^{-5} 4.2×10^{-5} 8.1×10^{-5} 4.9×10^{-5} 4.4×10^{-5} 7.9×10^{-5} 4.7×10^{-5} 4.5×10^{-5} 8.2×10^{-5} 5.2×10^{-5} 5.0×10^{-5}	AC-DC Active Current Shunt/ SICT-CP-40102

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Transconductance amplifiers				
AC Current	40102	(5 ~ 10) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 20) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (20 ~ 50) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (50 ~ 100) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz (100 ~ 120) A 60 Hz	8.6×10^{-5} 5.9×10^{-5} 7.8×10^{-5} 9.3×10^{-5} 6.8×10^{-5} 7.8×10^{-5} 1.0×10^{-4} 8.3×10^{-5} 1.1×10^{-4} 1.2×10^{-4} 9.7×10^{-5} 1.3×10^{-5} 4.6×10^{-4}	AC-DC Active Current Shunt/ SICT-CP-40102
DC voltage/current calibrators				
DC Voltage	40103	(±) 0 mV (0 ~ 1) mV (1 ~ 2) mV (2 ~ 5) mV (5 ~ 10) mV (10 ~ 20) mV (20 ~ 50) mV (50 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V	$0.29 \mu\text{V}$ 4.8×10^{-4} 2.4×10^{-4} 9.6×10^{-5} 4.9×10^{-5} 2.5×10^{-5} 1.1×10^{-5} 6.9×10^{-6} 3.7×10^{-6} 3.3×10^{-6} 5.4×10^{-6} 5.7×10^{-6}	Reference Multimeter/ SICT-CP-40103
DC Current		(±) 0 pA (0 ~ 1) pA (1 ~ 10) pA (10 ~ 100) pA (0.1 ~ 100) nA 100 nA ~ 10 A (10 ~ 100) A	9.6 fA 2.1×10^{-2} 6.8×10^{-3} 2.2×10^{-3} 8.5×10^{-4} 1.4×10^{-5} 4.5×10^{-5}	
Electrical temperature calibrators				
TEMPERATURE(SOURCE)	40104			
T/C		(-9.835 ~ 0.000) mV 0.000 mV (0.000 ~ 13.421) mV (13.421 ~ 37.006) mV (37.006 ~ 61.017) mV (61.017 ~ 76.373) mV	$0.42 \mu\text{V}$ $0.24 \mu\text{V}$ $0.42 \mu\text{V}$ $0.48 \mu\text{V}$ $0.53 \mu\text{V}$ $0.57 \mu\text{V}$	디지털 멀티미터/ SICT-CP-40104

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrical temperature calibrators	40104			
RTD		0.999 Ω (0.999 ~ 2.499) Ω (2.499 ~ 4.322) Ω (4.322 ~ 100.000) Ω (100.000 ~ 177.155) Ω (177.155 ~ 313.708) Ω (313.708 ~ 627.422) Ω (627.422 ~ 3 233.3) Ω (±) 0 mV (1 ~ 2) mV (2 ~ 3) mV (3 ~ 4) mV (4 ~ 5) mV (5 ~ 10) mV (10 ~ 50) mV (50 ~ 100) mV (0.1 ~ 0.2) V (0.2 ~ 0.3) V (0.3 ~ 0.6) V (0.6 ~ 1) V (1 ~ 6) V (6 ~ 10) V (10 ~ 70) V (70 ~ 100) V DC Current(SOURCE) (±) 0 mA (0 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 7) mA (7 ~ 10) mA (10 ~ 20) mA (20 ~ 30) mA (30 ~ 40) mA (40 ~ 100) mA Resistance(SOURCE) 0 Ω (0 ~ 0.6) Ω (0.6 ~ 1) Ω (1 ~ 10) Ω (10 ~ 20) Ω (20 ~ 30) Ω (30 ~ 50) Ω (50 ~ 70) Ω (70 ~ 100) Ω (0.1 ~ 0.2) kΩ (0.2 ~ 0.3) kΩ (0.3 ~ 0.5) kΩ (0.5 ~ 0.8) kΩ (0.8 ~ 1) kΩ	0.063 mΩ 3.0×10^{-5} 1.9×10^{-5} 9.9×10^{-6} 8.8×10^{-6} 1.1×10^{-5} 9.2×10^{-6} 1.1×10^{-5} 0.24 μV 4.0×10^{-4} 2.0×10^{-4} 1.3×10^{-4} 1.0×10^{-4} 8.2×10^{-5} 4.2×10^{-5} 6.3×10^{-6} 6.2×10^{-5} 3.1×10^{-5} 3.1×10^{-5} 9.5×10^{-6} 3.1×10^{-5} 9.3×10^{-6} 6.2×10^{-5} 9.2×10^{-6} 0.064 μA 9.2×10^{-5} 6.2×10^{-5} 3.5×10^{-5} 2.3×10^{-5} 1.9×10^{-5} 3.3×10^{-5} 8.2×10^{-5} 7.0×10^{-5} 6.3×10^{-5} 0.061 mΩ 6.1×10^{-4} 8.9×10^{-5} 6.7×10^{-5} 3.3×10^{-5} 2.3×10^{-5} 1.8×10^{-5} 1.4×10^{-5} 1.2×10^{-5} 3.2×10^{-5} 2.3×10^{-5} 1.8×10^{-5} 1.4×10^{-5} 1.1×10^{-5}	디지털 멀티미터/ SICT-CP-40104

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrical temperature calibrators				
Resistance(SOURCE)	40104	(1 ~ 2) kΩ	3.2×10^{-5}	디지털 멀티미터/ SICT-CP-40104
		(2 ~ 3) kΩ	2.3×10^{-5}	
		(3 ~ 5) kΩ	1.8×10^{-5}	
		(5 ~ 8) kΩ	1.4×10^{-5}	
		(8 ~ 10) kΩ	1.1×10^{-5}	
		(10 ~ 20) kΩ	3.2×10^{-5}	
		(20 ~ 30) kΩ	2.4×10^{-5}	
		(30 ~ 40) kΩ	1.9×10^{-5}	
		(40 ~ 50) kΩ	1.6×10^{-5}	
		(50 ~ 100) kΩ	1.1×10^{-5}	
TEMPERATURE(MEASURE)				
T/C		(-9.835 ~ 0.000) mV	0.59 μV	
		0.000 mV	0.50 μV	
		(0.000 ~ 13.422) mV	0.62 μV	
		(13.422 ~ 28.947) mV	0.75 μV	
		(28.947 ~ 45.094) mV	0.88 μV	
		(45.094 ~ 53.113) mV	0.95 μV	
		(53.113 ~ 76.374) mV	1.1 μV	
RTD		0.998 Ω	0.24 mΩ	
		(0.998 ~ 2.496) Ω	1.0×10^{-4}	
		(2.496 ~ 4.315) Ω	7.1×10^{-5}	
		(4.315 ~ 16.994) Ω	3.9×10^{-5}	
		(16.994 ~ 249.580) Ω	3.5×10^{-5}	
		(249.580 ~ 317.988) Ω	4.3×10^{-5}	
		(317.988 ~ 390.474) Ω	4.0×10^{-5}	
		(390.474 ~ 3 233.2) Ω	3.5×10^{-5}	
DC Voltage(MEASURE)		(±)		
		0 mV	0.50 μV	
		(1 ~ 5) mV	5.2×10^{-4}	
		(5 ~ 10) mV	9.3×10^{-5}	
		(10 ~ 100) mV	5.9×10^{-5}	
		(0.1 ~ 0.5) V	6.3×10^{-5}	
		(0.5 ~ 0.8) V	1.3×10^{-5}	
		(0.8 ~ 1) V	1.6×10^{-5}	
		(1 ~ 10) V	6.6×10^{-6}	
		(10 ~ 20) V	9.1×10^{-6}	
		(20 ~ 40) V	7.9×10^{-6}	
		(40 ~ 70) V	6.9×10^{-6}	
		(70 ~ 100) V	6.4×10^{-6}	
		(100 ~ 200) V	7.8×10^{-6}	
		(200 ~ 300) V	2.2×10^{-5}	

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrical temperature calibrators DC Current(MEASURE)	40104	(±) 0 mA (0 ~ 0.1) mA (0.1 ~ 0.2) mA (0.2 ~ 0.3) mA (0.3 ~ 0.7) mA (0.7 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 10) mA (10 ~ 20) mA (20 ~ 30) mA (30 ~ 40) mA (40 ~ 50) mA (50 ~ 100) mA (100 ~ 130) mA	0.062 μ A 6.4×10^{-4} 3.2×10^{-4} 2.2×10^{-4} 1.7×10^{-4} 9.3×10^{-5} 9.9×10^{-5} 7.6×10^{-5} 5.8×10^{-5} 9.9×10^{-5} 8.2×10^{-5} 7.4×10^{-5} 7.0×10^{-5} 6.7×10^{-5} 8.7×10^{-5}	디지털 멀티미터/ SICT-CP-40104
Resistance(MEASURE)		0 Ω (0 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) k Ω (1 ~ 10) k Ω (10 ~ 20) k Ω (20 ~ 30) k Ω (30 ~ 40) k Ω (40 ~ 50) k Ω (50 ~ 100) k Ω	0.098 m Ω 6.4×10^{-5} 1.1×10^{-5} 9.6×10^{-6} 6.5×10^{-5} 6.1×10^{-5} 4.7×10^{-5} 4.0×10^{-5} 4.2×10^{-5} 3.9×10^{-5} 3.4×10^{-5}	
DC current shunts	40105	Resistance 1 μ Ω (0.001 ~ 0.01) m Ω (0.01 ~ 0.2) m Ω (0.2 ~ 1) m Ω (1 ~ 10) m Ω (10 ~ 100) m Ω (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) k Ω (1 ~ 10) k Ω (10 ~ 100) k Ω (0.1 ~ 1) M Ω (1 ~ 10) M Ω (10 ~ 100) M Ω	0.32 n Ω 2.8×10^{-4} 1.8×10^{-4} 1.5×10^{-4} 1.1×10^{-6} 2.8×10^{-6} 8.1×10^{-7} 1.3×10^{-6} 6.7×10^{-7} 6.2×10^{-7} 7.9×10^{-7} 2.0×10^{-6} 1.4×10^{-6} 5.3×10^{-6} 9.8×10^{-6}	Trans Conductance Amplifier/ SICT-CP-40105
Galvanometers/null detectors	40106	DC Voltage (±) (100 ~ 300) μ V (0.3 ~ 1) mV 1 mV ~ 1 000 V	1.4×10^{-2} 1.2×10^{-2} 6.8×10^{-3}	Calibrator/ SICT-CP-40106
Potentiometers	40107	DC Voltage (100 ~ 300) μ V (0.3 ~ 1) mV (1 ~ 3) mV 3 mV ~ 1 000 V	5.7×10^{-3} 2.2×10^{-3} 6.0×10^{-4} 3.0×10^{-4}	Calibrator/ SICT-CP-40107

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC power supplies	40108	(±)		DC Electronics Load/ SICT-CP-40108
		0 mV	5.8 μ V	
		(0 ~ 10) mV	5.8 $\times 10^{-4}$	
		(10 ~ 100) mV	5.8 $\times 10^{-5}$	
		(0.1 ~ 1) V	7.5 $\times 10^{-6}$	
		(1 ~ 10) V	3.3 $\times 10^{-6}$	
		(10 ~ 100) V	7.7 $\times 10^{-6}$	
		(100 ~ 600) V	1.3 $\times 10^{-5}$	
		(600 ~ 1 000) V	6.6 $\times 10^{-5}$	
DC Current	40108	(1 ~ 10) mA	5.8 $\times 10^{-3}$	
		(10 ~ 100) mA	5.9 $\times 10^{-4}$	
		(0.1 ~ 1) A	2.4 $\times 10^{-4}$	
		(1 ~ 10) A	3.1 $\times 10^{-4}$	
		(10 ~ 300) A	2.4 $\times 10^{-4}$	
		(300 ~ 500) A	2.6 $\times 10^{-4}$	
		(500 ~ 1 000) A	4.7 $\times 10^{-5}$	
		(1 000 ~ 3 000) A	5.1 $\times 10^{-4}$	
Load regulation	40108	(0 ~ 2) mV	0.16 mV	
		(2 ~ 20) mV	7.8 $\times 10^{-2}$	
		(20 ~ 200) mV	8.2 $\times 10^{-3}$	
Ripple	40108	(0.1 ~ 0.4) mV	3.8 $\times 10^{-1}$	
		(0.4 ~ 0.6) mV	1.1 $\times 10^{-1}$	
		(0.6 ~ 1) mV	7.3 $\times 10^{-2}$	
		(1 ~ 10) mV	4.4 $\times 10^{-2}$	
		(10 ~ 50) mV	7.1 $\times 10^{-2}$	
DC voltage dividers	40110	(±)		Calibrator/ SICT-CP-40110
		(0.01 ~ 1 000) V	4.5 $\times 10^{-6}$	
		(1 ~ 50) kV	8.8 $\times 10^{-5}$	
		(50 ~ 100) kV	8.4 $\times 10^{-5}$	
DC voltage standards	40111	1 V	1.6 μ V	Null Detector/ SICT-CP-40111
		1.018 V	0.8 μ V	
		10 V	3.1 μ V	
DC voltmeters	40112	(±)		Calibrator/ SICT-CP-40112
		0 mV	0.51 μ V	
		(0 ~ 1) mV	5.0 $\times 10^{-4}$	
		(1 ~ 2) mV	2.5 $\times 10^{-4}$	
		(2 ~ 5) mV	1.0 $\times 10^{-4}$	
		(5 ~ 10) mV	6.2 $\times 10^{-5}$	
		(10 ~ 20) mV	5.0 $\times 10^{-5}$	
		(20 ~ 50) mV	2.5 $\times 10^{-5}$	
		(50 ~ 80) mV	1.0 $\times 10^{-5}$	
		(80 ~ 100) mV	6.2 $\times 10^{-6}$	
Static/ionic voltmeters	40113	(0.1 ~ 1 000) V	8.0 $\times 10^{-6}$	DC Power Supply/ SICT-CP-40113
		(±)		
		0 V	68 mV	
		0 V ~ 50 kV	1.3 $\times 10^{-2}$	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Capacitance bridges/indicators	40201			Standard Capacitance Set/ SICT-CP-40201
Frequency		50 Hz ~ 100 MHz	6.8×10^{-8}	
Capacitance		(1 pF)		
		50 Hz ~ 1 kHz	3.6×10^{-4}	
		1 kHz ~ 1 MHz	3.7×10^{-4}	
		2 MHz	4.2×10^{-4}	
		3 MHz	5.4×10^{-4}	
		4 MHz	7.2×10^{-4}	
		5 MHz	9.5×10^{-4}	
		10 MHz	2.5×10^{-3}	
		13 MHz	3.7×10^{-3}	
		(10 pF)		
		50 Hz ~ 5 MHz	3.6×10^{-4}	
		10 MHz	3.7×10^{-4}	
		13 MHz	3.9×10^{-4}	
		(100 pF)		
		(50 ~ 120) Hz	3.5×10^{-4}	
		120 Hz ~ 4 MHz	3.6×10^{-4}	
		5 MHz	3.8×10^{-4}	
		10 MHz	4.9×10^{-4}	
		13 MHz	6.1×10^{-4}	
		(1 000 pF)		
		50 Hz ~ 1 MHz	3.6×10^{-4}	
		2 MHz	3.8×10^{-4}	
		3 MHz	4.5×10^{-4}	
		4 MHz	5.7×10^{-4}	
		5 MHz	7.2×10^{-4}	
		10 MHz	2.0×10^{-3}	
		13 MHz	3.0×10^{-3}	
		(10 nF)		
		(50 ~ 100) Hz	3.0×10^{-4}	
		100 Hz ~ 100 kHz	8.2×10^{-5}	
		(100 nF)		
		(50 ~ 100) Hz	3.0×10^{-4}	
		100 Hz ~ 100 kHz	8.2×10^{-5}	
		(1 μF)		
		(50 ~ 100) Hz	7.0×10^{-4}	
		100 Hz ~ 10 kHz	8.2×10^{-5}	
		(10 ~ 100) kHz	1.1×10^{-4}	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Capacitance bridges/indicators Capacitance	40201	(10 μ F) 120 Hz ~ 1 kHz	1.3×10^{-3}	Standard Capacitance Set/ SICT-CP-40201
		(100 μ F) 120 Hz	1.4×10^{-3}	
		(1 mF) 120 Hz	1.5×10^{-3}	
		(3 mF) 120 Hz	1.5×10^{-3}	
		(10 mF) 120 Hz	1.5×10^{-3}	
		(30 mF) 120 Hz	2.9×10^{-3}	
		(50 ~ 60) Hz		
		100 pF	7.7×10^{-5}	
		1 000 pF	3.9×10^{-5}	
		$\tan \delta$	1.7×10^{-4}	SICT-CP-40201-1
Decade capacitors Capacitance	40202	0.000 0 ~ 0.001 0		
		0.001 0 ~ 0.003 0		
		0.003 0 ~ 0.005 0		
		0.005 0 ~ 0.008 0		
		0.008 0 ~ 0.010 0		
		0.010 0 ~ 0.030 0		
		0.030 0 ~ 0.050 0		
		0.050 0 ~ 0.080 0		
		0.080 0 ~ 0.100 0		
		(50 Hz ~ 20 kHz)		
		1 pF	5.7×10^{-5}	Standard Capacitance Set/ SICT-CP-40202
		(1 ~ 10) pF	4.6×10^{-5}	
		(10 ~ 100) pF	3.8×10^{-5}	
		(100 ~ 1 000) pF	4.6×10^{-5}	
		1 000 pF ~ 100 nF	2.9×10^{-4}	
		100 nF ~ 1 μ F	5.1×10^{-4}	
		(1 kHz)		
		1 pF	2.5×10^{-5}	
		(1 ~ 1 000) pF	2.4×10^{-5}	
		1 000 pF ~ 100 nF	5.5×10^{-5}	
		100 nF ~ 1 μ F	9.3×10^{-5}	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Standard capacitors	40204	(50 Hz ~ 20 kHz) 1 pF (1 ~ 10) pF (10 ~ 100) pF (100 ~ 1 000) pF 1 000 pF ~ 100 nF 100 nF ~ 1 μF (1 kHz) 1 pF (1 ~ 10) pF (10 ~ 100) pF (100 ~ 1 000) pF 1 000 pF ~ 100 nF 100 nF ~ 1 μF (1 pF) 1 kHz 1 kHz ~ 1 MHz (1 ~ 2) MHz (2 ~ 3) MHz (3 ~ 4) MHz (4 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz (1 ~ 10) pF 1 kHz ~ 3 MHz (3 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz (10 ~ 100) pF 1 kHz ~ 1 MHz (1 ~ 3) MHz (3 ~ 4) MHz (4 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz (100 pF ~ 1 nF) 1 kHz 1 kHz ~ 1 MHz (1 ~ 2) MHz (2 ~ 3) MHz (3 ~ 4) MHz (4 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz	5.2 × 10 ⁻⁵ 4.0 × 10 ⁻⁵ 3.0 × 10 ⁻⁵ 4.0 × 10 ⁻⁵ 2.9 × 10 ⁻⁴ 5.1 × 10 ⁻⁴ 9.1 × 10 ⁻⁶ 7.1 × 10 ⁻⁶ 6.1 × 10 ⁻⁶ 7.1 × 10 ⁻⁶ 5.0 × 10 ⁻⁵ 9.0 × 10 ⁻⁵ 2.4 × 10 ⁻⁴ 2.5 × 10 ⁻⁴ 3.3 × 10 ⁻⁴ 4.7 × 10 ⁻⁴ 6.7 × 10 ⁻⁴ 9.1 × 10 ⁻⁴ 2.5 × 10 ⁻³ 3.7 × 10 ⁻³ 2.3 × 10 ⁻⁴ 2.4 × 10 ⁻⁴ 2.6 × 10 ⁻⁴ 2.8 × 10 ⁻⁴ 2.3 × 10 ⁻⁴ 2.4 × 10 ⁻⁴ 2.5 × 10 ⁻⁴ 2.7 × 10 ⁻⁴ 4.0 × 10 ⁻⁴ 5.4 × 10 ⁻⁴ 2.3 × 10 ⁻⁴ 2.4 × 10 ⁻⁴ 2.8 × 10 ⁻⁴ 3.6 × 10 ⁻⁴ 5.0 × 10 ⁻⁴ 6.6 × 10 ⁻⁴ 1.9 × 10 ⁻³ 2.8 × 10 ⁻³	Standard Capacitance Set/ Capacitance Bridge SICT-CP-40204

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Standard capacitors	40204	(1 ~ 100) nF 120 Hz ~ 100 kHz	2.3×10^{-4}	Standard Capacitance Set/ Capacitance Bridge SICT-CP-40204
		(100 nF ~ 1 μ F) 120 Hz	2.4×10^{-4}	
		120 Hz ~ 10 kHz (10 ~ 100) kHz	2.3×10^{-4} 2.4×10^{-4}	
		(1 ~ 10) μ F 120 Hz ~ 1 kHz	1.2×10^{-3}	
		(30 μ F) 120 Hz	1.3×10^{-3}	
		(100 μ F) 120 Hz	1.3×10^{-3}	
		(300 μ F) 120 Hz	1.5×10^{-3}	
		(1 mF) 120 Hz	1.4×10^{-3}	
		(3 mF) 120 Hz	1.5×10^{-3}	
		(10 mF) 120 Hz	1.4×10^{-3}	
		(30 mF) 120 Hz	2.9×10^{-3}	
Earth testers	40205	1 V	6.4×10^{-3}	Decade Resistor/ SICT-CP-40205
		(1 ~ 10) V	6.4×10^{-4}	
		(10 ~ 50) V	1.3×10^{-3}	
		(50 ~ 100) V	6.4×10^{-4}	
		(100 ~ 500) V	1.6×10^{-4}	
		(500 ~ 1 000) V	6.4×10^{-4}	
		1 $\text{m}\Omega$	8.6×10^{-4}	
		(1 ~ 10) $\text{m}\Omega$	7.2×10^{-4}	
		10 $\text{m}\Omega$ ~ 100 k Ω	6.8×10^{-4}	
		1 A	1.2×10^{-3}	
AC Current out	40205	(1 ~ 3) A	1.5×10^{-3}	
		(3 ~ 20) A	9.7×10^{-4}	
		(20 ~ 30) A	1.0×10^{-3}	
		(30 ~ 60) A	8.4×10^{-4}	
		(60 ~ 100) A	1.0×10^{-3}	
		(100 ~ 150) A	4.6×10^{-3}	
		(150 ~ 200) A	3.7×10^{-3}	
		1 s	5.8×10^{-6}	
		(1 ~ 100) s	5.8×10^{-6}	
		(100 ~ 1 000) s	8.2×10^{-6}	
		(1 000 ~ 10 000) s	5.8×10^{-5}	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Inductance bridges/indicators	40206			
Frequency		50 Hz ~ 100 MHz	7.0×10^{-8}	Standard Inductor/ SICT-CP-40206
Inductance		(1 kHz) 100 µH 1 mH 10 mH 100 mH 1 H 10 H	1.9 × 10 ⁻⁴ 1.3 × 10 ⁻⁴	
Inductors	40208			
Standard Inductance		(1 kHz) 100 µH 1 mH 10 mH 100 mH 1 H 10 H	28 nH 0.24 µH 2.4 µH 24 µH 0.24 mH 2.5 mH	Standard Inductor/ SICT-CP-40208
Decade Inductance		(1 kHz) 100 µH ~ 10 H	3.5×10^{-3}	
Insulation testers	40210			High Resistance Decade/ SICT-CP-40210
AC Voltage		1 V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V	8.8×10^{-5} 9.0×10^{-5} 1.0×10^{-4} 1.1×10^{-4}	
Insulation Voltage		1 V (1 ~ 10) V (10 ~ 25) V (25 ~ 50) V (50 ~ 100) V (100 ~ 250) V (250 ~ 500) V (500 ~ 1 000) V (1 000 ~ 5 000) V (5 000 ~ 10 000) V	6.4×10^{-4} 6.4×10^{-5} 2.5×10^{-4} 1.3×10^{-4} 6.4×10^{-5} 2.5×10^{-4} 1.3×10^{-4} 6.4×10^{-5} 6.5×10^{-3} 6.1×10^{-3}	
Insulation Resistance		1 kΩ (1 ~ 10) kΩ (10 ~ 100) kΩ (0.1 ~ 1) MΩ (1 ~ 10) MΩ (10 ~ 100) MΩ (0.1 ~ 1) GΩ (1 ~ 10) GΩ (10 ~ 100) GΩ (0.1 ~ 1) TΩ 10 TΩ	7.1×10^{-5} 3.7×10^{-5} 2.5×10^{-5} 3.1×10^{-5} 9.5×10^{-5} 2.4×10^{-5} 3.1×10^{-5} 6.1×10^{-5} 1.3×10^{-4} 2.6×10^{-4} 6.3×10^{-4}	
Q-meters	40211			Universal Counter/ SICT-CP-40211
Frequency Test		60 Hz ~ 100 MHz	7.0×10^{-8}	
Quality Factor		0 ~ 1 000	9.7×10^{-4}	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistance bridges & similar instruments	40213			Standard Resistance Set/ SICT-CP-40213
Resistance(Rheostat Arm)		1 mΩ (1 ~ 10) mΩ (10 ~ 100) mΩ (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) kΩ (1 ~ 10) kΩ (10 ~ 100) kΩ (0.1 ~ 1) MΩ (1 ~ 10) MΩ (10 ~ 100) MΩ	1.4 × 10 ⁻⁶ 7.5 × 10 ⁻⁷ 4.0 × 10 ⁻⁶ 4.7 × 10 ⁻⁷ 5.1 × 10 ⁻⁷ 4.4 × 10 ⁻⁷ 2.4 × 10 ⁻⁷ 5.1 × 10 ⁻⁷ 1.1 × 10 ⁻⁶ 1.3 × 10 ⁻⁶ 5.3 × 10 ⁻⁶ 1.1 × 10 ⁻⁵	
Resistance(Ratio Arm)		1 mΩ (1 ~ 10) mΩ (10 ~ 100) mΩ (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) kΩ (1 ~ 10) kΩ (10 ~ 100) kΩ (0.1 ~ 1) MΩ (1 ~ 10) MΩ (10 ~ 100) MΩ	1.4 × 10 ⁻⁶ 7.5 × 10 ⁻⁷ 4.0 × 10 ⁻⁶ 4.7 × 10 ⁻⁷ 5.1 × 10 ⁻⁷ 4.4 × 10 ⁻⁷ 2.4 × 10 ⁻⁷ 5.1 × 10 ⁻⁷ 1.1 × 10 ⁻⁶ 1.3 × 10 ⁻⁶ 5.3 × 10 ⁻⁶ 1.1 × 10 ⁻⁵	
Resistance Ratio		0.1 ~ 1.3	33 × 10 ⁻⁹	
Resistance meters	40214			Standard Resistance Set/ SICT-CP-40214
DC Resistance		1 μΩ 5 μΩ 10 μΩ 25 μΩ 100 μΩ 1 mΩ (1 ~ 10) mΩ (10 ~ 100) mΩ (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) kΩ (1 ~ 10) kΩ (10 ~ 100) kΩ (0.1 ~ 1) MΩ (1 ~ 10) MΩ (10 ~ 100) MΩ (0.1 ~ 1) GΩ (1 ~ 10) GΩ (10 ~ 100) GΩ (0.1 ~ 1) TΩ 10 TΩ	8.7 × 10 ⁻⁴ 5.8 × 10 ⁻⁴ 4.8 × 10 ⁻⁴ 5.2 × 10 ⁻⁴ 3.0 × 10 ⁻⁴ 1.2 × 10 ⁻⁶ 7.5 × 10 ⁻⁷ 4.0 × 10 ⁻⁶ 4.7 × 10 ⁻⁷ 5.1 × 10 ⁻⁷ 4.4 × 10 ⁻⁷ 2.4 × 10 ⁻⁷ 5.1 × 10 ⁻⁷ 1.1 × 10 ⁻⁶ 1.3 × 10 ⁻⁶ 5.3 × 10 ⁻⁶ 1.1 × 10 ⁻⁵ 2.5 × 10 ⁻⁴ 7.1 × 10 ⁻⁴ 9.4 × 10 ⁻⁴ 1.5 × 10 ⁻³ 7.1 × 10 ⁻⁴	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistance meters	40214			Standard Resistance Set/ SICT-CP-40214
Frequency		1 kHz	6.8×10^{-8}	
AC Voltage		10 mV (10 ~ 100) mV (0.1 ~ 10) V	2.4×10^{-4} 8.0×10^{-5} 8.2×10^{-5}	
AC Resistance		(1 kHz) 1 mΩ (1 ~ 10) mΩ (10 ~ 100) mΩ (0.1 ~ 1) Ω 1 Ω ~ 10 kΩ (10 ~ 100) kΩ (0.1 ~ 1) MΩ (1 ~ 10) MΩ	5.1×10^{-3} 5.2×10^{-4} 3.3×10^{-4} 1.5×10^{-4} 1.3×10^{-4} 1.6×10^{-4} 3.1×10^{-4} 3.0×10^{-3}	
Resistors	40215			Standard Resistance Set/ SICT-CP-40215
DC Resistance		1 mΩ (1 ~ 10) mΩ (10 ~ 100) mΩ (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) kΩ (1 ~ 10) kΩ (10 ~ 100) kΩ (0.1 ~ 1) MΩ (1 ~ 10) MΩ (10 ~ 100) MΩ (0.1 ~ 1) GΩ (1 ~ 10) GΩ (10 ~ 100) GΩ (0.1 ~ 1) TΩ (1 ~ 10) TΩ (10 ~ 100) TΩ	1.6×10^{-6} 1.1×10^{-6} 2.8×10^{-6} 8.1×10^{-7} 1.3×10^{-6} 6.7×10^{-7} 6.2×10^{-7} 7.9×10^{-7} 2.0×10^{-6} 1.4×10^{-6} 5.2×10^{-6} 9.7×10^{-6} 2.3×10^{-4} 6.9×10^{-4} 9.3×10^{-4} 1.4×10^{-3} 4.1×10^{-3} 7.6×10^{-3}	
AC Resistance		(50 Hz ~ 1 kHz) 1 mΩ (1 ~ 10) mΩ (10 ~ 100) mΩ (0.1 ~ 1) Ω (1 ~ 100) Ω (10 Ω) 1 kHz 1 MHz 2 MHz 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz	1.0×10^{-3} 5.9×10^{-4} 3.9×10^{-4} 2.4×10^{-4} 2.5×10^{-4} 2.4×10^{-4} 4.0×10^{-4} 5.6×10^{-4} 6.5×10^{-4} 7.5×10^{-4} 1.0×10^{-3} 4.0×10^{-3} 6.0×10^{-3}	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistors	40215	(10 ~ 100) Ω		Standard Resistance Set/ SICT-CP-40215
AC Resistance		1 kHz	2.4×10^{-4}	
		1 MHz	4.0×10^{-4}	
		2 MHz	4.8×10^{-4}	
		3 MHz	5.6×10^{-4}	
		4 MHz	5.6×10^{-4}	
		5 MHz	5.6×10^{-4}	
		10 MHz	2.0×10^{-3}	
		13 MHz	3.0×10^{-3}	
		(100 Ω ~ 1 k Ω)		
		1 kHz	2.4×10^{-4}	
		100 kHz	4.0×10^{-4}	
		1 MHz	4.0×10^{-4}	
		2 MHz	4.0×10^{-4}	
		3 MHz	4.0×10^{-4}	
		4 MHz	4.8×10^{-4}	
		5 MHz	5.6×10^{-4}	
		10 MHz	2.0×10^{-3}	
		13 MHz	3.0×10^{-3}	
		(1 ~ 10) k Ω		
		1 kHz	2.4×10^{-4}	
		100 kHz	3.3×10^{-4}	
		1 MHz	4.0×10^{-4}	
		(10 ~ 100) k Ω		
		1 kHz	2.4×10^{-4}	
		100 kHz	4.0×10^{-4}	
		1 MHz	4.0×10^{-4}	
		(100 k Ω ~ 1 M Ω)		
		1 kHz	3.8×10^{-4}	
		(1 ~ 10) M Ω		
		1 kHz	3.0×10^{-3}	
Decade Resistance		0 Ω	$64 \mu\Omega$	
		(0 ~ 10) m Ω	$6.5 \mu\Omega$	
		(10 ~ 100) m Ω	$64 \mu\Omega$	
		(0.1 ~ 1) Ω	$66 \mu\Omega$	
		(1 ~ 7) Ω	3.9×10^{-5}	
		(7 ~ 10) Ω	1.3×10^{-5}	
		(10 ~ 70) Ω	2.0×10^{-5}	
		(70 ~ 100) Ω	9.8×10^{-6}	
		(100 ~ 700) Ω	1.9×10^{-5}	
		(0.7 ~ 1) k Ω	9.6×10^{-6}	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistors	40215	(1 ~ 7) kΩ (7 ~ 10) kΩ (10 ~ 70) kΩ (70 ~ 100) kΩ (100 ~ 600) kΩ (0.6 ~ 1) MΩ (1 ~ 7) MΩ (7 ~ 10) MΩ (10 ~ 100) MΩ (0.1 ~ 1) GΩ (1 ~ 10) GΩ (10 ~ 100) GΩ (0.1 ~ 1) TΩ (1 ~ 10) TΩ	3.5×10^{-5} 1.2×10^{-5} 2.0×10^{-5} 9.8×10^{-6} 2.9×10^{-5} 1.2×10^{-5} 8.0×10^{-5} 2.7×10^{-5} 2.0×10^{-4} 2.5×10^{-4} 7.0×10^{-4} 1.0×10^{-3} 1.5×10^{-3} 4.3×10^{-3}	Standard Resistance Set/ SICT-CP-40215
Impedance bridges/LCR meters	40217	50 Hz ~ 100 MHz	7.0×10^{-8}	Standard Capacitor Set, Standard Resistor Set, Standard Inductor/ SICT-CP-40217
	Frequency	1 mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 10) V (10 ~ 100) V	2.1×10^{-3} 4.4×10^{-4} 8.8×10^{-5} 8.2×10^{-5} 8.9×10^{-5}	
	AC Voltage	(1 pF) 50 Hz ~ 1 kHz 1 kHz ~ 1 MHz 2 MHz 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz (10 pF) 50 Hz ~ 5 MHz 10 MHz 13 MHz (100 pF) 50 Hz ~ 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz	3.5×10^{-4} 3.6×10^{-4} 4.2×10^{-4} 5.4×10^{-4} 7.2×10^{-4} 9.4×10^{-4} 2.5×10^{-3} 3.6×10^{-3} 3.5×10^{-4} 3.7×10^{-4} 3.8×10^{-4} 3.5×10^{-4} 3.6×10^{-4} 3.7×10^{-4} 4.8×10^{-4} 6.0×10^{-4}	
	Capacitance			

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Impedance bridges/LCR meters	40217			Standard Capacitor Set, Standard Resistor Set, Standard Inductor/ SICT-CP-40217
Capacitance		(1 000 pF)		
		50 Hz ~ 1 MHz	3.5×10^{-4}	
		2 MHz	3.8×10^{-4}	
		3 MHz	4.5×10^{-4}	
		4 MHz	5.6×10^{-4}	
		5 MHz	7.2×10^{-4}	
		10 MHz	2.0×10^{-3}	
		13 MHz	2.9×10^{-3}	
		(10 nF)		
		(50 ~ 100) Hz	3.0×10^{-4}	
		100 Hz ~ 100 kHz	8.1×10^{-5}	
		(100 nF)		
		(50 ~ 100) Hz	3.0×10^{-4}	
		100 Hz ~ 100 kHz	8.1×10^{-5}	
		(1 μF)		
		(50 ~ 100) Hz	5.1×10^{-4}	
		100 Hz ~ 10 kHz	8.1×10^{-5}	
		(10 ~ 100) kHz	1.0×10^{-4}	
		(10 μF)		
		120 Hz ~ 1 kHz	1.2×10^{-3}	
		(100 μF)		
		120 Hz	1.3×10^{-3}	
		(1 mF)		
		120 Hz	1.4×10^{-3}	
		(3 mF)		
		120 Hz	1.4×10^{-3}	
		(10 mF)		
		120 Hz	1.4×10^{-3}	
		(30 mF)		
		120 Hz	2.9×10^{-3}	
		(1 pF)		
		1 kHz ~ 1 MHz	0.000 12	
		1 MHz ~ 5 MHz	0.000 23	
		5 MHz ~ 13 MHz	0.000 84	
		(10 pF)		
		1 kHz ~ 13 MHz	0.000 15	
		(100 pF)		
		1 kHz ~ 5 MHz	0.000 13	
		5 MHz ~ 13 MHz	0.000 27	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Impedance bridges/LCR meters	40217			Standard Capacitor Set, Standard Resistor Set, Standard Inductor/ SICT-CP-40217
Dissipation Factor		(1 pF) 1 kHz ~ 1 MHz (1 ~ 5) MHz (5 ~ 13) MHz	0.000 12 0.000 23 0.000 84	
		(10 pF) 1 kHz ~ 13 MHz	0.000 15	
		(100 pF) 1 kHz ~ 5 MHz (5 ~ 13) MHz	0.000 13 0.000 27	
		(1 000 pF) 1 kHz ~ 1 MHz (1 ~ 5) MHz (5 ~ 13) MHz	0.000 12 0.000 24 0.000 86	
		(10 nF) 120 Hz ~ 100 kHz	0.000 46	
		(100 nF) 120 Hz ~ 100 kHz	0.000 58	
		(1 μF) 120 Hz ~ 100 kHz	0.000 81	
Resistance		(1 mΩ) 50 Hz 50 Hz ~ 1 kHz	6.0 × 10 ⁻³ 5.0 × 10 ⁻³	
		(10 mΩ) 50 Hz 50 Hz ~ 1 kHz	1.0 × 10 ⁻³ 5.2 × 10 ⁻⁴	
		(100 mΩ) 50 Hz 50 Hz ~ 1 kHz	7.1 × 10 ⁻⁴ 3.3 × 10 ⁻⁴	
		(1 Ω) 50 Hz 50 Hz ~ 1 kHz	6.8 × 10 ⁻⁴ 1.3 × 10 ⁻⁴	
		(10 Ω) 50 Hz 50 Hz ~ 1 kHz 1 kHz ~ 1 MHz (1 ~ 2) MHz (2 ~ 3) MHz (3 ~ 4) MHz (4 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz	9.1 × 10 ⁻⁴ 1.3 × 10 ⁻⁴ 3.3 × 10 ⁻⁴ 5.2 × 10 ⁻⁴ 6.1 × 10 ⁻⁴ 7.1 × 10 ⁻⁴ 1.0 × 10 ⁻³ 4.0 × 10 ⁻³ 6.0 × 10 ⁻³	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Impedance bridges/LCR meters Resistance	40217	(100 Ω)		Standard Capacitor Set, Standard Resistor Set, Standard Inductor/ SICT-CP-40217
		50 Hz	6.2×10^{-4}	
		50 Hz ~ 1 kHz	1.3×10^{-4}	
		1 kHz ~ 1 MHz	3.3×10^{-4}	
		(1 ~ 2) MHz	4.2×10^{-4}	
		(2 ~ 5) MHz	5.2×10^{-4}	
		(5 ~ 10) MHz	2.0×10^{-3}	
		(10 ~ 13) MHz	3.0×10^{-3}	
		(1 k Ω)		
		1 kHz	1.3×10^{-4}	
Inductance		1 kHz ~ 3 MHz	3.3×10^{-4}	
		(3 ~ 4) MHz	4.2×10^{-4}	
		(4 ~ 5) MHz	5.2×10^{-4}	
		(5 ~ 10) MHz	2.0×10^{-3}	
		(10 ~ 13) MHz	3.0×10^{-3}	
		(10 k Ω)		
		1 kHz	1.3×10^{-4}	
		(1 ~ 100) kHz	2.4×10^{-4}	
		100 kHz ~ 1 MHz	3.3×10^{-4}	
		(100 k Ω)		
DC Bias		1 kHz	1.4×10^{-4}	
		100 kHz ~ 1 MHz	3.3×10^{-4}	
		(1 M Ω)		
		1 kHz	3.0×10^{-4}	
		(10 M Ω)		
		1 kHz	2.9×10^{-3}	
DC Current		(1 kHz)		
		100 μ H	1.9×10^{-4}	
		1 mH	1.3×10^{-4}	
		10 mH	1.3×10^{-4}	
		100 mH	1.3×10^{-4}	
		1 H	1.3×10^{-4}	
DC Bias		10 H	1.3×10^{-4}	
		(\pm)		
		0 μ V	0.68 μ V	
		0 μ V ~ 100 mV	1.1×10^{-5}	
		(0.1 ~ 1) V	7.5×10^{-6}	
		(1 ~ 10) V	7.2×10^{-6}	
DC Current		(10 ~ 100) V	8.2×10^{-6}	
		0 μ A	5.8 μ A	
		0 μ A ~ 200 mA	4.1×10^{-5}	
		(0.2 ~ 2) A	3.6×10^{-5}	
		(2 ~ 20) A	1.9×10^{-4}	
		(20 ~ 100) A	1.4×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC ammeters	40301	(100 μA) 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (100 μA ~ 1 mA) 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (1 ~ 10) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (10 ~ 100) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (100 mA ~ 1 A) 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (1 ~ 10) A (40 ~ 100) Hz 100 Hz ~ 5 kHz (10 ~ 20) A (40 ~ 100) Hz 100 Hz ~ 5 kHz (20 ~ 100) A (40 ~ 100) Hz 100 Hz ~ 5 kHz (100 ~ 200) A 60 Hz	2.2 × 10 ⁻⁴ 4.9 × 10 ⁻⁴ 2.1 × 10 ⁻³ 1.8 × 10 ⁻⁴ 3.8 × 10 ⁻⁴ 2.1 × 10 ⁻³ 1.8 × 10 ⁻⁴ 3.4 × 10 ⁻⁴ 1.9 × 10 ⁻³ 1.7 × 10 ⁻⁴ 3.2 × 10 ⁻⁴ 1.5 × 10 ⁻³ 3.5 × 10 ⁻⁴ 6.7 × 10 ⁻⁴ 8.3 × 10 ⁻³ 2.1 × 10 ⁻⁴ 5.8 × 10 ⁻⁴ 2.0 × 10 ⁻⁴ 5.3 × 10 ⁻⁴ 1.9 × 10 ⁻⁴ 5.9 × 10 ⁻⁴ 8.5 × 10 ⁻⁴	Power Calibrator, Calibrator/ SICT-CP-40301
Clamp ammeters/voltmeters	40302	(100 μA) 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (100 ~ 300) μA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (300 ~ 900) μA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	6.5 × 10 ⁻⁴ 7.8 × 10 ⁻⁴ 2.1 × 10 ⁻³ 3.8 × 10 ⁻⁴ 7.9 × 10 ⁻⁴ 4.0 × 10 ⁻³ 3.1 × 10 ⁻⁴ 6.4 × 10 ⁻⁴ 3.3 × 10 ⁻³	Power Calibrator, Calibrator/ SICT-CP-40302

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters	40302	(900 μ A ~ 1 mA)		Power Calibrator, Calibrator/ SICT-CP-40302
AC Current		40 Hz ~ 1 kHz	6.3×10^{-4}	
		(1 ~ 5) kHz	7.2×10^{-4}	
		(5 ~ 10) kHz	2.1×10^{-3}	
		(1 ~ 3) mA		
		40 Hz ~ 1 kHz	3.8×10^{-4}	
		(1 ~ 5) kHz	7.2×10^{-4}	
		(5 ~ 10) kHz	3.8×10^{-3}	
		(3 ~ 9) mA		
		40 Hz ~ 1 kHz	3.1×10^{-4}	
		(1 ~ 5) kHz	5.8×10^{-4}	
		(5 ~ 10) kHz	3.1×10^{-3}	
		(9 ~ 10) mA		
		40 Hz ~ 1 kHz	6.3×10^{-4}	
		(1 ~ 5) kHz	7.0×10^{-4}	
		(5 ~ 10) kHz	2.0×10^{-3}	
		(10 ~ 30) mA		
		40 Hz ~ 1 kHz	3.6×10^{-4}	
		(1 ~ 5) kHz	6.8×10^{-4}	
		(5 ~ 10) kHz	2.6×10^{-3}	
		(30 ~ 90) mA		
		40 Hz ~ 1 kHz	2.9×10^{-4}	
		(1 ~ 5) kHz	5.4×10^{-4}	
		(5 ~ 10) kHz	2.2×10^{-3}	
		(90 ~ 100) mA		
		40 Hz ~ 1 kHz	6.3×10^{-4}	
		(1 ~ 5) kHz	6.9×10^{-4}	
		(5 ~ 10) kHz	1.6×10^{-3}	
		(100 ~ 300) mA		
		40 Hz	3.3×10^{-4}	
		40 Hz ~ 1 kHz	5.7×10^{-4}	
		(1 ~ 5) kHz	1.2×10^{-3}	
		(5 ~ 10) kHz	9.3×10^{-3}	
		(300 ~ 900) mA		
		40 Hz	5.1×10^{-4}	
		40 Hz ~ 1 kHz	4.8×10^{-4}	
		(1 ~ 5) kHz	1.0×10^{-3}	
		(5 ~ 10) kHz	8.9×10^{-3}	
		(900 mA ~ 1 A)		
		40 Hz ~ 1 kHz	7.0×10^{-4}	
		(1 ~ 5) kHz	9.0×10^{-4}	
		(5 ~ 10) kHz	5.1×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters	40302			Power Calibrator, Calibrator/ SICT-CP-40302
		(1 ~ 2) A		
		40 Hz ~ 1 kHz	4.3×10^{-4}	
		(1 ~ 5) kHz	6.6×10^{-4}	
		(5 ~ 10) kHz	4.6×10^{-3}	
		(2 ~ 3) A		
		(40 ~ 100) Hz	5.7×10^{-4}	
		100 Hz ~ 5 kHz	9.7×10^{-4}	
		(5 ~ 10) kHz	4.5×10^{-3}	
		(3 ~ 9) A		
		(40 ~ 100) Hz	4.4×10^{-4}	
		100 Hz ~ 5 kHz	8.2×10^{-4}	
		(5 ~ 10) kHz	4.4×10^{-3}	
		(9 ~ 10) A		
		(40 ~ 100) Hz	6.4×10^{-4}	
		100 Hz ~ 5 kHz	8.5×10^{-4}	
		(5 ~ 10) kHz	4.3×10^{-3}	
		(10 ~ 30) A		
		(40 ~ 100) Hz	3.6×10^{-4}	
		100 Hz ~ 5 kHz	7.0×10^{-4}	
		(5 ~ 10) kHz	6.2×10^{-2}	
		(30 ~ 90) A		
		(40 ~ 100) Hz	2.5×10^{-4}	
		100 Hz ~ 5 kHz	6.4×10^{-4}	
		(5 ~ 10) kHz	5.5×10^{-2}	
		(90 ~ 100) A		
		(40 ~ 100) Hz	1.8×10^{-4}	
		100 Hz ~ 5 kHz	5.4×10^{-4}	
		(5 ~ 10) kHz	4.3×10^{-2}	
		(100 ~ 1 000) A		
		40 Hz	1.3×10^{-3}	
		40 Hz ~ 1 kHz	2.5×10^{-3}	
		(1 000 ~ 2 500) A		
		(40 ~ 60) Hz	1.3×10^{-3}	
		(2 500 ~ 3 000) A		
		60 Hz	1.3×10^{-3}	
		(3 000 ~ 10 000) A		
		60 Hz	3.6×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters	40302	0 μA	9.2 nA	Power Calibrator, Calibrator/ SICT-CP-40302
		(0 ~ 0.1) μA	9.2 × 10 ⁻²	
		(0.1 ~ 0.2) μA	4.6 × 10 ⁻²	
		(0.2 ~ 0.3) μA	3.1 × 10 ⁻²	
		(0.3 ~ 0.9) μA	2.3 × 10 ⁻²	
		(0.9 ~ 1) μA	9.3 × 10 ⁻³	
		(1 ~ 2) μA	4.7 × 10 ⁻³	
		(2 ~ 3) μA	3.1 × 10 ⁻³	
		(3 ~ 9) μA	2.3 × 10 ⁻³	
		(9 ~ 10) μA	9.6 × 10 ⁻⁴	
		(10 ~ 20) μA	5.1 × 10 ⁻⁴	
		(20 ~ 30) μA	3.5 × 10 ⁻⁴	
		(30 ~ 50) μA	2.7 × 10 ⁻⁴	
		(50 ~ 90) μA	1.9 × 10 ⁻⁴	
		(90 ~ 100) μA	6.2 × 10 ⁻⁴	
		(100 ~ 200) μA	3.2 × 10 ⁻⁴	
		(200 ~ 700) μA	2.2 × 10 ⁻⁴	
		(700 ~ 900) μA	9.3 × 10 ⁻⁵	
		(0.9 ~ 1) mA	6.1 × 10 ⁻⁴	
		(1 ~ 2) mA	3.1 × 10 ⁻⁴	
		(2 ~ 7) mA	2.2 × 10 ⁻⁴	
		(7 ~ 9) mA	9.1 × 10 ⁻⁵	
		(9 ~ 10) mA	6.1 × 10 ⁻⁴	
		(10 ~ 20) mA	3.1 × 10 ⁻⁴	
		(20 ~ 70) mA	2.2 × 10 ⁻⁴	
		(70 ~ 90) mA	9.9 × 10 ⁻⁵	
		(90 ~ 100) mA	6.1 × 10 ⁻⁴	
		(100 ~ 200) mA	3.1 × 10 ⁻⁴	
		(200 ~ 700) mA	2.5 × 10 ⁻⁴	
		(700 ~ 900) mA	1.3 × 10 ⁻⁴	
		(0.9 ~ 1) A	6.4 × 10 ⁻⁴	
		(1 ~ 2) A	3.4 × 10 ⁻⁴	
		(2 ~ 3) A	4.5 × 10 ⁻⁴	
		(3 ~ 7) A	3.6 × 10 ⁻⁴	
		(7 ~ 9) A	2.2 × 10 ⁻⁴	
		(9 ~ 10) A	6.4 × 10 ⁻⁴	
		(10 ~ 30) A	3.4 × 10 ⁻⁴	
		(30 ~ 70) A	2.6 × 10 ⁻⁴	
		(70 ~ 100) A	1.7 × 10 ⁻⁴	
		(100 ~ 2 500) A	1.3 × 10 ⁻³	
AC Voltage		(1 mV)		
		40 Hz ~ 10 kHz	4.8 × 10 ⁻³	
		(10 ~ 50) kHz	5.0 × 10 ⁻³	
		(50 ~ 100) kHz	6.5 × 10 ⁻³	
		(1 ~ 2) mV		
		40 Hz ~ 10 kHz	2.4 × 10 ⁻³	
		(10 ~ 50) kHz	2.6 × 10 ⁻³	
		(50 ~ 100) kHz	3.5 × 10 ⁻³	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters	40302			Power Calibrator, Calibrator/ SICT-CP-40302
		(2 ~ 5) mV		
		40 Hz ~ 10 kHz	1.7×10^{-3}	
		(10 ~ 50) kHz	1.9×10^{-3}	
		(50 ~ 100) kHz	2.7×10^{-3}	
		(5 ~ 7) mV		
		40 Hz ~ 10 kHz	8.9×10^{-4}	
		(10 ~ 50) kHz	1.0×10^{-3}	
		(50 ~ 100) kHz	1.6×10^{-3}	
		(7 ~ 9) mV		
		40 Hz ~ 10 kHz	6.9×10^{-4}	
		(10 ~ 50) kHz	8.4×10^{-4}	
		(50 ~ 100) kHz	1.4×10^{-3}	
		(9 ~ 10) mV		
		40 Hz ~ 10 kHz	8.3×10^{-4}	
		(10 ~ 50) kHz	9.4×10^{-4}	
		(50 ~ 100) kHz	1.3×10^{-3}	
		(10 ~ 30) mV		
		40 Hz ~ 10 kHz	4.5×10^{-4}	
		(10 ~ 50) kHz	5.6×10^{-4}	
		(50 ~ 100) kHz	1.1×10^{-3}	
		(30 ~ 50) mV		
		40 Hz ~ 10 kHz	3.2×10^{-4}	
		(10 ~ 50) kHz	4.0×10^{-4}	
		(50 ~ 100) kHz	9.0×10^{-4}	
		(50 ~ 70) mV		
		40 Hz ~ 10 kHz	2.3×10^{-4}	
		(10 ~ 50) kHz	3.1×10^{-4}	
		(50 ~ 100) kHz	7.1×10^{-4}	
		(70 ~ 90) mV		
		40 Hz ~ 10 kHz	1.9×10^{-4}	
		(10 ~ 50) kHz	2.6×10^{-4}	
		(50 ~ 100) kHz	6.2×10^{-4}	
		(90 ~ 100) mV		
		40 Hz ~ 10 kHz	1.6×10^{-4}	
		(10 ~ 50) kHz	2.4×10^{-4}	
		(50 ~ 100) kHz	5.7×10^{-4}	
		(100 ~ 200) mV		
		40 Hz ~ 10 kHz	1.1×10^{-4}	
		(10 ~ 50) kHz	1.8×10^{-4}	
		(50 ~ 100) kHz	4.6×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters	40302	(200 ~ 500) mV		Power Calibrator, Calibrator/ SICT-CP-40302
AC Voltage		40 Hz ~ 10 kHz	8.8×10^{-5}	
		(10 ~ 50) kHz	1.4×10^{-4}	
		(50 ~ 100) kHz	2.2×10^{-4}	
		(500 ~ 700) mV		
		40 Hz ~ 10 kHz	7.1×10^{-5}	
		(10 ~ 50) kHz	1.0×10^{-4}	
		(50 ~ 100) kHz	1.6×10^{-4}	
		(700 ~ 900) mV		
		40 Hz ~ 10 kHz	6.5×10^{-5}	
		(10 ~ 50) kHz	9.5×10^{-5}	
		(50 ~ 100) kHz	1.4×10^{-4}	
		(900 mV ~ 1 V)		
		40 Hz ~ 10 kHz	8.6×10^{-5}	
		(10 ~ 50) kHz	1.1×10^{-4}	
		(50 ~ 100) kHz	1.5×10^{-4}	
		(1 ~ 2) V		
		40 Hz ~ 10 kHz	6.4×10^{-5}	
		(10 ~ 50) kHz	9.0×10^{-5}	
		(50 ~ 100) kHz	1.2×10^{-4}	
		(2 ~ 5) V		
		40 Hz	1.3×10^{-4}	
		40 Hz ~ 10 kHz	9.7×10^{-5}	
		(10 ~ 50) kHz	1.5×10^{-4}	
		(50 ~ 100) kHz	2.2×10^{-4}	
		(5 ~ 7) V		
		40 Hz	8.2×10^{-5}	
		40 Hz ~ 10 kHz	6.8×10^{-5}	
		(10 ~ 50) kHz	1.1×10^{-4}	
		(50 ~ 100) kHz	1.5×10^{-4}	
		(7 ~ 9) V		
		40 Hz	7.0×10^{-5}	
		40 Hz ~ 10 kHz	6.2×10^{-5}	
		(10 ~ 50) kHz	9.9×10^{-5}	
		(50 ~ 100) kHz	1.3×10^{-4}	
		(9 ~ 10) V		
		40 Hz	8.9×10^{-5}	
		40 Hz ~ 10 kHz	8.4×10^{-5}	
		(10 ~ 50) kHz	1.1×10^{-4}	
		(50 ~ 100) kHz	1.4×10^{-4}	
		(10 ~ 20) V		
		40 Hz ~ 10 kHz	6.2×10^{-5}	
		(10 ~ 50) kHz	9.0×10^{-5}	
		(50 ~ 100) kHz	1.1×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters	40302			Power Calibrator, Calibrator/ SICT-CP-40302
AC Voltage		(20 ~ 50) V		
		40 Hz ~ 10 kHz	1.6×10^{-4}	
		(10 ~ 50) kHz	1.7×10^{-4}	
		(50 ~ 100) kHz	3.4×10^{-4}	
		(50 ~ 70) V		
		40 Hz	9.8×10^{-5}	
		40 Hz ~ 10 kHz	8.3×10^{-5}	
		(10 ~ 50) kHz	1.2×10^{-4}	
		(50 ~ 100) kHz	2.4×10^{-4}	
		(70 ~ 90) V		
		40 Hz	8.5×10^{-5}	
		40 Hz ~ 10 kHz	7.5×10^{-5}	
		(10 ~ 50) kHz	1.1×10^{-4}	
		(50 ~ 100) kHz	2.2×10^{-4}	
		(90 ~ 100) V		
		40 Hz ~ 10 kHz	9.9×10^{-5}	
		(10 ~ 50) kHz	1.2×10^{-4}	
		(50 ~ 100) kHz	2.2×10^{-4}	
		(100 ~ 200) V		
		40 Hz ~ 10 kHz	7.3×10^{-5}	
		(10 ~ 50) kHz	1.0×10^{-4}	
		(50 ~ 100) kHz	1.9×10^{-4}	
		(200 ~ 500) V		
		40 Hz ~ 1 kHz	1.3×10^{-4}	
		(1 ~ 10) kHz	2.2×10^{-4}	
		(10 ~ 20) kHz	6.6×10^{-4}	
		(500 ~ 1 000) V		
		40 Hz ~ 1 kHz	1.5×10^{-4}	
		(1 ~ 10) kHz	2.2×10^{-4}	
		(10 ~ 20) kHz	9.9×10^{-4}	
DC Voltage		0 mV	$61 \mu\text{V}$	
		(0 ~ 10) mV	6.1×10^{-3}	
		(10 ~ 20) mV	3.1×10^{-3}	
		(20 ~ 30) mV	2.0×10^{-3}	
		(30 ~ 60) mV	1.5×10^{-3}	
		(60 ~ 70) mV	8.7×10^{-4}	
		(70 ~ 80) mV	7.6×10^{-4}	
		(80 ~ 100) mV	6.8×10^{-4}	
		(100 ~ 200) mV	3.3×10^{-5}	
		(200 ~ 300) mV	2.2×10^{-5}	
		(300 ~ 800) mV	1.7×10^{-5}	
		(800 ~ 900) mV	9.6×10^{-6}	
		(0.9 ~ 1) V	6.1×10^{-5}	
		(1 ~ 2) V	3.1×10^{-5}	
		(2 ~ 3) V	2.1×10^{-5}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Clamp ammeters/voltmeters	DC Voltage	40302	$(3 \sim 6) \text{ V}$ $(6 \sim 7) \text{ V}$ $(7 \sim 9) \text{ V}$ $(9 \sim 10) \text{ V}$ $(10 \sim 20) \text{ V}$ $(20 \sim 30) \text{ V}$ $(30 \sim 80) \text{ V}$ $(80 \sim 90) \text{ V}$ $(90 \sim 100) \text{ V}$ $(100 \sim 200) \text{ V}$ $(200 \sim 300) \text{ V}$ $(300 \sim 500) \text{ V}$ $(500 \sim 900) \text{ V}$ $(900 \sim 1\,000) \text{ V}$	1.6×10^{-5} 9.8×10^{-6} 8.8×10^{-6} 6.1×10^{-5} 3.1×10^{-5} 2.2×10^{-5} 1.7×10^{-5} 9.3×10^{-6} 6.1×10^{-5} 3.2×10^{-5} 2.3×10^{-5} 1.8×10^{-5} 1.3×10^{-5} 6.2×10^{-5}	
		0 Ω	0.61 $m\Omega$	Power Calibrator, Calibrator/ SICT-CP-40302	
		$(0 \sim 9) \Omega$	0.66 $m\Omega$		
		$(9 \sim 100) \Omega$	6.2 $m\Omega$		
		$(100 \sim 900) \Omega$	9.2 $m\Omega$		
		$(0.9 \sim 9) k\Omega$	92 $m\Omega$		
		$(9 \sim 90) k\Omega$	1.1 Ω		
		$(0.090 \sim 1) M\Omega$	63 Ω		
		$(1 \sim 10) M\Omega$	0.77 $k\Omega$		
		$(10 \sim 100) M\Omega$	13 $k\Omega$		
AC voltage/current calibrators	AC Voltage	40303	(1 mV) $10 \text{ Hz} \sim 10 \text{ kHz}$ $(10 \sim 100) \text{ kHz}$ $100 \text{ kHz} \sim 1 \text{ MHz}$ $(1 \sim 2) \text{ mV}$ $10 \text{ Hz} \sim 10 \text{ kHz}$ $(10 \sim 100) \text{ kHz}$ $100 \text{ kHz} \sim 1 \text{ MHz}$ $(2 \sim 5) \text{ mV}$ 10 Hz $10 \text{ Hz} \sim 10 \text{ kHz}$ $(10 \sim 100) \text{ kHz}$ $100 \text{ kHz} \sim 1 \text{ MHz}$ $(5 \sim 10) \text{ mV}$ 10 Hz $10 \text{ Hz} \sim 10 \text{ kHz}$ $(10 \sim 100) \text{ kHz}$ $100 \text{ kHz} \sim 1 \text{ MHz}$ $(10 \sim 20) \text{ mV}$ 10 Hz $10 \text{ Hz} \sim 10 \text{ kHz}$ $(10 \sim 100) \text{ kHz}$ $100 \text{ kHz} \sim 1 \text{ MHz}$	1.7×10^{-3} 3.0×10^{-3} 1.2×10^{-2} 1.1×10^{-3} 1.7×10^{-3} 7.7×10^{-3} 6.4×10^{-4} 5.8×10^{-4} 1.0×10^{-3} 5.4×10^{-3} 4.2×10^{-4} 3.5×10^{-4} 5.8×10^{-4} 3.9×10^{-3} 1.8×10^{-4} 1.4×10^{-4} 2.2×10^{-4} 2.2×10^{-3}	Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators	40303			Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
AC Voltage		(20 ~ 50) mV		
		10 Hz	1.4×10^{-4}	
		10 Hz ~ 10 kHz	9.2×10^{-5}	
		(10 ~ 100) kHz	1.6×10^{-4}	
		100 kHz ~ 1 MHz	1.4×10^{-3}	
		(50 ~ 100) mV		
		10 Hz	1.1×10^{-4}	
		10 Hz ~ 10 kHz	6.6×10^{-5}	
		(10 ~ 100) kHz	1.2×10^{-4}	
		100 kHz ~ 1 MHz	1.3×10^{-3}	
		(100 ~ 200) mV		
		10 Hz	8.2×10^{-5}	
		10 Hz ~ 10 kHz	3.9×10^{-5}	
		(10 ~ 100) kHz	7.6×10^{-5}	
		100 kHz ~ 1 MHz	1.1×10^{-3}	
		(200 ~ 500) mV		
		10 Hz	7.8×10^{-5}	
		10 Hz ~ 10 kHz	3.6×10^{-5}	
		(10 ~ 100) kHz	7.1×10^{-5}	
		100 kHz ~ 1 MHz	1.1×10^{-3}	
		(500 mV ~ 1 V)		
		10 Hz	7.6×10^{-5}	
		10 Hz ~ 10 kHz	3.3×10^{-5}	
		(10 ~ 100) kHz	6.6×10^{-5}	
		100 kHz ~ 1 MHz	1.1×10^{-3}	
		(1 ~ 2) V		
		10 Hz	7.1×10^{-5}	
		10 Hz ~ 10 kHz	2.7×10^{-5}	
		(10 ~ 100) kHz	5.8×10^{-5}	
		100 kHz ~ 1 MHz	1.0×10^{-3}	
		(2 ~ 5) V		
		10 Hz	7.2×10^{-5}	
		10 Hz ~ 10 kHz	2.6×10^{-5}	
		10 kHz ~ 100 kHz	7.5×10^{-5}	
		100 kHz ~ 1 MHz	1.4×10^{-3}	
		(5 ~ 20) V		
		10 Hz	7.2×10^{-5}	
		10 Hz ~ 10 kHz	2.8×10^{-5}	
		(10 ~ 100) kHz	7.5×10^{-5}	
		100 kHz ~ 1 MHz	1.4×10^{-3}	
		(20 ~ 50) V		
		10 Hz	7.2×10^{-5}	
		10 Hz ~ 10 kHz	3.0×10^{-5}	
		(10 ~ 100) kHz	8.0×10^{-5}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators	40303			Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
AC Voltage		(50 ~ 200) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	7.4×10^{-5} 3.3×10^{-5} 8.5×10^{-5}	
		(200 ~ 1 000) V 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	7.7×10^{-5} 3.3×10^{-5} 5.8×10^{-4}	
AC Current		(10 µA) 10 Hz ~ 10 kHz	2.6×10^{-3}	
		(10 ~ 100) µA 10 Hz ~ 1 kHz (1 ~ 10) kHz	3.6×10^{-4} 6.4×10^{-4}	
		(100 µA ~ 1 mA) 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	9.8×10^{-5} 7.5×10^{-5} 9.4×10^{-5}	
		(1 ~ 100) mA 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	7.8×10^{-5} 4.6×10^{-5} 4.2×10^{-5}	
		(100 mA ~ 1 A) 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	8.1×10^{-5} 4.9×10^{-5} 4.4×10^{-5}	
		(1 ~ 2) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	7.9×10^{-5} 4.7×10^{-5} 4.5×10^{-5}	
		(2 ~ 5) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	8.2×10^{-5} 5.2×10^{-5} 5.0×10^{-5}	
		(5 ~ 10) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	8.6×10^{-5} 5.9×10^{-5} 7.8×10^{-5}	
		(10 ~ 20) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	9.3×10^{-5} 6.8×10^{-5} 7.8×10^{-5}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators	40303			Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
AC Current		(20 ~ 50) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	1.0×10^{-4} 8.3×10^{-5} 1.1×10^{-4}	
		(50 ~ 100) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	1.2×10^{-4} 9.7×10^{-5} 1.3×10^{-4}	
		(100 ~ 200) A 60 Hz	4.5×10^{-4}	
Wattmeter calibrators	40304			Power Standard, Counter/ SICT-CP-40304
AC Voltage		(1 mV) 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	1.7×10^{-3} 3.0×10^{-3} 1.2×10^{-2}	
		(1 ~ 2) mV 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	1.1×10^{-3} 1.7×10^{-3} 7.7×10^{-3}	
		(2 ~ 5) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	6.4×10^{-4} 5.8×10^{-4} 1.0×10^{-3} 5.4×10^{-3}	
		(5 ~ 10) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	4.2×10^{-4} 3.5×10^{-4} 5.8×10^{-4} 3.9×10^{-3}	
		(10 ~ 20) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	1.8×10^{-4} 1.4×10^{-4} 2.2×10^{-4} 2.2×10^{-3}	
		(20 ~ 50) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	1.4×10^{-4} 9.2×10^{-5} 1.6×10^{-4} 1.4×10^{-3}	
		(50 ~ 100) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	1.1×10^{-4} 6.6×10^{-5} 1.2×10^{-4} 1.3×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Wattmeter calibrators	40304			
AC Voltage		(100 ~ 200) mV		
		10 Hz	8.2×10^{-5}	
		10 Hz ~ 10 kHz	3.9×10^{-5}	
		(10 ~ 100) kHz	7.6×10^{-5}	
		100 kHz ~ 1 MHz	1.1×10^{-3}	
		(200 ~ 500) mV		
		10 Hz	7.8×10^{-5}	
		10 Hz ~ 10 kHz	3.6×10^{-5}	
		(10 ~ 100) kHz	7.1×10^{-5}	
		100 kHz ~ 1 MHz	1.1×10^{-3}	
		(500 mV ~ 1 V)		
		10 Hz	7.6×10^{-5}	
		10 Hz ~ 10 kHz	3.3×10^{-5}	
		(10 ~ 100) kHz	6.6×10^{-5}	
		100 kHz ~ 1 MHz	1.1×10^{-3}	
		(1 ~ 2) V		
		10 Hz	7.1×10^{-5}	
		10 Hz ~ 10 kHz	2.7×10^{-5}	
		(10 ~ 100) kHz	5.8×10^{-5}	
		100 kHz ~ 1 MHz	1.0×10^{-3}	
		(2 ~ 5) V		
		10 Hz	7.2×10^{-5}	
		10 Hz ~ 10 kHz	2.6×10^{-5}	
		(10 ~ 100) kHz	7.5×10^{-5}	
		100 kHz ~ 1 MHz	1.4×10^{-3}	
		(5 ~ 20) V		
		10 Hz	7.2×10^{-5}	
		10 Hz ~ 10 kHz	2.8×10^{-5}	
		(10 ~ 100) kHz	7.5×10^{-5}	
		100 kHz ~ 1 MHz	1.4×10^{-3}	
		(20 ~ 50) V		
		10 Hz	7.2×10^{-5}	
		10 Hz ~ 10 kHz	3.0×10^{-5}	
		(10 ~ 100) kHz	8.0×10^{-5}	
		(50 ~ 200) V		
		10 Hz	7.4×10^{-5}	
		10 Hz ~ 10 kHz	3.3×10^{-5}	
		(10 ~ 100) kHz	8.5×10^{-5}	
		(200 ~ 1 000) V		
		10 Hz	7.7×10^{-5}	
		10 Hz ~ 10 kHz	3.3×10^{-5}	
		(10 ~ 100) kHz	5.8×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Wattmeter calibrators	40304			
		(10 µA) 10 Hz ~ 10 kHz	2.6×10^{-3}	
		(10 ~ 100) µA 10 Hz ~ 1 kHz (1 ~ 10) kHz	3.6×10^{-4} 6.4×10^{-4}	
		(100 µA ~ 1 mA) 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	9.8×10^{-5} 7.5×10^{-5} 9.4×10^{-5}	
		(1 ~ 100) mA 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	7.8×10^{-5} 4.6×10^{-5} 4.2×10^{-5}	
		(100 mA ~ 1 A) 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	8.1×10^{-5} 4.9×10^{-5} 4.4×10^{-5}	
		(1 ~ 2) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	7.9×10^{-5} 4.7×10^{-5} 4.5×10^{-5}	
		(2 ~ 5) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	8.2×10^{-5} 5.2×10^{-5} 5.0×10^{-5}	
		(5 ~ 10) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	8.6×10^{-5} 5.9×10^{-5} 7.8×10^{-5}	
		(10 ~ 20) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	9.3×10^{-5} 6.8×10^{-5} 7.8×10^{-5}	
		(20 ~ 50) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	1.0×10^{-4} 8.3×10^{-5} 1.1×10^{-4}	
		(50 ~ 100) A 10 Hz 10 Hz ~ 1 kHz (1 ~ 10) kHz	1.2×10^{-4} 9.7×10^{-5} 1.3×10^{-4}	
		(100 ~ 200) A 60 Hz	4.5×10^{-4}	
				Power Standard, Counter/ SICT-CP-40304

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Wattmeter calibrators	40304			
AC Power		(50 ~ 60) Hz 0 mW (0 ~ 0.22) mW (0.22 ~ 1.1) mW (1.1 ~ 2.2) mW (2.2 ~ 11) mW (11 ~ 22) mW (22 ~ 44) mW (44 ~ 66) mW (66 ~ 88) mW (88 ~ 110) mW (110 ~ 480) mW (480 ~ 550) mW (0.55 ~ 1.1) W (1.1 ~ 5.5) W 5.5 W ~ 1.1 kW (1.1 ~ 2.2) kW (2.2 ~ 24) kW	0.05 mW 3.2×10^{-1} 6.3×10^{-2} 3.1×10^{-2} 6.3×10^{-3} 3.1×10^{-3} 1.6×10^{-3} 1.1×10^{-3} 8.0×10^{-4} 6.4×10^{-4} 2.6×10^{-4} 2.3×10^{-4} 1.4×10^{-4} 1.6×10^{-4} 1.4×10^{-4} 1.5×10^{-4} 1.4×10^{-4}	Power Standard, Counter/ SICT-CP-40304
Power Factor		(50 ~ 60) Hz (0 ~ 550) mW -1 ~ 1	2.3×10^{-4}	
		550 mW ~ 24 kW -1 ~ 1	1.5×10^{-4}	
Harmonic Voltage		(50 ~ 60) Hz (0.5 ~ 3) % (3 ~ 10) % (10 ~ 20) %	0.042 % 0.052 % 0.081 %	
Harmonic Current		(50 ~ 60) Hz (0.5 ~ 3) % (3 ~ 10) % (10 ~ 20) %	0.042 % 0.052 % 0.055 %	
Flicker		P_{st} (0.25 ~ 5), (50 Hz) Modulation Frequency 8.333 mHz 16.667 mHz 58.333 mHz 325.000 mHz 916.667 mHz 13.500 Hz 33.333 Hz	2.7×10^{-3} 2.7×10^{-3} 2.7×10^{-3} 2.7×10^{-3} 2.7×10^{-3} 2.7×10^{-3} 2.7×10^{-3}	
Frequency		(10 ~ 100) Hz (100 ~ 400) Hz 400 Hz ~ 1 MHz	7.0×10^{-7} 3.8×10^{-7} 7.0×10^{-7}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC current shunts AC Resistance	40305	(100 ~ 200) A (60 Hz) 1 mΩ	8.4×10^{-4}	Reference Multimeter, Calibrator/ SICT-CP-40305
		(40 ~ 60) Hz (1 ~ 10) mΩ (10 ~ 100) mΩ	2.1×10^{-4} 3.5×10^{-4}	
		(60 Hz ~ 1 kHz) (1 ~ 10) mΩ (10 ~ 100) mΩ	5.7×10^{-4} 3.4×10^{-4}	
		(40 Hz ~ 1 kHz) 100 mΩ ~ 100 Ω 100 Ω ~ 10 kΩ	1.8×10^{-4} 2.3×10^{-4}	
		(50 Hz) (1 ~ 100) kV	2.6×10^{-4}	
		(60 Hz) (1 ~ 100) kV	1.9×10^{-4}	
Phase angle generators, synchro resolve generators	40306	(-180 ~ 180) ° 50 Hz (50 ~ 500) Hz (500 ~ 1 000) Hz	0.001 6 ° 0.003 1 ° 0.010 °	전력 교정기/ SICT-CP-40307
Voltage/current phase angle meters/synchro resolve meters	40307	(50 ~ 60) Hz (-180 ~ 180) °	0.008 8 °	Power Calibrator/ SICT-CP-40307
Potential transformer test sets	40308	(110 ~ 1 100) V (-19.99 ~ 19.99) % (-680 ~ 680) '	0.020 % 0.70 '	Standard Potential transforme, Ratio transformers/ SICT-CP-40308
		(1 100 ~ 100 000) V (-19.99 ~ 19.99) %	0.016 %	
		(-680 ~ 680) '	0.50 '	
	40309	(110 ~ 1 100) V (-19.99 ~ 19.99) % (-680 ~ 680) '	0.020 % 0.70 '	Standard Potential transforme/ SICT-CP-40309
		(1 100 ~ 100 000) V (-19.99 ~ 19.99) %	0.016 %	
		(-680 ~ 680) '	0.50 '	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Power factor meters AC Power Factor	40310	(50 Hz, 60 Hz) -1 ~ 1	1.1×10^{-4}	Power Calibrator/ SICT-CP-40310
AC power meters AC Voltage	40311	(1 mV) 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (1 ~ 2) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (2 ~ 5) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (5 ~ 7) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (7 ~ 9) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (9 ~ 10) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (10 ~ 30) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (30 ~ 60) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz (60 ~ 200) mV 40 Hz ~ 10 kHz (10 ~ 50) kHz (50 ~ 100) kHz	4.8 $\times 10^{-3}$ 5.0 $\times 10^{-3}$ 6.5 $\times 10^{-3}$ 2.4 $\times 10^{-3}$ 2.6 $\times 10^{-3}$ 3.5 $\times 10^{-3}$ 1.7 $\times 10^{-3}$ 1.9 $\times 10^{-3}$ 2.7 $\times 10^{-3}$ 8.9 $\times 10^{-4}$ 1.0 $\times 10^{-3}$ 1.6 $\times 10^{-3}$ 6.9 $\times 10^{-4}$ 8.4 $\times 10^{-4}$ 1.4 $\times 10^{-3}$ 5.7 $\times 10^{-4}$ 7.1 $\times 10^{-4}$ 1.2 $\times 10^{-3}$ 3.6 $\times 10^{-4}$ 4.7 $\times 10^{-4}$ 1.1 $\times 10^{-3}$ 2.9 $\times 10^{-4}$ 3.7 $\times 10^{-4}$ 8.8 $\times 10^{-4}$ 1.9 $\times 10^{-4}$ 2.7 $\times 10^{-4}$ 6.5 $\times 10^{-4}$	Power Calibrator, Calibrator/ SICT-CP-40311

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311	(200 ~ 300) mV		Power Calibrator, Calibrator/ SICT-CP-40311
		40 Hz ~ 10 kHz	8.6×10^{-5}	
		(10 ~ 50) kHz	1.3×10^{-4}	
		(50 ~ 100) kHz	2.2×10^{-4}	
		(300 ~ 600) mV		
		40 Hz ~ 10 kHz	7.6×10^{-5}	
		(10 ~ 50) kHz	1.2×10^{-4}	
		(50 ~ 100) kHz	1.9×10^{-4}	
		(600 mV ~ 1 V)		
		40 Hz ~ 10 kHz	6.7×10^{-5}	
		(10 ~ 50) kHz	9.8×10^{-5}	
		(50 ~ 100) kHz	1.5×10^{-4}	
		(1 ~ 2) V		
		40 Hz ~ 10 kHz	5.6×10^{-5}	
		(10 ~ 50) kHz	8.5×10^{-5}	
		(50 ~ 100) kHz	1.2×10^{-4}	
		(2 ~ 3) V		
		40 Hz	1.3×10^{-4}	
		40 Hz ~ 10 kHz	9.5×10^{-5}	
		(10 ~ 50) kHz	1.5×10^{-4}	
		(50 ~ 100) kHz	2.2×10^{-4}	
		(3 ~ 5) V		
		40 Hz	1.1×10^{-4}	
		40 Hz ~ 10 kHz	8.0×10^{-5}	
		(10 ~ 50) kHz	1.3×10^{-4}	
		(50 ~ 100) kHz	1.8×10^{-4}	
		(5 ~ 7) V		
		40 Hz	8.1×10^{-5}	
		40 Hz ~ 10 kHz	6.7×10^{-5}	
		(10 ~ 50) kHz	1.1×10^{-4}	
		(50 ~ 100) kHz	1.5×10^{-4}	
		(7 ~ 20) V		
		40 Hz	7.0×10^{-5}	
		40 Hz ~ 10 kHz	6.1×10^{-5}	
		(10 ~ 50) kHz	9.9×10^{-5}	
		(50 ~ 100) kHz	1.3×10^{-4}	
		(20 ~ 60) V		
		40 Hz	1.6×10^{-4}	
		40 Hz ~ 10 kHz	1.2×10^{-4}	
		(10 ~ 50) kHz	1.6×10^{-4}	
		(50 ~ 100) kHz	3.4×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311	(60 ~ 100) V		Power Calibrator, Calibrator/ SICT-CP-40311
		40 Hz	9.0×10^{-5}	
		40 Hz ~ 10 kHz	7.8×10^{-5}	
		(10 ~ 50) kHz	1.2×10^{-4}	
		(50 ~ 100) kHz	2.3×10^{-4}	
		(100 ~ 200) V		
		40 Hz	6.7×10^{-5}	
		40 Hz ~ 10 kHz	6.5×10^{-5}	
		(10 ~ 50) kHz	9.9×10^{-5}	
		(50 ~ 100) kHz	1.9×10^{-4}	
		(200 ~ 400) V		
		40 Hz	1.4×10^{-4}	
		40 Hz ~ 1 kHz	1.1×10^{-4}	
		(1 ~ 10) kHz	2.2×10^{-4}	
		(10 ~ 20) kHz	6.6×10^{-4}	
		(400 ~ 500) V		
		40 Hz	1.3×10^{-4}	
		40 Hz ~ 1 kHz	1.1×10^{-4}	
		(1 ~ 10) kHz	2.1×10^{-4}	
		(10 ~ 20) kHz	5.4×10^{-4}	
		(500 ~ 600) V		
		40 Hz	1.5×10^{-4}	
		40 Hz ~ 1 kHz	1.3×10^{-4}	
		(1 ~ 10) kHz	2.2×10^{-4}	
		(10 ~ 20) kHz	9.9×10^{-4}	
		(600 ~ 700) V		
		40 Hz	1.4×10^{-4}	
		40 Hz ~ 1 kHz	1.2×10^{-4}	
		(1 ~ 10) kHz	2.2×10^{-4}	
		(10 ~ 20) kHz	8.5×10^{-4}	
		(700 ~ 900) V		
		40 Hz	1.3×10^{-4}	
		40 Hz ~ 1 kHz	1.1×10^{-4}	
		(1 ~ 10) kHz	2.1×10^{-4}	
		(10 ~ 20) kHz	7.5×10^{-4}	
		(900 ~ 1 000) V		
		40 Hz	1.2×10^{-4}	
		40 Hz ~ 1 kHz	1.0×10^{-4}	
		(1 ~ 10) kHz	2.1×10^{-4}	
		(10 ~ 20) kHz	6.1×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311			Power Calibrator, Calibrator/ SICT-CP-40311
AC Current		(100 μ A) 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	2.3×10^{-4} 4.9×10^{-4} 2.1×10^{-3}	
		(100 ~ 300) μ A 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	2.3×10^{-4} 4.6×10^{-4} 1.8×10^{-3}	
		(300 μ A ~ 2 mA) 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	1.9×10^{-4} 3.8×10^{-4} 2.1×10^{-3}	
		(2 ~ 4) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	3.2×10^{-4} 6.9×10^{-4} 3.8×10^{-3}	
		(4 ~ 7) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	2.3×10^{-4} 4.8×10^{-4} 2.7×10^{-3}	
		(7 ~ 20) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	1.9×10^{-4} 3.7×10^{-4} 2.1×10^{-3}	
		(20 ~ 30) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	2.9×10^{-4} 6.5×10^{-4} 2.6×10^{-3}	
		(30 ~ 60) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	2.4×10^{-4} 5.2×10^{-4} 2.2×10^{-3}	
		(60 ~ 200) mA 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	1.8×10^{-4} 3.7×10^{-4} 1.7×10^{-3}	
		(200 ~ 300) mA 40 Hz (40 ~ 60) Hz 60 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	5.8×10^{-4} 1.2×10^{-4} 5.3×10^{-4} 1.2×10^{-3} 9.3×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311	(300 ~ 500) mA		Power Calibrator, Calibrator/ SICT-CP-40311
AC Current		40 Hz	4.9×10^{-4}	
		(40 ~ 60) Hz	1.1×10^{-4}	
		60 Hz ~ 1 kHz	4.6×10^{-4}	
		(1 ~ 5) kHz	1.0×10^{-3}	
		(5 ~ 10) kHz	8.9×10^{-3}	
		(500 ~ 800) mA		
		40 Hz	4.0×10^{-4}	
		(40 ~ 60) Hz	1.3×10^{-4}	
		60 Hz ~ 1 kHz	3.9×10^{-4}	
		(1 ~ 5) kHz	8.3×10^{-4}	
		(5 ~ 10) kHz	8.6×10^{-3}	
		(800 mA ~ 2 A)		
		40 Hz	3.5×10^{-4}	
		(40 ~ 60) Hz	9.5×10^{-5}	
		60 Hz ~ 1 kHz	3.5×10^{-4}	
		(1 ~ 5) kHz	7.2×10^{-4}	
		(5 ~ 10) kHz	8.4×10^{-3}	
		(2 ~ 3) A		
		40 Hz	5.3×10^{-4}	
		(40 ~ 60) Hz	1.5×10^{-4}	
		(60 ~ 100) Hz	5.3×10^{-4}	
		100 Hz ~ 5 kHz	9.4×10^{-4}	
		(5 ~ 10) kHz	4.5×10^{-3}	
		(3 ~ 6) A		
		40 Hz	4.1×10^{-4}	
		(40 ~ 60) Hz	1.3×10^{-4}	
		(60 ~ 100) Hz	4.1×10^{-4}	
		100 Hz ~ 5 kHz	8.1×10^{-4}	
		(5 ~ 10) kHz	4.4×10^{-3}	
		(6 ~ 10) A		
		40 Hz	2.6×10^{-4}	
		(40 ~ 60) Hz	1.1×10^{-4}	
		(60 ~ 100) Hz	2.6×10^{-4}	
		100 Hz ~ 5 kHz	6.4×10^{-4}	
		(5 ~ 10) kHz	4.3×10^{-3}	
		(10 ~ 50) A		
		40 Hz	2.4×10^{-4}	
		(40 ~ 60) Hz	1.1×10^{-4}	
		(60 ~ 100) Hz	2.4×10^{-4}	
		100 Hz ~ 5 kHz	6.6×10^{-4}	
		(5 ~ 10) kHz	6.2×10^{-2}	
		(50 ~ 100) A		
		(40 ~ 100) Hz	1.9×10^{-4}	
		100 Hz ~ 5 kHz	5.7×10^{-4}	
		(5 ~ 10) kHz	4.8×10^{-2}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311	(100 ~ 1 000) A (40 ~ 100) Hz 100 Hz ~ 1 kHz	1.3×10^{-3} 2.7×10^{-3}	Power Calibrator, Calibrator/ SICT-CP-40311
AC Current		(1 000 ~ 2 500) A (40 ~ 60) Hz	1.2×10^{-3}	
		(2 500 ~ 3 000) A 60 Hz	1.3×10^{-3}	
AC Wattage		(50 ~ 60) Hz 0 mW (0 ~ 0.22) mW (0.22 ~ 1.1) mW (1.1 ~ 2.2) mW (2.2 ~ 11) mW (11 ~ 22) mW (22 ~ 44) mW (44 ~ 66) mW (66 ~ 88) mW (88 ~ 110) mW (110 ~ 480) mW 480 mW ~ 12 kW (12 ~ 24) kW (24 ~ 300) kW (300 ~ 600) kW	70 μ W 2.1×10^{-1} 4.1×10^{-2} 2.1×10^{-2} 4.1×10^{-3} 2.1×10^{-3} 1.0×10^{-3} 7.0×10^{-4} 5.3×10^{-4} 4.3×10^{-4} 2.1×10^{-4} 1.2×10^{-4} 6.8×10^{-4} 1.2×10^{-3} 1.4×10^{-3}	
DC Voltage		0 mV (0 ~ 1) mV (1 ~ 2) mV (2 ~ 3) mV (3 ~ 4) mV (4 ~ 8) mV (8 ~ 9) mV (9 ~ 10) mV (10 ~ 20) mV (20 ~ 30) mV (30 ~ 60) mV (60 ~ 200) mV (200 ~ 300) mV (300 ~ 400) mV (400 ~ 700) mV (700 ~ 900) mV (0.9 ~ 1) V (1 ~ 2) V (2 ~ 3) V (3 ~ 9) V (9 ~ 10) V (10 ~ 20) V (20 ~ 40) V (40 ~ 90) V (90 ~ 200) V (200 ~ 400) V (400 ~ 900) V (900 ~ 1 000) V	0.78 μ V 8.0×10^{-4} 4.0×10^{-4} 2.7×10^{-4} 2.0×10^{-4} 1.6×10^{-4} 9.4×10^{-5} 8.5×10^{-5} 4.5×10^{-5} 3.2×10^{-5} 2.6×10^{-5} 1.8×10^{-5} 9.3×10^{-6} 8.3×10^{-6} 7.8×10^{-6} 6.9×10^{-6} 9.0×10^{-6} 7.2×10^{-6} 5.6×10^{-6} 5.1×10^{-6} 7.5×10^{-6} 6.0×10^{-6} 7.9×10^{-6} 6.9×10^{-6} 8.8×10^{-6} 9.9×10^{-6} 8.7×10^{-6} 1.0×10^{-5}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311	0 μA	9.0 nA	Power Calibrator, Calibrator/ SICT-CP-40311
		(0 ~ 0.1) μA	9.0 × 10 ⁻²	
		(0.1 ~ 0.2) μA	4.5 × 10 ⁻²	
		(0.2 ~ 0.3) μA	3.0 × 10 ⁻²	
		(0.3 ~ 0.4) μA	2.3 × 10 ⁻²	
		(0.4 ~ 0.9) μA	1.8 × 10 ⁻²	
		(0.9 ~ 1) μA	9.1 × 10 ⁻³	
		(1 ~ 2) μA	4.5 × 10 ⁻³	
		(2 ~ 3) μA	3.0 × 10 ⁻³	
		(3 ~ 4) μA	2.3 × 10 ⁻³	
		(4 ~ 9) μA	1.8 × 10 ⁻³	
		(9 ~ 10) μA	9.4 × 10 ⁻⁴	
		(10 ~ 20) μA	5.0 × 10 ⁻⁴	
		(20 ~ 30) μA	3.4 × 10 ⁻⁴	
		(30 ~ 50) μA	2.7 × 10 ⁻⁴	
		(50 ~ 90) μA	1.9 × 10 ⁻⁴	
		(90 ~ 100) μA	1.3 × 10 ⁻⁴	
		(100 ~ 200) μA	8.6 × 10 ⁻⁵	
		(200 ~ 300) μA	6.5 × 10 ⁻⁵	
		(300 ~ 500) μA	5.5 × 10 ⁻⁵	
		(0.5 ~ 2) mA	5.1 × 10 ⁻⁵	
		(2 ~ 3) mA	7.3 × 10 ⁻⁵	
		(3 ~ 4) mA	6.3 × 10 ⁻⁵	
		(4 ~ 7) mA	5.7 × 10 ⁻⁵	
		(7 ~ 20) mA	4.9 × 10 ⁻⁵	
		(20 ~ 30) mA	8.2 × 10 ⁻⁵	
		(30 ~ 50) mA	7.4 × 10 ⁻⁵	
		(50 ~ 100) mA	6.6 × 10 ⁻⁵	
		(100 ~ 200) mA	5.7 × 10 ⁻⁵	
		(200 ~ 600) mA	1.5 × 10 ⁻⁴	
		(0.6 ~ 2) A	1.1 × 10 ⁻⁴	
		(2 ~ 3) A	4.0 × 10 ⁻⁴	
		(3 ~ 4) A	3.2 × 10 ⁻⁴	
		(4 ~ 7) A	2.8 × 10 ⁻⁴	
		(7 ~ 10) A	2.1 × 10 ⁻⁴	
		(10 ~ 20) A	1.5 × 10 ⁻⁴	
		(20 ~ 30) A	2.5 × 10 ⁻⁴	
		(30 ~ 80) A	2.1 × 10 ⁻⁴	
		(80 ~ 100) A	1.5 × 10 ⁻⁴	
		(100 ~ 2 500) A	1.3 × 10 ⁻³	
DC Wattage	40311	0 mW	61 nW	Power Calibrator, Calibrator/ SICT-CP-40311
		(0 ~ 1) mW	7.7 × 10 ⁻⁵	
		(1 ~ 10) mW	4.8 × 10 ⁻⁵	
		(10 ~ 100) mW	6.1 × 10 ⁻⁵	
		(0.1 ~ 100) W	1.1 × 10 ⁻⁴	
		(0.1 ~ 10) kW	1.9 × 10 ⁻⁴	
		(10 ~ 20) kW	1.5 × 10 ⁻⁴	
		(20 ~ 1 000) kW	1.0 × 10 ⁻³	
		(1 000 ~ 2 500) kW	1.3 × 10 ⁻³	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311			Power Calibrator, Calibrator/ SICT-CP-40311
Harmonic Voltage		(50 ~ 60) Hz (0.5 ~ 3) % (3 ~ 5) % (5 ~ 10) % (10 ~ 20) %	0.030 % 0.033 % 0.042 % 0.065 %	
Harmonic Current		(50 ~ 60) Hz (0.5 ~ 3) % (3 ~ 10) % (5 ~ 20) %	0.030 % 0.032 % 0.038 %	
Flicker		F_{st} (0.25 ~ 5), (50 Hz) Modulation Frequency 8.333 mHz 16.667 mHz 58.333 mHz 325 mHz 916.667 mHz 13.5 Hz 33.333 Hz	3.5×10^{-2} 3.5×10^{-2} 3.5×10^{-2} 3.5×10^{-2} 3.5×10^{-2} 3.5×10^{-2} 3.5×10^{-2}	
Frequency		10 Hz ~ 10 MHz	1.3×10^{-4}	
Power Factor		(50 ~ 60) Hz -1 ~ 1	1.1×10^{-4}	
Current burden		(50 ~ 60) Hz 1.25 VA (1.25 ~ 3.75) VA (3.75 ~ 5) VA (5 ~ 10) VA (10 ~ 100) VA	2.9×10^{-3} 1.5×10^{-3} 8.0×10^{-4} 5.1×10^{-4} 4.0×10^{-4}	
Current burden factor		0.5 ~ 1	2.2×10^{-4}	
Voltage burden		(50 ~ 60) Hz 1.25 VA (1.25 ~ 3.75) VA (3.75 ~ 5) VA (5 ~ 10) VA (10 ~ 100) VA	7.4×10^{-4} 4.6×10^{-4} 3.7×10^{-4} 3.0×10^{-4} 3.1×10^{-4}	
Voltage burden factor		0.5 ~ 1	1.6×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power supplies	40312	(10 mV) 40 Hz ~ 5 kHz	2.2×10^{-4}	Voltage Standard, Multimeter, Current Shunt/ SICT-CP-40312
		(10 ~ 100) mV 40 Hz ~ 5 kHz	7.9×10^{-5}	
		(100 mV ~ 1 V) 40 Hz ~ 5 kHz	6.7×10^{-5}	
		(1 ~ 10) V 40 Hz ~ 5 kHz	6.8×10^{-5}	
		(10 ~ 100) V 40 Hz ~ 5 kHz	7.0×10^{-5}	
		(100 ~ 600) V 40 Hz ~ 5 kHz	4.1×10^{-5}	
		(600 ~ 1 000) V 40 Hz ~ 5 kHz	7.1×10^{-5}	
		Frequency		
		10 Hz	9.5×10^{-5}	
		(10 ~ 50) Hz	1.9×10^{-5}	
		(50 ~ 100) Hz	7.7×10^{-6}	
		(0.1 ~ 1) kHz	8.4×10^{-7}	
		(1 ~ 5) kHz	3.8×10^{-7}	
		AC Current		
		(1 mA) (50 ~ 60) Hz	6.4×10^{-4}	
		(1 ~ 10) mA (50 ~ 60) Hz	3.6×10^{-4}	
		(10 ~ 100) mA (50 ~ 60) Hz	2.4×10^{-4}	
		(100 mA ~ 1 A) (50 ~ 60) Hz	2.1×10^{-4}	
		(1 ~ 10) A (50 ~ 60) Hz	2.3×10^{-4}	
		(10 ~ 20) A (50 ~ 60) Hz	4.0×10^{-4}	
		(20 ~ 30) A (50 ~ 60) Hz	6.4×10^{-4}	
		(30 ~ 50) A (50 ~ 60) Hz	4.2×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power supplies	40312	(±)		Voltage Standard, Multimeter, Current Shunt/ SICT-CP-40312
		0 mV	5.8 μ V	
		(0 ~ 10) mV	5.8 $\times 10^{-4}$	
		(10 ~ 100) mV	5.8 $\times 10^{-5}$	
		(0.1 ~ 100) V	7.7 $\times 10^{-6}$	
		(100 ~ 600) V	1.3 $\times 10^{-5}$	
		(600 ~ 1 000) V	6.6 $\times 10^{-5}$	
		(1 ~ 10) mA	5.8 $\times 10^{-3}$	
		(10 ~ 100) mA	5.9 $\times 10^{-4}$	
		(0.1 ~ 1) A	2.4 $\times 10^{-4}$	
DC Current	40312	(1 ~ 10) A	3.1 $\times 10^{-4}$	
		(10 ~ 300) A	2.4 $\times 10^{-4}$	
		(300 ~ 500) A	2.6 $\times 10^{-4}$	
		(500 ~ 1 000) A	4.7 $\times 10^{-5}$	
		(1 000 ~ 3 000) A	5.1 $\times 10^{-4}$	
		(0 ~ 2) mV	0.16 mV	
		(2 ~ 20) mV	7.8 $\times 10^{-2}$	
		(20 ~ 200) mV	8.2 $\times 10^{-3}$	
		(0.1 ~ 0.4) mV	3.8 $\times 10^{-1}$	
		(0.4 ~ 0.6) mV	1.1 $\times 10^{-1}$	
Load Regulation	40312	(0.6 ~ 1) mV	7.3 $\times 10^{-2}$	
		(1 ~ 10) mV	4.4 $\times 10^{-2}$	
		(10 ~ 50) mV	7.1 $\times 10^{-2}$	
		(50 ~ 60) Hz		
		0.5 %	0.050 %	
		(0.5 ~ 10) %	0.051 %	
		(10 ~ 20) %	0.082 %	
		(50 ~ 60) Hz		
		0.5 %	0.050 %	
		(0.5 ~ 20) %	0.051 %	
Puncture/safety testers	40313	(±)		AC/DC Kilovoltmeter, High Voltage Digital Meter, Reference Multimeter/ SICT-CP-40313
		0 kV	0.58 V	
		(0 ~ 0.5) kV	1.2 $\times 10^{-3}$	
		(0.5 ~ 1) kV	6.1 $\times 10^{-4}$	
		(1 ~ 2) kV	3.0 $\times 10^{-4}$	
		(2 ~ 100) kV	2.3 $\times 10^{-4}$	
		(100 ~ 200) kV	1.2 $\times 10^{-2}$	
		(50 ~ 60) Hz		
		0.01 kV	0.58 V	
		(0.01 ~ 0.5) kV	1.2 $\times 10^{-3}$	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Puncture/safety testers	40313	(50 ~ 60) Hz		AC/DC Kilovoltmeter, High Voltage Digital Meter, Reference Multimeter/ SICT-CP-40313
		0.1 mA	5.3×10^{-4}	
		(0.1 ~ 0.5) mA	4.4×10^{-4}	
		(0.5 ~ 1) mA	7.3×10^{-4}	
		(1 ~ 2) mA	7.1×10^{-4}	
		(2 ~ 5) mA	4.4×10^{-4}	
		(5 ~ 10) mA	3.6×10^{-4}	
		(10 ~ 20) mA	7.1×10^{-4}	
		(20 ~ 50) mA	4.4×10^{-4}	
		(50 ~ 100) mA	7.3×10^{-4}	
DC Breaking Current		0.1 mA	3.9×10^{-4}	
		(0.1 ~ 0.5) mA	1.3×10^{-4}	
		(0.5 ~ 1) mA	6.4×10^{-4}	
		(1 ~ 2) mA	3.2×10^{-4}	
		(2 ~ 5) mA	1.3×10^{-4}	
		(5 ~ 10) mA	6.5×10^{-5}	
		(10 ~ 20) mA	3.3×10^{-4}	
		(20 ~ 50) mA	1.4×10^{-4}	
		(50 ~ 100) mA	6.4×10^{-4}	
Resistance		1 mΩ	8.6×10^{-4}	
		(1 ~ 10) mΩ	7.2×10^{-4}	
		10 mΩ ~ 100 kΩ	6.8×10^{-4}	
Insulation Voltage		1 V	6.4×10^{-4}	
		(1 ~ 10) V	6.4×10^{-5}	
		(10 ~ 25) V	2.5×10^{-4}	
		(25 ~ 50) V	1.3×10^{-4}	
		(50 ~ 100) V	6.4×10^{-5}	
		(100 ~ 250) V	2.5×10^{-4}	
		(250 ~ 500) V	1.3×10^{-4}	
		(500 ~ 1 000) V	6.4×10^{-5}	
		(1 000 ~ 5 000) V	6.5×10^{-3}	
		(5 000 ~ 10 000) V	6.1×10^{-3}	
Insulation Resistance		1 kΩ	7.1×10^{-5}	
		(1 ~ 10) kΩ	3.7×10^{-5}	
		(10 ~ 100) kΩ	2.5×10^{-5}	
		(0.1 ~ 1) MΩ	3.1×10^{-5}	
		(1 ~ 10) MΩ	9.5×10^{-5}	
		(10 ~ 100) MΩ	2.4×10^{-5}	
		(0.1 ~ 1) GΩ	3.1×10^{-5}	
		(1 ~ 10) GΩ	6.1×10^{-5}	
		(10 ~ 100) GΩ	1.3×10^{-4}	
		(0.1 ~ 1) TΩ	2.6×10^{-4}	
		10 TΩ	6.3×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Puncture/safety testers	40313			AC/DC Kilovoltmeter, High Voltage Digital Meter, Reference Multimeter/ SICT-CP-40313
Leakage current(DC)		0 μA	7.0 nA	
		(0 ~ 1) μA	7.0×10^{-3}	
		(1 ~ 2) μA	3.6×10^{-3}	
		(2 ~ 5) μA	1.4×10^{-3}	
		(5 ~ 10) μA	7.4×10^{-4}	
		(10 ~ 20) μA	4.0×10^{-4}	
		(20 ~ 50) μA	1.8×10^{-4}	
		(50 ~ 100) μA	1.3×10^{-4}	
		(100 ~ 200) μA	8.5×10^{-5}	
		(0.2 ~ 100) mA	6.1×10^{-4}	
Leakage current(AC)		(20 μA)		
		10 Hz	1.3×10^{-3}	
		(10 ~ 20) Hz	8.5×10^{-4}	
		20 Hz ~ 1 kHz	7.0×10^{-4}	
		(1 ~ 5) kHz	1.3×10^{-3}	
		(5 ~ 10) kHz	5.5×10^{-3}	
		(20 ~ 50) μA		
		10 Hz	6.8×10^{-4}	
		(10 ~ 20) Hz	4.4×10^{-4}	
		20 Hz ~ 1 kHz	3.4×10^{-4}	
		(1 ~ 5) kHz	6.8×10^{-4}	
		(5 ~ 10) kHz	2.8×10^{-3}	
		(50 ~ 100) μA		
		10 Hz	4.9×10^{-4}	
		(10 ~ 20) Hz	3.2×10^{-4}	
		20 Hz ~ 1 kHz	2.3×10^{-4}	
		(1 ~ 5) kHz	4.9×10^{-4}	
		(5 ~ 10) kHz	4.0×10^{-4}	
		(100 ~ 200) μA		
		10 Hz	3.9×10^{-4}	
		(10 ~ 20) Hz	2.5×10^{-4}	
		20 Hz ~ 1 kHz	1.7×10^{-4}	
		(1 ~ 5) kHz	4.0×10^{-4}	
		(5 ~ 10) kHz	1.7×10^{-3}	
		(200 ~ 500) μA		
		10 Hz	4.4×10^{-4}	
		(10 ~ 20) Hz	3.2×10^{-4}	
		20 Hz ~ 1 kHz	2.4×10^{-4}	
		(1 ~ 5) kHz	5.4×10^{-4}	
		(5 ~ 10) kHz	2.8×10^{-3}	
		500 μA ~ 1 mA		
		10 Hz	7.0×10^{-4}	
		(10 ~ 20) Hz	6.6×10^{-4}	
		20 Hz ~ 1 kHz	6.3×10^{-4}	
		(1 ~ 5) kHz	7.2×10^{-4}	
		(5 ~ 10) kHz	2.1×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Puncture/safety testers	40313	(1 ~ 100) mA		AC/DC Kilovoltmeter, High Voltage Digital Meter, Reference Multimeter/ SICT-CP-40313	
		10 Hz	7.0×10^{-4}		
		(10 ~ 20) Hz	6.6×10^{-4}		
		20 Hz ~ 1 kHz	6.3×10^{-4}		
		(1 ~ 5) kHz	7.0×10^{-4}		
		(5 ~ 10) kHz	2.7×10^{-3}		
		(60 Hz)			
		1 A	1.2×10^{-3}		
		(1 ~ 3) A	1.5×10^{-3}		
		(3 ~ 20) A	9.7×10^{-4}		
Output AC Current		(20 ~ 30) A	1.0×10^{-3}		
		(30 ~ 60) A	8.4×10^{-4}		
		(60 ~ 100) A	1.0×10^{-3}		
		(100 ~ 150) A	4.6×10^{-3}		
		(150 ~ 200) A	3.7×10^{-3}		
		1 s	5.8×10^{-6}		
		(1 ~ 100) s	5.8×10^{-6}		
		(100 ~ 1 000) s	8.2×10^{-6}		
		(1 000 ~ 10 000) s	5.8×10^{-5}		
Power recorders	40314	(50 ~ 60) Hz		Power Energy Calibrator/ SICT-CP-40314	
		0 mW	$70 \mu\text{W}$		
		(0 ~ 0.22) mW	2.1×10^{-1}		
		(0.22 ~ 1.1) mW	4.1×10^{-2}		
		(1.1 ~ 2.2) mW	2.1×10^{-2}		
		(2.2 ~ 11) mW	4.1×10^{-3}		
		(11 ~ 22) mW	2.1×10^{-3}		
		(22 ~ 44) mW	1.0×10^{-3}		
		(44 ~ 66) mW	7.0×10^{-4}		
		(66 ~ 88) mW	5.3×10^{-4}		
		(88 ~ 110) mW	4.3×10^{-4}		
		(110 ~ 480) mW	2.1×10^{-4}		
		480 mW ~ 12 kW	1.2×10^{-4}		
		(12 ~ 24) kW	6.8×10^{-4}		
		(24 ~ 300) kW	1.2×10^{-3}		
		(300 ~ 600) kW	1.4×10^{-3}		
		0 mW	61nW		
		(0 ~ 1) mW	7.7×10^{-5}		
		(1 ~ 10) mW	4.8×10^{-5}		
		(10 ~ 100) mW	6.1×10^{-5}		
		(0.1 ~ 100) W	1.1×10^{-4}		
		(0.1 ~ 10) kW	1.9×10^{-4}		
		(10 ~ 20) kW	1.5×10^{-4}		
		(20 ~ 1 000) kW	1.0×10^{-3}		
		(1 000 ~ 2 500) kW	1.3×10^{-3}		

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Current transformer test sets	40315	(5 ~ 1 500) A (-19.99 ~ 19.99) %	0.020 %	Current transforme, Ratio transformers/ SICT-CP-40315
		(-680 ~ 680) '	0.70'	
Current/turn current coil transformers transducers	40316	(5 ~ 10 000) A (-19.99 ~ 19.99) %	0.020 %	Current transforme/ SICT-CP-40316
		(-680 ~ 680) '	0.70'	
	Current Coil transducers	(AC) 2 ~ 50	0.10 %	
		(DC) 2 ~ 50	0.10 %	
		(±) (10 A) 50 : 1 ~ 5 000 : 1	1.2×10^{-4}	
		(10 ~ 1 000) A 50 : 1 ~ 5 000 : 1	1.9×10^{-4}	
		(1 000 ~ 2 000) A 50 : 1 ~ 5 000 : 1	2.5×10^{-4}	
AC voltmeters	40318	(600 µV) 1 kHz	7.8×10^{-3}	Reference Multimeter, Calibrator/ SICT-CP-40318
		(600 µV ~ 1 mV) 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	5.0×10^{-3} 4.8×10^{-3} 6.5×10^{-3}	
		(1 ~ 3) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	2.0×10^{-3} 1.7×10^{-3} 2.8×10^{-3}	
		(3 ~ 10) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	7.7×10^{-4} 5.7×10^{-4} 1.2×10^{-3}	
		(10 ~ 30) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	8.0×10^{-4} 3.7×10^{-4} 1.1×10^{-3}	
		(30 ~ 100) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	4.3×10^{-4} 1.7×10^{-4} 5.7×10^{-4}	
		(100 mV ~ 10 V) 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	4.9×10^{-4} 1.1×10^{-4} 2.6×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltmeters	40318			Reference Multimeter, Calibrator/ SICT-CP-40318
	AC Voltage	(10 ~ 100) V		
		10 Hz	5.3×10^{-4}	
		10 Hz ~ 10 kHz	1.3×10^{-4}	
		(10 ~ 100) kHz	3.6×10^{-4}	
		(100 ~ 1 000) V		
		50 Hz	3.7×10^{-4}	
		50 Hz ~ 1 kHz	1.1×10^{-4}	
		(1 mV)		
		100 kHz	6.5×10^{-3}	
		100 kHz ~ 1 MHz	2.7×10^{-2}	
		(1 ~ 10) mV		
		100 kHz	1.2×10^{-3}	
		100 kHz ~ 1 MHz	5.6×10^{-3}	
		(10 ~ 100) mV		
		100 kHz	5.7×10^{-4}	
		100 kHz ~ 1 MHz	3.7×10^{-3}	
		(100 mV ~ 1 V)		
		100 kHz	1.5×10^{-4}	
		100 kHz ~ 1 MHz	2.3×10^{-3}	
		(1 ~ 10) V		
		100 kHz	4.7×10^{-5}	
		100 kHz ~ 1 MHz	7.0×10^{-4}	
		(10 ~ 20) V		
		100 kHz	7.6×10^{-5}	
		100 kHz ~ 1 MHz	1.3×10^{-3}	
		(25 mV)		
		1 MHz	1.9×10^{-2}	
		(1 ~ 30) MHz	2.3×10^{-2}	
		(25 ~ 100) mV		
		1 MHz	2.3×10^{-2}	
		(1 ~ 30) MHz	2.8×10^{-2}	
		(100 ~ 300) mV		
		1 MHz	3.7×10^{-2}	
		(1 ~ 30) MHz	4.0×10^{-2}	
		(300 mV ~ 1 V)		
		1 MHz	2.4×10^{-2}	
		(1 ~ 30) MHz	2.7×10^{-2}	
		(1 ~ 2) V		
		1 MHz	1.5×10^{-2}	
		(1 ~ 30) MHz	1.8×10^{-2}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltmeters	40318			Reference Multimeter, Calibrator/ SICT-CP-40318
AC Output Voltage		(1 mV) 10 Hz ~ 10 kHz (10 ~ 100) kHz (1 ~ 10) mV 10 Hz ~ 10 kHz (10 ~ 100) kHz (10 ~ 100) mV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz (100 mV ~ 1 V) 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	1.9×10^{-3} 3.1×10^{-3} 2.6×10^{-4} 3.7×10^{-4} 9.2×10^{-5} 4.8×10^{-5} 9.1×10^{-5} 7.2×10^{-5} 1.8×10^{-5} 5.9×10^{-5}	
DC Output Voltage		1 mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V	7.5×10^{-4} 7.6×10^{-5} 9.4×10^{-6} 2.2×10^{-5}	
Watt hour meters	40319			Power Calibrator/ SICT-CP-40319
Watt Hour		(50 ~ 60) Hz 0 mWh (0 ~ 480) mWh 480 mWh ~ 12 kWh (12 ~ 24) kWh (24 ~ 300) kWh (300 ~ 600) kWh (DC) 0 mWh (0 ~ 1) mWh (1 ~ 100) mWh 100 mWh ~ 100 Wh 100 Wh ~ 10 kWh (10 ~ 20) kWh (20 ~ 1 000) kWh (1 000 ~ 2 500) kWh	$44 \mu\text{Wh}$ 4.0×10^{-4} 3.6×10^{-4} 7.6×10^{-4} 1.3×10^{-3} 1.4×10^{-3} 61nWh 3.5×10^{-4} 3.4×10^{-4} 3.5×10^{-4} 3.8×10^{-4} 3.7×10^{-4} 1.1×10^{-3} 1.3×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Ratio transformers	Ratio	40321	PT (±) (0.1 ~ 1.000) %	0.006 %
			(1.000 ~ 19.00) %	0.01 %
			CT (±) (0.1 ~ 1.000) %	0.019 %
			(1.000 ~ 19.00) %	0.02 %
	Phase	PT (±) (0.040 ~ 1.999) '	0.060 '	
		(1.999 ~ 19.99) '	0.06 '	
		(19.99 ~ 199.9) '	0.2 '	
		(199.9 ~ 600) '	1 '	
		CT (±) (0.040 ~ 1.999) '	0.060 '	
		(1.999 ~ 19.99) '	0.06 '	
		(19.99 ~ 199.9) '	0.2 '	
		(199.9 ~ 600) '	1 '	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF amplifiers	40401			Reference Multimeter/ SICT-CP-40401
Amplifier		(DC) 1 mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 1 000) V (10 Hz ~ 10 kHz) 1 mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 1 000) V (10 ~ 100) kHz 1 mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 1 000) V	0.4 μ V 7.4×10^{-5} 6.1×10^{-5} 6.0×10^{-5} 1.0×10^{-4} 1.7 μ V 2.6×10^{-4} 1.1×10^{-4} 9.0×10^{-5} 1.0×10^{-4} 3.1 μ V 3.7×10^{-4} 1.1×10^{-4} 8.0×10^{-5} 1.0×10^{-4}	
DC/LF attenuators	40402	10 Hz ~ 100 kHz (0 ~ -20) dB (-20 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB		Reference Multimeter/ SICT-CP-40402
Multimeter calibrators	40403	(\pm) 0 mV (0 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V DC Current (\pm) 1 nA (1 ~ 100) nA 100 nA ~ 10 A (10 ~ 50) A (50 ~ 100) A AC Voltage (1 mV) 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (1 ~ 2) mV 10 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz	0.001 9 dB 0.001 7 dB 0.005 5 dB 0.008 7 dB 0.05 μ V 3.3×10^{-6} 1.4×10^{-6} 1.0×10^{-6} 1.6×10^{-6} 2.1×10^{-6} 7.0 pA 4.7×10^{-3} 1.2×10^{-5} 4.0×10^{-5} 4.4×10^{-5} 1.7 $\times 10^{-3}$ 3.0×10^{-3} 1.2×10^{-2} 1.1 $\times 10^{-3}$ 1.7×10^{-3} 7.7×10^{-3}	Reference Multimeter/ SICT-CP-40403

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators	40403			Reference Multimeter/ SICT-CP-40403
AC Voltage		(2 ~ 5) mV		
		10 Hz	6.4×10^{-4}	
		10 Hz ~ 10 kHz	5.8×10^{-4}	
		(10 ~ 100) kHz	1.0×10^{-3}	
		100 kHz ~ 1 MHz	5.4×10^{-3}	
		(5 ~ 10) mV		
		10 Hz	4.2×10^{-4}	
		10 Hz ~ 10 kHz	3.5×10^{-4}	
		(10 ~ 100) kHz	5.8×10^{-4}	
		100 kHz ~ 1 MHz	3.9×10^{-3}	
		(10 ~ 20) mV		
		10 Hz	1.8×10^{-4}	
		10 Hz ~ 10 kHz	1.4×10^{-4}	
		(10 ~ 100) kHz	2.2×10^{-4}	
		100 kHz ~ 1 MHz	2.2×10^{-3}	
		(20 ~ 50) mV		
		10 Hz	1.4×10^{-4}	
		10 Hz ~ 10 kHz	9.2×10^{-5}	
		(10 ~ 100) kHz	1.6×10^{-4}	
		100 kHz ~ 1 MHz	1.4×10^{-3}	
		(50 ~ 100) mV		
		10 Hz	1.1×10^{-4}	
		10 Hz ~ 10 kHz	6.6×10^{-5}	
		(10 ~ 100) kHz	1.2×10^{-4}	
		100 kHz ~ 1 MHz	1.3×10^{-3}	
		(100 ~ 200) mV		
		10 Hz	8.2×10^{-5}	
		10 Hz ~ 10 kHz	3.9×10^{-5}	
		(10 ~ 100) kHz	7.6×10^{-5}	
		100 kHz ~ 1 MHz	1.1×10^{-3}	
		(200 ~ 500) mV		
		10 Hz	7.8×10^{-5}	
		10 Hz ~ 10 kHz	3.6×10^{-5}	
		(10 ~ 100) kHz	7.1×10^{-5}	
		100 kHz ~ 1 MHz	1.1×10^{-3}	
		(0.5 ~ 1) V		
		10 Hz	7.6×10^{-5}	
		10 Hz ~ 10 kHz	3.3×10^{-5}	
		(10 ~ 100) kHz	6.6×10^{-5}	
		100 kHz ~ 1 MHz	1.1×10^{-3}	
		(1 ~ 2) V		
		10 Hz	7.1×10^{-5}	
		10 Hz ~ 10 kHz	2.7×10^{-5}	
		(10 ~ 100) kHz	5.8×10^{-5}	
		100 kHz ~ 1 MHz	1.0×10^{-3}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators	40403			Reference Multimeter/ SICT-CP-40403
AC Voltage		(2 ~ 5) V		
		10 Hz	7.2×10^{-5}	
		10 Hz ~ 10 kHz	2.6×10^{-5}	
		(10 ~ 100) kHz	7.5×10^{-5}	
		100 kHz ~ 1 MHz	1.4×10^{-3}	
		(5 ~ 20) V		
		10 Hz	7.2×10^{-5}	
		10 Hz ~ 10 kHz	2.8×10^{-5}	
		(10 ~ 100) kHz	7.5×10^{-5}	
		100 kHz ~ 1 MHz	1.4×10^{-3}	
		(20 ~ 50) V		
		10 Hz	7.2×10^{-5}	
		10 Hz ~ 10 kHz	3.0×10^{-5}	
		(10 ~ 100) kHz	8.0×10^{-5}	
		(50 ~ 200) V		
		10 Hz	7.4×10^{-5}	
		10 Hz ~ 10 kHz	3.3×10^{-5}	
		(10 ~ 100) kHz	8.5×10^{-5}	
		(200 ~ 1 000) V		
		10 Hz	7.7×10^{-5}	
		10 Hz ~ 10 kHz	3.3×10^{-5}	
		(10 ~ 100) kHz	5.8×10^{-4}	
AC Current		(10 µA)		
		10 Hz ~ 10 kHz	2.6×10^{-3}	
		(10 ~ 100) µA		
		10 Hz ~ 1 kHz	3.6×10^{-4}	
		(1 ~ 10) kHz	6.4×10^{-4}	
		(0.1 ~ 1) mA		
		10 Hz	9.8×10^{-5}	
		10 Hz ~ 1 kHz	7.5×10^{-5}	
		(1 ~ 10) kHz	9.4×10^{-5}	
		(1 ~ 100) mA		
		10 Hz	7.8×10^{-5}	
		10 Hz ~ 1 kHz	4.6×10^{-5}	
		(1 ~ 10) kHz	4.2×10^{-5}	
		(0.1 ~ 1) A		
		10 Hz	8.1×10^{-5}	
		10 Hz ~ 1 kHz	4.9×10^{-5}	
		(1 ~ 10) kHz	4.4×10^{-5}	
		(1 ~ 2) A		
		10 Hz	7.9×10^{-5}	
		10 Hz ~ 1 kHz	4.7×10^{-5}	
		(1 ~ 10) kHz	4.5×10^{-5}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators	40403			Reference Multimeter/ SICT-CP-40403
AC Current		(2 ~ 5) A		
		10 Hz	8.2×10^{-5}	
		10 Hz ~ 1 kHz	5.2×10^{-5}	
		(1 ~ 10) kHz	5.0×10^{-5}	
		(5 ~ 10) A		
		10 Hz	8.6×10^{-5}	
		10 Hz ~ 1 kHz	5.9×10^{-5}	
		(1 ~ 10) kHz	7.8×10^{-5}	
		(10 ~ 20) A		
		10 Hz	9.3×10^{-5}	
		10 Hz ~ 1 kHz	6.8×10^{-5}	
		(1 ~ 10) kHz	7.8×10^{-5}	
		(20 ~ 50) A		
		10 Hz	1.0×10^{-4}	
		10 Hz ~ 1 kHz	8.3×10^{-5}	
		(1 ~ 10) kHz	1.1×10^{-4}	
		(50 ~ 100) A		
		10 Hz	1.2×10^{-4}	
		10 Hz ~ 1 kHz	9.7×10^{-5}	
		(1 ~ 10) kHz	1.3×10^{-4}	
		(100 ~ 200) A		
		60 Hz	4.5×10^{-4}	
Resistance		0 Ω	$0.14 \mu\Omega$	
		(0 ~ 1) Ω	6.6×10^{-6}	
		(1 ~ 1.9) Ω	8.4×10^{-6}	
		(1.9 ~ 10) Ω	3.6×10^{-6}	
		(10 ~ 19) Ω	2.6×10^{-6}	
		(19 ~ 100) Ω	2.8×10^{-6}	
		(0.1 ~ 1) kΩ	2.5×10^{-6}	
		(1 ~ 1.9) kΩ	3.8×10^{-6}	
		(1.9 ~ 10) kΩ	2.0×10^{-6}	
		(10 ~ 19) kΩ	1.3×10^{-6}	
		(19 ~ 100) kΩ	1.9×10^{-6}	
		(100 ~ 190) kΩ	2.0×10^{-6}	
		(0.19 ~ 1) MΩ	2.9×10^{-6}	
		(1 ~ 1.9) MΩ	3.1×10^{-6}	
		(1.9 ~ 10) MΩ	3.6×10^{-6}	
		(10 ~ 19) MΩ	2.9×10^{-6}	
		(19 ~ 100) MΩ	1.5×10^{-5}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators	40403			Reference Multimeter/ SICT-CP-40403
Multimeter calibrators(property) (Digital sampling)				
AC Voltage		(1 mV) 0.1 Hz ~ 3 kHz	8.4×10^{-4}	
		(1 ~ 2) mV 0.1 Hz ~ 3 kHz	4.2×10^{-4}	
		(2 ~ 3) mV 0.1 Hz ~ 3 kHz	2.8×10^{-4}	
		(3 ~ 5) mV 0.1 Hz ~ 3 kHz	1.7×10^{-4}	
		(5 ~ 10) mV 0.1 Hz ~ 3 kHz	8.8×10^{-5}	
		(10 ~ 20) mV 0.1 Hz ~ 3 kHz	4.8×10^{-5}	
		(20 ~ 30) mV 0.1 Hz ~ 3 kHz	3.6×10^{-5}	
		(30 ~ 50) mV 0.1 Hz ~ 3 kHz	3.0×10^{-5}	
		(50 ~ 100) mV 0.1 Hz ~ 3 kHz	2.6×10^{-5}	
		(100 ~ 200) mV 0.1 Hz ~ 3 kHz	4.8×10^{-5}	
		(200 ~ 300) mV 0.1 Hz ~ 3 kHz	3.6×10^{-5}	
		(300 ~ 500) mV 0.1 Hz ~ 3 kHz	2.8×10^{-5}	
		(500 mV ~ 1 V) 0.1 Hz ~ 3 kHz	2.4×10^{-5}	
		(1 ~ 2) V 0.1 Hz ~ 3 kHz	4.8×10^{-5}	
		(2 ~ 3) V 0.1 Hz ~ 3 kHz	3.6×10^{-5}	
		(3 ~ 5) V 0.1 Hz ~ 3 kHz	3.0×10^{-5}	
		(5 ~ 10) V 0.1 Hz ~ 3 kHz	2.6×10^{-5}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators	40403			Reference Multimeter/ SICT-CP-40403
Multimeter calibrators(property) (Digital sampling)				
AC Voltage		(10 ~ 30) V 10 Hz ~ 3 kHz	3.6×10^{-5}	
		(30 ~ 50) V 10 Hz ~ 3 kHz	2.8×10^{-5}	
		(50 ~ 100) V 10 Hz ~ 3 kHz	2.4×10^{-5}	
		(100 ~ 200) V 10 Hz ~ 3 kHz	4.8×10^{-5}	
		(200 ~ 1 000) V 50 Hz ~ 1 kHz	2.4×10^{-5}	
Oscilloscope calibrators	40404			Calibrator/ SICT-CP-40404
DC Voltage Amplitude		(±) 0 mV (0 ~ 1) mV (1 ~ 2) mV (2 ~ 5) mV (5 ~ 10) mV (10 ~ 20) mV (20 ~ 50) mV (50 ~ 100) mV (100 ~ 200) mV (200 ~ 500) mV (0.5 ~ 1) V (1 ~ 2) V (2 ~ 5) V (5 ~ 10) V (10 ~ 20) V (20 ~ 50) V (50 ~ 100) V (100 ~ 200) V (200 ~ 500) V	0.50 μV 4.2×10^{-4} 2.1×10^{-5} 8.5×10^{-5} 4.3×10^{-5} 2.1×10^{-5} 8.5×10^{-6} 7.2×10^{-6} 3.8×10^{-6} 2.8×10^{-6} 6.3×10^{-6} 3.7×10^{-6} 2.9×10^{-6} 6.3×10^{-6} 3.8×10^{-6} 3.3×10^{-6} 6.4×10^{-6} 3.9×10^{-6} 3.8×10^{-6}	
AC Voltage Amplitude		(10 Hz ~ 10 kHz) 1 mV (1 ~ 2) mV (2 ~ 10) mV (10 ~ 500) mV (0.5 ~ 100) V (100 ~ 200) V	0.76 μV 8.3×10^{-5} 8.4×10^{-5} 6.0×10^{-5} 5.8×10^{-5} 4.0×10^{-5}	
Sine Wave Generator		(100 ~ 600) mV 50 kHz (50 ~ 500) kHz 0.5 MHz ~ 1 GHz (1 ~ 6) GHz	0.58 mV 1.0×10^{-3} 1.7×10^{-2} 1.9×10^{-2}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Oscilloscope calibrators	40404			Calibrator/ SICT-CP-40404
Sine Wave Generator		(600 mV ~ 1 V) 50 kHz (50 ~ 500) kHz 0.5 MHz ~ 1 GHz (1 ~ 6) GHz	0.58 mV 1.0×10^{-3} 1.7×10^{-2} 1.9×10^{-2}	
Time Marker Generator		(0.1 ~ 1) ns (1 ~ 10) ns (10 ~ 100) ns 0.1 μ s ~ 10 ms (10 ~ 100) ms (0.1 ~ 1) s (1 ~ 5) s	5.8×10^{-8} 6.5×10^{-9} 3.1×10^{-9} 5.8×10^{-8} 6.1×10^{-9} 5.8×10^{-8} 1.2×10^{-8}	
Impedance Mesurement		(50 ~ 75) Ω 75 Ω ~ 1 M Ω	1.7×10^{-4} 2.1×10^{-4}	
CD/DVD meters/analyzers	40405			Modulation Domain Analyzer/ SICT-CP-40405
Jitter		(1.0 ~ 60.0) ns 1 % 2 % 4 % 8 % 10 % 15 %	1.7×10^{-3} 0.05 % 0.09 % 0.19 % 0.36 % 0.44 % 0.67 %	
Video signal generators	40406			Video Measurement/ SICT-CP-40406
NTSC, PAL Multiburst		(0.1 ~ 1) MHz (1 ~ 2) MHz (2 ~ 6) MHz	6.0×10^{-2} 6.2×10^{-3} 3.1×10^{-3}	
NTSC, PAL, SECAM Pulse and Bar		(0 ~ 300) ns (0 ~ 1 000) mV	4.2×10^{-4} 3.5×10^{-3}	
NTSC, PAL , SECAM Frequency		1 Hz ~ 10 MHz	1.6×10^{-9}	
Video frequency		(10 ~ 100) Hz 100 Hz ~ 500 MHz	6.2×10^{-8} 6.2×10^{-9}	
Video level		(30 ~ 600) mV (600 ~ 1 200) mV	2.6×10^{-3} 2.3×10^{-3}	
TTL Sync level		(1 ~ 5) V	2.7×10^{-3}	
D-TV Level		(30 ~ 600) mV (600 ~ 1 200) mV	2.6×10^{-3} 2.3×10^{-3}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Video signal generators NTSC, PAL, H-Timming(Level1) (Time)	40406	(0 ~ 100) mV	2.6×10^{-3}	Video Measurement/ SICT-CP-40406
		(100 ~ 1 000) mV	3.4×10^{-3}	
		(0 ~ 254) ns	1.2×10^{-2}	
		(254 ~ 300) ns	3.8×10^{-3}	
		300 ns ~ 3 μ s	3.2×10^{-3}	
		(3 ~ 7) μ s	7.4×10^{-3}	
		(7 ~ 10) μ s	4.2×10^{-3}	
		NTSC, PAL Color Bar(Luminance Level)	0.06 mV	
		(100 ~ 1 000) mV	3.4×10^{-3}	
		NTSC, PAL Color Bar(Chrominance Level)	0.06 mV	
		(100 ~ 1 000) mV	3.4×10^{-3}	
NTSC, PAL Color Bar(Phase)		(0 ~ 360)°	0.13°	
SECAM Color Bar Frequency		(D'R & D'B) (3 ~ 5) MHz	1.2×10^{-3}	
RF Output frequency		10 kHz ~ 10 MHz	6.0×10^{-4}	
		(10 ~ 100) MHz	6.0×10^{-5}	
		(100 ~ 1 000) MHz	6.0×10^{-6}	
RF Output level		(0.1 ~ 10) mV	1.4×10^{-2}	
		(10 ~ 500) mV	1.3×10^{-2}	
Sound Frequency		10 Hz ~ 100 kHz	6.1×10^{-8}	
		100 kHz ~ 1 MHz	6.1×10^{-7}	
Audio distortion analyzers/meters Input Frequency Input Level Flatness Test Input DC Voltage Input Distortion	40407	1 Hz ~ 200 kHz	6.1×10^{-7}	Calibrator/ SICT-CP-40407
		(10 ~ 100) kHz	0.008 3 dB	
		0 mV	0.27 μ V	
		(0 ~ 1) mV	5.8×10^{-3}	
		(1 ~ 10) mV	5.8×10^{-4}	
		10 mV ~ 300 V	5.8×10^{-4}	
		(100 Hz ~ 10 kHz)		
		(-10 ~ -40) dB	0.005 8 dB	
		(-40 ~ -50) dB	0.006 0 dB	
		(-50 ~ -60) dB	0.006 8 dB	
		(-60 ~ -70) dB	0.012 dB	
		(-70 ~ -80) dB	0.028 dB	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Audio distortion analyzers/meters				
Input Distortion	40407	(10 kHz ~ 50 kHz) (-10 ~ -40) dB (-40 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	0.005 9 dB 0.006 3 dB 0.008 2 dB 0.019 dB 0.052 dB	Calibrator/ SICT-CP-40407
Input AC Voltage		(10 ~ 100) Hz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 100 V (100 ~ 300) V	9.0×10^{-4} 4.0×10^{-4} 4.2×10^{-4} 5.3×10^{-4}	
		(100 Hz ~ 1 kHz) (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 10 V (10 ~ 100) V (100 ~ 300) V	8.4×10^{-4} 1.8×10^{-4} 1.1×10^{-4} 1.0×10^{-4} 2.3×10^{-4}	
		(1 ~ 10) kHz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 10 V (10 ~ 100) V	8.4×10^{-4} 1.8×10^{-4} 1.5×10^{-4} 2.7×10^{-4}	
		(10 ~ 100) kHz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 1 V (1 ~ 10) V (10 ~ 100) V	1.4×10^{-3} 7.6×10^{-4} 4.8×10^{-4} 4.1×10^{-4} 3.4×10^{-4}	
Input Attenuation		(10 Hz) (30 ~ -50) dB (-50 ~ -60) dB (-60 ~ -80) dB	0.006 8 dB 0.016 dB 0.052 dB	
		(10 Hz ~ 10 kHz) (30 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	0.008 3 dB 0.014 dB 0.042 dB	
		(10 ~ 100) kHz (30 ~ -50) dB (-50 ~ -70) dB (-70 ~ -80) dB	0.009 1 dB 0.023 dB 0.057 dB	
Input Impedance		300 Ω ~ 200 kΩ	3.1×10^{-4}	
Input Filter		(10 Hz ~ 100 kHz) 1 V	8.3×10^{-4}	
(Distortion meter calibrator)				
Distortion		(400 Hz , 1 kHz) (-10 ~ -20) dB (-20 ~ -40) dB (-40 ~ -60) dB (-60 ~ -80) dB	0.15 dB 0.14 dB 0.17 dB 0.26 dB	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF filters	40408			Audio Analyzer/ SICT-CP-40408
Filter		10 Hz ~ 50 kHz (50 ~ 100) kHz (100 ~ 150) kHz	5.8 × 10 ⁻⁴ 1.2 × 10 ⁻³ 5.8 × 10 ⁻³	
LF/audio signal analyzers	40409			Calibrator, Reference Multimeter/ SICT-CP-40409
Output Frequency		1 Hz ~ 200 kHz	5.8 × 10 ⁻⁶	
AC Output Level		(10 ~ 100) Hz (1 ~ 10) mV 10 mV ~ 30 V (-20 ~ 10) dBm	8.7 × 10 ⁻⁴ 9.4 × 10 ⁻⁵ 0.005 8 dB	
		(100 Hz ~ 10 kHz) (1 ~ 10) mV 10 mV ~ 30 V (-20 ~ 10) dBm	8.7 × 10 ⁻⁴ 6.5 × 10 ⁻⁵ 0.005 8 dB	
		(10 ~ 100) kHz (1 ~ 10) mV 10 mV ~ 30 V (-20 ~ 10) dBm	8.7 × 10 ⁻⁴ 9.4 × 10 ⁻⁵ 0.005 8 dB	
AC Output Level Flatness		10 Hz ~ 100 kHz	0.007 1 dB	
Output Attenuation		(0 ~ -60) dB	0.005 8 dB	
Output DC Offset		(±) 0 mV (0 ~ 1) mV (1 ~ 10) mV 10 mV ~ 50 V	0.7 μV 1.0 × 10 ⁻³ 1.0 × 10 ⁻⁴ 7.0 × 10 ⁻⁵	
Output Impedance		5 Ω (10 ~ 600) Ω	1.2 × 10 ⁻³ 6.0 × 10 ⁻⁴	
Input Frequency		1 Hz ~ 200 kHz	6.1 × 10 ⁻⁷	
AC Input Level Flatness		10 Hz ~ 100 kHz	0.008 3 dB	
DC Input Level		(±) 0 mV (0 ~ 1) mV (1 ~ 10) mV 10 mV ~ 300 V	0.27 μV 5.8 × 10 ⁻³ 5.8 × 10 ⁻⁴ 5.8 × 10 ⁻⁴	
Input Distortion		(100 Hz ~ 10 kHz) (-10 ~ -40) dB (-40 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	0.005 8 dB 0.006 0 dB 0.006 8 dB 0.012 dB 0.028 dB	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF/audio signal analyzers	40409			Calibrator, Reference Multimeter/ SICT-CP-40409
Input Distortion		(10 kHz ~ 50 kHz) (-10 ~ -40) dB (-40 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	0.005 9 dB 0.006 3 dB 0.008 2 dB 0.019 dB 0.052 dB	
AC Input Level		(10 ~ 100) Hz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 100 V (100 ~ 300) V (100 Hz ~ 1 kHz) (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 10 V (10 ~ 100) V (100 ~ 300) V (1 ~ 10) kHz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 10 V (10 ~ 100) V (10 ~ 100) kHz (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 1 V (1 ~ 10) V (10 ~ 100) V (10 ~ 100) kHz (10 Hz ~ 10 kHz) (30 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	9.0 × 10 ⁻⁴ 4.0 × 10 ⁻⁴ 4.2 × 10 ⁻⁴ 5.3 × 10 ⁻⁴ 8.4 × 10 ⁻⁴ 1.8 × 10 ⁻⁴ 1.1 × 10 ⁻⁴ 1.0 × 10 ⁻⁴ 2.3 × 10 ⁻⁴ 8.4 × 10 ⁻⁴ 1.8 × 10 ⁻⁴ 1.5 × 10 ⁻⁴ 2.7 × 10 ⁻⁴ 1.4 × 10 ⁻³ 7.6 × 10 ⁻⁴ 4.1 × 10 ⁻⁴ 3.4 × 10 ⁻⁴ 2.6 × 10 ⁻⁴ 0.006 8 dB 0.016 dB 0.052 dB 0.008 3 dB 0.014 dB 0.042 dB	
Input Attenuation		(10 Hz) (30 ~ -50) dB (-50 ~ -60) dB (-60 ~ -80) dB (10 Hz ~ 10 kHz) (30 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB (10 ~ 100) kHz (30 ~ -50) dB (-50 ~ -70) dB (-70 ~ -80) dB	0.009 1 dB 0.023 dB 0.057 dB	
Input Impedance		300 Ω ~ 200 kΩ	3.1 × 10 ⁻⁴	
Input Filter		(10 Hz ~ 100 kHz) 1 V	8.3 × 10 ⁻⁴	
Line frequency meters	40410	16 Hz ~ 1 kHz	1.3 × 10 ⁻⁴	Calibrator/ SICT-CP-40410
Frequency				

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Function generators	Frequency	40411		
		(0.01 ~ 0.1) Hz	5.8×10^{-6}	Audio Analyzer, Digital Multimeter/ SICT-CP-40411
		(0.1 ~ 1) Hz	5.8×10^{-7}	
		1 Hz ~ 1 GHz	5.8×10^{-9}	
	Output Level	(1 ~ 4) GHz	1.5×10^{-8}	
		(10 ~ 100) Hz		
		1 mV	1.0×10^{-3}	
		(1 ~ 10) mV	1.0×10^{-4}	
		10 mV ~ 100 V	7.0×10^{-5}	
		(100 Hz ~ 10 kHz)		
DC Offset	DC Offset	1 mV	1.0×10^{-3}	
		(1 ~ 10) mV	1.0×10^{-4}	
		10 mV ~ 100 V	3.0×10^{-5}	
		(10 ~ 100) kHz		
		1 mV	1.0×10^{-3}	
	Level Flatness	(1 ~ 10) mV	1.0×10^{-4}	
		10 mV ~ 20 V	6.0×10^{-5}	
		(100 mV)		
		(10 ~ 100) Hz	0.099 dB	
		100 Hz ~ 10 kHz	0.083 dB	
Attenuation	Level Flatness	(10 ~ 100) kHz	0.095 dB	
		(100 mV ~ 1 V)		
		(10 ~ 100) Hz	0.005 4 dB	
		100 Hz ~ 10 kHz	0.001 1 dB	
	Attenuation	(10 ~ 100) kHz	0.007 2 dB	
		(1 ~ 30) V		
		(10 ~ 100) Hz	0.021 dB	
		100 Hz ~ 10 kHz	0.015 dB	
		(10 ~ 100) kHz	0.027 dB	
		(10 Hz ~ 100 kHz)		
Distortion	Attenuation	(0 ~ 80) dB	0.006 1 dB	
		(20 Hz ~ 1 kHz)		
	Distortion	(3.16 ~ 0.010) %	1.5×10^{-1}	
		(1 ~ 100) kHz		
		(3.16 ~ 0.010) %	3.2×10^{-1}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Function generators	40411	100 μ s ~ 100 ns	7.0×10^{-4}	Audio Analyzer, Digital Multimeter/ SICT-CP-40411
		(100 ~ 10) ns	7.8×10^{-4}	
		(10 ~ 1) ns	4.7×10^{-3}	
		1 ns ~ 100 ps	4.6×10^{-2}	
		Duty cycle (1 ~ 99) %	0.006 1 %	
High Frequency Flatness Test		(100 kHz ~ 80 MHz)		
		(0 ~ 20) dBm	0.11 dB	
		(0.1 ~ 400) kHz	1.2×10^{-2}	
AM Modulation		(0.1 ~ 100) %	1.2×10^{-2}	
Genescopes	40412	9 kHz ~ 10 MHz	2.8×10^{-6}	Signal Generator/ SICT-CP-40412
		(10 ~ 200) MHz	6.4×10^{-7}	
		9 kHz ~ 200 MHz		
		(100 ~ 50) dB μ V	0.31 dB	
AC/DC high voltages voltmeters	40413	(\pm)		Calibrator/ SICT-CP-40413
		0 kV	0.58 V	
		(0 ~ 0.5) kV	1.2×10^{-3}	
		(0.5 ~ 1) kV	6.1×10^{-4}	
		(1 ~ 2) kV	4.4×10^{-4}	
		(2 ~ 100) kV	3.4×10^{-4}	
		AC Voltage (50 Hz)		
		0.01 kV	0.58 V	
		(0.01 ~ 0.5) kV	1.2×10^{-3}	
		(0.5 ~ 1) kV	6.2×10^{-4}	
		(1 ~ 2) kV	5.5×10^{-4}	
		(2 ~ 3) kV	5.3×10^{-4}	
		(3 ~ 15 kV	5.0×10^{-4}	
		(15 ~ 100) kV	5.7×10^{-4}	
		AC Voltage (60 Hz)		
		0.01 kV	0.58 V	
		(0.01 ~ 0.5) kV	1.2×10^{-3}	
		(0.5 ~ 1) kV	6.2×10^{-4}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Jitter meters	40415	(1 ~ 20) ns	1.7×10^{-3}	Modulation Domain Analyzer/ SICT-CP-40415
		(20 ~ 60) ns	1.6×10^{-3}	
		0.05 μs	0.66 ns	
		(0.05 ~ 0.1) μs	0.77 ns	
		(0.1 ~ 0.2) μs	1.2 ns	
		(0.2 ~ 0.5) μs	2.8 ns	
		(0.5 ~ 0.7) μs	4.3 ns	
		1 %	0.05 %	
		2 %	0.09 %	
		4 %	0.19 %	
Leakage current testers	40416	8 %	0.36 %	Calibrator/ SICT-CP-40416
		10 %	0.44 %	
		15 %	0.67 %	
		0 μA	7.0 nA	
		(0 ~ 1) μA	2.4×10^{-3}	
		(1 ~ 2) μA	3.6×10^{-3}	
		(2 ~ 5) μA	1.4×10^{-3}	
		(5 ~ 10) μA	7.4×10^{-4}	
		(10 ~ 20) μA	4.0×10^{-4}	
		(20 ~ 50) μA	1.8×10^{-4}	
AC Current	40416	(50 ~ 100) μA	1.3×10^{-4}	
		(100 ~ 200) μA	8.5×10^{-5}	
		(0.2 ~ 100) mA	6.1×10^{-4}	
		(20 μA)		
		10 Hz	26 nA	
		(10 ~ 20) Hz	8.5×10^{-4}	
		(0.02 ~ 1) kHz	7.0×10^{-4}	
		(1 ~ 5) kHz	1.3×10^{-3}	
		(5 ~ 10) kHz	5.5×10^{-3}	
		(20 ~ 50) μA		
		10 Hz	6.8×10^{-4}	
		(10 ~ 20) Hz	4.4×10^{-4}	
		(0.02 ~ 1) kHz	3.4×10^{-4}	
		(1 ~ 5) kHz	6.8×10^{-4}	
		(5 ~ 10) kHz	2.8×10^{-3}	
		(50 ~ 100) μA		
		10 Hz	4.9×10^{-4}	
		(10 ~ 20) Hz	3.2×10^{-4}	
		(0.02 ~ 1) kHz	2.3×10^{-4}	
		(1 ~ 5) kHz	4.9×10^{-4}	
		(5 ~ 10) kHz	4.0×10^{-4}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers				
AC Current	40416	(100 ~ 200) μ A 10 Hz (10 ~ 20) Hz (0.02 ~ 1) kHz (1 ~ 5) kHz (5 ~ 10) kHz (200 ~ 500) μ A 10 Hz (10 ~ 20) Hz (0.02 ~ 1) kHz (1 ~ 5) kHz (5 ~ 10) kHz (0.5 ~ 1) mA 10 Hz (10 ~ 20) Hz (0.02 ~ 1) kHz (1 ~ 5) kHz (5 ~ 10) kHz (1 ~ 100) mA 10 Hz (10 ~ 20) Hz (0.02 ~ 1) kHz (1 ~ 5) kHz (5 ~ 10) kHz	3.9×10^{-4} 2.5×10^{-4} 1.7×10^{-4} 4.0×10^{-4} 1.7×10^{-3} 4.4×10^{-4} 3.2×10^{-4} 2.4×10^{-4} 5.4×10^{-4} 2.8×10^{-3} 7.0×10^{-4} 6.6×10^{-4} 6.3×10^{-4} 7.2×10^{-4} 2.1×10^{-3} 7.0×10^{-4} 6.6×10^{-4} 6.3×10^{-4} 7.0×10^{-4} 2.7×10^{-3}	Calibrator/ SICT-CP-40416
DC Voltage		0 V (0 ~ 0.1) V (0.1 ~ 0.2) V (0.2 ~ 0.5) V (0.5 ~ 1) V (1 ~ 2) V (2 ~ 5) V (5 ~ 10) V (10 ~ 20) V (20 ~ 50) V (50 ~ 100) V (100 ~ 200) V (200 ~ 300) V (300 ~ 500) V (500 ~ 1 000) V	0.06 mV 6.0×10^{-4} 3.0×10^{-4} 1.2×10^{-4} 6.0×10^{-5} 3.1×10^{-4} 1.2×10^{-4} 6.1×10^{-5} 3.1×10^{-5} 1.4×10^{-5} 8.8×10^{-6} 3.1×10^{-5} 2.3×10^{-5} 1.2×10^{-4} 5.8×10^{-5}	
AC Voltage		0.1 V 10 Hz (0.01 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz	0.074 mV 6.5×10^{-4} 8.3×10^{-4} 1.2×10^{-3} 2.0×10^{-3} 3.6×10^{-3}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers	40416	(0.1 ~ 0.2) V		Calibrator/ SICT-CP-40416
		10 Hz	4.5×10^{-4}	
		(0.01 ~ 50) kHz	3.5×10^{-4}	
		(50 ~ 100) kHz	5.5×10^{-4}	
		(100 ~ 300) kHz	9.5×10^{-4}	
		(300 ~ 500) kHz	1.8×10^{-3}	
		(0.5 ~ 1) MHz	3.4×10^{-3}	
		(0.2 ~ 0.5) V		
		10 Hz	4.0×10^{-4}	
		(10 ~ 20) Hz	2.0×10^{-4}	
		(0.02 ~ 50) kHz	1.7×10^{-4}	
		(50 ~ 100) kHz	2.2×10^{-4}	
		(100 ~ 300) kHz	6.0×10^{-4}	
		(300 ~ 500) kHz	1.7×10^{-3}	
		(0.5 ~ 1) MHz	2.8×10^{-3}	
		(0.5 ~ 1) V		
		10 Hz	3.3×10^{-4}	
		(10 ~ 20) Hz	1.4×10^{-4}	
		(0.02 ~ 20) kHz	9.1×10^{-5}	
		(20 ~ 100) kHz	1.5×10^{-4}	
		(100 ~ 300) kHz	4.9×10^{-4}	
		(300 ~ 500) kHz	1.4×10^{-3}	
		(0.5 ~ 1) MHz	2.4×10^{-3}	
		(1 ~ 2) V		
		10 Hz	4.3×10^{-4}	
		(0.01 ~ 100) kHz	3.3×10^{-4}	
		(100 ~ 300) kHz	5.5×10^{-4}	
		(300 ~ 500) kHz	1.3×10^{-3}	
		(0.5 ~ 1) MHz	2.1×10^{-3}	
		(2 ~ 5) V		
		10 Hz	4.0×10^{-4}	
		(0.01 ~ 100) kHz	2.0×10^{-4}	
		(100 ~ 300) kHz	5.0×10^{-4}	
		(300 ~ 500) kHz	1.7×10^{-3}	
		(0.5 ~ 1) MHz	2.8×10^{-3}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers				
	40416	(5 ~ 10) V		Calibrator/ SICT-CP-40416
		10 Hz	3.3×10^{-4}	
		(10 ~ 20) Hz	1.4×10^{-4}	
		(0.02 ~ 20) kHz	8.9×10^{-5}	
		(20 ~ 100) kHz	1.4×10^{-4}	
		(100 ~ 300) kHz	3.9×10^{-4}	
		(300 ~ 500) kHz	1.4×10^{-3}	
		(0.5 ~ 1) MHz	2.2×10^{-3}	
		(10 ~ 20) V		
		10 Hz	3.1×10^{-4}	
		(10 ~ 20) Hz	1.2×10^{-4}	
		(0.02 ~ 20) kHz	6.0×10^{-5}	
		(20 ~ 50) kHz	9.0×10^{-5}	
		(50 ~ 100) kHz	1.1×10^{-4}	
		(20 ~ 50) V		
		10 Hz	4.2×10^{-4}	
		(10 ~ 20) Hz	2.2×10^{-4}	
		(0.02 ~ 50) kHz	1.8×10^{-4}	
		(50 ~ 100) kHz	2.8×10^{-4}	
		(50 ~ 100) V		
		10 Hz	3.4×10^{-4}	
		(0.01 ~ 50) kHz	1.4×10^{-4}	
		(50 ~ 100) kHz	2.2×10^{-4}	
		(100 ~ 1 000) V		
		(0.05 ~ 1) kHz	1.1×10^{-4}	
Resistance		100 mΩ	7.7 μΩ	
		1 Ω ~ 10 kΩ	6.2×10^{-5}	
Input Voltage to Output Current Display(U1)		20 Hz		
		(4.75 ~ 5.25) mA	0.006 3 mA	
		50 Hz		
		(4.77 ~ 5.28) mA	0.006 1 mA	
		60 Hz		
		(4.77 ~ 5.28) mA	0.006 1 mA	
		100 Hz		
		(4.85 ~ 5.36) mA	0.006 1 mA	
		200 Hz		
		(5.11 ~ 5.65) mA	0.006 1 mA	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers Input Voltage to Output Current Display(U1)	40416	500 Hz (6.64 ~ 7.34) mA	0.006 1 mA	Calibrator/ SICT-CP-40416
		1 kHz (9.70 ~ 10.73) mA	0.006 1 mA	
		2 kHz (14.07 ~ 15.56) mA	0.006 2 mA	
		5 kHz (17.82 ~ 19.70) mA	0.006 2 mA	
		10 kHz (18.66 ~ 20.63) mA	0.006 2 mA	
		20 kHz (18.92 ~ 20.92) mA	0.006 2 mA	
		50 kHz (19.00 ~ 21.00) mA	0.006 4 mA	
		100 kHz (19.00 ~ 21.00) mA	0.006 7 mA	
		200 kHz (19.00 ~ 21.00) mA	0.010 mA	
		500 kHz (19.00 ~ 21.00) mA	0.030 mA	
		1 MHz (19.00 ~ 21.00) mA	0.046 mA	
Input Voltage to Output Current Display(U2)		20 Hz (4.75 ~ 5.25) mA	0.006 3 mA	
		50 Hz (4.77 ~ 5.28) mA	0.006 1 mA	
		60 Hz (4.77 ~ 5.28) mA	0.006 1 mA	
		100 Hz (4.80 ~ 5.30) mA	0.006 1 mA	
		200 Hz (4.92 ~ 5.44) mA	0.006 1 mA	
		500 Hz (5.37 ~ 5.93) mA	0.006 1 mA	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers Input Voltage to Output Current Display(U2)	40416	1 kHz (5.56 ~ 6.14) mA	0.006 1 mA	Calibrator/ SICT-CP-40416
		2 kHz (4.68 ~ 5.17) mA	0.006 1 mA	
		5 kHz (2.53 ~ 2.80) mA	0.000 63 mA	
		10 kHz (1.35 ~ 1.49) mA	0.000 62 mA	
		20 kHz (0.683 ~ 0.755) mA	0.000 61 mA	
		50 kHz (274.57 ~ 303.47) mA	0.029 μ A	
		100 kHz (137.48 ~ 151.95) μ A	0.020 μ A	
		200 kHz (68.82 ~ 76.06) μ A	0.030 μ A	
		500 kHz (27.43 ~ 30.32) μ A	0.042 μ A	
		1 MHz (13.71 ~ 15.16) μ A	0.033 μ A	
Input Voltage to Output Current Display(U3)		20 Hz (4.75 ~ 5.25) mA	0.006 3 mA	
		50 Hz (4.77 ~ 5.28) mA	0.006 1 mA	
		60 Hz (4.77 ~ 5.28) mA	0.006 1 mA	
		100 Hz (4.80 ~ 5.30) mA	0.006 1 mA	
		200 Hz (4.95 ~ 5.47) mA	0.006 1 mA	
		500 Hz (5.65 ~ 6.25) mA	0.006 1 mA	
		1 kHz (6.60 ~ 7.29) mA	0.006 1 mA	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers Input Voltage to Output Current Display(U3)	40416	2 kHz (7.14 ~ 7.89) mA	0.0061 mA	Calibrator/ SICT-CP-40416
		5 kHz (5.31 ~ 5.87) mA	0.0061 mA	
		10 kHz (3.12 ~ 3.45) mA	0.00064 mA	
		20 kHz (1.63 ~ 1.81) mA	0.00062 mA	
		50 kHz (0.664 ~ 0.734) mA	0.00062 mA	
		100 kHz (322.16 ~ 367.12) µA	0.046 µA	
		200 kHz (166.03 ~ 183.81) µA	0.070 µA	
		500 kHz (66.37 ~ 73.35) µA	0.10 µA	
		1 MHz (33.14 ~ 36.63) µA	0.08 µA	
Input Voltage to Output Voltage Ratio(U1)		4.00 (20 Hz)	1.3×10^{-4}	
		3.98 (50 Hz)	6.5×10^{-5}	
		3.97 (60 Hz)	6.5×10^{-5}	
		3.92 (100 Hz)	6.5×10^{-5}	
		3.72 (200 Hz)	6.5×10^{-5}	
		2.87 (500 Hz)	6.4×10^{-5}	
		1.96 (1 kHz)	6.4×10^{-5}	
		1.96 (2 kHz)	6.4×10^{-5}	
		1.96 (5 kHz)	6.4×10^{-5}	
		1.96 (10 kHz)	6.4×10^{-5}	
		1.00 (20 kHz)	6.7×10^{-5}	
		1.00 (50 kHz)	9.6×10^{-5}	
		1.00 (100 kHz)	1.2×10^{-4}	
		1.00 (200 kHz)	4.2×10^{-4}	
		1.00 (500 kHz)	1.5×10^{-3}	
		1.00 (1 MHz)	2.6×10^{-3}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers				
Input Voltage to Output Voltage Ratio(U2)	40416	4.00 (20 Hz) 3.99 (50 Hz) 3.99 (60 Hz) 3.96 (100 Hz) 3.87 (200 Hz) 3.54 (500 Hz) 3.43 (1 kHz) 4.06 (2 kHz) 7.50 (5 kHz) 14.1 (10 kHz) 27.8 (20 kHz) 69.2 (50 kHz) 138 (100 kHz) 272 (200 kHz) 691 (500 kHz) 1 382 (1 MHz)	1.3×10^{-4} 6.5×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.6×10^{-5} 7.0×10^{-5} 3.6×10^{-5} 4.8×10^{-5} 1.0×10^{-4} 2.7×10^{-4} 1.1×10^{-3} 3.0×10^{-3}	Calibrator/ SICT-CP-40416
Input Voltage to Output Voltage Ratio(U3)		4.00 (20 Hz) 3.99 (50 Hz) 3.98 (60 Hz) 3.95 (100 Hz) 3.83 (200 Hz) 2.36 (500 Hz) 2.87 (1 kHz) 2.65 (2 kHz) 3.57 (5 kHz) 6.09 (10 kHz) 11.6 (20 kHz) 28.7 (50 kHz) 57.2 (100 kHz) 114 (200 kHz) 286 (500 kHz) 572 (1 MHz)	1.3×10^{-4} 6.5×10^{-5} 6.5×10^{-5} 6.6×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.4×10^{-5} 6.4×10^{-5} 6.5×10^{-5} 6.5×10^{-5} 6.9×10^{-5} 3.8×10^{-5} 7.9×10^{-5} 1.9×10^{-4} 6.1×10^{-4} 2.3×10^{-3}	
mAs Meter		1 mAs (1 ~ 2 000) mAs (2 000 ~ 9 999) mAs	1.2×10^{-3} 1.0×10^{-3} 1.1×10^{-3}	
Electronic AC/DC loads	40417			
DC Voltage		0 mV (0 ~ 5) mV (5 ~ 20) mV (20 ~ 100) mV (0.1 ~ 1) V (1 ~ 2) V (2 ~ 4) V (4 ~ 7) V (7 ~ 9) V (9 ~ 10) V (10 ~ 50) V (50 ~ 100) V (100 ~ 200) V (200 ~ 400) V (400 ~ 1 000) V	0.058 mV 5.8×10^{-2} 5.8×10^{-3} 1.2×10^{-3} 6.2×10^{-5} 3.2×10^{-5} 2.1×10^{-5} 1.3×10^{-5} 9.1×10^{-6} 7.9×10^{-6} 3.1×10^{-5} 1.0×10^{-5} 3.4×10^{-5} 2.5×10^{-5} 1.6×10^{-5}	Calibrator/ SICT-CP-40417

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electronic AC/DC loads	40417	1 mA	5.8 μ A	Calibrator/ SICT-CP-40417
		(1 ~ 2) mA	2.9 $\times 10^{-3}$	
		(2 ~ 5) mA	1.2 $\times 10^{-3}$	
		(5 ~ 20) mA	5.8 $\times 10^{-4}$	
		(20 ~ 50) mA	1.2 $\times 10^{-4}$	
		(50 ~ 100) mA	5.9 $\times 10^{-5}$	
		(0.1 ~ 0.2) A	2.9 $\times 10^{-4}$	
		(0.2 ~ 0.4) A	1.9 $\times 10^{-4}$	
		(0.4 ~ 0.6) A	1.2 $\times 10^{-4}$	
		(0.6 ~ 0.8) A	8.4 $\times 10^{-5}$	
		(0.8 ~ 1) A	6.6 $\times 10^{-5}$	
		(1 ~ 3) A	5.1 $\times 10^{-5}$	
		(3 ~ 6) A	2.6 $\times 10^{-5}$	
		(6 ~ 10) A	1.6 $\times 10^{-5}$	
		(10 ~ 40) A	4.0 $\times 10^{-5}$	
		(40 ~ 100) A	2.9 $\times 10^{-5}$	
		(100 ~ 1 000) A	1.4 $\times 10^{-4}$	
		(1 000 ~ 2 000) A	4.9 $\times 10^{-4}$	
		0 mV	0.058 mV	
		(0 ~ 5) mV	5.8 $\times 10^{-2}$	
		(5 ~ 20) mV	5.8 $\times 10^{-3}$	
		(20 ~ 100) mV	1.2 $\times 10^{-3}$	
		(0.1 ~ 1) V	6.2 $\times 10^{-5}$	
		(1 ~ 2) V	3.2 $\times 10^{-5}$	
		(2 ~ 4) V	2.1 $\times 10^{-5}$	
		(4 ~ 7) V	1.3 $\times 10^{-5}$	
		(7 ~ 9) V	9.1 $\times 10^{-6}$	
		(9 ~ 10) V	7.9 $\times 10^{-6}$	
		(10 ~ 50) V	3.1 $\times 10^{-5}$	
		(50 ~ 100) V	1.0 $\times 10^{-5}$	
		(100 ~ 200) V	3.4 $\times 10^{-5}$	
		(200 ~ 400) V	2.5 $\times 10^{-5}$	
		(400 ~ 1 000) V	1.6 $\times 10^{-5}$	
		(1 000 ~ 1 200) V	1.0 $\times 10^{-3}$	
		(1 200 ~ 1 400) V	9.2 $\times 10^{-4}$	
		(1 400 ~ 1 500) V	8.7 $\times 10^{-4}$	
Charge and Discharge Current		(±)		
		1 mA	5.8 μ A	
		(1 ~ 2) mA	2.9 $\times 10^{-3}$	
		(2 ~ 5) mA	1.2 $\times 10^{-3}$	
		(5 ~ 20) mA	5.8 $\times 10^{-4}$	
		(20 ~ 50) mA	1.2 $\times 10^{-4}$	
		(50 ~ 100) mA	5.9 $\times 10^{-5}$	
		(0.1 ~ 0.2) A	2.9 $\times 10^{-4}$	
		(0.2 ~ 0.4) A	1.9 $\times 10^{-4}$	
		(0.4 ~ 0.6) A	1.2 $\times 10^{-4}$	
		(0.6 ~ 0.8) A	8.4 $\times 10^{-5}$	
		(0.8 ~ 1) A	6.6 $\times 10^{-5}$	
		(1 ~ 3) A	5.1 $\times 10^{-5}$	
		(3 ~ 6) A	2.6 $\times 10^{-5}$	
		(6 ~ 10) A	1.6 $\times 10^{-5}$	
		(10 ~ 40) A	4.0 $\times 10^{-5}$	
		(40 ~ 100) A	2.9 $\times 10^{-5}$	
		(100 ~ 1 000) A	1.4 $\times 10^{-4}$	
		(1 000 ~ 3 000) A	4.9 $\times 10^{-4}$	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electronic AC/DC loads	40417			Calibrator/ SICT-CP-40417
Resistance		0.1 Ω (0.1 ~ 1) Ω (1 ~ 2) Ω (2 ~ 4) Ω (4 ~ 500) Ω (0.5 ~ 2) kΩ (2 ~ 10) kΩ	0.58 mΩ 2.1×10^{-3} 4.0×10^{-3} 2.3×10^{-3} 1.6×10^{-3} 9.0×10^{-4} 2.0×10^{-4}	
AC Voltage		(0.001 V) (40 ~ 400) Hz (0.001 ~ 0.1) V (40 ~ 400) Hz (0.1 ~ 0.2) V (40 ~ 400) Hz (0.2 ~ 0.5) V (40 ~ 400) Hz (0.5 ~ 2) V (40 ~ 400) Hz (2 ~ 3) V (40 ~ 400) Hz (3 ~ 7) V (40 ~ 400) Hz (7 ~ 20) V (40 ~ 50) Hz (50 ~ 400) Hz (20 ~ 80) V (40 ~ 50) Hz (50 ~ 400) Hz (80 ~ 200) V (40 ~ 400) Hz (200 ~ 500) V (50 ~ 400) Hz	0.61 mV 6.1×10^{-3} 3.1×10^{-3} 1.2×10^{-3} 6.2×10^{-4} 2.8×10^{-4} 2.3×10^{-4} 1.5×10^{-4} 9.8×10^{-5} 2.1×10^{-4} 1.2×10^{-4} 1.3×10^{-4} 1.8×10^{-4}	
AC Current		(1 mA) (40 ~ 400) Hz (1 ~ 100) mA (40 ~ 400) Hz (100 mA ~ 0.2 A) (40 ~ 400) Hz	0.58 mA 5.8×10^{-2} 5.8×10^{-3}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electronic AC/DC loads	AC Current	(0.2 ~ 0.6) A (40 ~ 400) Hz	2.0×10^{-3}	Calibrator/ SICT-CP-40417
		(0.6 ~ 2) A (40 ~ 400) Hz	9.4×10^{-4}	
		(2 ~ 5) A (40 ~ 400) Hz	1.2×10^{-3}	
		(5 ~ 20) A (40 ~ 400) Hz	1.0×10^{-3}	
	AC Resistance	(1 Ω) (40 ~ 400) Hz	1.0 mΩ	
		(1 ~ 50) Ω (40 ~ 400) Hz	1.5×10^{-3}	
		(50 ~ 100) Ω (40 ~ 400) Hz	1.1×10^{-3}	
		(100 Ω ~ 10 kΩ) (40 ~ 400) Hz	1.7×10^{-3}	
Modulation meters	Frequency Modulation	0 kHz (0 ~ 400) kHz	1 Hz 1.2×10^{-2}	Measuring Receiver/ SICT-CP-40418
		0 % (0 ~ 100) %	0.01 % 1.2×10^{-2}	
		0 rad (0 ~ 400) rad	1.2 mrad 1.2×10^{-2}	
	Amplitude Modulation	(±)		
		0 mV (0 ~ 1) mV	0.43 μV 5.0×10^{-4}	
		(1 ~ 2) mV (2 ~ 5) mV (5 ~ 10) mV (10 ~ 15) mV (15 ~ 20) mV (20 ~ 50) mV (0.05 ~ 0.2) V (0.2 ~ 0.5) V (0.5 ~ 1) V (1 ~ 2) V (2 ~ 5) V (5 ~ 10) V (10 ~ 20) V (20 ~ 50) V (50 ~ 100) V (100 ~ 200) V (200 ~ 500) V (500 ~ 1 000) V	2.5 × 10 ⁻⁴ 1.0 × 10 ⁻⁴ 5.0 × 10 ⁻⁵ 3.3 × 10 ⁻⁵ 2.5 × 10 ⁻⁶ 1.2 × 10 ⁻⁵ 8.0 × 10 ⁻⁶ 4.8 × 10 ⁻⁶ 3.8 × 10 ⁻⁶ 4.0 × 10 ⁻⁶ 2.6 × 10 ⁻⁶ 2.3 × 10 ⁻⁶ 6.0 × 10 ⁻⁶ 4.0 × 10 ⁻⁶ 3.5 × 10 ⁻⁶ 8.0 × 10 ⁻⁶ 5.2 × 10 ⁻⁶ 4.5 × 10 ⁻⁶	
Analogue/digital multimeters	DC Voltage	(±)		Calibrator/ SICT-CP-40419
		0 mV (0 ~ 1) mV	0.43 μV 5.0×10^{-4}	
		(1 ~ 2) mV (2 ~ 5) mV (5 ~ 10) mV (10 ~ 15) mV (15 ~ 20) mV (20 ~ 50) mV (0.05 ~ 0.2) V (0.2 ~ 0.5) V (0.5 ~ 1) V (1 ~ 2) V (2 ~ 5) V (5 ~ 10) V (10 ~ 20) V (20 ~ 50) V (50 ~ 100) V (100 ~ 200) V (200 ~ 500) V (500 ~ 1 000) V	2.5 × 10 ⁻⁴ 1.0 × 10 ⁻⁴ 5.0 × 10 ⁻⁵ 3.3 × 10 ⁻⁵ 2.5 × 10 ⁻⁶ 1.2 × 10 ⁻⁵ 8.0 × 10 ⁻⁶ 4.8 × 10 ⁻⁶ 3.8 × 10 ⁻⁶ 4.0 × 10 ⁻⁶ 2.6 × 10 ⁻⁶ 2.3 × 10 ⁻⁶ 6.0 × 10 ⁻⁶ 4.0 × 10 ⁻⁶ 3.5 × 10 ⁻⁶ 8.0 × 10 ⁻⁶ 5.2 × 10 ⁻⁶ 4.5 × 10 ⁻⁶	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/digital multimeters	40419	(0.6 mV)		Calibrator/ SICT-CP-40419
AC Voltage		1 kHz	4.1 μ V	
		(1 mV)		
		10 Hz	4.2 μ V	
		(10 ~ 40) Hz	4.2 μ V	
		(0.04 ~ 20) kHz	4.1 μ V	
		(20 ~ 50) kHz	4.2 μ V	
		(50 ~ 100) kHz	5.5 μ V	
		(100 ~ 300) kHz	11 μ V	
		(300 ~ 500) kHz	21 μ V	
		(0.5 ~ 1) MHz	23 μ V	
		(1 ~ 2) mV		
		10 Hz	2.2 $\times 10^{-3}$	
		(10 ~ 40) Hz	2.2 $\times 10^{-3}$	
		(0.04 ~ 20) kHz	2.1 $\times 10^{-3}$	
		(20 ~ 50) kHz	2.2 $\times 10^{-3}$	
		(50 ~ 100) kHz	3.0 $\times 10^{-3}$	
		(100 ~ 300) kHz	6.0 $\times 10^{-3}$	
		(300 ~ 500) kHz	1.1 $\times 10^{-2}$	
		(0.5 ~ 1) MHz	1.3 $\times 10^{-2}$	
		(2 ~ 5) mV		
		10 Hz	1.1 $\times 10^{-3}$	
		(10 ~ 40) Hz	9.2 $\times 10^{-4}$	
		(0.04 ~ 20) kHz	9.0 $\times 10^{-4}$	
		(20 ~ 50) kHz	1.0 $\times 10^{-3}$	
		(50 ~ 100) kHz	1.5 $\times 10^{-3}$	
		(100 ~ 300) kHz	3.0 $\times 10^{-3}$	
		(300 ~ 500) kHz	5.2 $\times 10^{-3}$	
		(0.5 ~ 1) MHz	6.8 $\times 10^{-3}$	
		(5 ~ 10) mV		
		10 Hz	6.3 $\times 10^{-4}$	
		(10 ~ 40) Hz	5.0 $\times 10^{-4}$	
		(0.04 ~ 20) kHz	4.9 $\times 10^{-4}$	
		(20 ~ 50) kHz	5.9 $\times 10^{-4}$	
		(50 ~ 100) kHz	9.5 $\times 10^{-4}$	
		(100 ~ 300) kHz	1.9 $\times 10^{-3}$	
		(300 ~ 500) kHz	3.2 $\times 10^{-3}$	
		(0.5 ~ 1) MHz	4.3 $\times 10^{-3}$	
		(10 ~ 15) mV		
		10 Hz	4.8 $\times 10^{-4}$	
		(10 ~ 40) Hz	3.6 $\times 10^{-4}$	
		(0.04 ~ 20) kHz	3.5 $\times 10^{-4}$	
		(20 ~ 50) kHz	4.5 $\times 10^{-4}$	
		(50 ~ 100) kHz	7.6 $\times 10^{-4}$	
		(100 ~ 300) kHz	1.5 $\times 10^{-3}$	
		(300 ~ 500) kHz	2.5 $\times 10^{-3}$	
		(0.5 ~ 1) MHz	3.7 $\times 10^{-3}$	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/digital multimeters	40419			Calibrator/ SICT-CP-40419
AC Voltage		(15 ~ 20) mV		
		10 Hz	4.1×10^{-4}	
		(10 ~ 40) Hz	2.9×10^{-4}	
		(0.04 ~ 20) kHz	2.8×10^{-4}	
		(20 ~ 50) kHz	3.8×10^{-4}	
		(50 ~ 100) kHz	6.7×10^{-4}	
		(100 ~ 300) kHz	1.4×10^{-3}	
		(300 ~ 500) kHz	2.2×10^{-3}	
		(0.5 ~ 1) MHz	3.4×10^{-3}	
		(20 ~ 50) mV		
		10 Hz	4.6×10^{-4}	
		(10 ~ 40) Hz	2.6×10^{-4}	
		(0.04 ~ 20) kHz	2.1×10^{-4}	
		(20 ~ 50) kHz	2.7×10^{-4}	
		(50 ~ 100) kHz	6.6×10^{-4}	
		(100 ~ 300) kHz	1.0×10^{-3}	
		(300 ~ 500) kHz	1.6×10^{-3}	
		(0.5 ~ 1) MHz	3.3×10^{-3}	
		(50 ~ 100) mV		
		10 Hz	3.3×10^{-4}	
		(10 ~ 40) Hz	1.6×10^{-4}	
		(0.04 ~ 20) kHz	1.3×10^{-4}	
		(20 ~ 50) kHz	1.9×10^{-4}	
		(50 ~ 100) kHz	4.8×10^{-4}	
		(100 ~ 300) kHz	7.6×10^{-4}	
		(300 ~ 500) kHz	1.3×10^{-3}	
		(0.5 ~ 1) MHz	2.8×10^{-3}	
		(100 ~ 150) mV		
		10 Hz	2.9×10^{-4}	
		(10 ~ 40) Hz	1.3×10^{-4}	
		(0.04 ~ 20) kHz	1.0×10^{-4}	
		(20 ~ 50) kHz	1.6×10^{-4}	
		(50 ~ 100) kHz	4.1×10^{-4}	
		(100 ~ 300) kHz	6.8×10^{-4}	
		(300 ~ 500) kHz	1.3×10^{-3}	
		(0.5 ~ 1) MHz	2.6×10^{-3}	
		(150 ~ 200) mV		
		10 Hz	2.7×10^{-4}	
		(10 ~ 40) Hz	1.2×10^{-4}	
		(0.04 ~ 20) kHz	9.2×10^{-5}	
		(20 ~ 50) kHz	1.4×10^{-4}	
		(50 ~ 100) kHz	3.9×10^{-4}	
		(100 ~ 300) kHz	6.4×10^{-4}	
		(300 ~ 500) kHz	1.2×10^{-3}	
		(0.5 ~ 1) MHz	2.6×10^{-3}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/digital multimeters	40419	(0.2 ~ 0.5) V		Calibrator/ SICT-CP-40419
AC Voltage		10 Hz	3.0×10^{-4}	
		(10 ~ 20) Hz	1.3×10^{-4}	
		(20 ~ 40) Hz	9.0×10^{-5}	
		(0.04 ~ 20) kHz	6.6×10^{-5}	
		(20 ~ 50) kHz	1.0×10^{-4}	
		(50 ~ 100) kHz	1.6×10^{-4}	
		(100 ~ 300) kHz	4.6×10^{-4}	
		(300 ~ 500) kHz	1.2×10^{-3}	
		(0.5 ~ 1) MHz	2.7×10^{-3}	
		(0.5 ~ 1) V		
		10 Hz	2.5×10^{-4}	
		(10 ~ 20) Hz	1.0×10^{-4}	
		(20 ~ 40) Hz	5.8×10^{-5}	
		(0.04 ~ 20) kHz	4.9×10^{-5}	
		(20 ~ 50) kHz	7.7×10^{-5}	
		(50 ~ 100) kHz	1.2×10^{-4}	
		(100 ~ 300) kHz	3.6×10^{-4}	
		(300 ~ 500) kHz	1.0×10^{-3}	
		(0.5 ~ 1) MHz	2.5×10^{-3}	
		(1 ~ 2) V		
		10 Hz	2.2×10^{-4}	
		(10 ~ 20) Hz	8.5×10^{-5}	
		(20 ~ 40) Hz	4.5×10^{-5}	
		(0.04 ~ 20) kHz	4.2×10^{-5}	
		(20 ~ 50) kHz	6.8×10^{-5}	
		(50 ~ 100) kHz	9.6×10^{-5}	
		(100 ~ 300) kHz	3.2×10^{-4}	
		(300 ~ 500) kHz	9.0×10^{-4}	
		(0.5 ~ 1) MHz	2.4×10^{-3}	
		(2 ~ 5) V		
		10 Hz	3.0×10^{-4}	
		(10 ~ 20) Hz	1.3×10^{-4}	
		(20 ~ 40) Hz	8.2×10^{-5}	
		(0.04 ~ 20) kHz	6.2×10^{-5}	
		(20 ~ 50) kHz	1.0×10^{-4}	
		(50 ~ 100) kHz	1.4×10^{-4}	
		(100 ~ 300) kHz	4.4×10^{-4}	
		(300 ~ 500) kHz	1.3×10^{-3}	
		(0.5 ~ 1) MHz	2.2×10^{-3}	
		(5 ~ 10) V		
		10 Hz	2.5×10^{-4}	
		(10 ~ 20) Hz	9.8×10^{-5}	
		(20 ~ 40) Hz	5.4×10^{-5}	
		(0.04 ~ 20) kHz	4.7×10^{-5}	
		(20 ~ 50) kHz	7.7×10^{-5}	
		(50 ~ 100) kHz	1.1×10^{-4}	
		(100 ~ 300) kHz	3.2×10^{-4}	
		(300 ~ 500) kHz	1.0×10^{-3}	
		(0.5 ~ 1) MHz	1.7×10^{-3}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/digital multimeters	40419			Calibrator/ SICT-CP-40419
AC Voltage		(10 ~ 15) V		
		10 Hz	2.3×10^{-4}	
		(10 ~ 20) Hz	8.7×10^{-5}	
		(20 ~ 40) Hz	4.6×10^{-5}	
		(0.04 ~ 20) kHz	4.3×10^{-5}	
		(20 ~ 50) kHz	7.1×10^{-5}	
		(50 ~ 100) kHz	9.5×10^{-5}	
		(100 ~ 300) kHz	2.9×10^{-4}	
		(300 ~ 500) kHz	9.4×10^{-4}	
		(0.5 ~ 1) MHz	1.5×10^{-3}	
		(15 ~ 20) V		
		10 Hz	2.2×10^{-4}	
		(10 ~ 20) Hz	8.5×10^{-5}	
		(20 ~ 40) Hz	4.3×10^{-5}	
		(0.04 ~ 20) kHz	4.1×10^{-5}	
		(20 ~ 50) kHz	6.8×10^{-5}	
		(50 ~ 100) kHz	9.0×10^{-5}	
		(100 ~ 300) kHz	2.8×10^{-4}	
		(300 ~ 500) kHz	9.1×10^{-4}	
		(0.5 ~ 1) MHz	1.4×10^{-3}	
		(20 ~ 50) V		
		10 Hz	3.2×10^{-4}	
		(10 ~ 20) Hz	1.4×10^{-4}	
		(20 ~ 40) Hz	9.8×10^{-5}	
		(0.04 ~ 20) kHz	7.4×10^{-5}	
		(20 ~ 50) kHz	1.1×10^{-5}	
		(50 ~ 100) kHz	2.1×10^{-4}	
		(50 ~ 100) V		
		10 Hz	2.5×10^{-4}	
		(10 ~ 20) Hz	1.0×10^{-4}	
		(20 ~ 40) Hz	6.5×10^{-5}	
		(0.04 ~ 20) kHz	5.6×10^{-5}	
		(20 ~ 50) kHz	8.5×10^{-5}	
		(50 ~ 100) kHz	1.6×10^{-4}	
		(100 ~ 200) V		
		10 Hz	2.3×10^{-4}	
		(10 ~ 20) Hz	9.3×10^{-5}	
		(20 ~ 40) Hz	5.6×10^{-5}	
		(0.04 ~ 20) kHz	5.1×10^{-5}	
		(20 ~ 50) kHz	7.9×10^{-5}	
		(50 ~ 100) kHz	1.4×10^{-4}	
		(200 ~ 500) V		
		50 Hz ~ 1 kHz	6.7×10^{-5}	
		(500 ~ 1 000) V		
		50 Hz ~ 1 kHz	6.3×10^{-5}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/digital multimeters	Resistance	0 Ω	0.001 0 mΩ	Calibrator/ SICT-CP-40419
		0 Ω ~ 10 kΩ	1.2 × 10 ⁻⁶	
		(10 ~ 100) kΩ	1.4 × 10 ⁻⁶	
	DC Current	(0.1 ~ 1) MΩ	7.2 × 10 ⁻⁶	
		(1 ~ 10) MΩ	7.7 × 10 ⁻⁶	
		(10 ~ 100) MΩ	1.2 × 10 ⁻⁵	
		(0.1 ~ 1) GΩ	3.2 × 10 ⁻⁴	
		(±)		
		0 nA	6.0 nA	
		(0 ~ 1) nA	6.9 × 10 ⁻³	
		(1 ~ 100) nA	4.6 × 10 ⁻³	
		(0.1 ~ 1) μA	6.0 × 10 ⁻³	
		(1 ~ 2) μA	3.1 × 10 ⁻³	
		(2 ~ 5) μA	1.2 × 10 ⁻³	
		(5 ~ 10) μA	6.3 × 10 ⁻⁴	
		(10 ~ 20) μA	3.5 × 10 ⁻⁴	
		(20 ~ 50) μA	1.6 × 10 ⁻⁴	
		(50 ~ 100) μA	9.4 × 10 ⁻⁵	
		(100 ~ 200) μA	6.3 × 10 ⁻⁵	
		(0.2 ~ 0.5) mA	4.8 × 10 ⁻⁵	
		(0.5 ~ 1) mA	3.5 × 10 ⁻⁵	
		(1 ~ 1.5) mA	3.1 × 10 ⁻⁵	
		(1.5 ~ 2) mA	3.0 × 10 ⁻⁵	
		(2 ~ 5) mA	4.4 × 10 ⁻⁵	
		(5 ~ 10) mA	3.2 × 10 ⁻⁵	
		(10 ~ 15) mA	2.9 × 10 ⁻⁵	
		(15 ~ 20) mA	2.8 × 10 ⁻⁵	
		(20 ~ 50) mA	5.6 × 10 ⁻⁵	
		(50 ~ 100) mA	4.4 × 10 ⁻⁵	
		(100 ~ 150) mA	4.1 × 10 ⁻⁵	
		(150 ~ 200) mA	3.9 × 10 ⁻⁵	
		(0.2 ~ 0.5) A	9.4 × 10 ⁻⁵	
		(0.5 ~ 1) A	6.9 × 10 ⁻⁵	
		(1 ~ 1.5) A	6.1 × 10 ⁻⁵	
		(1.5 ~ 2) A	5.8 × 10 ⁻⁵	
		(2 ~ 3) A	3.3 × 10 ⁻⁴	
		(3 ~ 5) A	2.4 × 10 ⁻⁴	
		(5 ~ 10) A	1.6 × 10 ⁻⁴	
		(10 ~ 20) A	1.2 × 10 ⁻⁴	
		(20 ~ 30) A	2.4 × 10 ⁻⁴	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/digital multimeters	40419			Calibrator/ SICT-CP-40419
AC Current		(20 μ A) 1 kHz 10 kHz (20 ~ 50) μ A 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (50 ~ 100) μ A 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (0.1 ~ 0.2) mA 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (0.2 ~ 0.5) mA 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (0.5 ~ 1) mA 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (1 ~ 2) mA 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (2 ~ 5) mA 10 Hz (10 ~ 20) Hz 20 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz		

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/digital multimeters	40419			Calibrator/ SICT-CP-40419
AC Current		(5 ~ 10) mA		
		10 Hz	1.4×10^{-4}	
		(10 ~ 20) Hz	1.0×10^{-4}	
		20 Hz ~ 1 kHz	7.3×10^{-5}	
		(1 ~ 5) kHz	1.4×10^{-4}	
		(5 ~ 10) kHz	7.2×10^{-4}	
		(10 ~ 20) mA		
		10 Hz	2.4×10^{-4}	
		(10 ~ 20) Hz	1.6×10^{-4}	
		20 Hz ~ 1 kHz	1.2×10^{-4}	
		(1 ~ 5) kHz	2.2×10^{-4}	
		(5 ~ 10) kHz	1.1×10^{-3}	
		(20 ~ 50) mA		
		10 Hz	1.0×10^{-4}	
		(10 ~ 20) Hz	8.0×10^{-5}	
		20 Hz ~ 1 kHz	4.8×10^{-5}	
		(1 ~ 5) kHz	1.1×10^{-4}	
		(5 ~ 10) kHz	4.0×10^{-4}	
		(50 ~ 100) mA		
		10 Hz	1.4×10^{-4}	
		(10 ~ 20) Hz	1.0×10^{-4}	
		20 Hz ~ 1 kHz	6.8×10^{-5}	
		(1 ~ 5) kHz	1.4×10^{-4}	
		(5 ~ 10) kHz	6.0×10^{-4}	
		(0.1 ~ 0.2) A		
		10 Hz	2.4×10^{-4}	
		(10 ~ 20) Hz	1.6×10^{-4}	
		20 Hz ~ 1 kHz	1.1×10^{-4}	
		(1 ~ 5) kHz	2.1×10^{-4}	
		(5 ~ 10) kHz	1.0×10^{-3}	
		(0.2 ~ 1) A		
		40 Hz	1.4×10^{-4}	
		40 Hz ~ 1 kHz	1.4×10^{-4}	
		(1 ~ 5) kHz	2.6×10^{-4}	
		(5 ~ 10) kHz	2.7×10^{-3}	
		(1 ~ 2) A		
		40 Hz ~ 1 kHz	2.4×10^{-4}	
		(1 ~ 5) kHz	4.2×10^{-4}	
		(5 ~ 10) kHz	5.2×10^{-3}	
		(2 ~ 3) A		
		(40 ~ 100) Hz	1.8×10^{-4}	
		100 Hz ~ 1 kHz	1.9×10^{-4}	
		(1 ~ 10) kHz	9.9×10^{-4}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/digital multimeters	40419			Calibrator/ SICT-CP-40419
AC Current		(3 ~ 5) A (40 ~ 100) Hz 100 Hz ~ 1 kHz (1 ~ 10) kHz (5 ~ 10) A (40 ~ 100) Hz 100 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 20) A (40 ~ 60) Hz (60 ~ 100) Hz 100 Hz ~ 1 kHz (1 ~ 10) kHz (20 ~ 30) A (40 ~ 60) Hz (60 ~ 100) Hz 100 Hz ~ 1 kHz (1 ~ 10) kHz	2.4×10^{-4} 2.5×10^{-4} 1.6×10^{-3} 4.2×10^{-4} 4.2×10^{-4} 3.1×10^{-3} 6.0×10^{-4} 7.0×10^{-4} 1.3×10^{-3} 2.3×10^{-3} 8.0×10^{-4} 1.2×10^{-3} 3.9×10^{-3} 4.5×10^{-3}	
Frequency		10 Hz ~ 10 MHz	6.4×10^{-7}	
(Digital Smapling) AC Voltage		(1 mV) 0.1 Hz ~ 3 kHz (1 mV ~ 2 mV) 0.1 Hz ~ 3 kHz (2 mV ~ 3 mV) 0.1 Hz ~ 3 kHz (3 mV ~ 5 mV) 0.1 Hz ~ 3 kHz (5 mV ~ 10 mV) 0.1 Hz ~ 3 kHz (10 mV ~ 20 mV) 0.1 Hz ~ 3 kHz (20 mV ~ 30 mV) 0.1 Hz ~ 3 kHz (30 mV ~ 50 mV) 0.1 Hz ~ 3 kHz (50 mV ~ 100 mV) 0.1 Hz ~ 3 kHz	8.4×10^{-4} 4.2×10^{-4} 2.8×10^{-4} 1.7×10^{-4} 8.8×10^{-5} 4.8×10^{-5} 3.6×10^{-5} 3.0×10^{-5} 2.6×10^{-5}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/digital multimeters (Digital Smapling)	40419			Calibrator/ SICT-CP-40419
AC Voltage		(100 mV ~ 200 mV) 0.1 Hz ~ 3 kHz	4.8×10^{-5}	
		(200 mV ~ 300 mV) 0.1 Hz ~ 3 kHz	3.6×10^{-5}	
		(300 mV ~ 500 mV) 0.1 Hz ~ 3 kHz	2.8×10^{-5}	
		(500 mV ~ 1 V) 0.1 Hz ~ 3 kHz	2.4×10^{-5}	
		(1 V ~ 2 V) 0.1 Hz ~ 3 kHz	4.8×10^{-5}	
		(2 V ~ 3 V) 0.1 Hz ~ 3 kHz	3.6×10^{-5}	
		(3 V ~ 5 V) 0.1 Hz ~ 3 kHz	3.0×10^{-5}	
		(5 V ~ 10 V) 0.1 Hz ~ 3 kHz	2.6×10^{-5}	
		(10 V ~ 30 V) 10 Hz ~ 3 kHz	3.6×10^{-5}	
		(30 V ~ 50 V) 10 Hz ~ 3 kHz	2.8×10^{-5}	
		(50 V ~ 100 V) 10 Hz ~ 3 kHz	2.4×10^{-5}	
		(100 V ~ 200 V) 10 Hz ~ 3 kHz	4.8×10^{-5}	
		(200 V ~ 1 000 V) 50 Hz ~ 1 kHz	2.4×10^{-5}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Noise meters				
AC Voltage Test	40420	(600 μ V) 1 kHz	7.8×10^{-3}	Calibrator/ SICT-CP-40420
		(600 μ V ~ 20 mV) (10 ~ 40) Hz 40 Hz ~ 20 kHz (20 ~ 100) kHz (100 ~ 300) kHz 300 kHz ~ 1 MHz	2.6×10^{-3} 2.5×10^{-3} 3.5×10^{-3} 7.0×10^{-3} 1.5×10^{-2}	
		(20 ~ 200) mV (10 ~ 40) Hz 40 Hz ~ 20 kHz (20 ~ 100) kHz 100 kHz ~ 1 MHz	5.2×10^{-4} 3.4×10^{-4} 8.8×10^{-4} 4.3×10^{-3}	
		(200 mV ~ 2 V) (10 ~ 40) Hz 40 Hz ~ 20 kHz (20 ~ 100) kHz (100 ~ 300) kHz 300 kHz ~ 1 MHz	3.5×10^{-4} 1.5×10^{-4} 4.6×10^{-4} 8.8×10^{-4} 3.4×10^{-3}	
		(2 ~ 20) V (10 ~ 40) Hz 40 Hz ~ 100 kHz (100 ~ 300) kHz 300 kHz ~ 1 MHz	3.0×10^{-4} 1.2×10^{-4} 4.4×10^{-4} 2.2×10^{-3}	
		(20 ~ 200) V (10 ~ 40) Hz 40 Hz ~ 20 kHz (20 ~ 100) kHz	3.0×10^{-4} 1.2×10^{-4} 1.8×10^{-4}	
		(200 ~ 500) V 50 Hz ~ 1 kHz	3.8×10^{-4}	
		(500 ~ 1 000) V 50 Hz ~ 1 kHz	3.7×10^{-4}	
		(25 ~ 500) mV (1 ~ 30) MHz	2.1×10^{-2}	
		(500 mV ~ 2 V) (0.1 ~ 30) MHz	2.1×10^{-2}	
Weighting Test		(DIN/NOISE) 31.5 Hz ~ 10 kHz (JIS A) 31.5 Hz ~ 16 kHz (CCIR) 31.5 Hz ~ 31.5 kHz (CCIR/ARM) 31.5 Hz ~ 31.5 kHz	0.12 dB 0.12 dB 0.12 dB 0.12 dB	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Noise meters	40420			
AC Voltage Output		(10 mV) 1 kHz (10 mV ~ 1 V) 20 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz	2.8×10^{-3} 8.5×10^{-5} 1.3×10^{-4} 6.9×10^{-4}	Calibrator/ SICT-CP-40420
DC Voltage Output		0 mV 100 mV ~ 1 V	0.99 μ V 1.1×10^{-5}	
Oscilloscopes	40421			Calibration Generator/ SICT-CP-40421
Impedance Measure		50 Ω 75 Ω 1 M Ω	3.5×10^{-5} 2.7×10^{-5} 2.5×10^{-5}	
DC Voltage		(\pm) 0 mV (0 ~ 1) mV (1 ~ 5) mV (5 ~ 10) mV (10 ~ 100) mV (100 ~ 900) mV (0.9 ~ 9) V (9 ~ 200) V	 0.79 μ V 8.0×10^{-4} 4.1×10^{-4} 1.7×10^{-4} 8.5×10^{-5} 1.5×10^{-5} 9.1×10^{-6} 9.5×10^{-6}	
AC Voltage(Square wave)		(1 kHz) 1 mV (1 ~ 25) mV (0.025 ~ 0.5) V (0.5 ~ 2.2) V (2.2 ~ 130) V	 6.5×10^{-3} 8.8×10^{-4} 9.1×10^{-4} 6.8×10^{-4} 8.4×10^{-4}	
Time Marker		100 ps (100 ~ 200) ps 200 ps ~ 20 ms 20 ms ~ 5 s	 6.2×10^{-7} 3.1×10^{-7} 1.7×10^{-6} 1.6×10^{-5}	
CAL Output Amplitude		(40 Hz ~ 20 kHz) 100 mV 100 mV ~ 12 V	 3.2×10^{-5} 1.9×10^{-5}	
CAL Output Frequency		100 Hz ~ 10 MHz	6.2×10^{-7}	
Sinewave Signal Generator Level		50 kHz 50 kHz ~ 1 MHz 1 MHz ~ 1 GHz (1 ~ 4) GHz (4 ~ 18) GHz (18 ~ 25) GHz (25 ~ 33) GHz (33 ~ 40) GHz	 2.3×10^{-2} 4.7×10^{-2} 1.5×10^{-2} 1.8×10^{-2} 3.2×10^{-2} 5.5×10^{-2} 5.8×10^{-2} 6.0×10^{-2}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Oscilloscopes	40421	(10 ~ 40) Hz		Calibration Generator/ SICT-CP-40421
AC Voltage(Sine wave)		2 mV	5.3 μ V	
		(2 ~ 20) mV	1.9×10^{-3}	
		(20 ~ 200) mV	6.8×10^{-4}	
		(0.2 ~ 20) V	4.3×10^{-4}	
		(20 ~ 100) V	4.5×10^{-4}	
		(40 Hz ~ 1 kHz)		
		2 mV	5.0 μ V	
		(2 ~ 20) mV	1.7×10^{-3}	
		(20 ~ 800) mV	3.3×10^{-4}	
		(0.8 ~ 20) V	1.7×10^{-4}	
		(20 ~ 200) V	1.8×10^{-4}	
		(1 ~ 50) kHz		
		2 mV	5.2 μ V	
		(2 ~ 20) mV	1.9×10^{-3}	
		(20 ~ 200) mV	4.8×10^{-4}	
		(0.2 ~ 20) V	1.3×10^{-4}	
		(20 ~ 100) V	1.7×10^{-4}	
		(50 ~ 100) kHz		
		2 mV	7.1 μ V	
		(2 ~ 20) mV	2.7×10^{-3}	
		(20 ~ 200) mV	9.0×10^{-4}	
		(0.2 ~ 20) V	1.9×10^{-4}	
		(20 ~ 100) V	3.0×10^{-4}	
LF phase meters	40422	(1 Hz ~ 200 kHz)		Multi Function Generator/ SICT-CP-40422
Phase Test		(-180 ~ 180) °	0.074 °	
Volt/Current recorders	40424	(±)		Calibrator/ SICT-CP-40424
DC Voltage		(0 ~ 100) μ V	0.51 μ V	
		(0.1 ~ 1) mV	5.2×10^{-4}	
		(1 ~ 10) mV	5.9×10^{-5}	
		(0.01 ~ 1) V	6.7×10^{-6}	
		(1 ~ 10) V	4.3×10^{-6}	
		(10 ~ 100) V	6.3×10^{-6}	
		(100 ~ 1 000) V	8.7×10^{-6}	
DC Current		(±)		
		(0 ~ 1) nA	6.9 pA	
		(1 ~ 100) nA	4.6×10^{-3}	
		(0.1 ~ 1) μ A	2.3×10^{-3}	
		(1 ~ 10) μ A	7.2×10^{-4}	
		(10 ~ 100) μ A	1.4×10^{-4}	
		(0.1 ~ 10) mA	7.6×10^{-5}	
		(10 ~ 100) mA	8.4×10^{-5}	
		(0.1 ~ 1) A	1.2×10^{-4}	
		(1 ~ 100) A	2.1×10^{-4}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Relay test sets	40425			Digital Multimeter/ SICT-CP-40425
DC Voltage		1 mV 1 mV ~ 1 V (1 ~ 100) V (100 ~ 1 000) V	6 µV 7.0×10^{-4} 7.0×10^{-5} 2.2×10^{-5}	
DC Current		1 mA 1 mA ~ 1 A (1 ~ 20) A (20 ~ 100) A	58 µA 6.0×10^{-4} 2.5×10^{-4} 4.0×10^{-4}	
AC Voltage		(1 mV) 20 Hz ~ 100 kHz (1 ~ 100) mV 20 Hz ~ 10 kHz (10 ~ 100) kHz (100 mV ~ 1 V) 20 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (1 ~ 10) V 20 Hz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1 MHz (10 ~ 100) V 20 Hz ~ 10 kHz (10 ~ 100) kHz (100 ~ 1 000) V 50 Hz ~ 10 kHz (10 ~ 30) kHz	58 µV 6.1×10^{-4} 1.2×10^{-3} 7.6×10^{-4} 1.3×10^{-3} 2.0×10^{-2} 2.2×10^{-4} 1.1×10^{-3} 2.1×10^{-2} 1.8×10^{-4} 1.1×10^{-3} 1.9×10^{-4} 4.5×10^{-4}	
AC Current		(1 mA) 40 Hz ~ 10 kHz (1 ~ 100) mA 40 Hz ~ 10 kHz (100 mA ~ 1 A) 40 Hz ~ 10 kHz (1 ~ 10) A 40 Hz ~ 10 kHz (10 ~ 100) A 40 Hz ~ 10 kHz	58 µA 8.6×10^{-4} 9.1×10^{-4} 9.9×10^{-4} 2.3×10^{-4}	
Timer		(1 ~ 100) s	5.8×10^{-6}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF signal generators	40426			Audio Analyzer, Digital Multimeter/ SICT-CP-40426
Frequency Test		(0.1 ~ 1) Hz 1 Hz ~ 100 MHz	5.8 × 10 ⁻⁷ 5.8 × 10 ⁻⁹	
Output Level Test		(10 ~ 100) Hz 1 mV (1 ~ 10) mV 10 mV ~ 100 V (100 Hz ~ 10 kHz) 1 mV (1 ~ 10) mV 10 mV ~ 100 V (10 ~ 100) kHz 1 mV (1 ~ 10) mV 10 mV ~ 100 V	1.0 × 10 ⁻³ 1.0 × 10 ⁻⁴ 7.0 × 10 ⁻⁵ 1.0 × 10 ⁻³ 1.0 × 10 ⁻⁴ 3.0 × 10 ⁻⁵ 1.0 × 10 ⁻³ 1.0 × 10 ⁻⁴ 8.0 × 10 ⁻⁵	
DC Offset		(±) 0 mV (0 ~ 1) mV (1 ~ 10) mV 10 mV ~ 20 V	0.7 μV 0.7 μV 1.0 × 10 ⁻⁴ 6.0 × 10 ⁻⁵	
Output Level Flatness Test		(100 mV) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz (100 mV ~ 1 V) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz (1 ~ 30) V (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz	0.099 dB 0.083 dB 0.095 dB 0.005 4 dB 0.001 1 dB 0.007 2 dB 0.021 dB 0.015 dB 0.027 dB	
Attenuator Test		(10 Hz ~ 100 kHz) (0 ~ 80) dB	0.006 1 dB	
Distortion		(20 Hz ~ 1 kHz) (3.16 ~ 0.010) %	1.5 × 10 ⁻¹	
		(1 ~ 100) kHz (3.16 ~ 0.010) %	3.2 × 10 ⁻¹	
Rise/Fall Time		100 μs ~ 100 ns (100 ~ 10) ns (10 ~ 1) ns 1 ns ~ 100 ps	7.0 × 10 ⁻⁴ 7.8 × 10 ⁻⁴ 4.7 × 10 ⁻³ 4.6 × 10 ⁻²	
Duty cycle		(1 ~ 99) %	0.006 1 %	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF spectrum analyzers	40427			Synthesizer Function
Reference Frequency		10 MHz	7.7×10^{-12}	Generator/ SICT-CP-40427
Center Frequency		10 Hz (10 ~ 100) Hz 100 Hz ~ 1 kHz 1 kHz ~ 1 MHz (1 ~ 100) MHz 100 MHz ~ 1 GHz	6.1×10^{-5} 6.1×10^{-6} 6.1×10^{-7} 6.1×10^{-8} 6.1×10^{-9} 6.1×10^{-8}	
Frequency Range		10 Hz (10 ~ 100) Hz 100 Hz ~ 1 kHz (1 ~ 100) kHz 100 kHz ~ 1 MHz (1 ~ 100) MHz 100 MHz ~ 1 GHz	1.1×10^{-3} 1.1×10^{-4} 1.1×10^{-5} 1.1×10^{-4} 1.1×10^{-6} 1.1×10^{-7} 1.1×10^{-9}	
Resolution Bandwidth		100 Hz 100 Hz ~ 3 kHz (3 ~ 300) kHz 300 kHz ~ 1 MHz	3.3×10^{-2} 3.1×10^{-2} 3.3×10^{-2} 3.4×10^{-2}	
Absolute Amplitude		(-60 dBV) 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz (-60 ~ -50) dBV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz (-50 ~ -40) dBV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz (-40 ~ -30) dBV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz (-30 ~ 30) dBV 10 Hz 10 Hz ~ 10 kHz (10 ~ 100) kHz	0.043 dB 0.042 dB 0.056 dB 0.017 dB 0.015 dB 0.022 dB 0.009 dB 0.009 8 dB 0.012 dB 0.009 dB 0.006 8 dB 0.011 dB 0.016 dB 0.006 3 dB 0.007 3 dB	
Referency Level		(-60 dBV) 10 Hz ~ 100 kHz (-60 ~ 30) dBV 10 Hz ~ 100 kHz	0.17 dB 0.16 dB	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF spectrum analyzers	40427			Synthesizer Function
Frequency Response		10 Hz	0.009 1 dB	Generator/ SICT-CP-40427
		10 Hz ~ 100 kHz	0.008 7 dB	
Logscale Fidelity		(0 ~ -60) dB	0.009 2 dB	
		(-60 ~ -70) dB	0.012 dB	
		(-70 ~ -80) dB	0.016 dB	
		(-80 ~ -90) dB	0.042 dB	
Output frequency		10 Hz ~ 300 MHz	6.1×10^{-11}	
Input Impedance		1 MΩ	0.000 12 MΩ	
		(50 ~ 75) Ω	0.000 7 Ω	
Output Voltage		10 mV	0.000 38 mV	
		10 mV ~ 5 V	8.8×10^{-5}	
Output Offset Voltage		(-30 ~ 30) V	6.7×10^{-6}	
Output Voltage Flatness		10 Hz ~ 100 kHz	0.000 67 dB	
Spot generators	40428			Audio Analyzer, Digital Multimeter/ SICT-CP-40428
Frequency		(0.1 ~ 1) Hz	5.8×10^{-7}	
		1 Hz ~ 100 kHz	5.8×10^{-9}	
Output Level		(10 ~ 100) Hz		
		1 mV	1.0×10^{-3}	
		(1 ~ 10) mV	1.0×10^{-4}	
		10 mV ~ 10 V	7.0×10^{-5}	
		(100 Hz ~ 10 kHz)		
		1 mV	1.0×10^{-3}	
		(1 ~ 10) mV	1.0×10^{-4}	
		10 mV ~ 10 V	3.0×10^{-5}	
		(10 ~ 100) kHz		
		1 mV	1.0×10^{-3}	
		(1 ~ 10) mV	1.0×10^{-4}	
		10 mV ~ 10 V	7.0×10^{-5}	
Output Level Flatness		(100 mV)		
		(10 ~ 100) Hz	0.099 dB	
		100 Hz ~ 10 kHz	0.083 dB	
		(10 ~ 100) kHz	0.095 dB	
		(100 mV ~ 1 V)		
		(10 ~ 100) Hz	0.005 4 dB	
		100 Hz ~ 10 kHz	0.001 1 dB	
		(10 ~ 100) kHz	0.007 2 dB	
		(1 ~ 10) V		
		(10 ~ 100) Hz	0.010 dB	
		100 Hz ~ 100 kHz	0.011 dB	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Spot generators	40428			Audio Analyzer, Digital Multimeter/ SICT-CP-40428
Attenuation		(10 Hz ~ 100 kHz) (0 ~ 80) dB	0.006 1 dB	
Distortion		(20 Hz ~ 1 kHz) (3.16 ~ 0.010) % (1 ~ 100) kHz (3.16 ~ 0.010) %	1.5×10^{-1} 3.2×10^{-1}	
Sweep generators	40429			Audio Analyzer, Digital Multimeter/ SICT-CP-40429
Frequency		(0.1 ~ 1) Hz 1 Hz ~ 100 kHz	5.8×10^{-7} 5.8×10^{-9}	
Output Level Test		(10 ~ 100) Hz 1 mV (1 ~ 10) mV 100 mV ~ 10 V (100 Hz ~ 10 kHz) 1 mV (1 ~ 100) mV 100 mV ~ 10 V (10 ~ 100) kHz 1 mV (1 ~ 100) mV 100 mV ~ 10 V	1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5} 1.0×10^{-3} 1.0×10^{-4} 3.0×10^{-5} 1.0×10^{-3} 1.0×10^{-4} 7.0×10^{-5}	
Output Level Flatness		(100 mV) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz (100 mV ~ 1 V) (10 ~ 100) Hz 100 Hz ~ 10 kHz (10 ~ 100) kHz (1 ~ 10) V (10 ~ 100) Hz 100 Hz ~ 100 kHz	0.099 dB 0.083 dB 0.095 dB 0.005 4 dB 0.001 1 dB 0.007 2 dB 0.010 dB 0.011 dB	
Attenuation		(10 Hz ~ 10 kHz) (0 ~ 80) dB	0.006 1 dB	
Distortion		(20 Hz ~ 1 kHz) (3.16 ~ 0.010) % (1 ~ 100) kHz (3.16 ~ 0.010) %	1.5×10^{-1} 3.2×10^{-1}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Signal transducers	40430	1 mV	5.2×10^{-4}	Digital Multimeter/ SICT-CP-40430
		(1 ~ 10) mV	8.8×10^{-5}	
		10 mV ~ 100 V	3.1×10^{-5}	
		(100 ~ 300) V	4.3×10^{-3}	
		10 µA	9.5×10^{-4}	
		(10 ~ 100) µA	9.7×10^{-5}	
		100 µA ~ 100 mA	7.0×10^{-5}	
		100 mA ~ 20 A	2.0×10^{-4}	
		(1 ~ 10) Hz	3.1×10^{-4}	
		10 Hz ~ 100 kHz	7.2×10^{-5}	
Transistor curve tracers	40432	0 mV	$1.0 \mu\text{V}$	Digital Multimeter/ SICT-CP-40432
		(0 ~ 100) mV	1.1×10^{-5}	
		100 mV ~ 1 V	1.0×10^{-5}	
		(1 ~ 10) V	9.8×10^{-6}	
		(10 ~ 100) V	1.1×10^{-5}	
		(100 ~ 1 000) V	8.0×10^{-6}	
		0 nA	0.12 nA	
		(0 ~ 1) nA	1.0×10^{-2}	
		(1 ~ 100) nA	8.0×10^{-3}	
		(0.1 ~ 1) µA	8.0×10^{-4}	
		(1 ~ 10) µA	9.0×10^{-5}	
		10 µA ~ 10 mA	1.6×10^{-5}	
		(10 ~ 100) mA	5.0×10^{-5}	
		100 mA ~ 10 A	2.5×10^{-4}	
		0 mV	0.78 µV	
		(0 ~ 100) mV	1.5×10^{-5}	
		100 mV ~ 1 V	6.0×10^{-4}	
		(1 ~ 100) V	6.4×10^{-4}	
		(100 ~ 1 000) V	6.2×10^{-5}	
AC/DC high voltage generators	40434	(±)		High Voltage Digital Meter/ SICT-CP-40434
		0 kV	0.58 V	
		(0 ~ 0.5) kV	1.2×10^{-3}	
		(0.5 ~ 1) kV	6.1×10^{-4}	
		(1 ~ 2) kV	3.0×10^{-4}	
		(2 ~ 100) kV	2.3×10^{-4}	
		(100 ~ 200) kV	1.2×10^{-2}	
		(50 Hz)		
		0.01 kV	0.58 V	
		(0.01 ~ 0.5) kV	1.2×10^{-3}	
		(0.5 ~ 1) kV	6.1×10^{-4}	
		(1 ~ 100) kV	5.7×10^{-4}	
		(100 ~ 200) kV	1.2×10^{-2}	
		(60 Hz)		
		0.01 kV	0.58 V	
		(0.01 ~ 0.5) kV	1.2×10^{-3}	
		(0.5 ~ 1) kV	6.2×10^{-4}	
		(1 ~ 100) kV	5.2×10^{-4}	
		(100 ~ 200) kV	1.2×10^{-2}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC/DC high voltage probes	40435			DC Power Supply/ SICT-CP-40435
DC Voltage		(±) 0 kV (0 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 100) kV	0.06 V 4.2×10^{-4} 2.1×10^{-4} 3.5×10^{-4}	
AC Voltage		(50 Hz) 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 3) kV (3 ~ 5) kV (5 ~ 10) kV (10 ~ 20) kV (20 ~ 100) kV	2.2 V 4.4×10^{-3} 2.3×10^{-3} 1.2×10^{-3} 7.5×10^{-4} 6.0×10^{-4} 4.0×10^{-4} 3.7×10^{-4}	
		(60 Hz) 0.01 kV (0.01 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 3) kV (3 ~ 5) kV (5 ~ 10) kV (10 ~ 20) kV (20 ~ 100) kV	2.2 V 4.4×10^{-3} 2.3×10^{-3} 1.2×10^{-3} 7.5×10^{-4} 6.0×10^{-4} 4.0×10^{-4} 3.7×10^{-4}	
Logic analyzers	40436			Calibrator/ SICT-CP-40436
DC Voltage		(0 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V	1.5 μ V 8.8×10^{-6} 7.2×10^{-6}	
Clock frequency		10 MHz	7.7×10^{-12}	
Telephone testers	40437			Tone Pulse Simulator/ SICT-CP-40437
L1, L2 Output Voltage		(1 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 1 000) V	9.9×10^{-4} 1.1×10^{-5} 7.1×10^{-6} 8.5×10^{-6}	
Loop Current		(0.1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A	2.5×10^{-5} 5.2×10^{-5} 2.2×10^{-4} 4.7×10^{-4}	
Ring Output Voltage		(10 Hz ~ 20 kHz) 100 mV ~ 1 V (1 ~ 100) V (100 ~ 1 000) V	4.7×10^{-4} 2.9×10^{-4} 3.1×10^{-4}	
Ring Frequency		(1 ~ 1 000) Hz	7.0×10^{-5}	
D.T.M.F & Pulse		(+10 ~ -39.9) dBm	0.09 dB	
D.T.M.F & Frequency		(697 ~ 1 477) Hz	0.59 Hz	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Video signal analyzers	40438			Video Amplitude Calibration
Color Bar Decoding Accuracy(Gain)		(0 ~ 5) mV (5 ~ 10) mV (10 ~ 100) mV (100 ~ 200) mV (200 ~ 500) mV (500 ~ 1 000) mV	2.5 × 10 ⁻¹ 5.0 × 10 ⁻² 8.2 × 10 ⁻² 4.1 × 10 ⁻² 2.1 × 10 ⁻² 8.5 × 10 ⁻³	Fixture/ SICT-CP-40438
Frequency		20 Hz ~ 5 MHz	5.8 × 10 ⁻⁶	
Color Bar Decoding Accuracy(Phase)		(0 ~ 360)°	0.70°	
Measure Square Wave		(0 ~ 5) mV (5 ~ 10) mV (10 ~ 100) mV (100 ~ 300) mV (300 ~ 400) mV (400 ~ 600) mV (600 ~ 999.9) mV	9.4 × 10 ⁻² 2.0 × 10 ⁻² 9.9 × 10 ⁻³ 2.1 × 10 ⁻³ 1.5 × 10 ⁻³ 1.2 × 10 ⁻³ 9.4 × 10 ⁻⁴	
Measure Sine Wave		No Filter, PAL NTS BW Lim, NTSC,PAL Chroma BP, NTSC,PAL (10 kHz ~ 10 MHz) 500 mV		
Burst Frequency		(3 ~ 5) MHz	7.0 × 10 ⁻³	
vertical Gain		(0 ~ 5) mV (5 ~ 10) mV (10 ~ 100) mV (100 ~ 300) mV (300 ~ 600) mV (600 ~ 999.9) mV	4.0 × 10 ⁻⁷ 9.4 × 10 ⁻² 2.0 × 10 ⁻² 9.9 × 10 ⁻³ 2.1 × 10 ⁻³ 1.5 × 10 ⁻³ 9.4 × 10 ⁻⁴	
Horizontal Frequency		(20 ~ 100) Hz 100 Hz ~ 10 kHz 10 kHz ~ 10 MHz	3.1 × 10 ⁻³ 6.1 × 10 ⁻⁴ 6.1 × 10 ⁻⁵	
Gain Frequency Response		Flat, Luminance, Chroma at (20 Hz ~ 20 MHz) 700 mV	7.0 × 10 ⁻³	
Transient Response (Video Noise)		(0 ~ 1 000) mV	1.3 × 10 ⁻²	
Luminance Volt Level		(0 ~ -30) dB	4.8 × 10 ⁻¹	
Chrominance AM/PM Level		(0 ~ -30) dB	6.7 × 10 ⁻¹	
Luminance Volt Level		(0 ~ 1 000) mV	1.7 × 10 ⁻⁵	
Luminance Inputt Level		(0 ~ 1 000) mV	1.8 × 10 ⁻⁵	
Chrominance Input Level		(0 ~ 1 000) mV	1.7 × 10 ⁻⁵	

405. Low frequency electric & magnetic fields

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Flux meters	40503			
Flux		0.1 mWb ~ 10 Wb	5.8×10^{-3}	Flux sources/ SICT-CP-40503
Flux sources	40504			
Flux		(0.1 ~ 50) mWb (0.05 ~ 0.1) Wb (0.1 ~ 10) Wb	6.6×10^{-4} 2.3×10^{-5} 1.4×10^{-5}	GPS receiver, Frequency counter/ SICT-CP-40504
Magnetometers	40508			
Gauss		(0 ~ 0.1) mT (0.1 ~ 0.5) mT (0.5 ~ 3) mT (3 ~ 5) mT (5 ~ 20) mT (20 ~ 30) mT (30 ~ 1 700) mT	7.1×10^{-2} 1.4×10^{-2} 7.0×10^{-3} 4.0×10^{-3} 3.0×10^{-3} 6.7×10^{-3} 6.4×10^{-3}	Helmholtz coil, Standard magnets/ SICT-CP-40508
Reference/standard magnets	40510			
Gauss		(1.5 ~ 30) mT (30 ~ 1 000) mT	7.3×10^{-3} 2.6×10^{-3}	Gaussmeters/ SICT-CP-40510

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF amplifiers	40601	10 Hz ~ 10 kHz (0 ~ 80) dB	0.07 dB	Power Sensor, Attenuator/ SICT-CP-40601
		10 kHz ~ 10 GHz (0 ~ 40) dB	0.08 dB	
		(40 ~ 80) dB	0.11 dB	
		(10 ~ 18) GHz (0 ~ 40) dB	0.10 dB	
		(40 ~ 80) dB	0.13 dB	
		(18 ~ 30) GHz (0 ~ 60) dB	0.20 dB	
		(60 ~ 70) dB	0.31 dB	
		(70 ~ 80) dB	0.75 dB	
		(30 ~ 40) GHz (0 ~ 60) dB	0.26 dB	
		(60 ~ 70) dB	0.35 dB	
		(70 ~ 80) dB	0.76 dB	
Harmonics		(40 ~ 50) GHz (0 ~ 60) dB	0.41 dB	
		(60 ~ 70) dB	0.47 dB	
		(70 ~ 80) dB	0.82 dB	
		(50 ~ 67) GHz (0 ~ 20) dB	0.36 dB	
		(20 ~ 45) dB	0.45 dB	
		(9 kHz ~ 40 GHz) (0 ~ 100) dBc	0.90 dB	
		(0 ~ 1)		
		10 Hz ~ 2 GHz	4.7×10^{-3}	
		(2 ~ 20) GHz	9.2×10^{-3}	
		(20 ~ 40) GHz	1.5×10^{-2}	
Reflection coefficient		(40 ~ 50) GHz	1.9×10^{-2}	
		(50 ~ 67) GHz	3.3×10^{-2}	
		(1 ~ ∞)		
		10 Hz ~ 2 GHz	9.5×10^{-3}	
		(2 ~ 20) GHz	1.9×10^{-2}	
		(20 ~ 40) GHz	3.1×10^{-2}	
SWR		(40 ~ 50) GHz	3.9×10^{-2}	
		(50 ~ 67) GHz	6.7×10^{-2}	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Coaxial attenuators	40602	(10 Hz ~ 1 kHz) (0 ~ 40) dB (40 ~ 50) dB (50 ~ 70) dB (1 kHz ~ 9 kHz) (0 ~ 40) dB (40 ~ 50) dB (50 ~ 70) dB (9 kHz ~ 26.5 GHz) (0 ~ 10) dB (10 ~ 30) dB (30 ~ 40) dB (40 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB (80 ~ 90) dB (90 ~ 110) dB (110 ~ 120) dB (26.5 ~ 45) GHz (0 ~ 20) dB (20 ~ 70) dB (45 ~ 67) GHz (0 ~ 10) dB (10 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB	0.063 dB 0.070 dB 0.12 dB 0.063 dB 0.068 dB 0.11 dB 0.064 dB 0.069 dB 0.073 dB 0.077 dB 0.081 dB 0.086 dB 0.090 dB 0.095 dB 0.11 dB 0.12 dB 0.21 dB 0.24 dB 0.23 dB 0.25 dB 0.27 dB 0.32 dB	Power Sensor, Directional Coupler/ SICT-CP-40602
Reflection coefficient		(0 ~ 1) 10 Hz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.004 7 0.009 2 0.015 0.019 0.033	
SWR		(1 ~ ∞) 10 Hz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.009 5 0.019 0.031 0.039 0.067	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Burst pulse generators	40605	50 Ω (±) 5 V (5 ~ 100) V (0.1 ~ 8) kV	2.0 × 10 ⁻² 1.6 × 10 ⁻² 2.0 × 10 ⁻²	Digital Oscilloscope/ SICT-CP-40605
		1 kΩ (±) 5 V ~ 8 kV	4.0 × 10 ⁻²	
		1 ns (1 ~ 2) ns (2 ~ 4) ns 4 ns ~ 1 μs (1 ~ 2) μs (2 ~ 4) μs (4 ~ 10) μs	2.0 × 10 ⁻² 6.8 × 10 ⁻³ 2.6 × 10 ⁻³ 1.5 × 10 ⁻³ 6.2 × 10 ⁻³ 2.6 × 10 ⁻³ 1.3 × 10 ⁻³	
		1 ns (1 ~ 2) ns 2 ns ~ 200 ms	6.0 × 10 ⁻³ 3.1 × 10 ⁻³ 1.5 × 10 ⁻³	
		1 ns (1 ~ 2) ns 2 ns ~ 400 ms 400 ms ~ 10 s	6.0 × 10 ⁻³ 3.1 × 10 ⁻³ 1.5 × 10 ⁻³ 1.2 × 10 ⁻³	
		1 Hz ~ 25 MHz	1.6 × 10 ⁻³	
		(0 ~ 10) dB (10 ~ 20) dB (20 ~ 30) dB (30 ~ 40) dB (40 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB (80 ~ 90) dB (90 ~ 100) dB (100 ~ 110) dB (110 ~ 120) dB	0.024 dB 0.025 dB 0.027 dB 0.029 dB 0.031 dB 0.034 dB 0.036 dB 0.039 dB 0.042 dB 0.045 dB 0.048 dB 0.052 dB	
Attenuator calibrators	40606	3 μW 10 μW 30 μW 100 μW 300 μW 1 mW 3 mW 10 mW 30 mW 100 mW	0.1 nW 0.2 nW 0.5 nW 1 nW 4 nW 0.18 μW 0.19 μW 0.2 μW 0.3 μW 1 μW	Verification Kit/ SICT-CP-40606
RF power meter calibrators	40607	3 μW 10 μW 30 μW 100 μW 300 μW 1 mW 3 mW 10 mW 30 mW 100 mW	0.1 nW 0.2 nW 0.5 nW 1 nW 4 nW 0.18 μW 0.19 μW 0.2 μW 0.3 μW 1 μW	Digital Multimeter/ SICT-CP-40607

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
EMC transducers ; current probes, absorbing clamps, etc.	40608			Power Senso, Network analyzer/ SICT-CP-40608
Transfer Impedance		10 Hz ~ 50 MHz (50 ~ 200) MHz 200 MHz ~ 3 GHz	0.60 dB 1.1 dB 1.9 dB	
Insertion Loss		30 MHz ~ 1 GHz	1.9 dB	
Electric Magnetic Near-Field		100 kHz ~ 1 GHz	1.9 dB	
Reflection coefficient		(0 ~ 1) 10 Hz ~ 1 GHz (1 ~ 3) GHz	4.2×10^{-3} 6.0×10^{-3}	
SWR		(1 ~ ∞) 10 Hz ~ 1 GHz (1 ~ 3) GHz	9.0×10^{-3} 1.3×10^{-2}	
Coaxial directional couplers/ splitters	40610			Power Sensor, Synthesized Sweeper/ SICT-CP-40610
Coupling Factor		(10 Hz ~ 10 kHz) (0 ~ 40) dB (40 ~ 50) dB (50 ~ 70) dB	0.063 dB 0.070 dB 0.12 dB	
		(10 kHz ~ 100 kHz) (0 ~ 30) dB (30 ~ 50) dB (50 ~ 70) dB	0.064 dB 0.087 dB 0.098 dB	
		(100 kHz ~ 15 GHz) (0 ~ 40) dB (40 ~ 50) dB (50 ~ 70) dB	0.090 dB 0.093 dB 0.11 dB	
		(15 GHz ~ 18 GHz) (0 ~ 70) dB	0.12 dB	
		(18 GHz ~ 26.5 GHz) (0 ~ 70) dB	0.17 dB	
		(26.5 GHz ~ 45 GHz) (0 ~ 20) dB (20 ~ 70) dB	0.21 dB 0.24 dB	
		(45 GHz ~ 67 GHz) (0 ~ 10) dB (10 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB	0.23 dB 0.25 dB 0.27 dB 0.32 dB	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Coaxial directional couplers/ splitters	40610			Power Sensor, Synthesized Sweeper/
Reflection coefficient		(0 ~ 1) 10 Hz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.004 7 0.009 2 0.015 0.019 0.033	
SWR		(1 ~ ∞) 10 Hz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.009 5 0.019 0.031 0.039 0.067	
Electrostatic discharge generators	40613			Digital Oscilloscope/ SICT-CP-40613
Peak Current(I_p)		(\pm) (3.75 ~ 7.5) A (7.5 ~ 15) A (15 ~ 22.5) A (22.5 ~ 56.3) A (56.3 ~ 93.8) A (93.8 ~ 112.5) A (112.5 ~ 150) A	5.1×10^{-2} 5.3×10^{-2} 4.6×10^{-2} 5.2×10^{-2} 4.9×10^{-2} 5.7×10^{-2} 5.2×10^{-2}	
Current I_1 (30 ~ 60) ns		(\pm) 2 A (2 ~ 4) A (4 ~ 8) A (8 ~ 16) A (16 ~ 36) A (36 ~ 50) A (50 ~ 60) A (60 ~ 80) A	4.5×10^{-2} 5.0×10^{-2} 5.3×10^{-2} 4.9×10^{-2} 5.0×10^{-2} 4.4×10^{-2} 5.7×10^{-2} 5.2×10^{-2}	
Current I_2 (60 ~ 130) ns		(\pm) 1 A (1 ~ 2) A (2 ~ 4) A (4 ~ 6) A (6 ~ 8) A (8 ~ 15) A (15 ~ 25) A (25 ~ 30) A (30 ~ 40) A	5.0×10^{-2} 5.4×10^{-2} 5.7×10^{-2} 4.9×10^{-2} 5.4×10^{-2} 6.5×10^{-2} 5.2×10^{-2} 6.7×10^{-2} 6.1×10^{-2}	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrostatic discharge generators Current I3 (360 ~ 800) ns	40613	(±) 0.275 A (0.275 ~ 0.55) A (0.55 ~ 1.1) A (1.1 ~ 1.65) A (1.65 ~ 4.13) A (4.13 ~ 6.88) A (6.88 ~ 8.25) A (8.25 ~ 11) A	1.5×10^{-1} 2.2×10^{-1} 1.9×10^{-1} 1.6×10^{-1} 2.3×10^{-1} 1.5×10^{-1} 2.4×10^{-1} 1.9×10^{-1}	Digital Oscilloscope/ SICT-CP-40613
Current I4 (30 ~ 65) ns		(±) 0.15 A (0.15 ~ 0.3) A (0.3 ~ 0.6) A (0.6 ~ 1.2) A (1.2 ~ 2.25) A (2.25 ~ 2.7) A (2.7 ~ 3.75) A (3.75 ~ 4.5) A (4.5 ~ 6) A	1.3×10^{-1} 1.9×10^{-1} 3.0×10^{-1} 2.1×10^{-1} 2.5×10^{-1} 2.0×10^{-1} 1.5×10^{-1} 2.9×10^{-1} 2.0×10^{-1}	
Semiconductor Peak Current HBM		(±) (20 ~ 83.3) mA (0.083 3 ~ 1.33) A (1.33 ~ 6.66) A	3.1×10^{-2} 3.6×10^{-2} 2.7×10^{-2}	
Semiconductor Peak Current MM		(±) (0.219 ~ 14) A (14 ~ 35) A	3.5×10^{-2} 3.1×10^{-2}	
Time		0.1 ns 0.1 ns ~ 1 ms	2.7×10^{-2} 2.4×10^{-2}	
Peak Voltage		(±) 1 kV (1 ~ 35) kV	3.0×10^{-2} 2.5×10^{-2}	
EMC receivers	40614	100 kHz ~ 1 GHz 9 kHz ~ 50 MHz 50 MHz ~ 8 GHz (8 ~ 19) GHz (19 ~ 26) GHz (26 ~ 40) GHz 1 Hz ~ 10 MHz 1 Hz ~ 10 MHz 1 Hz ~ 1 GHz	6.1×10^{-10} 0.011 0.028 0.035 0.045 0.064 65 mHz 6.7×10^{-4} 0.12 dB	Network Analyzer, Pulse Generator/ SICT-CP-40614

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
EMC receivers	40614			Network Analyzer, Pulse Generator/ SICT-CP-40614
Frequency Response		10 Hz ~ 5 kHz 5 kHz ~ 1 GHz (1 ~ 10) GHz (10 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz (40 ~ 50) GHz	0.01 dB 0.07 dB 0.08 dB 0.09 dB 0.11 dB 0.16 dB 0.20 dB	
Frequency Response (CISPR)		9 kHz ~ 1 GHz	0.80 dB	
Display linearity accuracy		(80 ~ 50) dB μ V (50 ~ 0) dB μ V	0.10 dB 0.14 dB	
Input Attenuation		(0 ~ 30) dB (30 ~ 70) dB (70 ~ 110) dB	0.15 dB 0.12 dB 0.10 dB	
Noise Indicator		DC ~ 26.5 GHz	0.16 dB	
Interference Immunity		9 kHz ~ 40 GHz	0.67 dB	
RF filters	40615			Network Analyzer/ SICT-CP-40615
Reject Frequency		(9 ~ 90) kHz (90 ~ 900) kHz 900 kHz ~ 900 MHz 900 MHz ~ 18 GHz (18 ~ 50) GHz	0.024 kHz 0.24 kHz 0.025 MHz 0.068 MHz 0.12 MHz	
Insertion Loss		(9 kHz ~ 8 GHz) (0 ~ 10) dB (10 ~ 20) dB (20 ~ 40) dB (40 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB (80 ~ 90) dB (90 ~ 100) dB (8 ~ 18) GHz (0 ~ 10) dB (10 ~ 30) dB (30 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB (80 ~ 90) dB (90 ~ 100) dB	0.13 dB 0.14 dB 0.15 dB 0.16 dB 0.18 dB 0.23 dB 0.66 dB 1.7 dB 4.3 dB 0.23 dB 0.24 dB 0.25 dB 0.26 dB 0.31 dB 0.73 dB 1.7 dB 4.3 dB	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
RF filters Insertion Loss	40615	(18 ~ 50) GHz (0 ~ 10) dB (10 ~ 20) dB (20 ~ 30) dB (20 ~ 40) dB (40 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB (80 ~ 90) dB (90 ~ 100) dB	0.48 dB 0.51 dB 0.52 dB 0.53 dB 0.54 dB 0.59 dB 0.78 dB 1.6 dB 2.8 dB 6.0 dB	Network Analyzer/ SICT-CP-40615	
RF impedance meters RF Level	40616	(100 kHz ~ 18 GHz) (35 ~ 20) dBm (20 Hz ~ 18 GHz) (20 ~ -70) dBm	0.11 dB 0.12 dB	Performance Kit/ SICT-CP-40616	
Frequency		9 kHz ~ 0.1 MHz 0.1 MHz ~ 18 GHz	6.8×10^{-10} 6.2×10^{-11}		
Load Measurement		DC 10 Hz ~ 100 MHz (100 ~ 500) MHz 500 MHz ~ 1.8 GHz (1.8 ~ 3.0) GHz (3.0 ~ 18) GHz	0.02 Ω 0.06 Ω 0.15 Ω 0.21 Ω 0.41 Ω 1.1 Ω		
RF impulse generators Pulse Level	40617	9 kHz ~ 1 GHz	0.33 dB	Digital Oscilloscope/ SICT-CP-40617	
Line impedance stabilization networks; LISN, CDN, ISN, etc.	40618	Impedance Phase Insertion Loss Decoupling attenuation(Isolation) Coupling/Decoupling network(Impedance) Coupling/Decoupling network (Insertion loss)	9 kHz ~ 1 000 MHz 9 kHz ~ 1 000 MHz (0 ~ 100) dB 9 kHz ~ 100 MHz (100 ~ 1 000) MHz (0 ~ 100) dB (9 ~ 30) kHz (0.03 ~ 20) MHz (20 ~ 1 000) MHz 9 kHz ~ 1 000 MHz (0 ~ 100) dB 9 kHz ~ 1 000 MHz	2.0×10^{-2} 1.2 ° 0.07 dB 0.08 dB 0.21 dB 0.20 dB 0.21 dB 2.0×10^{-2} 0.10 dB	Impedance/Gain-Phase Analyzer, Calibration Kit/ SICT-CP-40618

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Coaxial standard mismatches	40619			Network Analyzer, Calibration Kit/ SICT-CP-40619-1
(Coaxial standard mismatches)				
Reflection coefficient		(0 ~ 1) 10 Hz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.004 7 0.009 2 0.015 0.019 0.033	
SWR		(1 ~ ∞) 10 Hz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.009 5 0.019 0.031 0.039 0.067	
(Calibration kit)				SICT-CP-40619-2
Magnitude of reflection coefficient		(Termination) 9 kHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.004 7 0.009 2 0.015 0.019 0.033	
		(Short, Open circuit) 9 kHz ~ 45 MHz 45 MHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.020 0.016 0.026 0.044 0.056 0.066	
SWR		(1 ~ 1.01) 9 kHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.009 5 0.019 0.031 0.039 0.067	
		(1.01 ~ 1.05) 9 kHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (50 ~ 67) GHz	0.011 0.020 0.032 0.040 0.070	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Coaxial standard mismatches	40619	(1.05 ~ 1.2)		Network Analyzer, Calibration
	SWR	9 kHz ~ 45 MHz	0.015	Kit /
		45 MHz ~ 2 GHz	0.014	SICT-CP-40619-2
		(2 ~ 20) GHz	0.024	
		(20 ~ 40) GHz	0.039	
		(40 ~ 50) GHz	0.049	
		(50 ~ 67) GHz	0.083	
		(1.2 ~ 1.5)		
		9 kHz ~ 45 MHz	0.023	
		45 MHz ~ 2 GHz	0.021	
		(2 ~ 20) GHz	0.035	
		(20 ~ 40) GHz	0.054	
		(40 ~ 50) GHz	0.066	
		(50 ~ 67) GHz	0.12	
		(1.5 ~ 2)		
		9 kHz ~ 45 MHz	0.040	
		45 MHz ~ 2 GHz	0.036	
		(2 ~ 20) GHz	0.055	
		(20 ~ 40) GHz	0.086	
		(40 ~ 50) GHz	0.11	
		(50 ~ 67) GHz	0.18	
		(2 ~ 3)		
		9 kHz ~ 45 MHz	0.096	
		45 MHz ~ 2 GHz	0.076	
		(2 ~ 20) GHz	0.12	
		(20 ~ 40) GHz	0.20	
		(40 ~ 50) GHz	0.24	
		(50 ~ 67) GHz	0.35	
		(3 ~ 9)		
		9 kHz ~ 45 MHz	0.80	
		45 MHz ~ 2 GHz	0.70	
		(2 ~ 20) GHz	1.1	
		(20 ~ 40) GHz	1.8	
		(40 ~ 50) GHz	2.2	
		(50 ~ 67) GHz	2.8	
Phase of reflection coefficient		(Termination)		
		9 kHz ~ 67 GHz	180°	
		(Short, Open circuit)		
		9 kHz ~ 45 MHz	1.2°	
		45 MHz ~ 2 GHz	0.87°	
		(2 ~ 20) GHz	1.5°	
		(20 ~ 40) GHz	2.6°	
		(40 ~ 50) GHz	3.2°	
		(50 ~ 67) GHz	3.8°	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Mobile communication test sets	40621			Measuring Receiver, RF Signal Generator/ SICT-CP-40621
RF Output Level		(35 ~ 20) dBm 100 kHz ~ 18 GHz	0.08 dB	
		(20 ~ -20) dBm 9 kHz ~ 1 GHz	0.06 dB	
		(1 ~ 18) GHz	0.11 dB	
		(18 ~ 40) GHz	0.14 dB	
		(40 ~ 50) GHz	0.18 dB	
		(-20 ~ -60) dBm 9 kHz ~ 1 GHz	0.06 dB	
		(1 ~ 18) GHz	0.11 dB	
		(18 ~ 40) GHz	0.16 dB	
		(40 ~ 50) GHz	0.21 dB	
		(-60 ~ -70) dBm 9 kHz ~ 1 GHz	0.08 dB	
		(1 ~ 18) GHz	0.10 dB	
		(18 ~ 40) GHz	0.18 dB	
		(40 ~ 50) GHz	0.21 dB	
		(-70 ~ -120) dBm 9 kHz ~ 26.5 GHz	0.15 dB	
Amplitude Modulation		(0.1 ~ 100) %	1.2×10^{-2}	
Frequency Modulation		(0.1 ~ 400) kHz	1.2×10^{-2}	
Phase Modulation		(0.1 ~ 400) rad	1.2×10^{-2}	
Distortion Harmonics of Modulation Rate Signal		≤ 20 %	2.3×10^{-2}	
Harmonics		(0 ~ -90) dB	0.36 dB	
Frequency Output Accuracy		9 kHz ~ 40 GHz	7.9×10^{-11}	
AC Output Level		(10 Hz ~ 100 kHz) (1 ~ 100) mV	5.2×10^{-4}	
		100 mV ~ 100 V	5.8×10^{-4}	
DC Output Level		1 mV ~ 1 V (1 ~ 10) V	1.5×10^{-6}	
		(10 ~ 100) V	1.3×10^{-6}	
			1.8×10^{-6}	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Mobile communication test sets	40621	(10 Hz ~ 100 kHz)		Measuring Receiver, RF Signal Generator/ SICT-CP-40621
AC Input Level		(1 ~ 10) mV	7.6×10^{-3}	
		(10 ~ 100) mV	1.3×10^{-3}	
		(0.1 ~ 1) V	6.7×10^{-4}	
		(1 ~ 10) V	1.7×10^{-4}	
		(10 ~ 100) V	2.0×10^{-4}	
DC Input Level		(1 ~ 10) mV	1.3×10^{-3}	
		(10 ~ 100) mV	3.1×10^{-4}	
		(0.1 ~ 1) V	6.2×10^{-5}	
		(1 ~ 100) V	6.1×10^{-5}	
RF Input Level		(9 kHz ~ 18 GHz)		
		(10 ~ -70) dBm	0.10 dB	
		(18 ~ 40) GHz		
		(10 ~ -70) dBm	0.16 dB	
Modulation meters	40622	0 kHz	0.01 %	Measuring Receiver/ SICT-CP-40622
Amplitude Modulation		(0 ~ 400) kHz	1.2×10^{-2}	
Frequency Modulation		0 %	1 Hz	
		(0 ~ 100) %	1.2×10^{-2}	
Phase Modulation		0 rad	1.2 mrad	
		(0 ~ 400) rad	1.2×10^{-2}	
Network analyzers	40623	10 Hz ~ 40 GHz	6.8×10^{-10}	Power Sensor, Verification Kit/ SICT-CP-40623
Frequency		(20 ~ -30) dBm		
Source Power Level		10 Hz ~ 100 Hz	0.01 dB	
		100 Hz ~ 500 MHz	0.08 dB	
		500 MHz ~ 10 GHz	0.09 dB	
		(10 ~ 18) GHz	0.11 dB	
		(18 ~ 27) GHz	0.13 dB	
		(27 ~ 40) GHz	0.16 dB	
		(40 ~ 50) GHz	0.21 dB	
		(50 ~ 80) GHz	0.45 dB	
		(80 ~ 110) GHz	0.53 dB	
		(-30 ~ -60) dBm		
		9 kHz ~ 500 MHz	0.08 dB	
		500 MHz ~ 10 GHz	0.09 dB	
		(10 ~ 18) GHz	0.11 dB	
		(18 ~ 27) GHz	0.13 dB	
		(27 ~ 40) GHz	0.16 dB	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Network analyzers	40623	(100 kHz ~ 18 GHz)		Power Sensor, Verification Kit/ SICT-CP-40623
Dynamic Range		(0 ~ 10) dB	0.086 dB	
		(10 ~ 20) dB	0.087 dB	
		(20 ~ 30) dB	0.088 dB	
		(30 ~ 40) dB	0.091 dB	
		(40 ~ 50) dB	0.099 dB	
		(50 ~ 60) dB	0.11 dB	
		(60 ~ 70) dB	0.12 dB	
		(70 ~ 80) dB	0.14 dB	
		(80 ~ 90) dB	0.15 dB	
		(90 ~ 100) dB	0.18 dB	
Attenuation		(20 dB)		
		300 kHz ~ 1.5 GHz	0.050 dB	
		(1.5 ~ 8) GHz	0.051 dB	
		(8 ~ 18) GHz	0.055 dB	
		(18 ~ 26.5) GHz	0.067 dB	
		(40 dB)		
		300 kHz ~ 1.5 GHz	0.054 dB	
		(1.5 ~ 8) GHz	0.055 dB	
		(8 ~ 18) GHz	0.059 dB	
		(18 ~ 26.5) GHz	0.082 dB	
Phase		(±180 °)		
		300 kHz ~ 45 MHz	0.04°	
		45 MHz ~ 2.0 GHz	0.09°	
		(2.0 ~ 3.0) GHz	0.10°	
		(3.0 ~ 4.5) GHz	0.11°	
		(4.5 ~ 6.0) GHz	0.15°	
		(6.0 ~ 7.5) GHz	0.19°	
		(7.5 ~ 8.0) GHz	0.21°	
		(8.0 ~ 9.0) GHz	0.22°	
		(9.0 ~ 10.5) GHz	0.29°	
		(10.5 ~ 12.0) GHz	0.28°	
		(12.0 ~ 13.5) GHz	0.26°	
		(13.5 ~ 15.0) GHz	0.27°	
		(15.0 ~ 16.5) GHz	0.26°	
		(16.5 ~ 18.0) GHz	0.29°	
		(18.0 ~ 21.0) GHz	0.31°	
		(21.0 ~ 22.5) GHz	0.28°	
		(22.5 ~ 24.0) GHz	0.39°	
		(24.0 ~ 25.5) GHz	0.33°	
		(25.5 ~ 26.5) GHz	0.44°	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Network analyzers	40623 SWR	(30 kHz ~ 2 GHz) 1.05 1.20 1.50 2.00 (2 ~ 18) GHz 1.05 1.20 1.50 2.00	0.021 0.021 0.021 0.021 0.018 0.018 0.018 0.024	Power Sensor, Verification Kit/ SICT-CP-40623	
Noise figure meters	40624	10 MHz ~ 26.5 GHz 10 MHz 9 kHz ~ 1 GHz (1 ~ 20) GHz (20 ~ 26.5) GHz DC voltage Range Noise Figure	4.3×10^{-5} 6.1×10^{-10} 0.008 0.019 0.03 0.000 18 V 0.052 dB 0.12 dB 0.16 dB 0.37 dB	Noise Source/ SICT-CP-40624	
Noise generators	40625	(-80 ~ -130) dBm/Hz (0 ~ 50) dB	0.10 dB 0.27 dB	Spectrum Analyzer/ SICT-CP-40625	
Noise impulse simulators	40626	Peak Voltage Rise/Fall Time Pulse Width	(±) 0.1 kV (0.1 ~ 5) kV 1 ns (1 ~ 2) ns (2 ~ 4) ns 10 ns (10 ~ 1 000) ns	4.0×10^{-2} 3.5×10^{-2} 6.0×10^{-3} 3.1×10^{-3} 1.5×10^{-3} 2.0×10^{-3} 1.5×10^{-3}	Digital Oscilloscope/ SICT-CP-40626
RF phase noise meters	40627	RF phase noise	(Carrier Frequency) 100 MHz ~ 18 GHz (Offset Frequency) 10 Hz ~ 100 MHz	1.0 dB 1.0 dB	RF Signal analyzer/ SICT-CP-40627

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Coaxial noise sources	ENR	40628 (4.5 ~ 16) dB (0.01 ~ 1) GHz (1 ~ 7) GHz (7 ~ 8) GHz (8 ~ 14) GHz (14 ~ 18) GHz (12 ~ 17) dB (0.01 ~ 1) GHz (1 ~ 2) GHz (2 ~ 6) GHz (6 ~ 7) GHz (7 ~ 12) GHz (12 ~ 18) GHz (18 ~ 26.5) GHz	0.28 dB 0.27 dB 0.30 dB 0.31 dB 0.32 dB 0.31 dB 0.28 dB 0.30 dB 0.29 dB 0.40 dB 0.41 dB 0.47 dB	Coaxial noise sources, Noise figure analyzer/ SICT-CP-40628
	SWR	(0 ~ 1) (0.01 ~ 3) GHz (3 ~ 20) GHz (20 ~ 26.5) GHz	0.006 8 0.010 0.015	
RF power meters	40635	High power (0.1 ~ 500) W 10 kHz ~ 250 MHz (0.1 ~ 150) W (80 ~ 1 000) MHz (0.1 ~ 10) W (1 000 ~ 4 200) MHz Zero Carryover Power Calibration Factor Power Ref. Output	2.6 × 10 ⁻² 2.6 × 10 ⁻² 2.7 × 10 ⁻² 3 nW 0.01 mW 1.6 × 10 ⁻³ 0.5 × 10 ⁻³ 8 μW	Range Calibrator/ SICT-CP-40635

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Diode power sensors	40636 Cal factor Reflection coefficient SWR	(1 ~ 10) μ W		Thermistor Mount, Synthesized Sweeper/ SICT-CP-40636
		9 kHz ~ 100 kHz	0.4×10^{-2}	
		100 kHz ~ 1 GHz	0.9×10^{-2}	
		(1 ~ 10) GHz	1.1×10^{-2}	
		(10 ~ 18) GHz	1.9×10^{-2}	
		(18 ~ 26.5) GHz	2.8×10^{-2}	
		(26.5 ~ 40) GHz	4.0×10^{-2}	
		 (10 μ W ~ 10 mW) 9 kHz ~ 100 kHz	 0.4×10^{-2}	
		100 kHz ~ 1 GHz	0.9×10^{-2}	
		(1 ~ 10) GHz	1.1×10^{-2}	
		(10 ~ 18) GHz	1.9×10^{-2}	
		(18 ~ 26.5) GHz	2.6×10^{-2}	
		(26.5 ~ 40) GHz	3.5×10^{-2}	
Thermocouple power sensors	40637 Cal Factor Reflection coefficient SWR	(0 ~ 1)		Thermistor Mount, Synthesized Sweeper/ SICT-CP-40637
		20 Hz ~ 1 GHz	4.2×10^{-3}	
		(1 ~ 20) GHz	9.4×10^{-3}	
		(20 ~ 40) GHz	1.5×10^{-2}	
		 (1 ~ ∞) 20 Hz ~ 1 GHz	 9.7×10^{-3}	
		(1 ~ 20) GHz	2.4×10^{-2}	
		(20 ~ 40) GHz	3.8×10^{-2}	
		 (1 ~ 10) μ W 9 kHz ~ 100 kHz	 0.4×10^{-2}	
		100 kHz ~ 1 GHz	0.9×10^{-2}	
		(1 ~ 10) GHz	1.1×10^{-2}	
		(10 ~ 18) GHz	1.9×10^{-2}	
		(18 ~ 26.5) GHz	2.8×10^{-2}	
		(26.5 ~ 40) GHz	4.0×10^{-2}	
		 (10 μ W ~ 10 mW) 9 kHz ~ 100 kHz	 0.4×10^{-2}	
		100 kHz ~ 1 GHz	0.9×10^{-2}	
		(1 ~ 10) GHz	1.1×10^{-2}	
		(10 ~ 18) GHz	1.9×10^{-2}	
		(18 ~ 26.5) GHz	2.6×10^{-2}	
		(26.5 ~ 40) GHz	3.5×10^{-2}	
		 (0 ~ 1) 20 Hz ~ 1 GHz	 4.2×10^{-3}	
		(1 ~ 20) GHz	9.4×10^{-3}	
		(20 ~ 40) GHz	1.5×10^{-2}	
		 (1 ~ ∞) 20 Hz ~ 1 GHz	 9.7×10^{-3}	
		(1 ~ 20) GHz	2.4×10^{-2}	
		(20 ~ 40) GHz	3.8×10^{-2}	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Pulse generators	40638			Digital Oscilloscope/ SICT-CP-40638
Period		100 ps ~ 1 s	5.8×10^{-9}	
Frequency		1 Hz ~ 1 GHz (1 ~ 3.35) GHz	5.8×10^{-9} 1.7×10^{-8}	
Width		100 ps ~ 1 s	1.2×10^{-3}	
Delay Time		100 ps ~ 1 s	1.2×10^{-3}	
Double Pulse		100 ps ~ 1 s	1.2×10^{-3}	
Duty Cycle		(1 ~ 99) %	0.006 2 %	
DC Level		$\pm(10 \text{ mV} \sim 100 \text{ V})$	5.8×10^{-4}	
Output Level		(100 Hz ~ 10 kHz) (10 ~ -20) dBm	0.018 dB	
Radar test sets	40639			Signal Analyzer, SART Generator, EPIRB Generator/ SICT-CP-40639
(Ship)				
RF Level		(20 ~ -20) dBm 20 Hz ~ 1 GHz (1 ~ 18) GHz	0.09 dB 0.13 dB	
		(-20 ~ -60) dBm 9 kHz ~ 1 GHz (1 ~ 18) GHz	0.06 dB 0.11 dB	
		(-60 ~ -120) dBm 10 MHz ~ 18 GHz	0.25 dB	
Amplitude Modulation		(0.1 ~ 100) %	1.2×10^{-2}	
Frequency Modulation		(0.1 ~ 400) kHz	1.2×10^{-2}	
Phase Modulation		(0.1 ~ 400) rad	1.2×10^{-2}	
Distortion of Modulation		(0 ~ 2) %	1.2×10^{-3}	
Harmonics		(9 kHz ~ 18 GHz) (0 ~ -110) dB	0.25 dB	
Frequency		9 kHz ~ 18 GHz	6.2×10^{-11}	
Pulse Period		1 ns ~ 10 ms	1.2×10^{-2}	
High power		(0.1 ~ 500) W 10 kHz ~ 250 MHz	2.6×10^{-2}	
		(0.1 ~ 150) W (80 ~ 1 000) MHz	2.6×10^{-2}	
		(0.1 ~ 10) W (1 000 ~ 4 200) MHz	2.7×10^{-2}	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Radar test sets (flight) Frequency(VOR/ILS/DME)	40639	(74.6 ~ 1 150) MHz	8.2×10^{-8}	Signal Analyzer, SART Generator, EPIRB Generator/ SICT-CP-40639
Amplitude Modulation(VOR/ILS)		Localizer (108.1 ~ 111.95) MHz (0.1 ~ 20) %	0.62 %	
		Glideslope (330.95 ~ 334.70) MHz (20 ~ 40) %	0.84 %	
		Marker Beacon (74.6 ~ 75.4) MHz (40 ~ 95) %	1.4 %	
		VOR (108 ~ 117.95) MHz (0.1 ~ 30) %	0.62 %	
고주파 레벨(VOR/ILS)		Localizer (108.1 ~ 111.95) MHz (10 ~ -30) dBm (-30 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -110) dBm (-110 ~ -120) dBm	0.19 dB 0.23 dB 0.24 dB 0.34 dB 0.35 dB	
		Glideslope (330.95 ~ 334.70) MHz (10 ~ -30) dBm (-30 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -110) dBm (-110 ~ -120) dBm	0.19 dB 0.23 dB 0.24 dB 0.34 dB 0.35 dB	
DDM(VOR/ILS)		Localizer (108.1 ~ 111.95) MHz LEFT(-0.200 ~ -0.155) LEFT(-0.155 ~ -0.093) CENTER(0.000) RIGHT(0.093 ~ 0.155) RIGHT(0.155 ~ 0.200)	0.000 5 0.000 5 0.000 5 0.000 5 0.000 5	
		Glideslope (330.95 ~ 334.70) MHz DOWN(0.400 ~ 0.175) DOWN(0.175 ~ 0.091) CENTER(0.000) UP(-0.091 ~ -0.175) UP(-0.175 ~ -0.400)	0.000 5 0.000 5 0.000 5 0.000 5 0.000 5	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Radar test sets	40639	Localizer (108.1 ~ 111.95) MHz (0 ~ 40) %	0.87 %	Signal Analyzer, SART Generator, EPIRB Generator/ SICT-CP-40639
SDM(VOR/ILS)		Glideslope (330.95 ~ 334.70) MHz (40 ~ 80) %	1.2 %	
BEARING ANGLE(VOR)		VOR (108 ~ 117.95) MHz (0 ~ 360) °	0.04 °	
RF signal generators	40640	(0.1 ~ 400) kHz	1.2×10^{-2}	Measuring Receiver/ SICT-CP-40640
Frequency Modulation		(0.1 ~ 100) %	1.2×10^{-2}	
Amplitude Modulation		(0.1 ~ 400) rad	1.2×10^{-2}	
Phase Modulation		(100 kHz ~ 12 000 MHz) Period (1 μs ~ 1 s)	1.2×10^{-3}	
Pulse Modulation		ton (100 ns ~ 100 μs)	1.2×10^{-3}	
		PRR	3.1×10^{-3}	
Distortion		≤ 20 %	2.3×10^{-2}	
RF Level		(54 ~ 57) dBm 100 kHz ~ 500 MHz	0.35 dB	
		(51 ~ 54) dBm 100 kHz ~ 2 GHz	0.32 dB	
		(35 ~ 51) dBm 100 kHz ~ 18 GHz	0.32 dB	
		(20 ~ 35) dBm 100 kHz ~ 10 GHz	0.26 dB	
		(10 ~ 18) GHz	0.27 dB	
		(-30 ~ 20) dBm 20 Hz ~ 10 GHz	0.09 dB	
		(10 ~ 18) GHz	0.11 dB	
		(18 ~ 28) GHz	0.12 dB	
		(28 ~ 40) GHz	0.16 dB	
		(40 ~ 50) GHz	0.20 dB	
		(50 ~ 70) GHz	0.29 dB	
		(70 ~ 110) GHz	0.38 dB	
		(-30 ~ -60) dBm 20 Hz ~ 10 GHz	0.10 dB	
		(10 ~ 18) GHz	0.11 dB	
		(18 ~ 28) GHz	0.12 dB	
		(28 ~ 40) GHz	0.16 dB	
		(40 ~ 50) GHz	0.21 dB	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF signal generators	40640	(-120 ~ -60) dBm 20 Hz ~ 4.2 GHz (4.2 ~ 8) GHz (8 ~ 12.4) GHz (12.4 ~ 18) GHz (18 ~ 26.5) GHz	0.21 dB 0.23 dB 0.24 dB 0.27 dB 0.31 dB	Measuring Receiver/ SICT-CP-40640
Harmonic Frequency		(-10 ~ -110) dBc 9 kHz ~ 40 GHz	0.37 dB 2.1×10^{-11}	
RF spectrum analyzers	40641	(3 ~ 100) Hz (100 ~ 500) Hz (500 ~ 900) Hz 900 Hz ~ 100 kHz 0.1 MHz ~ 40 GHz	2.0×10^{-4} 6.1×10^{-6} 1.2×10^{-6} 6.8×10^{-7} 6.2×10^{-9}	Power Sensor, Synthesized Sweeper/ SICT-CP-40641
Center Frequency		(3 ~ 100) Hz (100 ~ 500) Hz (500 ~ 900) Hz 900 Hz ~ 100 kHz 0.1 MHz ~ 40 GHz	2.0×10^{-4} 6.1×10^{-6} 1.2×10^{-6} 6.8×10^{-7} 6.2×10^{-9}	
Frequency Counter		10 Hz ~ 100 kHz 0.1 MHz ~ 40 GHz	7.6×10^{-3} 7.7×10^{-6}	
Span		1 Hz ~ 100 MHz	6.2×10^{-6}	
RBW		1 Hz ~ 100 MHz	3.2×10^{-2}	
RBW Selectivity		1 Hz ~ 100 MHz	0.022 dB	
RBW Switching		1 dB ~ 10 dB scale/div	0.022 dB	
Scale Switching		(0 ~ -30) dB (-30 ~ -40) dB (-40 ~ -50) dB (-50 ~ -60) dB (-60 ~ -80) dB (-80 ~ -100) dB	0.073 dB 0.077 dB 0.082 dB 0.095 dB 0.13 dB 0.17 dB	
Scale Fidelity		(10 ~ 100) Hz 100 Hz ~ 1 GHz (1 ~ 6) GHz (6 ~ 10) GHz (10 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 30) GHz (30 ~ 35) GHz (35 ~ 40) GHz	0.01 dB 0.15 dB 0.16 dB 0.17 dB 0.19 dB 0.21 dB 0.23 dB 0.28 dB 0.32 dB	
Frequency Response		DC ~ 40 GHz	0.17 dB	
Average Noise Level		(-30 ~ 30) kHz	0.33 dB	
Sideband Noise Level		DC ~ 1 GHz	6.2×10^{-9}	
CAL Output Freq. & Int. Frequency		(-20 ~ 20) dBm	0.09 dB	
CAL Output Level		(5 ~ 3 000) m/s	0.01 m/s	Signal Generator/ SICT-CP-40642
RF speed guns	40642			

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Surge generators	40643	(±)		Digital Oscilloscope/ SICT-CP-40643
		5 mV	3.2×10^{-2}	
		5 mV ~ 2 V	3.0×10^{-2}	
		(2 ~ 5) V	4.0×10^{-2}	
		5 V ~ 200 kV	3.5×10^{-2}	
		(±)		
		5 A	3.3×10^{-2}	
		5 A ~ 200 kA	3.6×10^{-2}	
		1 ns	6.0×10^{-3}	
		(1 ~ 2) ns	3.0×10^{-3}	
Time measurement by section		2 ns ~ 10 s	2.0×10^{-3}	
		1 ns	6.0×10^{-3}	
		(1 ~ 2) ns	3.0×10^{-3}	
		2 ns ~ 10 s	2.0×10^{-3}	
		1 Hz ~ 25 MHz	1.6×10^{-3}	
		at 50 Hz		
		(0 ~ 360)°	1.2°	
		at 60 Hz		
		(0 ~ 360)°	1.4°	
SWR meters	40644	9 kHz ~ 18 GHz	6.4×10^{-5}	Coaxial Mismatch/ SICT-CP-40644
		30 kHz ~ 100 MHz	0.06 dB	
		100 MHz ~ 10 GHz	0.08 dB	
		(10 ~ 18) GHz	0.09 dB	
		(30 kHz ~ 30 MHz)		
		1.05	0.019	
		1.20	0.019	
		1.50	0.019	
		2.00	0.020	
		(30 MHz ~ 2 GHz)		
		1.05	0.021	
		1.20	0.021	
		1.50	0.021	
		2.00	0.021	
		(2 ~ 18) GHz		
		1.05	0.018	
		1.20	0.018	
		1.50	0.018	
		2.00	0.024	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF terminations (Open, Short ,Phase)	40645	(± 180 °) 10 Hz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.49° 0.61° 0.95° 1.2°	Network Analyzer, Coaxial Mismatch/ SICT-CP-40645
(Reflection coefficient)		(0 ~ 1) 10 Hz ~ 100 kHz 100 kHz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.004 3 0.006 0 0.009 5 0.016 0.019	
(SWR)		(1 ~ ∞) 10 Hz ~ 100 kHz 100 kHz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.008 6 0.012 0.019 0.032 0.038	
(Impedance)		(0.000 0 ~ 0.047 6) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.64 Ω 1.0 Ω 1.6 Ω 2.0 Ω	
		(0.047 6 ~ 0.090 9) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.71 Ω 1.1 Ω 1.8 Ω 2.2 Ω	
		(0.090 9 ~ 0.166 7) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.84 Ω 1.3 Ω 2.1 Ω 2.7 Ω	
		(0.166 7 ~ 0.230 8) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.99 Ω 1.6 Ω 2.5 Ω 3.1 Ω	
		(0.230 8 ~ 0.285 7) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	1.1 Ω 1.8 Ω 2.9 Ω 3.6 Ω	
		(0.285 7 ~ 0.333 4) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	1.3 Ω 2.1 Ω 3.3 Ω 4.2 Ω	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF terminations (Impedance Phase)	40645	(0.000 0 ~ 0.047 6, ±180 °) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.68° 1.1° 1.7° 2.2°	Network Analyzer, Coaxial Mismatch/ SICT-CP-40645
		(0.047 6 ~ 0.090 9, ±180 °) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.67° 1.1° 1.7° 2.1°	
		(0.090 9 ~ 0.166 7, ±180 °) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.66° 1.1° 1.7° 2.1°	
		(0.166 7 ~ 0.230 8, ±180 °) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.65° 1.0° 1.6° 2.1°	
		(0.230 8 ~ 0.285 7, ±180 °) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.64° 1.0° 1.6° 2.0°	
		(0.285 7 ~ 0.333 4, ±180 °) 10 Hz ~ 500 MHz 500 MHz ~ 20 GHz (20 ~ 40) GHz (40 ~ 50) GHz	0.62° 1.0° 1.6° 2.0°	
Coaxial thermistor mounts	40646	(1 ~ 10) μW (9 ~ 100) kHz 100 kHz ~ 1 GHz (1 ~ 10) GHz (10 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz	0.4×10^{-2} 0.9×10^{-2} 1.1×10^{-2} 1.9×10^{-2} 2.8×10^{-2} 4.0×10^{-2}	Thermistor Mount, Synthesized Sweeper/ SICT-CP-40646
	Cal Factor	(10 μW ~ 10 mW) (9 ~ 100) kHz 100 kHz ~ 1 GHz (1 ~ 10) GHz (10 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz	0.4×10^{-2} 0.9×10^{-2} 1.1×10^{-2} 1.9×10^{-2} 2.6×10^{-2} 3.5×10^{-2}	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Coaxial thermistor mounts Reflection coefficient	40646	(0 ~ 1) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	4.2×10^{-3} 9.4×10^{-3} 1.5×10^{-2}	Synthesized Sweeper/ SICT-CP-40646	
SWR		(1 ~ ∞) 20 Hz ~ 1 GHz (1 ~ 20) GHz (20 ~ 40) GHz	9.7×10^{-3} 2.4×10^{-2} 3.8×10^{-2}		
RF voltmeters	40650	3 V 1 V 300 mV 270 mV 240 mV 210 mV 180 mV 150 mV 120 mV 100 mV 90 mV 60 mV 30 mV 10 mV 3 mV 1 mV	4.2 mV 1.4 mV 0.36 mV 0.32 mV 0.28 mV 0.26 mV 0.24 mV 0.22 mV 0.17 mV 0.16 mV 0.13 mV 0.11 mV 0.048 mV 0.020 mV 0.018 mV 0.013 mV	RF Millivolt Meter Calibrator/ SICT-CP-40650	
Vector voltmeters	40651	3 V 1 V 300 mV 100 mV 30 mV 10 mV 3 mV 1 mV	4.2 mV 1.4 mV 0.36 mV 0.16 mV 0.048 mV 0.046 mV 0.022 mV 0.024 mV	Signal Generator/ SICT-CP-40651	
		RF Phase	(0 ~ 270)°	0.006°	
Field strength meters	40652	Center frequency Scale Fidelity Frequency response	(9 ~ 100) kHz 0.1 MHz ~ 18 GHz (0 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -100) dB 9 kHz ~ 500 MHz 500 MHz ~ 18 GHz	6.8×10^{-8} 6.2×10^{-9} 0.11 dB 0.12 dB 0.13 dB 0.18 dB 0.05 dB 0.08 dB	Signal Generator/ SICT-CP-40652
AM/FM test sources	40653	Output frequency	(10 ~ 560) MHz	6.2×10^{-10}	Measuring Receiver/ SICT-CP-40653

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Dip simulators	40654			Digital Oscilloscope/ SICT-CP-40654
DC Voltage		1 V (1 ~ 5) V (5 ~ 10) V (10 ~ 50) V (50 ~ 100) V (100 ~ 500) V	1.1×10^{-5} 6.4×10^{-6} 5.4×10^{-6} 9.3×10^{-6} 8.0×10^{-6} 1.1×10^{-5}	
AC Voltage		(50 ~ 60) Hz 50 V (50 ~ 100) V (100 ~ 500) V	5.0×10^{-5} 2.5×10^{-5} 1.6×10^{-4}	
Frequency		50 Hz 60 Hz	8.4×10^{-6} 8.3×10^{-6}	
Dip DC Voltage		(0 ~ 50) V 0 % (0 ~ 120) %	0.2 V 3.4×10^{-2}	
Dip AC Voltage		(50 Hz ~ 60 Hz, 0 V ~ 400 V) 0 % (0 ~ 120) %	0.9 V 3.4×10^{-2}	
Time measurement by section		100 ns ~ 2 μ s (2 ~ 4) μ s (4 ~ 400) μ s (0.4 ~ 2) ms 2 ms ~ 5 s	1.8×10^{-3} 2.0×10^{-3} 1.6×10^{-3} 2.0×10^{-3} 1.6×10^{-3}	
Inrush Current		(5 ~ 1 000) A	3.6×10^{-2}	
Phase Shifting		at 50 Hz (0 ~ 360) $^{\circ}$ at 60 Hz (0 ~ 360) $^{\circ}$	1.2 $^{\circ}$ 1.4 $^{\circ}$	

407. Field strength & antenna

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Probes				
Electric Field Probe	40702	9 kHz ~ 300 MHz (1 ~ 750) V/m	0.12	RF Power Meter, Multi Meter/ SICT-CP-40702
		300 MHz ~ 400 MHz (1 ~ 300) V/m	0.12	
		400 MHz ~ 1 GHz (1 ~ 150) V/m	0.12	
		1 GHz ~ 6 GHz (1 ~ 150) V/m	0.12	
		6 GHz ~ 10 GHz (1 ~ 100) V/m	0.13	
		10 GHz ~ 18 GHz (2 ~ 100) V/m	0.13	
		18 GHz ~ 40 GHz (2 ~ 100) V/m	0.15	
Magnetic Field Probe		10 Hz ~ 1 kHz (0.39 ~ 1 000) A/m	0.07	
		1 kHz ~ 10 kHz (0.39 ~ 600) A/m	0.07	
		10 kHz ~ 30 kHz (2.65 ~ 390) mA/m (0.39 ~ 100) A/m	0.12 0.07	
		30 kHz ~ 150 kHz (2.65 ~ 390) mA/m (0.39 ~ 20) A/m	0.12 0.07	
		150 kHz ~ 200 kHz (2.65 ~ 390) mA/m (0.39 ~ 10) A/m	0.12 0.07	
		200 kHz ~ 300 MHz 2.65 mA/m ~ 1.98 A/m	0.12	
		300 MHz ~ 400 MHz 2.65 mA/m ~ 0.79 A/m	0.12	
		400 MHz ~ 1 GHz 2.65 mA/m ~ 0.39 A/m	0.12	

407. Field strength & antenna

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Dipole antennas (Dipole Antenna) Antenna Factor	40703	20 MHz ~ 18 GHz (1 ~ ∞) 20 MHz ~ 1 GHz 1 GHz ~ 18 GHz	1.1 dB 0.022 0.039	Network analyzer/ SICT-CP-40703 ancillary facilities(Chungju)
Voltage Standing Wave Ratio Antenna Pattern		700 MHz ~ 18 GHz	1.4 dB	
(Biconical Antenna) Antenna Factor		20 MHz ~ 18 GHz	1.2 dB	
Voltage Standing Wave Ratio Antenna Pattern		(1 ~ ∞) 20 MHz ~ 1 GHz 1 GHz ~ 18 GHz	0.022 0.039	
(Log Periodic Antenna) Antenna Factor		700 MHz ~ 18 GHz	1.4 dB	
Voltage Standing Wave Ratio Antenna Pattern		20 MHz ~ 18 GHz (1 ~ ∞) 20 MHz ~ 1 GHz 1 GHz ~ 18 GHz	1.2 dB 0.022 0.039	
Loop antennas Antenna Factor	40704	10 Hz ~ 30 MHz	1.2 dB	Signal generator1, Signal analyzer/ SICT-CP-40704
Monopole antennas Antenna Factor	40705	10 Hz ~ 30 MHz	1.4 dB	Signal generator1, Signal analyzer/ SICT-CP-40705
Horn antennas Antenna Factor	40707	200 MHz ~ 18 MHz 18 GHz ~ 40 GHz	1.1 dB 1.4 dB	Network analyzer/ SICT-CP-40707 ancillary facilities(Chungju)
Voltage Standing Wave Ratio Antenna Pattern		(1 ~ ∞) 200 MHz ~ 1 GHz 1 GHz ~ 18 GHz 18 GHz ~ 40 GHz 700 MHz ~ 18 GHz	0.022 0.039 0.041 1.4 dB	

501. Contact thermometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Temperature generators: ovens, furnaces, isothermal liquid baths, ice-point baths, dry-block calibrators	50101	0 °C (-196 ~ -95) °C (-95 ~ -90) °C (-90 ~ 250) °C (250 ~ 550) °C (550 ~ 660) °C (660 ~ 1 100) °C (1 100 ~ 1 600) °C	0.010 °C 0.060 °C 0.030 °C 0.017 °C 0.020 °C 0.060 °C 0.7 °C 1.7 °C	SPRT, STANDARD TC/ SICT-CP-50101
Temperature indicators/recorders /controllers, temperature calibrators (Temperature indicators/recorders/controllers)	50102	With Sensor (-196 ~ 500) °C (500 ~ 660) °C (660 ~ 700) °C (700 ~ 900) °C (900 ~ 1 100) °C (1 100 ~ 1 400) °C (1 400 ~ 1 600) °C Without Sensor (-196 ~ 0) °C (0 ~ 100) °C (100 ~ 200) °C (200 ~ 300) °C (300 ~ 400) °C (400 ~ 500) °C (500 ~ 600) °C (600 ~ 700) °C (700 ~ 800) °C (800 ~ 1 300) °C (1 300 ~ 1 600) °C (temperature calibrators) Output (-196 ~ 500) °C (500 ~ 600) °C (600 ~ 800) °C (800 ~ 1 300) °C (1 300 ~ 1 600) °C Input (-196 ~ 500) °C (500 ~ 600) °C (600 ~ 800) °C (800 ~ 1 300) °C (1 300 ~ 1 600) °C	0.020 °C 0.045 °C 0.59 °C 0.60 °C 0.61 °C 2.0 °C 2.1 °C 0.010 °C 0.013 °C 0.018 °C 0.022 °C 0.025 °C 0.029 °C 0.033 °C 0.040 °C 0.044 °C 0.07 °C 0.09 °C 0.005 °C 0.006 °C 0.007 °C 0.08 °C 0.10 °C 0.03 °C 0.04 °C 0.05 °C 0.07 °C 0.09 °C	SPRT, STANDARD TC/ SICT-CP-50102
Glass thermometers; liquid-in-glass, Beckmann liquid-in-glass	50103	(-90 ~ -58) °C (-58 ~ 400) °C (400 ~ 500) °C	0.15 °C 0.04 °C 0.15 °C	SPRT/ SICT-CP-50103

501. Contact thermometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistance thermometers; SPRT, IPRT, thermistors, etc.	50104	(-196 ~ 100) °C (100 ~ 300) °C (300 ~ 500) °C (500 ~ 660) °C (Fixed point) Ar TP Hg TP H ₂ O TP Ga MP In FP Sn FP Zn FP Al FP Ag FP	0.019 °C 0.020 °C 0.021 °C 0.044 °C -189.344 2 °C -38.834 4 °C 0.01 °C 29.764 6 °C 156.598 5 °C 231.928 °C 419.527 °C 660.323 °C 961.78 °C	SPRT, Fixed point/ SICT-CP-50104
Thermal expansion thermometers; bimetal, gas or liquid type	50105	bimetal (-196 ~ -70) °C (-70 ~ 100) °C (100 ~ 200) °C (200 ~ 500) °C (500 ~ 650) °C	0.6 °C 0.2 °C 0.3 °C 0.6 °C 1.2 °C	SPRT/ SICT-CP-50105
Thermocouples: noble metal, base metal, pure metal, special type, etc.	50106	Base metal (0 ~ 1 100) °C (1 100 ~ 1 300) °C (1 300 ~ 1 600) °C Noble metal (-196 ~ -90) °C (-90 ~ 300) °C (300 ~ 500) °C (500 ~ 660) °C (660 ~ 900) °C (900 ~ 1 100) °C (1 100 ~ 1 300) °C (Fixed point) H ₂ O ICE Point Sn FP Zn FP Al FP Ag FP Cu FP Co-C MP Fe MP	0.5 °C 1.7 °C 1.8 °C 0.4 °C 0.2 °C 0.3 °C 0.4 °C 1.1 °C 1.4 °C 1.8 °C 0.2 °C 0.2 °C 0.2 °C 0.2 °C 0.3 °C 0.3 °C 1.1 °C 1.6 °C	SPRT, Fixed point, STANDARD TC/ SICT-CP-50106

501. Contact thermometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Temperature transducers	50107	(-196 ~ 400) °C (400 ~ 500) °C (500 ~ 660) °C (660 ~ 800) °C (800 ~ 1 100) °C (1 100 ~ 1 300) °C (1 300 ~ 1 600) °C	0.031 °C 0.043 °C 0.072 °C 0.6 °C 0.7 °C 2.1 °C 2.2 °C	SPRT, THERMOCOUPLE, MULTIMETER SICT-CP-50107
Primary fixed-point cells and apparatus H ₂ O TP	50108	0.01 °C	0.24 mK	Triple-Point Cell SICT-CP-50108

502. non contact thermometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Optical pyrometers	50203	(900 ~ 1 800) °C	5 °C	Standard Lamp/ SICT-CP-50203
Standard radiation thermometers	50204	(-40 ~ -20) °C (-20 ~ 0) °C (0 ~ 10) °C (10 ~ 50) °C (50 ~ 100) °C (100 ~ 200) °C (200 ~ 300) °C (300 ~ 400) °C (400 ~ 500) °C (500 ~ 600) °C (600 ~ 700) °C (700 ~ 800) °C (800 ~ 900) °C (900 ~ 1 200) °C (1 200 ~ 1 400) °C (1 400 ~ 1 500) °C (1 500 ~ 1 600) °C (1 600 ~ 1 800) °C (1 800 ~ 2 000) °C (2 000 ~ 2 100) °C (2 100 ~ 2 200) °C (2 200 ~ 2 300) °C (2 300 ~ 2 400) °C	0.9 °C 0.7 °C 0.4 °C 0.3 °C 0.4 °C 0.5 °C 0.7 °C 0.8 °C 1.0 °C 1.1 °C 1.3 °C 1.6 °C 1.7 °C 1.8 °C 1.9 °C 2.0 °C 2.1 °C 2.5 °C 2.6 °C 4.2 °C 4.4 °C 4.6 °C 4.7 °C	Transfer Standard Pyrometer/ SICT-CP-50204
Thermal image apparatus	50205	(-40 ~ -20) °C (-20 ~ 0) °C (0 ~ 10) °C (10 ~ 50) °C (50 ~ 100) °C (100 ~ 200) °C (200 ~ 300) °C (300 ~ 400) °C (400 ~ 500) °C (500 ~ 600) °C (600 ~ 700) °C (700 ~ 800) °C (800 ~ 900) °C (900 ~ 1 200) °C (1 200 ~ 1 400) °C (1 400 ~ 1 500) °C (1 500 ~ 1 600) °C (1 600 ~ 1 800) °C (1 800 ~ 2 000) °C (2 000 ~ 2 100) °C (2 100 ~ 2 200) °C (2 200 ~ 2 300) °C (2 300 ~ 2 400) °C	0.9 °C 0.7 °C 0.4 °C 0.3 °C 0.4 °C 0.5 °C 0.7 °C 0.8 °C 1.0 °C 1.1 °C 1.3 °C 1.6 °C 1.7 °C 1.8 °C 1.9 °C 2.0 °C 2.1 °C 2.5 °C 2.6 °C 4.2 °C 4.4 °C 4.6 °C 4.7 °C	Transfer Standard Pyrometer/ SICT-CP-50205 SICT-CP-50205

502. non contact thermometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Blackbody furnaces	50206	(-40 ~ 0) °C (0 ~ 10) °C (10 ~ 50) °C (50 ~ 100) °C (100 ~ 200) °C (200 ~ 300) °C (300 ~ 400) °C (400 ~ 500) °C (500 ~ 600) °C (600 ~ 700) °C (700 ~ 1 100) °C (1 100 ~ 1 300) °C (1 300 ~ 1 500) °C (1 500 ~ 1 700) °C (1 700 ~ 1 800) °C (1 800 ~ 1 900) °C (1 900 ~ 2 000) °C (2 000 ~ 2 100) °C (2 100 ~ 2 200) °C (2 200 ~ 2 300) °C (2 300 ~ 2 400) °C	0.6 °C 0.4 °C 0.3 °C 0.4 °C 0.5 °C 0.7 °C 0.8 °C 0.9 °C 1.0 °C 1.1 °C 1.4 °C 1.5 °C 1.6 °C 1.8 °C 1.9 °C 2.0 °C 2.1 °C 3.7 °C 3.9 °C 4.1 °C 4.3 °C	Transfer Standard Pyrometer/ SICT-CP-50206
Others; ear thermometers, etc.	50207	(30 ~ 45) °C	0.07 °C	Standard prt/ SICT-CP-50207

503. Humidity

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Dew-point hygrometers; chilled mirror, alumina thin film, etc.	50301	(-90 ~ -80) °C D.P. (-80 ~ -70) °C D.P. (-70 ~ -50) °C D.P. (-50 ~ -20) °C D.P. (-20 ~ 90) °C D.P. (90 ~ 95) °C D.P.	0.60 °C D.P. 0.32 °C D.P. 0.20 °C D.P. 0.19 °C D.P. 0.13 °C D.P. 0.15 °C D.P.	Dewpoint Meter/ SICT-CP-50301	
Relative humidity hygrometers; polimer thin film, hair, etc.	50302	humidity Temperature	(3 ~ 60) % R.H. (60 ~ 90) % R.H. (90 ~ 98) % R.H. (-80 ~ 0) °C (0 ~ 80) °C (80 ~ 100) °C (100 ~ 180) °C	1.3 % R.H. 1.4 % R.H. 1.5 % R.H. 0.6 °C 0.3 °C 0.5 °C 1.5 °C	Dewpoint Meter/ SICT-CP-50302
Psychrometers; Assmann ventilated, PRT type, etc.	50303	assmann ventilated (humidity) (Temperature) PRT type (humidity) (Temperature)	(10 ~ 50) % R.H. (50 ~ 70) % R.H. (70 ~ 90) % R.H. (90 ~ 95) % R.H. (0 ~ 50) °C (10 ~ 50) % R.H. (50 ~ 80) % R.H. (80 ~ 98) % R.H. (0 ~ 80) °C (80 ~ 100) °C	1.3 % R.H. 1.4 % R.H. 1.5 % R.H. 1.6 % R.H. 0.3 °C 1.3 % R.H. 1.4 % R.H. 1.5 % R.H. 0.3 °C 0.5 °C	Dewpoint Meter/ SICT-CP-50303
Temperature humidity recorders; hygrothermograph, etc.	50304	Humidity Temperature	(5 ~ 70) % R.H. (70 ~ 95) % R.H. (-20 ~ 80) °C	2.1 % R.H. 2.2 % R.H. 0.7 °C	Dewpoint Meter/ SICT-CP-50304

503. Humidity

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Transducers; dew-point /relative humidity (Dew-point Transducers)	50305			Dewpoint Meter/ SICT-CP-50305
Dew point		(-90 ~ -80) °C D.P. (-80 ~ -70) °C D.P. (-70 ~ -60) °C D.P. (-60 ~ -40) °C D.P. (-40 ~ -20) °C D.P. (-20 ~ 0) °C D.P. (0 ~ 50) °C D.P. (50 ~ 90) °C D.P. (90 ~ 95) °C D.P.	0.60 °C D.P. 0.33 °C D.P. 0.22 °C D.P. 0.21 °C D.P. 0.20 °C D.P. 0.15 °C D.P. 0.14 °C D.P. 0.15 °C D.P. 0.17 °C D.P.	
(Relative humidity Transducers)				
Humidity		(3 ~ 50) % R.H. (50 ~ 80) % R.H. (80 ~ 98) % R.H.	1.3 % R.H. 1.4 % R.H. 1.5 % R.H.	
Temperature		(-80 ~ 0) °C (0 ~ 80) °C (80 ~ 100) °C (100 ~ 180) °C	0.7 °C 0.3 °C 0.5 °C 1.5 °C	
Humidity generators; two-pressure, two-temperature, flow mixing humidity generator, constant temperature and humidity chamber, etc.	50306			Dewpoint Meter/ SICT-CP-50306
Dew point		(-90 ~ -80) °C D.P. (-80 ~ -70) °C D.P. (-70 ~ -50) °C D.P. (-50 ~ -30) °C D.P. (-30 ~ -10) °C D.P. (-10 ~ 60) °C D.P. (60 ~ 80) °C D.P. (80 ~ 95) °C D.P.	0.60 °C D.P. 0.32 °C D.P. 0.19 °C D.P. 0.17 °C D.P. 0.16 °C D.P. 0.13 °C D.P. 0.14 °C D.P. 0.15 °C D.P.	
Humidity		(3 ~ 20) % R.H. (20 ~ 30) % R.H. (30 ~ 40) % R.H. (40 ~ 50) % R.H. (50 ~ 70) % R.H. (70 ~ 90) % R.H. (90 ~ 98) % R.H.	1.8 % R.H. 1.7 % R.H. 1.4 % R.H. 1.5 % R.H. 1.6 % R.H. 1.8 % R.H. 1.9 % R.H.	
Temperature		(-90 ~ 50) °C (50 ~ 100) °C (100 ~ 200) °C	0.4 °C 0.5 °C 0.6 °C	

504. Moisture

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Cereal moisture meters Moisture	50401	(9 ~ 20) % M.C.	0.7 % M.C.	Balance/ SICT-CP-50401
Wood moisture meters Moisture	50402	(8 ~ 25) % M.C.	2.5 % M.C.	Balance/ SICT-CP-50402
Paper moisture meters Moisture	50403	(8 ~ 20) % M.C.	3.4 % M.C.	Balance/ SICT-CP-50403

601. Sound in air

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Sound calibrators Pistonphones	60102	(200 ~ 300) Hz	0.08 dB	Reference microphone/ SICT-CP-60102
Sound Pressure Level Calibrators		(200 ~ 300) Hz	0.08 dB	
		(900 ~ 1 100) Hz	0.08 dB	
Multifunction Acoustic Calibrators		(28 ~ 35) Hz	0.12 dB	
		(35 ~ 90) Hz	0.09 dB	
		(90 ~ 4 500) Hz	0.08 dB	
		(4 500 ~ 9 000) Hz	0.09 dB	
		(9 000~ 14 000) Hz	0.13 dB	
		(14 000~ 17 000) Hz	0.21 dB	
Microphones	60104	20 Hz	0.15 dB	Reference microphone/ SICT-CP-60104
		(20 ~ 25) Hz	0.13 dB	
		(25 ~ 31.5) Hz	0.12 dB	
		(31.5 ~ 50) Hz	0.10 dB	
		(50 ~ 63) Hz	0.09 dB	
		(63 ~ 8 000) Hz	0.08 dB	
		(8 000 ~ 10 000) Hz	0.09 dB	
		(10 000 ~ 12 500) Hz	0.10 dB	
		(12 500 ~ 16 000) Hz	0.12 dB	
		(16 000 ~ 20 000) Hz	0.16 dB	
Sound level meters	60106	20 Hz	0.5 dB	Reference microphone/ SICT-CP-60106
		(20 ~ 50) Hz	0.4 dB	
		(50 ~ 160) Hz	0.3 dB	
		(160 ~ 2 000) Hz	0.2 dB	
		(2 000 ~ 8 000) Hz	0.3 dB	
		(8 000 ~ 16 000) Hz	0.4 dB	
		(16 000 ~ 20 000) Hz	0.5 dB	

603. Vibration

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Vibration calibrators	60301	(10 ~ 5 000) Hz	1.6×10^{-2}	Reference Accelerometer/ SICT-CP-60301
Vibration transducers	60302	(0.2 ~ 0.3) Hz (0.3 ~ 20) Hz (20 ~ 1 250) Hz (1 250 ~ 2 500) Hz (2 500 ~ 5 000) Hz (5 000 ~ 10 000) Hz (10 000 ~ 15 000) Hz (15 000 ~ 20 000) Hz	1.4×10^{-2} 1.3×10^{-2} 1.1×10^{-2} 1.2×10^{-2} 2.4×10^{-2} 2.7×10^{-2} 3.0×10^{-2} 3.8×10^{-2}	Reference Accelerometer(Shock)/ SICT-CP-60302
Shock transducers		at Pulse duration : (0.1 ~ 5) ms (200 ~ 2 000) m/s ² (2 000 ~ 20 000) m/s ² (20 000 ~ 100 000) m/s ²	1.0×10^{-2} 1.9×10^{-2} 3.3×10^{-2}	
Vibration measuring instruments	60303			Reference Accelerometer/ SICT-CP-60303
Acceleration		(10 ~ 20) Hz (20 ~ 1 250) Hz (1 250 ~ 2 500) Hz (2 500 ~ 5 000) Hz	1.6×10^{-2} 1.5×10^{-2} 1.6×10^{-2} 1.7×10^{-2}	
Velocity		(10 ~ 20) Hz (20 ~ 1 250) Hz (1 250 ~ 2 500) Hz	1.6×10^{-2} 1.5×10^{-2} 1.6×10^{-2}	
Displacement		(10 ~ 160) Hz (160 ~ 315) Hz (315 ~ 630) Hz	1.4×10^{-2} 2.1×10^{-2} 5.9×10^{-2}	

701. Photometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Illuminance meters Illuminance	70101	(0.5 ~ 20 000) lx	1.7×10^{-2}	Illuminance Meters/ SICT-CP-70101
Luminance meters Luminance	70102	1 cd/m ² (1 ~ 10) cd/m ² (10 ~ 3 000) cd/m ² (3 000 ~ 15 000) cd/m ²	2.1 × 10 ⁻² 1.7 × 10 ⁻² 1.4 × 10 ⁻² 1.6 × 10 ⁻²	Luminance Standard Sources/ SICT-CP-70102
Total luminous flux meters Total luminous flux	70103	70 lm (70 ~ 4 650) lm	3.2 × 10 ⁻² 1.5 × 10 ⁻²	Total Luminous Flux Standard Lamps/ SICT-CP-70103
Luminous intensity meters Luminance	70104	(72 ~ 3 200) cd	3.7×10^{-2}	Luminous Intensity Standard Lamps, Illuminance Meters / SICT-CP-70104

702. Property of detectors & sources

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color temperature meters Color temperature Chromaticity	70202	(2 677 ~ 3 333) K x y	25 K 0.004 0.004	Color Temperature Standard Lamps/ SICT-CP-70202
Color temperature standard lamps Color temperature Chromaticity	70203	(2 677 ~ 3 333) K x y	27 K 0.005 0.005	Spectroradiometers/ SICT-CP-70203
Colorimeters; source color Luminance Chromaticity	70204	1 cd/m ² (1 ~ 10) cd/m ² (10 ~ 3 000) cd/m ² (3 000 ~ 15 000) cd/m ² (WHITE) x y (RED) x y (GREEN) x y (BLUE) x y (CIE Standard Illuminant A) x y	2.1 × 10 ⁻² 1.7 × 10 ⁻² 1.4 × 10 ⁻² 1.6 × 10 ⁻² 0.004 0.004 0.004 0.004 0.004 0.004 0.004 0.004 0.003 0.003	Luminance Standard Sources/ SICT-CP-70204
Laser power meters	70207	(405 nm) (0.75 ~ 9) mW (660 nm) (0.7 ~ 47) mW (785 nm) (0.7 ~ 46) mW (1 080 nm) (1 ~ 40) W	1.2 × 10 ⁻² 1.2 × 10 ⁻² 1.2 × 10 ⁻² 3.3 × 10 ⁻²	Optical Power Meters/ SICT-CP-70207
Standard LED light sources Total luminous flux	70208	(68.4 ~ 72.6) lm	3.8 × 10 ⁻²	Total Spectral Radiant Flux Meters/ SICT-CP-70208
Total luminous flux standard lamps Total luminous flux	70209	(320 ~ 10 000) lm	4.7 × 10 ⁻²	Total Luminous Flux Standard lamps/ SICT-CP-70209

702. Property of detectors & sources

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Optical detectors Relative spectral responsivity	70210	(0 ~ 1) 300 nm (300 ~ 305) nm (305 ~ 310) nm (310 ~ 320) nm (320 ~ 370) nm (370 ~ 440) nm (440 ~ 525) nm (525 ~ 630) nm (630 ~ 955) nm (955 ~ 990) nm (990 ~ 1 100) nm	8.1 × 10 ⁻² 7.0 × 10 ⁻² 6.1 × 10 ⁻² 5.2 × 10 ⁻² 3.9 × 10 ⁻² 3.0 × 10 ⁻² 1.4 × 10 ⁻² 1.0 × 10 ⁻² 1.2 × 10 ⁻² 2.9 × 10 ⁻² 4.0 × 10 ⁻²	Photodiodes/ SICT-CP-70210
Pyranometers and pyrheliometers Irradiance responsivity	70211	(250 ~ 2 500) nm (1 000 ± 150) W/m ²	2.9 × 10 ⁻²	Standard pyranometers/ SICT-CP-70211
Display color analyzers; luminance, chromaticity, white balance, etc. Luminance Chromaticity	70213	1 cd/m ² (1 ~ 5) cd/m ² (5 ~ 200) cd/m ² (WHITE) x y (RED) x y (GREEN) x y (BLUE) x y	3.8 × 10 ⁻² 1.8 × 10 ⁻² 1.7 × 10 ⁻² 0.004 4 0.006 1 0.003 6 0.003 3 0.003 5 0.004 2 0.003 5 0.003 2	Luminance Meters/ SICT-CP-70213
Luminous intensity standard lamps Luminous intensity	70214	(10 ~ 20 000) cd	4.0 × 10 ⁻²	Spectroradiometers/ SICT-CP-70214
Spectral irradiance standard lamps Illuminance Spectral irradiance	70215	(792 ~ 7 105) lx (Deuterium arc lamp) 200 nm (200 ~ 205) nm (205 ~ 400) nm (Tungsten halogen lamp) 250 nm (250 ~ 270) nm (270 ~ 295) nm (295 ~ 375) nm (375 ~ 2 295) nm (2 295 ~ 2 345) nm (2 345 ~ 2 400) nm	2.8 × 10 ⁻² 5.4 × 10 ⁻² 5.3 × 10 ⁻² 4.9 × 10 ⁻² 5.5 × 10 ⁻² 5.2 × 10 ⁻² 5.0 × 10 ⁻² 4.4 × 10 ⁻² 3.8 × 10 ⁻² 4.0 × 10 ⁻² 4.6 × 10 ⁻²	Spectral Irradiance Standard Lamps/ SICT-CP-70215

702. Property of detectors & sources

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Total spectral radiant flux standard lamps	70216			Total Spectral Radiant Flux Standard Lamps/ SICT-CP-70216
Total spectral radiant flux		350 nm (350 ~ 365) nm (365 ~ 380) nm (380 ~ 400) nm (400 ~ 455) nm (455 ~ 850) nm	6.7 × 10 ⁻² 6.3 × 10 ⁻² 5.8 × 10 ⁻² 4.2 × 10 ⁻² 3.9 × 10 ⁻² 3.6 × 10 ⁻²	
Luminance standard sources	70217			Luminance Standard Sources/ SICT-CP-70217
Luminance		1 cd/m ² (1 ~ 10) cd/m ² (10 ~ 3 000) cd/m ² (3 000 ~ 15 000) cd/m ²	2.2 × 10 ⁻² 1.8 × 10 ⁻² 1.5 × 10 ⁻² 1.8 × 10 ⁻²	
Chromaticity		(WHITE) x y (RED) x y (GREEN) x y (BLUE) x y (CIE Standard Illuminant A) x y	0.005 0.005 0.005 0.005 0.005 0.005 0.005 0.005 0.004 0.004	
Spectral radiance standard sources	70218			Spectral Radiance Standard Sources/ SICT-CP-70218
Spectral radiance		300 nm (300 ~ 305) nm (305 ~ 310) nm (310 ~ 315) nm (315 ~ 320) nm (320 ~ 330) nm (330 ~ 340) nm (340 ~ 425) nm (425 ~ 470) nm (470 ~ 1 050) nm (1 050 ~ 1 600) nm	2.0 × 10 ⁻¹ 1.7 × 10 ⁻¹ 1.2 × 10 ⁻¹ 8.9 × 10 ⁻² 7.4 × 10 ⁻² 4.8 × 10 ⁻² 4.1 × 10 ⁻² 3.5 × 10 ⁻² 3.0 × 10 ⁻² 2.8 × 10 ⁻² 3.0 × 10 ⁻²	
UV irradiance meters	70219			UV Meter Standard Detectors/ SICT-CP-70219
Irradiance (UV Meter)		(254 nm) 50 µW/cm ² ~ 3 mW/cm ² (365 nm) 10 µW/cm ² ~ 230 mW/cm ² (405 nm) 10 µW/cm ² ~ 230 mW/cm ²	3.7 × 10 ⁻² 3.8 × 10 ⁻² 3.8 × 10 ⁻²	

702. Property of detectors & sources

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Spectral irradiance meters	70220			Spectral Irradiance Standard
Wavelength		(250 ~ 2 030) nm	0.3 nm	Lamps/ SICT-CP-70220
Illuminance		(813 ~ 6 879) lx	2.1×10^{-2}	
Spectral Irradiance		200 nm	4.9×10^{-2}	
		(200 ~ 245) nm	4.8×10^{-2}	
		(245 ~ 345) nm	4.4×10^{-2}	
		(345 ~ 375) nm	3.8×10^{-2}	
		(375 ~ 895) nm	3.4×10^{-2}	
		(895 ~ 1 050) nm	3.0×10^{-2}	
		(1 050 ~ 2 295) nm	3.4×10^{-2}	
		(2 295 ~ 2 345) nm	3.7×10^{-2}	
		(2 345 ~ 2 400) nm	4.3×10^{-2}	
Total spectral radiant flux meters	70221			Total Spectral Radiant Flux Standard Lamps/ SICT-CP-70221
Wavelength		(350 ~ 850) nm	0.25 nm	
Total spectral radiant flux		350 nm	2.0×10^{-2}	
		(350 ~ 365) nm	1.8×10^{-2}	
		(365 ~ 375) nm	1.7×10^{-2}	
		(375 ~ 390) nm	1.6×10^{-2}	
		(390 ~ 445) nm	1.5×10^{-2}	
		(445 ~ 850) nm	1.4×10^{-2}	
Spectral radiance meters	70222			Spectral Radiance Standard Sources/ SICT-CP-70222
Wavelength		(350 ~ 1 694) nm	0.25 nm	
Spectral radiance		300 nm	2.0×10^{-1}	
		(300 ~ 305) nm	1.7×10^{-1}	
		(305 ~ 310) nm	1.2×10^{-1}	
		(310 ~ 315) nm	8.8×10^{-2}	
		(315 ~ 320) nm	7.2×10^{-2}	
		(320 ~ 325) nm	5.5×10^{-2}	
		(325 ~ 335) nm	4.6×10^{-2}	
		(335 ~ 345) nm	3.7×10^{-2}	
		(345 ~ 405) nm	3.5×10^{-2}	
		(405 ~ 455) nm	3.0×10^{-2}	
		(455 ~ 755) nm	2.6×10^{-2}	
		(755 ~ 1 400) nm	2.7×10^{-2}	
		(1 400 ~ 1 525) nm	3.0×10^{-2}	
		(1 525 ~ 1 600) nm	2.8×10^{-2}	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color	70301			Color Standard Tiles/ SICT-CP-70301
Included Reflectance Std. Light Source T_{rec} A(2°, 10°), C(2°, 10°), D65(2°, 10°)				
1. White	X Y Z		1.0×10^{-2} 1.0×10^{-2} 1.0×10^{-2}	
2. I,Gray	X Y Z		1.0×10^{-2} 1.0×10^{-2} 1.0×10^{-2}	
3. m,Gray	X Y Z		1.0×10^{-2} 1.0×10^{-2} 1.0×10^{-2}	
4. d,Gray	X Y Z		1.1×10^{-2} 1.0×10^{-2} 1.0×10^{-2}	
5. Red	X Y Z		1.1×10^{-2} 1.2×10^{-2} 1.8×10^{-2}	
6. Yellow	X Y Z		1.0×10^{-2} 1.0×10^{-2} 1.6×10^{-2}	
7. Green	X Y Z		1.0×10^{-2} 1.0×10^{-2} 1.1×10^{-2}	
8. Cyan	X Y Z		1.0×10^{-2} 1.0×10^{-2} 1.0×10^{-2}	
Included Reflectance Std. Light Source T_{rec} A(2°, 10°), C(2°, 10°), D65(2°, 10°)				
1. White	L* a* b*		0.36 0.09 0.07	
2. I,Gray	L* a* b*		0.32 0.08 0.07	
3. m,Gray	L* a* b*		0.24 0.06 0.06	
4. d,Gray	L* a* b*		0.17 0.04 0.04	
5. Red	L* a* b*		0.25 0.27 0.26	
6. Yellow	L* a* b*		0.34 0.17 0.43	
7. Green	L* a* b*		0.24 0.13 0.12	
8. Cyan	L* a* b*		0.24 0.14 0.15	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color	70301			Color Standard Tiles/ SICT-CP-70301
Included Reflectance Std. Light Source T_{vma} A(2°, 10°), C(2°, 10°), D65(2°, 10°)				
1. White		x	1.3×10^{-3}	
		y	1.3×10^{-3}	
2. I,Gray		x	1.3×10^{-3}	
		y	1.3×10^{-3}	
3. m,Gray		x	1.4×10^{-3}	
		y	1.3×10^{-3}	
4. d,Gray		x	1.4×10^{-3}	
		y	1.3×10^{-3}	
5. Red		x	2.7×10^{-3}	
		y	1.3×10^{-3}	
6. Yellow		x	1.2×10^{-3}	
		y	1.3×10^{-3}	
7. Green		x	1.4×10^{-3}	
		y	1.0×10^{-3}	
8. Cyan		x	2.0×10^{-3}	
		y	1.7×10^{-3}	
Exclude Reflectance Std. Light Source T_{vma} A(2°, 10°), C(2°, 10°), D65(2°, 10°)				
1. White		X	1.0×10^{-2}	
		Y	1.0×10^{-2}	
		Z	1.0×10^{-2}	
2. I,Gray		X	1.0×10^{-2}	
		Y	1.0×10^{-2}	
		Z	1.0×10^{-2}	
3. m,Gray		X	1.0×10^{-2}	
		Y	1.0×10^{-2}	
		Z	1.0×10^{-2}	
4. d,Gray		X	1.1×10^{-2}	
		Y	1.0×10^{-2}	
		Z	1.4×10^{-2}	
5. Red		X	1.2×10^{-2}	
		Y	1.3×10^{-2}	
		Z	3.1×10^{-2}	
6. Yellow		X	1.0×10^{-2}	
		Y	1.0×10^{-2}	
		Z	2.3×10^{-2}	
7. Green		X	1.0×10^{-2}	
		Y	1.0×10^{-2}	
		Z	1.2×10^{-2}	
8. Cyan		X	1.0×10^{-2}	
		Y	1.0×10^{-2}	
		Z	1.0×10^{-2}	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Exclude Reflectance Std. Light Source T_{vma} A(2°, 10°), C(2°, 10°), D65(2°, 10°)	70301			Color Standard Tiles/ SICT-CP-70301
1. White		L*	0.36	
		a*	0.09	
		b*	0.07	
2. I,Gray		L*	0.31	
		a*	0.08	
		b*	0.07	
3. m,Gray		L*	0.23	
		a*	0.06	
		b*	0.05	
4. d,Gray		L*	0.15	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.25	
		a*	0.32	
		b*	0.66	
6. Yellow		L*	0.33	
		a*	0.18	
		b*	0.58	
7. Green		L*	0.23	
		a*	0.14	
		b*	0.14	
8. Cyan		L*	0.23	
		a*	0.16	
		b*	0.16	
Exclude Reflectance Std. Light Source T_{vma} A(2°, 10°), C(2°, 10°), D65(2°, 10°)				
1. White	x		1.3×10^{-3}	
	y		1.3×10^{-3}	
2. I,Gray	x		1.3×10^{-3}	
	y		1.3×10^{-3}	
3. m,Gray	x		1.4×10^{-3}	
	y		1.3×10^{-3}	
4. d,Gray	x		1.3×10^{-3}	
	y		1.3×10^{-3}	
5. Red	x		5.3×10^{-3}	
	y		1.3×10^{-3}	
6. Yellow	x		1.3×10^{-3}	
	y		1.5×10^{-3}	
7. Green	x		1.5×10^{-3}	
	y		1.0×10^{-3}	
8. Cyan	x		2.1×10^{-3}	
	y		1.8×10^{-3}	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color	70301			Color Standard Tiles/ SICT-CP-70301
Transmittance Std. Light Source Type A(2°, 10°), C(2°, 10°), D65(2°, 10°)				
1. E-LA200		X	4.0 × 10 ⁻³	
		Y	4.0 × 10 ⁻³	
		Z	9.0 × 10 ⁻³	
2. G533		X	7.0 × 10 ⁻³	
		Y	5.0 × 10 ⁻³	
		Z	1.3 × 10 ⁻²	
3. B460		X	5.0 × 10 ⁻³	
		Y	4.0 × 10 ⁻³	
		Z	3.0 × 10 ⁻³	
4. ND 25		X	3.0 × 10 ⁻³	
		Y	3.0 × 10 ⁻³	
		Z	5.0 × 10 ⁻³	
5. ND 40		X	3.0 × 10 ⁻³	
		Y	3.0 × 10 ⁻³	
		Z	3.0 × 10 ⁻³	
6. ND 50		X	3.0 × 10 ⁻³	
		Y	3.0 × 10 ⁻³	
		Z	3.0 × 10 ⁻³	
7. ND 70		X	3.0 × 10 ⁻³	
		Y	2.0 × 10 ⁻³	
		Z	3.0 × 10 ⁻³	
Transmittance Std. Light Source Type A(2°, 10°), C(2°, 10°), D65(2°, 10°)				
1. E-LA200		L*	0.11	
		a*	0.09	
		b*	0.22	
2. G533		L*	0.14	
		a*	0.12	
		b*	0.19	
3. B460		L*	0.12	
		a*	0.08	
		b*	0.16	
4. ND 25		L*	0.07	
		a*	0.04	
		b*	0.06	
5. ND 40		L*	0.07	
		a*	0.04	
		b*	0.05	
6. ND 50		L*	0.08	
		a*	0.05	
		b*	0.05	
7. ND 70		L*	0.09	
		a*	0.05	
		b*	0.05	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Colorimeters; material color Transmittance Std. Light Source Type A(2°, 10°), C(2°, 10°), D65(2°, 10°)	70301			Color Standard Tiles/ SICT-CP-70301
1. E-LA200		x	1.0 × 10 ⁻³	
		y	1.0 × 10 ⁻³	
2. G533		x	1.4 × 10 ⁻³	
		y	1.6 × 10 ⁻³	
3. B460		x	2.0 × 10 ⁻³	
		y	1.6 × 10 ⁻³	
4. ND 25		x	1.3 × 10 ⁻³	
		y	1.2 × 10 ⁻³	
5. ND 40		x	1.3 × 10 ⁻³	
		y	1.3 × 10 ⁻³	
6. ND 50		x	1.3 × 10 ⁻³	
		y	1.3 × 10 ⁻³	
7. ND 70		x	1.3 × 10 ⁻³	
		y	1.3 × 10 ⁻³	
Color standard tiles Included Reflectance Std. Light Source A(2°, 10°), C(2°, 10°), D65(2°, 10°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White		X	0.95	
		Y	0.86	
		Z	0.99	
2. I,Gray		X	0.66	
		Y	0.60	
		Z	0.70	
3. m,Gray		X	0.29	
		Y	0.26	
		Z	0.31	
4. d,Gray		X	0.11	
		Y	0.10	
		Z	0.11	
5. Red		X	0.36	
		Y	0.22	
		Z	0.14	
6. Yellow		X	0.78	
		Y	0.68	
		Z	0.21	
7. Green		X	0.19	
		Y	0.23	
		Z	0.20	
8. Cyan		X	0.20	
		Y	0.24	
		Z	0.49	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles	70304			Color Standard Tiles/ SICT-CP-70304
Included Reflectance Std. Light Source T_{rms} A(2°, 10°), C(2°, 10°), D65(2°, 10°)				
1. White		L*	0.37	
		a*	0.09	
		b*	0.07	
2. I,Gray		L*	0.32	
		a*	0.08	
		b*	0.07	
3. m,Gray		L*	0.24	
		a*	0.06	
		b*	0.06	
4. d,Gray		L*	0.17	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.25	
		a*	0.27	
		b*	0.26	
6. Yellow		L*	0.34	
		a*	0.17	
		b*	0.43	
7. Green		L*	0.24	
		a*	0.13	
		b*	0.12	
8. Cyan		L*	0.24	
		a*	0.14	
		b*	0.15	
Included Reflectance Std. Light Source T_{rms} A(2°, 10°), C(2°, 10°), D65(2°, 10°)				
1. White		x	0.000 7	
		y	0.000 6	
2. I,Gray		x	0.000 7	
		y	0.000 6	
3. m,Gray		x	0.000 7	
		y	0.000 6	
4. d,Gray		x	0.000 7	
		y	0.000 6	
5. Red		x	0.001 4	
		y	0.000 6	
6. Yellow		x	0.000 7	
		y	0.000 8	
7. Green		x	0.000 6	
		y	0.000 7	
8. Cyan		x	0.000 6	
		y	0.000 6	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles	70304			Color Standard Tiles/ SICT-CP-70304
Exclude Reflectance Std. Light Source T_{rec} A(2°, 10°), C(2°, 10°), D65(2°, 10°)				
1. White		X	0.91	
		Y	0.82	
		Z	0.95	
2. I,Gray		X	0.62	
		Y	0.56	
		Z	0.66	
3. m,Gray		X	0.25	
		Y	0.23	
		Z	0.26	
4. d,Gray		X	0.07	
		Y	0.06	
		Z	0.07	
5. Red		X	0.32	
		Y	0.18	
		Z	0.12	
6. Yellow		X	0.74	
		Y	0.64	
		Z	0.19	
7. Green		X	0.15	
		Y	0.20	
		Z	0.16	
8. Cyan		X	0.17	
		Y	0.21	
		Z	0.45	
Exclude Reflectance Std. Light Source T_{rec} A(2°, 10°), C(2°, 10°), D65(2°, 10°)				
1. White		L*	0.36	
		a*	0.09	
		b*	0.07	
2. I,Gray		L*	0.31	
		a*	0.08	
		b*	0.07	
3. m,Gray		L*	0.23	
		a*	0.06	
		b*	0.05	
4. d,Gray		L*	0.15	
		a*	0.04	
		b*	0.04	
5. Red		L*	0.25	
		a*	0.32	
		b*	0.66	
6. Yellow		L*	0.33	
		a*	0.18	
		b*	0.58	
7. Green		L*	0.23	
		a*	0.14	
		b*	0.14	
8. Cyan		L*	0.23	
		a*	0.16	
		b*	0.16	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Color standard tiles Exclude Reflectance Std. Light Source T_{rma} A(2°, 10°), C(2°, 10°), D65(2°, 10°)	70304			Color Standard Tiles/ SICT-CP-70304
1. White x y			0.000 6 0.000 6	
2. I,Gray x y			0.000 6 0.000 6	
3. m,Gray x y			0.000 6 0.000 6	
4. d,Gray x y			0.000 6 0.000 6	
5. Red x y			0.002 9 0.000 6	
6. Yellow x y			0.000 8 0.000 9	
7. Green x y			0.000 6 0.000 7	
8. Cyan x y			0.000 5 0.000 5	
Absolute Spectral Reflectance White Plate (Include, Exclude Reflectance)		360 nm (360 ~ 830) nm	0.009 4 0.012	
Gloss meters Gloss	70306	20 ° 60 ° 85 °	8.9×10^{-3} 9.5×10^{-3} 8.0×10^{-3}	Gloss Standard/ SICT-CP-70306
Gloss standard plates Gloss	70307	20° 60° 85°	9.5×10^{-3} 9.8×10^{-3} 8.3×10^{-3}	Gloss Meter/ SICT-CP-70307
Haze meters Haze	70308	H-1 H-5 H-10 H-20 H-30	0.30 0.26 0.4 0.6 0.8	Haze Standard Plate, Transmittance Standard Plates/ SICT-CP-70308
		T-30 T-50 T-70 T-90	0.50 0.50 0.50 0.50	
Lens meters Vertex diopter	70312	-25 D ~ 25 D	0.03 D	Reference Lens/ SICT-CP-70312
Optical densitometers Density	70315	1 Step ~ 10 Step 11 Step 11 Step ~ 15 Step	0.03 0.06 0.11	Density CRM/ SICT-CP-70315

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Reflectance meters	70319			Absolute Spectral Reflectance White Standard Plates/ SICT-CP-70319
Reflectance		380 nm ~ 780 nm	1.1×10^{-2}	
Refractometers	70321			Reference Refracto CRM/ SICT-CP-70321
Refracto		(1.332 99 ~ 1.505 80) nD 1.51 nD 1.62 nD	0.000 04 nD 0.000 2 nD 0.000 2 nD	
Transmittance meters	70323			Transmittance Filter/ SICT-CP-70323
		(0.1) (250 ~ 750) nm	6.1×10^{-3}	
		(0.5) (250 ~ 750) nm	3.8×10^{-3}	
		(0.9) (250 ~ 750) nm	2.2×10^{-3}	
Spectrophotometers including FT-IR spectrophotometers	70325			Wavelength Filters, Transmittance Filters, Absolute Spectral Reflectance White Standard Plates, Wavenumber Filters/ SICT-CP-70325
Wavelength		(240.7 ~ 976.1) nm (990.2 ~ 2 536.5) nm	0.4 nm 0.5 nm	
Transmittance		(Filter N0 10) 250 nm 300 nm 350 nm 400 nm 450 nm 500 nm 550 nm 600 nm 650 nm 700 nm 750 nm 900 nm 1 000 nm 1 200 nm 1 400 nm 1 600 nm 1 800 nm 2 000 nm 2 200 nm 2 400 nm 2 500 nm	8.7×10^{-3} 9.2×10^{-3} 8.2×10^{-3} 6.6×10^{-3} 6.8×10^{-3} 6.5×10^{-3} 6.6×10^{-3} 6.8×10^{-3} 6.4×10^{-3} 6.7×10^{-3} 6.6×10^{-3} 8.0×10^{-3} 8.1×10^{-3} 8.4×10^{-3} 7.7×10^{-3} 8.0×10^{-3} 8.2×10^{-3} 8.1×10^{-3} 8.6×10^{-3} 9.2×10^{-3} 1.7×10^{-2}	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Spectrophotometers including FT-IR spectrophotometers	70325			Wavelength Filters, Transmittance Filters, Absolute Spectral Reflectance White Standard Plates, Wavenumber Filters/ SICT-CP-70325

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Spectrophotometers including FT-IR spectrophotometers	70325			Wavelength Filters, Transmittance Filters, Absolute Spectral Reflectance White Standard Plates, Wavenumber Filters/ SICT-CP-70325
Transmittance		(Filter NO 1) 440 nm 465 nm 546 nm 590 nm 635 nm	1.4×10^{-2} 9.1×10^{-3} 9.4×10^{-3} 1.1×10^{-2} 1.0×10^{-2}	
		(Filter NO 3) 440 nm 465 nm 546 nm 590 nm 635 nm	7.9×10^{-3} 5.8×10^{-3} 6.1×10^{-3} 6.3×10^{-3} 6.2×10^{-3}	
Absorbance		(Filter NO 10) 250 nm 300 nm 350 nm 400 nm 450 nm 500 nm 550 nm 600 nm 650 nm 700 nm 750 nm 900 nm 1 000 nm 1 200 nm 1 400 nm 1 600 nm 1 800 nm 2 000 nm 2 200 nm 2 400 nm 2 500 nm	0.003 7 0.003 8 0.003 5 0.002 7 0.002 6 0.002 5 0.002 5 0.002 5 0.002 8 0.002 6 0.002 4 0.003 3 0.003 3 0.003 3 0.003 2 0.003 3 0.003 3 0.003 4 0.003 7 0.007 2	
		(Filter NO 30, 40, 50) 250 nm 300 nm 350 nm 400 nm 450 nm 500 nm 550 nm 600 nm 650 nm 700 nm 750 nm	0.003 8 0.003 7 0.003 6 0.002 5 0.002 5 0.002 6 0.002 6 0.002 6 0.002 5 0.002 5 0.002 5	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Spectrophotometers including FT-IR spectrophotometers	70325			Wavelength Filters, Transmittance Filters, Absolute Spectral Reflectance White Standard Plates, Wavenumber Filters/ SICT-CP-70325
Absorbance		(Filter NO 30)		
		900 nm	0.003 5	
		1 000 nm	0.003 4	
		1 200 nm	0.003 2	
		1 400 nm	0.003 3	
		1 600 nm	0.003 4	
		1 800 nm	0.003 2	
		2 000 nm	0.003 3	
		2 200 nm	0.003 5	
		2 400 nm	0.003 3	
		2 500 nm	0.003 9	
		(Filter NO 90)		
		250 nm	0.003 6	
		300 nm	0.003 6	
		350 nm	0.003 5	
		400 nm	0.002 5	
		450 nm	0.002 5	
		500 nm	0.002 5	
		550 nm	0.002 6	
		600 nm	0.002 6	
		650 nm	0.002 6	
		700 nm	0.002 5	
		750 nm	0.002 6	
		900 nm	0.003 3	
		1 000 nm	0.003 3	
		1 200 nm	0.003 3	
		1 400 nm	0.003 3	
		1 600 nm	0.003 2	
		1 800 nm	0.003 4	
		2 000 nm	0.003 3	
		2 200 nm	0.003 3	
		2 400 nm	0.003 3	
		2 500 nm	0.003 6	
		(Filter NO 1)		
		440 nm	0.004 1	
		465 nm	0.002 8	
		546 nm	0.002 8	
		590 nm	0.003 0	
		635 nm	0.003 5	
		(Filter NO 3)		
		440 nm	0.003 2	
		465 nm	0.002 4	
		546 nm	0.002 6	
		590 nm	0.002 5	
		635 nm	0.002 7	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Spectrophotometers including FT-IR spectrophotometers	70325			Wavelength Filters, Transmittance Filters, Absolute Spectral Reflectance White Standard Plates, Wavenumber Filters/ SICT-CP-70325
Spectral Reflectance		250 nm (250 ~ 2 500) nm	1.5×10^{-2} 1.5×10^{-2}	
Wavenumber		544.9 cm ⁻¹ 842.1 cm ⁻¹ 906.82 cm ⁻¹ 1 028.42 cm ⁻¹ 1 069.27 cm ⁻¹ 1 154.62 cm ⁻¹ 1 583.04 cm ⁻¹ 1 601.38 cm ⁻¹ 2 850.20 cm ⁻¹ 3 001.40 cm ⁻¹ 3 026.44 cm ⁻¹ 3 060.14 cm ⁻¹ 3 082.22 cm ⁻¹	2.5 cm ⁻¹ 1.3 cm ⁻¹ 0.12 cm ⁻¹ 0.28 cm ⁻¹ 0.78 cm ⁻¹ 0.12 cm ⁻¹ 0.12 cm ⁻¹ 0.13 cm ⁻¹ 0.14 cm ⁻¹ 0.12 cm ⁻¹ 0.12 cm ⁻¹ 0.12 cm ⁻¹ 0.12 cm ⁻¹	
Wavelength reference materials; absorption cell, bandpass filter, etc.	70326			Spectrophotometers, Absolute Spectral Reflectance White Standard Plates/ SICT-CP-70326
Wavelength		(240 ~750) nm	0.5 nm	
Transmittance		(0.1 ~ 0.3) 250 nm 300 nm 350 nm 400 nm 450 nm 500 nm 550 nm 600 nm 650 nm 700 nm 750 nm (0.3 ~ 0.5) 250 nm 300 nm 350 nm 400 nm 450 nm 500 nm 550 nm 600 nm 650 nm 700 nm 750 nm	8.5×10^{-3} 8.1×10^{-3} 8.1×10^{-3} 5.9×10^{-3} 5.7×10^{-3} 8.3×10^{-3} 8.1×10^{-3} 8.0×10^{-3} 5.7×10^{-3} 5.7×10^{-3} 5.7×10^{-3} 5.7×10^{-3} 5.7×10^{-3} 5.7×10^{-3}	

703. Property of materials

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Wavelength reference materials; absorption cell, bandpass filter, etc.	70326			Spectrophotometers, Absolute Spectral Reflectance White Standard Plates/ SICT-CP-70326
Transmittance		(0.5 ~ 0.9) 250 nm 300 nm 350 nm 400 nm 450 nm 500 nm 550 nm 600 nm 650 nm 700 nm 750 nm	8.2×10^{-3} 8.1×10^{-3} 8.0×10^{-3} 5.7×10^{-3} 5.7×10^{-3} 5.7×10^{-3} 5.7×10^{-3} 5.6×10^{-3} 5.6×10^{-3} 5.7×10^{-3} 5.9×10^{-3}	
Absorbance		(0.1 ~ 0.3) 250 nm 300 nm 350 nm 400 nm 450 nm 500 nm 550 nm 600 nm 650 nm 700 nm 750 nm	0.003 7 0.003 7 0.003 8 0.002 7 0.002 7 0.002 8 0.002 8 0.002 7 0.002 7 0.002 7 0.002 8	
		(0.3 ~ 0.5) 250 nm 300 nm 350 nm 400 nm 450 nm 500 nm 550 nm 600 nm 650 nm 700 nm 750 nm	0.003 6 0.003 6 0.003 6 0.002 4 0.002 4 0.002 4 0.002 4 0.002 4 0.002 4 0.002 4 0.002 5	
		(0.5 ~ 0.9) 250 nm 300 nm 350 nm 400 nm 450 nm 500 nm 550 nm 600 nm 650 nm 700 nm 750 nm	0.003 6 0.003 5 0.003 5 0.002 4 0.002 4 0.002 4 0.002 4 0.002 4 0.002 4 0.002 4 0.002 4	
Reflectance		(360 ~ 830) nm	1.0×10^{-2}	

704. Fiber optics

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Broadband light sources Wavelength output	70402	1 310 nm, 1 550 nm	0.058 nm	Optical spectrum analyzer, Optical powermeter/ SICT-CP-70402
Optical power output		1 310 nm, 1 550 nm (0 ~ -60) dBm	0.070 dB	
Optical attenuators Optical Attenuation	70410	1 310 nm, 1 550 nm (-60 ~ 0) dB	0.08 dB	Optical powermeter, Optical power stabilized lasers and LDs/ SICT-CP-70410
Absolute optical power		1 310 nm, 1 550 nm (0 ~ -60) dBm	0.072 dB	
Fiber-optic power meters Optical Linearity	70412	1 310 nm, 1 550 nm (0 ~ -60) dB	0.03 dB	Optical powermeter, Optical power stabilized lasers and LDs, Optical attenuator/ SICT-CP-70412
Optical loss testers Optical Attenuation		1 310 nm, 1 550 nm (0 ~ -60) dB	0.03 dB	
Optical multimeters Absolute optical power measure	70415	1 310 nm, 1 550 nm (0 ~ -60) dBm	0.072 dB	Optical powermeter, Optical power stabilized lasers and LDs, Optical attenuator/ SICT-CP-70415
Linearity measure		1 310 nm, 1 550 nm (0 ~ -60) dB	0.03 dB	
Optical network analyzer (Optical multimeter) Absolute optical power (Optical spectrum analyzer) Wavelength measure Resolution measure Absolute optical power measure (Optical attenuator) Optical Attenuation Return loss (Optical time domain reflectometer) Wavelength output Optical Length measure	70416	1 310 nm, 1 550 nm (-60 ~ 0) dBm	0.072 dB	Optical powermeter, OTDR, Fiber reference, Wavelength meter Optical spectrum analyzer Optical attenuator Optical Returnloss generator/ SICT-CP-70416
1 310 nm		0.058 nm	0.058 nm	
1 550 nm		0.058 nm	0.058 nm	
Resolution: (0.1 ~ 1) nm		0.058 nm	0.058 nm	
1 310 nm		0.058 nm	0.058 nm	
1 550 nm		0.058 nm	0.058 nm	
1 310 nm, 1 550 nm (-60 ~ 0) dBm		0.072 dB	0.072 dB	
1 310 nm, 1 550 nm (-60 ~ 0) dB		0.07 dB	0.07 dB	
1 310 nm, 1 550 nm (20 ~ 40) dB		0.8 dB	0.8 dB	
1 310 nm, 1 550 nm		0.082 nm	0.082 nm	
1 310 nm		0.081 m	0.081 m	
3.3 km Fiber		0.34 m	0.34 m	
13.4 km Fiber		0.080 m	0.080 m	
1 550 nm		0.34 m	0.34 m	
3.3 km Fiber		0.080 m	0.080 m	
13.4 km Fiber		0.34 m	0.34 m	

704. Fiber optics

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Optical network analyzer Optical loss measure	70416	1 310 nm 7.20 dB Fiber 2.90 dB Fiber 1 550 nm 4.20 dB Fiber 1.60 dB Fiber	0.13 dB 0.05 dB 0.05 dB 0.05 dB	Optical powermeter, OTDR, Fiber reference, Wavelength meter Optical spectrum analyzer Optical attenuator Optical Returnloss generator/ SICT-CP-70416
Optical spectrum analyzers Wavelength measure	70417	1 310 nm 1 550 nm Resolution : (0.1 ~ 1) nm 1 310 nm 1 550 nm	0.058 nm 0.058 nm 0.058 nm 0.058 nm	Optical powermeter, Optical power stabilized lasers and LDs, Optical attenuator, Optical spectrum analyzer/ SICT-CP-70417
Absolute optical power measure		1 310 nm, 1 550 nm (-60 ~ 0) dBm	0.072 dB	
Linearity measure		1 310 nm, 1 550 nm (-60 ~ 0) dB	0.03 dB	
Optical time domain reflectometers; OTDR Wavelength output	70418	1 310 nm, 1 550 nm	0.08 nm	Optical length fiber reference, Optical fiberloss reference, Optical spectrum analyzer/ SICT-CP-70418
Optical Length measure		1 310 nm 3.3 km Fiber 13.4 km Fiber 1 550 nm 3.3 km Fiber 13.4 km Fiber	0.081 m 0.34 m 0.080 m 0.34 m	
Optical loss measure		1 310 nm 7.20 dB Fiber 2.90 dB Fiber 1 550 nm 4.20 dB Fiber 1.60 dB Fiber	0.13 dB 0.05 dB 0.05 dB 0.05 dB	
Return loss meters Return loss measure	70423	1 310 nm, 1 550 nm 20 dB ~ 40 dB	0.8 dB	Optical Returnloss generator SICT-CP-70423
Frequency stabilized lasers and LDs Wavelength optical power	70429	1 310 nm 1 550 nm 1 310 nm, 1 550 nm (-60 ~ 0) dBm	4 pm 4 pm 0.07 dB	Wavelength meter, Optical powermeter/ SICT-CP-70429

704. Fiber optics

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
ASE light sources Wavelength output	70430	1 310 nm, 1 550 nm	0.058 nm	Optical spectrum analyzer, Optical powermeter/ SICT-CP-70430
Optical power output		1 310 nm, 1 550 nm (-60 ~ 0) dBm	0.07 dB	
Optical power stabilized lasers and LDs Wavelength output	70433	1 310 nm	4 pm	Wavelength meter, Optical powermeter/ SICT-CP-70433
		1 550 nm	4 pm	
Optical power output		1 310 nm, 1 550 nm (-60 ~ 0) dBm	0.07 dB	

901. Chemical analysis

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Breath alcohol analyzers	90101	(0.000 ~ 0.080) %BAC	3.3×10^{-2}	Standard gas/ SICT-CP-90101
		(0.080 ~ 0.190) %BAC	2.1×10^{-2}	
	90101	(0.000 ~ 0.080) %BAC	2.9×10^{-2}	
		(0.080 ~ 0.150) %BAC (0.150 ~ 0.400) %BAC	1.6×10^{-2} 1.3×10^{-2}	
Environmental air quality monitoring instruments	90102	(0 ~ 700) $\mu\text{mol/mol}$	1.0×10^{-2}	Standard gas/ SICT-CP-90102
		(0.07 ~ 1.5) cmol/mol	2.0×10^{-2}	
		(1.5 ~ 20) cmol/mol	1.2×10^{-2}	
		(0 ~ 850) $\mu\text{mol/mol}$	2.2×10^{-2}	
		Carbon monoxide	2.0×10^{-2}	
		Carbon dioxide	1.5×10^{-2}	
		(0 ~ 0.50) cmol/mol	2.1×10^{-2}	
		(0.50 ~ 5.00) cmol/mol	2.0×10^{-2}	
		(5.00 ~ 19.00) cmol/mol	1.4×10^{-2}	
		Nitrogen monoxide	2.1×10^{-2}	
		Isobutane	2.2×10^{-2}	
		Methane	1.4×10^{-2}	
		Hydrogen sulfide	3.6×10^{-2}	
		Propane	3.0×10^{-2}	
		Isobutylene	1.0×10^{-2}	
		Ammonia	4.9×10^{-2}	
		Sulfur dioxide	2.2×10^{-2}	
Gas analyzers	90103	Nitrogen dioxide	1.0×10^{-2}	Standard gas/ SICT-CP-90103
		Hydrogen	2.3×10^{-2}	
		(0 ~ 500) $\mu\text{mol/mol}$	2.1×10^{-2}	
		(0.05 ~ 2.0) cmol/mol	2.1×10^{-2}	
		Hydrogen chloride	4.8×10^{-2}	
		Sulfur hexafluoride	0.1×10^{-2}	
		Ozone	2.2 nmol/mol	
		(0.0 ~ 1 000.0) nmol/mol	2.5×10^{-2}	
		Oxygen	1.0×10^{-2}	
		(0 ~ 700) $\mu\text{mol/mol}$	2.0×10^{-2}	
		(0.07 ~ 1.5) cmol/mol	1.2×10^{-2}	
		(1.5 ~ 20) cmol/mol	2.2×10^{-2}	
		Carbon monoxide	2.0×10^{-2}	
		Carbon dioxide	1.5×10^{-2}	
		(0 ~ 0.50) cmol/mol	2.1×10^{-2}	
		(0.50 ~ 5.00) cmol/mol	2.0×10^{-2}	
		(5.00 ~ 19.00) cmol/mol	1.4×10^{-2}	
		Nitrogen monoxide	2.1×10^{-2}	
		Isobutane	2.2×10^{-2}	
		Methane	1.4×10^{-2}	
		Hydrogen sulfide	3.6×10^{-2}	
		Propane	3.0×10^{-2}	
		Isobutylene	1.0×10^{-2}	
		Ammonia	4.9×10^{-2}	
		Sulfur dioxide	2.2×10^{-2}	
		Nitrogen dioxide	1.0×10^{-2}	
		Hydrogen	2.3×10^{-2}	
		(0 ~ 500) $\mu\text{mol/mol}$	2.1×10^{-2}	
		(0.05 ~ 2.0) cmol/mol	2.1×10^{-2}	
		Hydrogen chloride	4.8×10^{-2}	
		Sulfur hexafluoride	0.1×10^{-2}	
		Ozone	2.2 nmol/mol	
		(0.0 ~ 1 000.0) nmol/mol	2.5×10^{-2}	

901. Chemical analysis

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Exhaust gas test instruments	90104	(0 ~ 1.5) $\mu\text{mol/mol}$	2.0×10^{-2}	Standard gas/ SICT-CP-90103
		(1.5 ~ 20) cmol/mol	1.1×10^{-2}	
		(0 ~ 5.0) cmol/mol	2.1×10^{-2}	
		(0 ~ 19) cmol/mol	2.0×10^{-2}	
		(0 ~ 2 000) $\mu\text{mol/mol}$	2.0×10^{-2}	
		(0 ~ 0.8) cmol/mol	2.2×10^{-2}	
		(0 ~ 2.0) cmol/mol	1.4×10^{-2}	
		(0 ~ 2 000) $\mu\text{mol/mol}$	3.0×10^{-2}	
		(0 ~ 50) $\mu\text{mol/mol}$	4.9×10^{-2}	
		(0 ~ 850) $\mu\text{mol/mol}$	2.2×10^{-2}	
Others: pH meter, Electrical conductivity meter	90104	(0 ~ 1 000) $\mu\text{mol/mol}$	1.0×10^{-2}	CRM/ SICT-CP-90199
		(0 ~ 500) $\mu\text{mol/mol}$	2.3×10^{-2}	
		(0.05 ~ 2.0) cmol/mol	2.1×10^{-2}	
		(4 ~ 10) pH	0.013 pH	
Electrical conductivity meter	90104	100 $\mu\text{S/cm}$	3.1 $\mu\text{S/cm}$	
		1 413 $\mu\text{S/cm}$	9.7 $\mu\text{S/cm}$	
		12.85 mS/cm	0.073 mS/cm	
		111.3 mS/cm	0.78 mS/cm	

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017 & KS Q ISO/IEC 17025:2017

Institute of Calibration Technology Co.,Ltd
47-22, Jingoksan 2beon-ro, Gwangsan-gu, Gwangju
Phone : 82-62-953-5353, Fax : 82-62-953-5454 e-mail : sictadmin@sict.co.kr

CALIBRATION

Valid To : Oct. 29. 2025

Accreditation No : KC01-018

In recognition of the successful completion of the KOLAS evaluation process,
accreditation is granted to this laboratory to perform the following calibrations

Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site	Field Code	Item of Calibration	On-Site
102. Linear dimension			201. Mass			40424	Volt/Current recorders	Y
10211	Filler gauges	Y	20109	Electric balances	Y	501. Contact thermometry		
10216	Height gauges/measuring machines	Y	20112	Platform scale balances	Y	50101	Temperature generators: ovens, furnaces, isothermal liquid baths, ice-point baths, dry-block calibrators	Y
10216			20116	Weights	Y			
10228	Cylindrical plug/pingauges, Thread measuring wire gauges	Y	401. DC Voltage & current			50102	Temperature indicators /recorders/controllers, temperature calibrators	Y
			40101	DC ammeters	Y	50103	Glass thermometers: liquid-in-glass, Beckmann	N
10234	Ultrasonic thickness gauges	Y	40103	DC voltage/current calibrators	Y			
10236	Coating thickness testers	Y	40104	Electrical temperature calibrators	Y	50104	Resistance thermometers: SPRT, IPRT, thermistors, etc.	Y
	104. Form					50105	Thermal expansion thermometers ; bimetal, gas or liquid type	Y
10401	Form testers	Y	40205	Earth testers	Y			
10407	Precision surface plates	Y	40210	Insulation testers	Y	50107	Temperature transducers	Y
10409	Roundness measurement instruments	N	40212	DC voltmeters	Y	503. Humidity		
			402. Resistance, Capacitance and Inductance			50302	Relative humidity hygrometers polymer thin film, hair, etc.	Y
10412	Straight edges	Y	40214	Resistance meters	Y	50304	Temperature humidity recorders ; Hygrothermograph, etc	N
			40215	Resistors	Y	50305	Transducers; dew-point /relative humidity	N
	105. Complex geometry		403. AC voltage, current & power			50306	Humidity generators; two-pressure, two-temperature, flow mixing humidity generator, constant temperature and humidity chamber, etc.	Y
10503	Contact coordinate measuring machines	Y	40301	AC ammeters	Y			
			40302	Clamp ammeters/voltmeters	Y			
10504	Non-contact coordinate measuring machines	Y	40303	AC voltage/current calibrators	Y			
10511	Measuring microscopes, Profile projectors	Y	40310	Power factor meters	Y			
10512	Microscopes, micro measuring	Y	40311	AC power meters	Y			
10517	Stylus type roughness testers	Y	40312	AC power supplies	Y			
			40313	Puncture/safety testers	Y			
10601	Inside/Outside/Geartooth calipers, Caliper gauges	Y	40314	Power recorders	Y			
10603	Cylinder/bore gauges	Y	40318	AC voltmeters	Y			
10604	Depth gauges, Depth micrometers	Y	404. Other DC & LF Measurements					
10605	Dial/digital gauges	Y	40410	Line frequency meters	Y			
10609	Microindicators, Test indicators	Y	40416	Leakage current testers	Y			
10611	3-points, Micrometers	Y	40417	Electronic AC/DC loads	Y			
10612	Inside micrometers	Y	40419	Analogue/Digital multimeters	Y			
10613	Outside micrometers	Y	40421	Oscilloscopes	Y			

Note

1. This laboratory provides calibration services in permanent standard laboratory and at on-site.
2. Laboratory conducts on-site calibration should meet requirements of KOLAS-SR-007.
3. On-site calibration is allowed to items with marking 'Y', not allowed to items with marking 'N'.
4. Measurement uncertainty normally is quoted as an expanded uncertainty at a coverage probability of 95%, which usually requires the use of a coverage factor of $k = 2$. It expresses the lowest uncertainty of measurement that can be provided by accredited calibration laboratories in normal conditions.
5. Due to the calibration environment such as reference standards or customers' facilities, it is noted that uncertainty of measurement on a calibration certificate may be expressed larger than measurement uncertainty on scope of accreditation in general.

102. Linear dimension

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Filler gauges	10211	(0.01 ~ 10) mm	1.2 μm	Micrometers/ SICT-CP-10211
Height gauges/measuring machines	10216	(0 ~ 1 010) mm	$\sqrt{1.1^2 + (0.003 7 \times l_0)^2} \mu\text{m}$	Gauge Block, Step gauge/ SICT-CP-10216
Cylindrical plug/pin gauges, Thread measuring wire gauges	10228			Laser Scan Micrometers/ SICT-CP-10228
Cylindrical plug/pin gauges		(1 ~ 20) mm	$\sqrt{0.48^2 + (0.003 4 \times l_0)^2} \mu\text{m}$	
Ultrasonic thickness gauges	10234	(2.5 ~ 100) mm	3 μm	Ultrasonic Tester Blocks/ SICT-CP-10234
Coating thickness testers	10236	(0 ~ 1.527) mm	1.1 μm	Thickness specimens/ SICT-CP-10236

Note 1. l_0 unit : mm

104. Form

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Form testers	10401			Form Standard Specimens
Height length		(0 ~ 50) mm	1.0 μm	Gage Block, Angle Gage Block/
Width		(0 ~ 45) mm	1.7 μm	SICT-CP-10401
Angle		15° ~ 45°	2'	
Precision surface plates	10407			Electronic Level/ SICT-CP-10407
Flatness		(2 000 ~ 20 000) cm^2	2.0 μm	
Roundness measurement instruments	10409			Roundness Standard Ball/
Dector accuracy		(0 ~ 15) μm	0.62 μm	SICT-CP-10409
Rotational accuracy of spindle		360°	0.09 μm	
Rotational accuracy of axis		360°	0.05 μm	
Straight edges	10412			Electronic Micrometers/ SICT-CP-10412
		(0 ~ 1 500) mm	1.9 μm	
		(0 ~ 1 500) mm	1.8 μm	

Note 1. l_0 unit : mm

105. Complex geometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Contact coordinate measuring machines	10503	(0 ~ 1 000) mm	$\sqrt{0.94^2 + (0.004 8 \times l_0)^2} \mu\text{m}$	Step Gauge/ SICT-CP-10503
Non-contact coordinate measuring machines	10504	(0 ~ 500) mm	$\sqrt{0.53^2 + (0.003 6 \times l_0)^2} \mu\text{m}$	Standard Scale/ SICT-CP-10504
Measuring microscopes, Profile projectors	10511	(0 ~ 300) mm	$\sqrt{0.56^2 + (0.003 6 \times l_0)^2} \mu\text{m}$	Standard Scale/ SICT-CP-10511
		(5 ~ 100) 배	0.05 %	
		(0 ~ 360) °	0.9'	
Microscopes, micro measuring	10512	(0 ~ 1) mm (1 ~ 50) mm	1.3 μm 3.0 μm	Standard Scale/ SICT-CP-10512
Stylus type roughness testers	10517	(0 ~ 4) μm	추후확인	Roughness Specimen/ SICT-CP-10517
		(0 ~ 12) μm	추후확인	

Note 1. l_0 unit : mm

106. Various dimensional

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Inside/Outside/Geartooth calipers, Caliper gauges	10601			Gauge Block/ SICT-CP-10601	
		(0 ~ 2 000) mm	$\sqrt{8.2^2 + (0.008 1 \times l_0)^2} \mu\text{m}$		
		(0 ~ 100) mm (100 ~ 300) mm	$\sqrt{2.5^2 + (0.004 2 \times l_0)^2} \mu\text{m}$ $\sqrt{8.2^2 + (0.004 2 \times l_0)^2} \mu\text{m}$		
Cylinder/bore gauges	10603			Dial Gauge Tester/ SICT-CP-10603	
		Cylinder gauges Hole gauges	(0.7 μm 3.5 μm)		
Depth gauges, Depth micrometers	10604			Gauge Block/ SICT-CP-10604	
		Depth micrometers	$\sqrt{0.87^2 + (0.003 3 \times l_0)^2} \mu\text{m}$		
		Depth gauges	$\sqrt{6.0^2 + (0.007 8 \times l_0)^2} \mu\text{m}$		
Dial/digital gauges	10605			Gauge Block/ SICT-CP-10605	
		Dial/Digital gauges	$\sqrt{0.33^2 + (0.006 8 \times l_0)^2} \mu\text{m}$		
		Digital thickness gauges	$\sqrt{0.82^2 + (0.006 8 \times l_0)^2} \mu\text{m}$		
Micro indicators, Test indicators	10609	(0 ~ 5) mm	0.29 μm	Dial Gauge Tester/ SICT-CP-10609	
3-points, Micrometers	10611	(6 ~ 100) mm	$\sqrt{1.2^2 + (0.004 1 \times l_0)^2} \mu\text{m}$	Ring Gauge/ SICT-CP-10611	
Inside micrometers	10612			Gauge Block/ SICT-CP-10612	
		Inside micrometer	$\sqrt{1.1^2 + (0.004 2 \times l_0)^2} \mu\text{m}$		
		bar type micrometer(Accuracy of scale)	$\sqrt{1.1^2 + (0.004 2 \times l_0)^2} \mu\text{m}$		
bar type micrometer(Length of extension bars)					
		(13 ~ 150) mm	$\sqrt{1.5^2 + (0.004 2 \times l_0)^2} \mu\text{m}$		
Outside micrometers	10613			Gauge Block, cylindrical plug gauges/ SICT-CP-10613	
		Outside micrometers	$\sqrt{0.22^2 + (0.003 \times l_0)^2} \mu\text{m}$		
		V-anvil micrometers	$\sqrt{0.83^2 + (0.003 1 \times l_0)^2} \mu\text{m}$ 0.97 μm		

Note 1. l_0 unit : mm

201. Mass

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electric balances	20109	(0 ~ 2) g (2 ~ 6) g (6 ~ 20) g (20 ~ 50) g (50 ~ 200) g (200 ~ 500) g (500 ~ 2 000) g (2 ~ 5) kg (5 ~ 10) kg (10 ~ 20) kg (20 ~ 30) kg (30 ~ 60) kg (60 ~ 100) kg (100 ~ 600) kg (600 ~ 1 000) kg	21 µg 40 µg 42 µg 61 µg 0.12 mg 0.33 mg 1.2 mg 3.2 mg 6.2 mg 12 mg 0.20 g 0.36 g 6.8 g 18 g 38 g	Weight / SICT-CP-20109
Platform scale balances	20112	(0 ~ 200) kg (200 ~ 500) kg (500 ~ 1 000) kg	46 g 0.092 kg 0.46 kg	Weight / SICT-CP-20112
Weights less than class F1	20116	(1 mg ~ 20 kg) 1 mg 2 mg 5 mg 10 mg 20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg	3.5 µg 3.5 µg 3.5 µg 3.5 µg 4.4 µg 5.3 µg 6.4 µg 7.5 µg 10 µg 12 µg 16 µg 19 µg 24 µg 31 µg 38 µg 64 µg 0.12 mg 0.39 mg 0.66 mg 1.2 mg 4.7 mg 7.1 mg 13 mg	Weight , Electronic Balance / SICT-CP-20116

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC ammeters	40101	(±)		Calibrator/ SICT-CP-40101
		0 μA	9.3 nA	
		(0.1 ~ 1) μA	9.4 × 10 ⁻³	
		(1 ~ 10) μA	9.9 × 10 ⁻⁴	
		(10 ~ 100) μA	1.5 × 10 ⁻⁴	
		(0.1 ~ 10) mA	6.9 × 10 ⁻⁵	
		(10 ~ 100) mA	8.0 × 10 ⁻⁵	
		(0.1 ~ 1) A	1.2 × 10 ⁻⁴	
		(1 ~ 10) A	6.5 × 10 ⁻⁴	
		(10 ~ 20) A	1.2 × 10 ⁻³	
DC voltage/current calibrators	40103	(±)		Reference Multimeter/ SICT-CP-40103
		0 mV	0.51 μV	
		(0 ~ 1) mV	6.1 × 10 ⁻⁴	
		(1 ~ 10) mV	6.2 × 10 ⁻⁵	
		(10 ~ 100) mV	1.3 × 10 ⁻⁵	
		(0.1 ~ 1) V	6.6 × 10 ⁻⁶	
		(1 ~ 10) V	5.2 × 10 ⁻⁶	
		(10 ~ 100) V	7.2 × 10 ⁻⁶	
		(100 ~ 1 000) V	9.0 × 10 ⁻⁶	
DC Current	40103	(±)		
		0 μA	6.9 nA	
		(0 ~ 10) μA	7.4 × 10 ⁻⁴	
		(10 ~ 100) μA	1.1 × 10 ⁻⁴	
		(0.1 ~ 1) mA	7.2 × 10 ⁻⁵	
		(1 ~ 10) mA	4.7 × 10 ⁻⁵	
		(10 ~ 100) mA	7.5 × 10 ⁻⁵	
		(0.1 ~ 1) A	2.4 × 10 ⁻⁴	
		(1 ~ 10) A	5.7 × 10 ⁻⁴	
		(10 ~ 100) A	2.8 × 10 ⁻⁴	
Electrical temperature calibrators	40104			Digital Multimeter/ SICT-CP-40104
		TEMPERATURE(SOURCE)		
		T/C		
		-9.835 mV	0.62 μV	
		(-9.835 ~ -5.237) mV	0.62 μV	
		(-5.237 ~ 0.000) mV	0.52 μV	
		(0.000 ~ 13.421) mV	0.62 μV	
		(13.421 ~ 68.788) mV	1.2 μV	
		(68.788 ~ 76.373) mV	1.3 μV	
RTD	40104	1.000 Ω	0.063 mΩ	
		(1.000 ~ 2.499) Ω	3.7 × 10 ⁻⁵	
		(2.499 ~ 16.996) Ω	3.0 × 10 ⁻⁵	
		(16.996 ~ 100.000) Ω	1.0 × 10 ⁻⁵	
		(100.000 ~ 249.584) Ω	2.8 × 10 ⁻⁵	
		(249.584 ~ 3 233.3) Ω	1.8 × 10 ⁻⁵	

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrical temperature calibrators				Digital Multimeter/ SICT-CP-40104
DC VOLTAGE(SOURCE)	40104	(±) 0 mV (0 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V	8.1 μ V 8.5×10^{-4} 1.2×10^{-4} 4.6×10^{-5} 4.0×10^{-5} 5.9×10^{-5}	
DC CURRENT(SOURCE)		(±) 0 mA (0 ~ 1) mA (1 ~ 10) mA (10 ~ 100) mA	0.14 μ A 7.0×10^{-4} 1.0×10^{-3} 7.0×10^{-4}	
RESISTANCE(SOURCE)		0 Ω (0 ~ 1) Ω (1 ~ 10) Ω 100 Ω ~ 50 k Ω	0.072 m Ω 6.7×10^{-5} 4.6×10^{-4} 1.2×10^{-4}	
TEMPERATURE(MEASURE)	T/C	-9.836 mV (-9.836 ~ -5.238) mV (-5.238 ~ 0.000) mV (0.000 ~ 0.002) mV (0.002 ~ 6.319) mV (6.319 ~ 13.421) mV (13.421 ~ 21.036) mV (21.036 ~ 28.946) mV (28.946 ~ 37.006) mV (37.006 ~ 53.112) mV (53.112 ~ 61.017) mV (61.017 ~ 76.373) mV	0.84 μ V 0.83 μ V 0.80 μ V 0.70 μ V 0.81 μ V 0.87 μ V 0.94 μ V 1.0 μ V 1.1 μ V 1.2 μ V 1.3 μ V 1.4 μ V	
	RTD	0.999 Ω (0.999 ~ 2.497) Ω (2.497 ~ 4.316) Ω (4.316 ~ 16.995) Ω (16.995 ~ 177.156) Ω (177.156 ~ 249.584) Ω (249.584 ~ 3 233.3) Ω	0.24 m Ω 1.0×10^{-4} 7.1×10^{-5} 3.9×10^{-5} 3.4×10^{-5} 3.5×10^{-5} 4.3×10^{-5}	
DC VOLTAGE(MEASURE)		(±) 0 mV (0 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 200) V (200 ~ 300) V	1.2 μ V 1.4×10^{-4} 3.5×10^{-5} 6.0×10^{-5} 1.7×10^{-5} 2.3×10^{-5} 2.9×10^{-5}	
DC CURRENT(MEASURE)		(±) 0 mA (0 ~ 1) mA (1 ~ 10) mA (10 ~ 50) mA (50 ~ 100) mA (100 ~ 130) mA	0.09 μ A 9.3×10^{-5} 7.0×10^{-5} 9.6×10^{-5} 8.1×10^{-5} 9.6×10^{-5}	

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrical temperature calibrators RESISTANCE(MEASURE)	40104	0 Ω (0 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) kΩ (1 ~ 10) kΩ (10 ~ 20) kΩ (20 ~ 30) kΩ (30 ~ 40) kΩ (40 ~ 50) kΩ	23 mΩ 1.3×10^{-4} 4.8×10^{-5} 3.4×10^{-5} 7.3×10^{-5} 7.0×10^{-5} 4.8×10^{-5} 4.0×10^{-5} 4.2×10^{-5} 3.9×10^{-5}	Digital Multimeter/ SICT-CP-40104
DC power supplies	40108	(±) 0 mV (0 ~ 10) mV (10 ~ 100) mV 100 mV ~ 1 V (1 ~ 10) V (10 ~ 100) V (100 ~ 600) V (600 ~ 1 000) V	5.8 μV 5.8×10^{-4} 5.9×10^{-5} 9.2×10^{-5} 5.3×10^{-5} 9.2×10^{-5} 1.7×10^{-5} 6.6×10^{-5}	DC Electronics Load/ SICT-CP-40108
		DC Current 1 mA (1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 5) A (5 ~ 10) A (10 ~ 20) A (20 ~ 200) A	5.8 μA 6.5×10^{-3} 6.5×10^{-4} 8.7×10^{-5} 1.3×10^{-4} 4.0×10^{-4} 3.1×10^{-4} 2.2×10^{-4}	
		Load regulation 0 mV (0 ~ 2) mV (2 ~ 20) mV (20 ~ 200) mV	0.97 mV 4.9×10^{-1} 3.3×10^{-1} 3.3×10^{-2}	
		Ripple 0.1 mV (0.1 ~ 0.2) mV (0.2 ~ 0.4) mV (0.4 ~ 0.6) mV (0.6 ~ 1) mV (1 ~ 10) mV (10 ~ 50) mV	0.097 mV 6.4×10^{-1} 2.5×10^{-1} 1.6×10^{-1} 9.6×10^{-2} 9.2×10^{-2} 1.9×10^{-1}	
DC voltmeters	40112	(±) 0 mV (0 ~ 1) mV (1 ~ 2) mV (2 ~ 8) mV (8 ~ 10) mV (10 ~ 20) mV (20 ~ 50) mV (50 ~ 80) mV (80 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V	0.69 μV 8.0×10^{-4} 4.0×10^{-4} 1.6×10^{-4} 8.0×10^{-5} 4.5×10^{-5} 2.4×10^{-5} 1.9×10^{-5} 1.6×10^{-5} 9.4×10^{-6} 8.6×10^{-6} 1.0×10^{-5} 1.1×10^{-5}	Calibrator/ SICT-CP-40112

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Earth testers	Test Voltage	40205		Decade Resistor/ SICT-CP-40205
		60 Hz		
		1 V	6.1×10^{-3}	
		(1 ~ 10) V	6.2×10^{-4}	
		(10 ~ 50) V	1.2×10^{-3}	
		(50 ~ 100) V	6.2×10^{-4}	
		(100 ~ 500) V	4.9×10^{-4}	
		(500 ~ 1 000) V	7.7×10^{-4}	
		1 mΩ	1.3×10^{-3}	
		1 mΩ ~ 1 Ω	1.3×10^{-3}	
	Resistance	1 Ω ~ 10 Ω	8.4×10^{-4}	
		10 Ω ~ 100 kΩ	6.5×10^{-4}	
		60 Hz		
		1 A	3.1×10^{-3}	
	AC Current out	(1 ~ 3) A	4.1×10^{-3}	
		(3 ~ 10) A	3.1×10^{-3}	
		(10 ~ 20) A	3.4×10^{-3}	
		(20 ~ 30) A	6.4×10^{-3}	
		(30 ~ 60) A	4.6×10^{-3}	
		(1 ~ 100) s	5.8×10^{-6}	
		(100 ~ 1 000) s	8.2×10^{-6}	
Insulation testers	AC Voltage	40210		High Resistance Decade/ SICT-CP-40210
		60 Hz		
		1 V	1.2×10^{-4}	
		(1 ~ 10) V	1.2×10^{-4}	
		(10 ~ 100) V	1.3×10^{-4}	
	Insulation Voltage	(100 ~ 1 000) V	4.9×10^{-4}	
		1 V	6.1×10^{-4}	
		(1 ~ 10) V	6.1×10^{-5}	
		(10 ~ 25) V	2.5×10^{-4}	
		(25 ~ 50) V	1.2×10^{-4}	
		(50 ~ 100) V	6.2×10^{-5}	
		(100 ~ 250) V	2.5×10^{-4}	
		(250 ~ 500) V	1.2×10^{-4}	
		(500 ~ 1 000) V	6.1×10^{-5}	
		(1 000 ~ 5 000) V	6.5×10^{-3}	
	Insulation Resistance	(5 000 ~ 10 000) V	6.1×10^{-3}	
		1 kΩ	1.4×10^{-4}	
		1 kΩ ~ 1 MΩ	1.2×10^{-4}	
		(1 ~ 10) MΩ	3.6×10^{-4}	
		(10 ~ 100) MΩ	1.2×10^{-3}	
		100 MΩ ~ 1 GΩ	2.3×10^{-3}	
		(1 ~ 10) GΩ	5.8×10^{-3}	
		(10 ~ 100) GΩ	5.9×10^{-3}	
Resistance meters	Resistance	40214		Standard Resistance Set/ SICT-CP-40214
		1 mΩ	2.2×10^{-5}	
		(1 ~ 10) mΩ	2.1×10^{-5}	
		(10 ~ 100) mΩ	2.2×10^{-5}	
		100 mΩ ~ 1 MΩ	2.1×10^{-5}	

402. Resistance, Capacitance and Inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistors Decade Resistance	40215	0 Ω (0 ~ 1) mΩ (1 ~ 2) mΩ (2 ~ 3) mΩ (3 ~ 6) mΩ (6 ~ 8) mΩ (8 ~ 10) mΩ (10 ~ 30) mΩ (30 ~ 60) mΩ 60 mΩ ~ 0.1 Ω (0.1 ~ 0.3) Ω (0.3 ~ 0.6) Ω (0.6 ~ 1) Ω (1 ~ 3) Ω (3 ~ 10) Ω (10 ~ 300) Ω (300 ~ 1000) Ω (1 ~ 10) kΩ (10 ~ 100) kΩ 100 kΩ ~ 1 MΩ (1 ~ 2) MΩ (2 ~ 10) MΩ (10 ~ 100) MΩ 100 MΩ ~ 1 GΩ	62 μΩ 3.1×10^{-3} 6.2×10^{-3} 2.1×10^{-3} 1.5×10^{-3} 8.9×10^{-4} 6.9×10^{-4} 3.1×10^{-3} 1.6×10^{-3} 8.9×10^{-4} 3.1×10^{-4} 1.6×10^{-4} 9.0×10^{-5} 4.6×10^{-5} 2.5×10^{-5} 2.9×10^{-5} 1.6×10^{-5} 3.7×10^{-5} 3.0×10^{-5} 4.0×10^{-5} 8.5×10^{-5} 5.9×10^{-5} 6.5×10^{-4} 6.6×10^{-3}	Standard Resistance Set/ SICT-CP-40215

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC ammeters	40301	(100 µA) 50 Hz ~ 1 kHz	2.7×10^{-3}	Power Calibrator, Calibrator/ SICT-CP-40301
		(0.1 ~ 1) mA 50 Hz ~ 1 kHz	1.4×10^{-3}	
		(1 ~ 10) mA (50 ~ 60) Hz 60 Hz ~ 1 kHz	6.1×10^{-4} 7.0×10^{-4}	
		(10 ~ 100) mA (50 ~ 60) Hz 60 Hz ~ 1 kHz	4.7×10^{-4} 7.0×10^{-4}	
		(0.1 ~ 1) A (50 ~ 60) Hz 60 Hz ~ 1 kHz	2.1×10^{-4} 7.0×10^{-4}	
		(1 ~ 5) A (50 ~ 60) Hz 60 Hz ~ 1 kHz	2.6×10^{-4} 1.3×10^{-3}	
		(5 ~ 10) A (50 ~ 60) Hz 60 Hz ~ 1 kHz	3.0×10^{-4} 1.4×10^{-3}	
		(10 ~ 20) A (50 ~ 60) Hz	4.8×10^{-4}	
		(20 ~ 40) A (50 ~ 60) Hz	5.2×10^{-4}	
Clamp ammeters/voltmeters	40302	(50 ~ 60) Hz 100 µA (100 ~ 200) µA (200 ~ 300) µA (300 ~ 400) µA (400 ~ 500) µA (500 ~ 600) µA (600 ~ 900) µA 900 µA ~ 1 mA (1 ~ 2) mA (2 ~ 3) mA (3 ~ 4) mA (4 ~ 5) mA (5 ~ 6) mA (6 ~ 7) mA (7 ~ 8) mA (8 ~ 9) mA (9 ~ 10) mA (10 ~ 20) mA (20 ~ 30) mA (30 ~ 40) mA (40 ~ 50) mA (50 ~ 60) mA (60 ~ 70) mA	0.27 µA 2.1×10^{-3} 1.9×10^{-3} 1.7×10^{-3} 1.6×10^{-3} 1.5×10^{-3} 1.4×10^{-3} 1.5×10^{-3} 1.3×10^{-3} 1.2×10^{-3} 1.1×10^{-3} 9.5×10^{-4} 8.7×10^{-4} 8.1×10^{-4} 7.6×10^{-4} 7.3×10^{-4} 9.3×10^{-4} 7.6×10^{-4} 6.5×10^{-4} 6.0×10^{-4} 5.7×10^{-4} 4.8×10^{-4} 4.6×10^{-4}	Power Calibrator, Calibrator/ SICT-CP-40302

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters	40302	(70 ~ 90) mA	4.4×10^{-4}	Power Calibrator, Calibrator/ SICT-CP-40302
		(90 ~ 100) mA	7.5×10^{-4}	
		(100 ~ 200) mA	3.9×10^{-4}	
		(200 ~ 300) mA	3.1×10^{-4}	
		(300 ~ 400) mA	2.7×10^{-4}	
		(400 ~ 500) mA	2.5×10^{-4}	
		(500 ~ 700) mA	2.4×10^{-4}	
		(700 ~ 900) mA	2.2×10^{-4}	
		900 mA ~ 1 A	6.4×10^{-4}	
		(1 ~ 2) A	3.7×10^{-4}	
		(2 ~ 3) A	4.0×10^{-4}	
		(3 ~ 4) A	3.3×10^{-4}	
		(4 ~ 5) A	2.9×10^{-4}	
		(5 ~ 6) A	3.4×10^{-4}	
		(6 ~ 7) A	3.3×10^{-4}	
		(7 ~ 9) A	3.1×10^{-4}	
		(9 ~ 10) A	6.8×10^{-4}	
		(10 ~ 20) A	6.4×10^{-4}	
		(20 ~ 30) A	5.6×10^{-4}	
		(30 ~ 40) A	5.2×10^{-4}	
		(40 ~ 500) A	1.2×10^{-3}	
		(500 ~ 700) A	1.4×10^{-3}	
		(700 ~ 1 000) A	1.3×10^{-3}	
DC Current	30 µA	30 µA	29 nA	
		(30 ~ 40) µA	7.7×10^{-4}	
		(40 ~ 50) µA	6.5×10^{-4}	
		(50 ~ 60) µA	5.7×10^{-4}	
		(60 ~ 70) µA	5.1×10^{-4}	
		(70 ~ 80) µA	4.7×10^{-4}	
		(80 ~ 90) µA	4.4×10^{-4}	
		(90 ~ 100) µA	7.3×10^{-4}	
		(100 ~ 200) µA	4.2×10^{-4}	
		(200 ~ 300) µA	3.2×10^{-4}	
		(300 ~ 400) µA	2.8×10^{-4}	
		(400 ~ 500) µA	2.5×10^{-4}	
		(500 ~ 600) µA	2.4×10^{-4}	
		(600 ~ 700) µA	2.3×10^{-4}	
		(700 ~ 800) µA	2.2×10^{-4}	
		(800 ~ 900) µA	2.1×10^{-4}	
		900 µA ~ 1 mA	6.4×10^{-4}	
		(1 ~ 2) mA	3.6×10^{-4}	
		(2 ~ 4) mA	2.7×10^{-4}	
		(4 ~ 9) mA	1.8×10^{-4}	
		(9 ~ 10) mA	6.2×10^{-4}	
		(10 ~ 20) mA	3.3×10^{-4}	
		(20 ~ 70) mA	2.4×10^{-4}	
		(70 ~ 90) mA	1.5×10^{-4}	
		(90 ~ 100) mA	6.2×10^{-4}	
		(100 ~ 200) mA	3.3×10^{-4}	
		(200 ~ 400) mA	3.9×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters	40302			
DC Current		(400 ~ 900) mA	3.5×10^{-4}	Power Calibrator, Calibrator/ SICT-CP-40302
		900 mA ~ 1 A	6.7×10^{-4}	
		(1 ~ 2) A	4.8×10^{-4}	
		(2 ~ 3) A	6.0×10^{-4}	
		(3 ~ 9) A	8.1×10^{-4}	
		(9 ~ 10) A	8.9×10^{-4}	
		(10 ~ 20) A	1.2×10^{-3}	
		(20 ~ 40) A	3.8×10^{-4}	
		(40 ~ 500) A	1.3×10^{-3}	
		(500 ~ 1 000) A	1.7×10^{-3}	
AC Voltage		(50 ~ 60) Hz		
		10 mV	11 μ V	
		(10 ~ 20) mV	6.3×10^{-4}	
		(20 ~ 30) mV	4.7×10^{-4}	
		(30 ~ 40) mV	4.6×10^{-4}	
		(40 ~ 50) mV	4.0×10^{-4}	
		(50 ~ 60) mV	3.6×10^{-4}	
		(60 ~ 80) mV	3.3×10^{-4}	
		(80 ~ 90) mV	2.9×10^{-4}	
		(90 ~ 100) mV	2.8×10^{-4}	
		(100 ~ 200) mV	2.2×10^{-4}	
		(200 ~ 300) mV	2.0×10^{-4}	
		(300 ~ 400) mV	3.6×10^{-4}	
		(400 ~ 600) mV	3.2×10^{-4}	
		(600 ~ 700) mV	2.8×10^{-4}	
		(700 ~ 800) mV	2.6×10^{-4}	
		800 mV ~ 1 V	2.5×10^{-4}	
		(1 ~ 3) V	2.1×10^{-4}	
		(3 ~ 4) V	3.5×10^{-4}	
		(4 ~ 5) V	3.2×10^{-4}	
		(5 ~ 6) V	2.9×10^{-4}	
		(6 ~ 7) V	2.8×10^{-4}	
		(7 ~ 8) V	2.6×10^{-4}	
		(8 ~ 10) V	2.5×10^{-4}	
		(10 ~ 20) V	2.3×10^{-4}	
		(20 ~ 40) V	1.9×10^{-4}	
		(40 ~ 70) V	2.2×10^{-4}	
		(70 ~ 90) V	1.9×10^{-4}	
		(90 ~ 100) V	1.8×10^{-4}	
		(100 ~ 200) V	1.5×10^{-4}	
		(200 ~ 500) V	1.9×10^{-4}	
		(500 ~ 700) V	2.6×10^{-4}	
		(700 ~ 1 000) V	3.7×10^{-4}	
DC Voltage		10 mV	61 μ V	
		(10 ~ 20) mV	3.1×10^{-3}	
		(20 ~ 30) mV	2.0×10^{-3}	
		(30 ~ 40) mV	1.5×10^{-3}	
		(40 ~ 50) mV	1.2×10^{-3}	
		(50 ~ 60) mV	1.0×10^{-3}	
		(60 ~ 70) mV	8.7×10^{-4}	
		(70 ~ 80) mV	7.6×10^{-4}	
		(80 ~ 90) mV	6.8×10^{-4}	
		(90 ~ 100) mV	7.0×10^{-5}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters	40302	(100 ~ 200) mV	4.2×10^{-5}	Power Calibrator, Calibrator/ SICT-CP-40302
		(200 ~ 300) mV	3.4×10^{-5}	
		(300 ~ 400) mV	2.4×10^{-5}	
		(400 ~ 500) mV	2.2×10^{-5}	
		(500 ~ 600) mV	2.0×10^{-5}	
		(600 ~ 800) mV	1.8×10^{-5}	
		(800 ~ 900) mV	1.7×10^{-5}	
		900 mV ~ 1 V	6.3×10^{-5}	
		(1 ~ 2) V	3.4×10^{-5}	
		(2 ~ 4) V	2.5×10^{-5}	
		(4 ~ 5) V	2.2×10^{-5}	
		(5 ~ 6) V	2.0×10^{-5}	
		(6 ~ 7) V	1.9×10^{-5}	
		(7 ~ 9) V	1.8×10^{-5}	
		(9 ~ 10) V	6.3×10^{-5}	
		(10 ~ 20) V	3.4×10^{-5}	
		(20 ~ 30) V	2.5×10^{-5}	
		(30 ~ 40) V	3.0×10^{-6}	
		(40 ~ 50) V	2.7×10^{-5}	
		(50 ~ 60) V	2.6×10^{-5}	
		(60 ~ 70) V	2.5×10^{-5}	
		(70 ~ 90) V	2.4×10^{-5}	
		(90 ~ 100) V	6.5×10^{-5}	
		(100 ~ 200) V	3.8×10^{-5}	
		(200 ~ 400) V	3.0×10^{-5}	
		(400 ~ 500) V	2.8×10^{-5}	
		(500 ~ 600) V	2.6×10^{-5}	
		(600 ~ 700) V	2.5×10^{-5}	
		(700 ~ 900) V	2.4×10^{-5}	
		(900 ~ 1 000) V	6.5×10^{-5}	
Resistance		1 Ω	0.62 mΩ	
		(1 ~ 9) Ω	3.1×10^{-4}	
		(9 ~ 100) Ω	6.1×10^{-4}	
		(100 ~ 900) Ω	4.7×10^{-5}	
		900 Ω ~ 90 kΩ	7.0×10^{-5}	
		90 kΩ ~ 1 MΩ	7.2×10^{-5}	
		(1 ~ 10) MΩ	1.9×10^{-4}	
		(10 ~ 100) MΩ	7.1×10^{-4}	
AC voltage/current calibrators	40303	(2 mV)		Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
		10 Hz	8.2×10^{-3}	
		(10 ~ 40) Hz	2.5×10^{-3}	
		40 Hz ~ 1 kHz	1.3×10^{-3}	
		(1 ~ 20) kHz	5.0×10^{-3}	
		(20 ~ 100) kHz	1.2×10^{-2}	
		(2 ~ 20) mV		
		10 Hz	9.5×10^{-4}	
		(10 ~ 40) Hz	3.5×10^{-4}	
		40 Hz ~ 1 kHz	2.3×10^{-4}	
		(1 ~ 20) kHz	8.1×10^{-4}	
		(20 ~ 100) kHz	2.0×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators	40303			Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
AC Voltage		(20 ~ 30) mV		
		10 Hz	6.8×10^{-4}	
		(10 ~ 40) Hz	2.7×10^{-4}	
		40 Hz ~ 1 kHz	1.9×10^{-4}	
		(1 ~ 20) kHz	6.5×10^{-4}	
		(20 ~ 100) kHz	1.6×10^{-3}	
		(30 ~ 40) mV		
		10 Hz	5.4×10^{-4}	
		(10 ~ 40) Hz	2.3×10^{-4}	
		40 Hz ~ 1 kHz	1.7×10^{-4}	
		(1 ~ 20) kHz	5.8×10^{-4}	
		(20 ~ 100) kHz	1.4×10^{-3}	
		(40 ~ 50) mV		
		10 Hz	4.6×10^{-4}	
		(10 ~ 40) Hz	2.1×10^{-4}	
		40 Hz ~ 1 kHz	1.6×10^{-4}	
		(1 ~ 20) kHz	5.3×10^{-4}	
		(20 ~ 100) kHz	1.3×10^{-3}	
		(50 ~ 60) mV		
		10 Hz	4.1×10^{-4}	
		(10 ~ 40) Hz	1.9×10^{-4}	
		40 Hz ~ 1 kHz	1.5×10^{-4}	
		(1 ~ 20) kHz	5.0×10^{-4}	
		(20 ~ 100) kHz	1.2×10^{-3}	
		(70 ~ 90) mV		
		10 Hz	3.7×10^{-4}	
		(10 ~ 40) Hz	1.8×10^{-4}	
		40 Hz ~ 1 kHz	1.5×10^{-4}	
		(1 ~ 20) kHz	4.8×10^{-4}	
		(20 ~ 100) kHz	1.1×10^{-3}	
		(90 ~ 100) mV		
		10 Hz	3.0×10^{-4}	
		(10 ~ 40) Hz	1.6×10^{-4}	
		40 Hz ~ 1 kHz	1.4×10^{-4}	
		(1 ~ 20) kHz	4.4×10^{-4}	
		(20 ~ 100) kHz	1.0×10^{-3}	
		(100 ~ 200) mV		
		10 Hz	2.3×10^{-4}	
		(10 ~ 40) Hz	1.4×10^{-4}	
		40 Hz ~ 1 kHz	1.3×10^{-4}	
		(1 ~ 20) kHz	3.9×10^{-4}	
		(20 ~ 100) kHz	9.3×10^{-4}	
		(200 ~ 300) mV		
		10 Hz	8.6×10^{-4}	
		(10 ~ 40) Hz	2.5×10^{-4}	
		40 Hz ~ 1 kHz	2.0×10^{-4}	
		(1 ~ 20) kHz	4.1×10^{-4}	
		(20 ~ 100) kHz	1.4×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators	40303			Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
AC Voltage		(300 mV ~ 0.4 V)		
		10 Hz	6.6×10^{-4}	
		(10 ~ 40) Hz	2.1×10^{-4}	
		40 Hz ~ 1 kHz	1.6×10^{-4}	
		(1 ~ 20) kHz	3.6×10^{-4}	
		(20 ~ 100) kHz	1.2×10^{-3}	
		(0.4 ~ 0.9) V		
		10 Hz	5.7×10^{-4}	
		(10 ~ 40) Hz	1.9×10^{-4}	
		40 Hz ~ 1 kHz	1.6×10^{-4}	
		(1 ~ 20) kHz	3.3×10^{-4}	
		(20 ~ 100) kHz	1.2×10^{-3}	
		(0.9 ~ 2) V		
		10 Hz	4.1×10^{-4}	
		(10 ~ 40) Hz	1.6×10^{-4}	
		40 Hz ~ 1 kHz	1.1×10^{-4}	
		(1 ~ 20) kHz	2.8×10^{-4}	
		(20 ~ 100) kHz	8.2×10^{-4}	
		(2 ~ 3) V		
		10 Hz	8.6×10^{-4}	
		(10 ~ 40) Hz	2.3×10^{-4}	
		40 Hz ~ 1 kHz	2.0×10^{-4}	
		(1 ~ 20) kHz	4.1×10^{-4}	
		(20 ~ 100) kHz	1.4×10^{-3}	
		(3 ~ 9) V		
		10 Hz	6.6×10^{-4}	
		(10 ~ 40) Hz	2.0×10^{-4}	
		40 Hz ~ 1 kHz	1.6×10^{-4}	
		(1 ~ 20) kHz	3.6×10^{-4}	
		(20 ~ 100) kHz	1.2×10^{-3}	
		(9 ~ 20) V		
		10 Hz	4.1×10^{-4}	
		(10 ~ 40) Hz	1.5×10^{-4}	
		40 Hz ~ 1 kHz	1.1×10^{-4}	
		(1 ~ 20) kHz	2.9×10^{-4}	
		(20 ~ 100) kHz	8.4×10^{-4}	
		(20 ~ 30) V		
		10 Hz	8.8×10^{-4}	
		(10 ~ 40) Hz	2.5×10^{-4}	
		40 Hz ~ 1 kHz	2.0×10^{-4}	
		(1 ~ 20) kHz	4.1×10^{-4}	
		(20 ~ 100) kHz	1.4×10^{-3}	
		(30 ~ 90) V		
		10 Hz	6.8×10^{-4}	
		(10 ~ 40) Hz	2.1×10^{-4}	
		40 Hz ~ 1 kHz	1.6×10^{-4}	
		(1 ~ 20) kHz	3.6×10^{-4}	
		(20 ~ 100) kHz	1.2×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators	40303			Alternating Voltage Measurement Standard, Reference Multimeter, Current Shunt/ SICT-CP-40303
AC Voltage		(90 ~ 200) V 10 Hz (10 ~ 40) Hz 40 Hz ~ 1 kHz (1 ~ 20) kHz (20 ~ 100) kHz	8.6×10^{-4} 2.3×10^{-4} 2.0×10^{-4} 4.1×10^{-4} 1.4×10^{-3}	
		(200 ~ 300) V 50 Hz ~ 1 kHz	1.7×10^{-4}	
		(300 ~ 1 000) V 50 Hz 50 Hz ~ 1 kHz	1.3×10^{-4} 1.2×10^{-4}	
AC Current		(100 μA) 10 Hz 10 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	7.1×10^{-4} 5.6×10^{-4} 7.1×10^{-4} 2.3×10^{-3}	
		(100 μA ~ 1 mA) 10 Hz 10 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	6.2×10^{-4} 5.5×10^{-4} 6.4×10^{-4} 2.1×10^{-3}	
		(1 ~ 10) mA 10 Hz 10 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	6.2×10^{-4} 5.5×10^{-4} 6.2×10^{-4} 3.2×10^{-3}	
		(10 ~ 100) mA 10 Hz 10 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	6.3×10^{-4} 5.4×10^{-4} 6.1×10^{-4} 1.5×10^{-3}	
		(100 mA ~ 1 A) 40 Hz ~ 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz	9.8×10^{-4} 1.2×10^{-3} 8.1×10^{-3}	
		(1 ~ 10) A (40 ~ 100) Hz 100 Hz ~ 1 kHz	1.3×10^{-3} 1.5×10^{-3}	
		(10 ~ 20) A (40 ~ 100) Hz 100 Hz ~ 1 kHz	1.6×10^{-3} 1.7×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Power factor meters	40310	(50 ~ 60) Hz		
AC Power Factor		-1 ~ 1	2.1×10^{-4}	Power Calibrator/ SICT-CP-40310
		-0.9 , 0.9	2.7×10^{-4}	
		-0.8 , 0.8	3.2×10^{-4}	
		-0.7 , 0.7	4.2×10^{-4}	
		-0.6 , 0.6	4.4×10^{-4}	
		-0.5 , 0.5	4.6×10^{-4}	
		-0.4 , 0.4	4.8×10^{-4}	
		-0.3 , 0.3	4.9×10^{-4}	
		-0.2 , 0.2	5.0×10^{-4}	
		-0.1 , 0.1	5.0×10^{-4}	
AC power meters	40311	(50 ~ 60) Hz		
AC Voltage		3 mV	$8 \mu\text{V}$	Power Calibrator, Calibrator/ SICT-CP-40311
		(3 ~ 10) mV	9.6×10^{-4}	
		(10 ~ 20) mV	5.6×10^{-4}	
		(20 ~ 40) mV	4.4×10^{-4}	
		(40 ~ 60) mV	3.8×10^{-4}	
		(60 ~ 90) mV	3.2×10^{-4}	
		(90 ~ 100) mV	2.4×10^{-4}	
		(100 ~ 200) mV	2.4×10^{-4}	
		(200 ~ 300) mV	1.6×10^{-4}	
		(300 ~ 400) mV	1.3×10^{-4}	
		(400 ~ 600) mV	1.2×10^{-4}	
		(600 ~ 700) mV	1.1×10^{-4}	
		(700 ~ 900) mV	1.0×10^{-4}	
		900 mV ~ 1 V	9.9×10^{-5}	
		(1 ~ 5) V	1.2×10^{-4}	
		(5 ~ 10) V	9.8×10^{-5}	
		(10 ~ 20) V	2.2×10^{-4}	
		(20 ~ 30) V	1.8×10^{-4}	
		(30 ~ 40) V	1.5×10^{-4}	
		(40 ~ 70) V	1.4×10^{-4}	
		(70 ~ 100) V	1.2×10^{-4}	
		(100 ~ 300) V	1.6×10^{-4}	
		(300 ~ 400) V	1.3×10^{-4}	
		(400 ~ 500) V	1.2×10^{-4}	
		(500 ~ 800) V	1.1×10^{-4}	
		(800 ~ 1 000) V	1.0×10^{-4}	
AC Current		(50 ~ 60) Hz		
		100 μA	$0.27 \mu\text{A}$	
		(100 ~ 200) μA	2.0×10^{-3}	
		(200 ~ 400) μA	1.8×10^{-3}	
		(400 ~ 500) μA	1.6×10^{-3}	
		(500 ~ 700) μA	1.5×10^{-3}	
		(700 ~ 900) μA	1.4×10^{-3}	
		900 μA ~ 1 mA	3.2×10^{-4}	
		(1 ~ 2) mA	3.5×10^{-4}	
		(2 ~ 5) mA	2.7×10^{-4}	
		(5 ~ 8) mA	2.5×10^{-4}	
		(8 ~ 10) mA	2.2×10^{-4}	
		(10 ~ 15) mA	4.4×10^{-4}	
		(15 ~ 20) mA	3.5×10^{-4}	
		(20 ~ 30) mA	3.6×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311	(30 ~ 40) mA (40 ~ 60) mA (60 ~ 70) mA (70 ~ 100) mA (100 ~ 150) mA (150 ~ 300) mA (300 ~ 800) mA (800 mA ~ 2 A) (2 ~ 3) A (3 ~ 4) A (4 ~ 5) A (5 ~ 6) A (6 ~ 10) A (10 ~ 15) A (15 ~ 20) A (20 ~ 30) A (30 ~ 40) A (40 ~ 500) A (500 ~ 800) A (800 ~ 1 000) A	3.0×10^{-4} 2.7×10^{-4} 2.4×10^{-4} 2.3×10^{-4} 2.7×10^{-4} 2.5×10^{-4} 2.2×10^{-4} 2.1×10^{-4} 3.4×10^{-4} 2.9×10^{-4} 2.6×10^{-4} 3.3×10^{-4} 3.1×10^{-4} 6.2×10^{-4} 5.7×10^{-4} 5.2×10^{-4} 5.0×10^{-4} 1.3×10^{-3} 1.4×10^{-3} 1.3×10^{-3}	Power Calibrator, Calibrator/ SICT-CP-40311
AC Current		(50 ~ 60) Hz 24 mW (24 ~ 72) mW (72 ~ 120) mW (120 ~ 480) mW 480 mW ~ 2.4 W (2.4 ~ 24) W (24 ~ 120) W (120 ~ 240) W (240 ~ 480) W (480 ~ 600) W (600 ~ 960) W 960 W ~ 1.2 kW (1.2 ~ 2.4) kW (2.4 ~ 4.8) kW (4.8 ~ 9.6) kW (9.6 ~ 120) kW	0.42 mW 5.8×10^{-3} 3.5×10^{-3} 2.9×10^{-3} 1.1×10^{-3} 4.9×10^{-4} 2.1×10^{-4} 2.3×10^{-4} 2.6×10^{-4} 2.0×10^{-4} 2.4×10^{-4} 2.0×10^{-4} 3.0×10^{-4} 6.4×10^{-4} 5.8×10^{-4} 1.3×10^{-3}	
AC Wattage		25 Hz (25 ~ 60) Hz (60 ~ 100) Hz (100 ~ 200) Hz (200 ~ 400) Hz (400 ~ 1 000) Hz	9.1 mHz 1.7×10^{-4} 6.0×10^{-4} 3.9×10^{-4} 2.7×10^{-4} 1.3×10^{-3}	
Frequency		(50 ~ 60) Hz 240 mW -1 ~ 1 -0.8, 0.8 -0.5, 0.5 -0.3, 0.3 -0.1, 0.1	3.1×10^{-4} 3.9×10^{-4} 4.6×10^{-4} 5.6×10^{-4} 5.7×10^{-4}	
Power Factor		240 mW ~ 120 kW -1 ~ 1 -0.8, 0.8 -0.5, 0.5 -0.3, 0.3 -0.1, 0.1	2.1×10^{-4} 3.2×10^{-4} 4.6×10^{-4} 4.9×10^{-4} 5.0×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power supplies	AC Voltage	40312 100 mV 40 Hz ~ 1 kHz (100 mV ~ 10 V) 40 Hz ~ 1 kHz (10 ~ 100) V 50 Hz ~ 1 kHz (100 ~ 500) V 40 Hz ~ 5 kHz (100 ~ 1 000) V 40 Hz ~ 5 kHz	2.7 × 10 ⁻⁴ 1.3 × 10 ⁻⁴ 1.4 × 10 ⁻⁴ 1.3 × 10 ⁻⁴ 1.4 × 10 ⁻⁴	Voltage Standard, Multimeter, Current Shunt/ SICT-CP-40312
		10 Hz 10 Hz ~ 5 kHz	1.3 × 10 ⁻⁴ 1.2 × 10 ⁻⁴	
		(50 ~ 60) Hz (1 mA) 1 mA ~ 1 A (1 ~ 5) A (5 ~ 8) A (8 ~ 10) A	3.7 × 10 ⁻³ 2.4 × 10 ⁻³ 2.6 × 10 ⁻³ 2.9 × 10 ⁻³ 2.5 × 10 ⁻³	
	Frequency			
Puncture/safety testers	40313	0.1 kV (0.1 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 2) kV (2 ~ 3) kV (3 ~ 4) kV (4 ~ 5) kV (5 ~ 6) kV (6 ~ 10) kV	2.8 V 1.0 × 10 ⁻² 6.6 × 10 ⁻³ 4.8 × 10 ⁻³ 4.2 × 10 ⁻³ 3.9 × 10 ⁻³ 3.7 × 10 ⁻³ 3.6 × 10 ⁻³ 3.4 × 10 ⁻³	AC/DC Kilovoltmeter, High Voltage Digital Meter, Reference Multimeter/ SICT-CP-40313
		60 Hz 0.1 kV (0.1 ~ 0.5) kV (0.5 ~ 1) kV (1 ~ 2) kV (2 ~ 4) kV (4 ~ 5) kV (5 ~ 10) kV	2.4 V 7.4 × 10 ⁻³ 4.8 × 10 ⁻³ 3.5 × 10 ⁻³ 3.0 × 10 ⁻³ 2.8 × 10 ⁻³ 2.6 × 10 ⁻³	
AC Breaking Current	AC Voltage	60 Hz 0.1 mA (0.1 ~ 0.5) mA (0.5 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 10) mA (10 ~ 20) mA (20 ~ 100) mA	0.67 μA 1.5 × 10 ⁻³ 8.2 × 10 ⁻⁴ 5.3 × 10 ⁻⁴ 8.3 × 10 ⁻⁴ 5.5 × 10 ⁻⁴ 5.3 × 10 ⁻³ 8.3 × 10 ⁻⁴	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Puncture/safety testers				
DC Breaking Current	40313	0.1 mA (0.1 ~ 0.5) mA (0.5 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 10) mA (10 ~ 20) mA (20 ~ 50) mA (50 ~ 100) mA	0.61 µA 1.2×10^{-3} 6.1×10^{-4} 3.1×10^{-4} 1.5×10^{-4} 7.7×10^{-5} 3.2×10^{-4} 1.8×10^{-4} 6.2×10^{-4}	AC/DC Kilovoltmeter, High Voltage Digital Meter, Reference Multimeter/ SICT-CP-40313
Resistance		1 mΩ 1 mΩ ~ 10 mΩ 10 mΩ ~ 100 kΩ	0.84 µΩ 7.0×10^{-4} 6.5×10^{-4}	
Insulation Voltage		1 V (1 ~ 10) V (10 ~ 25) V (25 ~ 50) V (50 ~ 100) V (100 ~ 250) V (250 ~ 500) V (500 ~ 1 000) V (1 000 ~ 2 000) V	0.62 mV 1.2×10^{-4} 2.8×10^{-4} 1.7×10^{-4} 1.2×10^{-4} 2.8×10^{-4} 1.7×10^{-4} 1.2×10^{-4} 6.9×10^{-3}	
Insulation Resistance		1 kΩ (1 ~ 100) kΩ 100 kΩ ~ 1 MΩ (1 ~ 10) MΩ (10 ~ 100) MΩ 100 MΩ ~ 1 GΩ (1 ~ 10) GΩ (10 ~ 100) GΩ 100 GΩ ~ 1 TΩ	0.14 Ω 1.3×10^{-4} 1.0×10^{-4} 9.1×10^{-4} 2.0×10^{-4} 2.6×10^{-4} 9.1×10^{-3} 1.5×10^{-3} 2.8×10^{-3}	
Leakage current(DC)		30 µA (30 ~ 100) µA 100 µA ~ 10 mA (10 ~ 50) mA	29 nA 7.3×10^{-4} 6.4×10^{-4} 2.2×10^{-4}	
Leakage current(AC)		60 Hz 30 µA (30 ~ 100) µA 100 µA ~ 1 mA (1 ~ 10) mA (10 ~ 50) mA	0.16 µA 2.7×10^{-3} 1.5×10^{-3} 9.3×10^{-4} 9.5×10^{-4}	
Timer		1 s (1 ~ 100) s (100 ~ 1 000) s (1 000 ~ 10 000) s	5.8 µs 5.8×10^{-6} 8.2×10^{-6} 5.8×10^{-5}	
Output AC Current		60 Hz 1 A (1 ~ 5) A (5 ~ 10) A (10 ~ 20) A (20 ~ 30) A (30 ~ 50) A (50 ~ 60) A	2.9 mA 2.1×10^{-3} 2.0×10^{-3} 3.0×10^{-3} 4.0×10^{-3} 3.1×10^{-3} 7.8×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Power recorders	40314			Power Energy Calibrator/ SICT-CP-40314
AC Wattage		(50 ~ 60) Hz 240 mW 240 mW ~ 1.2 W (1.2 ~ 60) W 60 W ~ 2.4 kW (2.4 ~ 9.6) kW (9.6 ~ 120) kW	1.4 mW 3.2×10^{-3} 4.9×10^{-4} 3.0×10^{-4} 6.4×10^{-4} 3.5×10^{-3}	
AC voltmeters	40318	1 mV 50 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz (1 ~ 3) mV 50 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz (3 ~ 10) mV 50 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz (10 ~ 30) mV 50 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz (30 ~ 100) mV 50 Hz ~ 10 kHz (10 ~ 100) kHz (100 ~ 300) mV 50 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz (300 mV ~ 1 V) 50 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz (1 ~ 3) V 50 Hz ~ 10 kHz (10 ~ 100) kHz (3 ~ 10) V 50 Hz ~ 1 kHz (1 ~ 10) kHz (10 ~ 100) kHz	7.4×10^{-3} 8.7×10^{-3} 3.2×10^{-2} 2.6×10^{-3} 3.0×10^{-3} 1.2×10^{-2} 9.6×10^{-4} 1.0×10^{-3} 6.1×10^{-3} 4.3×10^{-4} 4.4×10^{-4} 4.6×10^{-3} 2.8×10^{-4} 1.3×10^{-3} 2.0×10^{-4} 2.4×10^{-4} 1.4×10^{-3} 2.5×10^{-4} 2.6×10^{-4} 1.0×10^{-3} 2.1×10^{-4} 1.0×10^{-3} 2.4×10^{-4} 2.5×10^{-4} 1.2×10^{-3}	Reference Multimeter, Calibrator/ SICT-CP-40318 SICT-CP-40318 SICT-CP-40318 SICT-CP-40318

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltmeters	AC Voltage	40318	(10 ~ 30) V 50 Hz ~ 1 kHz (1 ~ 10) kHz 50 Hz ~ 100 kHz	2.0×10^{-4} 2.2×10^{-4} 1.2×10^{-3}
			(30 ~ 100) V 50 Hz ~ 1 kHz (1 ~ 10) kHz 50 Hz ~ 100 kHz	2.4×10^{-4} 2.5×10^{-4} 2.9×10^{-3}
			(100 ~ 1 000) V 50 Hz ~ 1 kHz	3.6×10^{-4}
			(50 Hz ~ 1 kHz) 1 mV (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 1 V	$5.4 \mu\text{V}$ 7.8×10^{-4} 2.0×10^{-4} 1.1×10^{-4}
			1 mV (1 ~ 10) mV (10 ~ 100) mV 100 mV ~ 1 V	$0.86 \mu\text{V}$ 8.7×10^{-5} 1.4×10^{-5} 9.0×10^{-5}
	AC Output Voltage	40410	10 Hz (10 ~ 60) Hz (60 ~ 400) Hz 400 Hz ~ 1 kHz	6.1 mHz 1.5×10^{-4} 1.2×10^{-4} 1.4×10^{-4}
			(±)	
			0 μA (0 ~ 5) μA (5 ~ 10) μA (10 ~ 30) μA (30 ~ 50) μA (50 ~ 100) μA (100 ~ 200) μA (200 ~ 500) μA (0.5 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 10) mA (10 ~ 15) mA (15 ~ 20) mA	9 nA 1.9×10^{-3} 9.9×10^{-4} 3.7×10^{-4} 2.4×10^{-4} 1.6×10^{-4} 1.3×10^{-4} 8.1×10^{-5} 6.1×10^{-4} 3.2×10^{-4} 1.5×10^{-4} 6.1×10^{-4} 4.1×10^{-4} 3.1×10^{-4}
			(30 μA) 50 Hz ~ 1 kHz	$0.16 \mu\text{A}$
Leakage current testers	DC Current	40416	(30 ~ 50) μA 50 Hz ~ 1 kHz	3.9×10^{-3}
			(50 ~ 100) μA 50 Hz ~ 1 kHz	2.7×10^{-3}
			(30 μA) 50 Hz ~ 1 kHz	
			(30 ~ 50) μA 50 Hz ~ 1 kHz	
			(50 ~ 100) μA 50 Hz ~ 1 kHz	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers				
AC Current	40416	(100 ~ 200) μ A 50 Hz ~ 1 kHz	2.1×10^{-3}	Calibrator/ SICT-CP-40416
		(200 ~ 500) μ A 50 Hz ~ 1 kHz	1.6×10^{-3}	
		(0.5 ~ 1) mA 50 Hz ~ 1 kHz	6.9×10^{-4}	
		(1 ~ 2) mA 50 Hz ~ 1 kHz	4.7×10^{-4}	
		(2 ~ 5) mA 50 Hz ~ 1 kHz	3.0×10^{-3}	
		(5 ~ 10) mA 50 Hz ~ 1 kHz	8.2×10^{-4}	
		(10 ~ 20) mA 50 Hz ~ 1 kHz	5.3×10^{-4}	
		(20 ~ 30) mA 50 Hz ~ 1 kHz	1.2×10^{-4}	
DC Voltage		0 V	61 μ V	
		(0 ~ 0.1) V	6.1×10^{-4}	
		(0.1 ~ 0.2) V	3.1×10^{-4}	
		(0.2 ~ 0.5) V	1.2×10^{-4}	
		(0.5 ~ 1) V	6.1×10^{-4}	
		(1 ~ 2) V	3.1×10^{-4}	
		(2 ~ 5) V	1.2×10^{-4}	
		(5 ~ 10) V	6.2×10^{-5}	
		(10 ~ 20) V	3.3×10^{-5}	
		(20 ~ 50) V	1.7×10^{-5}	
		(50 ~ 100) V	1.2×10^{-5}	
		(100 ~ 200) V	3.4×10^{-5}	
		(200 ~ 300) V	2.5×10^{-5}	
		(300 ~ 500) V	1.2×10^{-4}	
		(500 ~ 1 000) V	6.2×10^{-5}	
AC Voltage		(0.1 V) 50 Hz ~ 1 kHz	6.5×10^{-4}	
		(1 ~ 20) kHz	6.6×10^{-4}	
		(20 ~ 50) kHz	7.8×10^{-4}	
		(50 ~ 100) kHz	1.4×10^{-3}	
		(0.1 ~ 1) V 50 Hz ~ 1 kHz	1.2×10^{-4}	
		(1 ~ 20) kHz	1.3×10^{-4}	
		(20 ~ 50) kHz	1.8×10^{-4}	
		(50 ~ 100) kHz	3.8×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Leakage current testers	40416	(1 ~ 10) V		Calibrator/ SICT-CP-40416	
		50 Hz ~ 1 kHz	1.2×10^{-4}		
		(1 ~ 20) kHz	1.3×10^{-4}		
		(20 ~ 50) kHz	1.8×10^{-4}		
		(50 ~ 100) kHz	3.4×10^{-4}		
		(10 ~ 100) V			
		50 Hz ~ 1 kHz	1.3×10^{-4}		
		(1 ~ 20) kHz	1.2×10^{-4}		
		(20 ~ 50) kHz	3.0×10^{-4}		
		(50 ~ 100) kHz	6.8×10^{-4}		
Output Resistance		(100 ~ 300) V			
		50 Hz ~ 1 kHz	3.1×10^{-4}		
		(1 ~ 10) kHz	3.4×10^{-4}		
Input Resistance		(10 ~ 20) kHz	5.5×10^{-4}		
		(300 ~ 1 000) V			
		50 Hz ~ 10 kHz	3.7×10^{-4}		
Electronic AC/DC loads	40417	100 mΩ	$8.6 \mu\Omega$	Calibrator/ SICT-CP-40417	
		100 mΩ ~ 1 Ω	6.3×10^{-5}		
		1 Ω ~ 10 kΩ	6.2×10^{-5}		
DC Voltage		100 mΩ	$24 \mu\Omega$		
		100 mΩ ~ 100 Ω	2.4×10^{-4}		
		100 Ω ~ 10 kΩ	6.3×10^{-5}		
		100 mV	$6.3 \mu V$		
		(0.1 ~ 1) V	6.2×10^{-5}		
		(1 ~ 2) V	3.2×10^{-5}		
		(2 ~ 4) V	2.3×10^{-5}		
		(4 ~ 6) V	1.5×10^{-5}		
		(6 ~ 8) V	1.2×10^{-5}		
		(8 ~ 10) V	1.1×10^{-5}		
		(10 ~ 50) V	3.3×10^{-5}		
		(50 ~ 100) V	1.2×10^{-5}		
DC Current		(100 ~ 200) V	3.6×10^{-5}		
		(200 ~ 400) V	2.6×10^{-5}		
		(400 ~ 1 000) V	1.7×10^{-5}		
		100 mA	$44 \mu A$		
		100 mA ~ 0.2 A	4.8×10^{-4}		
		(0.2 ~ 0.4) A	4.2×10^{-4}		
		(0.4 ~ 0.6) A	3.8×10^{-4}		
		(0.6 ~ 2) A	3.7×10^{-4}		

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electronic AC/DC loads	AC Voltage	40417	(50 ~ 60) Hz	Calibrator/ SICT-CP-40417
			0.1 V	
			(0.1 ~ 0.2) V	
			(0.2 ~ 0.5) V	
			(0.5 ~ 1) V	
			(1 ~ 10) V	
			(10 ~ 100) V	
			(100 ~ 200) V	
			(200 ~ 500) V	
	AC Current		(50 ~ 60) Hz	
			100 mA	
			100 mA ~ 0.3 A	
			(0.3 ~ 0.4) A	
			(0.4 ~ 0.5) A	
			(0.5 ~ 0.6) A	
			(0.6 ~ 0.8) A	
			(0.8 ~ 2) A	
			(2 ~ 4) A	
			(4 ~ 5) A	
			(5 ~ 7) A	
			(7 ~ 9) A	
			(9 ~ 10) A	
			(10 ~ 13) A	
			(13 ~ 16) A	
			(16 ~ 20) A	
Analogue/Digital multimeters	DC Voltage	40419	(±)	Calibrator/ SICT-CP-40419
			0 mV	
			(0 ~ 1) mV	
			(1 ~ 2) mV	
			(2 ~ 5) mV	
			(5 ~ 10) mV	
			(10 ~ 20) mV	
			(20 ~ 50) mV	
			(50 ~ 100) mV	
			100 mV ~ 0.2 V	
			(0.2 ~ 0.5) V	
			(0.5 ~ 1) V	
			(1 ~ 2) V	
			(2 ~ 5) V	
			(5 ~ 10) V	
			(10 ~ 20) V	
			(20 ~ 50) V	
			(50 ~ 100) V	
			(100 ~ 200) V	
			(200 ~ 500) V	
			(500 ~ 1 000) V	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters	40419	(3 mV)		Calibrator/ SICT-CP-40419
AC Voltage		50 Hz ~ 1 kHz	10 μ V	
		(3 ~ 10) mV		
		50 Hz ~ 10 kHz	9.0×10^{-4}	
		(10 ~ 20) kHz	1.3×10^{-3}	
		(20 ~ 50) kHz	1.3×10^{-3}	
		(50 ~ 100) kHz	2.1×10^{-3}	
		(10 ~ 100) mV		
		50 Hz ~ 10 kHz	2.5×10^{-4}	
		(10 ~ 20) kHz	2.7×10^{-4}	
		(20 ~ 50) kHz	5.4×10^{-4}	
		(50 ~ 100) kHz	1.4×10^{-3}	
		(0.1 ~ 1) V		
		50 Hz ~ 10 kHz	1.1×10^{-4}	
		(10 ~ 20) kHz	1.3×10^{-4}	
		(20 ~ 50) kHz	2.0×10^{-4}	
		(50 ~ 100) kHz	4.2×10^{-4}	
		(1 ~ 10) V		
		50 Hz ~ 10 kHz	1.1×10^{-4}	
		(10 ~ 20) kHz	1.3×10^{-4}	
		(20 ~ 50) kHz	2.0×10^{-4}	
		(50 ~ 100) kHz	3.8×10^{-4}	
		(10 ~ 100) V		
		50 Hz ~ 20 kHz	1.2×10^{-4}	
		(20 ~ 50) kHz	3.4×10^{-4}	
		(50 ~ 100) kHz	8.1×10^{-4}	
		(100 ~ 300) V		
		50 Hz ~ 1 kHz	2.3×10^{-4}	
		(1 ~ 10) kHz	2.7×10^{-4}	
		(10 ~ 20) kHz	5.1×10^{-4}	
		(300 ~ 1 000) V		
		50 Hz ~ 10 kHz	3.6×10^{-4}	
Resistance		0 Ω	85 μ Ω	
		(0 ~ 1) Ω	1.3×10^{-4}	
		(1 ~ 10) Ω	3.4×10^{-5}	
		(10 ~ 100) Ω	2.0×10^{-5}	
		100 Ω ~ 1 k Ω	1.6×10^{-5}	
		(1 ~ 10) k Ω	1.4×10^{-5}	
		(10 ~ 100) k Ω	1.7×10^{-5}	
		100 k Ω ~ 1 M Ω	2.4×10^{-5}	
		(1 ~ 10) M Ω	5.0×10^{-5}	
		(10 ~ 100) M Ω	1.3×10^{-4}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters	40419	(±)		Calibrator/ SICT-CP-40419
		10 μA	10 nA	
		(10 ~ 20) μA	5.2×10^{-4}	
		(20 ~ 50) μA	2.5×10^{-4}	
		(50 ~ 100) μA	1.5×10^{-4}	
		(0.1 ~ 0.2) mA	1.0×10^{-4}	
		(0.2 ~ 0.5) mA	8.1×10^{-5}	
		(0.5 ~ 1) mA	6.8×10^{-5}	
		(1 ~ 2) mA	9.0×10^{-5}	
		(2 ~ 5) mA	8.1×10^{-5}	
		(5 ~ 10) mA	6.8×10^{-5}	
		(10 ~ 20) mA	9.0×10^{-5}	
		(20 ~ 50) mA	9.2×10^{-5}	
		(50 ~ 100) mA	8.0×10^{-5}	
		(0.1 ~ 0.2) A	9.8×10^{-5}	
		(0.2 ~ 0.5) A	1.5×10^{-4}	
		(0.5 ~ 1) A	1.2×10^{-4}	
		(1 ~ 2) A	4.6×10^{-4}	
		(2 ~ 5) A	7.4×10^{-4}	
		(5 ~ 10) A	6.5×10^{-4}	
		(10 ~ 20) A	1.2×10^{-3}	
AC Current	AC Current	(30 μA)		
		50 Hz ~ 1 kHz	0.16 μA	
		(30 ~ 50) μA		
		50 Hz ~ 1 kHz	3.9×10^{-3}	
		(50 ~ 100) μA		
		50 Hz ~ 1 kHz	2.7×10^{-3}	
		(0.1 ~ 0.2) mA		
		50 Hz ~ 1 kHz	3.6×10^{-5}	
		(0.2 ~ 0.5) mA		
		50 Hz ~ 1 kHz	4.9×10^{-5}	
		(0.5 ~ 1) mA		
		50 Hz ~ 1 kHz	2.4×10^{-4}	
		(1 ~ 2) mA		
		50 Hz ~ 1 kHz	3.0×10^{-4}	
		(2 ~ 5) mA		
		50 Hz ~ 1 kHz	1.4×10^{-4}	
		(5 ~ 10) mA		
		50 Hz ~ 1 kHz	7.2×10^{-5}	
		(10 ~ 20) mA		
		50 Hz ~ 1 kHz	3.0×10^{-4}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/Digital multimeters				
AC Current	40419	(20 ~ 50) mA 50 Hz ~ 1 kHz	1.4×10^{-4}	Calibrator/ SICT-CP-40419
		(50 ~ 100) mA 50 Hz ~ 1 kHz	7.2×10^{-5}	
		100 mA ~ 0.2 A 50 Hz ~ 1 kHz	4.5×10^{-4}	
		(0.2 ~ 0.5) A 50 Hz ~ 1 kHz	2.6×10^{-4}	
		(0.5 ~ 1) A 50 Hz ~ 1 kHz	1.3×10^{-4}	
		(1 ~ 2) A 50 Hz ~ 1 kHz	7.6×10^{-4}	
		(2 ~ 5) A 50 Hz ~ 1 kHz	1.7×10^{-3}	
		(5 ~ 10) A 50 Hz ~ 1 kHz	1.4×10^{-3}	
		(10 ~ 15) A 50 Hz ~ 1 kHz	2.1×10^{-3}	
		(15 ~ 20) A 50 Hz ~ 1 kHz	2.0×10^{-3}	
Frequency		10 Hz ~ 1 MHz	6.1×10^{-7}	
Oscilloscopes				
Impedance Measure	40421	50 Ω 75 Ω 1 MΩ	0.7 mΩ 0.8 mΩ 13 Ω	Calibration Generator/ SICT-CP-40421
DC Voltage		1 mV (1 ~ 2) mV (2 ~ 5) mV (5 ~ 10) mV (10 ~ 20) mV (20 ~ 50) mV (50 ~ 100) mV 100 mV ~ 0.5 V (0.5 ~ 1) V (1 ~ 2) V (2 ~ 5) V (5 ~ 10) V (10 ~ 50) V (50 ~ 200) V	1.0 μV 4.9×10^{-4} 2.0×10^{-4} 1.0×10^{-4} 5.5×10^{-5} 2.7×10^{-5} 1.8×10^{-5} 1.6×10^{-5} 9.4×10^{-6} 1.1×10^{-5} 9.2×10^{-6} 8.6×10^{-6} 1.2×10^{-5} 1.5×10^{-5}	

404. Other DC & LF measurements

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Oscilloscopes	40421	(1 kHz) 1 mV (1 ~ 2) mV (2 ~ 5) mV (5 ~ 10) mV (10 ~ 20) mV (20 ~ 50) mV (50 ~ 100) mV (100 ~ 200) mV (0.2 ~ 0.5) V (0.5 ~ 120) V	47 μ V 2.4×10^{-2} 1.0×10^{-2} 5.8×10^{-3} 3.5×10^{-3} 2.1×10^{-3} 1.6×10^{-3} 1.4×10^{-3} 1.3×10^{-3} 1.2×10^{-3}	Calibration Generator/ SICT-CP-40421
AC Voltage(Square wave)		Time Marker 5 ns 5 ns ~ 10 ms (10 ~ 20) ms 20 ms ~ 5 s	16 fs 3.0×10^{-6} 3.3×10^{-6} 3.3×10^{-5}	
CAL Output Amplitude		(50 Hz ~ 10 kHz) 100 mV (100 ~ 200) mV 200 mV ~ 1.2 V (1.2 ~ 2) V (2 ~ 4) V (4 ~ 12) V	36 μ V 2.2×10^{-4} 1.7×10^{-4} 1.4×10^{-4} 1.2×10^{-4} 1.7×10^{-4}	
CAL Output Frequency		100 Hz ~ 10 MHz	6.5×10^{-7}	
Sine Wave Signal Generator Level		(600 mV) 50 kHz (0.05 ~ 100) MHz (100 ~ 500) MHz (0.5 ~ 1) GHz (1 ~ 3) GHz	25 mV 4.5×10^{-2} 7.2×10^{-2} 1.9×10^{-2} 2.3×10^{-2}	
Vol/Current recorders	40424	DC Voltage (±) 0 μ V 0 μ V ~ 10 mV (10 ~ 100) mV 100 mV ~ 1 V (1 ~ 10) V (10 ~ 100) V (100 ~ 500) V (500 ~ 1 000) V	0.70 μ V 8.1×10^{-5} 1.6×10^{-5} 9.4×10^{-6} 8.6×10^{-6} 1.0×10^{-5} 1.2×10^{-5} 1.1×10^{-5}	Calibrator/ SICT-CP-40424
		DC Current (±) 0 μ A (0 ~ 10) μ A (10 ~ 100) μ A 100 μ A ~ 10 mA 10 mA ~ 1 A (1 ~ 10) A (10 ~ 20) A	9.3 nA 9.9×10^{-4} 1.6×10^{-4} 9.2×10^{-5} 2.0×10^{-4} 6.5×10^{-4} 1.2×10^{-3}	

501. Contact thermometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Temperature generators: ovens, furnaces, isothermal liquid baths, ice-point baths, dry-block calibrators	50101	0 °C (-90 ~ 250) °C (250 ~ 500) °C (500 ~ 650) °C (650 ~ 700) °C (700 ~ 1 300) °C	0.01 °C 0.020 °C 0.024 °C 0.028 °C 1.3 °C 2.6 °C	SPRT, STANDARD TC/ SICT-CP-50101
Temperature indicators/recorders /controllers, temperature calibrators	50102			SPRT, STANDARD TC/ SICT-CP-50102
Temperature indicators/recorders/controllers (With Sensor)		(-45 ~ 0) °C (0 ~ 250) °C (250 ~ 650) °C (650 ~ 900) °C (900 ~ 1 100) °C (1 100 ~ 1 300) °C	0.024 °C 0.031 °C 0.13 °C 1.4 °C 1.5 °C 2.6 °C	
Temperature indicators/recorders/controllers (Without Sensor)		(-45 ~ 0) °C (0 ~ 100) °C (100 ~ 200) °C (200 ~ 300) °C (300 ~ 400) °C (400 ~ 500) °C (500 ~ 600) °C (600 ~ 700) °C (700 ~ 800) °C (800 ~ 1 100) °C (1 100 ~ 1 300) °C	0.013 °C 0.018 °C 0.022 °C 0.027 °C 0.031 °C 0.035 °C 0.048 °C 0.052 °C 0.057 °C 0.08 °C 0.09 °C	
Glass thermometers; liquid-in-glass, Beckmann	50103			SPRT/ SICT-CP-50103
liquid-in-glass		(-45 ~ 0) °C (0 ~ 100) °C (100 ~ 200) °C	0.048 °C 0.058 °C 0.062 °C	
Resistance thermometers; SPRT, IPRT, thermistors, etc.	50104			SPRT, Fixed point/ SICT-CP-50104
IPRT		(-45 ~ 50) °C (50 ~ 250) °C	0.024 °C 0.028 °C	
Thermal expansion thermometers; bimetal, gas or liquid type	50105			SPRT/ SICT-CP-50105
bimetal		(-45 ~ 100) °C (100 ~ 300) °C (300 ~ 400) °C (400 ~ 650) °C	0.4 °C 0.6 °C 1.5 °C 3.2 °C	

501. Contact thermometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Thermocouples: noble metal, base metal, pure metal, special type, etc.	50106			SPRT, Fixed point, STANDARD TC/ SICT-CP-50106
Base metal		(-45 ~ 300) °C	0.4 °C	
		(300 ~ 400) °C	0.6 °C	
		(400 ~ 500) °C	0.7 °C	
		(500 ~ 650) °C	0.8 °C	
		(650 ~ 1 000) °C	1.8 °C	
		(1 000 ~ 1 100) °C	2.0 °C	
		(1 100 ~ 1 200) °C	3.1 °C	
		(1 200 ~ 1 300) °C	3.6 °C	
Temperature transducers	50107	(-45 ~ 100) °C	0.05 °C	SPRT, THERMOCOUPLE, MULTIMETER SICT-CP-50107
		(100 ~ 200) °C	0.06 °C	
		(200 ~ 250) °C	0.09 °C	
		(250 ~ 300) °C	0.14 °C	
		(300 ~ 500) °C	0.16 °C	
		(500 ~ 650) °C	0.18 °C	
		(650 ~ 700) °C	1.4 °C	
		(700 ~ 800) °C	1.6 °C	
		(800 ~ 1 000) °C	1.7 °C	
		(1 000 ~ 1 100) °C	2.8 °C	
		(1 100 ~ 1 300) °C	3.0 °C	

503. Humidity

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Relative humidity hygrometers; polimer thinfilm, hair, etc.	50302			Dewpoint Meter/ SICT-CP-50302
humidity		(20 ~ 30) % R.H. (30 ~ 50) % R.H. (50 ~ 60) % R.H. (60 ~ 70) % R.H. (70 ~ 95) % R.H.	1.9 % R.H. 1.6 % R.H. 1.7 % R.H. 1.9 % R.H. 2.1 % R.H.	
Temperature		(-20 ~ 50) °C (50 ~ 80) °C (80 ~ 100) °C	0.7 °C 0.9 °C 1.2 °C	
Temperature humidity recorders; Hygrothermograph, etc	50304			Dewpoint Meter/ SICT-CP-50304
Humidity		(20 ~ 30) % R.H. (30 ~ 95) % R.H.	2.4 % R.H. 2.6 % R.H.	
Temperature		(-20 ~ 50) °C	1.1 °C	
Transducers; dew-point /relative humidity	50305			Dewpoint Meter/ SICT-CP-50305
Humidity		(20 ~ 30) % R.H. (30 ~ 40) % R.H. (40 ~ 50) % R.H. (50 ~ 60) % R.H. (60 ~ 70) % R.H. (70 ~ 80) % R.H. (80 ~ 90) % R.H. (90 ~ 95) % R.H.	1.9 % R.H. 1.6 % R.H. 1.7 % R.H. 1.8 % R.H. 1.9 % R.H. 2.0 % R.H. 2.1 % R.H. 2.2 % R.H.	
Temperature		(-20 ~ 50) °C (50 ~ 80) °C (80 ~ 100) °C	0.7 °C 0.9 °C 1.2 °C	
Humidity generators; two-pressure, two-temperature, flow mixing humidity generator, constant temperature and humidity chamber, etc.	50306			Dewpoint Meter/ SICT-CP-50306
ssure/two-temperature/flow mixing humidity (humidity)		(10 ~ 40) % R.H. (40 ~ 50) % R.H. (50 ~ 60) % R.H. (60 ~ 70) % R.H. (70 ~ 80) % R.H. (80 ~ 90) % R.H. (90 ~ 95) % R.H.	1.8 % R.H. 1.6 % R.H. 1.8 % R.H. 1.9 % R.H. 2.1 % R.H. 2.3 % R.H. 2.4 % R.H.	
(Temperature)		(-75 ~ 200) °C	0.4 °C	