

CERTIFICATE OF ACCREDITATION

Korea Testing Instrument Calibration Center

Accreditation No. : KC00-005

Corporation Registration No. : 110111-2326216

Address of Laboratory : (Guro-dong, Samsung IT-valy), 708, 27, Digital-ro 33-gil,
Guro-gu, Seoul, Republic of Korea

Date of Initial Accreditation : October 05, 2000.

Validity of Accreditation : October 28, 2021. ~ October 27, 2025.

Scope of Accreditation : Attached Annex

Date of issue : November 02, 2023.

This calibration laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to Joint ISO-ILAC-IAF Communiqué).



CHIN CHONGWOOK

Head

Korea Laboratory Accreditation Scheme

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017 & KS Q ISO/IEC 17025:2017

Korea Testing Instrument Calibration Center
(Guro-dong, Samsung IT Valley), 708, 27, Digital-ro33-gil, Guro-gu, Seoul, Korea
Phone : 02-717-5400, FAX : 02-2028-1515, e-mail : deokseop.kim@trescal.com

CALIBRATION

Valid No : Oct. 27, 2025.

Accreditation No. : KC00-005

In recognition of the successful completion of the KOLAS evaluation process,
accreditation is granted to this laboratory to perform the following calibrations

Field Code	Item of Calibration	on-site	Field Code	Item of Calibration	on-site	Field Code	Item of Calibration	on-site
102. Linear dimension			10511	Measuring microscopes, Profile projectors	Y	206. Volume		
10201	Balls	N				20601	Volumetric glasswares	N
10206	Dial/ cylinder gauge testers	N	10512	Micro measuring microscopes	Y	20602	Pycnometers	N
10207	Doctor blades	Y				20605	Concrete air content meters	N
10209	End bars	N	10527	Thread ring gauges	N	20606	Piston type volume meters	N
10210	Extensometers, linear displacement transducers	Y	106. Various dimensional			301. Time/frequency		
			10601	Inside/ outside/ gear tooth calipers, caliper gauges	Y	30102	Frequency standards	N
						30103	General frequency sources	Y
10211	Filler gauges	Y				30104	Frequency meters/ counters	Y
10212	Film applicators	Y						
10213	Gap gauges	Y	10603	Cylinder/bore gauges	Y	30105	Time interval sources	Y
10214	Gauge blocks, by comparison	N	10604	Depth gauges, depth micrometers	Y	30106	Time interval meters/ stop watches/timers	Y
10216	Height gauges /measuring machines	Y	10605	Dial/digital gauges	Y	302. Velocity & revolution		
			10608	Grind gauges	Y	30201	Standard RPM generators	Y
			10609	Micro indicators, test indicators	Y	30202	Contact type tachometers	Y
10223	Electronic micrometers	N	10610	Micrometer heads	Y	30203	Photo tachometers/ stroboscopes	Y
10224	Height micrometers, riser blocks	N	10611	3-points micrometers	Y	401. DC voltage & current		
			10612	Inside micrometers	Y	40101	DC ammeters	Y
			10613	Outside micrometers	Y	40102	Transconductance amplifiers	Y
10227	Standard tape rules, peripheral gauges	N	10617	Standard sieves	N	40103	DC voltage/current calibrators	Y
10228	Cylindrical plug/ pin gauges, thread measuring wire gauges	N	10620	Welding gauges	N			
			201. Mass			40105	DC current shunts	Y
			20106	Dial platform scale balances	Y	40106	Galvanometers null detectors	Y
10229	Radius gauges	N	20109	Electric balances	Y	40107	Potentiometers	Y
10230	Cylindrical ring gauges	N	20112	Platform scale balances	Y			
10232	Step gauges	N	20113	Spring scale balances	Y	40111	DC voltage standards	N
10233	Taper thickness gauges	N	20116	Weights	N	40112	DC voltmeters	Y
10234	Ultrasonic thickness gauges	Y	202. Force			40113	Static/ionic voltmeters	N
10235	Ultrasonic/ coating thickness specimens	Y	20203	Tension/compression testing machines	Y	402. Resistance, capacitance and inductance		
			20204	Push-pull gauges	N	40201	Capacitance bridges/ indicators	Y
203. Torque			20302	Torque measuring devices	N			
10236	Coating thickness testers	Y	20303	Torque wrenches/drivers	Y	40204	Standard capacitors	Y
104. Form			204. Pressure			40205	Earth testers	Y
10404	Optical flats	N	20402	Manometers	Y	40206	Inductance bridges/ indicators	Y
10406	Parallel blocks	Y	20406	Absolute pressure gauges	Y	40208	Inductors	Y
10407	Precision surface plates	Y	20407	Blood pressure gauges	Y			
10412	Straight edges	Y	20408	Compound pressure gauges	Y	40213	Resistance bridges & similar instruments	Y
10413	Straight rules	N	20409	Differential pressure gauges	Y			
105. Complex geometry			20411	Gauge pressure gauges	Y			
10503	Contact coordinate measuring machines	Y	20412	Pressure transducers /transmitters	Y			
10504	Non-contact coordinate measuring machines	Y	20413	Dial type vacuum gauges	Y			
			20414	Water depth meters	Y			

Field Code	Item of Calibration	on-site	Field Code	Item of Calibration	on-site	Field Code	Item of Calibration	on-site	
40214	Resistance meters	Y	40602	Coaxial attenuators	Y	50102	Temperature indicators/ recorders/ controllers, temperature calibrators	Y	
40215	Resistors	Y	40604	BER(Bit Error Rate) testers	Y				
40217	Impedance bridges/ LCR meters	Y	40605	Burst pulse generators	Y				
			40607	RF power meter calibrators	Y				
403. AC voltage, current & power			40608	EMC transduces; current probes, absorbing clamps, etc.	Y	50103	Glass thermometers; liquid-in-glass, Beckmann	N	
40301	AC ammeters	Y			40610	Coaxial directional couplers/splitters	Y	50104	Resistance thermometers; SPRT, IPRT, thermistors, etc.
40302	Clamp ammeters/ voltmeters	Y	40612	DS1/DS3 communications systems			Y	50105	Thermal expansion thermometers; bimetal, gas or liquid type
40303	AC voltage/ current calibrators	Y			40613	Electrostatic discharge generators	N		
	AC current shunts	Y	40614	EMC receivers	Y	50106	Thermocouples; noble metal, base metal pure metal, special type, etc.	N	
40307	Voltage/current phase angle meters /synchro resolve meters	Y			40615				RF filters
40310	Power factor meters	Y	40616	RF impedance meters	Y				
40311	AC power meters	Y	40617	RF impulse generators	Y				
40312	AC power supplies	Y	40618	Line impedance stabilization networks; LISN, CDN, ISN, etc.	Y	50107	Temperature transducers	Y	
40313	Puncture/ safety testers	Y			50109	Others; quartz, semiconductor, optical fiber, etc.	Y		
40314	Power recorders	Y	40619	Coaxial standard mismatches			Y		
40318	AC voltmeters	Y							
404. Other DC & LF measurements			40621	Mobile communication test sets	Y	502. Non contact thermometry			
40401	LF amplifiers	Y			50204	Standard radiation thermometers	N		
40402	DC/LF attenuators	Y	40622	Modulation meters	Y	50205	Thermal image apparatus	N	
40403	Multimeter calibrators	Y	40623	Network analyzers	Y	50206	Blackbody furnaces	Y	
40404	Oscilloscope calibrators	Y	40624	Noise figure meters	Y	50207	Others; ear thermometers, etc.	N	
40406	Video signal generators	Y	40625	Noise generators	Y				
40407	Audio distortion analyzers/meters	Y	40626	Noise impulse simulators	Y				
40408	LF filters	Y	40628	Coaxial noise sources	Y	503. Humidity			
40409	LF/audio signal analyzers	Y	40631	RF phase meters	Y	50302	Relative humidity hygrometers; polimer thin film, hair, etc.	Y	
40410	Line frequency meters	Y	40635	RF power meters	Y				
40411	Function generators	Y	40636	Diode power sensors	Y	50304	Temperature humidity recorders; Hygrothermograph, etc.	N	
40413	AC/DC high voltage voltmeters	Y	40637	Thermocouple power sensors	Y				
40414	LF impulse generators	Y	40638	Pulse generators	Y	50305	Transducers; dew-point/ relative humidity	N	
40416	Leakage current testers	Y	40639	Radar test sets	Y				
40417	Electronic AC/DC loads	Y	40640	RF signal generators	Y	50306	Humidity generators; two-pressure, two-temperature, flow mixing humidity generator, constant temperature and humidity chamber, etc.	Y	
40418	Modulation meters	Y	40641	RF spectrum analyzers	Y				
40419	Analogue/ digital multimeters	Y	40643	Surge generators	Y	601. Sound in air	60106	Sound level meters	Y
40420	Noise meters	Y	40644	SWR meters	Y				
40421	Oscilloscopes	Y	40645	RF terminations	Y				
40422	LF phase meters	Y	40646	Coaxial thermistor mounts	Y				
40423	Random wave generators	Y	40648	Transmission trouble testers	Y	603. Vibration			
40424	Voltage/current recorders	Y			40650	RF voltmeters	Y		
40425	Relay test sets	Y	40651	Vector voltmeters	Y	60301	Vibration calibrators	N	
40426	LF signal generators	Y	40652	Field strength meters	Y	60302	Vibration transducers	N	
40427	LF spectrum analyzers	Y	40653	AM/FM test sources	Y	60303	Vibration measuring instruments	N	
40429	Sweep generators	Y	40654	Dip simulators	Y				
40430	Signal transducers	Y	407. Field strength & antenna			701. Photometry			
40432	Transistor curve tracers	Y	501. Contact thermometry			704. Fiber optics			
40433	Waveform analyzers	Y	40704	Loop antennas	N	70101	Illuminance meters	N	
40434	AC/DC high voltage generators	Y	40705	Monopole antennas	N	70402	Broadband light sources	Y	
40435	AC/DC high voltage probes	Y	50101	Temperature generators; ovens, furnaces, isothermal liquid baths, ice-point baths, dry-block calibrators	Y	70410	Optical attenuators	Y	
40436	Logic analyzers	Y							
40437	Telephone testers	Y							
40438	Video signal analyzers	Y							
406. Radio frequency measurement									
40601	RF amplifiers	Y							

Field Code	Item of Calibration	on-site	Field Code	Item of Calibration	on-site	Field Code	Item of Calibration	on-site
70411	Optical couplers	Y						
70412	Fiber-optic power meters	Y						
70413	Optical loss testers	Y						
70415	Optical multimeters	Y						
70416	Optical network analyzers	Y						
70417	Optical spectrum analyzers	Y						
70418	Optical time domain reflectometers; OTDR	Y						
70419	PDH/SDH analyzers	Y						
70423	Return loss meters	Y						
70424	SDH/SONET analyzers	Y						
70426	Multi-laser wavelength meters	Y						
70427	Wavelength sweep multichannel measuring systems	Y						
70429	Frequency stabilized lasers and LDs	Y						
70430	ASE light sources	Y						
70431	CW-laser wavelength meters	Y						
70433	Optical power stabilized lasers and LDs	Y						
901. Chemical analysis								
90103	Gas analyzers	N						
		</						

Note

1. This laboratory provides calibration services in permanent standard laboratory and at on-site.
2. Laboratory conducts on-site calibration should meet requirements of KOLAS-SR-007.
3. On-site calibration is allowed to items with marking 'Y', not allowed to items with marking 'N'.
4. Measurement uncertainty normally is quoted as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k=2$. It expresses the lowest uncertainty of measurement that can be provided by accredited calibration laboratories in normal conditions.
5. Due to the calibration environment such as reference standards or customers' facilities, it is note that uncertainty of measurement on a calibration certificate may be expressed larger than measurement uncertainty on scope of accreditation in general.

102. Linear dimension

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Balls	10201	(0 ~ 100) mm	$\sqrt{(0.60 \mu\text{m})^2 + (4.2 \times 10^{-6} \times l)^2}$	Standard measuring machine /KTICC-CI-10201
Dial/cylinder gauge testers	10206	(0 ~ 100) mm	$\sqrt{(0.65 \mu\text{m})^2 + (2 \times 10^{-6} \times l)^2}$	Gauge block /KTICC-CI-10206
Doctor blades	10207	(0 ~ 10) mm	2.6 μm	Electronic micrometer /KTICC-CI-10207
End bars	10209	(0 ~ 100) mm	$\sqrt{(1.2 \mu\text{m})^2 + (2 \times 10^{-6} \times l)^2}$	Gauge block/KTICC-CI-10209
Extensometers, linear displacement transducers	10210	(0 ~ 500) mm	$\sqrt{(1.3 \mu\text{m})^2 + (2 \times 10^{-6} \times l)^2}$	Gauge block/KTICC-CI-10210
Filler gauges	10211	(0 ~ 5) mm	1.4 μm	Outside micrometer /KTICC-CI-10211
Film applicators	10212	(0 ~ 1) mm	2.7 μm	Electronic micrometer /KTICC-CI-10212
Gap gauges	10213	(1 ~ 150) mm	$\sqrt{(2.6 \mu\text{m})^2 + (2 \times 10^{-6} \times l)^2}$	Height micrometer /KTICC-CI-10213
Gauge blocks, by comparison	10214	(0.5 ~ 100) mm	$\sqrt{(74 \text{ nm})^2 + (1.3 \times 10^{-6} \times l)^2}$	Gauge block /KTICC-CI-10214
Height gauges/ measuring machines	10216	(0 ~ 1 000) mm	$\sqrt{(1.5 \mu\text{m})^2 + (3 \times 10^{-6} \times l)^2}$	Gauge block /KTICC-CI-10216
Standard measuring machines	10220	(0 ~ 300) mm	$\sqrt{(0.4 \mu\text{m})^2 + (2 \times 10^{-6} \times l)^2}$	Gauge block/KTICC-CI-10220
Electronic micrometers	10223	(0 ~ 10) mm	0.15 μm	Gauge block/KTICC-CI-10223
Height micrometers, riser blocks Height micrometers Block Head Riser blocks	10224	(0 ~ 610) mm (0 ~ 25) mm (0 ~ 600) mm	$\sqrt{(1.2 \mu\text{m})^2 + (3 \times 10^{-6} \times l)^2}$ $\sqrt{(0.68 \mu\text{m})^2 + (2 \times 10^{-6} \times l)^2}$ $\sqrt{(1.2 \mu\text{m})^2 + (3 \times 10^{-6} \times l)^2}$	Gauge block /KTICC-CI-10224
Laser scan micrometers	10225	(1 ~ 25) mm	0.41 μm	Cylindrical plug gauges /KTICC-CI-10225
Standard tape rules, peripheral gauges	10227	(0 ~ 5) m (5 ~ 20) m (20 ~ 50) m	$\sqrt{(0.07 \text{ mm})^2 + (10 \times 10^{-6} \times l)^2}$ $\sqrt{(0.22 \text{ mm})^2 + (10 \times 10^{-6} \times l)^2}$ $\sqrt{(0.62 \text{ mm})^2 + (10 \times 10^{-6} \times l)^2}$	Standard rules /KTICC-CI-10227
Cylindrical plug/pin gauges, thread measuring wire gauges	10228	(0 ~ 25) mm (0 ~ 10) mm	0.9 μm 0.62 μm	Laserscan micrometer Standard measuring machine /KTICC-CI-10228
Radius gauges	10229	(0.4 ~ 100) mm	1.3 μm	Non-contact coordinate measuring machine /KTICC-CI-10229
Cylindrical ring gauges	10230	(5 ~ 200) mm	$\sqrt{(0.74 \mu\text{m})^2 + (3 \times 10^{-6} \times D)^2}$	Standard measuring machines /KTICC-CI-10230
Step gauges	10232	(0 ~ 670) mm	$\sqrt{(1.2 \mu\text{m})^2 + (3 \times 10^{-6} \times l)^2}$	Gauge block/KTICC-CI-10232
Taper thickness gauges	10233	(0 ~ 90) mm	1.6 μm	Non-contact coordinate measuring machine /KTICC-CI-10233
Ultrasonic thickness gauges	10234	(0 ~ 300) mm	3.4 μm	Ultrasonic test block /KTICC-CI-10234

102. Linear dimension

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Ultrasonic/ coating thickness specimens Ultrasonic test blocks	10235	(0 ~ 25) mm (0 ~ 500) mm	0.6 μm $\sqrt{(1.3 \mu\text{m})^2 + (2 \times 10^{-6} \times l)^2}$	Outside micrometer, Gauge block KTICC-CI-10235
Coating thickness testers	10236	(0 ~ 1.5) mm (1.5 ~ 7.8) mm (7.8 ~ 25) mm	1.6 μm 7.6 μm 71 μm	Coating standard specimens /KTICC-CI-10236

104. Form

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Optical flats	10404	(0 ~ 60) mm	0.035 μm	Optical flats /KTICC-CI-10404
Parallel blocks Parallelism Flates Height and width difference	10406	(0 ~ 300) mm (0 ~ 300) mm (0 ~ 300) mm	1.2 μm 1.2 μm 1.7 μm	Electronic micrometer /KTICC-CI-10406
Precision surface plates Diagonal length	10407	(0 ~ 5 000) mm	3.3 μm	Electronic levels /KTICC-CI-10407
Straight edges Straightness Parallelism	10412	(0 ~ 2 500) mm (0 ~ 2 500) mm	3.6 μm 3.5 μm	Electronic levels /KTICC-CI-10412
Straight rules	10413	(0 ~ 2 000) mm	$\sqrt{(0.06 \text{ mm})^2 + (10 \times 10^{-6} \times l)^2}$	Standard rules /KTICC-CI-10413

105. Complex geometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Contact coordinate measuring machines Axis accuracy	10503	(0 ~ 1 000) mm	$\sqrt{(0.8 \mu\text{m})^2 + (3 \times 10^{-6} \times l)^2}$	Step gauge /KTICC-CI-10503
Non-contact coordinate measuring machines Axis accuracy Squireness	10504	(0 ~ 500) mm (0 ~ 490) mm	$\sqrt{(0.5 \mu\text{m})^2 + (4 \times 10^{-6} \times l)^2}$ 2.2 μm	Standard scales /KTICC-CI-10504
Measuring microscopes, Profile projectors Axis accuracy Squireness	10511	(0 ~ 500) mm (0 ~ 490) mm	$\sqrt{(0.86 \mu\text{m})^2 + (3 \times 10^{-6} \times l)^2}$ 2.2 μm	Standard scales /KTICC-CI-10511

105. Complex geometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Micro measuring microscopes	10512	(0 ~ 30) mm	$\sqrt{(5 \mu\text{m})^2 + (2 \times 10^{-6} \times l)^2}$	Standard scales /KTICC-CI-10512
Thread plug gauges Outside diameter Effective diameter Pitch Half angle	10525	(0 ~ 200) mm (0 ~ 200) mm (0.2 ~ 5) mm (0 ~ 45)°	1.8 μm 2.1 μm 1.4 μm 2'	Standard measuring machine /KTICC-CI-10525
Thread ring gauges Bore diameter Effective diameter Pitch	10527	(5 ~ 100) mm (5 ~ 100) mm (0.5 ~ 5) mm	2.1 μm 1.6 μm 0.7 μm	Standard measuring machine /KTICC-CI-10527

106. Various dimensional

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Inside/outside/gear tooth calipers, Caliper gauges	10601	(0 ~ 1 000) mm (0 ~ 120) mm	$\sqrt{(9.2 \mu\text{m})^2 + (2 \times 10^{-6} \times l)^2}$ $\sqrt{(6.0 \mu\text{m})^2 + (2 \times 10^{-6} \times l)^2}$	Caliper checker, Gauge block /KTICC-CI-10601
Cylinder/bore gauges	10603	(0 ~ 500) mm	0.8 μm	Dial gage tester /KTICC-CI-10603
Depth gauges, depth micrometers	10604	(0 ~ 600) mm (0 ~ 300) mm	$\sqrt{(7.4 \mu\text{m})^2 + (3 \times 10^{-6} \times l)^2}$ $\sqrt{(1.8 \mu\text{m})^2 + (2 \times 10^{-6} \times l)^2}$	Gauge block/KTICC-CI-10604
Dial/digital gauges	10605	(0 ~ 100) mm	0.80 μm	Dial gage tester /KTICC-CI-10605
Grind gauges Slope depth Scraper straightness	10608	(0 ~ 1) mm (0 ~ 1) mm	2.8 μm 1.4 μm	Electronic micrometer /KTICC-CI-10608
Micro indicators, Test indicators	10609	(0 ~ 2) mm (0 ~ 2) mm	0.74 μm 0.74 μm	Dial gage tester /KTICC-CI-10609
Micrometer heads	10610	(0 ~ 50) mm	$\sqrt{(1.0 \mu\text{m})^2 + (2 \times 10^{-6} \times l)^2}$	Gauge block /KTICC-CI-10610
3-points micrometers	10611	$\varnothing$ (2 ~ 100) mm	1.8 μm	Ring gauge /KTICC-CI-10611
Inside micrometers Bar type micrometers	10612	(0 ~ 300) mm (0 ~ 300) mm	$\sqrt{(1.2 \mu\text{m})^2 + (2 \times 10^{-6} \times l)^2}$ $\sqrt{(1.2 \mu\text{m})^2 + (2 \times 10^{-6} \times l)^2}$	Gauge block /KTICC-CI-10612
Outside micrometers V-anvil micrometers	10613	(0 ~ 300) mm (300 ~ 1 000) mm (1 ~ 25) mm	$\sqrt{(0.9 \mu\text{m})^2 + (2 \times 10^{-6} \times l)^2}$ $\sqrt{(1.9 \mu\text{m})^2 + (2 \times 10^{-6} \times l)^2}$ 1.2 μm	Gauge block, Pluge gauge /KTICC-CI-10613
Standard sieves Wire rod diameter Sieve opening	10617	(0 ~ 10) mm (0 ~ 150) mm	5.4 μm 7.8 μm	Non-contact coordinate measuring machine /KTICC-CI-10617
Welding gauges Length Angle	10620	(0 ~ 90) mm (0 ~ 180)°	6.1 μm 7.8'	Non-contact coordinate measuring machine /KTICC-CI-10620

201. MASS

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Dial platform scale balances	20106	(0 ~ 1) kg (1 ~ 10) kg (10 ~ 50) kg (50 ~ 100) kg (100 ~ 200) kg	2.9 g 15 g 62 g 0.15 kg 0.29 kg	Weight/KTICC-CI-20106
Electric balances	20109	(0 ~ 6) g (6 ~ 30) g (30 ~ 50) g (50 ~ 100) g (100 ~ 200) g (200 ~ 500) g (500 ~ 1 200) g (1 200 ~ 2 000) g (2 ~ 5) kg (5 ~ 10) kg (10 ~ 30) kg (30 ~ 60) kg (60 ~ 150) kg (150 ~ 300) kg (300 ~ 600) kg	0.048 mg 0.090 mg 0.096 mg 0.15 mg 0.24 mg 0.48 mg 1.4 mg 2.1 mg 4.8 mg 9.6 mg 21 mg 43 mg 6.5 g 13 g 0.13 kg	Weight/KTICC-CI-20109
Platform scale balances	20112	(0 ~ 50) kg (50 ~ 100) kg (100 ~ 200) kg	8 g 18 g 35 g	Weight/KTICC-CI-20112
Spring scale balances	20113	(0 ~ 1) kg (1 ~ 10) kg (10 ~ 50) kg (50 ~ 100) kg	2.9 g 15 g 62 g 0.15 kg	Weight/KTICC-CI-20113
Weights	20116	1 mg ~ 20 kg 1 mg 2 mg 5 mg 10 mg 20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg	Class F1 4.6 µg 4.6 µg 4.6 µg 4.6 µg 5.1 µg 6.5 µg 6.5 µg 7.3 µg 9.0 µg 16 µg 18 µg 23 µg 26 µg 31 µg 37 µg 0.10 mg 0.13 mg 0.28 mg 0.54 mg 1.0 mg 2.8 mg 5.4 mg 10 mg	Weight/KTICC-CI-20116

202. Force

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Tension/compression testing machines	20203			Load Cell /KTICC-CI-20203
Tension		(30 ~ 5 000) N (5 ~ 20) kN	1.5×10^{-3} 1.6×10^{-3}	
Compression		(30 ~ 300) N (300 ~ 500) N (0.5 ~ 1) kN (1 ~ 10) kN (10 ~ 30) kN (30 ~ 50) kN (50 ~ 100) kN (100 ~ 300) kN (300 ~ 500) kN (500 ~ 1 000) kN (1 000 ~ 2 000) kN	1.6×10^{-3} 1.5×10^{-3} 1.6×10^{-3} 1.5×10^{-3} 2.0×10^{-3} 2.1×10^{-3} 1.7×10^{-3} 2.0×10^{-3} 1.6×10^{-3} 1.5×10^{-3} 3.5×10^{-3}	
Push-pull gauges	20204			Weight/KTICC-CI-20204
Push, Pull		(1 ~ 500) N	1.3×10^{-3}	

203. Torque

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Torque measuring devices	20302			Torque calibration system /KTICC-CI-20302
		(0.1 ~ 1) N·m (1 ~ 5) N·m (5 ~ 10) N·m (10 ~ 20) N·m (20 ~ 200) N·m	2.4×10^{-3} 2.6×10^{-3} 1.3×10^{-3} 1.7×10^{-3} 1.1×10^{-3}	
Torque wrenches/drivers	20303			Torque calibration system Standard weights, Torque cell /KTICC-CI-20303
		(0.05 ~ 0.5) N·m (0.5 ~ 1) N·m (1 ~ 5) N·m (5 ~ 10) N·m (10 ~ 25) N·m (25 ~ 50) N·m (50 ~ 100) N·m (100 ~ 250) N·m (250 ~ 500) N·m (500 ~ 1 000) N·m	3.0×10^{-2} 3.0×10^{-2} 1.0×10^{-2} 1.1×10^{-2} 4.8×10^{-3} 4.7×10^{-3} 5.2×10^{-3} 2.5×10^{-3} 3.7×10^{-3} 3.9×10^{-3}	

204. Pressure

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Manometers Well, U type Inclined tube	20402	(0 ~ 100) kPa (0 ~ 10) kPa	2.0×10^{-3} 2.4×10^{-3}	Pressure generator & controller /KTICC-CI-20402
Absolute pressure gauges	20406	(75 ~ 135) kPa abs. (0.135 ~ 7) MPa abs.	9.4×10^{-5} 6.5×10^{-4}	Pressure generator & controller /KTICC-CI-20406
Blood pressure gauges	20407	(0 ~ 40) kPa	9.7×10^{-4}	Pressure generator & controller /KTICC-CI-20407
Compound pressure gauges	20408	(-0.095 ~ 7) MPa	6.1×10^{-4}	Deadweight tester Pressure generator & controller /KTICC-CI-20408
Differential pressure gauges	20409	(0 ~ 3) MPa	5.9×10^{-4}	Pressure generator & controller /KTICC-CI-20409
Gauge pressure gauges Gauge pressure gauges	20411	(0 ~ 40) kPa (0.04 ~ 10) MPa (10 ~ 110) MPa	2.7×10^{-4} 1.1×10^{-4} 1.3×10^{-4}	Deadweight tester /KTICC-CI-20411
Pressure transducers/transmitters	20412	(-95 ~ 0) kPa (0 ~ 40) kPa (0.04 ~ 110) MPa	9.7×10^{-4} 6.8×10^{-4} 3.8×10^{-4}	Deadweight tester /KTICC-CI-20412
Dial type vacuum gauges	20413	(-95 ~ 0) kPa	3.0×10^{-3}	Pressure generator & controller /KTICC-CI-20413
Water depth meters	20414	(0 ~ 100) m	2.5×10^{-3}	Pressure generator & controller /KTICC-CI-20414

206. Volume

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Volumetric glasswares	20601	(0 ~ 1) ml (1 ~ 2) ml (2 ~ 5) ml (5 ~ 10) ml (10 ~ 25) ml (25 ~ 50) ml (50 ~ 100) ml (100 ~ 200) ml (200 ~ 250) ml (250 ~ 500) ml (500 ~ 1 000) ml (1 000 ~ 2 000) ml	1.8 μ l 2.4 μ l 3.5 μ l 4.7 μ l 7.8 μ l 13 μ l 23 μ l 32 μ l 50 μ l 74 μ l 0.13 ml 0.24 ml	Electric balances /KTICC-CI-20601
Pycnometers	20602	(0 ~ 50) ml (50 ~ 100) ml (100 ~ 500) ml	4.5 μ l 9.3 μ l 45 μ l	Electric balances /KTICC-CI-20602

206. Volume

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Concrete air content meters	20605	(0 ~ 10) %	0.06 %	Electric balances /KTICC-CI-20605
Piston type volume meters	20606	(0 ~ 10) μ l (10 ~ 20) μ l (20 ~ 50) μ l (50 ~ 100) μ l (100 ~ 200) μ l (200 ~ 500) μ l (500 ~ 1 000) μ l (1 ~ 2) ml (2 ~ 5) ml (5 ~ 10) ml (10 ~ 20) ml	0.012 μ l 0.029 μ l 0.070 μ l 0.090 μ l 0.14 μ l 0.31 μ l 0.60 μ l 1.2 μ l 3.0 μ l 5.9 μ l 12 μ l	Electric balances /KTICC-CI-20606

301. Time/frequency

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Frequency standards	30102	(1 ~ 10) MHz	5.9×10^{-13}	Frequency standard /KTICC-CI-30102
General frequency sources	30103	(1 ~ 10) MHz	5.9×10^{-13}	Frequency standard /KTICC-CI-30103
Frequency meters/counters Compared to the frequency standard Input Frequency Reference Output	30104	10 MHz 1 Hz ~ 5 GHz (5 ~ 40) GHz (1 ~ 10) MHz	7.6×10^{-12} 1.2×10^{-10} 1.3 Hz 5.9×10^{-13}	Frequency Standard /KTICC-CI-30104
Time interval sources	30105	1 ns ~ 1 s (1 ~ 5) s	6.1×10^{-6} 1.2×10^{-6}	Frequency Counter /KTICC-CI-30105
Time interval meters/stop watches/timers Stop watches Time interval meters/timers	30106	(0.1 ~ 86 400) s (1 ~ 10 000) s	1.4×10^{-7} 6.1×10^{-5}	Watch test equipment, Oscilloscope /KTICC-CI-30106

302. Velocity & revolution

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Standard RPM generators	30201	$(1 \sim 30) \text{ min}^{-1}$ $(30 \sim 90) \text{ min}^{-1}$ $(90 \sim 300) \text{ min}^{-1}$ $(300 \sim 900) \text{ min}^{-1}$ $(900 \sim 3\,000) \text{ min}^{-1}$ $(3\,000 \sim 9\,000) \text{ min}^{-1}$ $(9\,000 \sim 30\,000) \text{ min}^{-1}$ $(30\,000 \sim 60\,000) \text{ min}^{-1}$ $(60\,000 \sim 99\,999) \text{ min}^{-1}$	0.02 min^{-1} 0.03 min^{-1} 0.2 min^{-1} 0.3 min^{-1} 2 min^{-1} 3 min^{-1} 8 min^{-1} 15 min^{-1} 24 min^{-1}	RPM Calibration system /KTICC-CI-30201
Contact type tachometers	30202	$(1 \sim 600) \text{ min}^{-1}$ $(600 \sim 4\,000) \text{ min}^{-1}$	0.10 min^{-1} 0.12 min^{-1}	RPM Calibration system /KTICC-CI-30202
Photo tachometers/storoboscopes Photo tachometers Stroboscope	30203	3 min^{-1} $(3 \sim 600) \text{ min}^{-1}$ $(600 \sim 6\,000) \text{ min}^{-1}$ $(6\,000 \sim 600\,000) \text{ min}^{-1}$ $(30 \sim 300\,000) \text{ min}^{-1}$	0.006 min^{-1} 0.06 min^{-1} 0.09 min^{-1} 0.6 min^{-1} 0.01 min^{-1}	RPM Calibration system /KTICC-CI-30203

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC ammeters DC Current	40101	(±) 0 nA (0 ~ 19) nA (19 ~ 190) nA (0.19 ~ 1.9) μA (1.9 ~ 19) μA (19 ~ 190) μA (0.19 ~ 1) mA (1 ~ 10) mA (10 ~ 19) mA (19 ~ 100) mA (100 ~ 190) mA (0.19 ~ 1) A (1 ~ 1.9) A (1.9 ~ 2) A (2 ~ 10) A (10 ~ 100) A	0.46 pA 2.3×10^{-4} 1.2×10^{-4} 5.2×10^{-5} 3.2×10^{-5} 1.8×10^{-5} 4.3×10^{-5} 4.0×10^{-5} 3.8×10^{-5} 5.3×10^{-5} 5.0×10^{-5} 9.4×10^{-5} 8.9×10^{-5} 9.0×10^{-5} 4.2×10^{-4} 1.5×10^{-4}	Meter calibrator /KTICC-CI-40101
Transconductance amplifiers DC Current AC Current	40102	(±) 0 mA (0 ~ 1) mA (1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A (10 ~ 100) A (0.04 ~ 1) kHz (0.009 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A (10 ~ 100) A	1.2 nA 1.4×10^{-5} 1.5×10^{-5} 1.8×10^{-5} 6.0×10^{-5} 8.9×10^{-5} 5.8×10^{-4} 4.5×10^{-4} 8.8×10^{-4} 1.1×10^{-3} 1.2×10^{-3}	DMM, Current shunt Meter calibrator /KTICC-CI-40102
DC voltage/current calibrators DC Voltage DC Current	40103	(±) 0 mV (0 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V (±) 0 nA (0 ~ 1) nA (1 ~ 10) nA (10 ~ 100) nA (0.1 ~ 1) μA (1 ~ 10) μA (10 ~ 100) μA (0.1 ~ 1) mA (1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A (10 ~ 20) A (20 ~ 100) A	0.38 μV 8.9×10^{-6} 3.8×10^{-6} 3.8×10^{-6} 6.0×10^{-6} 6.1×10^{-6} 0.01 pA 2.3×10^{-4} 1.2×10^{-4} 1.0×10^{-5} 7.0×10^{-6} 6.0×10^{-6} 6.5×10^{-6} 7.8×10^{-6} 8.3×10^{-6} 1.3×10^{-5} 5.9×10^{-5} 8.8×10^{-5} 1.2×10^{-4} 5.8×10^{-4}	DMM, Current shunt /KTICC-CI-40103

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrical temperature calibrators (Sensor not included)	40104			Meter calibrator, DMM /KTICC-CI-40104
Output DC Voltage		($\pm$) 0 mV (0 ~ 1) mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V	0.10 μ V 1.1×10^{-4} 1.6×10^{-5} 8.9×10^{-6} 3.8×10^{-6} 3.5×10^{-6} 4.6×10^{-6}	
Output DC Current		($\pm$) 0 mA (0 ~ 1) mA (1 ~ 10) mA (10 ~ 110) mA	0.012 nA 6.5×10^{-6} 8.3×10^{-6} 1.3×10^{-5}	
Output Resistance		0 Ω (0 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 10) k Ω	4.1 $\mu\Omega$ 1.6×10^{-5} 1.1×10^{-5} 8.7×10^{-6} 7.4×10^{-6}	
Input DC Voltage		($\pm$) 0 mV (0 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 300) V	0.41 μ V 4.8×10^{-5} 1.1×10^{-5} 5.9×10^{-6} 4.0×10^{-6} 5.9×10^{-6} 8.3×10^{-6}	
Input DC Current		($\pm$) 0 mA (0 ~ 1) mA (1 ~ 10) mA (10 ~ 110) mA	6.2 nA 4.3×10^{-5} 4.0×10^{-5} 5.3×10^{-5}	
Input Resistance		0 Ω (0 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) k Ω (1 ~ 10) k Ω	0.61 $\mu\Omega$ 1.1×10^{-5} 2.4×10^{-5} 8.0×10^{-6} 1.0×10^{-5} 6.9×10^{-6}	

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC current shunts Resistance	40105	(0.1 ~ 10) k Ω (10 ~ 100) Ω (1 ~ 10) Ω (0.1 ~ 1) Ω (0.01 ~ 0.1) Ω (0.001 ~ 0.01) Ω (0.5 ~ 1) m Ω (0 ~ 0.5) m Ω	7.4×10^{-6} 8.7×10^{-6} 1.1×10^{-5} 1.6×10^{-5} 9.5×10^{-5} 4.2×10^{-4} 1.3×10^{-4} 1.4×10^{-4}	Meter calibrator, DMM /KTICC-CI-40105
Galvanometers/null detectors DC Voltage	40106	($\pm$) 0 μ V (0 ~ 1) μ V (1 ~ 3) μ V (3 ~ 10) μ V (10 ~ 30) μ V (30 ~ 100) μ V (100 ~ 300) μ V (0.3 ~ 1) mV (1 ~ 3) mV (3 ~ 10) mV (10 ~ 30) mV (30 ~ 100) mV (100 ~ 300) mV (0.3 ~ 1) V (1 ~ 3) V (3 ~ 10) V (10 ~ 30) V (30 ~ 100) V (100 ~ 300) V (300 ~ 1 000) V	5.8 nV 5.8×10^{-3} 9.7×10^{-3} 5.8×10^{-3} 9.7×10^{-3} 5.8×10^{-3} 9.7×10^{-3} 5.8×10^{-3} 9.7×10^{-3} 5.8×10^{-3} 9.7×10^{-3} 5.8×10^{-3} 9.7×10^{-3} 5.8×10^{-3} 9.7×10^{-3} 5.8×10^{-3} 9.7×10^{-3} 5.8×10^{-3} 9.7×10^{-3} 5.8×10^{-3}	Meter calibrator /KTICC-CI-40106
Potentiometers DC Voltage	40107	($\pm$) 0 mV (0 ~ 1) mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 1000) V	0.41 μ V 4.7×10^{-4} 5.3×10^{-5} 1.3×10^{-5} 8.5×10^{-6} 7.3×10^{-6} 8.5×10^{-6} 9.5×10^{-6}	Meter calibrator /KTICC-CI-40107
DC power supplies DC Voltage	40108	($\pm$) 0 V (0 ~ 0.1) V (0.1 ~ 1) V (1 ~ 10) V (10 ~ 20) V (20 ~ 30) V (30 ~ 40) V (40 ~ 100) V (100 ~ 1 000) V	0.62 μ V 1.1×10^{-5} 7.2×10^{-6} 7.2×10^{-6} 8.0×10^{-6} 7.0×10^{-6} 6.8×10^{-6} 8.6×10^{-6} 8.6×10^{-6}	DMM, Current shunt, C.T /KTICC-CI-40108

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC power supplies DC Current	40108	($\pm$) 0 A (0 ~ 0.1) A (0.1 ~ 1) A (1 ~ 1.5) A (1.5 ~ 1.9) A (1.9 ~ 15) A (15 ~ 20) A (20 ~ 100) A (100 ~ 1 000) A	0.62 μ A 7.7×10^{-5} 2.0×10^{-4} 1.9×10^{-4} 1.8×10^{-4} 2.4×10^{-4} 2.5×10^{-4} 5.9×10^{-4} 3.6×10^{-4}	DMM, Current shunt, C.T /KTICC-CI-40108
DC voltage standards DC Voltage	40111	1.018 V 10 V	1.6×10^{-6} 1.6×10^{-6}	Null detector, DC standard /KTICC-CI-40111
DC voltmeters DC Voltage	40112	($\pm$) 0 mV (0 ~ 1) mV (1 ~ 1.9) mV (1.9 ~ 10) mV (10 ~ 19) mV (19 ~ 100) mV (100 ~ 190) mV (0.19 ~ 1) V (1 ~ 1.9) V (1.9 ~ 10) V (10 ~ 19) V (19 ~ 100) V (100 ~ 190) V (190 ~ 1 000) V	0.41 μ V 4.2×10^{-4} 2.2×10^{-4} 4.9×10^{-5} 2.9×10^{-5} 1.2×10^{-5} 1.0×10^{-5} 5.9×10^{-6} 5.8×10^{-6} 4.0×10^{-6} 3.8×10^{-6} 5.9×10^{-6} 5.8×10^{-6} 7.3×10^{-6}	Meter calibrator /KTICC-CI-40112

401. DC voltage & current

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Static/ionicvoltmeters DC Voltage	40113	(±) 0 kV (0 ~ 1) kV (1 ~ 2) kV (2 ~ 3) kV (3 ~ 4) kV (4 ~ 5) kV (5 ~ 6) kV (6 ~ 8) kV (8 ~ 15) kV (15 ~ 50) kV	0.61 V 1.3×10^{-3} 1.2×10^{-3} 2.4×10^{-3} 1.9×10^{-3} 1.7×10^{-3} 1.6×10^{-3} 1.4×10^{-3} 1.3×10^{-3} 1.2×10^{-3}	High voltage power supply, High voltage test equipment, DMM /KTICC-CI-40113
Ion Voltage		(±) 0 V (0 ~ 10) V (10 ~ 20) V (20 ~ 40) V (40 ~ 60) V (60 ~ 80) V (80 ~ 100) V (100 ~ 150) V (150 ~ 200) V (200 ~ 400) V (400 ~ 600) V (600 ~ 800) V (800 ~ 1 000) V	0.61 mV 6.1×10^{-4} 3.1×10^{-4} 1.5×10^{-4} 1.0×10^{-4} 7.6×10^{-5} 6.1×10^{-4} 4.1×10^{-4} 3.1×10^{-4} 1.5×10^{-4} 1.0×10^{-4} 7.6×10^{-5} 6.1×10^{-4}	

402. Resistance, capacitance and inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Capacitance bridges/indicators	40201	10 Hz ~ 100 MHz	6.1×10^{-7}	Standard Capacitor, DMM, Frequency Counter /KTICC-CI-40201
Frequency				
Voltage		1 mV		
		(0.04 ~ 10) kHz	1.1×10^{-3}	
		(10 ~ 50) kHz	1.6×10^{-3}	
		(50 ~ 100) kHz	2.0×10^{-3}	
		(0.1 ~ 1) MHz	9.7×10^{-3}	
		(1 ~ 10) mV		
		40 Hz	1.6×10^{-4}	
		(0.04 ~ 10) kHz	1.5×10^{-4}	
		(10 ~ 50) kHz	2.2×10^{-4}	
		(50 ~ 100) kHz	2.7×10^{-4}	
		(0.10 ~ 1) MHz	2.6×10^{-3}	
		(10 ~ 100) mV		
		40 Hz	7.4×10^{-5}	
		(0.04 ~ 10) kHz	7.2×10^{-5}	
		(10 ~ 50) kHz	7.7×10^{-5}	
		(50 ~ 100) kHz	1.0×10^{-4}	
		(0.10 ~ 1) MHz	1.0×10^{-3}	
		(0.1 ~ 1) V		
		40 Hz	6.7×10^{-5}	
		(0.04 ~ 10) kHz	6.5×10^{-5}	
		(10 ~ 50) kHz	6.7×10^{-5}	
		(50 ~ 100) kHz	7.7×10^{-5}	
		(0.1 ~ 1) MHz	9.6×10^{-4}	
		(1 ~ 2) V		
		40 Hz	4.1×10^{-5}	
		(0.04 ~ 10) kHz	3.8×10^{-5}	
		(10 ~ 50) kHz	4.1×10^{-5}	
		(50 ~ 100) kHz	5.5×10^{-5}	
		(0.10 ~ 1) MHz	9.5×10^{-4}	
		(2 ~ 5) V		
		40 Hz	3.2×10^{-5}	
		(0.04 ~ 1) kHz	2.4×10^{-5}	
		(1 ~ 10) kHz	2.6×10^{-5}	
		(10 ~ 50) kHz	3.0×10^{-5}	
		(50 ~ 100) kHz	5.8×10^{-5}	
		(0.10 ~ 1) MHz	1.2×10^{-3}	
Capacitance		1 pF		
		1 kHz	4.6×10^{-4}	
		(0.001 ~ 1) MHz	4.3×10^{-4}	
		(1 ~ 2) MHz	4.8×10^{-4}	
		(2 ~ 3) MHz	5.8×10^{-4}	
		(3 ~ 4) MHz	7.5×10^{-4}	
		(4 ~ 5) MHz	9.7×10^{-4}	
		(5 ~ 10) MHz	3.4×10^{-3}	
		(10 ~ 13) MHz	4.3×10^{-3}	

402. Resistance, capacitance and inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Capacitance bridges/indicators Capacitance	40201	(1 ~ 10) pF 1 kHz (0.001 ~ 5) MHz (5 ~ 13) MHz (10 ~ 100) pF 1 kHz (0.001 ~ 2) MHz (2 ~ 4) MHz (4 ~ 5) MHz (5 ~ 13) MHz (0.1 ~ 1) nF 1 kHz (0.001 ~ 1) MHz (1 ~ 2) MHz (2 ~ 3) MHz (3 ~ 4) MHz (4 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz (1 ~ 10) nF 120 Hz (0.12 ~ 1) kHz (1 ~ 100) kHz (10 ~ 100) nF 120 Hz (0.12 ~ 1) kHz (1 ~ 100) kHz (0.1 ~ 1) μF 120 Hz (0.12 ~ 1) kHz (1 ~ 10) kHz (10 ~ 100) kHz (1 ~ 10) μF (0.12 ~ 1) kHz (10 ~ 100) μF 120 Hz (0.12 ~ 1) kHz (0.1 ~ 10) mF (0.12 ~ 1) kHz	3.4×10^{-5} 4.2×10^{-4} 2.4×10^{-3} 3.4×10^{-5} 4.2×10^{-4} 4.3×10^{-4} 4.4×10^{-4} 2.4×10^{-3} 3.4×10^{-5} 4.3×10^{-4} 4.5×10^{-4} 5.1×10^{-4} 6.1×10^{-4} 7.5×10^{-4} 3.1×10^{-3} 3.8×10^{-3} 2.5×10^{-4} 1.3×10^{-4} 2.5×10^{-4} 2.5×10^{-4} 1.5×10^{-4} 2.5×10^{-4} 2.5×10^{-4} 2.5×10^{-4} 1.5×10^{-4} 2.5×10^{-4} 2.6×10^{-4} 5.9×10^{-4} 6.1×10^{-4} 6.0×10^{-4} 2.9×10^{-3}	Standard Capacitor, DMM, Frequency Counter /KTICC-CI-40201
Decade capacitors Capacitance	40202	1 kHz (1 ~ 100) pF (0.1 ~ 1) nF (1 ~ 10) nF (0.01 ~ 1) μF (1 ~ 10) μF (10 ~ 100) μF	3.5×10^{-4} 2.7×10^{-4} 3.0×10^{-4} 3.1×10^{-4} 6.6×10^{-4} 7.0×10^{-4}	LCR Meter /KTICC-CI-40202

402. Resistance, capacitance and inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Standard capacitors Capacitance	40204	1 pF 1 kHz (0.001 ~ 1) MHz (1 ~ 3) MHz (3 ~ 4) MHz (4 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz (1 ~ 10) pF 1 kHz (0.001 ~ 5) MHz (5 ~ 13) MHz (10 ~ 100) pF 1 kHz (0.001 ~ 2) MHz (2 ~ 5) MHz (5 ~ 13) MHz (0.1 ~ 1) nF 1 kHz (0.001 ~ 1) MHz (1 ~ 2) MHz (2 ~ 3) MHz (3 ~ 4) MHz (4 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz (1 ~ 10) nF 120 Hz (0.12 ~ 1) kHz (1 ~ 100) kHz (10 ~ 100) nF 120 Hz (0.12 ~ 1) kHz (1 ~ 100) kHz (0.1 ~ 1) μF 120 Hz (0.12 ~ 1) kHz (1 ~ 10) kHz (10 ~ 100) kHz (1 ~ 10) μF (0.12 ~ 1) kHz (10 ~ 100) μF 1 kHz	 5.8×10^{-4} 5.5×10^{-4} 1.1×10^{-3} 1.2×10^{-3} 1.4×10^{-3} 3.4×10^{-3} 4.5×10^{-3} 3.5×10^{-4} 5.0×10^{-4} 2.4×10^{-3} 3.5×10^{-4} 5.0×10^{-4} 5.1×10^{-4} 2.4×10^{-3} 2.7×10^{-4} 5.1×10^{-4} 5.2×10^{-4} 5.8×10^{-4} 6.7×10^{-4} 8.0×10^{-4} 3.1×10^{-3} 3.8×10^{-3} 3.6×10^{-4} 3.0×10^{-4} 3.6×10^{-4} 3.6×10^{-4} 3.1×10^{-4} 3.6×10^{-4} 3.6×10^{-4} 3.1×10^{-4} 3.6×10^{-4} 6.3×10^{-4} 6.6×10^{-4} 6.9×10^{-4}	LCR Meter /KTICC-CI-40204

402. Resistance, capacitance and inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Earth testers	40205			Meter calibrator, DMM Current shunt /KTICC-CI-40205
Input AC Voltage		(50 ~ 60) Hz 0.22 mV ~ 1 V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V	9.1×10^{-5} 8.9×10^{-5} 9.4×10^{-5} 1.0×10^{-4}	
Input DC Voltage		($\pm$) 0 V (0 ~ 100) V (100 ~ 1 000) V	61 μ V 6.1×10^{-5} 6.2×10^{-5}	
Input DC Resistance		(0.1 ~ 1) m Ω (1 ~ 10) m Ω (10 ~ 100) m Ω (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω 100 Ω ~ 1 M Ω (1 ~ 10) M Ω	1.3×10^{-4} 1.1×10^{-4} 8.5×10^{-5} 6.2×10^{-5} 6.1×10^{-5} 6.2×10^{-5} 6.1×10^{-5} 6.2×10^{-5}	
Output AC Voltage		(50 ~ 60) Hz (0.22 ~ 100) mV (0.1 ~ 10) V	6.5×10^{-4} 6.2×10^{-4}	
Output AC Current		(50 ~ 60) Hz 9 μ A ~ 100 A	1.4×10^{-3}	
Output DC Current		($\pm$) 0 A (0 ~ 0.1) A (0.1 ~ 1) A (1 ~ 2) A (2 ~ 5) A (5 ~ 10) A (10 ~ 20) A (20 ~ 40) A (40 ~ 60) A (60 ~ 80) A (80 ~ 100) A	2.2 μ A 6.2×10^{-4} 6.5×10^{-4} 3.9×10^{-4} 2.8×10^{-4} 6.6×10^{-4} 3.9×10^{-4} 6.0×10^{-4} 5.8×10^{-4} 5.9×10^{-4} 8.4×10^{-4}	
Input AC Current		(50 ~ 60) Hz (0.009 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A (10 ~ 20) A (20 ~ 50) A	6.4×10^{-4} 6.9×10^{-4} 8.1×10^{-4} 2.3×10^{-3} 2.4×10^{-3}	
Input AC Resistance		(50 ~ 60) Hz 0.5 m Ω ~ 100 Ω	1.3×10^{-3}	
Timer		(1 ~ 1 000) s	6.1×10^{-5}	

402. Resistance, capacitance and inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Inductance bridges/indicators	40206	10 Hz ~ 1 MHz	6.1×10^{-7}	Inductors, DMM, Frequency counter /KTICC-CI-40206	
Frequency		1 mV			
AC Voltage		(40 ~ 100) Hz	4.2×10^{-3}		
		(0.1 ~ 1) kHz	2.1×10^{-3}		
		(1 ~ 10) kHz	4.2×10^{-3}		
		(10 ~ 100) kHz	2.1×10^{-2}		
		(1 ~ 10) mV			
		40 Hz	5.6×10^{-4}		
		(40 ~ 100) Hz	5.4×10^{-4}		
		(0.1 ~ 1) kHz	3.5×10^{-4}		
		(1 ~ 10) kHz	5.4×10^{-4}		
		(10 ~ 100) kHz	2.8×10^{-3}		
		(10 ~ 100) mV			
		40 Hz	2.4×10^{-4}		
		(40 ~ 100) Hz	2.2×10^{-4}		
		(0.1 ~ 1) kHz	2.1×10^{-4}		
		(1 ~ 10) kHz	2.2×10^{-4}		
		(10 ~ 50) kHz	9.6×10^{-4}		
		(50 ~ 100) kHz	1.1×10^{-3}		
Inductance		(0.1 ~ 1) V			
		40 Hz	1.5×10^{-4}		
		(40 ~ 100) Hz	1.4×10^{-4}		
		(0.1 ~ 1) kHz	1.2×10^{-4}		
		(1 ~ 10) kHz	1.4×10^{-4}		
		(10 ~ 50) kHz	7.2×10^{-4}		
		(50 ~ 100) kHz	7.3×10^{-4}		
		(1 ~ 2) V			
		40 Hz	2.2×10^{-4}		
		(40 ~ 100) Hz	2.0×10^{-4}		
		(0.1 ~ 1) kHz	1.8×10^{-4}		
		(1 ~ 10) kHz	2.0×10^{-4}		
		(10 ~ 100) kHz	1.5×10^{-3}		
		(2 ~ 5) V			
		40 Hz	1.6×10^{-4}		
		(40 ~ 100) Hz	1.4×10^{-4}		
		(0.1 ~ 1) kHz	1.2×10^{-4}		
		(1 ~ 10) kHz	1.4×10^{-4}		
		(10 ~ 100) kHz	9.2×10^{-4}		
</					

402. Resistance, capacitance and inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Insulation testers	40210			Decade resistance, DMM, Meter calibrator /KTICC-CI-40210
Resistance		(0 ~ 1) Ω	6.2×10^{-5}	
		(1 ~ 10) Ω	6.1×10^{-5}	
		(10 ~ 100) Ω	6.2×10^{-5}	
		(0.1 ~ 100) $k\Omega$	6.1×10^{-5}	
		(0.1 ~ 1) $M\Omega$	1.3×10^{-4}	
		(1 ~ 10) $M\Omega$	3.6×10^{-4}	
		(10 ~ 100) $M\Omega$	3.8×10^{-4}	
		(0.1 ~ 10) $G\Omega$	8.4×10^{-4}	
		(10 ~ 100) $G\Omega$	1.3×10^{-3}	
		(0.1 ~ 1) $T\Omega$	2.4×10^{-3}	
Output DC Voltage		($\pm$)		
		(0.1 ~ 100) V	6.1×10^{-4}	
		(100 ~ 500) V	1.2×10^{-4}	
		(0.5 ~ 1) kV	6.1×10^{-4}	
		(1 ~ 2) kV	7.5×10^{-3}	
		(2 ~ 3) kV	7.0×10^{-3}	
		(3 ~ 4) kV	6.8×10^{-3}	
		(4 ~ 5) kV	6.6×10^{-3}	
		(5 ~ 10) kV	6.2×10^{-3}	
Input AC Voltage		(50 ~ 60) Hz		
		0.22 mV ~ 100 V	6.1×10^{-4}	
		(100 ~ 1 000) V	6.2×10^{-4}	
Input DC Voltage		($\pm$)		
		0 V	0.61 mV	
		(0 ~ 1 000) V	6.1×10^{-4}	
Timer		(1 ~ 1 000) s	6.1×10^{-5}	
Resistance bridges & similar instruments	40213			DMM, Decade resistance, Standard resistance /KTICC-CI-40213
Resistance ARM		0 $m\Omega$	7.3 $\mu\Omega$	
		(1 ~ 10) $m\Omega$	7.4×10^{-4}	
		(10 ~ 100) $m\Omega$	8.0×10^{-5}	
		(0.1 ~ 1) Ω	1.8×10^{-5}	
		(1 ~ 10) Ω	1.0×10^{-5}	
		(10 ~ 100) Ω	8.7×10^{-6}	
		(0.1 ~ 100) $k\Omega$	8.2×10^{-6}	
		(0.1 ~ 1) $M\Omega$	9.6×10^{-6}	
		(1 ~ 10) $M\Omega$	2.1×10^{-5}	
Ratio ARM		0.001	1.1×10^{-5}	
		(0.001 ~ 0.01)	1.3×10^{-5}	
		(0.01 ~ 0.1)	1.1×10^{-5}	
		(0.1 ~ 1)	1.2×10^{-5}	
		(1 ~ 10)	1.1×10^{-5}	
		(10 ~ 100)	1.2×10^{-5}	
		(100 ~ 1 000)	1.4×10^{-5}	

402. Resistance, capacitance and inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistance bridges & similar instruments Resistance Measured	40213	0.5 m Ω ~ 1 Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) k Ω (1 ~ 10) k Ω (10 ~ 100) k Ω (0.1 ~ 1) M Ω (1 ~ 10) M Ω (10 ~ 100) M Ω	1.1×10^{-5} 6.7×10^{-6} 9.6×10^{-6} 6.7×10^{-6} 8.1×10^{-6} 7.5×10^{-6} 7.0×10^{-6} 9.9×10^{-6} 1.0×10^{-4}	DMM, Decade resistance, Standard resistance /KTICC-CI-40213
Resistance meters DC Voltage DC Resistance Output Frequency Output AC Voltage AC Resistance	40214	($\pm$) 0.1 V (0.1 ~ 10) V (10 ~ 25) V (25 ~ 50) V (50 ~ 100) V (100 ~ 250) V (250 ~ 500) V (500 ~ 1 000) V 0.5 m Ω (0.5 ~ 1) m Ω (1 ~ 10) m Ω (10 ~ 100) m Ω (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) k Ω (1 ~ 10) k Ω (10 ~ 100) k Ω (0.1 ~ 1) M Ω (1 ~ 10) M Ω (10 ~ 100) M Ω (0.1 ~ 1) G Ω (1 ~ 10) G Ω (10 ~ 100) G Ω (0.1 ~ 1) T Ω 50 Hz (50 ~ 60) Hz (0.06 ~ 1) kHz (0.04 ~ 1) kHz 10 mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 100) V (0.04 ~ 1) kHz 0.5 m Ω (0.5 ~ 100) m Ω 100 m Ω ~ 1 k Ω (1 ~ 100) k Ω (0.1 ~ 1) M Ω	6.2×10^{-5} 6.1×10^{-5} 2.6×10^{-5} 1.4×10^{-5} 6.1×10^{-5} 2.6×10^{-5} 1.4×10^{-5} 6.1×10^{-5} 2.6×10^{-3} 1.2×10^{-4} 8.8×10^{-5} 5.8×10^{-5} 9.6×10^{-6} 2.8×10^{-6} 7.4×10^{-6} 2.9×10^{-6} 5.4×10^{-6} 3.1×10^{-6} 4.2×10^{-6} 8.3×10^{-6} 1.1×10^{-4} 6.2×10^{-4} 6.3×10^{-4} 1.2×10^{-3} 2.3×10^{-3} 1.2×10^{-6} 1.0×10^{-6} 6.1×10^{-7} 3.5×10^{-4} 2.1×10^{-4} 1.2×10^{-4} 1.1×10^{-4} 1.1×10^{-2} 1.2×10^{-3} 3.7×10^{-4} 2.6×10^{-4} 3.3×10^{-4}	DMM, Decade resistance, High resistance, Standards resistance, Frequency counter /KTICC-CI-40214

402. Resistance, capacitance and inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistors	40215			DMM, Megohmmeter, Meter calibrator, Standard resistance /KTICC-CI-40215
Decade Resistance		0 mΩ (0 ~ 10) mΩ (10 ~ 100) mΩ (0.1 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 100) kΩ (0.1 ~ 1) MΩ (1 ~ 10) MΩ (10 ~ 100) MΩ (0.1 ~ 1) GΩ (1 ~ 10) GΩ (10 ~ 100) GΩ (0.1 ~ 1) TΩ	7.3 μΩ 7.4×10^{-4} 8.0×10^{-5} 1.9×10^{-5} 1.1×10^{-5} 8.7×10^{-6} 8.3×10^{-6} 9.6×10^{-6} 2.1×10^{-5} 3.6×10^{-5} 1.5×10^{-4} 6.6×10^{-4} 5.3×10^{-3} 6.0×10^{-3}	
DC Resistance		0.5 mΩ (0.5 ~ 1) mΩ (1 ~ 10) mΩ (10 ~ 100) mΩ (0.1 ~ 1) Ω (1 ~ 1.9) Ω (1.9 ~ 10) Ω (10 ~ 19) Ω (19 ~ 100) Ω (100 ~ 190) Ω (0.19 ~ 1) kΩ (1 ~ 1.9) kΩ (1.9 ~ 10) kΩ (10 ~ 19) kΩ (19 ~ 100) kΩ (100 ~ 190) kΩ (0.19 ~ 1) MΩ (1 ~ 1.9) MΩ (1.9 ~ 19) MΩ (19 ~ 100) MΩ (0.1 ~ 1) GΩ (1 ~ 10) GΩ (10 ~ 100) GΩ (0.1 ~ 1) TΩ	2.2×10^{-4} 1.2×10^{-4} 8.9×10^{-5} 6.0×10^{-5} 1.8×10^{-5} 1.6×10^{-5} 1.1×10^{-5} 1.0×10^{-5} 8.7×10^{-6} 8.5×10^{-6} 8.3×10^{-6} 8.0×10^{-6} 8.3×10^{-6} 8.0×10^{-6} 8.3×10^{-6} 8.0×10^{-6} 9.6×10^{-6} 9.2×10^{-6} 1.1×10^{-5} 3.6×10^{-5} 1.5×10^{-4} 1.6×10^{-3} 5.3×10^{-3} 6.0×10^{-3}	
AC Resistance		(0.04 ~ 1) kHz 0.5 mΩ (0.5 ~ 1) mΩ (1 ~ 10) mΩ (10 ~ 100) mΩ (0.1 ~ 1) Ω	6.8×10^{-4} 5.1×10^{-4} 5.6×10^{-4} 3.9×10^{-4} 2.8×10^{-4}	

402. Resistance, capacitance and inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Resistors AC Resistance	40215	(1 ~ 10) Ω (0.04 ~ 1) kHz (0.001 ~ 1) MHz (1 ~ 2) MHz (2 ~ 4) MHz (4 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz (10 ~ 100) Ω (0.04 ~ 1) kHz (0.001 ~ 1) MHz (1 ~ 2) MHz (2 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz (0.1 ~ 1) k Ω (0.04 ~ 1) kHz (0.001 ~ 3) MHz (3 ~ 4) MHz (4 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz (1 ~ 10) k Ω 1 kHz 1 kHz ~ 1 MHz (10 ~ 100) k Ω 1 kHz 1 kHz ~ 1 MHz (0.1 ~ 1) M Ω 1 kHz	2.8×10^{-4} 5.7×10^{-4} 7.5×10^{-4} 8.4×10^{-4} 1.0×10^{-3} 4.0×10^{-3} 6.0×10^{-3} 2.8×10^{-4} 5.7×10^{-4} 6.6×10^{-4} 7.5×10^{-4} 2.0×10^{-3} 3.0×10^{-3} 4.6×10^{-4} 5.7×10^{-4} 6.6×10^{-4} 7.5×10^{-4} 2.0×10^{-3} 3.0×10^{-3} 3.7×10^{-4} 5.7×10^{-4} 3.7×10^{-4} 5.7×10^{-4} 4.8×10^{-4}	DMM, Megohmmeter, Meter calibrator, Standard resistance /KTICC-CI-40215
Impedance bridges/LCR meters Frequency Voltage	40217	10 Hz ~ 100 MHz (0 ~ 1) mV (0.04 ~ 10) kHz (10 ~ 50) kHz (50 ~ 100) kHz (0.1 ~ 1) MHz (1 ~ 10) mV 40 Hz (0.04 ~ 10) kHz (10 ~ 50) kHz (50 ~ 100) kHz (0.1 ~ 1) MHz (10 ~ 100) mV 40 Hz (0.04 ~ 10) kHz (10 ~ 50) kHz (50 ~ 100) kHz (0.1 ~ 1) MHz	6.1×10^{-7} 1.1×10^{-3} 1.6×10^{-3} 2.0×10^{-3} 9.7×10^{-3} 1.6×10^{-4} 1.5×10^{-4} 2.2×10^{-4} 2.7×10^{-4} 2.6×10^{-3} 7.4×10^{-5} 7.2×10^{-5} 7.7×10^{-5} 1.0×10^{-4} 1.0×10^{-3}	Standard Capacitors, Standard Inductors, Standard Resistors, Frequency Counter, DMM /KTICC-CI-40217

402. Resistance, capacitance and inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Impedance bridges/LCR meters	40217	(0.1 ~ 1) V		Standard Capacitors, Standard Inductors, Standard Resistors, Frequency Counter, DMM /KTICC-CI-40217
Voltage		40 Hz	6.7×10^{-5}	
		(0.04 ~ 10) kHz	6.5×10^{-5}	
		(10 ~ 50) kHz	6.7×10^{-5}	
		(50 ~ 100) kHz	7.7×10^{-5}	
		(0.1 ~ 1) MHz	9.6×10^{-4}	
		(1 ~ 2) V		
		40 Hz	4.1×10^{-5}	
		(0.04 ~ 10) kHz	3.7×10^{-5}	
		(10 ~ 50) kHz	4.0×10^{-5}	
		(50 ~ 100) kHz	5.5×10^{-5}	
		(0.1 ~ 1) MHz	9.5×10^{-4}	
		(2 ~ 5) V		
		40 Hz	3.2×10^{-5}	
		(0.04 ~ 1) kHz	2.4×10^{-5}	
		(1 ~ 10) kHz	2.6×10^{-5}	
		(10 ~ 50) kHz	3.0×10^{-5}	
		(50 ~ 100) kHz	5.8×10^{-5}	
		(0.1 ~ 1) MHz	1.2×10^{-3}	
		(5 ~ 10) V		
		40 Hz	3.0×10^{-5}	
		(0.04 ~ 10) kHz	2.3×10^{-5}	
		(10 ~ 50) kHz	2.9×10^{-5}	
		(50 ~ 100) kHz	5.5×10^{-5}	
		(0.1 ~ 1) MHz	1.2×10^{-3}	
		(10 ~ 20) V		
		40 Hz	3.2×10^{-5}	
		(0.04 ~ 10) kHz	2.4×10^{-5}	
		(10 ~ 50) kHz	3.0×10^{-5}	
		(50 ~ 100) kHz	5.5×10^{-5}	
		(0.1 ~ 1) MHz	1.2×10^{-3}	
DC Bias Voltage		(±)		
		0 V	1.0 μV	
		(0 ~ 0.1) V	1.1×10^{-5}	
		(0.1 ~ 0.5) V	1.3×10^{-5}	
		(0.5 ~ 1) V	7.2×10^{-6}	
		(1 ~ 5) V	1.3×10^{-5}	
		(5 ~ 10) V	7.2×10^{-6}	
		(10 ~ 20) V	7.0×10^{-6}	
		(20 ~ 40) V	5.7×10^{-6}	
DC Bias Current		(±)		
		0 A	6.1 μA	
		(0 ~ 0.1) A	1.3×10^{-4}	
		(0.1 ~ 18) A	2.5×10^{-4}	
		(18 ~ 40) A	5.8×10^{-4}	

402. Resistance, capacitance and inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Impedance bridges/LCR meters Capacitance	40217	1 pF 1 kHz (0.001 ~ 1) MHz (1 ~ 2) MHz (2 ~ 3) MHz (3 ~ 4) MHz (4 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz (1 ~ 10) pF 1 kHz (0.001 ~ 5) MHz (5 ~ 13) MHz (10 ~ 100) pF 1 kHz (0.001 ~ 2) MHz (2 ~ 4) MHz (4 ~ 5) MHz (5 ~ 13) MHz (0.1 ~ 1) nF 1 kHz (0.001 ~ 1) MHz (1 ~ 2) MHz (2 ~ 3) MHz (3 ~ 4) MHz (4 ~ 5) MHz (5 ~ 10) MHz (10 ~ 13) MHz (1 ~ 10) nF 120 Hz (0.12 ~ 1) kHz (1 ~ 100) kHz (10 ~ 100) nF 120 Hz (0.12 ~ 1) kHz (1 ~ 100) kHz (0.1 ~ 1) μF 120 Hz (0.12 ~ 1) kHz (1 ~ 10) kHz (10 ~ 100) kHz (1 ~ 10) μF (0.12 ~ 1) kHz (10 ~ 100) μF 120 Hz (0.12 ~ 1) kHz (0.1 ~ 10) mF (0.12 ~ 1) kHz (1 kHz) 100 μH (0.1 ~ 10) mH	 4.6×10^{-4} 4.3×10^{-4} 4.8×10^{-4} 5.8×10^{-4} 7.5×10^{-4} 9.7×10^{-4} 3.4×10^{-3} 4.3×10^{-3} 3.4×10^{-5} 4.2×10^{-4} 2.4×10^{-3} 3.4×10^{-5} 4.2×10^{-4} 4.3×10^{-4} 4.4×10^{-4} 2.4×10^{-3} 3.4×10^{-5} 4.3×10^{-4} 4.5×10^{-4} 5.1×10^{-4} 6.1×10^{-4} 7.5×10^{-4} 3.1×10^{-3} 3.8×10^{-3} 2.5×10^{-4} 1.3×10^{-4} 2.5×10^{-4} 2.5×10^{-4} 1.5×10^{-4} 2.5×10^{-4} 2.5×10^{-4} 1.5×10^{-4} 2.5×10^{-4} 2.6×10^{-4} 5.9×10^{-4} 6.1×10^{-4} 6.0×10^{-4} 2.9×10^{-3} 4.5×10^{-4} 3.2×10^{-4}	Standard Capacitors, Standrd Inductors, Standard Resistors, Frequency Counter, DMM /KTICC-CI-40217
Inductance				

402. Resistance, capacitance and inductance

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Impedance bridges/LCR meters	40217	(0.01 ~ 1) H	2.7×10^{-4}	Standard Capacitors, Standrd Inductors, Standard Resistors, Frequency Counter, DMM /KTICC-CI-40217
Inductance		(1 ~ 10) H	3.2×10^{-4}	
AC Resistance		1 m Ω		
		1 kHz	3.6×10^{-3}	
		(1 ~ 10) m Ω		
		1 kHz	8.1×10^{-4}	
		(10 ~ 100) m Ω		
		1 kHz	1.0×10^{-3}	
		(0.1 ~ 1) Ω		
		1 kHz	3.3×10^{-4}	
		(1 ~ 10) Ω		
		1 kHz	3.3×10^{-4}	
		(0.001 ~ 1) MHz	4.2×10^{-4}	
		(1 ~ 4) MHz	7.1×10^{-4}	
		(4 ~ 5) MHz	1.0×10^{-3}	
		(5 ~ 10) MHz	4.0×10^{-3}	
		(10 ~ 13) MHz	6.0×10^{-3}	
		(10 ~ 100) Ω		
		1 kHz	3.7×10^{-4}	
		(0.001 ~ 1) MHz	4.2×10^{-4}	
		(1 ~ 2) MHz	5.2×10^{-4}	
		(2 ~ 5) MHz	6.1×10^{-4}	
		(5 ~ 10) MHz	2.0×10^{-3}	
		(10 ~ 13) MHz	3.0×10^{-3}	
		(0.1 ~ 1) k Ω		
		1 kHz	3.7×10^{-4}	
		(0.001 ~ 3) MHz	4.2×10^{-4}	
		(3 ~ 4) MHz	5.2×10^{-4}	
		(4 ~ 5) MHz	6.1×10^{-4}	
		(5 ~ 10) MHz	2.0×10^{-3}	
		(10 ~ 13) MHz	3.0×10^{-3}	
		(1 ~ 10) k Ω		
		1 kHz	2.6×10^{-4}	
		1 kHz ~ 1 MHz	4.2×10^{-4}	
		(10 ~ 100) k Ω		
		1 kHz	2.6×10^{-4}	
		1 kHz ~ 1 MHz	4.2×10^{-4}	
		(0.1 ~ 1) M Ω		
		1 kHz	3.3×10^{-4}	
DC Resistance		(0 ~ 1) Ω	1.1×10^{-5}	
		(1 ~ 10) Ω	6.7×10^{-6}	
		(10 ~ 100) Ω	9.3×10^{-6}	
		(0.1 ~ 1) k Ω	6.7×10^{-6}	
		(1 ~ 10) k Ω	7.8×10^{-6}	
		(10 ~ 100) k Ω	7.5×10^{-6}	
		(0.1 ~ 1) M Ω	8.1×10^{-6}	
		(1 ~ 10) M Ω	1.1×10^{-5}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC ammeters AC Current	40301	9 μ A ~ 190 mA 10 Hz (10 ~ 40) Hz (0.04 ~ 1) kHz (1 ~ 10) kHz (0.19 ~ 1) A (0.01 ~ 1) kHz (1 ~ 10) kHz (1 ~ 1.9) A (0.01 ~ 1) kHz (1 ~ 10) kHz (1.9 ~ 10) A (45 ~ 100) Hz (0.1 ~ 1) kHz (10 ~ 20) A (45 ~ 100) Hz (0.1 ~ 1) kHz (20 ~ 100) A (40 ~ 60) Hz (0.06 ~ 1) kHz	3.1×10^{-4} 2.1×10^{-4} 1.7×10^{-4} 1.9×10^{-3} 3.3×10^{-4} 7.3×10^{-3} 3.1×10^{-4} 7.4×10^{-3} 1.4×10^{-3} 1.8×10^{-3} 2.2×10^{-3} 2.5×10^{-3} 3.6×10^{-4} 1.8×10^{-3}	Meter calibrator, Current calibrator /KTICC-CI-40301
Clamp ammeters/voltmeters DC Voltage AC Voltage	40302	(±) 0 mV (0 ~ 10) mV (0.01 ~ 1 000) V (0.22 ~ 10) mV 10 Hz (10 ~ 40) Hz (0.04 ~ 1) kHz (10 ~ 100) mV 10 Hz (10 ~ 40) Hz (0.04 ~ 1) kHz (0.1 ~ 1) V 10 Hz (10 ~ 40) Hz (0.04 ~ 1) kHz (1 ~ 10) V 10 Hz (10 ~ 40) Hz (0.04 ~ 1) kHz (10 ~ 100) V 10 Hz (10 ~ 40) Hz (0.04 ~ 1) kHz (100 ~ 1 000) V 50 Hz (0.05 ~ 1) kHz	0.76 μ V 7.8×10^{-5} 6.2×10^{-5} 6.6×10^{-4} 5.1×10^{-4} 5.0×10^{-4} 3.8×10^{-4} 1.8×10^{-4} 1.7×10^{-4} 3.0×10^{-4} 1.3×10^{-4} 9.1×10^{-5} 3.0×10^{-4} 1.3×10^{-4} 8.9×10^{-5} 3.0×10^{-4} 1.3×10^{-4} 9.4×10^{-5} 3.3×10^{-4} 1.0×10^{-4}	Meter calibrator, Turn coil, Current calibrator /KTICC-CI-40302

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Clamp ammeters/voltmeters	40302	(±)		Meter calibrator, Turn coil, Current calibrator /KTICC-CI-40302
DC Current		0 μA	61 nA	
		(0 ~ 1) A	6.2×10^{-4}	
		(1 ~ 10) A	9.9×10^{-4}	
		(10 ~ 20) A	1.1×10^{-3}	
		(20 ~ 100) A	1.6×10^{-3}	
		(100 ~ 200) A	1.7×10^{-3}	
		(200 ~ 500) A	1.5×10^{-3}	
		(500 ~ 1 000) A	1.7×10^{-3}	
		(1 000 ~ 2 500) A	1.2×10^{-3}	
AC Current		9 μA ~ 100 mA		
		(0.01 ~ 1) kHz	6.8×10^{-4}	
		(1 ~ 10) kHz	1.9×10^{-3}	
		(0.1 ~ 1) A		
		(0.01 ~ 1) kHz	6.9×10^{-4}	
		(1 ~ 10) kHz	7.3×10^{-3}	
		(1 ~ 10) A		
		(0.01 ~ 1) kHz	3.0×10^{-3}	
		(10 ~ 20) A		
		(0.01 ~ 1) kHz	3.1×10^{-3}	
		(20 ~ 100) A		
		(40 ~ 60) Hz	4.0×10^{-3}	
		(100 ~ 200) A		
		(40 ~ 60) Hz	3.7×10^{-3}	
		(200 ~ 300) A		
		(40 ~ 60) Hz	4.7×10^{-3}	
		(300 ~ 400) A		
		(40 ~ 60) Hz	4.3×10^{-3}	
		(400 ~ 500) A		
		(40 ~ 60) Hz	4.0×10^{-3}	
		(500 ~ 1 000) A		
		(40 ~ 60) Hz	3.7×10^{-3}	
		(1 000 ~ 2 000) A		
		(40 ~ 60) Hz	1.6×10^{-3}	
		(2 000 ~ 5 000) A		
		(40 ~ 60) Hz	3.0×10^{-3}	
Resistance		0 Ω	0.61 mΩ	
		(0 ~ 10) MΩ	6.1×10^{-4}	
Turn Current Coil				
DC Ratio		2	9.5×10^{-4}	
		(2 ~ 10)	1.4×10^{-3}	
		(10 ~ 25)	1.5×10^{-3}	
		(25 ~ 50)	1.4×10^{-3}	
AC Ratio		(60 Hz)		
		2	1.2×10^{-3}	
		(2 ~ 10)	2.2×10^{-3}	
		(10 ~ 25)	2.4×10^{-3}	
		(25 ~ 50)	2.2×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators AC Voltage	40303	(0.5 ~ 2) mV 10 Hz (0.01 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (2 ~ 20) mV 10 Hz (0.01 ~ 0.02) kHz (0.02 ~ 0.04) kHz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (20 ~ 200) mV 10 Hz (0.01 ~ 0.02) kHz (0.02 ~ 0.04) kHz (0.04 ~ 1) kHz (1 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (0.2 ~ 2) V 10 Hz (0.01 ~ 0.02) kHz (0.02 ~ 0.04) kHz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (2 ~ 20) V 10 Hz (0.01 ~ 0.02) kHz (0.02 ~ 0.04) kHz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz	6.0×10^{-4} 6.0×10^{-4} 9.0×10^{-4} 1.1×10^{-3} 1.9×10^{-3} 4.1×10^{-3} 6.5×10^{-3} 1.3×10^{-4} 1.3×10^{-4} 1.0×10^{-4} 9.5×10^{-5} 1.4×10^{-4} 1.8×10^{-4} 3.5×10^{-4} 8.5×10^{-4} 2.3×10^{-3} 6.5×10^{-5} 6.0×10^{-5} 3.6×10^{-5} 3.3×10^{-5} 3.4×10^{-5} 4.1×10^{-5} 7.5×10^{-5} 1.7×10^{-4} 3.0×10^{-4} 1.0×10^{-3} 5.5×10^{-5} 5.0×10^{-5} 2.8×10^{-5} 2.2×10^{-5} 2.7×10^{-5} 4.7×10^{-5} 1.3×10^{-4} 2.2×10^{-4} 9.5×10^{-4} 5.5×10^{-5} 5.0×10^{-5} 3.0×10^{-5} 2.3×10^{-5} 2.9×10^{-5} 5.5×10^{-5} 1.4×10^{-4} 3.5×10^{-4} 1.2×10^{-3}	DMM, Current shunt /KTICC-CI-40303

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators AC Voltage	40303	(20 ~ 200) V 10 Hz (0.01 ~ 0.02) kHz (0.02 ~ 0.04) kHz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (200 ~ 700) V 20 Hz (0.02 ~ 0.04) kHz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (700 ~ 1 000) V 40 Hz (0.04 ~ 1) kHz (1 ~ 20) kHz (20 ~ 30) kHz	6.5×10^{-5} 5.5×10^{-5} 3.3×10^{-5} 2.8×10^{-5} 4.0×10^{-5} 6.5×10^{-5} 5.3×10^{-5} 3.4×10^{-5} 2.9×10^{-5} 1.1×10^{-4} 6.6×10^{-4} 3.6×10^{-5} 2.9×10^{-5} 3.0×10^{-5} 1.1×10^{-4}	DMM, Current shunt /KTICC-CI-40303
AC Current		(0.009 ~ 1) mA 20 Hz (0.02 ~ 1) kHz (1 ~ 10) kHz (1 ~ 1.9) mA 20 Hz (0.02 ~ 1) kHz (1 ~ 10) kHz (1.9 ~ 10) mA 20 Hz (0.02 ~ 1) kHz (1 ~ 10) kHz (10 ~ 19) mA 20 Hz (0.02 ~ 1) kHz (1 ~ 10) kHz (19 ~ 100) mA 20 Hz (0.02 ~ 1) kHz (1 ~ 10) kHz (100 ~ 190) mA 20 Hz (0.02 ~ 1) kHz (1 ~ 10) kHz (0.19 ~ 1) A 20 Hz (0.02 ~ 1) kHz (1 ~ 10) kHz (1 ~ 1.9) A 20 Hz (0.02 ~ 1) kHz (1 ~ 10) kHz	4.4×10^{-4} 4.4×10^{-4} 1.7×10^{-3} 3.6×10^{-4} 3.6×10^{-4} 1.6×10^{-3} 4.4×10^{-4} 4.4×10^{-4} 1.5×10^{-3} 3.6×10^{-4} 3.6×10^{-4} 1.4×10^{-3} 4.4×10^{-4} 4.4×10^{-4} 1.2×10^{-3} 3.6×10^{-4} 3.6×10^{-4} 1.2×10^{-3} 8.7×10^{-4} 8.7×10^{-4} 6.3×10^{-3} 7.9×10^{-4} 7.9×10^{-4} 6.3×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltage/current calibrators AC Current	40303	(1.9 ~ 10) A 40 Hz (0.04 ~ 1) kHz (10 ~ 30) A 40 Hz (0.04 ~ 1) kHz (30 ~ 100) A 40 Hz (0.04 ~ 1) kHz	1.0×10^{-3} 1.0×10^{-3} 1.3×10^{-3} 1.3×10^{-3} 1.2×10^{-3} 1.2×10^{-3}	DMM, Current shunt /KTICC-CI-40303
Frequency		10 Hz ~ 1 MHz	6.1×10^{-7}	
AC current shunts AC Resistance	40305	(100 ~ 1 000) Ω 10 Hz (10 ~ 40) Hz (0.04 ~ 1) kHz (1 ~ 10) kHz (10 ~ 100) Ω 10 Hz (10 ~ 40) Hz (0.04 ~ 1) kHz (1 ~ 10) kHz (1 ~ 10) Ω 10 Hz (10 ~ 40) Hz (0.04 ~ 1) kHz (1 ~ 10) kHz (0.1 ~ 1) Ω 10 Hz (10 ~ 40) Hz (0.04 ~ 1) kHz (1 ~ 10) kHz (0.01 ~ 0.1) Ω 10 Hz (0.01 ~ 1) kHz (1 ~ 10) kHz (0.001 ~ 0.01) Ω 45 Hz (45 ~ 100) Hz (0.1 ~ 1) kHz (0.000 5 ~ 0.001) Ω 60 Hz 0.000 5 Ω 60 Hz	4.3×10^{-4} 2.4×10^{-4} 2.2×10^{-4} 1.8×10^{-3} 3.2×10^{-4} 2.2×10^{-4} 1.8×10^{-4} 1.8×10^{-3} 3.2×10^{-4} 2.2×10^{-4} 1.8×10^{-4} 1.6×10^{-3} 3.2×10^{-4} 2.3×10^{-4} 1.8×10^{-4} 1.2×10^{-3} 3.4×10^{-4} 3.3×10^{-4} 7.3×10^{-3} 1.4×10^{-3} 1.4×10^{-3} 1.8×10^{-3} 3.4×10^{-4} 3.4×10^{-4}	DMM, Meter calibrator Current calibrator /KTICC-CI-40305
Voltage/current phase angle meters /synchro resolve meters Phase	40307	(50 ~ 60) Hz -180° ~ 180°	0.043°	Power calibrator /KTICC-CI-40307

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Power factor meters Power Factor	40310	(50 ~ 60) Hz -1 ~ 1	(Absolute) 0.000 52	Power calibrator /KTICC-CI-40310
AC power meters DC Voltage	40311	($\pm$) 0 mV (0.1 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V	0.76 μ V 1.3×10^{-5} 8.5×10^{-6} 7.3×10^{-6} 8.5×10^{-6} 9.5×10^{-6}	Meter calibrator, Power calibrator /KTICC-CI-40311
AC Voltage		(0.04 ~ 1) kHz (0.22 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V	1.6×10^{-4} 6.8×10^{-5} 6.5×10^{-5} 7.2×10^{-5} 8.5×10^{-5}	
DC Current		($\pm$) 0 μ A 1 μ A ~ 1 mA (1 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 2) A (2 ~ 10) A (10 ~ 100) A (100 ~ 200) A (200 ~ 400) A (400 ~ 500) A (500 ~ 1 000) A (1 000 ~ 2 500) A	9.7 nA 4.4×10^{-5} 4.1×10^{-5} 5.4×10^{-5} 9.4×10^{-5} 8.8×10^{-5} 7.8×10^{-4} 1.4×10^{-4} 1.4×10^{-3} 1.5×10^{-3} 1.4×10^{-3} 1.7×10^{-3} 1.2×10^{-3}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311			Meter calibrator, Power calibrator /KTICC-CI-40311
AC Current		(0.04 ~ 1) kHz 9 μ A ~ 100 mA (0.1 ~ 1) A (1 ~ 2) A (2 ~ 10) A (10 ~ 20) A 60 Hz (20 ~ 100) A (100 ~ 200) A (200 ~ 300) A (300 ~ 400) A (400 ~ 500) A (500 ~ 800) A (800 ~ 2 000) A (2 000 ~ 5 000) A	1.8×10^{-4} 3.3×10^{-4} 3.1×10^{-4} 2.9×10^{-3} 3.1×10^{-3} 1.3×10^{-3} 3.6×10^{-3} 4.7×10^{-3} 4.3×10^{-3} 4.0×10^{-3} 3.8×10^{-3} 1.6×10^{-3} 3.0×10^{-3}	
AC Power		(50 ~ 60) Hz 0.24 mW (0.24 ~ 0.48) mW (0.000 48 ~ 1.2) W (1.2 ~ 2.4) W (2.4 ~ 12) W (12 ~ 24) W (24 ~ 60) W (60 ~ 120) W (120 ~ 240) W (240 ~ 600) W (0.6 ~ 1.2) kW (1.2 ~ 2.4) kW (2.4 ~ 3.6) kW (3.6 ~ 4.8) kW (4.8 ~ 6.0) kW (6.0 ~ 7.2) kW (7.2 ~ 9.6) kW (9.6 ~ 12) kW	2.6×10^{-2} 1.4×10^{-2} 1.6×10^{-3} 1.0×10^{-3} 5.8×10^{-4} 4.7×10^{-4} 1.7×10^{-4} 1.8×10^{-4} 1.7×10^{-4} 1.8×10^{-4} 3.3×10^{-4} 6.3×10^{-4} 5.8×10^{-4} 6.3×10^{-4} 5.3×10^{-4} 5.7×10^{-4} 5.4×10^{-4} 5.3×10^{-4}	
Power Factor		(50 ~ 60) Hz -1 ~ 1	(absolute) 0.000 52	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC power meters	40311			Meter calibrator, Power calibrator /KTICC-CI-40311
Frequency		10 Hz ~ 1 MHz	6.1×10^{-5}	
Harmonic Voltage		(50 ~ 3000) Hz 0.5 % (0.5 ~ 1) % (1 ~ 3) % (3 ~ 5) % (5 ~ 10) % (10 ~ 20) %	8.2×10^{-2} 4.1×10^{-2} 1.4×10^{-2} 8.2×10^{-3} 4.2×10^{-3} 2.2×10^{-3}	
Harmonic Current		(50 ~ 3000) Hz 0.5 % (0.5 ~ 1) % (1 ~ 3) % (3 ~ 5) % (5 ~ 10) % (10 ~ 20) %	8.4×10^{-2} 4.2×10^{-2} 1.4×10^{-2} 8.4×10^{-3} 4.2×10^{-3} 2.1×10^{-3}	
DC Power		1 mW (0.001 ~ 1) W (1 ~ 10) W (10 ~ 20) W (20 ~ 100) W (100 ~ 200) W (0.2 ~ 1) kW (1 ~ 2) kW (2 ~ 10) kW (10 ~ 20) kW	2.7×10^{-4} 2.7×10^{-4} 8.6×10^{-4} 1.0×10^{-3} 8.6×10^{-4} 1.0×10^{-3} 8.6×10^{-4} 1.0×10^{-3} 8.6×10^{-4} 1.0×10^{-3}	
AC power supplies	40312			DMM, Current shunt /KTICC-CI-40312
AC Voltage		(45 ~ 100) Hz (0 ~ 10) V (10 ~ 20) V (20 ~ 40) V (40 ~ 150) V (0.1 ~ 5) kHz (0 ~ 10) V (10 ~ 20) V (20 ~ 40) V (40 ~ 150) V (0.045 ~ 5) kHz (150 ~ 200) V (200 ~ 250) V (250 ~ 300) V (300 ~ 350) V (350 ~ 400) V (400 ~ 500) V	1.4×10^{-4} 2.0×10^{-4} 1.5×10^{-4} 1.4×10^{-4} 1.2×10^{-4} 1.8×10^{-4} 1.3×10^{-4} 1.2×10^{-4} 2.1×10^{-4} 1.8×10^{-4} 1.7×10^{-4} 2.4×10^{-4} 2.2×10^{-4} 2.0×10^{-4}	
AC Current		(0.045 ~ 1) kHz 9 μ A ~ 100 mA (0.1 ~ 1) A (1 ~ 100) A	4.6×10^{-4} 9.7×10^{-4} 1.2×10^{-3}	
Frequency		(10 ~ 5 000) Hz	6.1×10^{-5}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Puncture/safety testers	40313			High voltage meter, DMM High voltage test equipment / KTICC-CI-40313
DC Voltage		($\pm$) 1 V ~ 100 kV	1.2×10^{-3}	
AC Voltage		(50 ~ 60) Hz 1 V ~ 100 kV	1.3×10^{-2}	
DC Cutoff Current		($\pm$) (0.01 ~ 0.5) mA (0.5 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 10) mA (10 ~ 20) mA	9.6×10^{-3} 7.3×10^{-3} 6.5×10^{-3} 6.0×10^{-3} 6.1×10^{-3} 6.0×10^{-3}	
AC Cutoff Current		(50 ~ 60) Hz (0.01 ~ 0.5) mA (0.5 ~ 1) mA (1 ~ 2) mA (2 ~ 5) mA (5 ~ 10) mA (10 ~ 20) mA (20 ~ 50) mA (50 ~ 100) mA	3.0×10^{-2} 1.8×10^{-2} 1.3×10^{-2} 1.0×10^{-2} 9.3×10^{-3} 9.0×10^{-3} 1.2×10^{-2} 1.1×10^{-2}	
Timer		(1 ~ 1 000) s	6.1×10^{-5}	
Power recorders	40314			Meter calibrator, Power calibrator /KTICC-CI-40314
AC Voltage		(0.04 ~ 1) kHz (0.22 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V	1.6×10^{-4} 6.8×10^{-5} 6.5×10^{-5} 7.2×10^{-5} 8.5×10^{-5}	
AC Current		(0.04 ~ 1) kHz 9 μ A ~ 100 mA (0.1 ~ 1) A (1 ~ 2) A (2 ~ 10) A (10 ~ 20) A 60 Hz (20 ~ 100) A	1.8×10^{-4} 3.3×10^{-4} 3.1×10^{-4} 1.8×10^{-3} 2.5×10^{-3} 3.1×10^{-4}	
AC Power		(50 ~ 60) Hz, 120 V (0.06 ~ 600) W (0.6 ~ 1.2) kW (1.2 ~ 2.4) kW (2.4 ~ 3.6) kW (3.6 ~ 4.8) kW (4.8 ~ 6.0) kW	1.8×10^{-4} 3.3×10^{-4} 6.3×10^{-4} 5.8×10^{-4} 5.4×10^{-4} 5.3×10^{-4}	

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Power recorders AC Power	40314	(50 ~ 60) Hz, 240 V 0.06 W ~ 1.2 kW (1.2 ~ 2.4) kW (2.4 ~ 4.8) kW (4.8 ~ 7.2) kW (7.2 ~ 9.6) kW (9.6 ~ 12) kW	1.8×10^{-4} 3.2×10^{-4} 6.3×10^{-4} 5.7×10^{-4} 5.4×10^{-4} 5.3×10^{-4}	Meter calibrator, Power calibrator /KTICC-CI-40314
AC voltmeters AC Voltage	40318	(0.1 ~ 1) mV (10 ~ 20) Hz (0.02 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 200) kHz (200 ~ 500) kHz (0.5 ~ 1) MHz (1 ~ 10) mV (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 200) kHz (200 ~ 500) kHz (0.5 ~ 1) MHz (10 ~ 100) mV (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 200) kHz (200 ~ 500) kHz (0.5 ~ 1) MHz (0.1 ~ 1) V (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 200) kHz (200 ~ 500) kHz (0.5 ~ 1) MHz	4.3×10^{-3} 4.2×10^{-3} 4.3×10^{-3} 5.6×10^{-3} 1.1×10^{-2} 2.2×10^{-2} 2.3×10^{-2} 6.6×10^{-4} 5.1×10^{-4} 5.0×10^{-4} 6.2×10^{-4} 1.0×10^{-3} 2.1×10^{-3} 3.5×10^{-3} 4.8×10^{-3} 3.8×10^{-4} 1.8×10^{-4} 1.7×10^{-4} 2.9×10^{-4} 6.5×10^{-4} 1.1×10^{-3} 1.7×10^{-3} 3.2×10^{-3} 3.0×10^{-4} 1.3×10^{-4} 9.1×10^{-5} 1.1×10^{-4} 1.6×10^{-4} 5.2×10^{-4} 1.2×10^{-3} 2.0×10^{-3}	Meter calibrator /KTICC-CI-40318

403. AC voltage, current & power

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
AC voltmeters AC Voltage	40318	(1 ~ 10) V (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 200) kHz (200 ~ 500) kHz (0.5 ~ 1) MHz (10 ~ 100) V (20 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 1 000) V (15 ~ 50) Hz (0.05 ~ 1) kHz (0.000 1 ~ 0.33) V (1 ~ 30) MHz (0.33 ~ 3.5) V (1 ~ 30) MHz	3.0×10^{-4} 1.3×10^{-4} 8.9×10^{-5} 1.1×10^{-4} 1.4×10^{-4} 3.6×10^{-4} 1.2×10^{-3} 1.9×10^{-3} 1.3×10^{-4} 9.4×10^{-5} 1.2×10^{-4} 2.0×10^{-4} 3.3×10^{-4} 1.0×10^{-4} 1.3×10^{-2} 1.4×10^{-2}	Meter calibrator /KTICC-CI-40318

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF amplifiers DC Gain	40401	1 (1 ~ 3.162 3) (3.162 3 ~ 10) (10 ~ 50) (50 ~ 1 000)	1.4×10^{-5} 5.7×10^{-5} 1.4×10^{-5} 6.3×10^{-5} 6.3×10^{-5}	Meter calibrator, DMM /KTICC-CI-40401
LF Gain		10 Hz 1 (1 ~ 31.623) (31.623 ~ 40) (40 ~ 50) (50 ~ 100) (100 ~ 1 000)	2.6×10^{-4} 3.8×10^{-4} 4.3×10^{-4} 2.8×10^{-4} 3.0×10^{-4} 3.8×10^{-4}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF amplifiers LF Gain	40401	(0.01 ~ 20) kHz 1 (1 ~ 3.162 3) (3.162 3 ~ 10) (10 ~ 20) (20 ~ 40) (40 ~ 50) (50 ~ 100) (100 ~ 316.23) (316.23 ~ 1 000) (20 ~ 50) kHz 1 (1 ~ 316.23) (316.23 ~ 1 000) (50 ~ 100) kHz (1 ~ 3.162 3) (3.162 3 ~ 20) (20 ~ 316.23) (316.23 ~ 1 000) (0.1 ~ 1) MHz (1 ~ 10) (10 ~ 20) (20 ~ 31.623) (31.623 ~ 40) (40 ~ 316.23) (316.23 ~ 600) (600 ~ 800) (800 ~ 1 000)	7.0×10^{-5} 9.8×10^{-5} 7.5×10^{-5} 9.5×10^{-5} 1.0×10^{-4} 9.2×10^{-5} 9.5×10^{-5} 1.1×10^{-4} 1.7×10^{-4} 1.1×10^{-4} 1.5×10^{-4} 2.9×10^{-4} 1.9×10^{-4} 1.8×10^{-4} 2.3×10^{-4} 6.5×10^{-4} 3.7×10^{-3} 4.7×10^{-3} 5.7×10^{-3} 6.8×10^{-3} 7.6×10^{-3} 9.8×10^{-3} 1.2×10^{-2} 1.4×10^{-2}	Meter calibrator, DMM /KTICC-CI-40401
DC Bias Voltage		(±) 0 V (0 ~ 10) V (10 ~ 20) V (20 ~ 40) V (40 ~ 60) V (60 ~ 80) V (80 ~ 100) V (100 ~ 150) V (150 ~ 200) V	61 μV 6.1×10^{-5} 3.2×10^{-5} 1.7×10^{-5} 1.2×10^{-5} 9.8×10^{-6} 6.1×10^{-5} 4.1×10^{-5} 3.1×10^{-5}	
DC/LF attenuators LF Attenuator	40402	(0.01 ~ 100) kHz (0 ~ 70) dB (70 ~ 80) dB (100 ~ 300) kHz (0 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB	0.008 dB 0.014 dB 0.008 dB 0.014 dB 0.038 dB	Audio analyzer, DMM Meter calibrator /KTICC-CI-40402

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DC/LF attenuators LF Attenuator	40402	(300 ~ 500) kHz (0 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB (0.5 ~ 1) MHz (0 ~ 50) dB (50 ~ 60) dB (60 ~ 70) dB (70 ~ 80) dB	0.008 dB 0.013 dB 0.028 dB 0.078 dB 0.013 dB 0.025 dB 0.045 dB 0.099 dB	Audio analyzer, DMM Meter calibrator /KTICC-CI-40402
Multimeter calibrators DC Voltage Resistance DC Current	40403	(±) 0 mV (0 ~ 100) mV (100 ~ 190) mV (0.19 ~ 1) V (1 ~ 10) V (10 ~ 19) V (19 ~ 100) V (100 ~ 190) V (190 ~ 1 000) V (0 ~ 1) Ω (1 ~ 1.9) Ω (1.9 ~ 10) Ω (10 ~ 19) Ω (19 ~ 100) Ω (100 ~ 190) Ω (0.19 ~ 19) kΩ (19 ~ 100) kΩ (100 ~ 190) kΩ (0.19 ~ 1) MΩ (1 ~ 1.9) MΩ (1.9 ~ 10) MΩ (10 ~ 19) MΩ (19 ~ 100) MΩ (±) 0 μA (0 ~ 19) μA (19 ~ 100) μA (100 ~ 190) μA (0.19 ~ 1) mA (1 ~ 1.9) mA (1.9 ~ 10) mA (10 ~ 19) mA (19 ~ 190) mA (0.19 ~ 1) A (1 ~ 1.9) A (1.9 ~ 10) A (10 ~ 30) A	0.12 μV 8.9×10^{-6} 8.4×10^{-6} 3.8×10^{-6} 3.8×10^{-6} 3.6×10^{-6} 6.0×10^{-6} 5.8×10^{-6} 6.1×10^{-6} 1.6×10^{-5} 1.4×10^{-5} 1.1×10^{-5} 1.0×10^{-5} 8.7×10^{-6} 8.4×10^{-6} 7.4×10^{-6} 7.5×10^{-6} 7.4×10^{-6} 9.6×10^{-6} 8.9×10^{-6} 2.1×10^{-5} 1.7×10^{-5} 1.4×10^{-4} 12 pA 1.0×10^{-5} 6.5×10^{-6} 6.3×10^{-6} 4.9×10^{-6} 4.8×10^{-6} 5.3×10^{-6} 5.0×10^{-6} 1.3×10^{-5} 5.9×10^{-5} 5.8×10^{-5} 8.9×10^{-5} 1.2×10^{-4}	DC standard, DMM AC/DC transfer standard, Current shunt Standard resistance Meter calibrator /KTICC-CI-40403

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators AC Voltage	40403	(0.5 ~ 2) mV 10 Hz (0.01 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (2 ~ 20) mV 10 Hz (0.01 ~ 0.02) kHz (0.02 ~ 0.04) kHz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (100 ~ 500) kHz (0.5 ~ 1) MHz (20 ~ 200) mV 10 Hz (0.01 ~ 0.02) kHz (0.02 ~ 0.04) kHz (0.04 ~ 1) kHz (1 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (0.2 ~ 2) V 10 Hz (0.01 ~ 0.02) (0.02 ~ 0.04) (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz (2 ~ 20) V 10 Hz (0.01 ~ 0.02) kHz (0.02 ~ 0.04) kHz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 300) kHz (300 ~ 500) kHz (0.5 ~ 1) MHz	6.0×10^{-4} 6.0×10^{-4} 9.0×10^{-4} 1.1×10^{-3} 1.9×10^{-3} 4.1×10^{-3} 6.5×10^{-3} 1.3×10^{-4} 1.3×10^{-4} 1.0×10^{-4} 9.5×10^{-5} 1.4×10^{-4} 1.8×10^{-4} 3.5×10^{-4} 8.5×10^{-4} 2.3×10^{-3} 6.5×10^{-5} 6.0×10^{-5} 3.6×10^{-5} 3.3×10^{-5} 3.4×10^{-5} 4.1×10^{-5} 7.5×10^{-5} 1.7×10^{-4} 3.0×10^{-4} 1.0×10^{-3} 5.5×10^{-5} 5.0×10^{-5} 2.8×10^{-5} 2.2×10^{-5} 2.7×10^{-5} 4.7×10^{-5} 1.3×10^{-4} 2.2×10^{-4} 9.5×10^{-4} 5.5×10^{-5} 5.0×10^{-5} 3.0×10^{-5} 2.3×10^{-5} 2.9×10^{-5} 5.5×10^{-5} 1.4×10^{-4} 3.5×10^{-4} 1.2×10^{-3}	DC standard, DMM AC/DC transfer standard, Current shunt Standard resistance Meter calibrator /KTICC-CI-40403

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators AC Voltage	40403	(20 ~ 200) V 10 Hz (0.01 ~ 0.02) kHz (0.02 ~ 0.04) kHz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (200 ~ 700) V 20 Hz (0.02 ~ 0.04) kHz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (700 ~ 1 000) V 40 Hz (0.04 ~ 1) kHz (1 ~ 20) kHz (20 ~ 30) kHz	6.5×10^{-5} 5.5×10^{-5} 3.3×10^{-5} 2.8×10^{-5} 4.0×10^{-5} 6.5×10^{-5} 5.3×10^{-5} 3.4×10^{-5} 2.9×10^{-5} 1.1×10^{-4} 6.6×10^{-4} 3.6×10^{-5} 2.9×10^{-5} 3.0×10^{-5} 1.1×10^{-4}	DC standard, DMM AC/DC transfer standard, Current shunt Standard resistance Meter calibrator /KTICC-CI-40403
AC Current		(0.009 ~ 1) mA 20 Hz (0.02 ~ 1) kHz (1 ~ 10) kHz (1 ~ 1.9) mA 20 Hz (0.02 ~ 1) kHz (1 ~ 10) kHz (1.9 ~ 10) mA 20 Hz (0.02 ~ 1) kHz (1 ~ 10) kHz (10 ~ 19) mA 20 Hz (0.02 ~ 1) kHz (1 ~ 10) kHz (19 ~ 100) mA 20 Hz (0.02 ~ 1) kHz (1 ~ 10) kHz (100 ~ 190) mA 20 Hz (0.02 ~ 1) kHz (1 ~ 10) kHz (0.19 ~ 1) A 20 Hz (0.02 ~ 1) kHz (1 ~ 10) kHz (1 ~ 1.9) A 20 Hz (0.02 ~ 1) kHz (1 ~ 10) kHz	4.4×10^{-4} 4.4×10^{-4} 1.7×10^{-3} 3.6×10^{-4} 3.6×10^{-4} 1.6×10^{-3} 4.4×10^{-4} 4.4×10^{-4} 1.5×10^{-3} 3.6×10^{-4} 3.6×10^{-4} 1.4×10^{-3} 4.4×10^{-4} 4.4×10^{-4} 1.2×10^{-3} 3.6×10^{-4} 3.6×10^{-4} 1.2×10^{-3} 8.7×10^{-4} 8.7×10^{-4} 6.6×10^{-3} 7.9×10^{-4} 7.9×10^{-4} 6.3×10^{-3}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multimeter calibrators	40403			DC standard, DMM
AC Current		(1.9 ~ 10) A		AC/DC transfer standard,
		40 Hz	1.1×10^{-3}	Current shunt
		(0.04 ~ 1) kHz	1.1×10^{-3}	Standard resistance
		(10 ~ 30) A		Meter calibrator
		40 Hz	1.3×10^{-3}	/KTICC-CI-40403
		(0.04 ~ 1) kHz	1.3×10^{-3}	
Frequency		10 Hz ~ 1 MHz	6.1×10^{-7}	
Widband Voltage		(0.5 ~ 1) mV		
		10 Hz	8.7×10^{-4}	
		(10 ~ 30) Hz	8.1×10^{-4}	
		(0.03 ~ 100) kHz	4.5×10^{-4}	
		(0.1 ~ 2) MHz	1.3×10^{-3}	
		(2 ~ 10) MHz	2.1×10^{-3}	
		(10 ~ 20) MHz	3.2×10^{-3}	
		(20 ~ 30) MHz	7.1×10^{-3}	
		(30 ~ 50) MHz	1.0×10^{-2}	
		(1 ~ 3.2) mV		
		10 Hz	8.1×10^{-4}	
		(10 ~ 30) Hz	7.8×10^{-4}	
		(0.03 ~ 100) kHz	4.1×10^{-4}	
		(0.1 ~ 2) MHz	8.1×10^{-4}	
		(2 ~ 10) MHz	1.1×10^{-3}	
		(10 ~ 20) MHz	1.7×10^{-3}	
		(20 ~ 30) MHz	3.4×10^{-3}	
		(30 ~ 50) MHz	4.7×10^{-3}	
		(3.2 ~ 32) mV		
		10 Hz	8.1×10^{-4}	
		(10 ~ 30) Hz	8.0×10^{-4}	
		(0.03 ~ 100) kHz	4.1×10^{-4}	
		(0.1 ~ 2) MHz	5.6×10^{-4}	
		(2 ~ 10) MHz	8.4×10^{-4}	
		(10 ~ 20) MHz	1.5×10^{-3}	
		(20 ~ 30) MHz	3.2×10^{-3}	
		(30 ~ 50) MHz	5.1×10^{-3}	
		(0.032 ~ 3.5) V		
		10 Hz	8.0×10^{-4}	
		(10 ~ 30) Hz	8.0×10^{-4}	
		(0.03 ~ 100) kHz	3.2×10^{-4}	
		(0.1 ~ 2) MHz	4.1×10^{-4}	
		(2 ~ 10) MHz	8.5×10^{-4}	
		(10 ~ 20) MHz	1.3×10^{-3}	
		(20 ~ 30) MHz	3.1×10^{-3}	
		(30 ~ 50) MHz	5.1×10^{-3}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Oscilloscope calibrators Output DC Voltage	40404	(±) 0 V (0 ~ 1) mV (1 ~ 10) mV (10 ~ 100) mV (100 ~ 300) mV (300 ~ 500) mV (0.5 ~ 1) V (1 ~ 3) V (3 ~ 5) V (5 ~ 10) V (10 ~ 30) V (30 ~ 50) V (50 ~ 100) V (100 ~ 130) V (130 ~ 200) V	0.12 μV 1.2×10^{-4} 1.6×10^{-5} 1.1×10^{-5} 5.0×10^{-6} 4.2×10^{-6} 7.2×10^{-6} 5.3×10^{-6} 5.0×10^{-6} 7.2×10^{-6} 7.0×10^{-6} 6.4×10^{-6} 8.6×10^{-6} 7.6×10^{-6} 8.5×10^{-6}	DMM, Frequency counter, RF power meter / KTICC-CI-40404
Output AC Voltage		(0.05 ~ 10) kHz (0.5 ~ 1) mV (1 ~ 3) mV (3 ~ 5) mV (5 ~ 10) mV (10 ~ 30) mV (30 ~ 50) mV (50 ~ 100) mV (100 ~ 300) mV (300 ~ 500) mV (0.5 ~ 1) V (1 ~ 5) V (5 ~ 10) V (10 ~ 30) V (30 ~ 50) V (50 ~ 100) V (100 ~ 200) V	1.1×10^{-3} 3.7×10^{-4} 2.4×10^{-4} 1.4×10^{-4} 7.3×10^{-5} 6.2×10^{-5} 3.8×10^{-5} 2.9×10^{-5} 2.8×10^{-5} 2.3×10^{-5} 2.2×10^{-5} 2.3×10^{-5} 2.7×10^{-5} 2.8×10^{-5} 2.9×10^{-5} 2.8×10^{-5}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Oscilloscope calibrators Time Marker	40404	1 ns	6.1×10^{-8}	DMM, Frequency counter, RF power meter / KTICC-CI-40404
		(1 ~ 2) ns	3.1×10^{-8}	
		(2 ~ 5) ns	1.2×10^{-8}	
		(5 ~ 10) ns	6.1×10^{-8}	
		(10 ~ 20) ns	3.1×10^{-8}	
		(20 ~ 50) ns	1.2×10^{-8}	
		(50 ~ 100) ns	6.1×10^{-8}	
		(100 ~ 200) ns	3.1×10^{-8}	
		(200 ~ 500) ns	1.2×10^{-8}	
		(0.5 ~ 1) μ s	6.1×10^{-8}	
		(1 ~ 2) μ s	3.1×10^{-8}	
		(2 ~ 5) μ s	1.2×10^{-8}	
		(5 ~ 10) μ s	6.1×10^{-8}	
		(10 ~ 20) μ s	3.1×10^{-8}	
		(20 ~ 50) μ s	1.2×10^{-8}	
		(50 ~ 100) μ s	6.1×10^{-8}	
		(100 ~ 200) μ s	3.1×10^{-8}	
		(200 ~ 500) μ s	1.2×10^{-8}	
		(0.5 ~ 1) ms	6.1×10^{-8}	
		(1 ~ 2) ms	3.1×10^{-8}	
		(2 ~ 5) ms	1.2×10^{-8}	
		(5 ~ 10) ms	6.1×10^{-8}	
		(10 ~ 20) ms	3.1×10^{-8}	
		(20 ~ 50) ms	1.2×10^{-7}	
		(50 ~ 100) ms	6.1×10^{-7}	
		(100 ~ 200) ms	3.1×10^{-7}	
		(200 ~ 500) ms	1.2×10^{-7}	
		(0.5 ~ 1) s	6.1×10^{-7}	
		(1 ~ 2) s	3.1×10^{-7}	
		(2 ~ 5) s	1.2×10^{-7}	
Level Sine Wave Amplitude		(1 ~ 50) kHz		
		5 mV	1.8×10^{-3}	
		(5 ~ 10) mV	7.6×10^{-4}	
		(10 ~ 30) mV	5.0×10^{-4}	
		(30 ~ 100) mV	5.1×10^{-4}	
		(100 ~ 600) mV	4.5×10^{-4}	
		(0.6 ~ 5.5) V	3.9×10^{-4}	
Level Sine Wave Flatness		50 kHz ~ 6 GHz		
		30 mV	2.0×10^{-2}	
		(0.03 ~ 3) V	1.8×10^{-2}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Video signal generators	40406			Video signal analyzers /KTICC-CI-40406
Frequency		10 Hz ~ 5 GHz	3.7×10^{-10}	
Luminance Level		NTSC, PAL (1 ~ 100) mV (100 ~ 200) mV (200 ~ 300) mV (300 ~ 400) mV (400 ~ 500) mV (500 ~ 600) mV (600 ~ 714.3) mV	0.70 mV 0.92 mV 1.2 mV 1.5 mV 1.8 mV 2.5 mV 2.5 mV	
Chrominance Level		NTSC, PAL (100 ~ 714.3) mV	8.4×10^{-3}	
Phase		NTSC, PAL (0 ~ 360)°	1.4°	
H Timing		H Blanking, Sync to Setup (8.8 ~ 13) µs Sync to Burst Start / End (4 ~ 10) µs Sync Width (1 ~ 8) µs Burst Width (1.4 ~ 3) µs Front Proch (0.1 ~ 3.5) µs Rise Time, Fall Time 80 ns ~ 1 µs	0.024 µs 0.036 µs 0.024 µs 0.036 µs 0.024 µs 12 ns	
Audio distortion analyzers/meters	40407			Meter calibrator, Frequency counter, Audio analyzer /KTICC-CI-40407
Frequency		(0.01 ~ 100) kHz (100 ~ 200) kHz	6.1×10^{-6} 3.1×10^{-6}	
AC Voltage		(0.22 ~ 1) mV 10 Hz (10 ~ 40) Hz (0.04 ~ 10) kHz (10 ~ 100) kHz (100 ~ 200) kHz (1 ~ 10) mV 10 Hz (0.01 ~ 10) kHz (10 ~ 100) kHz (100 ~ 200) kHz (10 ~ 100) mV 10 Hz (0.01 ~ 10) kHz (10 ~ 100) kHz (100 ~ 200) kHz	4.4×10^{-3} 5.0×10^{-3} 4.2×10^{-3} 5.6×10^{-3} 1.1×10^{-2} 6.6×10^{-4} 5.1×10^{-4} 1.0×10^{-3} 2.1×10^{-3} 3.8×10^{-4} 1.8×10^{-4} 6.5×10^{-4} 1.1×10^{-3}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Audio distortion analyzers/meters AC Voltage	40407	(0.1 ~ 1) V 10 Hz (10 ~ 40) Hz (0.04 ~ 10) kHz (10 ~ 100) kHz (100 ~ 200) kHz (1 ~ 10) V 10 Hz (10 ~ 40) Hz (0.04 ~ 10) kHz (10 ~ 100) kHz (100 ~ 200) kHz (10 ~ 100) V 10 Hz (10 ~ 40) Hz (0.04 ~ 10) kHz (10 ~ 100) kHz (100 ~ 200) V 10 Hz (10 ~ 40) Hz (0.04 ~ 10) kHz (10 ~ 100) kHz (200 ~ 300) V 50 Hz (0.05 ~ 1) kHz (±) 0 V (0 ~ 100) mV (0.1 ~ 100) V (100 ~ 300) V	3.0×10^{-4} 1.3×10^{-4} 9.1×10^{-5} 1.6×10^{-4} 5.2×10^{-4} 3.0×10^{-4} 1.3×10^{-4} 8.9×10^{-5} 1.5×10^{-4} 3.6×10^{-4} 3.0×10^{-4} 1.3×10^{-4} 9.4×10^{-5} 2.0×10^{-4} 2.7×10^{-4} 1.1×10^{-4} 7.5×10^{-5} 1.9×10^{-4} 3.7×10^{-4} 9.3×10^{-5} $6.1 \mu\text{V}$ 6.2×10^{-5} 6.1×10^{-5} 2.2×10^{-5}	Meter calibrator, Frequency counter, Audio analyzer /KTICC-CI-40407
DC Voltage				
Level		10 Hz (40 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm (-60 ~ -70) dBm (0.01 ~ 1) kHz (50 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm (-60 ~ -70) dBm (1 ~ 10) kHz (40 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm (-60 ~ -70) dBm	0.009 dB 0.018 dB 0.048 dB 0.15 dB 0.009 dB 0.016 dB 0.047 dB 0.14 dB 0.008 dB 0.016 dB 0.047 dB 0.14 dB	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Audio distortion analyzers/meters Level	40407	(10 ~ 100) kHz (40 ~ -20) dBm (-20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm (-60 ~ -70) dBm (100 ~ 200) kHz (20 ~ -10) dBm (-10 ~ -20) dBm (-20 ~ -30) dBm (-30 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm (-60 ~ -70) dBm	0.009 dB 0.012 dB 0.023 dB 0.062 dB 0.18 dB 0.009 dB 0.012 dB 0.017 dB 0.022 dB 0.046 dB 0.12 dB 0.36 dB	Meter calibrator, Frequency counter, Audio analyzer /KTICC-CI-40407
Frequency Responses (Voltage)		100 mV (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 10) kHz (10 ~ 100) kHz (100 ~ 200) kHz (0.1 ~ 1) V (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 10) kHz (10 ~ 100) kHz (100 ~ 200) kHz (1 ~ 10) V (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 10) kHz (10 ~ 100) kHz (100 ~ 200) kHz	3.8×10^{-4} 1.8×10^{-4} 1.7×10^{-4} 6.5×10^{-4} 1.1×10^{-3} 3.0×10^{-4} 1.3×10^{-4} 9.1×10^{-5} 1.6×10^{-4} 5.2×10^{-4} 3.0×10^{-4} 1.3×10^{-4} 8.9×10^{-5} 1.5×10^{-4} 3.6×10^{-4}	
Frequency Responses (Level)		(10 ~ -10) dBm (10 ~ 40) Hz (0.04 ~ 100) kHz (100 ~ 200) kHz	0.008 dB 0.007 dB 0.010 dB	
Distortion		(0.02 ~ 100) kHz (0 ~ -60) dB (-60 ~ -80) dB	0.19 dB 0.50 dB	
Filter(Frequency) (Weight, Low, High Pass)		(0.01 ~ 100) kHz	6.1×10^{-6}	
Filter(Level) (Weight, Low, High Pass)		(0.01 ~ 100) kHz (20 ~ -63) dB	0.007 dB	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF filters Filter(Frequency) (Weight, Low, High Pass etc.) Filter(Level) (Weight, Low, High Pass etc.)	40408	10 Hz ~ 1 MHz 20 Hz (0 ~ -30) dB (-30 ~ -50) dB (-50 ~ -60) dB (0.02 ~ 100) kHz (0 ~ -30) dB (-30 ~ -50) dB (-50 ~ -60) dB (0.1 ~ 1) MHz (0 ~ -20) dB (-20 ~ -30) dB (-30 ~ -40) dB (-40 ~ -50) dB (-50 ~ -60) dB	6.1×10^{-6} 0.10 dB 0.20 dB 0.30 dB 0.054 dB 0.10 dB 0.21 dB 0.076 dB 0.086 dB 0.20 dB 0.22 dB 0.38 dB	Meter calibrator, LF signal generator, Frequency counter, Audio analyzer /KTICC-CI-40408
LF/audio signal analyzers Input Frequency Input AC Voltage	40409	(0.01 ~ 100) kHz (100 ~ 200) kHz (0.22 ~ 1) mV 10 Hz (10 ~ 40) Hz (0.04 ~ 10) kHz (10 ~ 100) kHz (100 ~ 200) kHz (1 ~ 10) mV 10 Hz (0.01 ~ 10) kHz (10 ~ 100) kHz (100 ~ 200) kHz (10 ~ 100) mV 10 Hz (0.01 ~ 10) kHz (10 ~ 100) kHz (100 ~ 200) kHz (0.01 ~ 1) V 10 Hz (10 ~ 40) Hz (0.04 ~ 10) kHz (10 ~ 100) kHz (100 ~ 200) kHz (1 ~ 10) V 10 Hz (10 ~ 40) Hz (0.04 ~ 10) kHz (10 ~ 100) kHz (100 ~ 200) kHz	6.1×10^{-6} 3.1×10^{-6} 4.4×10^{-3} 5.0×10^{-3} 4.2×10^{-3} 5.6×10^{-3} 1.1×10^{-2} 6.6×10^{-4} 5.1×10^{-4} 1.0×10^{-3} 2.1×10^{-3} 3.8×10^{-4} 1.8×10^{-4} 6.5×10^{-4} 1.1×10^{-3} 3.0×10^{-4} 1.3×10^{-4} 9.1×10^{-5} 1.6×10^{-4} 5.2×10^{-4} 3.0×10^{-4} 1.3×10^{-4} 8.9×10^{-5} 1.5×10^{-4} 3.6×10^{-4}	Meter calibrator, DMM Frequency counter, Audio analyzer /KTICC-CI-40409

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF/audio signal analyzers	40409	(10 ~ 100) V		Meter calibrator, DMM Frequency counter, Audio analyzer /KTICC-CI-40409
Input AC Voltage		10 Hz	3.0×10^{-4}	
		(10 ~ 40) Hz	1.3×10^{-4}	
		(0.04 ~ 10) kHz	9.4×10^{-5}	
		(10 ~ 100) kHz	2.0×10^{-4}	
		(100 ~ 200) V		
		10 Hz	2.7×10^{-4}	
		(10 ~ 40) Hz	1.1×10^{-4}	
		(0.04 ~ 10) kHz	7.5×10^{-5}	
		(10 ~ 100) kHz	1.9×10^{-4}	
		(200 ~ 300) V		
		50 Hz	3.7×10^{-4}	
		(0.05 ~ 1) kHz	9.3×10^{-5}	
Input DC Voltage		(±)		
		0 V	6.1 μV	
		(0 ~ 100) mV	6.2×10^{-5}	
		(0.1 ~ 100) V	6.1×10^{-5}	
		(100 ~ 300) V	2.2×10^{-5}	
Input Level		10 Hz		
		(40 ~ -40) dBm	0.009 dB	
		(-40 ~ -50) dBm	0.018 dB	
		(-50 ~ -60) dBm	0.048 dB	
		(-60 ~ -70) dBm	0.15 dB	
		(0.01 ~ 1) kHz		
		(50 ~ -40) dBm	0.009 dB	
		(-40 ~ -50) dBm	0.016 dB	
		(-50 ~ -60) dBm	0.047 dB	
		(-60 ~ -70) dBm	0.14 dB	
		(1 ~ 10) kHz		
		(40 ~ -40) dBm	0.008 dB	
		(-40 ~ -50) dBm	0.016 dB	
		(-50 ~ -60) dBm	0.047 dB	
		(-60 ~ -70) dBm	0.14 dB	
		(10 ~ 100) kHz		
		(40 ~ -20) dBm	0.009 dB	
		(-20 ~ -40) dBm	0.012 dB	
		(-40 ~ -50) dBm	0.023 dB	
		(-50 ~ -60) dBm	0.062 dB	
		(-60 ~ -70) dBm	0.18 dB	
		(100 ~ 200) kHz		
		(20 ~ -10) dBm	0.009 dB	
		(-10 ~ -20) dBm	0.012 dB	
		(-20 ~ -30) dBm	0.017 dB	
		(-30 ~ -40) dBm	0.022 dB	
		(-40 ~ -50) dBm	0.046 dB	
		(-50 ~ -60) dBm	0.12 dB	
		(-60 ~ -70) dBm	0.36 dB	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF/audio signal analyzers	40409	100 mV		Meter calibrator, DMM Frequency counter, Audio analyzer /KTICC-CI-40409
Frequency Responses (Voltage)		(10 ~ 20) Hz	3.8×10^{-4}	
		(20 ~ 40) Hz	1.8×10^{-4}	
		(0.04 ~ 10) kHz	1.7×10^{-4}	
		(10 ~ 100) kHz	6.5×10^{-4}	
		(100 ~ 200) kHz	1.1×10^{-3}	
		(0.1 ~ 1) V		
		(10 ~ 20) Hz	3.0×10^{-4}	
		(20 ~ 40) Hz	1.3×10^{-4}	
		(0.04 ~ 10) kHz	9.1×10^{-5}	
		(10 ~ 100) kHz	1.6×10^{-4}	
		(100 ~ 200) kHz	5.2×10^{-4}	
		(1 ~ 10) V		
		(10 ~ 20) Hz	3.0×10^{-4}	
		(20 ~ 40) Hz	1.3×10^{-4}	
		(0.04 ~ 10) kHz	8.9×10^{-5}	
		(10 ~ 100) kHz	1.5×10^{-4}	
		(100 ~ 200) kHz	3.6×10^{-4}	
Frequency Responses (Level)		(10 ~ -10) dBm		
		(10 ~ 40) Hz	0.008 dB	
		(0.04 ~ 100) kHz	0.007 dB	
		(100 ~ 200) kHz	0.010 dB	
Distortion		(0.02 ~ 100) kHz		
		(0 ~ -60) dB	0.19 dB	
		(-60 ~ -80) dB	0.50 dB	
Filter(Frequency) (Weight, Low, High Pass)		(0.01 ~ 100) kHz	6.1×10^{-6}	
Filter(Level) (Weight, Low, High Pass)		(0.01 ~ 100) kHz (20 ~ -63) dB	0.007 dB	
Output Frequency		(0.001 ~ 100) kHz (100 ~ 200) kHz	6.1×10^{-7} 1.0×10^{-7}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF/audio signal analyzers Output AC Voltage	40409	(0.5 ~ 1) mV 10 Hz (0.01 ~ 10) kHz (10 ~ 50) kHz (50 ~ 100) kHz (100 ~ 200) kHz (1 ~ 10) mV 10 Hz (0.01 ~ 10) kHz (10 ~ 50) kHz (50 ~ 100) kHz (100 ~ 200) kHz (10 ~ 100) mV 10 Hz (0.01 ~ 10) kHz (10 ~ 50) kHz (50 ~ 100) kHz (100 ~ 200) kHz (0.1 ~ 1) V 10 Hz (0.01 ~ 50) kHz (50 ~ 100) kHz (100 ~ 200) kHz (1 ~ 10) V 10 Hz (0.01 ~ 50) kHz (50 ~ 100) kHz (100 ~ 200) kHz (10 ~ 100) V 10 Hz (0.01 ~ 10) kHz (10 ~ 50) kHz (50 ~ 100) kHz Output Level 10 Hz (20 ~ -50) dBm (-50 ~ -60) dBm (0.01 ~ 10) kHz (20 ~ -50) dBm (-50 ~ -60) dBm (10 ~ 100) kHz (20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm (100 ~ 200) kHz (20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm	1.3×10^{-3} 1.2×10^{-3} 1.8×10^{-3} 2.1×10^{-3} 3.5×10^{-3} 1.8×10^{-4} 1.6×10^{-4} 2.2×10^{-4} 2.7×10^{-4} 7.5×10^{-4} 9.3×10^{-5} 7.4×10^{-5} 7.7×10^{-5} 1.0×10^{-4} 1.9×10^{-4} 8.3×10^{-5} 6.7×10^{-5} 7.7×10^{-5} 1.4×10^{-4} 8.2×10^{-5} 6.8×10^{-5} 8.2×10^{-5} 1.5×10^{-4} 8.8×10^{-5} 6.9×10^{-5} 7.3×10^{-5} 9.1×10^{-5} 0.008 dB 0.014 dB 0.008 dB 0.013 dB 0.007 dB 0.010 dB 0.023 dB 0.008 dB 0.014 dB 0.038 dB	Meter calibrator, DMM Frequency counter, Audio analyzer /KTICC-CI-40409

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF/audio signal analyzers Flatness (Voltage)	40409	100 mV 10 Hz (0.01 ~ 10) kHz (10 ~ 50) kHz (50 ~ 100) kHz (100 ~ 200) kHz (0.1 ~ 10) V 10 Hz (0.01 ~ 50) kHz (50 ~ 100) kHz (100 ~ 200) kHz	9.3×10^{-5} 7.4×10^{-5} 7.7×10^{-5} 1.0×10^{-4} 1.9×10^{-4} 8.2×10^{-5} 6.8×10^{-5} 8.2×10^{-5} 1.5×10^{-4}	Meter calibrator, DMM Frequency counter, Audio analyzer /KTICC-CI-40409
Flatness (Level)		(-10 ~ 10) dBm 10 Hz (0.01 ~ 200) kHz	0.007 dB 0.007 dB	
Line frequency meters Frequency	40410	25 Hz (25 ~ 60) Hz (60 ~ 100) Hz (100 ~ 500) Hz (500 ~ 1 000) Hz	4.4×10^{-4} 2.0×10^{-4} 2.4×10^{-4} 1.7×10^{-4} 2.4×10^{-4}	Power calibrator /KTICC-CI-40410
Function generators Frequency	40411	1 Hz ~ 100 MHz (100 ~ 400) MHz	6.1×10^{-9} 1.6×10^{-9}	DMM, Frequency counter, Measuring receiver, Oscilloscope, Audio Analyzer /KTICC-CI-40411
AC Voltage		(0.22 ~ 1) mV 10 Hz (0.01 ~ 10) kHz (10 ~ 100) kHz (0.1 ~ 1) MHz (1 ~ 10) mV 10 Hz (0.01 ~ 10) kHz (10 ~ 50) kHz (50 ~ 100) kHz (0.1 ~ 1) MHz (10 ~ 100) mV 10 Hz (0.01 ~ 10) kHz (10 ~ 50) kHz (50 ~ 100) kHz (0.1 ~ 1) MHz (0.1 ~ 1) V 10 Hz (0.01 ~ 50) kHz (50 ~ 100) kHz (0.1 ~ 1) MHz	1.3×10^{-3} 1.2×10^{-3} 2.1×10^{-3} 1.0×10^{-2} 1.8×10^{-4} 1.6×10^{-4} 2.2×10^{-4} 2.7×10^{-4} 2.7×10^{-3} 9.3×10^{-5} 7.4×10^{-5} 7.7×10^{-5} 1.0×10^{-4} 1.0×10^{-3} 8.3×10^{-5} 6.7×10^{-5} 7.7×10^{-5} 9.6×10^{-4}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Function generators	40411	(1 ~ 10) V		DMM, Frequency counter, Measuring receiver, Oscilloscope, Audio Analyzer /KTICC-CI-40411
AC Voltage		10 Hz (0.01 ~ 10) kHz (50 ~ 100) kHz (0.1 ~ 1) MHz	8.2×10^{-5} 6.8×10^{-5} 8.2×10^{-5} 1.2×10^{-3}	
		(10 ~ 100) V		
		10 Hz (0.01 ~ 10) kHz (10 ~ 50) kHz (50 ~ 100) kHz	8.8×10^{-5} 6.9×10^{-5} 7.3×10^{-5} 9.1×10^{-5}	
Level		10 Hz (20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm	0.008 dB 0.010 dB 0.49 dB	
		(10 ~ 40) Hz (20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm	0.007 dB 0.010 dB 0.30 dB	
		(0.04 ~ 10) kHz (20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm	0.007 dB 0.010 dB 0.21 dB	
		(0.01 ~ 10) MHz (20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm	0.008 dB 0.011 dB 0.14 dB	
		(10 ~ 50) MHz (20 ~ -50) dBm (-50 ~ -60) dBm	0.024 dB 0.14 dB	
		(50 ~ 400) MHz (20 ~ -30) dBm (-30 ~ -60) dBm	0.12 dB 0.14 dB	
Flatness (Voltage)		100 mV		
		10 Hz (0.01 ~ 100) kHz (0.1 ~ 1) MHz (1 ~ 10) MHz (10 ~ 30) MHz (30 ~ 50) MHz (50 ~ 400) MHz	8.7×10^{-4} 4.5×10^{-4} 5.1×10^{-4} 8.9×10^{-4} 3.0×10^{-3} 5.1×10^{-3} 2.6×10^{-2}	
		(0.1 ~ 1) V 10 Hz (0.01 ~ 100) kHz (0.1 ~ 1) MHz (1 ~ 10) MHz (10 ~ 30) MHz (30 ~ 50) MHz (50 ~ 400) MHz	8.9×10^{-4} 4.0×10^{-4} 5.1×10^{-4} 8.9×10^{-4} 3.0×10^{-3} 5.0×10^{-3} 2.7×10^{-2}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Function generators	40411	(1 ~ 3) V 10 Hz (0.01 ~ 100) kHz (0.1 ~ 1) MHz (1 ~ 10) MHz (10 ~ 30) MHz (30 ~ 50) MHz (50 ~ 400) MHz	9.0×10^{-4} 4.3×10^{-4} 5.3×10^{-4} 9.0×10^{-4} 3.0×10^{-3} 5.0×10^{-3} 2.8×10^{-2}	DMM, Frequency counter, Measuring receiver, Oscilloscope, Audio Analyzer /KTICC-CI-40411
Flatness (Voltage)				
Flatness (Level)		(10 ~ -10) dBm 10 Hz ~ 10 MHz (10 ~ 30) MHz (30 ~ 50) MHz (50 ~ 400) MHz	0.008 dB 0.015 dB 0.023 dB 0.11 dB	
DC Offset Voltage		($\pm$) 0 V (0 ~ 1) V (1 ~ 5) V (5 ~ 10) V	$6.2 \mu\text{V}$ 6.2×10^{-5} 1.3×10^{-5} 7.2×10^{-6}	
Attenuator		(0.04 ~ 100) kHz (30 ~ -40) dB (-40 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB (0.1 ~ 400) MHz (30 ~ -40) dB (-40 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	0.054 dB 0.10 dB 0.20 dB 0.30 dB 0.061 dB 0.068 dB 0.072 dB 0.076 dB 0.081 dB	
Frequency Modulation		(0 ~ 400) kHz	2.8×10^{-2}	
Amplitude Modulation		(0 ~ 99) %	2.7×10^{-2}	
Phase		(0 ~ 360)°	0.061°	
Duty cycle		(1 ~ 99) %	0.006 1 %	
Rise/Fall Time		0.4 ns (0.4 ~ 1) ns 1 ns ~ 1 ms	4.8×10^{-1} 9.0×10^{-2} 8.2×10^{-3}	
Distortion		(0.02 ~ 100) kHz (-30 ~ -80) dB	2.3 dB	
AC/DC high voltage voltmeters	40413	($\pm$) 0 kV (0 ~ 1) kV (1 ~ 100) kV	0.61 V 1.3×10^{-3} 1.2×10^{-3}	High voltage supply High volt meter, DMM High voltage test equipment /KTICC-CI-40413
AC Voltage		(50 ~ 60) Hz (0.1 ~ 10) kV	1.3×10^{-2}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF impulse generators Output Voltage Pulse Width Frequency	40414	($\pm$) (0.001 ~ 20) kV 1 ns (1 ~ 10) ns 10 ns ~ 100 ms 100 Hz (100 ~ 200) Hz (200 ~ 500) Hz (0.5 ~ 1) kHz (1 ~ 2) kHz (2 ~ 5) kHz (5 ~ 10) kHz (10 ~ 20) kHz	1.3×10^{-2} 9.0×10^{-2} 3.6×10^{-3} 3.4×10^{-3} 3.4×10^{-3} 3.9×10^{-3} 5.6×10^{-3} 3.4×10^{-3} 3.9×10^{-3} 5.6×10^{-3} 3.4×10^{-3} 3.9×10^{-3}	Oscilloscope, High voltage probe /KTICC-CI-40414
Leakage current testers DC Voltage AC Voltage DC Current AC Current	40416	($\pm$) 0 V (0 ~ 0.1) V (0.1 ~ 100) V (100 ~ 1 000) V (0.02 ~ 1) kHz (0.22 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 1 000) V ($\pm$) 0 A (0 ~ 10) μ A (10 ~ 100) μ A (10 ~ 500) μ A (0.5 ~ 1) mA (1 ~ 5) mA (5 ~ 10) mA (10 ~ 50) mA (50 ~ 100) mA (0.04 ~ 1) kHz (0.009 ~ 1) mA (1 ~ 5) mA (5 ~ 10) mA (10 ~ 50) mA (50 ~ 100) mA	6.1μ V 6.2×10^{-5} 6.1×10^{-5} 6.2×10^{-5} 1.7×10^{-4} 9.1×10^{-5} 8.9×10^{-5} 9.4×10^{-5} 1.0×10^{-4} 8.7 nA 8.9×10^{-4} 6.2×10^{-4} 1.3×10^{-4} 6.1×10^{-4} 1.3×10^{-4} 6.2×10^{-4} 1.4×10^{-4} 6.2×10^{-4} 6.4×10^{-4} 2.4×10^{-4} 7.6×10^{-4} 6.4×10^{-4} 7.6×10^{-4}	Meter calibrator, DMM /KTICC-CI-40416

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Leakage current testers	40416	Unweighted touch current measuring network		Meter calibrator, DMM /KTICC-CI-40416
Output voltage to input voltage ratios		ratio (1 ~ 4)		
		0.02 kHz	3.0×10^{-4}	
		0.05 kHz	1.4×10^{-4}	
		0.06 kHz	1.4×10^{-4}	
		0.1 kHz	1.4×10^{-4}	
		0.2 kHz	1.3×10^{-4}	
		0.5 kHz	1.1×10^{-4}	
		1 kHz	9.2×10^{-5}	
		2 kHz	8.1×10^{-5}	
		5 kHz	7.4×10^{-5}	
		10 kHz	7.4×10^{-5}	
		20 kHz	7.3×10^{-5}	
		50 kHz	1.0×10^{-4}	
		100 kHz	1.4×10^{-4}	
		200 kHz	3.5×10^{-4}	
		500 kHz	1.2×10^{-3}	
		1 000 kHz	2.1×10^{-3}	
		Perception or reaction measuring network		
		ratio (3.43 ~ 1 382)		
		0.02 kHz	3.0×10^{-4}	
		0.05 kHz	1.4×10^{-4}	
		0.06 kHz	1.4×10^{-4}	
		0.1 kHz	1.4×10^{-4}	
		0.2 kHz	1.3×10^{-4}	
		0.5 kHz	1.2×10^{-4}	
		1 kHz	1.1×10^{-4}	
		2 kHz	1.3×10^{-4}	
		5 kHz	7.9×10^{-5}	
		10 kHz	4.3×10^{-5}	
		20 kHz	8.6×10^{-5}	
		50 kHz	2.2×10^{-4}	
		100 kHz	4.2×10^{-4}	
		200 kHz	2.9×10^{-4}	
		500 kHz	7.2×10^{-4}	
		1 000 kHz	1.4×10^{-3}	
		Foil around lamp network		
		ratio (1 000 ~ 5 503)		
		0.06 kHz	3.2×10^{-4}	
		0.1 kHz	3.3×10^{-4}	
		0.2 kHz	3.9×10^{-4}	
		0.5 kHz	4.3×10^{-4}	
		1 kHz	5.3×10^{-4}	
		2 kHz	8.4×10^{-4}	
		5 kHz	1.4×10^{-3}	
		10 kHz	1.6×10^{-3}	
		20 kHz	1.7×10^{-3}	
		50 kHz	1.7×10^{-3}	
		100 kHz	1.8×10^{-3}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electronic AC/DC loads DC Voltage	40417	($\pm$) 0 V (0 ~ 1 000) V	61 μ V 6.2×10^{-5}	DC/AC power supply, Meter calibrator, Current shunt, DMM /KTICC-CI-40417
DC Current		($\pm$) 0 A (0 ~ 100) mA (0.1 ~ 1) A (1 ~ 10) A (10 ~ 100) A	61 nA 1.3×10^{-4} 2.4×10^{-4} 2.5×10^{-4} 5.9×10^{-4}	
AC Voltage		(40 ~ 100) Hz 0.22 mV ~ 10 V (10 ~ 100) V (100 ~ 1 000) V	9.5×10^{-5} 1.0×10^{-4} 2.0×10^{-4}	
AC Current		(40 ~ 100) Hz 0.009 mA ~ 30 A	1.3×10^{-3}	
Modulation meters Frequency Modulation	40418	(0 ~ 400) kHz	2.8×10^{-2}	RF signal generator, Measuring receiver, /KTICC-CI-40418
Amplitude Modulation		(0 ~ 99) %	2.7×10^{-2}	
Phase Modulation		(0 ~ 400) rad	4.2×10^{-2}	
Frequency		(0.1 ~ 1 000) MHz	6.1×10^{-9}	
Analogue/digital multimeters DC Voltage	40419	($\pm$) 0 V (0 ~ 10) mV (10 ~ 19) mV (19 ~ 100) mV (100 ~ 190) mV (0.19 ~ 1) V (1 ~ 1.9) V (1.9 ~ 10) V (10 ~ 19) V (19 ~ 100) V (100 ~ 190) V (190 ~ 1 000) V	0.41 μ V 4.9×10^{-5} 2.9×10^{-5} 1.2×10^{-5} 1.0×10^{-5} 5.9×10^{-6} 5.8×10^{-6} 4.0×10^{-6} 3.9×10^{-6} 5.9×10^{-6} 5.8×10^{-6} 7.3×10^{-6}	Meter calibrator, Frequency generator, Standard resistance, DC reference standard /KTICC-CI-40419
AC Voltage		(0.22 ~ 100) mV (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 200) kHz (200 ~ 500) kHz (0.5 ~ 1) MHz	3.7×10^{-4} 1.7×10^{-4} 1.6×10^{-4} 2.8×10^{-4} 6.5×10^{-4} 1.1×10^{-3} 1.7×10^{-3} 3.2×10^{-3}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/digital multimeters AC Voltage	40419	(100 ~ 190) mV (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 200) kHz (200 ~ 500) kHz (0.5 ~ 1) MHz (0.19 ~ 1) V (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 200) kHz (200 ~ 500) kHz (0.5 ~ 1) MHz (1 ~ 1.9) V (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 200) kHz (200 ~ 500) kHz (0.5 ~ 1) MHz (1.9 ~ 10) V (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 200) kHz (200 ~ 500) kHz (0.5 ~ 1) MHz (10 ~ 19) V (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz (100 ~ 200) kHz (200 ~ 500) kHz (0.5 ~ 1) MHz (19 ~ 100) V (20 ~ 40) Hz (0.04 ~ 20) kHz (20 ~ 50) kHz (50 ~ 100) kHz	3.2×10^{-4} 1.4×10^{-4} 1.3×10^{-4} 2.5×10^{-4} 5.8×10^{-4} 1.1×10^{-3} 1.6×10^{-3} 3.0×10^{-3} 2.9×10^{-4} 1.1×10^{-4} 6.8×10^{-5} 9.6×10^{-5} 1.5×10^{-4} 5.2×10^{-4} 1.2×10^{-3} 2.0×10^{-3} 2.7×10^{-4} 1.1×10^{-4} 6.3×10^{-5} 8.9×10^{-5} 1.4×10^{-4} 4.8×10^{-4} 1.2×10^{-3} 1.9×10^{-3} 2.9×10^{-4} 1.1×10^{-4} 6.5×10^{-5} 1.0×10^{-4} 1.3×10^{-4} 3.5×10^{-4} 1.2×10^{-3} 1.9×10^{-3} 2.7×10^{-4} 1.1×10^{-4} 6.3×10^{-5} 8.9×10^{-5} 1.2×10^{-4} 3.3×10^{-4} 1.2×10^{-3} 1.7×10^{-3} 1.1×10^{-4} 7.2×10^{-5} 1.0×10^{-4} 1.9×10^{-4}	Meter calibrator, Frequency generator, Standard resistance, DC reference standard /KTICC-CI-40419

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/digital multimeters	40419	(100 ~ 190) V		Meter calibrator, Frequency generator, Standard resistance, DC reference standard /KTICC-CI-40419
AC Voltage		(20 ~ 40) Hz	1.1×10^{-4}	
		(0.04 ~ 20) kHz	6.8×10^{-5}	
		(20 ~ 50) kHz	1.0×10^{-4}	
		(50 ~ 100) kHz	1.8×10^{-4}	
		(190 ~ 1 000) V		
		(15 ~ 50) Hz	3.2×10^{-4}	
		(50 ~ 500) Hz	8.5×10^{-5}	
		(0.5 ~ 1) kHz	8.5×10^{-5}	
		(1 ~ 20) kHz	1.8×10^{-4}	
DC Current		(±)		
		0 A	6.1 nA	
		(0 ~ 100) μA	1.0×10^{-4}	
		(100 ~ 190) μA	7.4×10^{-5}	
		(0.19 ~ 1) mA	4.3×10^{-5}	
		(1 ~ 10) mA	4.0×10^{-5}	
		(10 ~ 19) mA	3.8×10^{-5}	
		(19 ~ 100) mA	5.4×10^{-5}	
		(100 ~ 190) mA	5.0×10^{-5}	
		(0.19 ~ 1) A	9.4×10^{-5}	
		(1 ~ 1.9) A	8.9×10^{-5}	
		(1.9 ~ 2) A	9.0×10^{-5}	
		(2 ~ 10) A	5.1×10^{-4}	
		(10 ~ 20) A	1.3×10^{-4}	
		(20 ~ 30) A	3.7×10^{-4}	
AC Current		(0.009 ~ 1) mA		
		(10 ~ 20) Hz	3.1×10^{-4}	
		(20 ~ 40) Hz	2.1×10^{-4}	
		(0.04 ~ 1) kHz	1.7×10^{-4}	
		(1 ~ 10) kHz	1.9×10^{-3}	
		(1 ~ 1.9) mA		
		(10 ~ 20) Hz	5.5×10^{-4}	
		(20 ~ 40) Hz	1.9×10^{-4}	
		(0.04 ~ 1) kHz	1.6×10^{-4}	
		(1 ~ 10) kHz	1.5×10^{-3}	
		(1.9 ~ 10) mA		
		(10 ~ 20) Hz	3.1×10^{-4}	
		(20 ~ 40) Hz	2.1×10^{-4}	
		(0.04 ~ 1) kHz	1.7×10^{-4}	
		(1 ~ 10) kHz	1.6×10^{-3}	
		(10 ~ 19) mA		
		(10 ~ 20) Hz	2.9×10^{-4}	
		(20 ~ 40) Hz	1.9×10^{-4}	
		(0.04 ~ 1) kHz	1.6×10^{-4}	
		(1 ~ 10) kHz	1.4×10^{-3}	
		(19 ~ 100) mA		
		(10 ~ 20) Hz	3.1×10^{-4}	
		(20 ~ 40) Hz	2.2×10^{-4}	
		(0.04 ~ 1) kHz	1.7×10^{-4}	
		(1 ~ 10) kHz	1.2×10^{-3}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Analogue/digital multimeters AC Current	40419	(100 ~ 190) mA (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 1) kHz (1 ~ 10) kHz (0.19 ~ 1) A (0.01 ~ 1) kHz (1 ~ 10) kHz (1 ~ 1.9) A (0.01 ~ 1) kHz (1 ~ 10) kHz (1.9 ~ 2) A (0.01 ~ 1) kHz (1 ~ 10) kHz (2 ~ 10) A (45 ~ 100) Hz (0.1 ~ 1) kHz (1 ~ 10) kHz (10 ~ 20) A (45 ~ 100) Hz (0.1 ~ 1) kHz (20 ~ 30) A (45 ~ 100) Hz (0.1 ~ 1) kHz	2.9×10^{-4} 2.1×10^{-4} 1.6×10^{-4} 1.2×10^{-3} 3.3×10^{-4} 7.3×10^{-3} 3.1×10^{-4} 7.4×10^{-3} 3.2×10^{-4} 7.0×10^{-3} 8.0×10^{-4} 8.0×10^{-4} 1.7×10^{-2} 8.5×10^{-4} 1.4×10^{-3} 1.3×10^{-3} 4.0×10^{-3}	Meter calibrator, Frequency generator, Standard resistance, DC reference standard /KTICC-CI-40419
Resistance		(0 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) k Ω (1 ~ 10) k Ω (10 ~ 100) k Ω (0.1 ~ 1) M Ω (1 ~ 10) M Ω (10 ~ 100) M Ω (0.1 ~ 1) G Ω	9.6×10^{-6} 2.4×10^{-5} 7.4×10^{-6} 1.0×10^{-5} 5.4×10^{-6} 1.5×10^{-5} 2.3×10^{-5} 4.6×10^{-5} 1.0×10^{-4} 6.2×10^{-4}	
Frequency		1 Hz ~ 10 MHz	6.1×10^{-7}	
Noise meters AC Voltage	40420	(0.22 ~ 0.3) mV (0.01 ~ 20) kHz (20 ~ 100) kHz (100 ~ 500) kHz (0.3 ~ 1) mV (10 ~ 20) Hz (0.02 ~ 20) kHz (20 ~ 100) kHz (100 ~ 500) kHz (1 ~ 3) mV (0.01 ~ 20) kHz (20 ~ 100) kHz (100 ~ 500) kHz	1.5×10^{-2} 1.8×10^{-2} 7.0×10^{-2} 5.2×10^{-3} 5.1×10^{-3} 6.3×10^{-3} 2.2×10^{-2} 5.0×10^{-3} 5.3×10^{-3} 9.7×10^{-3}	Meter calibrator /KTICC-CI-40420

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Noise meters	40420	(3 ~ 10) mV		Meter calibrator /KTICC-CI-40420
AC Voltage		(10 ~ 20) Hz	3.0×10^{-3}	
		(0.02 ~ 20) kHz	2.9×10^{-3}	
		(20 ~ 100) kHz	3.1×10^{-3}	
		(100 ~ 500) kHz	4.5×10^{-3}	
		(10 ~ 30) mV		
		(10 ~ 20) Hz	5.0×10^{-3}	
		(0.02 ~ 20) kHz	4.7×10^{-3}	
		(20 ~ 100) kHz	5.0×10^{-3}	
		(100 ~ 500) kHz	5.3×10^{-3}	
		(30 ~ 100) mV		
		(0.01 ~ 20) kHz	2.9×10^{-3}	
		(20 ~ 100) kHz	3.0×10^{-3}	
		(100 ~ 500) kHz	3.4×10^{-3}	
		(100 ~ 300) mV		
		(0.01 ~ 100) kHz	4.7×10^{-3}	
		(100 ~ 500) kHz	5.0×10^{-3}	
		(0.3 ~ 1) V		
		(0.01 ~ 100) kHz	2.9×10^{-3}	
		(100 ~ 500) kHz	3.1×10^{-3}	
		(1 ~ 3) V		
		(0.01 ~ 100) kHz	4.7×10^{-3}	
		(100 ~ 500) kHz	5.0×10^{-3}	
		(3 ~ 10) V		
		(0.01 ~ 100) kHz	2.9×10^{-3}	
		(100 ~ 500) kHz	3.1×10^{-3}	
		(10 ~ 30) V		
		(0.01 ~ 100) kHz	4.7×10^{-3}	
		(30 ~ 100) V		
		(0.01 ~ 100) kHz	2.9×10^{-3}	
		(100 ~ 300) V		
		(0.015 ~ 1) kHz	4.7×10^{-3}	
Level		(0.01 ~ 1) kHz		
		(40 ~ -40) dBm	0.036 dB	
		(-40 ~ -50) dBm	0.039 dB	
		(-50 ~ -60) dBm	0.059 dB	
		(-60 ~ -70) dBm	0.15 dB	
		(1 ~ 20) kHz		
		(40 ~ -40) dBm	0.035 dB	
		(-40 ~ -50) dBm	0.038 dB	
		(-50 ~ -60) dBm	0.058 dB	
		(-60 ~ -70) dBm	0.15 dB	
		(20 ~ 100) kHz		
		(40 ~ -10) dBm	0.035 dB	
		(-10 ~ -20) dBm	0.036 dB	
		(-20 ~ -40) dBm	0.037 dB	
		(-40 ~ -50) dBm	0.042 dB	
		(-50 ~ -60) dBm	0.071 dB	
		(-60 ~ -70) dBm	0.19 dB	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Noise meters Level	40420	(100 ~ 500) kHz (20 ~ -20) dBm (-20 ~ -30) dBm (-30 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm (-60 ~ -70) dBm	0.039 dB 0.041 dB 0.050 dB 0.091 dB 0.24 dB 0.71 dB	Meter calibrator /KTICC-CI-40420
Filter (Weight, Low, High Pass etc.)		(0.01 ~ 31.5) kHz (20 ~ -63) dB	0.035 dB	
Oscilloscopes Vertical(DC Voltage)	40421	(±) 0 V (0 ~ 1) mV (1 ~ 2) mV (2 ~ 5) mV (5 ~ 10) mV (10 ~ 20) mV (20 ~ 50) mV (50 ~ 100) mV (100 ~ 200) mV (200 ~ 500) mV (0.5 ~ 1) V (1 ~ 2) V (2 ~ 10) V (10 ~ 20) V (20 ~ 50) V (50 ~ 100) V (100 ~ 130) V	29 μV 2.9×10^{-2} 1.5×10^{-2} 6.0×10^{-3} 3.2×10^{-3} 1.8×10^{-3} 8.6×10^{-4} 5.8×10^{-4} 4.4×10^{-4} 3.4×10^{-4} 3.2×10^{-4} 3.1×10^{-4} 3.0×10^{-4} 2.9×10^{-4} 2.8×10^{-4} 3.0×10^{-4} 2.9×10^{-4}	Scope calibrator, DMM, Frequency counter, RF signal generator /KTICC-CI-40421
Vertical(Square Wave)		1 kHz 1 mV (1 ~ 2) mV (2 ~ 5) mV (5 ~ 10) mV (10 ~ 20) mV (20 ~ 50) mV (50 ~ 100) mV (100 ~ 200) mV (200 ~ 500) mV (0.5 ~ 1) V (1 ~ 2) V (2 ~ 10) V (10 ~ 20) V (20 ~ 130) V	6.4×10^{-3} 3.5×10^{-3} 1.7×10^{-3} 1.2×10^{-3} 8.5×10^{-4} 7.0×10^{-4} 6.4×10^{-4} 6.0×10^{-4} 5.8×10^{-4} 5.9×10^{-4} 6.0×10^{-4} 5.8×10^{-4} 6.0×10^{-4} 5.8×10^{-4}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Oscilloscopes	40421			Scope calibrator, DMM, Frequency counter, RF signal generator /KTICC-CI-40421
Time(Period)		1 ns	6.1×10^{-5}	
		(1 ~ 2) ns	3.1×10^{-5}	
		(2 ~ 5) ns	1.2×10^{-5}	
		(5 ~ 10) ns	6.1×10^{-5}	
		(10 ~ 20) ns	3.1×10^{-5}	
		(20 ~ 50) ns	1.2×10^{-5}	
		(50 ~ 100) ns	6.1×10^{-5}	
		(100 ~ 200) ns	3.1×10^{-5}	
		(200 ~ 500) ns	1.2×10^{-5}	
		(0.5 ~ 1) μ s	6.1×10^{-5}	
		(1 ~ 2) μ s	3.1×10^{-5}	
		(2 ~ 5) μ s	1.2×10^{-5}	
		(5 ~ 10) μ s	6.1×10^{-5}	
		(10 ~ 20) μ s	3.1×10^{-5}	
		(20 ~ 50) μ s	1.2×10^{-5}	
		(50 ~ 100) μ s	6.1×10^{-5}	
		(100 ~ 200) μ s	3.1×10^{-5}	
		(200 ~ 500) μ s	1.2×10^{-5}	
		(0.5 ~ 1) ms	6.1×10^{-5}	
		(1 ~ 2) ms	3.1×10^{-5}	
		(2 ~ 5) ms	1.2×10^{-5}	
		(5 ~ 10) ms	6.1×10^{-5}	
		(10 ~ 20) ms	3.1×10^{-5}	
		(20 ~ 50) ms	1.4×10^{-5}	
		(50 ~ 100) ms	6.1×10^{-5}	
		(100 ~ 200) ms	3.1×10^{-5}	
		(200 ~ 500) ms	1.3×10^{-5}	
		(0.5 ~ 1) s	6.1×10^{-5}	
		(1 ~ 2) s	3.1×10^{-5}	
		(2 ~ 5) s	1.3×10^{-5}	
Bandwidth		30 mV (≤ 2 GHz)		
		50 kHz	2.5×10^{-2}	
		(0.05 ~ 100) MHz	2.9×10^{-2}	
		(100 ~ 300) MHz	3.3×10^{-2}	
		(300 ~ 500) MHz	5.0×10^{-2}	
		(500 ~ 600) MHz	5.3×10^{-2}	
		(0.6 ~ 1) GHz	6.3×10^{-2}	
		(1 ~ 2) GHz	7.7×10^{-2}	
		30 mV (≥ 2 GHz)		
		(0.01 ~ 5) GHz	5.0×10^{-2}	
		(5 ~ 10) GHz	5.3×10^{-2}	
		(10 ~ 18) GHz	5.7×10^{-2}	
		600 mV (≤ 2 GHz)		
		50 kHz	1.5×10^{-2}	
		(0.05 ~ 100) MHz	2.0×10^{-2}	
		(100 ~ 300) MHz	2.5×10^{-2}	
		(300 ~ 500) MHz	4.2×10^{-2}	
		(500 ~ 600) MHz	4.7×10^{-2}	
		(0.6 ~ 1) GHz	5.8×10^{-2}	
		(1 ~ 2) GHz	7.0×10^{-2}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Oscilloscopes	40421	600 mV (≥ 2 GHz)		Scope calibrator, DMM, Frequency counter, RF signal generator /KTICC-CI-40421
Bandwidth		(0.01 ~ 2) GHz	4.7×10^{-2}	
		(2 ~ 10) GHz	5.0×10^{-2}	
		(10 ~ 18) GHz	5.5×10^{-2}	
		3 V (≤ 2 GHz)		
		50 kHz	1.5×10^{-2}	
		(0.05 ~ 100) MHz	2.0×10^{-2}	
		(100 ~ 300) MHz	2.5×10^{-2}	
		(300 ~ 500) MHz	4.0×10^{-2}	
		(500 ~ 600) MHz	4.7×10^{-2}	
		(0.6 ~ 1) GHz	6.0×10^{-2}	
		(1 ~ 2) GHz	7.0×10^{-2}	
		3 V (≥ 2 GHz)		
		(0.01 ~ 2) GHz	4.7×10^{-2}	
		(2 ~ 10) GHz	5.0×10^{-2}	
		(10 ~ 18) GHz	5.3×10^{-2}	
CAL Output (DC Voltage)		0 V	0.12 μ V	
		(0 ~ 1) mV	1.2×10^{-4}	
		(1 ~ 10) mV	6.3×10^{-5}	
		(10 ~ 100) mV	6.2×10^{-5}	
		(0.1 ~ 1) V	6.1×10^{-5}	
		(1 ~ 2) V	3.2×10^{-5}	
CAL Output (AC Voltage)		0.22 mV ~ 0.1 V		
		100 Hz	2.2×10^{-4}	
		(0.1 ~ 1) kHz	2.1×10^{-4}	
		(1 ~ 10) kHz	2.2×10^{-4}	
		(10 ~ 100) kHz	1.1×10^{-3}	
		(0.1 ~ 0.5) V		
		100 Hz	1.9×10^{-4}	
		(0.1 ~ 1) kHz	1.7×10^{-4}	
		(1 ~ 10) kHz	1.9×10^{-4}	
		(10 ~ 100) kHz	9.4×10^{-4}	
		(0.5 ~ 1) V		
		100 Hz	1.4×10^{-4}	
		(0.1 ~ 1) kHz	1.2×10^{-4}	
		(1 ~ 10) kHz	1.4×10^{-4}	
		(10 ~ 100) kHz	7.3×10^{-4}	
		(1 ~ 5) V		
		100 Hz	1.4×10^{-4}	
		(0.1 ~ 1) kHz	1.2×10^{-4}	
		(1 ~ 10) kHz	1.4×10^{-4}	
		(10 ~ 100) kHz	9.3×10^{-4}	
CAL Output (Frequency)		(0.1 ~ 100) kHz	6.1×10^{-6}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF phase meters	40422	(0.01 ~ 10) kHz (0 ~ 360)° (10 ~ 100) kHz (0 ~ 360)° (0.1 ~ 1) MHz (0 ~ 360)° (1 ~ 10) MHz (0 ~ 360)°	0.061° 0.064° 0.19° 1.8°	Function generator, Frequency counter /KTICC-CI-40422
Random wave generators Frequency AC Voltage	40423	1 Hz ~ 100 MHz (100 ~ 400) MHz (0.22 ~ 1) mV 10 Hz (0.01 ~ 10) kHz (10 ~ 100) kHz (0.1 ~ 1) MHz (1 ~ 10) mV 10 Hz (0.01 ~ 10) kHz (10 ~ 50) kHz (50 ~ 100) kHz (0.1 ~ 1) MHz (10 ~ 100) mV 10 Hz (0.01 ~ 10) kHz (10 ~ 50) kHz (50 ~ 100) kHz (0.1 ~ 1) MHz (0.1 ~ 1) V 10 Hz (0.01 ~ 50) kHz (50 ~ 100) kHz (0.1 ~ 1) MHz (1 ~ 10) V 10 Hz (0.01 ~ 50) kHz (50 ~ 100) kHz (0.1 ~ 1) MHz (10 ~ 100) V 10 Hz (0.01 ~ 10) kHz (10 ~ 50) kHz (50 ~ 100) kHz	6.1×10^{-9} 2.6×10^{-9} 1.3×10^{-3} 1.2×10^{-3} 2.1×10^{-3} 1.0×10^{-2} 1.8×10^{-4} 1.6×10^{-4} 2.2×10^{-4} 2.7×10^{-4} 2.7×10^{-3} 9.3×10^{-5} 7.4×10^{-5} 7.7×10^{-5} 1.0×10^{-4} 1.0×10^{-3} 8.3×10^{-5} 6.7×10^{-5} 7.7×10^{-5} 9.6×10^{-4} 8.2×10^{-5} 6.8×10^{-5} 8.2×10^{-5} 1.2×10^{-3} 8.8×10^{-5} 6.9×10^{-5} 7.3×10^{-5} 9.1×10^{-5}	DMM, Frequency counter, Measuring receiver, Oscilloscope /KTICC-CI-40423

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Random wave generators Level	40423	10 Hz (20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm (10 ~ 40) Hz (20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm (0.04 ~ 10) kHz (20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm (0.01 ~ 10) MHz (20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm (10 ~ 50) MHz (20 ~ -50) dBm (-50 ~ -60) dBm (50 ~ 400) MHz (20 ~ -30) dBm (-30 ~ -60) dBm	0.008 dB 0.010 dB 0.49 dB 0.007 dB 0.010 dB 0.30 dB 0.007 dB 0.010 dB 0.21 dB 0.008 dB 0.011 dB 0.14 dB 0.024 dB 0.14 dB 0.12 dB 0.14 dB	DMM, Frequency counter, Measuring receiver, Oscilloscope /KTICC-CI-40423
Flatness (Voltage)		100 mV 10 Hz (0.01 ~ 100) kHz (0.1 ~ 1) MHz (1 ~ 10) MHz (10 ~ 30) MHz (30 ~ 50) MHz (50 ~ 400) MHz (0.1 ~ 1) V 10 Hz (0.01 ~ 100) kHz (0.1 ~ 1) MHz (1 ~ 10) MHz (10 ~ 30) MHz (30 ~ 50) MHz (50 ~ 400) MHz (1 ~ 3) V 10 Hz (0.01 ~ 100) kHz (0.1 ~ 1) MHz (1 ~ 10) MHz (10 ~ 30) MHz (30 ~ 50) MHz (50 ~ 400) MHz	8.7×10^{-4} 4.5×10^{-4} 5.1×10^{-4} 8.9×10^{-4} 3.0×10^{-3} 5.1×10^{-3} 2.6×10^{-2} 8.9×10^{-4} 4.0×10^{-4} 5.1×10^{-4} 8.9×10^{-4} 3.0×10^{-3} 5.0×10^{-3} 2.7×10^{-2} 9.0×10^{-4} 4.3×10^{-4} 5.3×10^{-4} 9.0×10^{-4} 3.0×10^{-3} 5.0×10^{-3} 2.8×10^{-2}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Random wave generators	40423	(10 ~ -10) dBm		DMM, Frequency counter, Measuring receiver, Oscilloscope /KTICC-CI-40423
Flatness (Level)		10 Hz ~ 10 MHz	0.008 dB	
		(10 ~ 30) MHz	0.015 dB	
		(30 ~ 50) MHz	0.023 dB	
		(50 ~ 400) MHz	0.11 dB	
DC Offset Voltage		(±)		
		0 V	6.1 μV	
		(0 ~ 1) V	6.2×10^{-5}	
		(1 ~ 5) V	1.3×10^{-5}	
		(5 ~ 10) V	7.2×10^{-6}	
Attenuator		(0.04 ~ 100) kHz		
		(30 ~ -40) dB	0.054 dB	
		(-40 ~ -60) dB	0.10 dB	
		(-60 ~ -70) dB	0.20 dB	
		(-70 ~ -80) dB	0.30 dB	
		(0.1 ~ 400) MHz		
		(30 ~ -40) dB	0.061 dB	
		(-40 ~ -50) dB	0.068 dB	
		(-50 ~ -60) dB	0.072 dB	
		(-60 ~ -70) dB	0.076 dB	
		(-70 ~ -80) dB	0.081 dB	
Frequency Modulation		(0 ~ 400) kHz	2.8×10^{-2}	
Amplitude Modulation		(0 ~ 99) %	2.7×10^{-2}	
Phase		(0 ~ 360)°	0.061°	
Duty cycle		(1 ~ 99) %	0.006 1 %	
Rise/Fall Time		0.4 ns	4.8×10^{-1}	
		(0.4 ~ 1) ns	9.0×10^{-2}	
		1 ns ~ 1 ms	8.2×10^{-3}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Voltage/current recorders	40424	($\pm$)		Meter calibrator
DC Voltage		0 V	0.41 μ V	/KTICC-CI-40424
		(0 ~ 10) mV	4.9×10^{-5}	
		(10 ~ 100) mV	1.2×10^{-5}	
		(0.1 ~ 1) V	5.9×10^{-6}	
		(1 ~ 10) V	4.0×10^{-6}	
		(10 ~ 100) V	5.9×10^{-6}	
		(100 ~ 1 000) V	7.3×10^{-6}	
AC Voltage		(0.22 ~ 100) mV		
		(20 ~ 40) Hz	1.7×10^{-4}	
		(0.04 ~ 20) kHz	1.6×10^{-4}	
		(20 ~ 50) kHz	2.8×10^{-4}	
		(50 ~ 100) kHz	6.5×10^{-4}	
		(100 ~ 300) kHz	1.1×10^{-3}	
		(0.1 ~ 1) V		
		(20 ~ 40) Hz	1.1×10^{-4}	
		(0.04 ~ 20) kHz	6.8×10^{-5}	
		(20 ~ 50) kHz	9.6×10^{-5}	
		(50 ~ 100) kHz	1.5×10^{-4}	
		(100 ~ 300) kHz	5.2×10^{-4}	
		(1 ~ 10) V		
		(20 ~ 40) Hz	1.1×10^{-4}	
		(0.04 ~ 20) kHz	6.5×10^{-5}	
		(20 ~ 50) kHz	1.0×10^{-4}	
		(50 ~ 100) kHz	1.3×10^{-4}	
		(100 ~ 200) kHz	3.5×10^{-4}	
		(200 ~ 300) kHz	3.6×10^{-4}	
		(10 ~ 100) V		
		(20 ~ 40) Hz	1.1×10^{-4}	
		(0.04 ~ 20) kHz	7.2×10^{-5}	
		(20 ~ 50) kHz	1.0×10^{-4}	
		(50 ~ 100) kHz	1.9×10^{-4}	
		(100 ~ 1 000) V		
		(15 ~ 50) Hz	3.2×10^{-4}	
		(0.05 ~ 1) kHz	8.5×10^{-5}	
DC Current		($\pm$)		
		0 A	6.1 nA	
		(0 ~ 100) μ A	1.0×10^{-4}	
		(0.1 ~ 1) mA	4.3×10^{-5}	
		(1 ~ 10) mA	4.0×10^{-5}	
		(10 ~ 100) mA	5.4×10^{-5}	
		(0.1 ~ 1) A	9.4×10^{-5}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Voltage/current recorders AC Current	40424	(0.009 ~ 1) mA (20 ~ 40) Hz (0.04 ~ 1) kHz (1 ~ 10) kHz (1 ~ 10) mA (20 ~ 40) Hz (0.04 ~ 1) kHz (1 ~ 10) kHz (10 ~ 100) mA (20 ~ 40) Hz (0.04 ~ 1) kHz (1 ~ 10) kHz (0.1 ~ 1) A (0.02 ~ 1) kHz (1 ~ 10) kHz	2.1×10^{-4} 1.7×10^{-4} 1.9×10^{-3} 2.1×10^{-4} 1.7×10^{-4} 1.6×10^{-3} 2.2×10^{-4} 1.7×10^{-4} 1.2×10^{-3} 3.3×10^{-4} 7.3×10^{-3}	Meter calibrator /KTICC-CI-40424
Resistance		(0 ~ 1) Ω (1 ~ 10) Ω (10 ~ 100) Ω (0.1 ~ 1) k Ω (1 ~ 10) k Ω (10 ~ 100) k Ω (0.1 ~ 1) M Ω (1 ~ 10) M Ω (10 ~ 100) M Ω	9.6×10^{-6} 2.4×10^{-5} 7.4×10^{-6} 1.0×10^{-5} 5.4×10^{-6} 1.5×10^{-5} 2.3×10^{-5} 4.6×10^{-5} 1.0×10^{-4}	
Frequency		(0.001 ~ 100) kHz	6.1×10^{-7}	
Relay test sets DC Voltage	40425	($\pm$) 0 V (0 ~ 100) mV (100 ~ 190) mV (0.19 ~ 1) V (1 ~ 1.9) V (1.9 ~ 10) V (10 ~ 19) V (19 ~ 100) V (100 ~ 190) V (190 ~ 1 000) V	0.62 μ V 1.1×10^{-5} 9.5×10^{-6} 7.2×10^{-6} 4.8×10^{-6} 7.2×10^{-6} 4.8×10^{-6} 8.6×10^{-6} 6.8×10^{-6} 8.6×10^{-6}	DMM, Current shunt, Function generator, Oscilloscope /KTICC-CI-40425

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Relay test sets	40425	(±)		DMM, Current shunt, Function generator, Oscilloscope /KTICC-CI-40425
DC Current		0 A (0 ~ 1) mA (1 ~ 1.9) mA (1.9 ~ 10) mA (10 ~ 19) mA (19 ~ 100) mA (100 ~ 190) mA (0.19 ~ 1) A (1 ~ 1.9) A (1.9 ~ 10) A (10 ~ 30) A (30 ~ 50) A (50 ~ 100) A	6.1 nA 8.9×10^{-6} 7.4×10^{-6} 1.0×10^{-5} 8.9×10^{-6} 1.4×10^{-5} 1.3×10^{-5} 5.9×10^{-5} 5.8×10^{-5} 8.9×10^{-5} 1.2×10^{-4} 2.4×10^{-4} 5.8×10^{-4}	
AC Voltage		(0.04 ~ 1) kHz (0.22 ~ 190) mV (0.19 ~ 1) V (1 ~ 1.9) V (1.9 ~ 19) V (19 ~ 190) V (190 ~ 1 000) V	4.0×10^{-5} 2.3×10^{-5} 2.2×10^{-5} 2.2×10^{-5} 2.6×10^{-5} 2.9×10^{-5}	
AC Current		(0.04 ~ 1) kHz (0.009 ~ 1) mA (1 ~ 1.9) mA (1.9 ~ 10) mA (10 ~ 19) mA (19 ~ 100) mA (100 ~ 190) mA (0.19 ~ 1) A (1 ~ 1.9) A (1.9 ~ 10) A (10 ~ 100) A	4.4×10^{-4} 3.6×10^{-4} 4.4×10^{-4} 3.6×10^{-4} 4.4×10^{-4} 3.6×10^{-4} 8.7×10^{-4} 8.9×10^{-4} 1.1×10^{-3} 1.2×10^{-3}	
Output Time Interval		(0 ~ 100) s	6.1×10^{-4}	
Input Time Interval		(0 ~ 100) s	6.1×10^{-4}	
LF signal generators	40426			DMM, Frequency counter, Audio analyzer /KTICC-CI-40426
Frequency		1 Hz ~ 10 MHz	6.1×10^{-7}	
AC Voltage		(0.22 ~ 1) mV (0.04 ~ 10) kHz (10 ~ 100) kHz (0.1 ~ 1) MHz (1 ~ 10) mV 40 Hz (0.04 ~ 10) kHz (10 ~ 50) kHz (0.05 ~ 1) MHz	1.2×10^{-3} 2.1×10^{-3} 9.7×10^{-3} 1.6×10^{-4} 1.5×10^{-4} 2.2×10^{-4} 2.7×10^{-3}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
LF signal generators	40426	(10 ~ 100) mV		DMM, Frequency counter, Audio analyzer /KTICC-CI-40426	
AC Voltage		40 Hz	7.4×10^{-5}		
		(0.04 ~ 10) kHz	7.2×10^{-5}		
		(10 ~ 50) kHz	7.7×10^{-5}		
		(50 ~ 100) kHz	1.0×10^{-4}		
		(0.1 ~ 1) MHz	1.0×10^{-3}		
		(0.1 ~ 1) V			
		40 Hz	6.7×10^{-5}		
		(0.04 ~ 50) kHz	6.7×10^{-5}		
		(50 ~ 100) kHz	7.7×10^{-5}		
		(0.1 ~ 1) MHz	9.6×10^{-4}		
		(1 ~ 10) V			
		40 Hz	6.8×10^{-5}		
		(0.04 ~ 50) kHz	6.7×10^{-5}		
		(50 ~ 100) kHz	8.2×10^{-5}		
		(0.1 ~ 1) MHz	1.2×10^{-3}		
		(10 ~ 100) V			
		40 Hz	6.9×10^{-5}		
		(0.04 ~ 10) kHz	6.7×10^{-5}		
		(10 ~ 50) kHz	7.3×10^{-5}		
		(50 ~ 100) kHz	9.1×10^{-5}		
		Level	40 Hz		
			(20 ~ -20) dBm		0.10 dB
			(-20 ~ -40) dBm		0.20 dB
			(-40 ~ -60) dBm		0.30 dB
			(0.04 ~ 10) kHz		
			(20 ~ -20) dBm		0.051 dB
			(-20 ~ -40) dBm		0.10 dB
			(-40 ~ -60) dBm		0.20 dB
			(10 ~ 100) kHz		
			(20 ~ 0) dBm		0.051 dB
			(0 ~ -20) dBm		0.054 dB
			(-20 ~ -40) dBm		0.10 dB
			(-40 ~ -60) dBm		0.21 dB
			(0.1 ~ 1) MHz		
			(20 ~ -20) dBm		0.076 dB
			(-20 ~ -40) dBm		0.20 dB
(-40 ~ -50) dBm			0.31 dB		
(-50 ~ -60) dBm			0.38 dB		
Flatness (Voltage)		(0 ~ 100) mV			
		(0.04 ~ 100) kHz	7.7×10^{-5}		
		(0.1 ~ 1) MHz	1.0×10^{-3}		
		(0.1 ~ 10) V			
		(0.04 ~ 100) kHz	8.2×10^{-5}		
		(0.1 ~ 1) MHz	1.2×10^{-3}		

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF signal generators	40426			DMM, Frequency counter, Audio analyzer /KTICC-CI-40426
Flatness (Level)		0 dB 40 Hz (0.04 ~ 100) kHz (0.1 ~ 1) MHz	0.10 dB 0.051 dB 0.073 dB	
Attenuator		(0.04 ~ 100) kHz (30 ~ -40) dB (-40 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB (0.1 ~ 1) MHz (30 ~ -40) dB (-40 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	0.054 dB 0.10 dB 0.20 dB 0.30 dB 0.076 dB 0.20 dB 0.31 dB 0.46 dB	
DC Offset		(±) 0 V (0 ~ 1) V (1 ~ 5) V (5 ~ 10) V	6.1 μV 6.2×10^{-5} 1.3×10^{-5} 7.2×10^{-6}	
Distortion		(0.02 ~ 100) kHz (-30 ~ -80) dB	2.3 dB	
LF spectrum analyzers	40427			Function generator RF signal generator, Measuring receiver, Frequency counter /KTICC-CI-40427
Frequency		10 Hz ~ 100 MHz	6.1×10^{-8}	
Reference Frequency		10 MHz	6.1×10^{-9}	
Amplitude		40 Hz ~ 100 MHz (0 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	0.18 dB 0.20 dB 0.22 dB 0.24 dB	
Display Scale		40 Hz ~ 100 MHz (0 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	0.18 dB 0.20 dB 0.22 dB 0.24 dB	
Frequency Response		(10 ~ -10) dBm 40 Hz ~ 1 MHz (1 ~ 100) MHz	0.18 dB 0.20 dB	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
LF spectrum analyzers	40427			Function generator RF signal generator, Measuring receiver, Frequency counter /KTICC-CI-40427
Frequency Span		8 Hz	1.1×10^{-4}	
		(8 ~ 16) Hz	5.3×10^{-4}	
		(16 ~ 40) Hz	2.1×10^{-4}	
		(40 ~ 80) Hz	1.1×10^{-4}	
		(80 ~ 160) Hz	5.3×10^{-4}	
		(160 ~ 400) Hz	2.1×10^{-4}	
		(0.4 ~ 0.8) kHz	1.1×10^{-4}	
		(0.8 ~ 1.6) kHz	5.3×10^{-4}	
		(1.6 ~ 4) kHz	2.1×10^{-4}	
		(4 ~ 8) kHz	1.1×10^{-4}	
		(8 ~ 16) kHz	5.3×10^{-4}	
		(16 ~ 40) kHz	2.1×10^{-4}	
		(40 ~ 80) kHz	1.1×10^{-4}	
		(80 ~ 160) kHz	5.3×10^{-4}	
		(160 ~ 400) kHz	2.1×10^{-4}	
		(0.4 ~ 0.8) MHz	1.1×10^{-4}	
		(0.8 ~ 1.6) MHz	5.3×10^{-4}	
		(1.6 ~ 4) MHz	2.1×10^{-4}	
		(4 ~ 8) MHz	1.1×10^{-4}	
		(8 ~ 16) MHz	5.3×10^{-4}	
Resolution Bandwidth		3 Hz	2.8×10^{-4}	
		(3 ~ 10) Hz	8.4×10^{-4}	
		(10 ~ 30) Hz	2.8×10^{-4}	
		(30 ~ 100) Hz	8.4×10^{-4}	
		(100 ~ 300) Hz	2.8×10^{-4}	
		(0.3 ~ 1) kHz	8.4×10^{-4}	
		(1 ~ 3) kHz	2.8×10^{-4}	
		(3 ~ 10) kHz	8.4×10^{-4}	
		(10 ~ 30) kHz	2.8×10^{-4}	
Output Level		(0.04 ~ 100) kHz		
		(10 ~ -20) dBm	0.059 dB	
		(-20 ~ -30) dBm	0.10 dB	
		(0.1 ~ 100) MHz		
		(10 ~ -30) dBm	0.11 dB	
Output Frequency		10 Hz ~ 100 MHz	6.1×10^{-9}	
Sweep generators	40429			DMM, Frequency counter, Measuring receiver, Oscilloscope /KTICC-CI-40429
Frequency		1 Hz ~ 100 MHz	6.1×10^{-9}	
		(100 ~ 400) MHz	2.6×10^{-9}	
AC Voltage		(0.22 ~ 1) mV		
		10 Hz	1.3×10^{-3}	
		(0.01 ~ 10) kHz	1.2×10^{-3}	
		(10 ~ 100) kHz	2.1×10^{-3}	
		(0.1 ~ 1) MHz	1.0×10^{-2}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Sweep generators AC Voltage	40429	(1 ~ 10) mV 10 Hz (0.01 ~ 10) kHz (10 ~ 50) kHz (50 ~ 100) kHz (0.1 ~ 1) MHz (10 ~ 100) mV 10 Hz (0.01 ~ 10) kHz (10 ~ 50) kHz (50 ~ 100) kHz (0.1 ~ 1) MHz (0.1 ~ 1) V 10 Hz (0.01 ~ 50) kHz (50 ~ 100) kHz (0.1 ~ 1) MHz (1 ~ 10) V 10 Hz (0.01 ~ 50) kHz (50 ~ 100) kHz (0.1 ~ 1) MHz (10 ~ 100) V 10 Hz (0.01 ~ 10) kHz (10 ~ 50) kHz (50 ~ 100) kHz	1.8×10^{-4} 1.6×10^{-4} 2.2×10^{-4} 2.7×10^{-4} 2.7×10^{-3} 9.3×10^{-5} 7.4×10^{-5} 7.7×10^{-5} 1.0×10^{-4} 1.0×10^{-3} 8.3×10^{-5} 6.7×10^{-5} 7.7×10^{-5} 9.6×10^{-4} 8.2×10^{-5} 6.8×10^{-5} 8.2×10^{-5} 1.2×10^{-3} 8.8×10^{-5} 6.9×10^{-5} 7.3×10^{-5} 9.1×10^{-5}	DMM, Frequency counter, Measuring receiver, Oscilloscope /KTICC-CI-40429
Level		10 Hz (20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm (10 ~ 40) Hz (20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm (0.04 ~ 10) kHz (20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm (0.01 ~ 10) MHz (20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm (10 ~ 50) MHz (20 ~ -50) dBm (-50 ~ -60) dBm (50 ~ 400) MHz (20 ~ -30) dBm (-30 ~ -60) dBm	0.008 dB 0.010 dB 0.49 dB 0.007 dB 0.010 dB 0.30 dB 0.007 dB 0.010 dB 0.21 dB 0.008 dB 0.011 dB 0.14 dB 0.024 dB 0.14 dB 0.12 dB 0.14 dB	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Sweep generators	40429	100 mV		DMM, Frequency counter, Measuring receiver, Oscilloscope /KTICC-CI-40429
Flatness (Voltage)		10 Hz (0.01 ~ 100) kHz (0.1 ~ 1) MHz (1 ~ 10) MHz (10 ~ 30) MHz (30 ~ 50) MHz (50 ~ 400) MHz (0.1 ~ 1) V 10 Hz (0.01 ~ 100) kHz (0.1 ~ 1) MHz (1 ~ 10) MHz (10 ~ 30) MHz (30 ~ 50) MHz (50 ~ 400) MHz (1 ~ 3) V 10 Hz (0.01 ~ 100) kHz (0.1 ~ 1) MHz (1 ~ 10) MHz (10 ~ 30) MHz (30 ~ 50) MHz (50 ~ 400) MHz	 8.7×10^{-4} 4.5×10^{-4} 5.1×10^{-4} 8.9×10^{-4} 3.0×10^{-3} 5.1×10^{-3} 2.6×10^{-2} 8.9×10^{-4} 4.0×10^{-4} 5.1×10^{-4} 8.9×10^{-4} 3.0×10^{-3} 5.0×10^{-3} 2.7×10^{-2} 9.0×10^{-4} 4.3×10^{-4} 5.3×10^{-4} 9.0×10^{-4} 3.0×10^{-3} 5.0×10^{-3} 2.8×10^{-2}	
Flatness (Level)		(10 ~ -10) dBm 10 Hz ~ 10 MHz (10 ~ 30) MHz (30 ~ 50) MHz (50 ~ 400) MHz	 0.008 dB 0.015 dB 0.023 dB 0.11 dB	
DC Offset Voltage		(±) 0 V (0 ~ 1) V (1 ~ 5) V (5 ~ 10) V	 6.1 μV 6.2×10^{-5} 1.3×10^{-5} 7.2×10^{-6}	
Attenuator		(0.04 ~ 100) kHz (30 ~ -40) dB (-40 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB (0.1 ~ 400) MHz (30 ~ -40) dB (-40 ~ -50) dB (-50 ~ -60) dB (-60 ~ -70) dB (-70 ~ -80) dB	 0.054 dB 0.10 dB 0.20 dB 0.30 dB 0.061 dB 0.068 dB 0.072 dB 0.076 dB 0.081 dB	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Sweep generators Frequency Modulation Amplitude Modulation Phase Duty cycle Rise/Fall Time	40429	(0 ~ 400) kHz (0 ~ 99) % (0 ~ 360)° (1 ~ 99) % 0.4 ns (0.4 ~ 1) ns 1 ns ~ 1 ms	2.8×10^{-2} 2.7×10^{-2} 0.061° 0.006 1 % 4.8×10^{-1} 9.0×10^{-2} 8.2×10^{-3}	DMM, Frequency counter, Measuring receiver, Oscilloscope /KTICC-CI-40429
Signal transducers DC Voltage DC Current Frequency	40430	(±) 0 V (0 ~ 1) mV (1 ~ 10) mV (10 ~ 100) mV (0.1 ~ 100) V (±) 0 A (0 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A 10 Hz ~ 100 kHz	0.12 μV 1.2×10^{-4} 6.3×10^{-5} 6.2×10^{-5} 6.1×10^{-5} 6.1 nA 6.5×10^{-5} 7.7×10^{-5} 2.0×10^{-4} 6.1×10^{-7}	DMM, Frequency counter /KTICC-CI-40430
Transistor curve tracers Input DC Voltage Input DC Current Output DC Voltage Output DC Current	40432	(±) 0 V (0 ~ 500) V (500 ~ 1 000) V (±) 0 A (0 ~ 5) mA (5 ~ 50) mA (50 ~ 500) mA (0.5 ~ 1) A (1 ~ 2) A (±) 0 V (0 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1 000) V (±) 0 A (0 ~ 10) mA (10 ~ 100) mA (0.1 ~ 1) A (1 ~ 2) A	6.1 μV 1.2×10^{-4} 6.1×10^{-4} 62 nA 1.3×10^{-4} 1.4×10^{-4} 1.6×10^{-4} 6.2×10^{-4} 3.2×10^{-4} 0.62 μV 6.3×10^{-5} 6.2×10^{-5} 6.1×10^{-5} 6.1 nA 6.5×10^{-5} 7.8×10^{-5} 2.0×10^{-4} 6.0×10^{-4}	Meter calibrator, DMM /KTICC-CI-40432

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Waveform analyzers	40433			Meter calibrator, Frequency counter, Audio analyzer /KTICC-CI-40433
Frequency		(0.01 ~ 100) kHz (100 ~ 200) kHz	6.1×10^{-6} 3.1×10^{-6}	
AC Voltage		(0.22 ~ 1) mV 10 Hz (10 ~ 40) Hz (0.04 ~ 10) kHz (10 ~ 100) kHz (100 ~ 200) kHz (1 ~ 10) mV 10 Hz (0.01 ~ 10) kHz (10 ~ 100) kHz (100 ~ 200) kHz (10 ~ 100) mV 10 Hz (0.01 ~ 10) kHz (10 ~ 100) kHz (100 ~ 200) kHz (0.01 ~ 1) V 10 Hz (10 ~ 40) Hz (0.04 ~ 10) kHz (10 ~ 100) kHz (100 ~ 200) kHz (1 ~ 10) V 10 Hz (10 ~ 40) Hz (0.04 ~ 10) kHz (10 ~ 100) kHz (100 ~ 200) kHz (10 ~ 100) V 10 Hz (10 ~ 40) Hz (0.04 ~ 10) kHz (10 ~ 100) kHz (100 ~ 200) V 10 Hz (10 ~ 40) Hz (0.04 ~ 10) kHz (10 ~ 100) kHz (200 ~ 300) V 50 Hz (50 ~ 1) kHz	4.4×10^{-3} 5.0×10^{-3} 4.2×10^{-3} 5.6×10^{-3} 1.1×10^{-2} 6.6×10^{-4} 5.1×10^{-4} 1.0×10^{-3} 2.1×10^{-3} 3.8×10^{-4} 1.8×10^{-4} 6.5×10^{-4} 1.1×10^{-3} 3.0×10^{-4} 1.3×10^{-4} 9.1×10^{-5} 1.6×10^{-4} 5.2×10^{-4} 3.0×10^{-4} 1.3×10^{-4} 8.9×10^{-5} 1.5×10^{-4} 3.6×10^{-4} 3.0×10^{-4} 1.3×10^{-4} 9.4×10^{-5} 2.0×10^{-4} 2.7×10^{-4} 1.1×10^{-4} 7.5×10^{-5} 1.9×10^{-4} 3.7×10^{-4} 9.3×10^{-5}	
DC Voltage		($\pm$) 0 V (0 ~ 100) mV (0.1 ~ 100) V (100 ~ 300) V	$6.1 \mu\text{V}$ 6.2×10^{-5} 6.1×10^{-5} 2.2×10^{-5}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Waveform analyzers Level	40433	10 Hz (40 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm (-60 ~ -70) dBm (0.01 ~ 1) kHz (50 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm (-60 ~ -70) dBm (1 ~ 10) kHz (40 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm (-60 ~ -70) dBm (10 ~ 100) kHz (40 ~ -20) dBm (-20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm (-60 ~ -70) dBm (100 ~ 200) kHz (20 ~ -10) dBm (-10 ~ -20) dBm (-20 ~ -30) dBm (-30 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -60) dBm (-60 ~ -70) dBm	0.009 dB 0.018 dB 0.048 dB 0.15 dB 0.009 dB 0.016 dB 0.047 dB 0.14 dB 0.008 dB 0.016 dB 0.047 dB 0.14 dB 0.009 dB 0.012 dB 0.023 dB 0.062 dB 0.18 dB 0.009 dB 0.012 dB 0.017 dB 0.022 dB 0.046 dB 0.12 dB 0.36 dB	Meter calibrator, Frequency counter, Audio analyzer /KTICC-CI-40433
Frequency Responses (Voltage)		100 mV (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 10) kHz (10 ~ 100) kHz (100 ~ 200) kHz (0.1 ~ 1) V (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 10) kHz (10 ~ 100) kHz (100 ~ 200) kHz (1 ~ 10) V (10 ~ 20) Hz (20 ~ 40) Hz (0.04 ~ 10) kHz (10 ~ 100) kHz (100 ~ 200) kHz	3.8×10^{-4} 1.8×10^{-4} 1.7×10^{-4} 6.5×10^{-4} 1.1×10^{-3} 3.0×10^{-4} 1.3×10^{-4} 9.1×10^{-5} 1.6×10^{-4} 5.2×10^{-4} 3.0×10^{-4} 1.3×10^{-4} 8.9×10^{-5} 1.5×10^{-4} 3.6×10^{-4}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Waveform analyzers Frequency Responses (Level) Distortion Filter(Frequency) (Weight, Low, High Pass etc.) Filter(Level) (Weight, Low, High Pass etc.)	40433	(10 ~ -10) dBm (10 ~ 40) Hz (0.04 ~ 100) kHz (100 ~ 200) kHz (0.02 ~ 100) kHz (0 ~ -60) dB (-60 ~ -80) dB (0.01 ~ 100) kHz (0.01 ~ 100) kHz (20 ~ -63) dB	0.008 dB 0.007 dB 0.010 dB 0.19 dB 0.50 dB 6.1×10^{-6} 0.007 dB	Meter calibrator, Frequency counter, Audio analyzer /KTICC-CI-40433
AC/DC high voltage generators DC Voltage AC Voltage	40434	($\pm$) 0 V (0 ~ 100) kV (50 ~ 60) Hz (0.001 ~ 100) kV	62 mV 1.2×10^{-3} 2.3×10^{-2}	High voltage test equipment, High voltage meter, DMM /KTICC-CI-40434
AC/DC high voltage probes DC Voltage AC Voltage Bandwidth	40435	($\pm$) (1 ~ 10 000) : 1 (0 ~ 1) kV (1 ~ 100) kV (1 ~ 10 000) : 1 (0.04 ~ 1) kHz (0.22 ~ 100) mV (0.1 ~ 100) V (0.1 ~ 1) kV (50 ~ 60) Hz (1 ~ 10) kV (5 ~ 600) mV 50 kHz (0.05 ~ 100) MHz (100 ~ 300) MHz (300 ~ 500) MHz (0.5 ~ 1) GHz (0.6 ~ 3) V 50 kHz (0.05 ~ 100) MHz (100 ~ 300) MHz (300 ~ 500) MHz (0.5 ~ 1) GHz	6.2×10^{-5} 1.2×10^{-3} 3.8×10^{-4} 2.3×10^{-4} 7.4×10^{-4} 1.3×10^{-2} 3.1×10^{-2} 3.4×10^{-2} 5.2×10^{-2} 6.2×10^{-2} 8.5×10^{-2} 5.8×10^{-2} 6.3×10^{-2} 7.8×10^{-2} 9.5×10^{-2} 1.2×10^{-1}	Meter calibrator, DMM, High voltage test equipment, Oscilloscope, High voltage power supply / KTICC-CI-4035

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Logic analyzers Vertical(Square Wave)	40436	1 kHz		Scope calibrator Meter calibrator /KTICC-CI-40436
		1 mV	6.4×10^{-3}	
		(1 ~ 2) mV	3.5×10^{-3}	
		(2 ~ 5) mV	1.7×10^{-3}	
		(5 ~ 10) mV	1.2×10^{-3}	
		(10 ~ 20) mV	8.5×10^{-4}	
		(20 ~ 50) mV	7.0×10^{-4}	
		(50 ~ 100) mV	6.4×10^{-4}	
		(100 ~ 200) mV	6.0×10^{-4}	
		(200 ~ 500) mV	5.8×10^{-4}	
		(0.5 ~ 1) V	5.9×10^{-4}	
		(1 ~ 2) V	6.0×10^{-4}	
		(2 ~ 10) V	5.8×10^{-4}	
		(10 ~ 20) V	6.0×10^{-4}	
		(20 ~ 130) V	5.8×10^{-4}	
Bandwidth		600 mV		
		50 kHz	1.5×10^{-2}	
		(0.05 ~ 100) MHz	2.0×10^{-2}	
		(100 ~ 300) MHz	2.5×10^{-2}	
		(300 ~ 500) MHz	4.2×10^{-2}	
		(0.5 ~ 1) GHz	5.8×10^{-2}	
		(1 ~ 2) GHz	7.0×10^{-2}	
Time(Period)		1 ns	6.1×10^{-5}	
		(1 ~ 2) ns	3.1×10^{-5}	
		(2 ~ 5) ns	1.2×10^{-5}	
		(5 ~ 10) ns	6.1×10^{-5}	
		(10 ~ 20) ns	3.1×10^{-5}	
		(20 ~ 50) ns	1.2×10^{-5}	
		(50 ~ 100) ns	6.1×10^{-5}	
		(100 ~ 200) ns	3.1×10^{-5}	
		(200 ~ 500) ns	1.2×10^{-5}	
		(0.5 ~ 1) μ s	6.1×10^{-5}	
		(1 ~ 2) μ s	3.1×10^{-5}	
		(2 ~ 5) μ s	1.2×10^{-5}	
		(5 ~ 10) μ s	6.1×10^{-5}	
		(10 ~ 20) μ s	3.1×10^{-5}	
		(20 ~ 50) μ s	1.2×10^{-5}	
		(50 ~ 100) μ s	6.1×10^{-5}	
		(100 ~ 200) μ s	3.1×10^{-5}	
		(200 ~ 500) μ s	1.2×10^{-5}	
		(0.5 ~ 1) ms	6.1×10^{-5}	
		(1 ~ 2) ms	3.1×10^{-5}	
		(2 ~ 5) ms	1.2×10^{-5}	
		(5 ~ 10) ms	6.1×10^{-5}	
		(10 ~ 20) ms	3.1×10^{-5}	
		(20 ~ 50) ms	1.4×10^{-5}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Logic analyzers Time(Period) Threshold Voltage	40436	(50 ~ 100) ms (100 ~ 200) ms (200 ~ 500) ms (0.5 ~ 1) s (1 ~ 2) s (2 ~ 5) s (-10 ~ 10) V	6.1×10^{-5} 3.1×10^{-5} 1.3×10^{-5} 6.1×10^{-5} 3.1×10^{-5} 1.3×10^{-5} 6.1×10^{-4}	Scope calibrator Meter calibrator /KTICC-CI-40436
Telephone testers Line Output Voltage Loop Current & Artificial Line Bell Output Voltage Bell Output Frequency Receiving Frequency DTMF Level	40437	($\pm$) 0 V (0 ~ 10) V (10 ~ 16) V (16 ~ 35) V (35 ~ 48) V (45 ~ 75) V (75 ~ 100) V (0 ~ 10) mA (10 ~ 20) mA (20 ~ 40) mA (40 ~ 60) mA (60 ~ 80) mA (80 ~ 100) mA (100 ~ 120) mA (120 ~ 150) mA (40 ~ 100) Hz (1 ~ 10) V (10 ~ 20) V (20 ~ 40) V (40 ~ 60) V (60 ~ 80) V (80 ~ 100) V (100 ~ 150) V 10 Hz (10 ~ 20) Hz (20 ~ 40) Hz (40 ~ 60) Hz (60 ~ 80) Hz (80 ~ 100) Hz 400 Hz (697 ~ 1 663) Hz (0 ~ -20) dBm (-20 ~ -30) dBm	61 mV 6.1×10^{-5} 3.9×10^{-5} 1.9×10^{-5} 1.4×10^{-5} 1.0×10^{-5} 6.1×10^{-5} 6.5×10^{-5} 8.5×10^{-5} 6.0×10^{-5} 5.2×10^{-5} 4.9×10^{-5} 7.7×10^{-5} 6.8×10^{-5} 6.0×10^{-5} 1.4×10^{-4} 2.0×10^{-4} 1.5×10^{-4} 1.4×10^{-4} 1.3×10^{-4} 1.4×10^{-4} 1.3×10^{-4} 6.1×10^{-5} 3.1×10^{-5} 1.6×10^{-5} 1.2×10^{-5} 9.6×10^{-6} 6.1×10^{-5} 1.6×10^{-5} 0.12 dB 0.15 dB	DMM, DTMF generator, Frequency counter /KTICC-CI-40437

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Telephone testers DTMF Frequency	40437	697 Hz	2.9×10^{-4}	DMM, DTMF generator, Frequency counter /KTICC-CI-40437
		(697 ~ 941) Hz	2.7×10^{-4}	
		(941 ~ 1 336) Hz	2.5×10^{-4}	
		(1 336 ~ 1 663) Hz	2.4×10^{-4}	
Output Frequency		(10 ~ 1 000) Hz	6.1×10^{-7}	
		(1 000 ~ 2 000) Hz	3.1×10^{-7}	
		(2 000 ~ 3 000) Hz	2.0×10^{-7}	
Output Level		(0.04 ~ 3) kHz		
		(0 ~ -30) dBm	0.051 dB	
		(-30 ~ -40) dBm	0.10 dB	
Video signal analyzers Frequency	40438	10 Hz ~ 1 GHz	3.7×10^{-10}	Video signal analyzer Video signal generator /KTICC-CI-40438
Squarewave Amplitude		NTSC, PAL		
		(1 ~ 10) mV	1.4×10^{-2}	
		(10 ~ 100) mV	1.9×10^{-3}	
		(100 ~ 200) mV	1.2×10^{-3}	
		(200 ~ 300) mV	1.0×10^{-3}	
		(300 ~ 400) mV	8.9×10^{-4}	
		(400 ~ 500) mV	8.8×10^{-4}	
		(500 ~ 600) mV	8.2×10^{-4}	
		(600 ~ 700) mV	7.8×10^{-4}	
		(700 ~ 800) mV	7.6×10^{-4}	
Sinewave Amplitude		(800 ~ 900) mV	7.3×10^{-4}	
		(900 ~ 999.9) mV	7.2×10^{-4}	
		(5 ~ 600) mV		
		50 kHz	4.2×10^{-3}	
		50 kHz ~ 10 MHz	8.8×10^{-3}	
Burst Frequency		(3 ~ 5) MHz	1.0 Hz	
Luminance Amplitude		NTSC, PAL		
		(1 ~ 10) mV	1.4×10^{-2}	
		(10 ~ 100) mV	1.9×10^{-3}	
		(100 ~ 200) mV	1.2×10^{-3}	
		(200 ~ 300) mV	1.0×10^{-3}	
		(300 ~ 400) mV	8.9×10^{-4}	
		(400 ~ 500) mV	8.8×10^{-4}	
		(500 ~ 600) mV	8.2×10^{-4}	
		(600 ~ 700) mV	7.8×10^{-4}	
		(700 ~ 800) mV	7.6×10^{-4}	
Chrominance Amplitude		(800 ~ 900) mV	7.3×10^{-4}	
		(900 ~ 999.9) mV	7.2×10^{-4}	
		NTSC, PAL		
		(1 ~ 714.3) mV	1.9×10^{-2}	

404. Other DC & LF measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Video signal analyzers	40438	(5 ~ 600) mV		Video signal analyzer Video signal generator /KTICC-CI-40438
Frequency Response		0.05 MHz	2.4×10^{-3}	
		(0.05 ~ 20) MHz	8.1×10^{-3}	
Horizontal Frequency		10 ns	7.0×10^{-4}	
		(10 ~ 20) ns	3.5×10^{-4}	
		(20 ~ 50) ns	1.4×10^{-4}	
		(50 ~ 100) ns	7.0×10^{-4}	
		(100 ~ 200) ns	3.5×10^{-4}	
		(200 ~ 500) ns	1.4×10^{-4}	
		(0.5 ~ 1) μ s	7.0×10^{-4}	
		(1 ~ 2) μ s	3.5×10^{-4}	
		(2 ~ 5) μ s	1.4×10^{-4}	
		(5 ~ 10) μ s	7.0×10^{-4}	
		(10 ~ 20) μ s	3.5×10^{-4}	
		(20 ~ 50) μ s	1.4×10^{-4}	
		(50 ~ 100) μ s	7.0×10^{-4}	
		(100 ~ 200) μ s	3.5×10^{-4}	
		(200 ~ 500) μ s	1.4×10^{-4}	
		(0.5 ~ 1) ms	7.0×10^{-4}	
Phase		NTSC, PAL (0 ~ 360)°	1.3°	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF amplifiers Gain	40601	(0 ~ 10) dB (5 ~ 100) Hz 100 Hz ~ 0.3 MHz 0.3 MHz ~ 1 GHz (1 ~ 2) GHz (2 ~ 8) GHz (8 ~ 18) GHz (18 ~ 40) GHz (10 ~ 20) dB (5 ~ 100) Hz 100 Hz ~ 0.3 MHz 0.3 MHz ~ 1 GHz (1 ~ 2) GHz (2 ~ 8) GHz (8 ~ 18) GHz (18 ~ 40) GHz (20 ~ 30) dB (5 ~ 100) Hz 100 Hz ~ 0.3 MHz 0.3 MHz ~ 1 GHz (1 ~ 2) GHz (2 ~ 8) GHz (8 ~ 18) GHz (18 ~ 40) GHz (30 ~ 40) dB (5 ~ 100) Hz 100 Hz ~ 0.3 MHz 0.3 MHz ~ 1 GHz (1 ~ 2) GHz (2 ~ 8) GHz (8 ~ 18) GHz (18 ~ 40) GHz (40 ~ 50) dB (5 ~ 100) Hz 100 Hz ~ 0.3 MHz 0.3 MHz ~ 1 GHz (1 ~ 2) GHz (2 ~ 8) GHz (8 ~ 18) GHz (18 ~ 40) GHz (50 ~ 60) dB (5 ~ 100) Hz 100 Hz ~ 0.3 MHz 0.3 MHz ~ 1 GHz (1 ~ 2) GHz (2 ~ 8) GHz (8 ~ 18) GHz (18 ~ 40) GHz	0.09 dB 0.06 dB 0.05 dB 0.06 dB 0.09 dB 0.12 dB 0.17 dB 0.11 dB 0.08 dB 0.07 dB 0.08 dB 0.11 dB 0.14 dB 0.19 dB 0.13 dB 0.09 dB 0.08 dB 0.10 dB 0.12 dB 0.16 dB 0.20 dB 0.15 dB 0.11 dB 0.10 dB 0.11 dB 0.14 dB 0.18 dB 0.22 dB 0.16 dB 0.14 dB 0.13 dB 0.12 dB 0.16 dB 0.19 dB 0.24 dB 0.19 dB 0.18 dB 0.17 dB 0.25 dB 0.28 dB 0.31 dB 0.36 dB	Network Analyzer, Calibration Kit /KTICC-CI-40601

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF amplifiers Gain	40601	(60 ~ 70) dB (5 ~ 100) Hz 100 Hz ~ 0.3 MHz 0.3 MHz ~ 1 GHz (1 ~ 2) GHz (2 ~ 8) GHz (8 ~ 18) GHz (18 ~ 40) GHz	0.27 dB 0.25 dB 0.24 dB 0.28 dB 0.31 dB 0.35 dB 0.39 dB	Network Analyzer, Calibration Kit /KTICC-CI-40601
Coaxial attenuators Attenuation	40602	(0 ~ 10) dB (5 ~ 100) Hz 100 Hz ~ 26.5 GHz (26.5 ~ 40) GHz (40 ~ 50) GHz (10 ~ 20) dB (5 ~ 100) Hz 100 Hz ~ 0.1 MHz 0.1 MHz ~ 26.5 GHz (26.5 ~ 40) GHz (40 ~ 50) GHz (20 ~ 30) dB (5 ~ 100) Hz 100 Hz ~ 0.1 MHz 0.1 MHz ~ 26.5 GHz (26.5 ~ 40) GHz (40 ~ 50) GHz (30 ~ 40) dB (5 ~ 100) Hz 100 Hz ~ 0.1 MHz 0.1 MHz ~ 26.5 GHz (26.5 ~ 40) GHz (40 ~ 50) GHz (40 ~ 50) dB (5 ~ 100) Hz 100 Hz ~ 0.1 MHz 0.1 MHz ~ 18 GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz (40 ~ 45) GHz (45 ~ 50) GHz (50 ~ 60) dB (5 ~ 100) Hz 100 Hz ~ 0.1 MHz 0.1 MHz ~ 18 GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz (40 ~ 45) GHz (45 ~ 50) GHz	0.09 dB 0.06 dB 0.17 dB 0.25 dB 0.12 dB 0.08 dB 0.06 dB 0.18 dB 0.27 dB 0.14 dB 0.09 dB 0.06 dB 0.20 dB 0.28 dB 0.20 dB 0.12 dB 0.07 dB 0.22 dB 0.32 dB 0.33 dB 0.17 dB 0.07 dB 0.08 dB 0.27 dB 0.37 dB 0.39 dB 0.73 dB 0.27 dB 0.08 dB 0.09 dB 0.36 dB 0.48 dB 0.55 dB	Network Analyzer Sensor Module, Measuring Receiver, Signal Generator, Calibration Kit /KTICC-CI-40602

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Coaxial attenuators Attenuation	40602	(60 ~ 70) dB (5 ~ 100) Hz 100 Hz ~ 0.1 MHz 0.1 MHz ~ 18 GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz (40 ~ 45) GHz (45 ~ 50) GHz (70 ~ 90) dB 0.1 MHz ~ 18 GHz (18 ~ 26.5) GHz (90 ~ 120) dB 0.1 MHz ~ 18 GHz (18 ~ 26.5) GHz	1.8 dB 0.55 dB 0.08 dB 0.09 dB 0.55 dB 0.74 dB 0.97 dB 0.09 dB 0.10 dB 0.10 dB 0.11 dB	Network Analyzer Sensor Module, Measuring Receiver, Signal Generator, Calibration Kit /KTICC-CI-40602
BER(Bit Error Rate) testers Output Bit rate Input Bit rate	40604	(0.05 ~ 5) GHz (5 ~ 30) GHz (0.05 ~ 5) GHz (5 ~ 30) GHz	1.2×10^{-10} 1.3 Hz 0.9 Hz 1.3 Hz	Frequency counter, Frequency standard Signal generator /KTICC-CI-40604
Burst pulse generators Burst Voltage Burst Duration Burst Cycle Repetition Frequency Rise Time	40605	(±) 0 kV (0 ~ 4) kV 10 μs ~ 10 ms (10 ~ 15) ms (15 ~ 20) ms (20 ~ 50) ms 100 μs 100 μs ~ 10 ms (10 ~ 50) ms (50 ~ 100) ms (100 ~ 300) ms (300 ~ 400) ms (400 ~ 1 000) ms 1 kHz (1 ~ 5) kHz (5 ~ 10) kHz (10 ~ 50) kHz (50 ~ 100) kHz 1 ns (1 ~ 5) ns (5 ~ 10) ns (10 ~ 100) ns 100 ns ~ 1 μs	0.1 V 1.3×10^{-2} 3.4×10^{-3} 5.5×10^{-3} 4.1×10^{-3} 3.8×10^{-3} 3.4×10^{-3} 4.8×10^{-3} 3.9×10^{-3} 3.4×10^{-3} 5.0×10^{-3} 4.6×10^{-3} 3.4×10^{-3} 3.4×10^{-3} 4.8×10^{-3} 3.4×10^{-3} 4.8×10^{-3} 3.4×10^{-3} 9.0×10^{-2} 7.6×10^{-3} 9.1×10^{-3} 4.2×10^{-3} 3.3×10^{-3}	Oscilloscope. Attenuator High voltage probe /KTICC-CI-40605

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Burst pulse generators Pulse Width Vibration Frequency	40605	10 ns (10 ~ 50) ns (50 ~ 100) ns (100 ~ 150) ns 90 kHz 90 kHz ~ 1 MHz (1 ~ 10) MHz (10 ~ 100) MHz	3.6×10^{-3} 4.4×10^{-3} 4.2×10^{-3} 2.8×10^{-3} 1.0×10^{-3} 1.1×10^{-3} 8.8×10^{-4} 1.4×10^{-3}	Oscilloscope, Attenuator High voltage probe /KTICC-CI-40605
RF power meter calibrators Power	40607	3 μ W 10 μ W 30 μ W 100 μ W 300 μ W 1 mW 3 mW 10 mW 30 mW 100 mW	2.5×10^{-5} 1.2×10^{-5} 2.4×10^{-5} 1.0×10^{-5} 1.9×10^{-5} 5.0×10^{-5} 1.6×10^{-5} 4.8×10^{-5} 1.5×10^{-5} 4.5×10^{-5}	Digital multimeter /KTICC-CI-40607
EMC transduces; current probes, absorbing clamps, etc. Current probes Transfer Impedance Absorbing clamps Insertion Loss	40608	(5 ~ 100) Hz 100 Hz ~ 1 GHz 30 MHz ~ 1 GHz	 2.0 dB 1.8 dB 1.8 dB	Network analyzer, Calibration kit /KTICC-CI-40608
Coaxial directional couplers /splitters Coupling Factor, Insertion Loss	40610	(0 ~ 10) dB (5 ~ 100) Hz 100 Hz ~ 0.3 MHz 0.3 MHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (10 ~ 20) dB (5 ~ 100) Hz 100 Hz ~ 0.3 MHz 0.3 MHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (20 ~ 30) dB (5 ~ 100) Hz 100 Hz ~ 0.3 MHz 0.3 MHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz	 0.09 dB 0.06 dB 0.05 dB 0.10 dB 0.17 dB 0.25 dB 0.12 dB 0.08 dB 0.07 dB 0.12 dB 0.18 dB 0.27 dB 0.14 dB 0.09 dB 0.08 dB 0.13 dB 0.20 dB 0.28 dB	Network Analyzer, Calibration Kit /KTICC-CI-40610

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Coaxial directional couplers /splitters Coupling Factor, Insertion Loss	40610	(30 ~ 40) dB (5 ~ 100) Hz 100 Hz ~ 0.3 MHz 0.3 MHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 50) GHz (40 ~ 50) dB (5 ~ 100) Hz 100 Hz ~ 0.3 MHz 0.3 MHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 45) GHz (45 ~ 50) GHz (50 ~ 60) dB (5 ~ 100) Hz 100 Hz ~ 0.3 MHz 0.3 MHz ~ 2 GHz (2 ~ 20) GHz (20 ~ 40) GHz (40 ~ 45) GHz (45 ~ 50) GHz (60 ~ 70) dB (5 ~ 100) Hz 100 Hz ~ 0.3 MHz 0.3 MHz ~ 2 GHz (2 ~ 10) GHz (10 ~ 20) GHz (20 ~ 40) GHz (40 ~ 45) GHz (45 ~ 50) GHz	0.20 dB 0.12 dB 0.10 dB 0.16 dB 0.22 dB 0.32 dB 0.33 dB 0.17 dB 0.13 dB 0.20 dB 0.27 dB 0.37 dB 0.39 dB 0.73 dB 0.27 dB 0.17 dB 0.27 dB 0.36 dB 0.48 dB 0.55 dB 1.8 dB 0.55 dB 0.27 dB 0.42 dB 0.44 dB 0.55 dB 0.74 dB 0.97 dB	Network Analyzer, Calibration Kit /KTICC-CI-40610

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
DS1/DS3 communications systems	40612			Frequency counter Transmission analyzer Audeo analyzer Frequency standard, Oscilloscope
Bit rate		50 Hz ~ 5 GHz	1.2×10^{-10}	/KTICC-CI-40612
Level & Amplitude		50 Hz ~ 200 kHz		
		(10 ~ -30) dBm	0.051 dB	
		(-30 ~ -50) dBm	0.10 dB	
		(-50 ~ -60) dBm	0.21 dB	
Frequency Response & Flatness		(20 ~ 100) Hz	0.10 dB	
		(0.1 ~ 100) kHz	0.051 dB	
		(0.1 ~ 1) MHz	0.081 dB	
Output Jitter		DS1 (10 Hz to 40 kHz)	0.023 UI	
		DS1 (8 kHz to 40 kHz)	0.023 UI	
		E1 (20 Hz to 100 kHz)	0.023 UI	
		E1 (18 kHz to 100 kHz)	0.023 UI	
		DS3 (10 Hz to 400 kHz)	0.035 UI	
		DS3 (30 kHz to 400 kHz)	0.035 UI	
Jitter Generator & Analyzer		DS1(1.544 MHz), 1 kHz		
		0.77 UI	0.09 UI	
		1.80 UI	0.26 UI	
		4.80 UI	0.50 UI	
		8.80 UI	0.83 UI	
		E1(2.048 MHz), 2.4 kHz		
		0.77 UI	0.09 UI	
		1.80 UI	0.26 UI	
		4.80 UI	0.50 UI	
		8.80 UI	0.83 UI	
		DS3(44.736 MHz), 4 kHz		
		0.77 UI	0.11 UI	
		1.80 UI	0.30 UI	
		4.80 UI	0.61 UI	
		8.80 UI	1.0 UI	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Electrostatic discharge generators	40613			Oscilloscope, Attenuator, ESD System /KTICC-CI-40613
Output Voltage		($\pm$) 0 kV (0 ~ 1) kV (1 ~ 2) kV (2 ~ 20) kV (20 ~ 30) kV	0.01 kV 8.7×10^{-3} 5.4×10^{-3} 3.8×10^{-3} 3.7×10^{-3}	
Peak Current		($\pm$) 0 A (0 ~ 7.5) A (7.5 ~ 15) A (15 ~ 22.5) A (22.5 ~ 30) A (30 ~ 56) A (56 ~ 112.5) A	0.01 A 1.3×10^{-2} 1.3×10^{-2} 1.3×10^{-2} 1.3×10^{-2} 1.3×10^{-2} 1.4×10^{-2}	
T1 Current(30 ns ~ 65 ns)		($\pm$) 0 A (0 ~ 4) A (4 ~ 8) A (8 ~ 12) A (12 ~ 16) A (16 ~ 30) A 30 A ~ 60 A	0.01 A 1.3×10^{-2} 1.3×10^{-2} 1.3×10^{-2} 1.3×10^{-2} 1.3×10^{-2} 1.3×10^{-2}	
T2 Current(60 ns ~ 130 ns)		($\pm$) 0 A (0 ~ 2) A (2 ~ 4) A (4 ~ 6) A (6 ~ 8) A (8 ~ 15) A (15 ~ 30) A	0.01 A 1.3×10^{-2} 1.3×10^{-2} 1.3×10^{-2} 1.3×10^{-2} 1.4×10^{-2} 1.3×10^{-2}	
T1 Current(180 ns ~ 400 ns)		($\pm$) 0 A (0 ~ 0.55) A (0.55 ~ 1.1) A (1.1 ~ 1.65) A (1.65 ~ 2.2) A (2.2 ~ 4.125) A (4.125 ~ 8.25) A	0.01 A 2.1×10^{-2} 1.5×10^{-2} 1.4×10^{-2} 1.3×10^{-2} 1.3×10^{-2} 1.6×10^{-2}	
T2 Current(360 ns ~ 800 ns)		($\pm$) 0 A (0 ~ 0.3) A (0.3 ~ 0.6) A (0.6 ~ 0.9) A (0.9 ~ 1.2) A (1.2 ~ 2.3) A (2.3 ~ 4.5) A	0.01 A 3.2×10^{-2} 2.0×10^{-2} 1.6×10^{-2} 1.5×10^{-2} 1.3×10^{-2} 1.3×10^{-2}	
Rjse/Fall Time		(0.5 ~ 1) ns	0.02 ns	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
EMC receivers	40614	9 kHz ~ 5 GHz	1.2×10^{-10}	Power sensor, Modulation meter, Power meter, Frequency counter, Measuring receiver, Signal generator /KTICC-CI-40614
Frequency		(5 ~ 40) GHz	1.3 Hz	
Input Level		(0.1 ~ 1 000) MHz		
		(20 ~ 10) dBm	0.21 dB	
		(10 ~ 0) dBm	0.20 dB	
		(0 ~ -40) dBm	0.21 dB	
		(-40 ~ -60) dBm	0.22 dB	
		(-60 ~ -90) dBm	0.23 dB	
		(-90 ~ -110) dBm	0.24 dB	
		(-110 ~ -120) dBm	0.25 dB	
		(1 ~ 4) GHz		
		(20 ~ -20) dBm	0.21 dB	
		(-20 ~ -40) dBm	0.22 dB	
		(-40 ~ -70) dBm	0.23 dB	
		(-70 ~ -100) dBm	0.24 dB	
		(-100 ~ -120) dBm	0.25 dB	
		(4 ~ 8) GHz		
		(20 ~ -20) dBm	0.22 dB	
		(-20 ~ -40) dBm	0.23 dB	
		(-40 ~ -70) dBm	0.24 dB	
		(-70 ~ -100) dBm	0.25 dB	
		(-100 ~ -120) dBm	0.26 dB	
		(8 ~ 10) GHz		
		(20 ~ -20) dBm	0.27 dB	
		(-20 ~ -50) dBm	0.28 dB	
		(-50 ~ -80) dBm	0.29 dB	
		(-80 ~ -110) dBm	0.30 dB	
		(-110 ~ -120) dBm	0.31 dB	
		(10 ~ 12) GHz		
		(20 ~ 10) dBm	0.28 dB	
		(-20 ~ 0) dBm	0.27 dB	
		(0 ~ -40) dBm	0.28 dB	
		(-40 ~ -60) dBm	0.29 dB	
		(-60 ~ -90) dBm	0.30 dB	
		(-90 ~ -120) dBm	0.31 dB	
		(12 ~ 18) GHz		
		(20 ~ -30) dBm	0.29 dB	
		(-30 ~ -50) dBm	0.30 dB	
		(-50 ~ -80) dBm	0.31 dB	
		(-80 ~ -100) dBm	0.32 dB	
		(-100 ~ -120) dBm	0.33 dB	
		(18 ~ 26.5) GHz		
		(20 ~ -20) dBm	0.37 dB	
		(-20 ~ -40) dBm	0.38 dB	
		(-40 ~ -50) dBm	0.39 dB	
		(-50 ~ -80) dBm	0.40 dB	
		(-80 ~ -110) dBm	0.41 dB	
		(-110 ~ -120) dBm	0.42 dB	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
EMC receivers	40614			Power sensor, Modulation meter, Power meter, Frequency counter, Measuring receiver, Signal generator /KTICC-CI-40614
Input Level		(26.5 ~ 40) GHz (20 ~ -20) dBm	0.50 dB	
Output Level		(0.1 ~ 1 000) MHz (20 ~ 10) dBm (10 ~ -10) dBm (-10 ~ -30) dBm (-30 ~ -40) dBm (-40 ~ -60) dBm (-60 ~ -70) dBm (-70 ~ -90) dBm (-90 ~ -110) dBm (-110 ~ -120) dBm	0.12 dB 0.11 dB 0.12 dB 0.13 dB 0.14 dB 0.15 dB 0.16 dB 0.17 dB 0.18 dB	
		(1 ~ 4) GHz (20 ~ 10) dBm (10 ~ -10) dBm (-10 ~ -30) dBm (-30 ~ -40) dBm (-40 ~ -60) dBm (-60 ~ -80) dBm (-80 ~ -100) dBm (-100 ~ -120) dBm	0.13 dB 0.12 dB 0.13 dB 0.14 dB 0.15 dB 0.16 dB 0.17 dB 0.18 dB	
		(4 ~ 8) GHz (20 ~ 10) dBm (10 ~ 0) dBm (0 ~ -20) dBm (-20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -90) dBm (-90 ~ -100) dBm (-100 ~ -120) dBm	0.14 dB 0.13 dB 0.14 dB 0.15 dB 0.16 dB 0.17 dB 0.18 dB 0.19 dB 0.20 dB	
		(8 ~ 10) GHz (20 ~ 10) dBm (10 ~ 0) dBm (0 ~ -20) dBm (-20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -90) dBm (-90 ~ -110) dBm (-110 ~ -120) dBm	0.15 dB 0.14 dB 0.15 dB 0.16 dB 0.17 dB 0.18 dB 0.19 dB 0.20 dB 0.21 dB	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
EMC receivers Output Level	40614	(10 ~ 12) GHz (20 ~ 10) dBm (10 ~ 0) dBm (0 ~ -20) dBm (-20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -90) dBm (-90 ~ -110) dBm (-110 ~ -120) dBm (12 ~ 18) GHz (20 ~ 10) dBm (10 ~ 0) dBm (0 ~ -20) dBm (-20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -90) dBm (-90 ~ -100) dBm (-100 ~ -120) dBm (18 ~ 26.5) GHz (20 ~ 10) dBm (10 ~ -10) dBm (-10 ~ -30) dBm (-30 ~ -40) dBm (-40 ~ -60) dBm (-60 ~ -80) dBm (-80 ~ -100) dBm (-100 ~ -120) dBm (26.5 ~ 40) GHz (20 ~ -20) dBm	0.16 dB 0.15 dB 0.16 dB 0.17 dB 0.18 dB 0.19 dB 0.20 dB 0.21 dB 0.22 dB 0.18 dB 0.17 dB 0.18 dB 0.19 dB 0.20 dB 0.21 dB 0.22 dB 0.23 dB 0.24 dB 0.25 dB 0.24 dB 0.25 dB 0.26 dB 0.28 dB 0.29 dB 0.30 dB 0.31 dB 0.21 dB	Power sensor, Modulation meter, Power meter, Frequency counter, Measuring receiver, Signal generator /KTICC-CI-40614
Frequency Modulation		Rate (0.01 ~ 100) kHz (0 ~ 400) kHz	2.7×10^{-2}	
Amplitude Modulation		Rate (0.01 ~ 50) kHz (0 ~ 99) %	2.7×10^{-2}	
Phase Modulation		Rate (0.05 ~ 100) kHz (0 ~ 400) rad	4.2×10^{-2}	
VSWR		(1 ~ 3) (9 ~ 100) kHz 100 kHz ~ 2 GHz (2 ~ 26.5) GHz (26.5 ~ 40) GHz	0.008 0.009 0.015 0.029	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
EMC receivers	40614			Power sensor, Modulation meter, Power meter, Frequency counter, Measuring receiver, Signal generator /KTICC-CI-40614
Frequency Response(CW)		10 Hz ~ 100 kHz	0.18 dB	
		100 kHz ~ 1 GHz	0.20 dB	
		(1 ~ 4) GHz	0.21 dB	
		(4 ~ 8) GHz	0.22 dB	
		(8 ~ 12) GHz	0.27 dB	
		(12 ~ 18) GHz	0.29 dB	
		(18 ~ 26.5) GHz	0.37 dB	
		(26.5 ~ 40) GHz	0.50 dB	
Frequency Response(Pulse)		9 kHz ~ 1 GHz	0.75 dB	
Frequency Response, Repetition(CISPR Band)		9 kHz ~ 1 GHz	0.75 dB	
Overall Selectivity		9 kHz ~ 1 GHz	0.07 dB	
IF Rejection		9 kHz ~ 100 kHz	0.06 dB	
		100 kHz ~ 1 GHz	0.11 dB	
		(1 ~ 4) GHz	0.12 dB	
		(4 ~ 8) GHz	0.13 dB	
		(8 ~ 10) GHz	0.14 dB	
		(10 ~ 12) GHz	0.15 dB	
		(12 ~ 18) GHz	0.18 dB	
		(18 ~ 26.5) GHz	0.24 dB	
(26.5 ~ 40) GHz		0.20 dB		
IF Image Frequency Rejection		9 kHz ~ 100 kHz	0.06 dB	
		100 kHz ~ 1 GHz	0.11 dB	
		(1 ~ 4) GHz	0.12 dB	
		(4 ~ 8) GHz	0.13 dB	
		(8 ~ 10) GHz	0.14 dB	
		(10 ~ 12) GHz	0.15 dB	
		(12 ~ 18) GHz	0.18 dB	
		(18 ~ 26.5) GHz	0.24 dB	
(26.5 ~ 40) GHz		0.20 dB		
Other Spurious Response		9 kHz ~ 100 kHz	0.06 dB	
		100 kHz ~ 1 GHz	0.11 dB	
		(1 ~ 4) GHz	0.12 dB	
		(4 ~ 8) GHz	0.13 dB	
		(8 ~ 10) GHz	0.14 dB	
		(10 ~ 12) GHz	0.15 dB	
		(12 ~ 18) GHz	0.18 dB	
		(18 ~ 26.5) GHz	0.24 dB	
(26.5 ~ 40) GHz		0.20 dB		

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
EMC receivers Random Noise	40614	9 kHz ~ 100 kHz 100 kHz ~ 1 GHz (1 ~ 4) GHz (4 ~ 8) GHz (8 ~ 10) GHz (10 ~ 12) GHz (12 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz	0.06 dB 0.11 dB 0.12 dB 0.13 dB 0.14 dB 0.15 dB 0.18 dB 0.24 dB 0.20 dB	Power sensor, Modulation meter, Power meter, Frequency counter, Measuring receiver, Signal generator /KTICC-CI-40614
Resolution Bandwidth		10 Hz ~ 10 MHz	1.1×10^{-3}	
RF filters Cutoff Frequency	40615	(5 ~ 100) Hz 100 Hz ~ 300 kHz 0.3 MHz ~ 1 GHz (1 ~ 2) GHz (2 ~ 8) GHz (8 ~ 18) GHz (18 ~ 40) GHz	0.60 Hz 72 Hz 72 kHz 0.12 MHz 0.16 MHz 0.31 MHz 0.42 MHz	Network analyzer Frequency standard, Calibration kit /KTICC-CI-40615
Insertion Loss		5 Hz ~ 9 kHz 9 kHz ~ 0.3 MHz 0.3 MHz ~ 1 GHz (1 ~ 2) GHz (2 ~ 8) GHz (8 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz	0.047 dB 0.041 dB 0.040 dB 0.041 dB 0.075 dB 0.11 dB 0.15 dB 0.16 dB	
RF impedance meters VSWR	40616	(1.00 ~ 1.05) (0.05 ~ 2) GHz (2 ~ 7) GHz (7 ~ 9) GHz (9 ~ 11) GHz (11 ~ 12) GHz (12 ~ 18) GHz (1.05 ~ 1.20) (0.05 ~ 1) GHz (1 ~ 8) GHz (8 ~ 12) GHz (12 ~ 18) GHz (1.20 ~ 1.50) (0.05 ~ 1) GHz (1 ~ 8) GHz (8 ~ 12) GHz (12 ~ 18) GHz (1.50 ~ 2.00) (0.05 ~ 1) GHz (1 ~ 4) GHz (4 ~ 12) GHz (12 ~ 17) GHz (17 ~ 18) GHz	0.059 0.062 0.063 0.062 0.063 0.095 0.071 0.074 0.085 0.12 0.095 0.10 0.11 0.20 0.14 0.15 0.16 0.26 0.27	Standard mismatch, Power sensor, Power meter, Spectrum analyzer, Frequency standard, Frequency counter, Calibration kit /KTICC-CI-40616

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF impedance meters	40616	9 kHz ~ 0.1 MHz (10 ~ -30) dBm	0.07 dB	Standard mismatch, Power sensor, Power meter, Spectrum analyzer, Frequency standard, Frequency counter, Calibration kit /KTICC-CI-40616
Output Power		(0.1 ~ 1 000) MHz (10 ~ -10) dBm (-10 ~ -30) dBm	0.11 dB 0.12 dB	
		(1 ~ 4) GHz (10 ~ -10) dBm (-10 ~ -30) dBm	0.12 dB 0.13 dB	
		(4 ~ 8) GHz (10 ~ 0) dBm 0 dBm (0 ~ -20) dBm (-20 ~ -30) dBm	0.14 dB 0.13 dB 0.14 dB 0.15 dB	
		(8 ~ 10) GHz (10 ~ 0) dBm 0 dBm (0 ~ -20) dBm (-20 ~ -30) dBm	0.15 dB 0.14 dB 0.15 dB 0.16 dB	
		(10 ~ 12) GHz (10 ~ 0) dBm 0 dBm (0 ~ -20) dBm (-20 ~ -30) dBm	0.16 dB 0.15 dB 0.16 dB 0.17 dB	
		(12 ~ 18) GHz (10 ~ 0) dBm 0 dBm (0 ~ -20) dBm (-20 ~ -30) dBm	0.18 dB 0.17 dB 0.18 dB 0.19 dB	
Frequency		10 Hz ~ 5 GHz	1.2×10^{-10}	
		(5 ~ 40) GHz	1.3 Hz	
Loss		(1 ~ 3) dB (0.05 ~ 1) GHz (1 ~ 3) GHz (3 ~ 8) GHz (8 ~ 10) GHz (10 ~ 15) GHz (15 ~ 18) GHz	0.040 dB 0.042 dB 0.040 dB 0.043 dB 0.048 dB 0.056 dB	
		(3 ~ 6) dB (0.05 ~ 1) GHz (1 ~ 3) GHz (3 ~ 8) GHz (8 ~ 10) GHz (10 ~ 15) GHz (15 ~ 18) GHz	0.041 dB 0.043 dB 0.041 dB 0.044 dB 0.048 dB 0.052 dB	
		(6 ~ 10) dB (0.05 ~ 1) GHz (1 ~ 3) GHz (3 ~ 8) GHz (8 ~ 10) GHz (10 ~ 15) GHz (15 ~ 18) GHz	0.041 dB 0.043 dB 0.042 dB 0.044 dB 0.042 dB 0.048 dB	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF impedance meters Impedance	40616	(0.05 ~ 2) GHz (2 ~ 18) GHz	0.6 Ω 1.1 Ω	Standard mismatch, Power sensor, Power meter, Calibration kit /KTICC-CI-40616
RF impulse generators Output Voltage Pulse Width	40617	($\pm$) 0 kV (0 ~ 20) kV 10 ns (10 ~ 100) ns 100 ns ~ 100 ms	0.1 V 1.3×10^{-2} 3.6×10^{-3} 4.2×10^{-3} 3.4×10^{-3}	Oscilloscope, High voltage probe /KTICC-CI-40617
Line impedance stabilization networks ; LISN, CDN, ISN, etc. Impedance Insertion Loss (Voltage Division Factor, Longitudinal conversion loss, Attenuation, Isolation) Phase Angle CDN Impedance Insertion Loss (Voltage Division Factor, Longitudinal conversion loss) Phase Angle	40618	(9 ~ 30) kHz 30 kHz ~ 1 GHz (9 ~ 30) kHz (30 ~ 50) kHz 50 kHz ~ 1 GHz (9 ~ 30) kHz 30 kHz ~ 1 GHz 9 kHz ~ 10 MHz 10 MHz ~ 1 GHz (9 ~ 300) kHz 300 kHz ~ 10 MHz 10 MHz ~ 1 GHz 9 kHz ~ 10 MHz 10 MHz ~ 1 GHz	0.45 Ω 0.40 Ω 0.060 dB 0.051 dB 0.041 dB 0.52° 0.46° 0.61 Ω 0.83 Ω 0.060 dB 0.046 dB 0.050 dB 0.23° 0.33°	Network analyzer Calibration kit /KTICC-CI-40618
Coaxial standard mismatches VSWR	40619	1.00 5 Hz ~ 0.1 MHz 0.1 MHz ~ 2 GHz (2 ~ 26.5) GHz (26.5 ~ 40) GHz (1 ~ 1.05) 5 Hz ~ 0.1 MHz 0.1 MHz ~ 2 GHz (2 ~ 26.5) GHz (26.5 ~ 40) GHz (1.05 ~ 1.10) 5 Hz ~ 0.1 MHz 0.1 MHz ~ 2 GHz (2 ~ 26.5) GHz (26.5 ~ 40) GHz	0.008 0.009 0.015 0.029 0.009 0.010 0.016 0.031 0.010 0.011 0.017 0.033	Network analyzer Calibration kit /KTICC-CI-40619

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Coaxial standard mismatches VSWR	40619	(1.10 ~ 1.20) 5 Hz ~ 0.1 MHz 0.1 MHz ~ 2 GHz (2 ~ 26.5) GHz (26.5 ~ 40) GHz (1.20 ~ 1.30) 5 Hz ~ 0.1 MHz 0.1 MHz ~ 2 GHz (2 ~ 26.5) GHz (26.5 ~ 40) GHz (1.30 ~ 1.50) 5 Hz ~ 0.1 MHz 0.1 MHz ~ 2 GHz (2 ~ 8) GHz (8 ~ 26.5) GHz (26.5 ~ 40) GHz (1.50 ~ 1.75) 5 Hz ~ 2 GHz (2 ~ 8) GHz (8 ~ 26.5) GHz (26.5 ~ 40) GHz (1.75 ~ 2.00) 5 Hz ~ 0.1 MHz 0.1 MHz ~ 2 GHz (2 ~ 8) GHz (8 ~ 26.5) GHz (26.5 ~ 40) GHz (2.00 ~ 2.50) 5 Hz ~ 0.1 MHz 0.1 MHz ~ 2 GHz (2 ~ 8) GHz (8 ~ 26.5) GHz (26.5 ~ 40) GHz (2.50 ~ 3.00) 5 Hz ~ 0.1 MHz 0.1 MHz ~ 2 GHz (2 ~ 8) GHz (8 ~ 26.5) GHz (26.5 ~ 40) GHz	0.011 0.013 0.020 0.038 0.014 0.015 0.023 0.043 0.018 0.019 0.030 0.031 0.054 0.025 0.041 0.043 0.070 0.033 0.032 0.054 0.058 0.090 0.052 0.048 0.087 0.10 0.14 0.075 0.067 0.13 0.14 0.20	Network analyzer Calibration kit /KTICC-CI-40619

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Mobile communication test sets	40621			Frequency counter, Measuring receiver, Power meter, Frequency counter, DMM, Meter calibrator, Function generator, Audio analyzer, RF spectrum analyzer /KTICC-CI-40621
Frequency		10 Hz ~ 5 GHz (5 ~ 40) GHz	1.2×10^{-10} 1.3 Hz	
Output level		(0.1 ~ 1 000) MHz (20 ~ 10) dBm (10 ~ -10) dBm (-10 ~ -30) dBm (-30 ~ -40) dBm (-40 ~ -60) dBm (-60 ~ -70) dBm (-70 ~ -90) dBm (-90 ~ -110) dBm (-110 ~ -120) dBm (1 ~ 4) GHz (20 ~ 10) dBm (10 ~ -10) dBm (-10 ~ -30) dBm (-30 ~ -40) dBm (-40 ~ -60) dBm (-60 ~ -80) dBm (-80 ~ -100) dBm (-100 ~ -120) dBm (4 ~ 8) GHz (20 ~ 10) dBm (10 ~ 0) dBm (0 ~ -20) dBm (-20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -90) dBm (-90 ~ -100) dBm (-100 ~ -120) dBm (8 ~ 10) GHz (20 ~ 10) dBm (10 ~ 0) dBm (0 ~ -20) dBm (-20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -90) dBm (-90 ~ -110) dBm (-110 ~ -120) dBm	0.12 dB 0.11 dB 0.12 dB 0.13 dB 0.14 dB 0.15 dB 0.16 dB 0.17 dB 0.18 dB 0.13 dB 0.12 dB 0.13 dB 0.14 dB 0.15 dB 0.16 dB 0.17 dB 0.18 dB 0.19 dB 0.20 dB 0.14 dB 0.13 dB 0.14 dB 0.15 dB 0.16 dB 0.17 dB 0.18 dB 0.19 dB 0.20 dB 0.15 dB 0.14 dB 0.15 dB 0.16 dB 0.17 dB 0.18 dB 0.19 dB 0.20 dB 0.21 dB	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Mobile communication test sets	40621	(10 ~ 12) GHz		Frequency counter, Measuring receiver, Power meter, Frequency counter, DMM, Meter calibrator, Function generator, Audio analyzer, RF spectrum analyzer /KTICC-CI-40621
Output level		(20 ~ 10) dBm	0.16 dB	
		(10 ~ 0) dBm	0.15 dB	
		(0 ~ -20) dBm	0.16 dB	
		(-20 ~ -40) dBm	0.17 dB	
		(-40 ~ -50) dBm	0.18 dB	
		(-50 ~ -70) dBm	0.19 dB	
		(-70 ~ -90) dBm	0.20 dB	
		(-90 ~ -110) dBm	0.21 dB	
		(-110 ~ -120) dBm	0.22 dB	
		(12 ~ 18) GHz		
		(20 ~ 10) dBm	0.18 dB	
		(10 ~ 0) dBm	0.17 dB	
		(0 ~ -20) dBm	0.18 dB	
		(-20 ~ -40) dBm	0.19 dB	
		(-40 ~ -50) dBm	0.20 dB	
		(-50 ~ -70) dBm	0.21 dB	
		(-70 ~ -90) dBm	0.22 dB	
		(-90 ~ -100) dBm	0.23 dB	
		(-100 ~ -120) dBm	0.24 dB	
		(18 ~ 26.5) GHz		
		(20 ~ 10) dBm	0.25 dB	
		(10 ~ -10) dBm	0.24 dB	
		(-10 ~ -30) dBm	0.25 dB	
		(-30 ~ -40) dBm	0.26 dB	
		(-40 ~ -60) dBm	0.28 dB	
		(-60 ~ -80) dBm	0.29 dB	
		(-80 ~ -100) dBm	0.30 dB	
		(-100 ~ -120) dBm	0.31 dB	
		(26.5 ~ 40) GHz		
		(20 ~ -20) dBm	0.21 dB	
Frequency Modulation		Rate (0.01 ~ 100) kHz		
		(0 ~ 400) kHz	2.7×10^{-2}	
Amplitude Modulation		Rate (0.01 ~ 50) kHz		
		(0 ~ 99) %	2.7×10^{-2}	
Phase Modulation		Rate (0.05 ~ 100) kHz		
		(0 ~ 400) rad	4.2×10^{-2}	
Harmonics		10 MHz ~ 3 GHz	0.4 dB	
		(3 ~ 8) GHz	0.5 dB	
		(8 ~ 12) GHz	0.6 dB	
		(12 ~ 16) GHz	0.7 dB	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Mobile communication test sets Input Level	40621	(0.1 ~ 1 000) MHz (20 ~ 10) dBm (10 ~ 0) dBm (0 ~ -40) dBm (-40 ~ -60) dBm (-60 ~ -90) dBm (-90 ~ -110) dBm (-110 ~ -120) dBm (1 ~ 4) GHz (20 ~ -20) dBm (-20 ~ -40) dBm (-40 ~ -70) dBm (-70 ~ -100) dBm (-100 ~ -120) dBm (4 ~ 8) GHz (20 ~ -20) dBm (-20 ~ -40) dBm (-40 ~ -70) dBm (-70 ~ -100) dBm (-100 ~ -120) dBm (8 ~ 10) GHz (20 ~ -20) dBm (-20 ~ -50) dBm (-50 ~ -80) dBm (-80 ~ -110) dBm (-110 ~ -120) dBm (10 ~ 12) GHz (20 ~ 10) dBm (-20 ~ 0) dBm (0 ~ -40) dBm (-40 ~ -60) dBm (-60 ~ -90) dBm (-90 ~ -120) dBm (12 ~ 18) GHz (20 ~ -30) dBm (-30 ~ -50) dBm (-50 ~ -80) dBm (-80 ~ -100) dBm (-100 ~ -120) dBm (18 ~ 26.5) GHz (20 ~ -20) dBm (-20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -80) dBm (-80 ~ -110) dBm (-110 ~ -120) dBm (26.5 ~ 40) GHz (20 ~ -20) dBm	0.21 dB 0.20 dB 0.21 dB 0.22 dB 0.23 dB 0.24 dB 0.25 dB 0.21 dB 0.22 dB 0.23 dB 0.24 dB 0.25 dB 0.22 dB 0.23 dB 0.24 dB 0.25 dB 0.26 dB 0.27 dB 0.28 dB 0.29 dB 0.30 dB 0.31 dB 0.28 dB 0.27 dB 0.28 dB 0.29 dB 0.30 dB 0.31 dB 0.32 dB 0.33 dB 0.29 dB 0.30 dB 0.31 dB 0.32 dB 0.33 dB 0.37 dB 0.38 dB 0.39 dB 0.40 dB 0.41 dB 0.42 dB 0.50 dB	Frequency counter, Measuring receiver, Power meter, Frequency counter, DMM, Meter calibrator, Function generator, Audio analyzer, RF spectrum analyzer /KTICC-CI-40621

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Mobile communication test sets	40621	1 mV		Frequency counter, Measuring receiver, Power meter, Frequency counter, DMM, Meter calibrator, Function generator, Audio analyzer, RF spectrum analyzer /KTICC-CI-40621
Audio Output Level		40 Hz ~ 10 kHz (10 ~ 100) kHz	4.2×10^{-3} 2.1×10^{-2}	
		(1 ~ 10) mV		
		40 Hz ~ 10 kHz (10 ~ 100) kHz	5.4×10^{-4} 2.8×10^{-3}	
		(10 ~ 100) mV		
		40 Hz ~ 10 kHz (10 ~ 100) kHz	2.2×10^{-4} 1.1×10^{-3}	
		(0.1 ~ 10) V		
		(40 ~ 100) Hz (0.1 ~ 1) kHz (1 ~ 10) kHz (10 ~ 100) kHz	1.4×10^{-4} 1.2×10^{-4} 1.4×10^{-4} 7.3×10^{-4}	
		(10 ~ 20) V		
		(40 ~ 100) Hz (0.1 ~ 1) kHz (1 ~ 10) kHz (10 ~ 100) kHz	2.0×10^{-4} 1.8×10^{-4} 2.0×10^{-4} 1.5×10^{-3}	
AC Input Voltage		10 mV		
		(0.04 ~ 1) kHz (1 ~ 10) kHz (10 ~ 100) kHz (100 ~ 500) kHz (0.5 ~ 1) MHz	5.1×10^{-4} 5.0×10^{-4} 1.0×10^{-3} 3.5×10^{-3} 4.8×10^{-3}	
		(10 ~ 100) mV		
		(0.04 ~ 1) kHz (1 ~ 10) kHz (10 ~ 100) kHz (100 ~ 500) kHz (0.5 ~ 1) MHz	1.8×10^{-4} 1.7×10^{-4} 6.5×10^{-4} 1.7×10^{-3} 3.2×10^{-3}	
		(0.1 ~ 1) V		
		(0.04 ~ 1) kHz (1 ~ 10) kHz (10 ~ 100) kHz (100 ~ 500) kHz (0.5 ~ 1) MHz	1.3×10^{-4} 9.1×10^{-5} 1.6×10^{-4} 1.2×10^{-3} 2.0×10^{-3}	
		(1 ~ 10) V		
		(0.04 ~ 1) kHz (1 ~ 10) kHz (10 ~ 100) kHz (100 ~ 500) kHz (0.5 ~ 1) MHz	1.3×10^{-4} 8.9×10^{-5} 1.4×10^{-4} 1.2×10^{-3} 1.9×10^{-3}	
DC Input Voltage		0 mV	$0.01 \mu V$	
		(0 ~ 10) mV (10 ~ 100) mV (0.1 ~ 1) V (1 ~ 10) V	4.9×10^{-5} 1.3×10^{-5} 5.9×10^{-6} 4.0×10^{-6}	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Mobile communication test sets DC Output Voltage	40621	0 V (0 ~ 0.1) V (0.1 ~ 0.2) V (0.2 ~ 0.5) V (0.5 ~ 1) V (1 ~ 2) V (2 ~ 5) V (5 ~ 10) V	0.01 mV 6.3×10^{-5} 3.3×10^{-5} 1.6×10^{-5} 1.1×10^{-5} 3.2×10^{-5} 1.5×10^{-5} 1.0×10^{-5}	Frequency counter, Measuring receiver, Power meter, Frequency counter, DMM, Meter calibrator, Function generator, Audio analyzer, RF spectrum analyzer /KTICC-CI-40621
Modulation meters Frequency Modulation Amplitude Modulation Phase Modulation Frequency Tuned RF Level ILS/VOR Analyzer Frequency(VOR/ILS) Amplitude Modulation(VOR/ILS) Input level(VOR/ILS)	40622	Rate (0.01 ~ 100) kHz (0 ~ 400) kHz Rate (0.01 ~ 50) kHz (0 ~ 99) % Rate (0.05 ~ 100) kHz (0 ~ 400) rad 10 Hz ~ 5 GHz (5 ~ 26.5) GHz (0 ~ -10) dB (-10 ~ -120) dB (70 ~ 350) MHz Localizer(108 ~ 112) MHz : (0 ~ 20) % Glideslope(320 ~ 340) MHz : (20 ~ 40) % Marker Beacon(74.7 ~ 75.3) MHz : (40 ~ 95) % VOR(108 ~ 118) MHz : (0 ~ 30) % Localizer(108 ~ 112) MHz : (10 ~ -10) dBm (-10 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -90) dBm (-90 ~ -110) dBm (-110 ~ -120) dBm	2.7×10^{-2} 2.7×10^{-2} 4.2×10^{-2} 1.2×10^{-10} 1.3 Hz 0.04 dB 0.05 dB 6.7 Hz 0.93 % 1.9 % 4.6 % 1.4 % 0.14 dB 0.15 dB 0.16 dB 0.17 dB 0.18 dB 0.19 dB 0.20 dB	Measuring receiver Frequency counter RF Signal Generator /KTICC-CI-40622

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Modulation meters Input level(VOR/ILS)	40622	Glideslope(320 ~ 340) MHz : (10 ~ -10) dBm (-10 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -90) dBm (-90 ~ -110) dBm (-110 ~ -120) dBm	0.14 dB 0.15 dB 0.16 dB 0.17 dB 0.18 dB 0.19 dB 0.20 dB	Measuring receiver Frequency counter RF Signal Generator /KTICC-CI-40622
DDM(VOR/ILS)		Localizer(108 ~ 112) MHz : 0 $\pm(0 \sim 0.155)$ Glideslope(320 ~ 340) MHz : 0 $\pm(0 \sim 0.175)$	0.000 7 0.003 2 0.001 1 0.003 2	
SDM(VOR/ILS)		Localizer(108 ~ 112) MHz : 40 % Glideslope(320 ~ 340) MHz : 80 %	0.34 % 1.5 %	
Azimuth(VOR/ILS)		VOR(108 ~ 118) MHz : (0 ~ 360)°	0.07°	
Network analyzers Output Frequency	40623	10 Hz ~ 5 GHz (5 ~ 46) GHz	1.2×10^{-10} 1.3 Hz	Frequency Standard, Frequency Counter, Power Sensor, Power Meter, Sensor Module, Measuring Receiver, Calibration Kit, Attenuator /KTICC-CI-40623
Output Level		9 kHz ~ 0.1 MHz (20 ~ -40) dBm (0.1 ~ 1 000) MHz (20 ~ 10) dBm (10 ~ -10) dBm (-10 ~ -30) dBm (-30 ~ -40) dBm (-40 ~ -60) dBm (-60 ~ -70) dBm (-70 ~ -90) dBm (-90 ~ -110) dBm (1 ~ 4) GHz (20 ~ 10) dBm (10 ~ -10) dBm (-10 ~ -30) dBm (-30 ~ -40) dBm (-40 ~ -60) dBm (-60 ~ -80) dBm (-80 ~ -100) dBm (-100 ~ -110) dBm	0.07 dB 0.12 dB 0.11 dB 0.12 dB 0.13 dB 0.14 dB 0.15 dB 0.16 dB 0.17 dB 0.13 dB 0.12 dB 0.13 dB 0.14 dB 0.15 dB 0.16 dB 0.17 dB 0.18 dB	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Network analyzers Output Level	40623	(4 ~ 8) GHz (20 ~ 10) dBm (10 ~ 0) dBm (0 ~ -20) dBm (-20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -90) dBm (-90 ~ -100) dBm (-100 ~ -110) dBm (8 ~ 10) GHz (20 ~ 10) dBm (10 ~ 0) dBm (0 ~ -20) dBm (-20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -90) dBm (-90 ~ -110) dBm (12 ~ 18) GHz (20 ~ 10) dBm (10 ~ 0) dBm (0 ~ -20) dBm (-20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -90) dBm (-90 ~ -100) dBm (-100 ~ -110) dBm (18 ~ 26.5) GHz (20 ~ 10) dBm (10 ~ -10) dBm (-10 ~ -30) dBm (-30 ~ -40) dBm (-40 ~ -60) dBm (-60 ~ -80) dBm (-80 ~ -100) dBm (-100 ~ -110) dBm (26.5 ~ 40) GHz (20 ~ -30) dBm (40 ~ 50) GHz (20 ~ -30) dBm	0.14 dB 0.13 dB 0.14 dB 0.15 dB 0.16 dB 0.17 dB 0.18 dB 0.19 dB 0.20 dB 0.15 dB 0.14 dB 0.15 dB 0.16 dB 0.17 dB 0.18 dB 0.19 dB 0.20 dB 0.18 dB 0.17 dB 0.18 dB 0.19 dB 0.20 dB 0.21 dB 0.22 dB 0.23 dB 0.24 dB 0.25 dB 0.24 dB 0.25 dB 0.26 dB 0.28 dB 0.29 dB 0.30 dB 0.31 dB 0.21 dB 0.25 dB	Frequency Standard, Frequency Counter, Power Sensor, Power Meter, Sensor Module, Measuring Receiver, Calibration Kit, Attenuator /KTICC-CI-40623

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Network analyzers Output Level Linearity	40623	9 kHz ~ 0.1 MHz (20 ~ -40) dBm (0.1 ~ 1 000) MHz (20 ~ -30) dBm (-30 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -100) dBm (-100 ~ -110) dBm (1 ~ 10) GHz (20 ~ -40) dBm (-40 ~ -60) dBm (-60 ~ -90) dBm (-90 ~ -110) dBm (10 ~ 18) GHz (20 ~ -60) dBm (-60 ~ -90) dBm (-90 ~ -110) dBm (18 ~ 22) GHz (20 ~ 10) dBm (10 ~ -10) dBm (-10 ~ -70) dBm (-70 ~ -110) dBm (22 ~ 26.5) GHz (20 ~ 10) dBm (10 ~ -10) dBm (-10 ~ -50) dBm (-50 ~ -90) dBm (-90 ~ -110) dBm (26.5 ~ 40) GHz (20 ~ -30) dBm (40 ~ 50) GHz (20 ~ -30) dBm	0.07 dB 0.06 dB 0.07 dB 0.08 dB 0.09 dB 0.10 dB 0.08 dB 0.09 dB 0.10 dB 0.11 dB 0.10 dB 0.11 dB 0.12 dB 0.15 dB 0.14 dB 0.15 dB 0.15 dB 0.16 dB 0.17 dB 0.21 dB 0.25 dB	Frequency Standard, Frequency Counter, Power Sensor, Power Meter, Sensor Module, Measuring Receiver, Calibration Kit, Attenuator /KTICC-CI-40623

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Network analyzers Frequency Response	40623	(10 ~ -10) dBm 9 kHz ~ 0.1 MHz (0.1 ~ 1 000) MHz (1 ~ 4) GHz (4 ~ 8) GHz (8 ~ 12) GHz (12 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz (40 ~ 50) GHz	0.18 dB 0.20 dB 0.21 dB 0.22 dB 0.27 dB 0.29 dB 0.37 dB 0.50 dB 0.52 dB	Frequency Standard, Frequency Counter, Power Sensor, Power Meter, Sensor Module, Measuring Receiver, Calibration Kit, Attenuator /KTICC-CI-40623
Attenuation		(0.05 ~ 3) GHz (0 ~ -80) dB (-80 ~ -90) dB (-90 ~ -100) dB (3 ~ 8) GHz (0 ~ -40) dB (-40 ~ -70) dB (-70 ~ -80) dB (-80 ~ -90) dB (-90 ~ -100) dB (8 ~ 10) GHz (0 ~ -40) dB (-40 ~ -70) dB (-70 ~ -80) dB (-80 ~ -90) dB (-90 ~ -100) dB (10 ~ 15) GHz (0 ~ -20) dB (-20 ~ -60) dB (-60 ~ -80) dB (-80 ~ -90) dB (-90 ~ -100) dB (15 ~ 18) GHz (0 ~ -70) dB (-70 ~ -80) dB (-80 ~ -90) dB (-90 ~ -100) dB	0.06 dB 0.07 dB 0.08 dB 0.05 dB 0.06 dB 0.07 dB 0.08 dB 0.09 dB 0.05 dB 0.06 dB 0.07 dB 0.09 dB 0.10 dB 0.06 dB 0.07 dB 0.08 dB 0.10 dB 0.11 dB 0.09 dB 0.10 dB 0.11 dB 0.12 dB	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Noise figure meters	40624			DMM, Frequency counter Attenuator, Noise source Network analyzer /KTICC-CI-40624
Reference Frequency		10 MHz	1.2 mHz	
Frequency		10 MHz ~ 26.5 GHz	0.58 kHz	
Noise Figure Range		(0 ~ -5) dB	0.040 dB	
		(-5 ~ -11) dB	0.041 dB	
		(-11 ~ -30) dB	0.042 dB	
		(-30 ~ -40) dB	0.044 dB	
		(-40 ~ -60) dB	0.045 dB	
Noise Figure		Noise source (6 dB)		
		(0.01 ~ 1) GHz	0.52 dB	
		(1 ~ 7) GHz	0.51 dB	
		(7 ~ 18) GHz	0.53 dB	
		Noise source (15 dB)		
		(0.01 ~ 1) GHz	0.52 dB	
		(1 ~ 3) GHz	0.50 dB	
		(3 ~ 7) GHz	0.51 dB	
		(7 ~ 9) GHz	0.55 dB	
		(9 ~ 17) GHz	0.56 dB	
		(17 ~ 18) GHz	0.57 dB	
		(18 ~ 26.5) GHz	0.74 dB	
VSWR		(1 ~ 3)		
		0.1 MHz	0.008	
		(0.1 ~ 2 000) MHz	0.009	
		(2 ~ 26.5) GHz	0.015	
Noise Source Voltage		(0 ~ 28) V	0.25 mV	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Noise generators	40625			Frequency counter, Power sensor, Power meter RF spectrum analyzer Sensor module, Measuring receiver /KTICC-CI-40625
Frequency		(0.1 ~ 5 000) MHz (5 ~ 18) GHz	1.2×10^{-10} 1.3 Hz	
Output Level		(0.1 ~ 1 000) MHz (20 ~ 10) dBm (10 ~ -10) dBm (-10 ~ -30) dBm (-30 ~ -40) dBm (-40 ~ -60) dBm (-60 ~ -70) dBm (-70 ~ -90) dBm (-90 ~ -110) dBm (-110 ~ -120) dBm (1 ~ 4) GHz (20 ~ 10) dBm (10 ~ -10) dBm (-10 ~ -30) dBm (-30 ~ -40) dBm (-40 ~ -60) dBm (-60 ~ -80) dBm (-80 ~ -100) dBm (-100 ~ -120) dBm (4 ~ 8) GHz (20 ~ 10) dBm (10 ~ 0) dBm (0 ~ -20) dBm (-20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -90) dBm (-90 ~ -100) dBm (-100 ~ -120) dBm (8 ~ 10) GHz (20 ~ 10) dBm (10 ~ 0) dBm (0 ~ -20) dBm (-20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -90) dBm (-90 ~ -110) dBm (-110 ~ -120) dBm	0.12 dB 0.11 dB 0.12 dB 0.13 dB 0.14 dB 0.15 dB 0.16 dB 0.17 dB 0.18 dB 0.13 dB 0.12 dB 0.13 dB 0.14 dB 0.15 dB 0.16 dB 0.17 dB 0.18 dB 0.19 dB 0.20 dB 0.14 dB 0.13 dB 0.14 dB 0.15 dB 0.16 dB 0.17 dB 0.18 dB 0.19 dB 0.20 dB 0.15 dB 0.14 dB 0.15 dB 0.16 dB 0.17 dB 0.18 dB 0.19 dB 0.20 dB 0.21 dB	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Noise generators Output Level	40625	(10 ~ 12) GHz (20 ~ 10) dBm (10 ~ 0) dBm (0 ~ -20) dBm (-20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -90) dBm (-90 ~ -110) dBm (-110 ~ -120) dBm (12 ~ 18) GHz (20 ~ 10) dBm (10 ~ 0) dBm (0 ~ -20) dBm (-20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -90) dBm (-90 ~ -100) dBm (-100 ~ -120) dBm	0.16 dB 0.15 dB 0.16 dB 0.17 dB 0.18 dB 0.19 dB 0.20 dB 0.21 dB 0.22 dB 0.18 dB 0.17 dB 0.18 dB 0.19 dB 0.20 dB 0.21 dB 0.22 dB 0.23 dB 0.24 dB	Frequency counter, Power sensor, Power meter RF spectrum analyzer Sensor module, Measuring receiver /KTICC-CI-40625
Noise impulse simulators Output Voltage Pulse Width Rise Time Repetition period	40626	(±) 0 kV (0 ~ 4) kV 10 ns (10 ~ 50) ns (50 ~ 100) ns (100 ~ 200) ns (200 ~ 250) ns (250 ~ 400) ns (400 ~ 500) ns (500 ~ 800) ns (800 ~ 1 000) ns 0.5 ns (0.5 ~ 1) ns (1 ~ 5) ns (10 ~ 1 000) ms	0.1 V 1.3×10^{-2} 3.6×10^{-3} 4.4×10^{-3} 4.2×10^{-3} 6.0×10^{-3} 5.3×10^{-3} 4.6×10^{-3} 3.9×10^{-3} 4.2×10^{-3} 3.4×10^{-3} 3.2×10^{-1} 9.1×10^{-2} 7.6×10^{-3} 3.4×10^{-3}	Oscilloscope, Attenuator High voltage probe /KTICC-CI-40626
Coaxial noise sources ENR Reflection coefficient	40628	(0.01 ~ 5) GHz (5 ~ 6) GHz (6 ~ 18) GHz (18 ~ 26.5) GHz (0 ~ 0.5) 0.01 GHz (0.01 ~ 2) GHz (2 ~ 26.5) GHz	0.47 dB 0.48 dB 0.62 dB 0.92 dB 0.004 1 0.004 6 0.007 3	Coaxial noise sources, Noise figure analyzer /KTICC-CI-40628

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF phase meters Phase	40631	(0 ~ 360)° (0.05 ~ 1) GHz (1 ~ 7) GHz (7 ~ 18) GHz	1.5° 3.8° 6.7°	RF signal generator Calibration kit Network analyzer /KTICC-CI-40631
RF power meters Instrument Accuracy Power Reference Accuracy Reference Frequency Calibration Factor Power linearity RF High Power	40635	0 μW 3 μW ~ 100 mW 1 mW (0.01 ~ 1) GHz (88 ~ 100) % (0.01 ~ 1) GHz (0 ~ -60) dBm (-60 ~ -80) dBm (-80 ~ -100) dBm (1 ~ 10) GHz (0 ~ -40) dBm (-40 ~ -70) dBm (-70 ~ -80) dBm (-80 ~ -90) dBm (-90 ~ -100) dBm (10 ~ 18) GHz (0 ~ -10) dBm (-10 ~ -40) dBm (-40 ~ -80) dBm (-80 ~ -100) dBm (0.08 ~ 1) GHz 0 W (0 ~ 100) W	0.001 μW 2.9×10^{-3} 5.8 μW 1.2×10^{-10} 3.0×10^{-3} 0.07 dB 0.08 dB 0.09 dB 0.08 dB 0.09 dB 0.10 dB 0.11 dB 0.12 dB 0.12 dB 0.11 dB 0.12 dB 0.14 dB 1 mW 2.2×10^{-2}	Range calibrator, Thermistor mount, Power meter, Frequency counter Fixed attenuator, Power sensor, RF amplifier, Signal generator /KTICC-CI-40635
Diode power sensors Calibration Factor	40636	0.1 μW ~ 1 mW 9 kHz ~ 10 MHz (0.01 ~ 1) GHz (1 ~ 10) GHz (10 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz	1.2×10^{-2} 1.1×10^{-2} 1.3×10^{-2} 1.5×10^{-2} 2.6×10^{-2} 3.6×10^{-2}	Thermistor mount, Power meter, Network analyzer /KTICC-CI-40636
Thermocouple power sensors Calibration Factor	40637	10 μW ~ 10 mW 9 kHz ~ 10 MHz (0.01 ~ 1) GHz (1 ~ 10) GHz (10 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz	1.2×10^{-2} 1.1×10^{-2} 1.3×10^{-2} 1.5×10^{-2} 2.6×10^{-2} 3.6×10^{-2}	Thermistor mount, Power meter, Network analyzer /KTICC-CI-40637

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.	
Pulse generators	40638			Frequency counter, DMM, Audio analyzer, Oscilloscope Measuring receiver Sensor module /KTICC-CI-40638	
Frequency		10 MHz	1.4×10^{-10}		
Period		0.2 ns ~ 1 s	1.4×10^{-10}		
Output Voltage		1 mV			
		20 Hz ~ 10 kHz	4.2×10^{-3}		
		(10 ~ 100) kHz	2.1×10^{-2}		
		100 kHz ~ 1 MHz	4.5×10^{-2}		
		(1 ~ 10) mV			
		20 Hz ~ 10 kHz	5.4×10^{-4}		
		(10 ~ 100) kHz	2.8×10^{-3}		
		100 kHz ~ 1 MHz	2.3×10^{-2}		
		(10 ~ 100) mV			
		20 Hz ~ 10 kHz	2.2×10^{-4}		
		(10 ~ 100) kHz	1.1×10^{-3}		
		100 kHz ~ 1 MHz	8.1×10^{-3}		
		(0.1 ~ 10) V			
		(40 ~ 100) Hz	1.4×10^{-4}		
		(0.1 ~ 1) kHz	1.2×10^{-4}		
		(1 ~ 10) kHz	1.4×10^{-4}		
		(10 ~ 100) kHz	7.3×10^{-4}		
		100 kHz ~ 1 MHz	8.1×10^{-3}		
		(10 ~ 20) V			
		(40 ~ 100) Hz	2.0×10^{-4}		
		(0.1 ~ 1) kHz	1.8×10^{-4}		
		(1 ~ 10) kHz	2.0×10^{-4}		
		(10 ~ 100) kHz	1.5×10^{-3}		
		100 kHz ~ 1 MHz	8.6×10^{-3}		
		(20 ~ 100) V			
		(40 ~ 100) Hz	1.4×10^{-4}		
		(0.1 ~ 1) kHz	1.2×10^{-4}		
		(1 ~ 10) kHz	1.4×10^{-4}		
		(10 ~ 100) kHz	7.4×10^{-4}		
		Output Level	(0.1 ~ 1 000) MHz		
			(20 ~ 10) dBm		0.12 dB
			(10 ~ -10) dBm		0.11 dB
			(-10 ~ -30) dBm		0.12 dB
			(-30 ~ -40) dBm		0.13 dB
			(-40 ~ -60) dBm		0.14 dB
			(-60 ~ -70) dBm		0.15 dB
(-70 ~ -90) dBm			0.16 dB		
(-90 ~ -110) dBm			0.17 dB		
(-110 ~ -120) dBm			0.18 dB		
Level Flatness		(0.1 ~ 1 000) MHz	0.11 dB		
		(1 ~ 4) GHz	0.12 dB		
		(4 ~ 8) GHz	0.13 dB		
		(8 ~ 10) GHz	0.14 dB		
		(10 ~ 12) GHz	0.15 dB		
		(12 ~ 18) GHz	0.17 dB		

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Pulse generators	40638			Frequency counter, DMM, Audio analyzer, Oscilloscope Measuring receiver Sensor module /KTICC-CI-40638
Pulse width		0.4 ns (0.4 ~ 1) ns (1 ~ 10) ns 10 ns ~ 1 ms 1 ms ~ 1 s	4.8×10^{-1} 9.0×10^{-2} 3.6×10^{-3} 3.4×10^{-3} 1.3×10^{-3}	
Pulse Time		0.4 ns (0.4 ~ 1) ns (1 ~ 10) ns 10 ns ~ 1 ms 1 ms ~ 1 s	4.8×10^{-1} 9.0×10^{-2} 3.6×10^{-3} 3.4×10^{-3} 1.3×10^{-3}	
Double Pulse		0.4 ns (0.4 ~ 1) ns (1 ~ 10) ns 10 ns ~ 1 ms 1 ms ~ 1 s	4.8×10^{-1} 9.0×10^{-2} 3.6×10^{-3} 3.4×10^{-3} 1.3×10^{-3}	
Pulse Delay		0.4 ns (0.4 ~ 1) ns (1 ~ 10) ns 10 ns ~ 1 ms 1 ms ~ 1 s	4.8×10^{-1} 9.0×10^{-2} 3.6×10^{-3} 3.4×10^{-3} 1.3×10^{-3}	
Transition		0.4 ns (0.4 ~ 1) ns (1 ~ 10) ns 10 ns ~ 1 ms 1 ms ~ 1 s	4.8×10^{-1} 9.0×10^{-2} 3.6×10^{-3} 3.4×10^{-3} 1.3×10^{-3}	
Duty cycle		(1 ~ 99) %	0.006 3 %	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Radar test sets	40639			Frequency Standard, Frequency Counter, Power Sensor, Power Meter, Signal Generator RF amplifiers attenuators /KTICC-CI-40639
Frequency		10 Hz ~ 5 GHz (5 ~ 10) GHz	1.2×10^{-10} 1.3 Hz	
Output Level		(0.1 ~ 1 000) MHz (20 ~ 10) dBm (10 ~ -10) dBm (-10 ~ -30) dBm (-30 ~ -40) dBm (-40 ~ -60) dBm (-60 ~ -70) dBm (-70 ~ -90) dBm (-90 ~ -110) dBm (-110 ~ -120) dBm	0.12 dB 0.11 dB 0.12 dB 0.13 dB 0.14 dB 0.15 dB 0.16 dB 0.17 dB 0.18 dB	
Frequency Modulation		Rate (0.01 ~ 100) kHz (0 ~ 400) kHz	2.7×10^{-2}	
Input level		(0.1 ~ 1 000) MHz (20 ~ 10) dBm (10 ~ 0) dBm (0 ~ -40) dBm (-40 ~ -60) dBm (-60 ~ -90) dBm (-90 ~ -110) dBm (-110 ~ -120) dBm	0.21 dB 0.20 dB 0.21 dB 0.22 dB 0.23 dB 0.24 dB 0.25 dB	
RF High Power		(0.08 ~ 1) GHz 0 W (0 ~ 100) W	1 mW 2.8×10^{-2}	
Pulse Time		0.4 ns (0.4 ~ 1) ns (1 ~ 10) ns 10 ns ~ 1 ms 1 ms ~ 1 s	4.8×10^{-1} 9.0×10^{-2} 3.6×10^{-3} 3.4×10^{-3} 1.3×10^{-3}	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF signal generators	40640	9 kHz ~ 5 GHz	1.2×10^{-10}	Frequency Counter, Power Meter, Power Sensor RF Spectrum Analyzer Sensor Module, Measuring Receiver Audio Analyzer /KTICC-CI-40640
Output Frequency		(5 ~ 46) GHz	1.3 Hz	
Output Level		9 kHz ~ 2.5 GHz		
		(30 ~ 57) dBm	0.25 dB	
		2.5 GHz ~ 18 GHz		
		(30 ~ 44) dBm	0.30 dB	
		9 kHz ~ 0.1 MHz		
		(20 ~ -40) dBm	0.07 dB	
		(0.1 ~ 1 000) MHz		
		(20 ~ 10) dBm	0.12 dB	
		(10 ~ -10) dBm	0.11 dB	
		(-10 ~ -30) dBm	0.12 dB	
		(-30 ~ -40) dBm	0.13 dB	
		(-40 ~ -60) dBm	0.14 dB	
		(-60 ~ -70) dBm	0.15 dB	
		(-70 ~ -90) dBm	0.16 dB	
		(-90 ~ -110) dBm	0.17 dB	
		(-110 ~ -120) dBm	0.18 dB	
		(1 ~ 4) GHz		
		(20 ~ 10) dBm	0.13 dB	
		(10 ~ -10) dBm	0.12 dB	
		(-10 ~ -30) dBm	0.13 dB	
		(-30 ~ -40) dBm	0.14 dB	
		(-40 ~ -60) dBm	0.15 dB	
		(-60 ~ -80) dBm	0.16 dB	
		(-80 ~ -100) dBm	0.17 dB	
		(-100 ~ -120) dBm	0.18 dB	
		(4 ~ 8) GHz		
		(20 ~ 10) dBm	0.14 dB	
		(10 ~ 0) dBm	0.13 dB	
		(0 ~ -20) dBm	0.14 dB	
		(-20 ~ -40) dBm	0.15 dB	
		(-40 ~ -50) dBm	0.16 dB	
		(-50 ~ -70) dBm	0.17 dB	
		(-70 ~ -90) dBm	0.18 dB	
		(-90 ~ -100) dBm	0.19 dB	
		(-100 ~ -120) dBm	0.20 dB	
		(8 ~ 10) GHz		
		(20 ~ 10) dBm	0.15 dB	
		(10 ~ 0) dBm	0.14 dB	
		(0 ~ -20) dBm	0.15 dB	
		(-20 ~ -40) dBm	0.16 dB	
		(-40 ~ -50) dBm	0.17 dB	
		(-50 ~ -70) dBm	0.18 dB	
		(-70 ~ -90) dBm	0.19 dB	
		(-90 ~ -110) dBm	0.20 dB	
		(-110 ~ -120) dBm	0.21 dB	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF signal generators Output Level	40640	(10 ~ 12) GHz (20 ~ 10) dBm (10 ~ 0) dBm (0 ~ -20) dBm (-20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -90) dBm (-90 ~ -110) dBm (-110 ~ -120) dBm (12 ~ 18) GHz (20 ~ 10) dBm (10 ~ 0) dBm (0 ~ -20) dBm (-20 ~ -40) dBm (-40 ~ -50) dBm (-50 ~ -70) dBm (-70 ~ -90) dBm (-90 ~ -100) dBm (-100 ~ -120) dBm (18 ~ 26.5) GHz (20 ~ 10) dBm (10 ~ -10) dBm (-10 ~ -30) dBm (-30 ~ -40) dBm (-40 ~ -60) dBm (-60 ~ -80) dBm (-80 ~ -100) dBm (-100 ~ -120) dBm (26.5 ~ 40) GHz (20 ~ -20) dBm (40 ~ 50) GHz (20 ~ -30) dBm	0.16 dB 0.15 dB 0.16 dB 0.17 dB 0.18 dB 0.19 dB 0.20 dB 0.21 dB 0.22 dB 0.18 dB 0.17 dB 0.18 dB 0.19 dB 0.20 dB 0.21 dB 0.22 dB 0.23 dB 0.24 dB 0.25 dB 0.24 dB 0.25 dB 0.26 dB 0.28 dB 0.29 dB 0.30 dB 0.31 dB 0.21 dB 0.25 dB	Frequency Counter, Power Meter, Power Sensor RF Spectrum Analyzer Sensor Module, Measuring Receiver Audio Analyzer /KTICC-CI-40640
Frequency Modulation		Rate (0.01 ~ 100) kHz (0 ~ 400) kHz	2.7×10^{-2}	
Amplitude Modulation		Rate (0.01 ~ 50) kHz (0 ~ 99) %	2.7×10^{-2}	
Phase Modulation		Rate (0.05 ~ 100) kHz (0 ~ 400) rad	4.2×10^{-2}	
Modulation distortion		FM, PM 20 Hz ~ 100 kHz AM 20 Hz ~ 100 kHz Depth (5 ~ 50) % Depth (50 ~ 95) %	0.12 % 0.35 % 0.69 %	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF signal generators Modulation Rate Harmonics	40640	10 Hz ~ 100 kHz 10 MHz ~ 3 GHz (3 ~ 8) GHz (8 ~ 12) GHz (12 ~ 20) GHz (20 ~ 26) GHz	6.8×10^{-10} 0.4 dB 0.5 dB 0.6 dB 0.8 dB 1.0 dB	Frequency Counter, Power Meter, Power Sensor RF Spectrum Analyzer Sensor Module, Measuring Receiver Audio Analyzer /KTICC-CI-40640
RF spectrum analyzers Reference Frequency Input Frequency Calibrator Output Input Amplitude Scale fidelity Scale fidelity Switching Frequency Response Absolute Amplitude Frequency Span Resolution Bandwidth Accuracy RBW Selectivity RBW Switching Accuracy	40641	(0.1 ~ 5 000) MHz (5 ~ 26.5) GHz 3 Hz ~ 5 GHz (5 ~ 40) GHz (0.1 MHz ~ 1 GHz) (10 ~ -10) dBm (-10 ~ -30) dBm (0 ~ -30) dB (-30 ~ -50) dB (-50 ~ -70) dB (-70 ~ -100) dB (-100 ~ -120) dB (0 ~ -30) dB (-30 ~ -50) dB (-50 ~ -70) dB (-70 ~ -100) dB (0.1 ~ 10) dB 10 Hz ~ 9 kHz 9 kHz ~ 1 GHz (1 ~ 4) GHz (4 ~ 8) GHz (8 ~ 12) GHz (12 ~ 18) GHz (18 ~ 26.5) GHz (26.5 ~ 40) GHz (0.1 ~ 1 000) MHz (0 ~ -10) dBm (-10 ~ -40) dBm (-40 ~ -50) dBm 1 kHz ~ 26.5 GHz 10 Hz ~ 10 MHz 10 Hz ~ 10 MHz 10 Hz ~ 10 MHz	1.2×10^{-10} 1.3 Hz 1.2×10^{-10} 1.3 Hz 0.11 dB 0.12 dB 0.06 dB 0.07 dB 0.08 dB 0.09 dB 0.10 dB 0.06 dB 0.07 dB 0.08 dB 0.09 dB 0.051 dB 0.18 dB 0.20 dB 0.21 dB 0.22 dB 0.27 dB 0.29 dB 0.37 dB 0.53 dB 0.20 dB 0.21 dB 0.22 dB 1.3×10^{-4} 1.1×10^{-3} 0.2×10^{-2} 0.06 dB	Frequency standard, Frequency counter, Power sensor, Power meter, Sensor module, Attenuator, Measuring receiver, Signal generator /KTICC-CI-40641

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.			
RF spectrum analyzers	40641			Frequency standard, Frequency counter, Power sensor, Power meter, Sensor module, Attenuator, Measuring receiver, Signal generator /KTICC-CI-40641			
Input Attenuator and Switching		(0 ~ -30) dB (-30 ~ -50) dB (-50 ~ -70) dB (-70 ~ -100) dB	0.06 dB 0.07 dB 0.08 dB 0.09 dB				
Noise Sidebands		0.1 kHz ~ 10 MHz	0.15 dB				
Display average noise level		30 Hz ~ 40 GHz	0.27 dB				
Surge generators	40643			Oscilloscope, High voltage probe, Current sensor /KTICC-CI-40643			
Output Voltage		(±) 0 kV (0 ~ 20) kV	0.1 V 1.3×10^{-2}				
Front Time (Rise Time)		0.1 μs (0.1 ~ 0.5) μs (0.5 ~ 1.2) μs (1.2 ~ 5) μs (5 ~ 10) μs (10 ~ 30) μs	1.4×10^{-2} 1.1×10^{-2} 1.2×10^{-2} 1.1×10^{-2} 1.4×10^{-2} 9.3×10^{-3}				
		Duration	10 μs (10 ~ 50) μs (50 ~ 700) μs (700 ~ 1 000) μs (1 000 ~ 3 000) μs (3 ~ 10) ms (10 ~ 50) ms (50 ~ 1 000) ms (1 000 ~ 3 000) ms		3.4×10^{-3} 3.8×10^{-3} 3.1×10^{-3} 3.4×10^{-3} 4.8×10^{-3} 3.4×10^{-3} 3.8×10^{-3} 3.4×10^{-3} 4.8×10^{-3}		
			Frequency (Ring Wave)		(1 ~ 100) kHz (100 ~ 200) kHz 200 kHz ~ 100 MHz	1.1×10^{-3} 1.0×10^{-3} 1.4×10^{-3}	
					Output Current	(±) (5.0 ~ 2 500) A (2 500 ~ 3 000) A	1.7×10^{-2} 1.8×10^{-2}
						Front Time (Rise Time)	1 μs (1 ~ 5) μs (5 ~ 10) μs
			Duration				10 μs (10 ~ 20) μs (20 ~ 100) μs (100 ~ 320) μs (320 ~ 400) μs

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
SWR meters	40644	(1.00 ~ 1.05)		Standard mismatch, Power sensor, Power meter, Spectrum analyzer, Frequency standard, Frequency counter /KTICC-CI-40644
VSWR		(0.05 ~ 2) GHz	0.059	
		(2 ~ 7) GHz	0.062	
		(7 ~ 9) GHz	0.063	
		(9 ~ 11) GHz	0.062	
		(11 ~ 12) GHz	0.063	
		(12 ~ 18) GHz	0.095	
		(1.05 ~ 1.20)		
		(0.05 ~ 1) GHz	0.071	
		(1 ~ 8) GHz	0.074	
		(8 ~ 12) GHz	0.085	
		(12 ~ 18) GHz	0.12	
		(1.20 ~ 1.50)		
		(0.05 ~ 1) GHz	0.095	
		(1 ~ 8) GHz	0.10	
		(8 ~ 12) GHz	0.11	
		(12 ~ 18) GHz	0.20	
		(1.50 ~ 2.00)		
		(0.05 ~ 1) GHz	0.14	
		(1 ~ 4) GHz	0.15	
		(4 ~ 12) GHz	0.16	
		(12 ~ 17) GHz	0.26	
		(17 ~ 18) GHz	0.27	
Output Power		(0.1 ~ 1 000) MHz		
		(10 ~ -10) dBm	0.11 dB	
		(-10 ~ -30) dBm	0.12 dB	
		(1 ~ 4) GHz		
		(10 ~ -10) dBm	0.12 dB	
		(-10 ~ -30) dBm	0.13 dB	
		(4 ~ 8) GHz		
		(10 ~ 0) dBm	0.14 dB	
		0 dBm	0.13 dB	
		(0 ~ -20) dBm	0.14 dB	
		(-20 ~ -30) dBm	0.15 dB	
		(8 ~ 10) GHz		
		(10 ~ 0) dBm	0.15 dB	
		0 dBm	0.14 dB	
		(0 ~ -20) dBm	0.15 dB	
		(-20 ~ -30) dBm	0.16 dB	
		(10 ~ 12) GHz		
		(10 ~ 0) dBm	0.16 dB	
		0 dBm	0.15 dB	
		(0 ~ -20) dBm	0.16 dB	
		(-20 ~ -30) dBm	0.17 dB	
		(12 ~ 18) GHz		
		(10 ~ 0) dBm	0.18 dB	
		0 dBm	0.17 dB	
		(0 ~ -20) dBm	0.18 dB	
		(-20 ~ -30) dBm	0.19 dB	
Frequency		(0.1 ~ 5 000) MHz	1.2×10^{-10}	
		(5 ~ 18) GHz	1.3 Hz	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
SWR meters Loss	40644	(1 ~ 3) dB (0.05 ~ 1) GHz (1 ~ 3) GHz (3 ~ 8) GHz (8 ~ 10) GHz (10 ~ 15) GHz (15 ~ 18) GHz (3 ~ 6) dB (0.05 ~ 1) GHz (1 ~ 3) GHz (3 ~ 8) GHz (8 ~ 10) GHz (10 ~ 15) GHz (15 ~ 18) GHz (6 ~ 10) dB (0.05 ~ 1) GHz (1 ~ 3) GHz (3 ~ 8) GHz (8 ~ 10) GHz (10 ~ 15) GHz (15 ~ 18) GHz	0.040 dB 0.042 dB 0.040 dB 0.043 dB 0.048 dB 0.056 dB 0.041 dB 0.043 dB 0.041 dB 0.044 dB 0.048 dB 0.052 dB 0.041 dB 0.043 dB 0.042 dB 0.044 dB 0.042 dB 0.048 dB	Standard mismatch, Power sensor, Power meter, Spectrum analyzer, Frequency standard, Frequency counter /KTICC-CI-40644
RF terminations VSWR Reflection coefficient	40645	(1 ~ 3) 5 Hz ~ 0.1 MHz 0.1 MHz ~ 2 GHz (2 ~ 26.5) GHz (26.5 ~ 40) GHz (40 ~ 50) GHz (0 ~ 0.5) 5 Hz ~ 0.1 MHz 0.1 MHz ~ 2 GHz (2 ~ 26.5) GHz (26.5 ~ 40) GHz (40 ~ 50) GHz	0.008 0.009 0.015 0.029 0.037 0.004 1 0.004 6 0.007 3 0.015 0.018	Network analyzer, Calibration kit /KTICC-CI-40645
Coaxial thermistor mounts Calibration Facor	40646	10 μ W ~ 10 mW (0.01 ~ 1) GHz (1 ~ 10) GHz (10 ~ 18) GHz	0.9×10^{-2} 1.1×10^{-2} 1.3×10^{-2}	Thermistor mount, Power meter, Network analyzer /KTICC-CI-40646
Transmission trouble testers Output Bit rate Input Bit rate	40648	(0.05 ~ 5) GHz (5 ~ 30) GHz (0.05 ~ 5) GHz (5 ~ 30) GHz	1.2×10^{-10} 1.3 Hz 0.9 Hz 1.3 Hz	Frequency counter, Transmission analyzer Frequency standard, Signal generator /KTICC-CI-40648

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
RF voltmeters Ratio of voltage and RF output voltage (F) DC Input Voltage	40650	(0.1 ~ 100) MHz (100 ~ 1 000) MHz 0 mV (0 ~ 10) mV (10 ~ 100) mV (0.1 ~ 100) V (100 ~ 400) V	2.6×10^{-2} 1.2×10^{-2} 0.61 μ V 7.8×10^{-5} 6.2×10^{-5} 6.1×10^{-5} 1.7×10^{-5}	Signal generator, TVC, Calibrator /KTICC-CI-40650
Vector voltmeters Ratio of voltage and RF output voltage (F) Phase	40651	(0.1 ~ 100) MHz (100 ~ 1 000) MHz (0 ~ 360)°	2.6×10^{-2} 1.2×10^{-2} 0.1°	Signal generator, TVC /KTICC-CI-40651
Field strength meters Frequency Input Level	40652	(0.1 ~ 5 000) MHz (5 ~ 18) GHz 9 kHz ~ 0.1 MHz (20 ~ -40) dBm (0.1 ~ 1 000) MHz (20 ~ 10) dBm (10 ~ 0) dBm (0 ~ -40) dBm (-40 ~ -60) dBm (-60 ~ -90) dBm (-90 ~ -110) dBm (-110 ~ -120) dBm (1 ~ 4) GHz (20 ~ -20) dBm (-20 ~ -40) dBm (-40 ~ -70) dBm (-70 ~ -100) dBm (-100 ~ -120) dBm (4 ~ 8) GHz (20 ~ -20) dBm (-20 ~ -40) dBm (-40 ~ -70) dBm (-70 ~ -100) dBm (-100 ~ -120) dBm (8 ~ 10) GHz (20 ~ -20) dBm (-20 ~ -50) dBm (-50 ~ -80) dBm (-80 ~ -110) dBm (-110 ~ -120) dBm (10 ~ 12) GHz (20 ~ 10) dBm (10 ~ 0) dBm (0 ~ -40) dBm (-40 ~ -60) dBm (-60 ~ -90) dBm (-90 ~ -120) dBm	1.2×10^{-10} 1.3 Hz 0.08 dB 0.21 dB 0.20 dB 0.21 dB 0.22 dB 0.23 dB 0.24 dB 0.25 dB 0.21 dB 0.22 dB 0.23 dB 0.24 dB 0.25 dB 0.22 dB 0.23 dB 0.24 dB 0.25 dB 0.26 dB 0.27 dB 0.28 dB 0.29 dB 0.30 dB 0.31 dB 0.28 dB 0.27 dB 0.28 dB 0.29 dB 0.30 dB 0.31 dB	Power sensor Power meter, Frequency counter Measuring receiver, Signal generator /KTICC-CI-40652

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Field strength meters Input Level	40652	(12 ~ 18) GHz (20 ~ -30) dBm (-30 ~ -50) dBm (-50 ~ -80) dBm (-80 ~ -100) dBm (-100 ~ -120) dBm	0.29 dB 0.30 dB 0.31 dB 0.32 dB 0.33 dB	Power sensor Power meter, Frequency counter Measuring receiver, Signal generator /KTICC-CI-40652
AM/FM test sources Output Frequency	40653	1 MHz ~ 1 GHz	1.2×10^{-10}	Frequency Counter, Measuring Receiver /KTICC-CI-40653
Residual FM		50 Hz ~ 3 kHz	5.6×10^{-2}	
Residual AM		50 Hz ~ 3 kHz	4.1×10^{-2}	
FM Distortion		(12.5 ~ 400) kHz	2.1×10^{-2}	
Dip simulators Output Frequency	40654	(50 ~ 60) Hz	6.3	Oscilloscope, DMM, High voltage probe /KTICC-CI-40654
Dip Voltage		(50 ~ 60) Hz (0 ~ 120) V 0 (0 ~ 40) % (40 ~ 80) % (80 ~ 100) % (100 ~ 120) % (120 ~ 240) V 0 (0 ~ 40) % (40 ~ 80) % (80 ~ 100) % (100 ~ 120) % (240 ~ 380) V 0 (0 ~ 40) % (40 ~ 70) % (70 ~ 80) % (80 ~ 120) %	0.01 2.1×10^{-2} 1.5×10^{-2} 1.8×10^{-2} 1.7×10^{-2} 0.01 2.1×10^{-2} 1.5×10^{-2} 1.8×10^{-2} 1.6×10^{-2} 0.01 2.2×10^{-2} 1.7×10^{-2} 1.6×10^{-2} 1.4×10^{-2}	
Dip Cycle		60 Hz 0.833 ms (0.833 ~ 1.67) ms (1.67 ~ 5.00) ms (5.00 ~ 8.33) ms (8.33 ~ 16.7) ms (16.7 ~ 83.3) ms (83.3 ~ 166.7) ms (166.7 ~ 417) ms (417 ~ 833) ms (0.833 ~ 1.67) s (1.67 ~ 5.00) s (5.00 ~ 8.33) s	4.1×10^{-3} 6.7×10^{-3} 3.8×10^{-3} 4.1×10^{-3} 6.7×10^{-3} 4.1×10^{-3} 6.7×10^{-3} 4.6×10^{-3} 4.1×10^{-3} 6.7×10^{-3} 3.8×10^{-3} 4.1×10^{-3}	

406. Radio frequency measurement

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Dip simulators	40654	50 Hz		Oscilloscope, DMM, High voltage probe /KTICC-CI-40654
Dip Cycle		1 ms	4.8×10^{-3}	
		(1 ~ 2) ms	6.0×10^{-3}	
		(2 ~ 6) ms	3.5×10^{-3}	
		(6 ~ 10) ms	3.4×10^{-3}	
		(10 ~ 20) ms	6.0×10^{-3}	
		(20 ~ 100) ms	3.4×10^{-3}	
		(100 ~ 200) ms	6.0×10^{-3}	
		(200 ~ 500) ms	3.8×10^{-3}	
		(0.5 ~ 1) s	3.4×10^{-3}	
		(1 ~ 2) s	6.0×10^{-3}	
		(2 ~ 6) s	3.5×10^{-3}	
		(6 ~ 10) s	3.4×10^{-3}	
Interval Cycle		60 Hz		
		8.33 ms	4.1×10^{-3}	
		(8.33 ~ 16.7) ms	6.7×10^{-3}	
		(16.7 ~ 83.3) ms	4.1×10^{-3}	
		(83.3 ~ 167) ms	6.7×10^{-3}	
		(167 ~ 417) ms	4.6×10^{-3}	
		(417 ~ 833) ms	4.1×10^{-3}	
		(0.833 ~ 1.67) s	6.7×10^{-3}	
		(1.67 ~ 5) s	3.8×10^{-3}	
		(5 ~ 8.33) s	4.1×10^{-3}	
		50 Hz		
		10 ms	3.4×10^{-3}	
		(10 ~ 20) ms	6.0×10^{-3}	
		(20 ~ 100) ms	3.4×10^{-3}	
		(100 ~ 200) ms	6.0×10^{-3}	
		(200 ~ 500) ms	3.8×10^{-3}	
		(0.5 ~ 1) s	3.4×10^{-3}	
		(1 ~ 2) s	6.0×10^{-3}	
		(2 ~ 6) s	3.5×10^{-3}	
		(6 ~ 10) s	3.4×10^{-3}	
Phase Time		(50 ~ 60) Hz		
		1 ms	3.8 μ s	
		(1 ~ 10) ms	3.8×10^{-3}	
		(10 ~ 20) ms	4.2×10^{-3}	
		(20 ~ 100) ms	3.8×10^{-3}	

407. Field strength &

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Loop antennas Antenna factor	40704	(160 ~ -10) dB(1/m) 20 Hz ~ 30 MHz	1.5 dB	Standard antenna Signal generator Spectrum analyzer, DMM /KTICC-CI-40704
Monopole antennas Antenna factor	40705	(120 ~ -30) dB(1/m) 1 kHz ~ 30 MHz	1.4 dB	Network Analyzer /KTICC-CI-40705

501. Contact thermometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Temperature generators; ovens, furnaces, isothermal liquid baths, ice-point baths, dry-block calibrators	50101	(-196 ~ -80) °C (-80 ~ 0) °C 0 °C (0 ~ 500) °C (500 ~ 1 100) °C	0.08 °C 0.02 0.017 0.02 2.4	Digital thermometer, SPRT, Noble metal thermocouples /KTICC-CI-50101
Temperature indicators /recorders/controllers, temperature calibrators With sensor Without sensor Temperature calibrators	50102	(-196 ~ 0) °C (0 ~ 200) °C (200 ~ 500) °C (500 ~ 1 100) °C (-196 ~ 500) °C (500 ~ 1 100) °C (-196 ~ 400) °C (400 ~ 1 100) °C	0.071 0.036 0.032 2.4 0.041 0.061 0.086 0.25	Temperature simulation, Calibrators, SPRT Noble metal thermocouples /KTICC-CI-50102
Glass thermometers; liquid-in-glass, Beckmann Glass thermometers	50103	(-50 ~ 50) °C (50 ~ 400) °C (400 ~ 500) °C	0.07 0.04 0.16	SPRT/KTICC-CI-50103
Resistance thermometers; SPRT, IPRT, thermistors, etc. SPRT, IPRT	50104	(-196 ~ 0) °C (0 ~ 500) °C	0.086 °C 0.069 °C	SPRT/KTICC-CI-50104

501. Contact thermometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Thermal expansion thermometers; bimetal, gas or liquid type	50105	(-50 ~ 100) °C (100 ~ 200) °C (200 ~ 500) °C	0.3 0.6 1.5	SPRT/KTICC-CI-50105
Thermocouples; noble metal, base metal pure metal, special type, etc. Base metal	50106	(-196 ~ 0) °C (0 ~ 200) °C (200 ~ 500) °C (500 ~ 1 100) °C	0.14 0.06 0.1 1.5	SPRT, Noble metal thermocouples /KTICC-CI-50106
Temperature transducers	50107	(-196 ~ -80) °C (-80 ~ 0) °C (0 ~ 200) °C (200 ~ 500) °C (500 ~ 1100) °C	0.11 °C 0.09 0.113 0.185 1.243	SPRT, Noble metal thermocouples /KTICC-CI-50107
Others: quartz, semiconductivity, optical fiber, etc. Thermistor thermometer	50109	(-80 ~ 500) °C	0.15	SPRT/KTICC-CI-50109

502. Non contact thermometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Standard radiation thermometers Infrared	50204	(0 ~ 300) °C (300 ~ 1000) °C	2.4 3.5	Infrared thermometer /KTICC-CI-50204
Thermal image apparatus	50205	(0 ~ 300) °C (300 ~ 1 000) °C	2.4 3.5	Infrared thermometer /KTICC-CI-50205
Blackbody furnaces	50206	(0 ~ 300) °C (300 ~ 1 000) °C	1.2 1.7	Infrared thermometer /KTICC-CI-50206
Others; ear thermometers, etc. Ear thermometer	50207	(20 ~ 40) °C	1.3	Black Body/KTICC-CI-50207

503. Humidity

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Relative humidity hygrometers; polimer thin film, hair, etc. Humidity Temperature	50302	(5 ~ 15) % R.H. (15 ~ 30) % R.H. (30 ~ 50) % R.H. (50 ~ 70) % R.H. (70 ~ 95) % R.H. (-40 ~ 80) °C	2.7 1.1 1.8 1.9 2 0.9	Dew point hygrometer /KTICC-CI-50302
Temperature humidity recorders; Hygrothermograph, etc. Humidity Temperature	50304	(5 ~ 15) % R.H. (15 ~ 30) % R.H. (30 ~ 50) % R.H. (50 ~ 70) % R.H. (70 ~ 95) % R.H. (-40 ~ 80) °C	2.9 2.1 2.2 2.3 2.4 1.1	Dew point hygrometer /KTICC-CI-50304
Transducers; dew-point/relative humidity Humidity	50305	(5 ~ 25) % R.H. (25 ~ 50) % R.H. (50 ~ 70) % R.H. (70 ~ 90) % R.H. (90 ~ 95) % R.H.	3 2.4 2.5 2.6 2.7	Dew point hygrometer /KTICC-CI-50305
Humidity generators; two-pressure, two-temperature, flow mixing humidity generator, constant temperature and humidity chamber, etc. Humidity Temperature	50306	(5 ~ 20) % R.H. (20 ~ 40) % R.H. (40 ~ 60) % R.H. (60 ~ 80) % R.H. (80 ~ 95) % R.H. (-75 ~ 100) °C (100 ~ 150) °C (150 ~ 180) °C	1.5 2.2 2.8 3.7 4.4 0.8 1.1 1.4	Dew point hygrometer /KTICC-CI-50306

601. Sound in air

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Sound level meters	60106	31.5 63 Hz 125 Hz 250 Hz 500 Hz 1 2 4 8 12.5 kHz	0.4 0.3 0.3 0.2 0.2 0.2 0.2 0.2 0.4 0.6	Sound calibrator /KTICC-CI-60106

603. Vibration

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Vibration calibrators Acceleration	60301	(10 ~ 160) Hz (160 ~ 315) Hz (315 ~ 1 250) Hz	1.6×10^{-2} 1.8×10^{-2} 2.2×10^{-2}	Accelerometer /KTICC-CI-60301
Vibration transducers Sensitivity	60302	(10 ~ 20) Hz (20 ~ 160) Hz (160 ~ 315) Hz (315 ~ 5 000) Hz	1.7×10^{-2} 1.6×10^{-2} 1.8×10^{-2} 2.2×10^{-2}	Accelerometer /KTICC-CI-60302
Vibration measuring instruments Acceleration Velocity Displacement	60303	(10 ~ 20) Hz (20 ~ 160) Hz (160 ~ 315) Hz (315 ~ 5 000) Hz (10 ~ 20) Hz (20 ~ 160) Hz (160 ~ 315) Hz (315 ~ 630) Hz (630 ~ 2 500) Hz (10 ~ 20) Hz (20 ~ 80) Hz (80 ~ 100) Hz (100 ~ 160) Hz (160 ~ 315) Hz (315 ~ 630) Hz (630 ~ 1 000) Hz (1 000 ~ 1 250) Hz	1.7×10^{-2} 1.6×10^{-2} 1.8×10^{-2} 2.2×10^{-2} 1.7×10^{-2} 1.6×10^{-2} 1.9×10^{-2} 2.3×10^{-2} 2.2×10^{-2} 1.6×10^{-2} 1.5×10^{-2} 1.6×10^{-2} 1.7×10^{-2} 1.8×10^{-2} 2.3×10^{-2} 3.2×10^{-2} 4.3×10^{-2}	Accelerometer /KTICC-CI-60303

701. Photometry

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Illuminance meters	70101	(0.5 ~ 10) lx (10 ~ 1 500) lx (1 500 ~ 20 000) lx	2.6×10^{-2} 2.2×10^{-2} 2.9×10^{-2}	Illuminance meters /KTICC-CI-70101

704. Fiber optics

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Broadband light sources Wavelength Optical power	70402	(600 ~ 1 700) nm 1 310 nm, 1 550 nm (-50 ~ 0) dBm	0.082 nm 0.05	Optical spectrum analyzers Optical multimeters /KTICC-CI-70402
Optical attenuators Optical attenuation	70410	1 310 nm, 1 550 nm (0 ~ -60) dB	0.06	Optical power stabilized Optical multimeters /KTICC-CI-70410
Optical couplers Division Ratio Fiber-optic power meters Absolute Optical Power Linearity	70411 70412	1 310 nm 1 550 nm 1 310 nm, 1 550 nm (0 ~ -60) dBm 1 310 nm, 1 550 nm (0 ~ -60) dB	1.1×10^{-2} 1.1×10^{-2} 0.06 0.06	Optical power stabilized Optical multimeters /KTICC-CI-70411 Optical power stabilized Optical multimeters Optical attenuators /KTICC-CI-70412
Optical loss Testers Wavelength Output Optical Power Input Optical Power Linearity	70413	(600 ~ 1 700) nm 1 310 nm, 1 550 nm (-50 ~ 0) dBm 1 310 nm, 1 550 nm (0 ~ -60) dBm 1 310 nm, 1 550 nm (0 ~ -60) dB	2.2×10^{-7} 0.05 0.06 0.06	Optical power stabilized Optical multimeters Optical attenuators Optical spectrum analyzers Multi-laser wavelength meters /KTICC-CI-70413
Optical multimeters Absolute Optical Power Linearity	70415	1 310 nm, 1 550 nm (0 ~ -60) dBm 1 310 nm, 1 550 nm (0 ~ -60) dB	0.06 0.06	Optical power stabilized Optical multimeters Optical attenuators /KTICC-CI-70415

704. Fiber optics

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Optical network analyzers	70416			Optical multimeters Optical spectrum analyzers Optical attenuators OTDR Wavelength meters, Multi laser Return loss test sets /KTICC-CI-70416
Optical multimeters				
Input Optical Power		1 310 nm, 1 550 nm (0 ~ -60) dBm	0.06	
Optical spectrum analyzers				
Wavelength		(1310 ~ 1575) nm	0.058 nm	
Resolution		1310 nm, 1550 nm (0.1 ~ 1) nm	0.058 nm	
Input Optical Power		1 310 nm, 1 550 nm (0 ~ -60) dBm	0.06	
Optical attenuators				
Attenuation		1 310 nm, 1 550 nm (0 ~ -60) dB	0.06	
OTDR				
Length		CRM(1 310 nm, 1550 nm) (2.4, 2.7) km 13 km	0.085 m 0.38 m	
Loss		CRM(1 310 nm, 1550 nm) (0 ~ 3) dB (3 ~ 10) dB	0.11 dB 0.19 dB	
Wavelength Meters, Multi laser				
Wavelength		1 310 nm 1 460 nm 1 500 nm 1 531.591 0 nm 1 531.587 8 nm 1 531.584 8 nm 1 550 nm 1 580 nm	2.4 pm 1.7 pm 1.7 pm 0.35 pm 0.35 pm 0.35 pm 2.4 pm 1.7 pm	
Input Optical Power		1 310 nm, 1 550 nm (0 ~ -60) dBm	0.06	
Return loss test				
Return Loss		1 310 nm, 1 550 nm (20 ~ 30) dB (30 ~ 40) dB	0.5 dB 0.6 dB	
Ethernet Tester				
Interface		(1 ~ 100) MHz	1.3×10^{-10}	
Wavelength		(600 ~ 1640) nm	0.082 nm	
Output Optical Power		1310 nm, 1550 nm (0 ~ -60) dBm	0.05 dB	
Sensitivity		1310 nm, 1550 nm	0.06 dB	

704. Fiber optics

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Optical spectrum analyzers Wavelength Resolution Input Optical Power Linearity	70417	(600 ~ 1 640) nm 1 310 nm, 1 550 nm (0.1 ~ 1) nm 1 310 nm, 1 550 nm (0 ~ -60) dBm 1 310 nm, 1 550 nm (0 ~ -60) dB	0.058 nm 0.058 nm 0.06 0.06	Optical power stabilized lasers and LDs Optical multimeters Optical attenuators /KTICC-CI-70417
Optical time domain reflectometers: OTDR Wavelength Length Loss	70418	1 310 nm, 1 550 nm CRM(1 310 nm, 1550 nm) (2.4, 2.7) km 13 km CRM(1 310 nm, 1550 nm) (0 ~ 3) dB (3 ~ 10) dB	0.082 nm 0.085 m 0.38 m 0.11 dB 0.19 dB	Optical length standard Optical fiber Optical spectrum analyzers Loss standard optical fiber /KTICC-CI-70418
PDH/SDH analyzers Bit rate Output Jitter Smsr Optical Power Extinction Ratio Sensitivity Reflectance	70419	DS1 (1.544 MHz) ~ STM-64 (9.953 28 GHz) DS1 (10 Hz to 40 kHz) DS1 (8 kHz to 40 kHz) E1 (20 Hz to 100 kHz) E1 (18 kHz to 100 kHz) DS3 (10 Hz to 400 kHz) DS3 (30 kHz to 400 kHz) STM-1 (12 kHz to 1.3 MHz) STM-4 (12 kHz to 5 MHz) STM-16 (13 kHz to 20 MHz) STM-1 (155.52 MHz) ~ STM-64 (9.953 28 GHz) STM-1 (155.52 MHz) ~ STM-64 (9.953 28 GHz) STM-1 (155.52 MHz) ~ STM-64 (9.953 28 GHz) STM-1 (155.52 MHz) ~ STM-64 (9.953 28 GHz) STM-1 (155.52 MHz) ~ STM-16 (2.488 32 GHz)	1.3×10^{-10} 15 ns 15 ns 11 ns 11 ns 0.78 ns 0.78 ns 0.37 ns 0.13 ns 0.032 ns 0.08 dB 0.05 dB 0.2 dB 0.05 dB 0.8	Optical multimeters Optical spectrum analyzers Optical attenuators PDH/SDH analyzers General frequency sources Frequency meters/counters /KTICC-CI-70419

704. Fiber optics

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
PDH/SDH analyzers Jitter Generator & Analyzers	70419	DS1 (1.544 MHz), 1 kHz 0.77 1.8 4.8 8.8 E1 (2.048 MHz), 2.4 kHz 0.77 1.8 4.8 8.8 DS3 (44.736 MHz), 4 kHz 0.77 1.8 4.8 8.8 STM-1 (155.52 MHz) 65 kHz / 0.20 UI 650 kHz / 0.20 UI 6.5 kHz / 0.91 UI 4 kHz / 3.52 UI 2.5 kHz / 7.52 UI STM-4 (622.08 MHz) 65 kHz / 0.20 UI 650 kHz / 0.20 UI 6.5 kHz / 0.91 UI 4 kHz / 3.52 UI 2.5 kHz / 7.52 UI STM-16 (2.488 32 GHz) 65 kHz / 0.20 UI 650 kHz / 0.20 UI 6.5 kHz / 0.91 UI 4 kHz / 3.52 UI 2.5 kHz / 7.52 UI	56 ns 0.17 μs 0.33 μs 0.54 μs 42 ns 0.13 μs 0.25 μs 0.40 μs 2.6 ns 6.7 ns 14 ns 23 ns 0.48 ns 0.49 ns 0.85 ns 2.6 ns 4.7 ns 0.16 ns 0.17 ns 0.25 ns 0.64 ns 1.7 ns 0.042 ns 0.051 ns 0.062 ns 0.18 ns 0.38 ns	Optical multimeters Optical spectrum analyzers Optical attenuators PDH/SDH analyzers General frequency sources Frequency meters/counters /KTICC-CI-70419
Return loss meters Return Loss	70423	1 310 nm, 1 550 nm (20 ~ 30) dB (30 ~ 40) dB	0.5 dB 0.6 dB	Return loss generator /KTICC-CI-70423

704. Fiber optics

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
SDH/SONET analyzers	70424	STM-1 (155.52 MHz) ~ STM-64 (9.953 28 GHz)	1.3×10^{-10}	Optical multimeters Optical spectrum analyzers Optical attenuators PDH/SDH analyzers General frequency sources Frequency meters/counters /KTICC-CI-70424
Bit Rate				
Output Jitter		STM-1 (12 kHz to 1.3 MHz) STM-4 (12 kHz to 5 MHz) STM-16 (13 kHz to 20 MHz)	0.37 ns 0.13 ns 0.032 ns	
Smsr		STM-1 (155.52 MHz) ~ STM-64 (9.953 28 GHz)	0.08 dB	
Optical Power		STM-1 (155.52 MHz) ~ STM-64 (9.953 28 GHz)	0.05 dB	
Extinction Ratio		STM-1 (155.52 MHz) ~ STM-64 (9.953 28 GHz)	0.2 dB	
Sensitivity		STM-1 (155.52 MHz) ~ STM-16 (2.488 32 GHz)	0.05 dB	
Reflectance		STM-1 (155.52 MHz) ~ STM-16 (2.488 32 GHz)	0.8	
Jitter Generator & Analyzers		STM-1 (155.52 MHz) 65 kHz / 0.20 UI 650 kHz / 0.20 UI 6.5 kHz / 0.91 UI 4 kHz / 3.52 UI 2.5 kHz / 7.52 UI STM-4 (622.08 MHz) 65 kHz / 0.20 UI 650 kHz / 0.20 UI 6.5 kHz / 0.91 UI 4 kHz / 3.52 UI 2.5 kHz / 7.52 UI STM-16 (2.488 32 GHz) 65 kHz / 0.20 UI 650 kHz / 0.20 UI 6.5 kHz / 0.91 UI 4 kHz / 3.52 UI 2.5 kHz / 7.52 UI	0.48 ns 0.49 ns 0.85 ns 2.6 ns 4.7 ns 0.16 ns 0.17 ns 0.25 ns 0.64 ns 1.7 ns 0.042 ns 0.051 ns 0.062 ns 0.18 ns 0.38 ns	

704. Fiber optics

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Multi-laser wavelength Wavelength Input Optical Power	70426	1 310 nm 1 460 nm 1 500 nm 1 531.591 0 nm 1 531.587 8 nm 1 531.584 8 nm 1 550 nm 1 580 nm 1 310 nm, 1 550 nm (0 ~ -60) dBm	2.4 pm 1.7 pm 1.7 pm 0.35 pm 0.35 pm 0.35 pm 2.4 pm 1.7 pm 0.06	Frequency stabilized lasers and LDs Optical multimeters /KTICC-CI-70426
Wavelength sweep multichannel measuring systems Input Optical Power	70427	1 310 nm, 1 550 nm (0 ~ -60) dBm	0.06	Optical power stabilized Optical multimeters Optical attenuators /KTICC-CI-70427
Frequency stabilized lasers and LDs Wavelength Output Optical Power Linearity	70429	(1 260 ~ 1 640) nm 1 310 nm, 1 550 nm (-50 ~ 0) dBm 1 310 nm, 1 550 nm (-50 ~ 0) dB	1.2 pm 0.05 dB 0.05	Optical spectrum analyzers Optical multimeters /KTICC-CI-70429
ASE light sources Wavelength Output Optical Power	70430	(600 ~ 1 700) nm 1 310 nm, 1 550 nm (-50 ~ 0) dBm	0.082 nm 0.05	Optical spectrum analyzers Optical multimeters /KTICC-CI-70430
CW-laser wavelength meters Wavelength Input Optical Power	70431	632.991 0 nm 1 310 nm 1 460 nm 1 500 nm 1 531.591 0 nm 1 531.587 8 nm 1 531.584 8 nm 1 550 nm 1 580 nm 1 640 nm 1 310 nm, 1 550 nm (0 ~ -60) dBm	0.06 pm 2.4 pm 1.7 pm 1.7 pm 0.35 pm 0.35 pm 0.35 pm 2.4 pm 1.7 pm 1.7 pm 0.06	Frequency stabilized lasers and LDs Optical attenuators Optical multimeters /KTICC-CI-70431

704. Fiber optics

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Optical power stabilized lasers and LDs	70433			Multi-laser wavelength meters Optical multimeters Optical spectrum analyzers /KTICC-CI-70433
Wavelength		(1 300 ~ 1 640) nm	2.2×10^{-7}	
Output Optical Power		1 310 nm, 1 550 nm (0 ~ -50) dBm	0.05	
Light sources, LED				
Wavelength		(600 ~ 1 640) nm	0.082 nm	
Output Optical Power		1 310 nm, 1 550 nm (-50 ~ 0) dBm	0.05	

901. Chemical analysis

Measured Quantity Instrument or Gauge	Field Code	Range	Measurement uncertainty (The Confidence Level is about 95 %)	Standard/Method of Measurement etc.
Gas analyzers	90103			Standard gas /KTICC-CI-90103
O ₂		(0 ~ 18) cmol/mol	0.37 cmol/mol	
CO		(0 ~ 100) μmol/mol	2.1 μmol/mol	
CH ₄		(0 ~ 2) cmol/mol	0.03 cmol/mol	
H ₂ S		(0 ~ 30) μmol/mol	0.89 μmol/mol	